

Report of Calibration
PSC 4CH-MSS-AR Slow XY Corr_S/N 0005
2026-01-23 19:16:58

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001102	1.004205	-1.011459	-1.010749	1.003621	-0.010089
-27.002111	27.001341	-26.818108	-26.823114	27.040604	0.028557

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003253	-0.017841	-0.016911	0.000000
Initial measured gains:	0.999851	0.992525	0.992745	1.001533
Gain corrections:	0.999851	1.007531	1.007308	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.002623
27.000000	27.038363

Measured offset: 0.001248
Measured gain: 1.001375
Gain correction: 0.998627

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001247	1.001026	-1.001391	-1.001376	1.001790	-0.012965
-27.002132	27.002275	-27.002352	-27.002274	27.003315	0.057525

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000235	-0.000141	-0.000128	0.000753
Final measured gains:	1.000014	1.000003	1.000000	1.000011

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000034	-0.000179	-0.000135	0.000820
Final measured offsets stdev:	0.000173	0.000152	0.000140	0.000259
Final measured gains mean:	0.999986	0.999986	0.999988	1.000014
Final measured gains stdev:	0.000018	0.000011	0.000011	0.000009

Saving channel 1 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000926	1.001485	-1.009905	-1.009832	1.000512	0.001254
-27.002016	26.994679	-26.814369	-26.829247	27.035002	0.052790

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.000863	-0.016548	-0.015900	0.000000
Initial measured gains:	0.999696	0.992438	0.993013	1.001589
Gain corrections:	0.999696	1.007620	1.007036	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999602
27.000000	27.032600

Measured offset: -0.001667

Measured gain: 1.001269

Gain correction: 0.998732

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001082	1.001059	-1.001162	-1.001147	1.000452	0.006794
-27.002187	27.002099	-27.001987	-27.002192	27.001900	0.029441

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000020	-0.000091	-0.000067	-0.000622
Final measured gains:	0.999998	0.999989	0.999998	1.000016

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000035	-0.000171	-0.000244	-0.000559
Final measured offsets stdev:	0.000052	0.000192	0.000093	0.000106
Final measured gains mean:	0.999983	0.999985	0.999982	1.000008
Final measured gains stdev:	0.000013	0.000003	0.000009	0.000010

Saving channel 2 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000945	1.002586	-1.010528	-1.009528	1.004328	0.029339
-27.002225	27.013758	-26.839409	-26.824390	27.057907	0.039377

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001260	-0.016220	-0.015759	0.000000
Initial measured gains:	1.000380	0.993370	0.992830	1.001630
Gain corrections:	1.000380	1.006675	1.007221	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.003391
27.000000	27.055685

Measured offset: 0.001379

Measured gain: 1.002011

Gain correction: 0.997993

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000921	1.000928	-1.001370	-1.001439	1.001631	0.011355
-27.002346	27.001687	-27.002079	-27.002094	27.002460	0.027596

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000033	-0.000477	-0.000548	0.000700
Final measured gains:	0.999974	0.999972	0.999970	1.000003

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000003	-0.000198	-0.000341	0.000598
Final measured offsets stdev:	0.000087	0.000188	0.000156	0.000077
Final measured gains mean:	0.999986	0.999984	0.999983	1.000001
Final measured gains stdev:	0.000010	0.000008	0.000009	0.000010

Saving channel 3 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000710	1.002464	-1.010051	-1.009499	0.999421	-0.025940
-27.002007	27.005509	-26.821964	-26.814371	27.023972	-0.003174

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001687	-0.016631	-0.016349	0.000000
Initial measured gains:	1.000067	0.992716	0.992446	1.000827
Gain corrections:	1.000067	1.007337	1.007612	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.998906
27.000000	27.021811

Measured offset: -0.001975

Measured gain: 1.000881

Gain correction: 0.999120

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000721	1.000616	-1.000881	-1.000962	1.001220	-0.019222
-27.001931	27.002355	-27.002327	-27.002205	27.002354	0.057055

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000126	-0.000151	-0.000239	0.000627
Final measured gains:	1.000020	1.000009	1.000001	0.999977

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000073	-0.000089	-0.000067	-0.000251
Final measured offsets stdev:	0.000114	0.000100	0.000156	0.000508
Final measured gains mean:	1.000003	0.999999	1.000000	1.000004
Final measured gains stdev:	0.000019	0.000012	0.000008	0.000015

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date _____