

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_S/N 0021
2026-01-21 12:08:54

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.003776	1.012578	-1.132040	-1.129996	1.007354	-0.015231
-224.040386	224.186936	-222.674225	-222.592194	224.824783	0.006372

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.008182	-0.134989	-0.133305	0.000000
Initial measured gains:	1.000618	0.993300	0.992941	1.002881
Gain corrections:	1.000618	1.006746	1.007109	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.006029
225.000000	225.787170

Measured offset: 0.002542
Measured gain: 1.003487
Gain correction: 0.996525

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.003609	1.002625	-1.005413	-1.005652	1.007308	-0.006482
-224.041298	224.039919	-224.041245	-224.041824	224.047501	0.017996

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000983	-0.001812	-0.002050	0.004671
Final measured gains:	0.999998	0.999992	0.999993	1.000013

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000841	-0.002255	-0.000887	0.004862
Final measured offsets stdev:	0.000209	0.000313	0.000650	0.000513
Final measured gains mean:	0.999995	0.999989	0.999992	1.000003
Final measured gains stdev:	0.000003	0.000004	0.000006	0.000011

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002109	1.032556	-1.144158	-1.145620	0.993283	0.004658
-224.046920	224.066818	-222.483109	-222.437805	224.664154	0.003143

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.030494	-0.149713	-0.151385	0.000000
Initial measured gains:	0.999953	0.992352	0.992142	1.002854
Gain corrections:	0.999953	1.007707	1.007920	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.990222
225.000000	225.621979

Measured offset: -0.012598
Measured gain: 1.002820
Gain correction: 0.997188

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002251	1.004472	-1.005275	-1.005036	1.001506	-0.004565
-224.047715	224.047787	-224.049591	-224.048035	224.046539	-0.005625

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.002231	-0.003030	-0.002796	-0.002974
Final measured gains:	0.999990	0.999995	0.999989	1.000008

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000816	-0.002728	-0.002543	-0.001711
Final measured offsets stdev:	0.000943	0.000204	0.000160	0.001134
Final measured gains mean:	0.999994	0.999991	0.999993	1.000004
Final measured gains stdev:	0.000004	0.000004	0.000004	0.000009

Saving channel 2 calibration constants to qspi

lab{2}Chan3
Burden resistor = 6.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001162	1.026621	-1.163368	-1.190018	1.010369	-0.010765
-300.082666	299.977467	-297.958618	-298.118164	300.831055	0.054217

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.025897	-0.169860	-0.196065	0.000000
Initial measured gains:	0.999563	0.992356	0.992800	1.002910
Gain corrections:	0.999563	1.007703	1.007252	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.010430
300.000000	300.749512

Measured offset: 0.007958
Measured gain: 1.002472
Gain correction: 0.997534

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001225	1.001067	-1.005618	-1.001659	1.010735	-0.011275
-300.082414	300.080492	-300.084167	-300.082703	300.098694	0.063943

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000153	-0.004401	-0.000434	0.009640
Final measured gains:	0.999994	0.999991	1.000000	1.000029

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000020	-0.004072	-0.001544	0.010022
Final measured offsets stdev:	0.000278	0.000250	0.001283	0.000914
Final measured gains mean:	0.999993	0.999989	0.999993	1.000020
Final measured gains stdev:	0.000002	0.000003	0.000006	0.000014

Saving channel 3 calibration constants to qspi

lab{2}Chan4
Burden resistor = 6.7900

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000481	1.023936	-1.159269	-1.138707	0.977269	0.017455
-266.072707	266.120535	-264.241974	-264.336151	266.766266	0.038297

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.023364	-0.166297	-0.145303	0.000000
Initial measured gains:	1.000092	0.992494	0.992927	1.002612
Gain corrections:	1.000092	1.007562	1.007123	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.977000
265.000000	265.690857

Measured offset: -0.025704
Measured gain: 1.002704
Gain correction: 0.997303

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000434	0.999221	-1.002123	-1.003684	0.994870	0.017395
-266.072976	266.071904	-266.072235	-266.073944	266.065613	0.056485

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.001214	-0.001698	-0.003258	-0.004344
Final measured gains:	1.000001	0.999991	0.999991	0.999993

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.001332	-0.001942	-0.003305	-0.004998
Final measured offsets stdev:	0.000231	0.000395	0.000310	0.000463
Final measured gains mean:	1.000001	0.999990	0.999986	1.000007
Final measured gains stdev:	0.000004	0.000005	0.000004	0.000015

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date_____