

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_S/N 0021
2026-01-23 13:00:41

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.003599	1.013054	-1.132611	-1.130658	1.008953	0.010151
-224.038596	224.205007	-222.692535	-222.607239	224.820160	-0.000431

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.008749	-0.135650	-0.134072	0.000000
Initial measured gains:	1.000704	0.993386	0.993013	1.002775
Gain corrections:	1.000704	1.006658	1.007036	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.005253
225.000000	225.785721

Measured offset: 0.001768
Measured gain: 1.003484
Gain correction: 0.996528

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.003712	1.003255	-1.004944	-1.004967	1.006806	-0.021699
-224.039100	224.038019	-224.038559	-224.039291	224.043610	0.011050

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000454	-0.001240	-0.001260	0.003542
Final measured gains:	0.999997	0.999992	0.999995	1.000009

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000231	-0.001243	-0.001414	0.004741
Final measured offsets stdev:	0.000232	0.000141	0.000677	0.000756
Final measured gains mean:	0.999988	0.999990	0.999992	0.999999
Final measured gains stdev:	0.000007	0.000005	0.000004	0.000012

Saving channel 1 calibration constants to qspi

Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002697	1.028986	-1.140754	-1.146990	0.988441	-0.003436
-224.043846	224.067808	-222.484665	-222.446365	224.646469	0.019469

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.026299	-0.145687	-0.152123	0.000000
Initial measured gains:	0.999990	0.992390	0.992191	1.002776
Gain corrections:	0.999990	1.007668	1.007871	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.984786
225.000000	225.605820

Measured offset: -0.017987

Measured gain: 1.002772

Gain correction: 0.997235

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002265	1.004187	-1.004853	-1.005584	1.003882	0.003686
-224.043715	224.043839	-224.044418	-224.046204	224.041321	-0.002600

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.001931	-0.002597	-0.003323	-0.000295
Final measured gains:	0.999992	0.999992	0.999996	0.999990

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000673	-0.001957	-0.003069	0.000726
Final measured offsets stdev:	0.000875	0.000554	0.000697	0.000705
Final measured gains mean:	0.999996	0.999993	0.999989	1.000000
Final measured gains stdev:	0.000003	0.000003	0.000004	0.000007

Saving channel 2 calibration constants to qspi

Burden resistor = 6.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001055	1.027426	-1.164099	-1.190856	1.012319	0.010917
-300.081670	299.993992	-297.974457	-298.138245	300.828735	0.064584

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.026753	-0.170643	-0.196941	0.000000
Initial measured gains:	0.999619	0.992409	0.992867	1.002843
Gain corrections:	0.999619	1.007649	1.007184	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.009516
300.000000	300.743408

Measured offset: 0.007062

Measured gain: 1.002454

Gain correction: 0.997552

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001619	1.002000	-1.001461	-1.002542	1.014024	0.009974
-300.081081	300.080684	-300.081482	-300.081787	300.090973	0.052360

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000384	0.000160	-0.000924	0.012030
Final measured gains:	0.999997	1.000002	0.999999	0.999994

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000760	-0.000406	-0.001337	0.009988
Final measured offsets stdev:	0.000249	0.001089	0.000241	0.001246
Final measured gains mean:	0.999993	0.999997	0.999992	1.000012
Final measured gains stdev:	0.000003	0.000003	0.000005	0.000011

Saving channel 3 calibration constants to qspi

Burden resistor = 6.7900

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000991	1.024911	-1.160291	-1.139501	0.979099	0.037087
-266.067564	266.125374	-264.249420	-264.342468	266.755432	0.026696

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.023792	-0.166768	-0.145548	0.000000
Initial measured gains:	1.000128	0.992540	0.992969	1.002549
Gain corrections:	1.000128	1.007516	1.007080	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.976300
265.000000	265.686676

Measured offset: -0.026391

Measured gain: 1.002691

Gain correction: 0.997316

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000878	0.999588	-1.002971	-1.004222	0.993470	0.008119
-266.068106	266.067480	-266.067474	-266.068756	266.066589	0.086798

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.001292	-0.002103	-0.003355	-0.006137
Final measured gains:	1.000003	0.999990	0.999990	1.000020

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000980	-0.002272	-0.003699	-0.004194
Final measured offsets stdev:	0.000203	0.000121	0.000179	0.000983
Final measured gains mean:	0.999996	0.999985	0.999985	1.000018
Final measured gains stdev:	0.000004	0.000004	0.000006	0.000008

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date _____