

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_S/N 9999
2026-01-24 15:08:09

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{1}Chan1
Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.004340	1.024986	-1.142125	-1.135466	1.004659	-0.003446
-224.043543	224.125942	-222.563843	-222.446228	224.280319	0.015853

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.020368	-0.145069	-0.138910	0.000000
Initial measured gains:	1.000277	0.992748	0.992251	1.000783
Gain corrections:	1.000277	1.007305	1.007810	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.000913
225.000000	225.236404

Measured offset: -0.000139
Measured gain: 1.001051
Gain correction: 0.998950

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.004077	1.003932	-1.004488	-1.007080	1.004887	0.002394
-224.044405	224.044446	-224.043793	-224.046616	224.050644	0.013093

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000146	-0.000415	-0.003006	0.000931
Final measured gains:	1.000001	0.999995	0.999996	1.000024

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000588	-0.000895	-0.002250	0.001178
Final measured offsets stdev:	0.000905	0.000333	0.000770	0.000427
Final measured gains mean:	0.999995	0.999994	0.999991	1.000013
Final measured gains stdev:	0.000006	0.000002	0.000005	0.000008

Saving channel 1 calibration constants to qspi

Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002754	1.024427	-1.131709	-1.151456	0.985197	0.008585
-224.043145	224.071313	-222.608063	-222.460251	224.548889	-0.004115

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.021643	-0.135986	-0.156487	0.000000
Initial measured gains:	1.000029	0.992988	0.992236	1.002317
Gain corrections:	1.000029	1.007062	1.007824	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.982684
225.000000	225.507080

Measured offset: -0.019657

Measured gain: 1.002341

Gain correction: 0.997664

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002634	1.001553	-1.004316	-1.004682	1.000227	0.003302
-224.044163	224.043786	-224.044083	-224.043518	224.049957	0.025099

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.001084	-0.001690	-0.002060	-0.001360
Final measured gains:	1.000003	0.999992	0.999988	1.000034

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000985	-0.001538	-0.002770	-0.001932
Final measured offsets stdev:	0.000163	0.000145	0.000851	0.000623
Final measured gains mean:	1.000001	0.999994	0.999989	1.000028
Final measured gains stdev:	0.000003	0.000001	0.000002	0.000009

Saving channel 2 calibration constants to qspi

Burden resistor = 6.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001637	1.012862	-1.206465	-1.191967	1.013902	-0.001970
-300.078780	299.986926	-297.905640	-297.882446	300.149078	0.055999

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.011571	-0.212792	-0.198324	0.000000
Initial measured gains:	0.999655	0.992049	0.992020	1.000539
Gain corrections:	0.999655	1.008015	1.008044	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.010126
300.000000	300.070984

Measured offset: 0.009922

Measured gain: 1.000204

Gain correction: 0.999797

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001761	1.002611	-1.005146	-1.006745	1.013232	0.000223
-300.078503	300.077166	-300.081116	-300.080536	300.091156	0.078927

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000858	-0.003388	-0.004994	0.010610
Final measured gains:	0.999993	0.999997	0.999990	1.000011

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000774	-0.003454	-0.003040	0.007738
Final measured offsets stdev:	0.000529	0.000166	0.001336	0.004153
Final measured gains mean:	0.999992	0.999991	0.999985	1.000015
Final measured gains stdev:	0.000006	0.000005	0.000006	0.000007

Saving channel 3 calibration constants to qspi

Burden resistor = 6.7900

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002715	1.032330	-1.171824	-1.172765	0.998530	0.016395
-266.068905	266.140453	-264.245422	-264.127502	266.943085	0.032224

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.029456	-0.176646	-0.178038	0.000000
Initial measured gains:	1.000158	0.992483	0.992034	1.003155
Gain corrections:	1.000158	1.007574	1.008030	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.995946
265.000000	265.868164

Measured offset: -0.007357

Measured gain: 1.003304

Gain correction: 0.996707

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002751	1.000667	-1.005231	-1.003886	0.999068	0.022484
-266.068455	266.066809	-266.068054	-266.067291	266.068848	0.029425

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.002085	-0.002492	-0.001144	-0.001612
Final measured gains:	1.000002	0.999989	0.999991	1.000014

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.001699	-0.002840	-0.001626	-0.002018
Final measured offsets stdev:	0.000301	0.000247	0.000398	0.000468
Final measured gains mean:	0.999998	0.999988	0.999988	1.000002
Final measured gains stdev:	0.000004	0.000004	0.000005	0.000009

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date _____