

Report of Calibration
PSC 4CH-MSS-BTA-Q16-Q15-Q9_S/N 0008
2026-01-21 14:54:13

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 4.7400

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.997850	1.034508	-1.112083	-1.108545	1.012009	-0.007348
-190.067777	189.987093	-188.683212	-188.787094	190.557159	0.080219

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.037277	-0.122143	-0.118039	0.000000
Initial measured gains:	0.999379	0.992073	0.992641	1.003136
Gain corrections:	0.999379	1.007991	1.007414	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.015171
190.000000	190.489273

Measured offset: 0.012662
Measured gain: 1.002508
Gain correction: 0.997498

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998162	0.997381	-0.999836	-0.999191	1.005301	-0.011131
-190.067176	190.065545	-190.066925	-190.065964	190.073135	0.091176

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000777	-0.001684	-0.001040	0.007922
Final measured gains:	0.999996	0.999990	0.999988	0.999998

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000569	-0.000973	-0.001194	0.007529
Final measured offsets stdev:	0.000235	0.000730	0.000170	0.000393
Final measured gains mean:	0.999992	0.999989	0.999987	0.999997
Final measured gains stdev:	0.000007	0.000004	0.000005	0.000008

Saving channel 1 calibration constants to qspi

Burden resistor = 4.7400

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.997905	1.022240	-1.110425	-1.103370	0.988338	0.011450
-190.065246	190.085836	-188.790512	-188.702271	190.553925	0.075497

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.024355	-0.119842	-0.113215	0.000000
Initial measured gains:	0.999980	0.992663	0.992233	1.002655
Gain corrections:	0.999980	1.007392	1.007828	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.990189
190.000000	190.489853

Measured offset: -0.012455

Measured gain: 1.002644

Gain correction: 0.997363

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998432	0.998191	-0.999904	-0.999345	0.994468	0.005901
-190.065455	190.062585	-190.065262	-190.063843	190.057755	0.077471

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000227	-0.001481	-0.000927	-0.003717
Final measured gains:	0.999986	0.999991	0.999987	0.999994

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000665	-0.001798	-0.000985	-0.000679
Final measured offsets stdev:	0.000622	0.000240	0.000077	0.002635
Final measured gains mean:	0.999986	0.999986	0.999986	0.999994
Final measured gains stdev:	0.000002	0.000004	0.000003	0.000007

Saving channel 2 calibration constants to qspi

Burden resistor = 4.7400

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998041	1.010345	-1.103967	-1.099457	1.007846	-0.006873
-190.063642	189.881254	-188.584579	-188.734329	190.443253	0.067343

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.013331	-0.114293	-0.108968	0.000000
Initial measured gains:	0.998970	0.991617	0.992433	1.002989
Gain corrections:	0.998970	1.008454	1.007625	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.011084
190.000000	190.379013

Measured offset: 0.009137

Measured gain: 1.001947

Gain correction: 0.998057

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998293	0.997521	-0.999846	-0.999711	1.002063	-0.007025
-190.064242	190.061658	-190.061630	-190.063370	190.067276	0.072195

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000763	-0.001575	-0.001430	0.004536
Final measured gains:	0.999990	0.999978	0.999988	1.000006

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000727	-0.001681	-0.001745	0.004230
Final measured offsets stdev:	0.000101	0.000140	0.000241	0.000454
Final measured gains mean:	0.999988	0.999983	0.999984	1.000015
Final measured gains stdev:	0.000002	0.000003	0.000003	0.000008

Saving channel 3 calibration constants to qspi

Test data reviewed by _____ Date _____

