

## DCCT Test Results for DCCT S/N: y

Technician: y, Life #: y

Test performed: 03/14/25, 11:10 PM

### FLT12 Fault Signal 1 and 2 Test Results

Test	Value	Status
FLT12 Overall Testing	N/A	Pass
FLT12 Positive Assertion	N/A	Pass
FLT12 Negative Assertion	N/A	Pass
FLT12 De-Assertion	N/A	Pass
FLT12 Initial Voltage	14.76 V	Pass
FLT12 Positive Fault Voltage	-0.00 V	Fail
FLT12 Negative Fault Voltage	-0.00 V	Fail


Pass/Fail Threshold Voltage for FLT12 InitialVoltage: 14.3

Pass/Fail Threshold Voltage for FLT12 FaultVoltage: 13


# Current Test Waveforms

ALSu DCCT Test Data


Channel 1, IOUT 1



Channel 2, IOUT 2



Channel 3, SIGNAL GEN



Channel 1 & Channel 2 (IOUT 1 and IOUT 2) Phase Shift

