

# DCCT Test Results for DCCT S/N: y

Technician: y, Life #: dut\_info['Life']

Test performed: 03/14/25, 03:00 PM

## FLT12 Fault Signal 1 and 2 Test Results

| Test                  | Value  | Status |
|-----------------------|--------|--------|
| FLT12 Overall Testing | N/A    | Fail   |
| FLT12 Assertion       | N/A    | Fail   |
| FLT12 De-Assertion    | N/A    | Fail   |
| FLT12 Initial Voltage | 0.00 V | Fail   |
| FLT12 Fault Voltage   | 0.00 V | Fail   |


Pass/Fail Threshold Voltage for FLT12 InitialVoltage: 14.3

Pass/Fail Threshold Voltage for FLT12 FaultVoltage: 13


# Current Test Waveforms

ALSu DCCT Test Data


Channel 1, IOUT 1



Channel 2, IOUT 2



Channel 3, SIGNAL GEN



Channel 1 & Channel 2 (IOUT 1 and IOUT 2) Phase Shift

