

DCCT Test Results for DCCT S/N: y

Technician: y, Life #: dut_info['Life']

Test performed: 03/14/25, 03:00 PM

FLT12 Fault Signal 1 and 2 Test Results

Test	Value	Status
FLT12 Overall Testing	N/A	Fail
FLT12 Assertion	N/A	Fail
FLT12 De-Assertion	N/A	Fail
FLT12 Initial Voltage	0.00 V	Fail
FLT12 Fault Voltage	0.00 V	Fail


Pass/Fail Threshold Voltage for FLT12 InitialVoltage: 14.3

Pass/Fail Threshold Voltage for FLT12 FaultVoltage: 13


Current Test Waveforms

ALSu DCCT Test Data


Channel 1, IOUT 1



Channel 2, IOUT 2



Channel 3, SIGNAL GEN



Channel 1 & Channel 2 (IOUT 1 and IOUT 2) Phase Shift

