Report of Calibration
PSC PSC-4CH-HSF-EIC_S/N 0007
2026-05-20 15:05:28

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001871	1.005120	-1.011279	-1.011401	1.000545	-0.051913
-27.001377	27.001374	-26.804577	-26.819235	27.022717	-0.045263

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003375	-0.017355	-0.016917	0.000000
Initial measured gains:	0.999875	0.992069	0.992628	1.000997
Gain corrections:	0.999875	1.007995	1.007427	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.998648
27.000000	27.021322

Measured offset: -0.002224
Measured gain: 1.000872
Gain correction: 0.999129

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001897	1.002121	-1.002012	-1.002074	1.002459	-0.048808
-27.001154	27.001282	-27.001064	-27.001127	27.001663	-0.041350

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000227	-0.000122	-0.000185	0.000336
Final measured gains:	0.999996	0.999992	0.999992	1.000002

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000231	-0.000102	-0.000116	0.000224
Final measured offsets stdev:	0.000246	0.000106	0.000117	0.000361
Final measured gains mean:	0.999993	0.999989	0.999990	0.999991
Final measured gains stdev:	0.000008	0.000005	0.000004	0.000020

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001710	1.003104	-1.009617	-1.011941	0.996817	-0.044391
-27.001403	27.002362	-26.815647	-26.842476	27.027689	-0.042726

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001411	-0.015368	-0.016749	0.000000
Initial measured gains:	0.999983	0.992551	0.993494	1.001216
Gain corrections:	0.999983	1.007505	1.006549	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.994915
27.000000	27.026115

Measured offset: -0.006285
Measured gain: 1.001200
Gain correction: 0.998801

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001770	1.001415	-1.001820	-1.001774	1.001061	-0.054685
-27.001385	27.001333	-27.001516	-27.001482	27.000914	-0.047101

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000366	-0.000047	-0.000000	-0.000352
Final measured gains:	1.000012	1.000003	1.000004	0.999998

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000068	-0.000043	-0.000024	-0.000062
Final measured offsets stdev:	0.000255	0.000085	0.000095	0.000342
Final measured gains mean:	1.000003	0.999997	0.999994	0.999991
Final measured gains stdev:	0.000011	0.000004	0.000005	0.000008

Saving channel 2 calibration constants to qspi

lab{2}Chan3
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001450	1.001165	-1.010327	-1.011330	1.000913	-0.047315
-27.001273	27.005511	-26.818676	-26.817804	27.018610	-0.045277

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	0.000459	-0.016252	-0.017328	0.000000
Initial measured gains:	1.000174	0.992636	0.992563	1.000513
Gain corrections:	1.000174	1.007419	1.007492	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999569
27.000000	27.017054

Measured offset: -0.001103
Measured gain: 1.000672
Gain correction: 0.999328

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001479	1.001246	-1.001557	-1.001644	1.001691	-0.054913
-27.001407	27.001433	-27.001678	-27.001797	27.002218	-0.054550

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000243	-0.000071	-0.000157	0.000432
Final measured gains:	1.000010	1.000007	1.000009	1.000013

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000216	0.000006	-0.000016	0.000308
Final measured offsets stdev:	0.000065	0.000070	0.000077	0.000078
Final measured gains mean:	1.000012	1.000003	1.000001	0.999999
Final measured gains stdev:	0.000005	0.000003	0.000004	0.000009

Saving channel 3 calibration constants to qspi

lab{2}Chan4
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001651	1.002541	-1.013211	-1.010983	0.998627	-0.057777
-27.001565	27.012329	-26.833454	-26.813190	27.015402	-0.058064

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.000510	-0.018483	-0.016950	0.000000
Initial measured gains:	1.000380	0.993090	0.992396	1.000269
Gain corrections:	1.000380	1.006959	1.007662	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.997004
27.000000	27.013517

Measured offset: -0.003631
Measured gain: 1.000635
Gain correction: 0.999365

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001650	1.001468	-1.001449	-1.001499	1.001192	-0.058389
-27.001678	27.002159	-27.001717	-27.001659	27.002174	-0.043890

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000207	0.000210	0.000156	-0.000288
Final measured gains:	1.000025	1.000009	1.000005	1.000011

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000061	-0.000013	-0.000009	-0.000130
Final measured offsets stdev:	0.000346	0.000136	0.000098	0.000204
Final measured gains mean:	1.000014	1.000007	1.000003	0.999999
Final measured gains stdev:	0.000014	0.000002	0.000002	0.000015

Saving channel 4 calibration constants to qspi

Test data reviewed by ______________________________ Date_____________