

Report of Calibration
PSC PSC-4CH-HSF-EIC_S/N 0001
2026-04-30 23:43:52

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001326	1.004205	-1.010650	-1.010015	1.001039	0.008359
-27.000889	27.020505	-26.832043	-26.816208	27.044441	0.022422

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.002234	-0.016186	-0.016136	0.000000
Initial measured gains:	1.000644	0.993147	0.992563	1.001042
Gain corrections:	1.000644	1.006900	1.007493	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999844
27.000000	27.043669

Measured offset: -0.001842
Measured gain: 1.001686
Gain correction: 0.998317

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001479	1.001211	-1.001541	-1.001517	1.001883	-0.019621
-27.001109	27.001301	-27.001247	-27.001137	27.001938	0.034090

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000285	-0.000059	-0.000038	0.000673
Final measured gains:	1.000018	1.000003	1.000000	0.999999

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000031	-0.000141	-0.000116	0.000547
Final measured offsets stdev:	0.000183	0.000045	0.000049	0.000363
Final measured gains mean:	0.999990	0.999989	0.999990	0.999981
Final measured gains stdev:	0.000016	0.000008	0.000005	0.000012

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000943	1.005238	-1.012850	-1.010397	1.000874	0.035735
-27.001032	27.023410	-26.833241	-26.815144	27.062616	0.019478

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003599	-0.018826	-0.016975	0.000000
Initial measured gains:	1.000695	0.993089	0.992487	1.001675
Gain corrections:	1.000695	1.006960	1.007570	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999986
27.000000	27.061354

Measured offset: -0.002374
Measured gain: 1.002360
Gain correction: 0.997645

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000903	1.000648	-1.000885	-1.000980	1.000606	0.001517
-27.000918	27.001336	-27.001080	-27.001127	27.001680	0.042372

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000281	0.000026	-0.000071	-0.000058
Final measured gains:	1.000026	1.000007	1.000005	1.000015

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000136	-0.000017	-0.000028	-0.000139
Final measured offsets stdev:	0.000150	0.000056	0.000048	0.000112
Final measured gains mean:	1.000007	0.999995	0.999994	1.000014
Final measured gains stdev:	0.000019	0.000012	0.000010	0.000008

Saving channel 2 calibration constants to qspi

lab{2}Chan3
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000887	1.003285	-1.010111	-1.009449	1.002837	-0.004605
-27.001110	26.995830	-26.802605	-26.820934	27.008051	0.056205

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.002693	-0.017220	-0.015827	0.000000
Initial measured gains:	0.999705	0.992011	0.992741	1.000487
Gain corrections:	0.999705	1.008054	1.007312	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.001724
27.000000	27.006966

Measured offset: 0.001522
Measured gain: 1.000202
Gain correction: 0.999798

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000933	1.000980	-1.000883	-1.000894	1.001998	0.018658
-27.001121	27.001012	-27.001129	-27.001060	27.002377	0.041069

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000053	0.000051	0.000038	0.001005
Final measured gains:	0.999994	1.000002	0.999999	1.000013

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000011	0.000079	0.000080	0.001099
Final measured offsets stdev:	0.000111	0.000066	0.000060	0.000097
Final measured gains mean:	0.999998	0.999997	0.999996	1.000009
Final measured gains stdev:	0.000006	0.000003	0.000005	0.000010

Saving channel 3 calibration constants to qspi

lab{2}Chan4
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001077	1.004316	-1.010997	-1.010650	1.001324	0.005751
-27.001156	27.007847	-26.824261	-26.819878	27.038521	0.041389

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003106	-0.017113	-0.016921	0.000000
Initial measured gains:	1.000133	0.992815	0.992660	1.001295
Gain corrections:	1.000133	1.007237	1.007395	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP DAC RB

1.000000 1.000370
27.000000 27.037617

Measured offset: -0.001063

Measured gain: 1.001433

Gain correction: 0.998569

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000947	1.000937	-1.000899	-1.000991	1.001214	0.012655
-27.001082	27.001089	-27.001015	-27.001034	27.000957	0.051573

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000011	0.000047	-0.000047	0.000293
Final measured gains:	1.000001	0.999999	0.999996	0.999984

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000009	-0.000004	-0.000029	0.000322
Final measured offsets stdev:	0.000117	0.000107	0.000103	0.000084
Final measured gains mean:	0.999998	0.999994	0.999992	0.999987
Final measured gains stdev:	0.000005	0.000006	0.000007	0.000005

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date_____