

Report of Calibration
PSC PSC-2CH-HSS-EIC_S/N 0004
2026-05-20 20:53:26

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002894	1.004415	-1.012550	-1.014522	0.999136	-0.002069
-27.001506	27.012135	-26.822754	-26.812645	27.018019	0.044780

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001170	-0.016924	-0.019362	0.000000
Initial measured gains:	1.000350	0.992753	0.992288	1.000429
Gain corrections:	1.000350	1.007300	1.007771	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.996412
27.000000	27.016270

Measured offset: -0.004352
Measured gain: 1.000764
Gain correction: 0.999237

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002848	1.002911	-1.002915	-1.002906	1.002881	0.011443
-27.001577	27.001331	-27.001419	-27.001356	27.001072	0.040314

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000075	-0.000076	-0.000069	-0.000022
Final measured gains:	0.999988	0.999991	0.999989	0.999991

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000061	-0.000045	-0.000051	0.000107
Final measured offsets stdev:	0.000065	0.000102	0.000100	0.000069
Final measured gains mean:	0.999987	0.999990	0.999990	0.999992
Final measured gains stdev:	0.000014	0.000011	0.000011	0.000015

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002510	1.004593	-1.012273	-1.012547	1.001505	0.015602
-27.001690	27.002206	-26.810619	-26.818645	27.030029	0.055332

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.002144	-0.017507	-0.017482	0.000000
Initial measured gains:	0.999940	0.992275	0.992574	1.001189
Gain corrections:	0.999940	1.007785	1.007482	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999092
27.000000	27.028286

Measured offset: -0.002030
Measured gain: 1.001123
Gain correction: 0.998878

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002327	1.002442	-1.002545	-1.002509	1.002272	0.008225
-27.001685	27.001521	-27.001472	-27.001465	27.001369	0.056989

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000126	-0.000234	-0.000198	-0.000170
Final measured gains:	0.999989	0.999983	0.999985	1.000001

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000156	-0.000014	-0.000037	-0.000518
Final measured offsets stdev:	0.000147	0.000123	0.000105	0.000242
Final measured gains mean:	0.999995	0.999984	0.999984	1.000006
Final measured gains stdev:	0.000015	0.000014	0.000014	0.000012

Saving channel 2 calibration constants to qspi

Test data reviewed by _____ Date_____