

Report of Calibration
PSC PSC-2CH-HSS-EIC_S/N 0007
2026-05-20 21:25:05

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002787	1.011528	-1.014078	-1.014679	1.003303	0.024200
-27.001606	27.010874	-26.822685	-26.825890	27.055275	0.027480

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.008720	-0.018628	-0.019128	0.000000
Initial measured gains:	1.000020	0.992684	0.992784	1.002024
Gain corrections:	1.000020	1.007370	1.007268	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.000452
27.000000	27.053631

Measured offset: -0.001593
Measured gain: 1.002045
Gain correction: 0.997959

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002722	1.002646	-1.002885	-1.002832	1.002168	0.009510
-27.001692	27.001117	-27.001064	-27.001091	27.000689	0.030662

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000056	-0.000194	-0.000138	-0.000480
Final measured gains:	0.999981	0.999970	0.999973	1.000002

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000036	-0.000063	-0.000061	-0.000020
Final measured offsets stdev:	0.000058	0.000174	0.000172	0.000341
Final measured gains mean:	0.999993	0.999989	0.999987	1.000004
Final measured gains stdev:	0.000010	0.000014	0.000013	0.000006

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002282	1.006892	-1.013796	-1.013526	1.006784	-0.001038
-27.001523	26.999601	-26.812737	-26.826433	27.073843	0.053744

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.004862	-0.019236	-0.018427	0.000000
Initial measured gains:	0.999749	0.992296	0.992833	1.002860
Gain corrections:	0.999749	1.007764	1.007219	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.004640
27.000000	27.072422

Measured offset: 0.002033
Measured gain: 1.002607
Gain correction: 0.997400

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002503	1.002328	-1.002199	-1.002263	1.003237	0.009362
-27.001566	27.001402	-27.001335	-27.001404	27.002239	0.046233

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000176	0.000307	0.000243	0.000912
Final measured gains:	1.000000	1.000003	1.000003	0.999997

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000021	-0.000009	-0.000019	0.000963
Final measured offsets stdev:	0.000105	0.000165	0.000149	0.000311
Final measured gains mean:	0.999991	0.999990	0.999990	0.999999
Final measured gains stdev:	0.000006	0.000007	0.000007	0.000006

Saving channel 2 calibration constants to qspi

Test data reviewed by _____ Date_____