

Report of Calibration
PSC PSC-2CH-HSS-EIC_S/N 0003
2026-05-20 20:23:14

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002426	1.004204	-1.013023	-1.011807	1.001099	-0.005654
-27.001658	26.998557	-26.817129	-26.819206	27.036470	0.043119

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001965	-0.018120	-0.016777	0.000000
Initial measured gains:	0.999812	0.992495	0.992622	1.001578
Gain corrections:	0.999812	1.007562	1.007433	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.998840
27.000000	27.034885

Measured offset: -0.002546
Measured gain: 1.001386
Gain correction: 0.998616

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002461	1.002523	-1.002428	-1.002418	1.002152	0.015862
-27.001515	27.001887	-27.001762	-27.001816	27.002081	0.055409

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000050	0.000044	0.000057	-0.000393
Final measured gains:	1.000012	1.000011	1.000013	1.000022

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000066	-0.000057	-0.000073	-0.000143
Final measured offsets stdev:	0.000157	0.000154	0.000142	0.000264
Final measured gains mean:	0.999989	0.999991	0.999992	1.000005
Final measured gains stdev:	0.000012	0.000011	0.000012	0.000013

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002007	1.003812	-1.011054	-1.010562	1.004350	0.004357
-27.001574	27.019593	-26.833149	-26.810150	27.091879	0.028442

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001181	-0.015887	-0.016262	0.000000
Initial measured gains:	1.000624	0.993174	0.992308	1.002758
Gain corrections:	1.000624	1.006873	1.007751	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.002261
27.000000	27.089895

Measured offset: -0.001109
Measured gain: 1.003371
Gain correction: 0.996641

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001854	1.002023	-1.001960	-1.001880	1.002502	0.023672
-27.000936	27.001516	-27.001482	-27.001413	27.001768	0.032288

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000154	-0.000089	-0.000009	0.000488
Final measured gains:	1.000016	1.000017	1.000017	0.999991

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000077	0.000052	0.000109	0.000397
Final measured offsets stdev:	0.000243	0.000172	0.000151	0.000292
Final measured gains mean:	1.000008	1.000004	1.000004	0.999995
Final measured gains stdev:	0.000020	0.000013	0.000014	0.000009

Saving channel 2 calibration constants to qspi

Test data reviewed by _____ Date_____