

Report of Calibration
PSC PSC-4CH-HSF-EIC_ S/N 0001
2026-05-10 23:15:00

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000258	1.001831	-1.008952	-1.008220	0.999630	-0.003594
-27.000904	27.012867	-26.825764	-26.808781	27.039728	0.057340

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001173	-0.015766	-0.015659	0.000000
Initial measured gains:	1.000400	0.992930	0.992305	1.001117
Gain corrections:	1.000400	1.007121	1.007755	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999334
27.000000	27.038948

Measured offset: -0.002190
Measured gain: 1.001524
Gain correction: 0.998479

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000320	1.000340	-1.000227	-1.000340	1.000699	0.007445
-27.000949	27.000773	-27.000957	-27.000961	27.001022	0.023765

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000027	0.000097	-0.000020	0.000363
Final measured gains:	0.999992	1.000004	1.000000	0.999996

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000013	-0.000041	-0.000129	0.000416
Final measured offsets stdev:	0.000090	0.000094	0.000080	0.000054
Final measured gains mean:	0.999992	0.999989	0.999989	0.999989
Final measured gains stdev:	0.000007	0.000009	0.000009	0.000014

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000150	1.003999	-1.011833	-1.008955	0.999772	0.021254
-27.000763	27.015859	-26.825954	-26.807030	27.057676	0.051076

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003417	-0.018857	-0.016596	0.000000
Initial measured gains:	1.000433	0.992827	0.992210	1.001770
Gain corrections:	1.000433	1.007224	1.007851	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999635
27.000000	27.057135

Measured offset: -0.002576

Measured gain: 1.002212

Gain correction: 0.997793

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000043	0.999801	-0.999839	-1.000094	0.999438	0.003454
-27.000623	27.000843	-27.000795	-27.000681	27.000151	0.029827

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000260	0.000218	-0.000050	-0.000350
Final measured gains:	1.000018	1.000014	1.000000	0.999987

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000171	0.000007	-0.000048	-0.000136
Final measured offsets stdev:	0.000147	0.000129	0.000045	0.000250
Final measured gains mean:	1.000008	1.000000	0.999998	0.999997
Final measured gains stdev:	0.000009	0.000008	0.000005	0.000007

Saving channel 2 calibration constants to qspi

lab{2}Chan3
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000046	1.002150	-1.008816	-1.007938	1.001653	0.016231
-27.000722	26.988937	-26.795017	-26.813103	27.002701	0.046725

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.002639	-0.017020	-0.015413	0.000000
Initial measured gains:	0.999466	0.991751	0.992480	1.000549
Gain corrections:	0.999466	1.008317	1.007576	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.001653
27.000000	27.001961

Measured offset: 0.001641

Measured gain: 1.000012

Gain correction: 0.999988

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000105	1.000061	-1.000262	-1.000185	1.000798	0.020655
-27.000739	27.000642	-27.000671	-27.000778	27.001484	0.041539

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000042	-0.000166	-0.000082	0.000732
Final measured gains:	0.999998	0.999991	0.999998	1.000004

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000030	-0.000030	-0.000039	0.000786
Final measured offsets stdev:	0.000110	0.000134	0.000137	0.000062
Final measured gains mean:	1.000002	0.999997	0.999998	1.000005
Final measured gains stdev:	0.000005	0.000012	0.000014	0.000009

Saving channel 3 calibration constants to qspi

lab{2}Chan4
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999470	1.002942	-1.009480	-1.009142	0.999805	0.011747
-27.000195	27.000209	-26.816357	-26.811670	27.032217	-0.007936

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003605	-0.017461	-0.017290	0.000000
Initial measured gains:	0.999867	0.992545	0.992377	1.001352
Gain corrections:	0.999867	1.007511	1.007681	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.000337
27.000000	27.031773

Measured offset: -0.000872

Measured gain: 1.001209

Gain correction: 0.998792

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999669	0.999859	-0.999631	-0.999707	0.999257	0.028221
-27.000414	27.000818	-27.000229	-27.000418	27.000929	0.056593

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000182	0.000033	-0.000039	-0.000630
Final measured gains:	1.000008	0.999994	0.999999	1.000027

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000033	-0.000002	-0.000072	-0.000153
Final measured offsets stdev:	0.000276	0.000165	0.000222	0.000318
Final measured gains mean:	1.000002	0.999994	0.999993	1.000011
Final measured gains stdev:	0.000005	0.000005	0.000011	0.000015

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date_____