

Report of Calibration
PSC PSC-4CH-HSF-EIC_ S/N 0001
2026-05-11 22:24:25

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000901	1.002466	1.009279	1.008602	1.000370	0.034309
-27.001578	27.014094	26.826977	26.810108	27.042046	0.032703

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001143	-0.015422	-0.015368	0.000000
Initial measured gains:	-1.000421	-0.992963	-0.992340	1.001155
Gain corrections:	-1.000421	-1.007087	-1.007719	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999493
27.000000	27.040527

Measured offset: -0.002086
Measured gain: 1.001578
Gain correction: 0.998424

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000925	1.000766	1.001119	1.001084	1.001083	-0.004823
-27.001573	27.001460	27.001328	27.001408	27.001549	0.049634

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000161	-0.000210	-0.000172	0.000326
Final measured gains:	-1.000002	-0.999983	-0.999988	0.999991

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000110	-0.000042	-0.000030	0.000297
Final measured offsets stdev:	0.000127	0.000122	0.000114	0.000322
Final measured gains mean:	-0.999998	-0.999989	-0.999991	0.999994
Final measured gains stdev:	0.000005	0.000004	0.000003	0.000008

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000148	1.004290	1.012058	1.009465	1.000359	0.004755
-27.000839	27.017152	26.827269	26.808716	27.060032	0.023829

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003675	-0.019045	-0.017066	0.000000
Initial measured gains:	-1.000468	-0.992866	-0.992253	1.001800
Gain corrections:	-1.000468	-1.007185	-1.007808	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.000145
27.000000	27.059130

Measured offset: -0.002124

Measured gain: 1.002269

Gain correction: 0.997736

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000170	1.000080	1.000105	1.000069	1.000216	0.010985
-27.001020	27.000843	27.000870	27.000933	27.001528	0.019208

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000087	0.000062	0.000102	0.000115
Final measured gains:	-0.999997	-0.999997	-1.000001	1.000021

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000012	-0.000055	-0.000036	0.000109
Final measured offsets stdev:	0.000130	0.000134	0.000133	0.000126
Final measured gains mean:	-0.999996	-0.999997	-0.999996	1.000007
Final measured gains stdev:	0.000009	0.000008	0.000011	0.000012

Saving channel 2 calibration constants to qspi

lab{2}Chan3
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000846	1.002889	1.009698	1.008728	1.002661	-0.008695
-27.001414	26.990482	26.796621	26.814474	27.004997	0.017676

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.002542	-0.017076	-0.015381	0.000000
Initial measured gains:	-0.999501	-0.991783	-0.992507	1.000567
Gain corrections:	-0.999501	-1.008285	-1.007550	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.001669
27.000000	27.003122

Measured offset: 0.001613

Measured gain: 1.000056

Gain correction: 0.999944

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000863	1.000740	1.000764	1.000860	1.001773	-0.000247
-27.001687	27.001507	27.001436	27.001453	27.003283	0.022952

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000120	0.000093	-0.000007	0.001005
Final measured gains:	-0.999998	-0.999994	-0.999991	1.000029

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000137	0.000050	0.000001	0.000706
Final measured offsets stdev:	0.000093	0.000084	0.000064	0.000311
Final measured gains mean:	-1.000006	-0.999998	-0.999996	1.000025
Final measured gains stdev:	0.000007	0.000004	0.000006	0.000011

Saving channel 3 calibration constants to qspi

lab{2}Chan4
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000891	1.003766	1.010336	1.009869	1.000913	0.002001
-27.001703	27.002236	26.817936	26.813385	27.035110	0.047370

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.002965	-0.016882	-0.016573	0.000000
Initial measured gains:	-0.999910	-0.992569	-0.992412	1.001374
Gain corrections:	-0.999910	-1.007487	-1.007646	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.000079
27.000000	27.033422

Measured offset: -0.001203

Measured gain: 1.001282

Gain correction: 0.998719

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000757	1.000821	1.000848	1.000947	1.000808	-0.007602
-27.001643	27.001581	27.001762	27.001766	27.002060	0.015641

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000068	-0.000090	-0.000193	-0.000031
Final measured gains:	-0.999995	-1.000001	-0.999997	1.000019

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000082	-0.000011	-0.000032	0.000001
Final measured offsets stdev:	0.000161	0.000071	0.000095	0.000136
Final measured gains mean:	-0.999994	-0.999994	-0.999993	0.999998
Final measured gains stdev:	0.000007	0.000004	0.000005	0.000012

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date_____