

Report of Calibration
PSC PSC-4CH-HSF-EIC_S/N 0001
2026-05-19 14:08:16

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002134	1.004211	-1.011351	-1.011165	1.001779	0.023016
-27.001049	27.014790	-26.827587	-26.827612	27.041689	0.052203

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001627	-0.016258	-0.016064	0.000000
Initial measured gains:	1.000449	0.992974	0.992982	1.001128
Gain corrections:	1.000449	1.007076	1.007068	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999602
27.000000	27.040516

Measured offset: -0.001971
Measured gain: 1.001574
Gain correction: 0.998429

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002083	1.001881	-1.002053	-1.002082	1.002442	0.007083
-27.000989	27.000880	-27.000942	-27.000990	27.001160	0.039925

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000206	0.000029	0.000002	0.000572
Final measured gains:	1.000004	0.999999	1.000000	0.999989

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000061	-0.000069	-0.000082	0.000454
Final measured offsets stdev:	0.000333	0.000085	0.000065	0.000214
Final measured gains mean:	0.999997	0.999993	0.999994	0.999999
Final measured gains stdev:	0.000009	0.000007	0.000006	0.000006

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001896	1.005391	-1.013423	-1.010625	1.001104	-0.006992
-27.001133	27.010829	-26.821861	-26.799139	27.052780	0.013247

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003256	-0.018880	-0.016849	0.000000
Initial measured gains:	1.000239	0.992661	0.991895	1.001778
Gain corrections:	1.000239	1.007393	1.008171	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999262
27.000000	27.051821

Measured offset: -0.002759

Measured gain: 1.002021

Gain correction: 0.997983

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002069	1.001908	-1.002223	-1.002130	1.002042	-0.037007
-27.001283	27.001929	-27.001709	-27.001572	27.002224	0.044991

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000192	-0.000143	-0.000052	0.000128
Final measured gains:	1.000031	1.000010	1.000009	1.000006

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000065	-0.000055	-0.000032	-0.000442
Final measured offsets stdev:	0.000093	0.000066	0.000051	0.000311
Final measured gains mean:	1.000006	0.999993	0.999994	1.000005
Final measured gains stdev:	0.000015	0.000010	0.000010	0.000005

Saving channel 2 calibration constants to qspi

lab{2}Chan3
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001365	1.004346	-1.010886	-1.009898	1.003824	0.017980
-27.000641	27.000858	-26.807045	-26.813667	27.014460	0.049934

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003087	-0.017343	-0.016062	0.000000
Initial measured gains:	0.999894	0.992188	0.992480	1.000543
Gain corrections:	0.999894	1.007874	1.007577	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.002393
27.000000	27.013741

Measured offset: 0.001956
Measured gain: 1.000436
Gain correction: 0.999564

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001401	1.001458	-1.001531	-1.001573	1.002080	0.006694
-27.000487	27.000075	-27.000189	-27.000257	27.000923	0.026097

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000075	-0.000147	-0.000188	0.000614
Final measured gains:	0.999982	0.999984	0.999985	1.000009

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000022	-0.000128	-0.000140	0.000769
Final measured offsets stdev:	0.000177	0.000094	0.000082	0.000121
Final measured gains mean:	0.999991	0.999989	0.999989	1.000001
Final measured gains stdev:	0.000020	0.000008	0.000008	0.000013

Saving channel 3 calibration constants to qspi

lab{2}Chan4
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001620	1.004751	-1.011366	-1.010974	1.001680	0.017581
-27.000917	27.003424	-26.819748	-26.814087	27.035374	0.054696

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003156	-0.017101	-0.016912	0.000000
Initial measured gains:	0.999976	0.992657	0.992454	1.001347
Gain corrections:	0.999976	1.007397	1.007603	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.000222
27.000000	27.034683

Measured offset: -0.001104
Measured gain: 1.001325
Gain correction: 0.998676

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001558	1.001739	-1.001855	-1.001890	1.001905	0.007935
-27.001001	27.000845	-27.000896	-27.000874	27.000404	0.040057

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000194	-0.000313	-0.000350	0.000189
Final measured gains:	0.999987	0.999985	0.999982	0.999977

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000009	-0.000107	-0.000094	0.000065
Final measured offsets stdev:	0.000235	0.000187	0.000201	0.000217
Final measured gains mean:	0.999993	0.999989	0.999988	0.999994
Final measured gains stdev:	0.000006	0.000006	0.000005	0.000013

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date_____