

Project Tetra

Table of Contents

| | |
|--|---|
| Stakeholder Concerns | 2 |
| 1: Volumes fill most compliant lung segment | 2 |
| 2: PEEP would be impossible to manage | 2 |
| 3: Monitoring pulmonary mechanics impossible | 2 |
| 4: Alarm monitoring infeasible | 2 |
| 5: Individualized management for clinical improvement | 2 |
| 6: Stop ventilation in the event of cardiac arrest | 3 |
| 7: Added circuit volume | 3 |
| 8: External monitoring required | 3 |
| 9: Patients deteriorate and recover at different rates | 3 |
| 10: Greatest risk with sudden deterioration | 4 |
| User Stories | 4 |

Stakeholder Concerns

The Project Tetra is guided by a series of stakeholder concerns, listed below.

1: Volumes fill most compliant lung segment

Volumes would go to the most compliant lung segments

Reference:

- [Joint Statement on Multiple Patients Per Ventilator](#)

2: PEEP would be impossible to manage

Positive end-expiratory pressure (PEEP), which is of critical importance in these patients, would be impossible to manage.

Reference:

- [Joint Statement on Multiple Patients Per Ventilator](#)

3: Monitoring pulmonary mechanics impossible

Monitoring patients and measuring pulmonary mechanics would be challenging, if not impossible.

Reference:

- [Joint Statement on Multiple Patients Per Ventilator](#)

4: Alarm monitoring infeasible

Alarm monitoring and management would not be feasible.

Reference:

- [Joint Statement on Multiple Patients Per Ventilator](#)

5: Individualized management for clinical improvement

Individualized management for clinical improvement or deterioration would be impossible.

Reference:

- [Joint Statement on Multiple Patients Per Ventilator](#)

6: Stop ventilation in the event of cardiac arrest

In the case of a cardiac arrest, ventilation to all patients would need to be stopped to allow the change to bag ventilation without aerosolizing the virus and exposing healthcare workers.

Reference:

- [Joint Statement on Multiple Patients Per Ventilator](#)

7: Added circuit volume

The added circuit volume defeats the operational self-test (the test fails).

Reference:

- [Joint Statement on Multiple Patients Per Ventilator](#)

8: External monitoring required

Additional external monitoring would be required. The ventilator monitors the average pressures and volumes.

Reference:

- [Joint Statement on Multiple Patients Per Ventilator](#)

9: Patients deteriorate and recover at different rates.

Patients deteriorate and recover at different rates, and distribution of gas to each patient would be unequal and unmonitored.

Reference:

- [Joint Statement on Multiple Patients Per Ventilator](#)

10: Greatest risk with sudden deterioration

The greatest risks occur with sudden deterioration of a single patient (e.g., pneumothorax, kinked endotracheal tube), with the balance of ventilation distributed to the other patients.

Reference:

- [Joint Statement on Multiple Patients Per Ventilator](#)

User Stories

The Project Tetra's stakeholder concerns are then used to identify a series of user stories to capture the functional requirements of the Project Tetra project.