

PT. Esco Utama  
Taman Tekno BSD CITY BLOK H8 No.1  
Setu, Setu, Kota Tangerang Selatan - Banten 15314  
Tel : 021-30446565, 021-30446566 | Fax 021-30446567



CALIBRATION CERTIFICATE

Certificate Number: CERT-B 00234/23

OWNER IDENTITY

Owner : Balai POM di Tarakan  
Address : Karang Balik, Kota Tarakan – Kalimantan

INSTRUMENT IDENTIFICATION

Name : Biological Safety Cabinet  
Manufacture : ESCO  
Type / Model : LA2-4S8-G4  
Serial Number : 2022-163363

METHOD OF CALIBRATION : NSF49:2016, EN 12469:2000

DATE OF CALIBRATION : 11-May-2023

RECOMENDED DUE DATE : 11-May-2024

REFERENCE STANDARD USED FOR RECALIBRATION

Test Equioment	Serial Number	Cal Due Date
Photometer	22702	23-Feb-24
Anemometer	9565P1326004	16-Aug-23
UV Meter	S.043170	12-Aug-23
Light Meter	Q882464	03-Aug-23
Sound Meter	I.301079	01-Nov-23
Particle Counter	93101509004	09-Dec-23



Agus Susanto  
Abriyadi  
Reviewed/Approved by  
Date of Issue : 11-May-2023

This certificate of calibration shall not be reproduced except in full without written approval of PT. Esco Utama  
This certificate is valid when sealed by PT Esco Utama