

PT. Esco Utama
Taman Tekno BSD CITY BLOK H8 No.1
Setu, Setu, Kota Tangerang Selatan - Banten 15314
Tel : 021-30446565, 021-30446566 | Fax 021-30446567



CALIBRATION CERTIFICATE

Certificate Number: CERT-B 00250/22

OWNER IDENTITY

Owner : BPOM Tarakan
Address : Jl. Yos Sudarso - Tarakan

INSTRUMENT IDENTIFICATION

Name : Biosafety Cabinet
Manufacture : ESCO
Type / Model : FC2-4A1
Serial Number : 2011-60859


METHOD OF CALIBRATION : NSF49:2016, EN 12469:2000

DATE OF CALIBRATION : 14-Jun-2022

RECOMENDED DUE DATE : 14-Jun-2023

REFERENCE STANDARD USED FOR RECALIBRATION

Test Equioment	Serial Number	Cal Due Date
Anemometer	T.95451509004	Mar-2023
UV Meter	S.043164	Sep-2023
Light Meter	S.028517	Jan-2023
Sound Meter	I.301132	Jul-2023



Abriyadi
Reviewed/Approved by
Date of Issue : 14-Jun-2023

This certificate of calibration shall not be reproduced except in full without written approval of PT. Esco Utama
This certificate is valid when sealed by PT Esco Utama