Method
Author: Beckman Coulter
Description:

Start

Instrument Setup
Deck: Multichannel
Pause to confirm setup.
Verify that the pod is set up in its default configuration.
Items:
P1: As Is.
P10: As Is.
P11: As Is.
P12: As Is.
P13: As Is.
P14: As Is.
P2: As Is.
P3: Nalgene_Reservoir named Media with nominal volume: 55000 µL of Water in all wells
P4: CellCarrier_96 named Cell Plate with known volume: 155 µL of Water in all wellsperkin_elmer_cellplate_lid named Cell Plate Lid
P5: As Is.
P6: As Is.
P7: Nothing
P8: ubottom_96_3795 named Waste with known volume: 0 µL of Water in all wells
P9: As Is.
Pelt1: Nothing
Shuttle_S8: Nothing
TL1: BC190F. Discard the tips to "<TipBox>". When done, move the box to "<Home>". Use these tips "999" times.
TL2: Nothing
TL3: Nothing
TR1: Nothing
WS1: Water
WS2: Water

Begin Holding Labware Step
Holding the top "1" plate(s) at "P4" using pod "Pod1" with "A1 near" the gripper...

Transfer
Using Pod1, execute the following transfer:
From: Cell Plate, section 1, Well Contents
Use the following custom technique:
  Use the following pipetting template: MC Pipetting Cells Conditioning Volume Calibration
  Calibration Offset: 0.522517707
  Calibration Slope: 1.109715355
  Minimum Pipetting Height: 0.4 mm
  Prewet: False
  Aspirate Blowout: True
  Follow Liquid: True
  Height: 3 mm from the bottom
  Mix: False
  Mix Aspirate Speed: 100µL/s
  Mix Aspirate Height: 0 mm from the bottom
  Mix Dispense Speed: 100µL/s
  Mix Dispense Height: 0 mm from the bottom
  Mix Count: 1
  Mix Volume: 10 µL
  Operation speed: 50µL/s
  Tip Touch: False
  Trailing Air Gap: False
  Override Liquid Type Settings
    Trailing Air Gap Volume: 0 µL
    Aspirate Delay: 0 ms
    Aspirate Speed: 5 µL/s
    Blowout Volume: 2 µL
    Blowout Delay: 0 ms
    Prewet Overage: 0 µL

```
    Prewet Delay: 0 ms
    Dispense Delay: 0 ms
    Dispense Speed: 10 µL/s
    Dispense Cutoff Velocity: 10 µL/s
    Tip Touch Height: -1 mm from the top
    Tip Touch Angle: 90
    Tip Touch Speed: 100 %
    Tip Touch Delay: 0 ms

To: Waste, 75µL, section 1, Tip Contents
Use the following custom technique:
  Use the following pipetting template: MC Pipetting Cells Conditioning Volume Calibration
  Calibration Offset: 0.52217707
  Calibration Slope: 1.109715355
  Minimum Pipetting Height: 0.4 mm
  Prewet: False
  Blowout: True
  Follow Liquid: True
  Height: 0.5 mm from the bottom
  Mix: False
  Mix Aspirate Speed: 100µL/s
  Mix Aspirate Height: 2 mm from the bottom
  Mix Dispense Speed: 100µL/s
  Mix Dispense Height: 5 mm from the bottom
  Mix Count: 4
  Mix Volume: 20 µL
  Operation speed: 50µL/s
  Tip Touch: False
  Override Liquid Type Settings
    Trailing Air Gap Volume: 0 µL
    Aspirate Delay: 0 ms
    Aspirate Speed: 50 µL/s
    Blowout Volume: 2 µL
    Blowout Delay: 0 ms
    Prewet Overage: 0 µL
    Prewet Delay: 0 ms
    Dispense Delay: 0 ms
    Dispense Speed: 50 µL/s
    Dispense Cutoff Velocity: 150 µL/s
    Tip Touch Height: -1 mm from the top
    Tip Touch Angle: 90
    Tip Touch Speed: 100 %
    Tip Touch Delay: 0 ms

Dispense up to 1 time(s) per draw.
Load BC190F onto the pod.
Keep the tips when finished.
Do not split transfer volumes that exceed tip capacity.
Stop when finished with Destinations.
--------------------------------------------------------------------------------
Transfer
Using Pod1, execute the following transfer:
From: Media, section 1, Well Contents
Use the following custom technique:
  Use the following pipetting template: MC Pipetting Cells Conditioning Volume Calibration
  Calibration Offset: 0.522517707
  Calibration Slope: 1.109715355
  Minimum Pipetting Height: 0.3 mm
  Prewet: False
  Aspirate Blowout: True
  Follow Liquid: True
  Height: 0.3 mm from the bottom
  Mix: False
  Mix Aspirate Speed: 20µL/s
  Mix Aspirate Height: 1.5 mm from the bottom
  Mix Dispense Speed: 20µL/s
  Mix Dispense Height: 1.5 mm from the bottom
  Mix Count: 1
```

```
  Mix Volume: 20 µL
  Operation speed: 50µL/s
  Tip Touch: False
  Trailing Air Gap: False
  Override Liquid Type Settings
    Trailing Air Gap Volume: 0 µL
    Aspirate Delay: 0 ms
    Aspirate Speed: 50 µL/s
    Blowout Volume: 2 µL
    Blowout Delay: 0 ms
    Prewet Overage: 0 µL
    Prewet Delay: 0 ms
    Dispense Delay: 0 ms
    Dispense Speed: 50 µL/s
    Dispense Cutoff Velocity: 100 µL/s
    Tip Touch Height: -1 mm from the top
    Tip Touch Angle: 90
    Tip Touch Speed: 100 %
    Tip Touch Delay: 0 ms

To: Cell Plate, 75µL, section 1, Tip Contents
Use the following custom technique:
  Use the following pipetting template: MC Pipetting Cells Conditioning Volume Calibration
  Calibration Offset: 0.522517707
  Calibration Slope: 1.109715355
  Minimum Pipetting Height: 0.8 mm
  Prewet: False
  Blowout: True
  Follow Liquid: True
  Height: 2 mm from the bottom
  Mix: False
  Mix Aspirate Speed: 100µL/s
  Mix Aspirate Height: 2 mm from the bottom
  Mix Dispense Speed: 100µL/s
  Mix Dispense Height: 5 mm from the bottom
  Mix Count: 4
  Mix Volume: 20 µL
  Operation speed: 50µL/s
  Tip Touch: False
  Override Liquid Type Settings
    Trailing Air Gap Volume: 0 µL
    Aspirate Delay: 0 ms
    Aspirate Speed: 5 µL/s
    Blowout Volume: 2 µL
    Blowout Delay: 0 ms
    Prewet Overage: 0 µL
    Prewet Delay: 0 ms
    Dispense Delay: 0 ms
    Dispense Speed: 5 µL/s
    Dispense Cutoff Velocity: 10 µL/s
    Tip Touch Height: -1 mm from the top
    Tip Touch Angle: 90
    Tip Touch Speed: 100 %
    Tip Touch Delay: 0 ms

Dispense up to 1 time(s) per draw.
Keep the current tips
Keep the tips when finished.
Do not split transfer volumes that exceed tip capacity.
Stop when finished with Destinations.
------------------------------------------------------------------------------
End Holding Labware
------------------------------------------------------------------------------
Finish
Clear current instrument setup of all labware.
Clear current device setup of all labware.
Unload disposable tips from all pods.
Move all pods and grippers to their park locations.
```

Clear all global variables.
