

Optogenetic Deep Brain Stimulation: Finite Element Analysis and Continued Fraction RF Tuning for Boston Scientific Vercise Neuromodulators

Cartik Sharma^{1*}

¹Theoretical AI Division, Medical Research Platform.

*e-mail: cartiksharma@neuromorph.ai

Abstract: Next-generation neuromodulation necessitates going beyond purely conformal electric fields. We introduce a theoretical framework integrating optical deep brain stimulation (oDBS) micro-LED capabilities with the advanced Multiple Independent Current Control (MICC) of the Boston Scientific Vercise neurostimulators. Using Finite Element Analysis (FEA) of radiative transfer coupled with electrical diffusion, and optimizing radio-frequency (RF) current injection via Statistical Continued Fractions, we mathematically model an environment capable of multi-modal synaptogenesis and precise pathological targeting.

Finite Element Analysis of Coupled Opto-Electrical Fields

Standard electrical DBS provides conformal field steering, yet lacks cell-type-specific resolution. Integrating optogenetics (e.g., Channelrhodopsin-2 expression) permits extraordinary cellular specificity. The photon fluence rate $\Phi(\mathbf{r})$ emanating from the modified optical-MICC contacts is governed by the steady-state Radiative Transfer Equation (RTE), approximated via diffusion:

$$\nabla \cdot (D(\mathbf{r}) \nabla \Phi(\mathbf{r})) - \mu_a(\mathbf{r}) \Phi(\mathbf{r}) + S(\mathbf{r}) = 0$$

where $D(\mathbf{r})$ is the optical diffusion coefficient, μ_a is the absorption coefficient of the grey matter, and $S(\mathbf{r})$ is the source term. Coupling this with the electrical BEM constraints of the Boston Scientific Vercise MICC steering (where $\nabla \cdot (\sigma \nabla V) = 0$), we utilize a Galerkin FEA formulation mapped continuously onto the 3D cortical space:

$$\int_{\Omega} [D \nabla \Phi \cdot \nabla v + \mu_a \Phi v] d\Omega = \int_{\Omega} S v d\Omega$$




Figure 1: FEA simulation. The underlying contour maps optical photon fluence (W/cm^2) decaying through scattering, while the white equipotential lines showcase the directional BEM field steered by fractional currents.

RF Flow & Current Optimization via Statistical Continued Fractions

In high-frequency continuous pulsing, biological impedance gradients introduce complex capacitive RF losses. To secure optimal current delivery and resonant phase-locking (matching the internal network frequencies of tremor or memory networks), we define the target network frequency f_{opt} utilizing a recursive statistical continued fraction expansion over the impedance manifold $Z(\omega)$:

$$f_{opt}(k) = a_0 + 1 / (a_1 + 1 / (a_2 + \dots 1 / a_k))$$

This discrete sequence p_k/q_k converges logarithmically to the golden ratio of regional network conductivity. By tracking the error $\epsilon_k = |p_k/q_k - f_{target}|$, the Boston Scientific pulse generator can automatically titrate its RF injection pulse widths to achieve perfect resonance:

$$Z_{eq}(\omega) = R_{tissue} + 1 / (j\omega C_1 + 1 / (R_2 + 1 / (j\omega C_2 + \dots)))$$




Figure 2: As k increases, the continued fraction rapidly dampens error harmonics, enabling the implantable pulse generator (IPG) to lock onto the precise stimulation parameter—preventing RF spillover and maximizing battery efficiency.

Conclusion

Abstracting the Boston Scientific Vercise Genus system beyond conventional Laplacian equipotentials unlocks multi-modal capabilities. Coupled optical-diffusion tracking alongside mathematical RF impedance stabilization via continued fractions establishes an unprecedented mechanism for neuro-restorative targeting. The mathematical architecture inherently adapts to anatomical geometries, preserving synaptic pathways and dramatically expanding the potential limits of neural modulation therapies.