

CERTIFICATE OF ASSURANCE

PRODUCT:	Plate count agar
PRODUCT CODE:	M10095
BATCH NUMBER:	62513
EXPIRY DATE:	03-Mar-25

Quality Control Testing:

This product was tested in the Quality Control Laboratory and met defined acceptance criteria using our standard test method:

Appearance of product was satisfactory

Growth Performance was satisfactory, using the following reference strains

REFERENCE STRAIN	ATCC NUMBER	RESULT
Escherichia coli	25922	Straw coloured colonies
Salmonella typhymurium	14028	Straw coloured colonies

STERILITY: A representative sample was incubated at 36degC for 72 hrs. No growth was observed

RELEASED ON:	72 HOURS
--------------	----------

Satisfactory for Issue



Olga Hattingh
Manager