



## 8°/Hemispherical Reflectance Calibration Certificate

**SERIAL NUMBER:** 99AA10-0614-9856

**DATE OF REPORT:** August 25, 2014

**PAGE 1 OF 6**

**Rendered To:**

**Authorization: Sales Order:**

**Calibration Laboratory:** Labsphere, Inc., Reflectance Calibration Laboratory  
231 Shaker Street  
North Sutton, NH 03260  
Tel 603-927-4266 Fax 603-927-4694

### Description of Calibrated Items

One white diffuse reflectance sample, model SRT-99-050 Reflectance Target, serial number 99AA10-0614-9856

### Calibration

8°/Hemispherical (8/h) Spectral Reflectance Calibration over the range 250 – 2500nm, reported at 50nm intervals

### Description of Calibration

The calibration items are measured using a Perkin Elmer Lambda 900 or Lambda 950 dual beam spectrophotometer with Labsphere PELA-1000 integrating sphere reflectance accessories which perform the 8/h reflectance measurements. The instrument used for the measurement of the calibrated items is identified with an X in Table I below.

Table I Measurement Instrument and Accessory

Spectrophotometer	Lambda 900A S/N 101N 3060902	Lambda 900B, S/N: 70038	Lambda 950C SN: 950N9031801
Accessory	PELA-1000 Serial No. 05220314341	PELA-1000 Serial No. 08220625262	PELA-1000 Serial No. 1203088776
Instrument Used	X		

The integrating sphere diameter is 150 mm, the sample port diameter is 25 mm, and the coating material is Spectralon. The reference standard and sample piece are placed sequentially in the sample port of the sphere. The reference beam of the instrument acts as an auxiliary beam to correct for substitution error in the sphere. A primary reference standard, calibrated by NIST is the basis of the calibration. Secondary standards are made from the primary standard and are used as the reference standard for the calibration. The calibration of the instrument is performed with pressed polytetrafluoroethylene (PTFE) reference standard<sup>1</sup>. The standards are listed below in the Table II. The primary standard is the same for all calibrations and is indicated in the table as well.

Table II. Reflectance Reference Standards

Primary Standard	REFL-01 7A21H-2267, NIST Report No. 685/281167-11, Calibrated August 18, 2011		
Secondary Standard	REFL-06-W 6	REFL-07-W 7	REFL-08-W 8
	S/N 99AA01-0214-8729	S/N 99AA01-0214-8730	S/N 99AA01-0214-8731
	Calibrated 1st of every month	Calibrated 1st of every month	Calibrated 1st of every month
Secondary Standard Referenced			X

Uncertainty values take into account the uncertainties generated by this transfer from master to working standards. The 8/h spectral reflectance of the sample  $R_s$  at each wavelength  $\lambda$  was calculated from:

Eq. 1 
$$R_s = \frac{M_s - M_H}{1 - M_H} \cdot R_R$$

where:

$M_s$  is the instrument's relative measurement of the reflectance of the calibration item

$M_H$  is the instrument's relative measurement of the reflectance of an open sample port and measures the effect of stray light overfilling the sample port

$R_R$  is the 8/h spectral reflectance of the reference standard.

<sup>1</sup>Wiedner V.R., and Hsia, J. J. "Reflection Properties of Pressed Polytetrafluoroethylene Powder", J.Opt.Soc.Am., Vol71, 1981, pp856-861



## 8°/Hemispherical Reflectance Calibration Certificate

**SERIAL NUMBER:** 99AA10-0614-9856

**DATE OF REPORT:** August 25, 2014

**PAGE 2 OF 6**

The instrument's relative measured reflectances,  $M_S$  and  $M_H$  are equivalent to:

Eq. 2 
$$M_S = \frac{S_S}{S_R}$$

and

Eq. 3 
$$M_H = \frac{S_H}{S_R}$$

where:

$S_R$  is the signal with the reference standard in place

$S_S$  is the signal with the calibration item in place

$S_H$  is the signal with an open port

Therefore, the 8/h spectral reflectance of the calibration item  $R_S$  at each wavelength  $\lambda$  is equivalent to:

Eq. 4 
$$R_S = \frac{S_S - S_H}{S_R - S_H} \cdot R_R$$

The final 8/h spectral reflectance is obtained by averaging the values from three scans.

Sources of uncertainty are:

- the 8/h spectral reflectance of the pressed PTFE reference standard
- nonuniformity across surfaces of the reference standard
- nonuniformity across the surface of the calibration item
- residual uncertainty in the correction for extraneous light
- nonlinearity of the instrument
- effect of wavelength error
- random noise in the measurements

The individual contributions to uncertainty are combined by adding in quadrature (root-sum-square). The quadrature sum is multiplied by a coverage factor ( $k$ ) to generate the expanded uncertainty. The coverage factor is chosen to provide a confidence level of 95%. For effective degrees of freedom of thirty or greater,  $k$  is 2.0. For lower degrees of freedom a larger coverage factor is used representing the coverage factor necessary to provide a confidence level of 95% for a t-distribution with the corresponding degrees of freedom. Reference Table IV for the expended uncertainty of the calibration results.

Calibrated by:

Title: Optical Calibration Technician

Approved by:

### General Information

1. The values in Table III apply only to the central 14 mm by 6 mm area of the items for measurement geometry 8/h.
2. The report of calibration may not be reproduced except in full without the written consent of this laboratory.
3. This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government.

## 8°/Hemispherical Reflectance Calibration Certificate

SERIAL NUMBER: 99AA10-0614-9856

DATE OF REPORT: August 25, 2014

PAGE 3 OF 6

Table III. 8/h spectral reflectance  $R_s$  as a function of wavelength  $\lambda$

Wavelength (nm)	Reflectance
250	0.9587
300	0.9761
350	0.9847
400	0.9881
450	0.9875
500	0.9888
550	0.9883
600	0.9883
650	0.9879
700	0.9886
750	0.9894
800	0.9887
850	0.9890
900	0.9894
950	0.9882
1000	0.9883
1050	0.9887
1100	0.9886
1150	0.9884
1200	0.9875
1250	0.9875
1300	0.9863
1350	0.9855

Wavelength (nm)	Reflectance
1400	0.9847
1450	0.9855
1500	0.9859
1550	0.9856
1600	0.9860
1650	0.9838
1700	0.9825
1750	0.9830
1800	0.9825
1850	0.9776
1900	0.9765
1950	0.9747
2000	0.9692
2050	0.9569
2100	0.9485
2150	0.9460
2200	0.9608
2250	0.962
2300	0.956
2350	0.946
2400	0.941
2450	0.937
2500	0.941


## 8°/Hemispherical Reflectance Calibration Certificate

SERIAL NUMBER: 99AA10-0614-9856

DATE OF REPORT: August 25, 2014

PAGE 4 OF 6

Graph I. 8/h spectral reflectance  $R_s$  as a function of wavelength



## 8°/Hemispherical Reflectance Calibration Certificate

**SERIAL NUMBER:** 99AA10-0614-9856

**DATE OF REPORT:** August 25, 2014

**PAGE 5 OF 6**

Table IV Uncertainty contributions and expanded uncertainty ( $k=2$ ) of the 8/h spectral reflectance for the measured sample

### CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <small>Note 3,5</small>	Remarks
<b>OPTICAL RADIATION</b>			
<b>Photometric (20/O02)</b>			
Relative Reflectance at Wavelength Shown Below: 250 nm to 600 nm	up to 0.02 > 0.02 to 0.05 > 0.05 to 0.10 > 0.10 to 0.20 > 0.20 to 0.50 > 0.50 to 0.80 > 0.80 to 0.99	0.0016 0.0029 0.012 0.012 0.0054 0.0054 0.0053	Relative reflectance is a dimensionless quantity
601 nm to 1500 nm	up to 0.02 > 0.02 to 0.05 > 0.05 to 0.10 > 0.10 to 0.20 > 0.20 to 0.50 > 0.50 to 0.80 > 0.80 to 0.99	0.0017 0.0022 0.0025 0.0052 0.0064 0.0064 0.0049	
1501 nm to 2200 nm	up to 0.02 > 0.02 to 0.05 > 0.05 to 0.10 > 0.10 to 0.20 > 0.20 to 0.50 > 0.50 to 0.80 > 0.80 to 0.99	0.0090 0.0090 0.015 0.015 0.0099 0.0083 0.0088	
2201 nm to 2500 nm	up to 0.02 > 0.02 to 0.05 > 0.05 to 0.10 > 0.10 to 0.20 > 0.20 to 0.50 > 0.50 to 0.80 > 0.80 to 0.99	0.054 0.054 0.043 0.043 0.035 0.028 0.032	
<b>END</b>			

### Notes

## 8°/Hemispherical Reflectance Calibration Certificate

SERIAL NUMBER: 99AA10-0614-9856

DATE OF REPORT: August 25, 2014

PAGE 6 OF 6

### Notes

**Note 1:** A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

**Note 2:** Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

**Note 3:** The uncertainty associated with a measurement in a CMC is an expanded uncertainty using a coverage factor,  $k = 2$ , with a level of confidence of approximately 95 %. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

**Note 3a:** The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

**Note 3b:** As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

**Note 3c:** As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under *normal conditions*. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.1.h. of NIST Handbook 150, Procedures and General Requirements.

**Note 4:** Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)

**Note 5:** Values listed with percent (%) are percent of reading or generated value unless otherwise noted.

**Note 6:** NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories.

**Note 7:** See [NIST Handbook 150](#) for further explanation of these notes.