

Sr No	Item No	Technical specification and scope of job
1	Examination Table	POWDER COATED WITH TABLE TOP ADJUSTMENT UPTO 40 DEGREE. 4 LEGS, FRAME MATERIAL MS WITH HEAD ADJUSTABLE.
2	ECG Machine	ITEM DESCRIPTION: ECG MACHINE; TYPE:DIGITAL DISPLAY; RANGE:50 HZ; DIMENSION:N/A; MODEL:N/A; 1. SIMULTANEOUS 12 CHANNEL ECG RECORDING WITH 12 LEAD SIMULTANEOUS ACQUISITION. 2. SHOULD HAVE VISUAL ALARM FOR OPEN LEAD. 3. SHOULD HAVE DIGITAL DISPLAY OF 12 CHANNEL ECG. 4. QWERTY ALPHANUMERIC KEYBOARD. 5. BUILT IN ECG PARAMETERS MEASUREMENTS AND INTERPRETATION 6. MINIMUM 40 ECG STORAGE INBUILT MEMORY. 7. 3 OPERATING MODES: AUTOMATIC, MANUAL AND RHYTHM 8. SHOULD HAVE A MAINTENANCE FREE DIGITAL THERMAL ARRAY PRINTER 9. PRINTER SHOULD WORK WITH STANDARD THERMAL PAPER (SHOULD BE AVAILABLE IN LOCAL MARKET) 10. SHOULD HAVE 12 LEAD ECG PREVIEW DISPLAY BEFORE TAKING PRINTOUTS AND SHOULD HAVE PRINTER ON/OFF SELECTION. 11. SHOULD HAVE ECG LEAD ANNOTATION FACILITY. 12. MACHINE SHOULD HAVE SUFFICIENT BATTERY BACKUP FOR TAKING AT LEAST 25 NOS. ECG ON A FULLY CHARGED BATTERY. 13. SHOULD SUPPLIED WITH TWO PATIENT CABLE SETS 8 CLIP ON ELECTRODES, 12 CHEST ELECTRODES WITH SILICON RUBBER BULB, 12 PACKETS OF RECORDING PAPER, 1 BOTTLES OF JELLY AND 12 NOS. REUSABLE BUTTON TYPE ELECTRODE. 14. SHOULD OPERATE ON MAINS (220V - 50 HZ) AND RECHARGEABLE BATTERY. 15. RECORDING SPEED SHOULD BE 25 MM PER SECOND AND 50 MM PER SECOND. 16. SHOULD HAVE DEFIBRILLATION PROTECTION. 17. CMRR SHOULD BE > 90DB OR ECG MACHINE SHOULD HAVE DIGITAL PROCESSING WITH ATLEAST 7000 SAMPLES PER SECOND FROM EACH LEAD WIRE. 18.FREQUENCY RESPONSE 0.05 HZ TO 150 HZ. 19. SHOULD HAVE A DIGITAL FILTER FOR AC AND EMG. 20. SHOULD HAVE SAFETY CERTIFICATE FROM A COMPETENT AUTHORITY CE/FDA (US)/STQC CB CERTIFICATE/STQC S CERTIFICATE OR VALID DETAILED ELECTRICAL AND FUNCTIONAL SAFETY TEST REPORT FORM ERTL. COPY OF THE CERTIFICATE/TEST REPORT SHALL BE PRODUCED ALONG WITH THE TECHNICAL BID
3	Crash Cart Trolley	ITEM NAME:CRASH CART TROLLEY; ISO 13485 CERTIFIED, MATERIAL OF CRASH CART CONSOLE ABS THERMOPLASTIC, OVER BRIDGE (NOS) 6 WIDTH OF TROLLEY (MM) 800 HEIGHT OF TROLLEY (MM) 1100 DEPTH OF TROLLEY (MM) 700 CASTERS 4 NOS CASTERS OF 4 TO 5 INCHES DIAMETER FOR EASY ROLLING AND WITH FOOT BRAKES CENTRAL LOCKING WITH ALL THE DRAWERS, USE OF DRAWERS TO KEEP DRUGS, RESUSCITATION INSTRUMENTS, IV FLUIDS, AIRWAYS DEVICES AS PER ITEM NAME:CRASH CART TROLLEY; ISO 13485 CERTIFIED, MATERIAL OF CRASH CART CONSOLE ABS THERMOPLASTIC, OVER BRIDGE (NOS) 6 WIDTH OF TROLLEY (MM) 800 HEIGHT OF TROLLEY (MM) 1100 DEPTH OF TROLLEY (MM) 700 CASTERS 4 NOS CASTERS OF 4 TO 5 INCHES DIAMETER FOR EASY ROLLING AND WITH FOOT BRAKES CENTRAL LOCKING WITH ALL THE DRAWERS, USE OF DRAWERS TO KEEP DRUGS, RESUSCITATION INSTRUMENTS, IV FLUIDS, AIRWAYS DEVICES AS PER ACLS GUIDELINES NUMBER OF DRAWERS (NOS) 6 DRAWERS OF DIFFERENT HEIGHT, UPPER THREE DRAWERS SHALL HAVE 25 TO 30 COMPARTMENTS EACH FOR STORING DRUGS, LENGTH OF SIDE STORAGE 322 MM DEPTH OF BET 380 MM, IV POLE WITH CLAMPS, PULL-OUT WRITING SURFACE TOP, SIDE STORAGE FOR KEEPING CAN, STORAGE BINS, GLOVES, DISPENSER, SHARP CONTAINER SET, A-TYPE OXYGEN CYLINDER, SHELF FOR DEFIBRILLATOR AND MONITOR WITH STRAPS AND FOR OXYGEN CYLINDER HOLDER, NUMBER OF UTILITY TRAYS IN OVER BRIDGE (NOS) 3 SHALL HAVE A OVER BRIDGE OF ABS THERMOPLASTIC AND UTILITY TRAYS FOR KEEPING CONSUMABLES , HANDLE ERGONOMIC HANDLE WITH EASY GRIP HEIGHT OF OVER BRIDGE FROM THE TROLLEY (MM) 1295, MEDICAL EQUIPMENT
4	Oxygen Cylinder Stand cum Trolley	CYLINDER STAND CUM TROLLEY FOR HOLDING OXYGEN CYLINDER; MATERIAL: IRON OR MILD STEEL (MS); LOAD CAPACITY: 50 KILOGRAMS; NUMBER OF WHEELS: 04; SHAPE: RECTANGLE; LENGTH:

Sr No	Item No	Technical specification and scope of job
5	Examination Bed	DESCRIPTION => ITEM DESCRIPTION:HOSPITAL BED; MATERIAL:N/A; TYPE:SEMI FOWLER; SIZE:600 MM; MANUFACTURER: MDT/SI SURGICAL/MIDMARK OR EQUIVALENT NUMBER OF BED PLATFORM SECTIONS: 2, AVAILABILITY OF BACKREST SECTION: YES AVAILABILITY OF LEG SECTION: YES ADJUSTMENTS: BACKREST THROUGH TWO SCREW SYSTEMS WITH THRUST BEARINGS INDIVIDUALLY MANOEUVRED BY HANDLE FOR SEMI FOWLER BED, SALINE ROD PROVISION: FOUR LOCATIONS ON THE BED PLATFORM TO HOLD STAINLESS STEEL TELESCOPIC SALINE ROD 12MM DIAMETER, NUMBER OF HOOKS IN TELESCOPIC I V POLE: 2, URINE BAG HOLDER: YES, MOULDED CHART HOLDER: YES, OVERALL LENGTH OF BED: 2050 TO 2100 +/- 5%, MATTRESS PLATFORM SIZE: LG: 1970 MM X WD: 790 MM +/- 5%, RAISED BACKREST ANGLE: 65 DEG
6	Pulse Oximeter	TYPE OF PULSE OXIMETER HAND HELD SHOULD WORK DURING MOTION AND VERY LOW PERFUSION CONDITIONS, SPECIAL SIGNAL PROCESSING TECHNOLOGY (MOTION ARTEFACT REJECTION) NELLORE COMPATIBLE SPO2 PROBES REUSABLE OXYGEN SATURATION RANGE 0-100% PULSE RATE 25-240 BPM DISPLAY OF SPO2 AND PR DIGITAL DISPLAY WITH BAR GRAPH ACCURACY $\pm 5\%$ TYPE OF PATIENT ADULT, PAEDIATRIC, NEONATAL SHOULD BE SHOCKPROOF , DISPLAY (WITH CONTRAST ADJUST-ABILITY) LCD , SATURATION DISPLAYED ON THE SCREEN , PULSE RATE DISPLAYED ON THE SCREEN , PERFUSION INDEX DISPLAYED ON THE SCREEN NO STATUS OF BATTERY CHARGING DISPLAYED ON THE SCREEN , SENSOR OFF /ON STATUS , AUDIBLE AND VISUAL ALARMS , ALARM OVERRIDE FACILITY SHOULD BE PRESENT , TRENDS 48 HRS TYPE OF TREND DISPLAY TABULAR DATA INTERVAL IN SEC 0 POWER SOURCE :BATTERY OPERATED WITH EXTERNAL CHARGER AUTOMATIC SWITCH FROM MAINS TO BATTERY IN CASE OF POWER FAILURE , OPERATING TEMPERATURE "AMBIENT TEMPERATURE OF 0 TO 40 DEG C AND RELATIVE HUMIDITY OF 15 TO 90%" WEIGHT OF THE MACHINE IN KG 0.175 PROTECTIVE SPLASH PROOF CASE FOR CLEAN AND STROGAGE AND SAFE TRANSPORT TO BE PROVIDED. DISPLAY MUST ALLOW EASY VIEWING IN ALL AMBIENT LIGHT LEVELS ,

Sr No	Item No	Technical specification and scope of job
7	Stethoscope	LENGTH OF Y TUBING (CM) 60 CENTIMETER RECOMMENDED APPLICATION PHYSICAL ASSESSMENT AND DIAGNOSIS DIAMETER OF THE BINAURAL TUBE 6 MM DIAPHRAGM TYPE FIXED CHEST PIECE MATERIAL STAINLESS STEEL MATERIAL OF SNAP ON RING PLASTIC LATEX FREE PRODUCT , NOISE ISOLATION EARPIECE , NON CHILLING SNAP ON RING PROVIDED ON CHEST PIECE , NUMBER OF SNAP ON RING PROVIDED 2 BINAURAL SPRING PROVIDED , WEIGHT OF THE PRODUCT (G) 350 DIAMETER OF CHEST PIECE(MM) 50 , NUMBER OF SPARE DIAPHRAGM PROVIDED 1 TYPE OF STETHOSCOPE ADULT STETHOSCOPES BINAURAL TUBE MATERIAL STAINLESS STEEL DIAMETER OF CHEST PIECE; DIAPHRAGM SIDE (MM) 50 TUNABLE DIAPHRAGM , FINISH OF BINAURAL TUBE CHROME PLATED OVER NICKLE CONFORMING TO GRADE C OF IS:1068-1958 OUTER DIAMETER OF Y TUBING (MM) 40 TYPE OF CHEST PIECE DIAPHRAGM DETECTABLE SOUND FREQUENCY RANGE (HZ) 900 EAR PIECES MATERIAL PVC NUMBER OF ID TAG PROVIDED 1 MATERIAL OF THE SPRING N/A OVER ALL DIMENSION; LENGTHXHEIGHTXWIDTH (MMXMMXMM) NOT AVAILABLE SPARE SNAP ON RING PROVIDED , STETHOSCOPE DESIGN AND MANUFACTURING COFIRMS TO IS 3391-1965 LATEST AMENDMENT , MATERIAL OF Y TUBING LATEX FREE PVC FINISH OF CHEST PIECE BRASS ANODIZED. DIAPHRAGM MATERIAL : PLASTIC, SPARE EAR PIECE TO BE PROVIDED. TYPE OF Y TUBING SINGLE LUMEN TEST REPORT NUMBER TO BE PROVIDED. A SET OF TWO SPARE EAR PIECE TO BE PROVIDED
8	Spine Board	PURPOSE- THE SPINAL BOARD IS A DEVICE FOR RECOVERY IMMOBILIZATION , AND TRANSFER TO THE STRETCHER OF PATIENT WITH SUSPECTED SPINAL INJURIES LENGTH IN MM WITH TOLERANCE ± 5 MM-1830 MM WIDTH IN MM WITH TOLERANCE ± 5 MM-420 MM WARRANTY IN YEARS-1 MATERIAL-PE OEM OF OFFERED PRODUCT IS ISO 9001:2015 CERTIFIED.
9	Double Bottle Suction Machine	MATERIAL OF CABINET (BASE, TOP AND PANEL) EPOXY POWDER COATED MILD STEEL (MS) RUST PROOF HOUSING CABINET AND CASTOR WHEELS , THE UNIT IS MOUNTED IN SUCH A WAY AS TO PRODUCE MINIMUM VIBRATIONS , MATERIAL OF TWIN JARS/BOTTLES GLASS SUCTION BOTTLES/JARS WITH SYNTHETIC RUBBER LID AND FITTED WITH ARRANGEMENT TO PREVENT OVERFLOW OF FLUID TO OUTSIDE , CAPACITY OF EACH SUCTION BOTTLE/JAR GRADUATED (ML) 2000 INTERCONNECTED JAR/BOTTLE WITH OVERFLOW PROTECTION DEVICE , DOUBLE OVERFLOW PROTECTION WITH SAFETY DEVICE PROVIDED , NON-COLLAPSIBLE SUCTION TUBING ID (MM) 8 MATERIAL OF SUCTION TUBE SILICON MINIMUM LENGTH OF SUCTION TUBE (MTR) 2 FLUTTER FREE VACUUM CONTROL KNOB PROVIDED , BACTERIAL SUCTION FILTER/AIR FILTERS AND WATER TRAPS BETWEEN THE SUCTION BOTTLE AND SUCTION PUMP PROVIDED , MATERIAL OF TRAY STAINLESS STEEL (SS) VACUUM GAUGE (BOURDON TUBE TYPE)

Sr No	Item No	Technical specification and scope of job
10	AMBU Bag	SUPPLY OF ADULT AMBU TYPE BAG- 1X ARTIFICIAL RESUSCITATOR, AMBU TYPE BAG - 1X NON-REBREATHING VALVE WITH PRESSURE RELEASE – 1X TRANSLUCENT FACE MASK (SIZE 4) 1X OXYGEN RESERVOIR WITH VALVE - 1X OXYGEN (O2) TUBE, 1.5 MTR. LONG (APPROX) 1X 360° SWIVELLING PATIENT CONNECTOR CAPACITY: AMBU TYPE BAG CAPACITY 1600 ML/2000 ML APPROX. FOLDING SILICONS BELOW -TUBE DIAMETER: STANDARD 15 MM INSIDE & 22 MM OUTSIDE DIAMETER MATERIAL: MADE FROM 100% LATEX FREE, SILICON MATERIAL (AUTOCLAVABLE) PRODUCT CONFORMS TO ISO 9001, CE-QUALITY MARK, ISO 13485 AND OTHER QUALITY STANDARDS
11	Medical Oxygen Cylinders	PRODUCT INFORMATION PRODUCT DESCRIPTION MEDICAL OXYGEN GAS CYLINDERS - STEEL (NORMALIZED) PURPOSE FILLING OF COMPRESSED MEDICAL OXYGEN (I.P) GAS FOR HOSPITAL PURPOSES CYLINDER TECHNICAL SPECIFICATION CONFORMITY TO STANDARD IS: 7285 (PART 1)/2004 (LATEST) PRODUCT TO BE ISI MARKED. FURTHER TO SUBMIT BIS LICENSE NO (CM/L NO) OF MANUFACTURER OR PESO APPROVED THIRD PARTY INSPECTION AGENCY LICENSE/CERTIFICATION NUMBER. MATERIAL OF CYLINDER BODY SHALL BE-NORMALIZED STEEL HIGH PRESSURE SEAMLESS COLOUR CODING OF CYLINDERS BODY - BLACK, SHOULDER - WHITE CYLINDER SIZE-B WATER CAPACITY OF CYLINDER (LITRES) 10 LITER HYDROSTATIC TEST PRESSURE (KG/CM2)-250 MAXIMUM WORKING PRESSURE AT NTP (NORMAL TEMPERATURE AND PRESSURE) (KG/CM2)-150 FILLING PRESSURE OF CYLINDER AT NTP (NORMAL TEMPERATURE AND PRESSURE)(KG/CM2)-150+/-5% OUTER DIAMETER, APPROX, (MM)-140 MILLIMETER MINIMUM WALL THICKNESS (MM)-4 MILLIMETER MINIMUM CYLINDER LENGTH (MM)-830 MILLIMETER TARE WEIGHT (KG)-14 KILOGRAM NECK RING PROVIDED-YES FOOT RING PROVIDED-NO TYPE OF VALVE FITTED WITH THE CYLINDER-TYPE B (KEY OPERATED) VALVE PROTECTION CAPS AVAILABLE-YES ISI MARKED VALVE CONFORMING TO IS 3224-2002 (LATEST) FITTED WITH THE CYLINDER-YES ALL FITTINGS SHOULD BE LEAK PROOF-YES STATUTORY REQUIREMENTS CYLINDER CONFORMS TO REQUIREMENTS PRESCRIBED IN GAS CYLINDER RULES, 2016 OF THE GOVT OF INDIA, WITH SUCH MODIFICATIONS AND AMENDMENTS AS MAY BE ORDERED/ISSUED FROM TIME TO TIME BY THE CHIEF CONTROLLER OF EXPLOSIVES, DEPT OF PETROLEUM, GOVT OF INDIA-YES ALL CYLINDERS MUST HAVE FILLING PERMISSION OR MUST HAVE HST (HYDRO STRETCH TEST) CERTIFICATION FROM TEST SHOP APPROVED BY CHIEF CONTROLLER OF EXPLOSIVES, DEPARTMENT OF PETROLEUM, GOVERNMENT OF INDIA YES TESTING OF CYLINDER AS PER STANDARD PROTOCOL IS THE RESPONSIBILITY OF CYLINDERS SUPPLIER-YES PACKING & MARKING EACH CYLINDER SHOULD BE ENGRAVED THE DETAILS VIZ (A) SERIAL NO OF CYLINDER/IDENTIFICATION (B) CODE OF MEG (C) WORKING PRESSURE
12	Emergency Shower and Eye Washer	CERTIFICATION-COMPLIANCE TO IS 10592 GENERAL CHARACTERISTICS- WATER PRESSURE IN PSI-30 STATION TYPE-EMERGENCY SHOWER, EYE WASHER EMERGENCY SHOWER SPECS DIAMETER OF SHOWER HEAD IN INCHES-10 MOUNTING-SURFACE MOUNTED FLOW RATE IN L/M-70 - 75 MATERIAL USED-ABS HEIGHT OF SHOWER IN MM-2100 – 2200 OPERATING MODE-HAND OPERATED EYE WASHER SPECS FLOW RATE OF EYW WASH IN L/M-9 - 11 HEIGHT OF EYE WASHER IN MM-800 - 1100 EYE WASHER OPERATING MODE- HAND OPERATED EYE WASHER MOUNTING SURFACE MOUNTED DIAMETER OF EYE WASH BOWL IN INCHES-10 MATERIAL-ABS