

बिड दस्तावेज़ / Bid Document

बिड विवरण / Bid Details	
बिड बंद होने की तारीख/समय / Bid End Date/Time	06-07-2026 11:00:00
बिड खुलने की तारीख/समय / Bid Opening Date/Time	06-07-2026 11:30:00
बिड पेशकश वैधता (बंद होने की तारीख से) / Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम / Ministry/State Name	Chhattisgarh
विभाग का नाम / Department Name	Health And Family Welfare Department Chhattisgarh
संगठन का नाम / Organisation Name	Chhattisgarh Medical Services Corporation Limited
कार्यालय का नाम / Office Name	Chhattisgarh Medical Services Corporation Limited
शिकायत निवारण के संपर्क विवरण / Contact details of Grievance redressal	HOD Email id : Buyer Email id: bme01cgmscl@gmail.com
कुल मात्रा / Total Quantity	589

बिड विवरण/Bid Details	
वस्तु श्रेणी /Item Category	Dental Chair (Q2) , Dental RadioVisioGraphy (RVG) Unit (V2) (Q2) , Dental X - Ray Machine (V2) (Q2) , Stereo microscope , High intensity light curing unit , Ultrasonic scaler , Conscious sedation unit , Hydrogen soldering unit , Spot welder with soldering attachment , Lab micro motor , Model trimmer double disc , Model Trimmer with Diamond Disc , High Speed lathe , Sand blaster micro macro , Autoclave-front loading , Ultrasonic cleaner capacity 3.5lts Type III , Plaster Dispenser , Vibrator , Acrylizer , Pressure molding machine , Induction Casting Machine with Vacuum pump , Preheating furnace , Die cutting machine , Endomotor , Physiodispensor , Implant Kit , Cryo Surgery Unit , Electro surgery unit , Peizo-surgery unit , Centrifuge for PRP PRF preparation , Electro surgical Cautery , Apex Locator , OPG radiography machine with Cephalostat , Cone Beam Computed Tomography CBCT , Dental Phantom Head Table , Dental Diode laser system with accessories , Automatic Intraoral film processor , Dental Operating Microscope , Dental Clinical Micro motor , Dental magnifying loupes with LED headlight , Glass bead sterilizer , Pulp Tester , Injectable GP condenser , Tissue floatation water bath , Biochemistry analyser , Smart dentin grinder for hospital, 300W- 500W , Tracheostomy kit , Cleft Surgery kit , Trauma kit with bone plating kit , Pedo extraction forceps sets , Rubber dam kit for pedo patients , Formalin chamber , Typhodonts , Adjustable mouth prop set , Double beak pliers anterior and posterior , Bracket Holding Pliers , Straight Hoe Pliers , Curved Hoe Pliers , Papoose Board for stabilization of childrens , Articulators Semi adjustable , Airtor Handpieces , Straight Handpiece for clinical micromotor , Supra gingival scalers , Sub gingival scaler , Surgery Trolleys
GeMARPTS में खोजी गई स्ट्रिंग्स / Searched Strings used in GeMARPTS	Stereo microscope, High intensity light curing unit, Ultrasonic scaler, Conscious sedation unit, Hydrogen soldering unit, Spot welder with soldering attachment, Lab micro motor, Model trimmer (double disc), Model Trimmer with Diamond Disc, High Speed lathe, Sand blaster (micro & macro), Autoclave- front loading, Ultrasonic cleaner capacity 3.5lts Type III, Plaster Dispenser, Vibrator, Acrylizer, Pressure molding machine, Induction Casting Machine with Vacuum pump, Preheating furnace, Die cutting machine, Endomotor, Physiodispensor, Implant Kit, Cryo Surgery Unit, Electro surgery unit, Peizo-surgery unit, Centrifuge for PRP/PRF preparation, Electro surgical Cautery, Apex Locator, OPG radiography machine with Cephalostat, Cone Beam Computed Tomography (CBCT) , Dental Phantom Head Table, Dental Diode laser system with accessories, Automatic Intraoral film processor, Dental Operating Microscope, Dental Clinical Micro motor, Dental magnifying loupes with LED headlight, Glass bead sterilizer, Pulp Tester, Injectable GP condenser, Tissue floatation water bath, Biochemistry analyser, Smart dentin grinder for hospital, 300W- 500W, Tracheostomy kit, Cleft Surgery kit, Trauma kit with bone plating kit, Pedo extraction forceps sets, Rubber dam kit for pedo patients, Formalin chamber, Typhodonts, Adjustable mouth prop set, Double beak pliers anterior and posterior, Bracket Holding Pliers, Straight Hoe Pliers, Curved Hoe Pliers, Papoose Board for stabilization of childrens, Articulators Semi adjustable, Airtor Handpiece s, Straight Handpiece for clinical micromotor, Supra gingival scalers, Sub gingival scaler, Surgery Trolleys

बिड विवरण/Bid Details	Searched String: Stereo microscope
	Stereo Microscope (V1), Microscopes - Pathological and Research as per IS 4381, IS 5204, IS 4381, IS 5204, Metallurgical Microscope, Advance Polarization Microscope, Inverted Research Microscope as per IS 5204, Ophthalmic Operating Microscope (V2), Ophthalmic Operating Microscopes, Microscope immersion oil, Surgical Operating ENT Microscope, Deca Head Microscope Searched String: High intensity light curing unit Dental Curing light gun, High Mast Lighting Tower for large area with LED Flood Lighting System, Emergency Light Unit (V2), High Frequency Radio Surgery Unit, LED Machine Work Lamp, High Flow Nasal Oxygen Therapy Unit (V2), High Mast Lighting Octagonal Tower with Solar Based LED Street Light, High Mast Lighting Octagonal Tower With Solar Based LED Street Light (V2), Lead Acid Storage Batteries For Motor Vehicles with Light Weight and High Cranking Performance (V2) Conforming to IS 14257, High Mast Lighting Octagonal Tower for Small area with LED Flood Lighting System (V2) Searched String: Ultrasonic scaler Ultrasonic Cleaners, Ultrasonic Probe Sonicator, Ultrasonic Sensor, Ultrasonic Interferometer, Ultrasonic Thickness Gauge, Nebulizer (V2), Ultrasonic Level Transmitter, Digital Ultrasonic Flaw Detector, Echo Cardiography Scanning Machine (V2), Ultrasonic pulse velocity tester Searched String: Conscious sedation unit Cryotherapy Unit, Sofa (V3), PHOTOTHERAPY Unit, Alarm Annunciator unit, Executive Table (V4), Driver Display Unit, Thyristor Power Unit, Metal Storage Cabinet, Ophthalmic Chair Unit, Argon Plasma Coagulator with ESU Searched String: Hydrogen soldering unit Fume extraction unit for filtration of soldering fumes, Sofa (V3), Electric Soldering Iron, Soldering Clamp Holder, Executive Table (V4), SMD Rework Station, Metal Storage Cabinet, Temperature controlled soldering station, Computer Table (V3), Hydrogen Gas as per IS 1090 Searched String: Spot welder with soldering attachment Resistance Welding Equipment - Rocker Arm Spot Welding Machine (V2), Hand Held Brush Cutter (V2), Soldering and Desoldering Station with digital display, Dried Blood Spot (DBS) Card, Network Attached Storage (V2), single texture white rubberized waterproof nylon fabric rain capes with attached hood, Temperature controlled soldering station, Lumbo Sacral Belt Kit (Size - Large) with Packing Box (Alimco), Steel Porta Cabin, Basketball Ring Searched String: Lab micro motor MOCHAI (LAB-LAB) Seeds, Ultra low temperature laboratory deep freezer, Dental Lab Micromotor (V2), Elisa Reader (ICMR), Micro/Semi-Micro/Analytical weighing balance, water baths, Wooden Lab Stools, Drone or Unmanned Aerial Vehicle with Customized Payload (Defence Forces - Navy, Army and Air Force), Lab Coating Machine, vortex mixers Searched String: Model trimmer (double disc)

बिड विवरण / Bid Details	Double Break Disconnectors, Agriculture Harrow (Disc Type), Double Side Lapping and Polishing Machine, Optical Storage Media, Disc polisher machine
	Searched String: Model Trimmer with Diamond Disc Manual Operated 66 kV, 800 A Isolators and with Solid Core Insulators (PSPCL), HEDGE TRIMMER or HEDGE CUTTER, Switch Disconnectors and Fuse Combination Units (Switch Fuse Unit) Conforming To IS/IEC 60947 (Part 3), Hand Held Brush Cutter (V2), Weightlifting Discs, Anatomy Structure Model (V3), Micro/Semi-Micro/Analytical weighing balance, Medical Training kit with Human mannequins for medical education or training, White LED Based Solar Home Lighting System with Batten Type Luminary as per MNRE Specifications, Anatomy Structure Model (V2) Searched String: High Speed lathe High speed stirrer, High End Desktop Computer, High speed tripping relay, High Speed Drill System for Neurosurgery & Spinal Surgery, High Steel Speed Cutters, Revolving Chair (V5), Lathe Bit, Lathe Machine as per IS 6893:1988, High Vacuum Suction Machine (V2), LED Luminaire for High Bay Lighting Searched String: Sand blaster (micro & macro) Category not available on GeM for the text string uploaded by the buyer Searched String: Autoclave- front loading Dental Autoclave, Autoclave (ICMR Specs), Vertical Autoclave, Table Top Autoclave, Autoclave (Horizontal), Pin Up Notice Board (V2), Horizontal Autoclave, Vertical Autoclave (FSSAI), DNA Gel loading dye, Cloth Dryer Stand Searched String: Ultrasonic cleaner capacity 3.5lts Type III Electronic Weighing Scales, Buses (V2), Ultrasonic Cleaners, Buses for State Road Transport (V2), Recirculating chiller, Gasket Sheet-IS:11149, Split Air Conditioner, Wall Mount Type (V3) ISI Marked to IS 1391 (Part 2), Split Air Conditioner Including Green AC, Wall Mount Type (V2), Welding Electrode - Covered Arc Welding Electrode - Rutile Type (BEML), Off - Grid Solar Photovoltaic (PV) Power Plant (V2) Searched String: Plaster Dispenser Soup Dispenser, Napkin Dispenser, Gypsum Building Plaster - Excluding Premixed as per IS 2547, Gypsum Building Plaster - Premixed Lightweight Plasters as per IS 2547, Touchless Sanitizer Dispenser, Plaster of Paris Bandage Conforming to IS 4738, Plaster Cutting Machine (V2), Bottle top dispenser, Non Carbonated Beverage Dispenser (V2), Touchless Soap Dispenser (V3) Searched String: Vibrator Centring Work Kit (Gujarat Govt), Nebulizer (V2), Vibratory Compactor (Soil Compactor and Tandem Roller), Vibro sifter, LED Luminaire for Road and Street Lights (V2) Conforming to IS 10322 (Part 5 / Section 3), LED Luminaire for Floodlight (V2) Conforming to IS 10322 (Part 5/Section 5), LED Luminaire For Road And Street Lights Conforming To IS 10322 (Part 5/Section 3) (V3) (Under BIS Scheme - II), Anti

बिड विवरण / Bid Details	Vibration Table, Whole Body Vibration Unit (Machine), LED Luminaire For Floodlight Conforming To IS 10322 (Part 5/Section 5) (V3) (Under BIS Scheme - II)
	Searched String: Acrylizer Table Top Elevator Writing Desk, Electronic Baby Weighing Scale - RBSK, Basketball backstop, Polymethyl Methacrylate (PMMA) (Acrylic) Sheets (V2) as per IS 14753 (Latest), Appointment Stand, Handloom Blanket (V2), Information Boards or Illustration Board (V2), Exam Pad (V2), External branding signage boards, Signage Board (V3) Searched String: Pressure molding machine Wax Molding Tray Set, Digital Blood Pressure Monitor, Tool Kit for Plastic Product - ODOP UP Gov. (Kanpur Village), CPAP Machine (V2), Sanitary Napkins Incinerator Machine with Smoke Control Unit, BiPAP Machine (V2), High Temperature High Pressure Table Top Sample Dyeing Machine, Sanitary Napkins Vending Machines, Truck Mounted Jetting Machine, Jetting, Grabbing and Rodding machine Searched String: Induction Casting Machine with Vacuum pump Vacuum Pump, Solar Photovoltaic Water Pumping Systems (V2) as per MNRE Specification Searched String: Preheating furnace Tubular Furnace, Muffle Furnace, Muffle/Box Furnace (V2), Benchtop Thermocouple Calibration Furnace (High Temperature Thermocouple Calibration Furnace)/Portable Calibration Furnace, Hot air oven, HIGH TEMPERATURE FURNACE WITH GAS (PROGRAMMABLE DUAL ZONE / SLIDING TUBE FURNACE) FACILITY, Dolomite, Coke - Blast Furnace (RINL) (In Metric tonne), Graphite Electrodes for Electric Arc Furnace, Cast copper tuyere - IS 8859 Searched String: Die cutting machine Laser cutting machine, GASKET CUTTING MACHINE, Wood cutter circular saw type, Programmable Paper Cutting Machine, Sanitary Napkins Incinerator Machine with Smoke Control Unit, CNC Plasma Cutting Machine, Sanitary Napkins Vending Machines, plasma air cutting machine, High Vacuum Suction Machine (V2), Motorized core cutting machine Searched String: Endomotor DENTAL ENDOMOTOR Searched String: Physiodispensor Category not available on GeM for the text string uploaded by the buyer Searched String: Implant Kit Cochlear Implant (Revision 1), Biochemistry Reagent Kit for Human Samples, Total Knee Replacement Implant, Sternal Closure Implant, personal safety kit reflective luminous garments, Dental Implant Motor (Physiodispenser), Total Knee Replacement Implant (V2), Networking Trainer

बिड विवरण / Bid Details	Educational Kit, Bipolar Hip Prosthesis (V2), Facial Kit
GeMARPTS में खोजा गया परिणाम / Searched Result generated in GeMARPTS	Searched String: Cryo Surgery Unit High Frequency Radio Surgery Unit, Sofa (V3), Thermotherapy or Cryo Cum Contrast Therapy Unit, Freezing / Cryo Box, Executive Table (V4), Cryotherapy Unit, Liquid Nitrogen Storage Equipment, Cryotherapy Unit (V2), Metal Storage Cabinet, Veterinary general surgery kit Searched String: Electro surgery unit High Frequency Radio Surgery Unit, Sofa (V3), Integrated Electrosurgical Unit and vessel sealer, Argon Plasma Coagulator with ESU, Bipolar Diathermy Electrosurgical Unit (V2), Executive Table (V4), Electro Etcher Marking machine, Metal Storage Cabinet, Electro Pneumatic Trainer - Basic, Computer Table (V3) Searched String: Peizo-surgery unit High Frequency Radio Surgery Unit, Sofa (V3), Veterinary general surgery kit, Laparoscopic Surgery System (V2), Executive Table (V4), Video Assisted Thoracoscopic Surgery Equipment, Pediatric General Surgery Instrument Set, Metal Storage Cabinet, Endoscopic Imaging System, Computer Table (V3) Searched String: Centrifuge for PRP/PRF preparation Refrigerated Centrifuge For General And Research Purpose, Non Refrigerated Centrifuge for General Purpose and research, Anhydrous Ammonia Conforming to IS 662, Welding Electrode - Covered Arc Welding Electrode - Basic Type (BEML), Welding Electrode - Covered Arc Welding Electrode - Rutile Type (BEML), WELDING ELECTRODE - Air arc gouging electrode (BEML) -as per C4117 BEML Welding Electrode Searched String: Electro surgical Cautery Electrocautery Machine (V2), Surgical Gowns Conforming to IS 17334, Bipolar Diathermy Electrosurgical Unit (V2), Electrosurgical Cautery Pencil (V2), Single Use Sterile Rubber Surgical Gloves Conforming to IS 13422, Disposable Cautery Plate / Pad (V3), Electro Etcher Marking machine, Surgical Drapes Conforming to IS 17334, Electro Pneumatic Trainer - Basic, Surgical Sutures Searched String: Apex Locator Power Cable Fault Locator, PIPE AND CABLE LOCATOR / ROUTE TRACER, D.C. Earth Fault Locator, Hot Air Oven (FSSAI), Handheld Telecom Cable Fault Locator, Bench Vice, Cable Fault Locator Surge Generator (V2), Walk in Cooler, Offline Fault Locator and Line Analyser, Autoclave (ICMR Specs) Searched String: OPG radiography machine with Cephalostat OPG Machine (V2), Sanitary Napkins Incinerator Machine with Smoke Control Unit, gmaw welding machine with accessories, Haemodialysis Machine with SLED / HDF, Reciprocal Shakers - Mechanical Shaker Machine with Timer, Vitrectomy Machine with Phacoemulsification (V2), Embroidery Work Kit (Gujarat Govt) as per IS 15449, Cardiac

बिड विवरण / Bid Details	TMT Machine (V2), Shoe Shining Machine, Screening Machine with Conveyor for processing Municipal Waste
	Searched String: Cone Beam Computed Tomography (CBCT) CT Scan Machine, SPECT CT, CT Scanners 128 Slice, CT Scanners 16 Slice, CT Scanners 64 Slice, PET- CT 64 Slice Scanning System, PET- CT 16 Slice Scanning System Searched String: Dental Phantom Head Table Dental Amalgam Conforming to IS 18495 Searched String: Dental Diode laser system with accessories Soft Tissue Diode Dental Laser, Dental RadioVisioGraphy (RVG) Unit (V2), Portable Emergency LED Lighting Tower System with Battery Bank Searched String: Automatic Intraoral film processor Automatic X - Ray Film Processor, Automatic Milk Collection Unit with Electronic Milk Analyzer and Adulteration testing, X - Ray Fixer for Radiology Searched String: Dental Operating Microscope Ophthalmic Operating Microscopes, Microscopes - Pathological and Research as per IS 4381, IS 5204, IS 4381, IS 5204, Ophthalmic Operating Microscope (V2), Surgical Operating ENT Microscope, Foot Operated Pedal Bin or Bucket for Bio - Medical Waste Collection, Metallurgical Microscope, General Operating Table, Advance Polarization Microscope, Emergency Foldable Rescue Stretcher, Neurosurgery Operating Table Searched String: Dental Clinical Micro motor Dental Implant Motor (Physiodispenser), Dental Applicator tips for bonding agent, DENTAL ENDOMOTOR, Dental Lab Micromotor (V2) Searched String: Dental magnifying loupes with LED headlight Category not available on GeM for the text string uploaded by the buyer Searched String: Glass bead sterilizer ETO Sterilizer, Single Use Sterile Rubber Surgical Gloves Conforming to IS 13422, Glass Beads, Bead Jewellery from Loding - ODOP, Shallow Sterilizer (Dressing Drum), Beads or bead activity sets for early math or Counting Beads on string, Atulya Microwave Sterilizer for (COVID - 19), Vertical Autoclave, Horizontal Autoclave, Non Thermal Plasma Air Sterilizer Searched String: Pulp Tester Pulp Devltalizer, Insulation Tester, Pulp Board as per IS 4664, Pulp Board (V2) Conforming to IS 4664, Compression Testing Machine, Battery Tester, Portable Hardness Tester, Line Tester, Pendulum Impact Tester, Portable Surface Roughness Tester

बिड विवरण/Bid Details	Searched String: Injectable GP condenser
	Condenser Coil, Laboratory Condenser, Injectable Contraceptive (Antara) for Family Planning Programme, Air Condenser, Tool Kit for Itra - ODOP UP Gov. (Kannauj), Welded Austenitic Stainless Steel Tubes For Condensers And Heat Exchangers- BHEL, Computer Graphics Card (GPU), Manikin for Various types of Injections Training, hand held gps, Spare / Add on Antenna for Transreceiver / Repeater - Mobile / Static / Handheld Searched String: Tissue floatation water bath Organ Bath, water baths, Serological Water Bath (FSSAI), Tissue Flotation Baths, Circulating/Shaking Water Baths (FSSAI), Bathroom Water Shower (V2), Stainless Steel Bain Marie, Waterless Bodywash (V2) Searched String: Biochemistry analyser Biochemistry Analyzer (Semi-Automated), Biochemistry Reagent Kit for Human Samples, Fully Automatic Biochemistry Analyzer (V2), Semi Automatic Bio Chemistry Analyser, Veterinary Biochemistry Analyzer (V2), Veterinary Biochemistry Analyzer, Gas Analyser, ESR Analyser, Breath Analyser, Power Analyser Searched String: Smart dentin grinder for hospital, 300W-500W foot step for hospitals, Internal Combustion Engine Crankcase Oils for Automotive Application (V2) Conforming to IS 13656, Internal Combustion Engine Crankcase Oils for Automotive Application (Diesel and gasoline) as per IS: 13656, Ballasts For Discharge Lamps Conforming To IS 15882, Emergency Foldable Rescue Stretcher, resuscitation crash cart console for hospitals, Instrument Trolley for Hospitals (V2), Fume extraction unit for filtration of soldering fumes, Air Mattress for Medical Use, personal safety kit reflective luminous garments Searched String: Tracheostomy kit Tracheostomy Mask, Biochemistry Reagent Kit for Human Samples, Disposable Tracheostomy Tube (V2), Percutaneous Tracheostomy Dilation Set, personal safety kit reflective luminous garments, Facial Kit, Waxing Kit, Networking Trainer Educational Kit, Ration Kit, Sewing Kit Searched String: Cleft Surgery kit Veterinary general surgery kit, Biochemistry Reagent Kit for Human Samples, Veterinary Periodontal Instrument or Surgery Basic Kit (V2), Laparoscopic Surgery System (V2), personal safety kit reflective luminous garments, High Frequency Radio Surgery Unit, Video Assisted Thoracoscopic Surgery Equipment, Networking Trainer Educational Kit, Pediatric General Surgery Instrument Set, Endoscopic Imaging System Searched String: Trauma kit with bone plating kit Secondary Stage Science Kit With Microscope (NCERT), Medical Training kit with Human mannequins for medical education or training, Secondary Stage Biology Laboratory Kit (With Microscope) (NCERT), Tool Kit for Seeng Evam Asthi Utpaad (Horn and Bone Products) - ODOP UP Gov.

बिड विवरण / Bid Details	(Sambhal), Middle Stage Science Kit with Microscope (NCERT), Secondary Stage Biology Laboratory Kit (Without Microscope) (NCERT), Secondary Stage Science Kit Without Microscope (NCERT), Transformer Oil BDV Test Kit, Dental Composite Kit with Etchant and Bonding Agent, Safety Equipment Tool Kit for Labours
	Searched String: Pedo extraction forceps sets Category not available on GeM for the text string uploaded by the buyer Searched String: Rubber dam kit for pedo patients Medical Training kit with Human mannequins for medical education or training, Gynecological Patient Education Model Searched String: Formalin chamber Inoculation Chamber, water baths, Humidity Chamber, Environmental Chamber, Muffle/Box Furnace (V2), Stability Chamber, Sterilization Chamber, Plant Growth Chamber, Neubauer Chamber / hemocytometer, Posterior Chamber Intraocular Lens (PCIOL) Searched String: Typhodonts Category not available on GeM for the text string uploaded by the buyer Searched String: Adjustable mouth prop set Category not available on GeM for the text string uploaded by the buyer Searched String: Double beak pliers anterior and posterior Artificial Teeth Set Searched String: Bracket Holding Pliers Bracket (Defence), Pipe Clamps (V2), Orthodontic Pliers, Orthopedic Instruments - Nose Pliers, Insulated pliers, Pliers, Pincers and Nippers as per IS 2615, IS 11526, Corner Bracket(BHEL), Sponge Holding Forcep, Groove Joint Pliers, Self - Centring Bone Holding Forceps Searched String: Straight Hoe Pliers Prong Hoe, Single Use Sterile Rubber Surgical Gloves Conforming to IS 13422, Double Wheel Hoe, Single Wheel Hoe, Orthodontic Pliers, Pliers, Pincers and Nippers as per IS 2615, IS 11526, Insulated pliers, Straight Grinders, Funnel, Groove Joint Pliers Searched String: Curved Hoe Pliers Prong Hoe, Single Use Sterile Rubber Surgical Gloves Conforming to IS 13422, Double Wheel Hoe, Single Wheel Hoe, Pliers, Pincers and Nippers as per IS 2615, IS 11526, Curved Member Apparatus, Orthodontic Pliers, cobweb brush, Insulated pliers, Mayo Scissors Searched String: Papoose Board for stabilization of childrens Medium Density Fiber Boards (MDF Boards) Marked to IS

विड विवरण / Bid Details	12406 (V2), Drawing Trussels / Drawing Board Stand, Pre School Uniform - Shorts (Half Pants) as per Government of Gujarat Specification
	Searched String: Articulators Semi adjustable Adjustable Spanner, Revolving Chair (V5), Adjustable Monitor Mount with Adjustable Keyboard Workstation, Box Strapping Machine, Semi Automated Nephelometer, Metal Shelving Racks (Adjustable Type) conforming To Is 1883 (V2), Semi-Rigid Ureteroscope, Semi Automatic Bio Chemistry Analyser, Height Adjustable Table, Patient Bed (Fowler and Semi Fowler) Searched String: Airotor Handpiece s Airotor Catridge For Airotor Handpiece- IS 1797, Dental Airotor Diamond Burs, S type load Cell, Parallel Bar - Outdoor Gym Equipment, S-Curve Urethral Dilator Set, Vertical Autoclave, 2,4,6-Tris(2-pyridyl)-s-triazine, Spirometer, Sternum Saw System, Books (V2) Searched String: Straight Handpiece for clinical micromotor Airotor Catridge For Airotor Handpiece- IS 1797, Cable Joint Kit Suitable for Cables Up to and Including 33 kV (V2), Clinical Apron, Dental Lab Micromotor (V2), Earthwire Fitting and Accessories for 66 kV Overhead Power Lines (PSPCL), Artificial Turf, Biochemistry Reagent Kit for Human Samples, Hair Straightener, Fabricated PVC - U Fittings for Portable Water Supplies as per IS 10124, High mast pole for national flags (V2) Searched String: Supra gingival scalers Dental Scalers, Online UPS (≤ 10 KVA) With Battery Conforming To IS 16242 (Part 1), Restorative Filling and Carving Dental Instruments Set (V2) - RBSK, Online UPS (> 10 KVA) With Battery Searched String: Sub gingival scaler 66 kV Potential Transformer for Sub-stations / Power Houses (PSPCL), Theft Proof Rifle Racks, 66 kV Current Transformer for Sub-stations / Power Houses (PSPCL)), 15 kV Neutral Current Transformer for Sub-stations / Power Houses (PSPCL), Handpump Sub-Assemblies (Deepwell Hand Pumps Components) Conforming IS 15500 (Part 3), Dental Scalers, Charge Fire Ext Mech Foam 9 L for High Altitude and Sub Zero Temp Conditions (Defence), Submersible Pumpsets (Three Phase) Conforming To IS 8034 (V2), PVC Insulated Cable For Submersible Pumps ISI Marked To IS 694, Submersible Pumps for Slurry Application Searched String: Surgery Trolleys Stretcher Trolleys, Instrument Trolley for Hospitals (V2), Serving Trolleys (Bamboo), trolleys or accessories, Serving Trolleys-handcrafted, Platform Trolley (V2), Passengers Baggage Trolleys, Bucket Mop Wringer Trolley, Platform Trolley (V3), Passengers Hand Baggage Trolleys

बिड विवरण/Bid Details	
अधिसूचना के लिए चयनित प्रासंगिक श्रेणियाँ / Relevant Categories selected for notification	- Dental RadioVisioGraphy (RVG) Unit (V2) - Dental X - Ray Machine (V2) - Dental Chair
बीओक्यू शीर्षक / BOQ Title	DENTAL COLLEGE EQUIPMENTS
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) / Minimum Average Annual Turnover of the bidder (For 3 Years)	249 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/ OEM Average Turnover (Last 3 Years)	499 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/ Years of Past Experience Required for same/similar service	3 Year (s)
वर्षों के अनुभव एवं टर्नओवर से एमएसई को छूट प्राप्त है / MSE Relaxation for Years Of Experience and Turnover	Yes Complete
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Relaxation for Years Of Experience and Turnover	Yes Complete
विक्रेता से मांगे गए दस्तावेज़/ Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC),Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/ Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	1
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	3
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
विगत प्रदर्शन / Past Performance	40 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	No

बिड विवरण/Bid Details	
बिड का प्रकार/Type of Bid	Two Packet Bid
प्राथमिक उत्पाद श्रेणी/Primary product category	Dental Chair
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Payment Timelines	Payments shall be made to the Seller within 45 days of issue of consignee receipt-cum-acceptance certificate (CRAC) and on-line submission of bills (This is in supersession of 10 days time as provided in clause 12 of GeM GTC)
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	HDFC Bank
ईएमडी राशि/EMD Amount	499927

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	HDFC Bank
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	3.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	42

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने है। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

(c). ईएमडी और संपादन जमानत राशि लाभार्थी के पक्ष में होनी चाहिए। / Earnest Money Deposit (EMD) shall also be accepted by the buyer in the form of a surety bond.

लाभार्थी /Beneficiary :

Managing Director

Chhattisgarh Medical Services Corporation Limited, Health and Family Welfare Department Chhattisgarh,
Chhattisgarh Medical Services Corporation Limited,
(Md Cgmisc)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No
--	----

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं/सेवा प्रदाता को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ / Purchase Preference to MSE OEMs/ Service Provider available upto price within $L1+X\%$	15
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माता/सेवा प्रदाता को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Percentage of Bid quantity/amount for MSE OEMs/ Service Provider Purchase preference	25

1. If the bidder is a Micro or Small Enterprise as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Experience Criteria" as defined above subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
2. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover, shall upload the supporting documents to prove his eligibility for Relaxation.
3. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Experience Criteria" as defined above subject to their meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
4. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Bidder Turnover" as defined above subject to their meeting of quality and technical specifications. If the bidder is DPIIT Registered OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover shall upload the supporting documents to prove his eligibility for Relaxation.
5. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
6. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
7. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid

document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

8. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs from the State of Bid Inviting Authority having valid Udyam Certificate and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry/ State of Bid Inviting Authority. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service, and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017 and its subsequent Orders/Notifications issued by concerned Ministry. Benefits of MSE will be allowed only if seller/service provider is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

9. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

10. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 40% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Pre Bid Detail(s)

मूल्य भिन्नता खंड दस्तावेज़/Pre-Bid Date and Time	प्री-बिड स्थान/Pre-Bid Venue
23-06-2026 12:00:00	Chhattisgarh Medical Services Corporation Limited 3rd Floor, South East Corner, Vyawshayk Parisar, Sector 27, Near Mini Market, Atal Nagar, Nawa Raipur - 492015, Chhattisgarh.

Dental Chair (50 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Chair along with complete unit is suitable for	Right handed operator, Left handed operator, Convertible chair for Left and Right handed operators
	Chair has provision for swivel of Dentist Arm Left/Right, with more than 300 degrees movement	Yes, No

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Chair has safety brake system at base and backrest while going down, for patient's exit position	Yes, No
	Lifting capacity of chair	150 to 180 Kg, 181 to 200 Kg, More than 200 Kg
	Number of patient positioning programs available	Upto 4, More than 4
	Headrest with multi-axis rotation and height adjustable	Yes (Mechanical), Yes (Pneumatic), Not applicable
	LED operating light	Yes (with manual operation), Yes (with touch free operation)
	LED operating light intensity in lux	25,000 to 30,000 lux, 30,001 to 40,000 lux, 40,001 to 50,000 lux, More than 50,000 lux
	Rinsing water / cup water heating facility available	Yes, No
	Comes with integrated LED light curing unit	Yes, No
	Number of programs in integrated LED light curing unit	3, More than 3, NA (Integrated LED light curing unit not provided)
	Dentist side syringe	3 Way Syringe, 6 Way Syringe
	Dentist side control has Provision for inbuilt brushless micromotor with speed range upto 40000 rpm	Yes (fibre optic), Yes (non-fibre optic), NA
	Dental Chair shall be provided with suitable oil free, Medical grade, noise free and with anti-bacterial coating Air Compressor	Yes, No
	Dental Compressor have capacity to run up to	Single dental chair unit, Two dental chair units, Three dental chair units, NA if not provided
	Airflow at 5 bar in litre per minute	? 65 litre per minute, NA
	Back-lit radiographic image viewer provided	Yes, No
	Operator stools provided with dental chair	1, 2
	Ultrasonic Scaler provided	Yes, No
	External Stabilizer	Yes, No
	Control panel with Display	With 3 digit display, With LED/LCD display, With Touch panel, Without display

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Comes with integrated medical grade flat panel mounted on an arm that should rotate and tilt for comfort viewing of doctor and patient	Yes (Without Touch Screen), Yes (With Touch Screen), No
	Size of flat panel of integrated medical grade screen in inches	18 Inches or less, 18 to 20 Inches, 21 to 24 Inches, More than 24 Inches, NA (if not provided)
	Resolution of screen	1920 x 1080, 1920 x 1200, 1360 x 768, NA if monitor not provided
	Comes with Intraoral Camera	Yes, No
	Comes with RVG	Yes, No
WARRANTY	Warranty in years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1, 2, 3, 4, 5 Or higher (year)

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	50	58

Dental Chair (5 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Chair along with complete unit is suitable for	Right handed operator, Left handed operator, Convertible chair for Left and Right handed operators
	Chair has provision for swivel of Dentist Arm Left/Right, with more than 300 degrees movement	Yes, No

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Chair has safety brake system at base and backrest while going down, for patient's exit position	Yes, No
	Lifting capacity of chair	150 to 180 Kg, 181 to 200 Kg, More than 200 Kg
	Number of patient positioning programs available	Upto 4, More than 4
	Headrest with multi-axis rotation and height adjustable	Yes (Mechanical), Yes (Pneumatic), Not applicable
	LED operating light	Yes (with manual operation), Yes (with touch free operation)
	LED operating light intensity in lux	25,000 to 30,000 lux, 30,001 to 40,000 lux, 40,001 to 50,000 lux, More than 50,000 lux
	Rinsing water / cup water heating facility available	Yes, No
	Comes with integrated LED light curing unit	Yes, No
	Number of programs in integrated LED light curing unit	3, More than 3, NA (Integrated LED light curing unit not provided)
	Dentist side syringe	3 Way Syringe, 6 Way Syringe
	Dentist side control has Provision for inbuilt brushless micromotor with speed range upto 40000 rpm	Yes (fibre optic), Yes (non-fibre optic), NA
	Dental Chair shall be provided with suitable oil free, Medical grade, noise free and with anti-bacterial coating Air Compressor	Yes, No
	Dental Compressor have capacity to run up to	Single dental chair unit, Two dental chair units, Three dental chair units, NA if not provided
	Airflow at 5 bar in litre per minute	? 65 litre per minute, NA
	Back-lit radiographic image viewer provided	Yes, No
	Operator stools provided with dental chair	1, 2
	Ultrasonic Scaler provided	Yes, No
	External Stabilizer	Yes, No
	Control panel with Display	With 3 digit display, With LED/LCD display, With Touch panel, Without display

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Comes with integrated medical grade flat panel mounted on an arm that should rotate and tilt for comfort viewing of doctor and patient	Yes (Without Touch Screen), Yes (With Touch Screen), No
	Size of flat panel of integrated medical grade screen in inches	18 Inches or less, 18 to 20 Inches, 21 to 24 Inches, More than 24 Inches, NA (if not provided)
	Resolution of screen	1920 x 1080, 1920 x 1200, 1360 x 768, NA if monitor not provided
	Comes with Intraoral Camera	Yes, No
	Comes with RVG	Yes, No
WARRANTY	Warranty in years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1, 2, 3, 4, 5 Or higher (year)

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	5	60

Dental Chair (10 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Chair along with complete unit is suitable for	Right handed operator, Left handed operator, Convertible chair for Left and Right handed operators
	Chair has provision for swivel of Dentist Arm Left/Right, with more than 300 degrees movement	Yes, No

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Chair has safety brake system at base and backrest while going down, for patient's exit position	Yes, No
	Lifting capacity of chair	150 to 180 Kg, 181 to 200 Kg, More than 200 Kg
	Number of patient positioning programs available	Upto 4, More than 4
	Headrest with multi-axis rotation and height adjustable	Yes (Mechanical), Yes (Pneumatic), Not applicable
	LED operating light	Yes (with manual operation), Yes (with touch free operation)
	LED operating light intensity in lux	25,000 to 30,000 lux, 30,001 to 40,000 lux, 40,001 to 50,000 lux, More than 50,000 lux
	Rinsing water / cup water heating facility available	Yes, No
	Comes with integrated LED light curing unit	Yes, No
	Number of programs in integrated LED light curing unit	3, More than 3, NA (Integrated LED light curing unit not provided)
	Dentist side syringe	3 Way Syringe, 6 Way Syringe
	Dentist side control has Provision for inbuilt brushless micromotor with speed range upto 40000 rpm	Yes (fibre optic), Yes (non-fibre optic), NA
	Dental Chair shall be provided with suitable oil free, Medical grade, noise free and with anti-bacterial coating Air Compressor	Yes, No
	Dental Compressor have capacity to run up to	Single dental chair unit, Two dental chair units, Three dental chair units, NA if not provided
	Airflow at 5 bar in litre per minute	? 65 litre per minute, NA
	Back-lit radiographic image viewer provided	Yes, No
	Operator stools provided with dental chair	1, 2
	Ultrasonic Scaler provided	Yes, No
	External Stabilizer	Yes, No
	Control panel with Display	With 3 digit display, With LED/LCD display, With Touch panel, Without display

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Comes with integrated medical grade flat panel mounted on an arm that should rotate and tilt for comfort viewing of doctor and patient	Yes (Without Touch Screen), Yes (With Touch Screen), No
	Size of flat panel of integrated medical grade screen in inches	18 Inches or less, 18 to 20 Inches, 21 to 24 Inches, More than 24 Inches, NA (if not provided)
	Resolution of screen	1920 x 1080, 1920 x 1200, 1360 x 768, NA if monitor not provided
	Comes with Intraoral Camera	Yes, No
	Comes with RVG	Yes, No
WARRANTY	Warranty in years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1, 2, 3, 4, 5 Or higher (year)

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	10	60

Dental RadioVisioGraphy (RVG) Unit (V2) (7 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Type of Sensor	Adult Sensor, Pediatric Sensor
	Size of the Adult sensor in mm x mm with tolerance +/- 2mm	20mm x 30mm, 25mm x 30mm, 25mm x 35mm, 37mm x 25 mm, 40mm x 30mm, 43mm x 31mm, Not Provided

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Sensor active area (Adult Sensor) in square millimeter)	less than 400 sq mm, 400 to 500 sq mm, 500 to 600 sq mm, 600 to 700 sq mm, more than 700 sq mm, Not Provided
	Size of the pediatric sensor in mm x mm with tolerance (+/- 2mm)	16mm x 24mm, 20mm x 30mm, 25mm x 30mm, 25mm x 35mm, 37mm x 25 mm, 40mm x 30mm, 43mm x 31mm, Not Provided
	Sensor active area (Pediatric Sensor) in square millimeter	less than 400 sq mm, 400 to 500 sq mm, 500 to 600 sq mm, 600 to 700 sq mm, more than 700 sq mm, Not Provided
	Quantity of Adult sensors provided	1, 2, NA
	Quantity of Pediatric sensors provided	1, 2, NA
	Sensor thickness in mm	Less than 5mm, 5mm to 7mm, 7mm to 9mm, more than 9mm
Accessories	Latest Generation Laptop / Desktop provided	Yes, No
	Compatible Color Printer provided	Yes, No
WARRANTY	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1, 2, 3, 4, 5 (year)

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	7	60

Dental RadioVisioGraphy (RVG) Unit (V2) (2 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Type of Sensor	Adult Sensor, Pediatric Sensor
	Size of the Adult sensor in mm x mm with tolerance +/- 2mm	20mm x 30mm, 25mm x 30mm, 25mm x 35mm, 37mm x 25 mm, 40mm x 30mm, 43mm x 31mm, Not Provided
	Sensor active area (Adult Sensor) in square millimeter)	less than 400 sq mm, 400 to 500 sq mm, 500 to 600 sq mm, 600 to 700 sq mm, more than 700 sq mm, Not Provided
	Size of the pediatric sensor in mm x mm with tolerance (+/- 2mm)	16mm x 24mm, 20mm x 30mm, 25mm x 30mm, 25mm x 35mm, 37mm x 25 mm, 40mm x 30mm, 43mm x 31mm, Not Provided
	Sensor active area (Pediatric Sensor) in square millimeter	less than 400 sq mm, 400 to 500 sq mm, 500 to 600 sq mm, 600 to 700 sq mm, more than 700 sq mm, Not Provided
	Quantity of Adult sensors provided	1, 2, NA
	Quantity of Pediatric sensors provided	1, 2, NA
	Sensor thickness in mm	Less than 5mm, 5mm to 7mm, 7mm to 9mm, more than 9mm
Accessories	Latest Generation Laptop / Desktop provided	Yes, No
	Compatible Color Printer provided	Yes, No
WARRANTY	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1, 2, 3, 4, 5 (year)

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	58

Dental X - Ray Machine (V2) (6 pieces)

Bis Required	Yes
--------------	-----

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
WARRANTY	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1, 2, 3, 4, 5 Or higher (year)

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	6	58

Dental X - Ray Machine (V2) (6 pieces)

Bis Required	Yes
--------------	-----

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
WARRANTY	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1, 2, 3, 4, 5 Or higher (year)

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	6	59

Stereo Microscope

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

High Intensity Light Curing Unit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	8	60

Ultrasonic Scaler

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	30	60

Conscious Sedation Unit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Hydrogen Soldering Unit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Spot Welder With Soldering Attachment

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Lab Micro Motor

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	50	60

Model Trimmer Double Disc

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	5	60

Model Trimmer With Diamond Disc

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

High Speed Lathe

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Sand Blaster Micro Macro

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Autoclave- Front Loading

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	10	60

Ultrasonic Cleaner Capacity 3.5lts Type III

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	3	60

Plaster Dispenser

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Vibrator

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Acrylizer

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Pressure Molding Machine

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Induction Casting Machine With Vacuum Pump

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Preheating Furnace

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Die Cutting Machine

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Endomotor

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	4	60

Physiodispensor

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	5	60

Implant Kit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	4	60

Cryo Surgery Unit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Electro Surgery Unit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Peizo-surgery Unit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Centrifuge For PRP PRF Preparation

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Electro Surgical Cautery

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Apex Locator

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	3	60

OPG Radiography Machine With Cephalostat

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Cone Beam Computed Tomography CBCT

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Dental Phantom Head Table

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	6	60

Dental Diode Laser System With Accessories

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Automatic Intraoral Film Processor

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Dental Operating Microscope

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Dental Clinical Micro Motor

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	200	60

Dental Magnifying Loupes With LED Headlight

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	8	60

Glass Bead Sterilizer

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Pulp Tester

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	4	60

Injectable GP Condenser

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Tissue Floatation Water Bath

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Biochemistry Analyser

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Smart Dentin Grinder For Hospital, 300W- 500W

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Tracheostomy Kit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Cleft Surgery Kit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Trauma Kit With Bone Plating Kit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Pedo Extraction Forceps Sets

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	4	60

Rubber Dam Kit For Pedo Patients

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Formalin Chamber

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Typhodonts

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	3	60

Adjustable Mouth Prop Set

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	3	60

Double Beak Pliers Anterior And Posterior

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Bracket Holding Pliers

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	3	60

Straight Hoe Pliers

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	3	60

Curved Hoe Pliers

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Papoose Board For Stabilization Of Childrens

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	2	60

Articulators Semi Adjustable

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	1	60

Airotor Handpieces

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	50	60

Straight Handpiece For Clinical Micromotor

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	20	60

Supra Gingival Scalers

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	5	60

Sub Gingival Scaler

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	5	60

Surgery Trolleys

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Virendra Vadher	492001,Near Marhi Mata Mandir, Rajbandha Maidan, Raipur	8	60

Special terms and conditions-Version:2 effective from 27-03-2026 for category Dental Chair

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts

manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.

11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

Special terms and conditions-Version:1 effective from 19-02-2025 for category Dental RadioVisioGraphy (RVG) Unit (V2)

1. 1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products

supplied are manufactured by OEM indicated and sourced from them.

9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination, Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

Special terms and conditions-Version:1 effective from 20-09-2023 for category Dental X - Ray Machine (V2)

1. 1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on the submission of valid Drug License and self declaration of product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of drug license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service

charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the

original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

अस्वीकरण/Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs and including the modification of technical specifications and / or terms and conditions governing the bid. All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM. If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code

on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and
- All operative provisions of the erstwhile Labour Laws until their complete substitution.

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.

However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---