

Govt. Dental College , Raipur (C.G.)

DENTAL DEPARTMENT (EQUIPMENT)

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
1	Electrical Dental Chairs and Units Type I	1.It should have double articulating headrest with seesaw movement. 2. It should be provided with soft cervical support. 3. Dental unit should have latest overhead delivery system. 4. It should have two 3 way syringes (Tip autoclavable, with spare tips) one on unit side and other on the assistant side. 5. It should have two high speed Air rotor terminals with two rotor hand pieces and accessories and one terminal for fiber optic. One for air motor/micro motor having straight and contra angle hand pieces and other for air rotor terminal with two air rotor hand pieces with two spare cartridges. 6. It should have LED light cure unit with minimum intensity 1200 mW/cm2. 7. It should have infection control system with non-retraction valves (Bio system/equivalent). 8. All hand pieces/terminals should be kept on Autoclavable pads. 8 spare autoclavable pads should be supplied. 9. Arm of unit should be pneumatically locked. 10. All air tubing of the delivery system can be disinfected internally after every dental procedure. 11. It should have one in built piezo ultrasonic scalar (max frequency should be 36 kHz) 12.Removable auxiliary tray (autoclavable) shall be supplied – 10 sets. 13. It should have integrated latest foot operated LED light (30000 - 50000 Lux). 14. It should have rotatable water system with removable spittoon. 15. It should have Medium Vacuum Suction and high suction (Motorized Suction). 16. Should have following multiple programmes. i. Two programmable working positions. ii. Splitting and last working position with light ON and OFF automatically. iii. Return to Zero position with light OFF automatically. iv. It should have emergency stop control with luminous indication. v. Programmable bowl water and cup filler water. 17. It should have LED based X-ray viewer (For I.P.G/O.P.G films). 18. It should be provided with right arm. 19. It should have multi functional foot control base. 20. It should be provided with two stool with adjustable backrest tilt including an adjustable ring for foot rest 21. Oil free medical grade compressor of 1HP (fully imported) 22 Weight of the base of the Chair should be minimum 160 kgs.		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	50
2	Dental chairs and unit Type II	Should be Electrically operated with shadow lamp, spittoon,3- way syringe, instrument tray and motorized suction		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	5
3	Dental Chairs and Units for Pediatric patient Type III	Should have Body contoured electrically operated PUmoulded ,with foldable handle, operating light should have two intensity, should have chair side porcelain spittoon , vaccum suction -high and low motorized, should have chair mount unit modular delivery system with hanging/over the patients,cords,air rotor- minimum2 points, 1 for fibre optic HP with cord, ultrasonic scalers with 3 tips and light cure LED light (turbo), 3 way syringe (2 no) , xray viewer , multifunction foot control. Operating (pneumatic)stool with arm rest and back rest, height adjusting should be provided.		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	10
4	Stereo microscope	It should have a Viewing Head- Trinocular Head, Inclined at 45°, with Interpupillary Distance 57- 75mm, 30:70 light distribution. Eyepiece should be- Extra wide field eyepiece EW10x/ 22mm WF15x/ 16mm WF20x/12.5mm. Zoom Objective should be- 0.8x-5× Zoom Ratio -6.3:1 .Working Distance -115mm Should have Frosted Glass Plate, Diameter should be 100mm. Stage Plate should be- White & Black round plate, with Diameter 100mm. Focusing Unit should be- Coarse Focus Knob with tension adjustable, Moving range should be 100mm. and Incident Illumination 100V-240V/LED Transmitted Illumination should be 100V-240V/LED.	Attachment- Photo Attachment (optional) Video Attachment (optional)	Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
5	High intensity light curing unit	Should have wavelength range – 430-480 nm, peak wave length – 44-452 nm, light intensity- 1470/mw/cm2, high emission Ar – 60- 65mm2, power supply – Lithium ion battery, Exposure time setting should be– 5, 10, 15, 20 sec, Continuous mode – 120 sec, Teck cure mode – 15 sec and Battery capacity should be – 2300 mAh		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	8
6	Ultrasonic scaler	It should be supplied with Cutting edge detachable and autoclavable handpiece that will operate at the optimal frequency, ensuring stable performance. The detachable handpiece should be sterilizable at high temperatures and pressure , preventing cross contamination., Power input should be– 110v 50/60 hz220- 240 v – 50/60 hz, Main unit input 24v – 1.3 A., output power 3 -20w, frequency - 28 khz, and water pressure 0.1- 5 bar		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	30
7	Conscious sedation unit	Conscious sedation equipmen have oxygen failure warning and should supply 65% maximum NO2 concentration, 35% minimum O2 concentration and should have Diameter index safety system		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1
8	Hydrogen soldering unit	It should be Electric, automatic, with voltage-230w, power consumption-1 unit per dayKW, should have supply ability =10unit/wk, should have Hypodermic needles for flame size adjustment, should have Fine pointed flame regulated by torch gas regulator and power knob.		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1
9	Spot welder with soldering attachment	Dental welder and solder that are perfect for orthodontic departments and labs., Adjustable upto 100W, Voltage : 220V, Heating time should be less than 2 secs, with stainless steel housing		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
10	Lab micro motor	Lab micro motor control unit should have hand piece, It should have both Forward and Reverse rotation with compact design and speed control option along with Foot control. Should have: Input voltage: DC 30V-0.5 A , power source :110/220V. Speed range:0-35000/0-45000rpm, , Should be supplied with, bur holder – 18 pcs and hand piece holder		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	50
11	Model trimmer (double disc)	Double carborundum wheels should have 3000-6000rpm,Sshould have Heavy base and can be mounted on platform. Should have Industrial grade scaled ball bearing which never needs lubrication		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	5
12	Model Trimmer with Diamond Disc	HP-1 motor with 2800 RPM for robust trimming power. ☐ Two ball bearings ensure stability and durability during use. ☐ Single-speed functionality for straightforward operation. ☐ Continuous rating and equipped with a buffer for enhanced performance. ☐ 12" Diamond Disc (Rough/Fine) from Indian craftsmanship for precise trimming. ☐ Specifically designed for orthodontic and prosthodontic work. ☐ Excludes suction, streamlining the trimming process without additional apparatus. ☐ Usage Rating: Continuous ☐ Disc Diameter: 12 inches ☐ Disc Type: Diamond (Rough/Fine)		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
13	High Speed lathe	Should be equipped with a suction box with light which can be connected to a dust collector for dust-free operations. Power Consumption: more than 14W Rotational speed: 3000rpm or more Max. Vacuum: -16kPa Max. Noise: 60dB		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
14	Sand blaster (micro & macro)	Micro sandblaster (intraoral) specification – ☐ Link way: 2 hole dental unit. ☐ Operation air pressure:0.3Mpa-0.4Mpa. ☐ Sand amount:1.5g/min-3.59g/min. ☐ Blasting hole: 0.5mm. ☐ Blasting head can rotate to 360°. ☐ Built-in antisuction system Macro sandblaster specification - It can be attached with vacuum, pneumatic foot controllers, two tungsten carbide tips for micro-sandblasting, one re-cycle for macro-sandblasting, air gun, LED light. Working Pressure should be in the range of 1-6 bar Maximum connection pressure 6-8 bar Minimum number of tank two Provision for lamp with lighting intensity should be more than 4000 lux Air supplied more than 80L/min		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
15	Autoclave- front loading	B type front loading Stainless steel chambers autoclave, should be able to sterilize instruments, glass wares and plastic wares. Operation should be simple with a large LED display, automatic process and automatic safety interlock that prevents door from opening until internal pressure has reached an equilibrium with the room. Capacity should be minimum- 23 L Sterilizing temperature 121°C, 134°C		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	10
16	Ultrasonic cleaner capacity 3.5lts Type III	It should have a Cleaner with a capacity of 3.5 litre, digital timer of 15 minutes, power – 100 w With digital control, timer, heating, programmable and liquid drain		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	3
17	Plaster Dispenser	It should have: High frequency dispensor. Rubber bucket for plaster storing. Should dispense plaster in dusty powder form. Bottom lever should start and stop flow of plaster instantly. Should be wall mountable. Capacity within range of 20- 30kg. Should have a stainless steel body		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
18	Vibrator	Shape should be Rectangular with Capacity of 3-4 Casting Cylinders Removable Plate Vibrational Frequencies should have Variable, controller-fitted Purpose: that Enhances precision in dental casting procedures		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
19	Acrylizer	Operation with distill water or saturated steam. -Meets all technical requirements for resin curing. -Enables muffle curing with adjustable pressure and temperatures and pressurization of cylinders. -Enables selection of manual and automatic cycles with selector for final discharge of air pressure only or total discharge (air & water) of the polymerization chamber. -Display of temperature and polymerization time. 10 storable programme. Operating temperature from 10 to 130 degree C. -Operating time ranges 01 min to 02 hrs. Mechanical safety lock, Acoustic warning signal and Pressure relief valve with automated setting.		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
20	Pressure molding machine	Heating Time: should reach moulding temperature in seconds. * Heating Element should have minimal power consumption. * Pressure Range: 5.5 to 7 kg/cm * Control Display: Programmable with LED digital display. * Indicators: for heating, cooling. * Interlock: Heating and cooling should operate one at a time. * Process Alerts: Acoustic signal for process end, overrun, or stoppage. * Auto Stoppage: Stop automatically if time exceeds. * Sheet Holder: Fits circular sheets up to 127 mm. * Pressure moulding machine		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1
21	Induction Casting Machine with Vacuum pump, capable of casting Titanium Chrome Cobalt precision Metal	Should permit the casting of precious and non-precious metals and titanium. Should allow to be used in either natural atmosphere,vaccum or inert atmosphere. Should have continous reading control and stabilization of the temperature from 800 to 2000 degree C. Should be equipped with powerful vaccum pump. Should have motor acceleration regulator to control the metal flow to the ring. Should have inbuilt cooling system. Accessories to be provided : 4 rings of stainless steel 1 universal rubber sprue former, 1 crucible per type, 1 ceramic stirring rod , 1 pair of safety goggles		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
22	Preheating furnace	Digital Furnace for Preheating and dewaxing. Heating upto 999 deg. C Double Body Furnace insulated with ceramic wool. Faster heating and cooling cycle that allows continuous Usage. Chamber and door fitted should be with high quality moulded ceramic Size should be more than width 130mm, H- 100mm, D-250mm.		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
23	Die cutting machine	Should have following features: High rotational speed, Adjustable model holder for cutting dental models Should be ideal equipment for cutting dental models. It should have high rotational speed, constant RPM and adjustable model holder. with the built-in lighting system and positioning laser beam. -The diamond plaster disc - Power Supply- 220-110V, 50-60HZ - Rotational Speed - 3,500RPM or more - Disc Size - 80mm/16mm/0.25mm - Optional Accessories - Dust Extractor		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1
24	Endomotor	It should have a speed range of 120-800 rpm, Torque- 0.8-5.2 Ncm with multiple adjustable program selections (at least 6) 16:1. Compact contra angle hand piece Unit should consist of 1 Control Unit, 1 Handpiece, 1 Contra Angle 16:1, One Handpiece Stand, 1 AC Adapter, 1 F-Type Spray Nozzle, Instruction Manual.		Minimum 01 years	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	4

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
25	Physiodispensor	Contra Handpiece 20:1 Automatic torque control Hand piece with LED Should have wireless foot control 100 disposable coolant hoses supplied KIT should be supplied with Resonance frequency analyzer Power Supply AC 100-240 V 50/60 Hz Max. Pump Output 75 mL/min Programs 8 Programs / Implant Systems Dimensions W 245 x D 235 x H 90 mm Torque 5-80 Ncm Motor Speed 200-40,000 min-1 Light Power (Optic Over 32,000 LUX Micromotor) Foot Control Functions Coolant Solution Flow Volume Button, PRG(Program) Butto Forward/Reverse Button, Speed Control Pedal		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	5
26	Implant Kit	Dental implat kit with different drills and torque wrench Sinus lift kit with provision of crestsal and lateral sinus lift Ridge split / spreader kit Prosthetic kit/ implant removal kit trephine bur kit bur kit for socket shield implant placement sterilization cassette compatible tissue punch		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	4
27	Cryo Surgery Unit	Cryo surgery Unit should have Nitrous Oxide It should be having: Nitrous oxide based cryotherapy unit 2. With probes (a). bevel edge derm probe (b). Round flat surface probe (c). Micro Tips skin warts probe (d). sharp Micro probe tags		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1
28	Electro surgery unit	General Specifications: Frequency: Operating frequency typically ranges between 300 kHz to 500 kHz. Power Output: It should have monopolar cutting with a maximum output of 300W, while coagulation modes which can reach 120W. Bipolar modes should have a maximum output of 70W or 80W. Operating Modes: Units should offer various cutting modes (e.g., SP cut, pure cut, blend cut) and coagulation modes (e.g., soft, fulgurate, spray). Bipolar modes should include options like micro and macro. Safety Standards: Electrosurgery units should comply with relevant safety standards such as IEC 60601-1-1, IEC 60601-2-2, and EMI/EMC compatibility.		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1
29	Peizo-surgery unit	It shoulh have following features: Power Input: Typically around 54VA to 100VA. Drive Frequency: 28-32kHz to 24-36kHz. Maximum Output: Can range up to 25W. Water Flow: 25- 110ml/min. Voltage: Standard AC230V 50/60Hz or 100-120V 50/60Hz. Other Features: features like LED lighting, touch screen displays, and ergonomic handpiece designs.		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1
30	Centrifuge for PRP/PRF preparation	It should have a maximum Speed of 5.000 rpm and maximum RCF (relative centrifuge force): 4.420 xg. Silent operation: less than 65 dB Dimensions approximately (LxWxH, in cm): 60 x 54 x 36. Chamber of medical grade stainless steel, LCD screen and should be equiped with smart navigation through touch- mode and buttons.		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
31	Electro surgical Cautery	It should have following features: Power Output: Ranges from 150W to 300W or higher. Operating Frequency: Typically between 300 kHz and 500 kHz. Monopolar vs. Bipolar: Monopolar should use a patient plate as the return electrode, allowing for greater power output and wider tissue contact. Bipolar should use the electrodes as both the active and return electrode, providing more precise coagulation and less risk of off-target effects.		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
32	Apex Locator	At least 3rd generation or more advanced generation apex locator with digital LCD display. It should consist of Main Unit, Measuring Wire, File Clip, Lip Hook, Measuring File, Adapter, Tester, Battery, Instruction Manual, Packing List		Minimum 01 years	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	3
33	RVG Type I	1. Operational Requirements: 1.1 RVG system required for taking radiography images with operational tools in the viewing software for diagnosis of Intra Oral Diseases 2. Sensor Requirement and Specifications 2.1 Sensor should have CCD or CMOS Sensor Technology 2.2. Sensor Cable Length has to be between 2.5m and 4m. 2.3. Sensor should be provided in a set of 1 Adult Size Sensor and relevant Accessories. 2.4. Should come strengthened twisted wire cable. 2.5. Should have rounded edges on all sides. 2.6. Should have a sensor thickness of up to 7mm ± 0.3mm 2.7. Should be compatible with all types of Dental X-Ray machine. 2.8. Should have detailed software that is easy to use and operate. 2.9.. Sensor Software should have lifetime licensing and the supplier shall ensure that the sensor software can be reinstalled and licensed even after multiple formatting of the PC during the duration of warranty and CMC. No additional charges or requests will be allowed for Re-Installation of Software or license during the set of installed period. 2.10.. Sensor Software should be compatible with the existing set of RVGs or should have a TWAIN module / DICOM Module to include the existing set of RVGs present at GDC Raipur without any reduction in operational characteristics. 3.12. The Sensor Software should have the following Features. A. Updated Diagnostic System Utilities like Pseudo 3D Mode, Relief Filter, Iso-Density, Densitometric Analysis and Measurement with Calibration. B. Should come with image filters like Perio, Endo, Dentin to Enamel Junction, Negative, Color, Etc. shall be preferred. 3.13. Sensor software should be such that it can be installed on Existing PC Units (Compatible with All) and must have a dedicated future for Database Management and Networking so that in future if any centralized data storage is required, the software can add and collect data from one single database. 3.14. Sensor should have true measured resolution of 23 lp/mm to 30lp/mm. 3.15. Sensors should come with a wall bracket that can be fixed on the wall for positioning the RVG. 3.16. Sensors should come with intra-oral positioners with both horizontal and vertical alignment. 3.17. Sensor should come with a set of 5000 Sleeves which are smoothened and not cause any discomfort to the patient. Sample needs to be submitted at the time of bidding. 4. Standards, Safety and Training 4.1. Should be FDA/ CE /ISO/BIS approved product.	Compatible computer system CompatiblePrinter Compatible UPS with minimum 30 mns back-up	Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	7
34	RVG for pediatric patient Type II	Super CMOS Scintillator Optical Fiber Pixel Size – 18.5 x 18.5 micron to 20 x 20 micron, , sensor size 0 (for pedo) Exposure life should be minimum 4 lakhs, Soft ware such as IOC , Sensor cable length should be 3 metre and reinforced for durability and reliability, Fibre optic and scintillator technique.,	Compatible computer system CompatiblePrinter Compatible UPS with minimum 30 mns back-up	Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
35	OPG radiography machine with Cephalostat	1. OPG Requirement and Specifications: 1.1. The OPG unit should be Wall Mounted / Floor Mounted 1.2. The OPG unit should have an extended Cephalometric Arm with separate sensors for 2D Panoramic Imaging and 2D Cephalometric Imaging. The Cephalometric modality should come with Auto Tracing and allow to customized and create templates. 2. OPG X Ray Tube Head and Sensor Specifications: 2.1. X Ray Generator for the OPG unit should be of Constant potential, high frequency up to 160 kHz and microprocessor controlled. The radiation mode for the generator should be continuous or pulsed. 2.2. X Ray Generator for the OPG should have an X-Ray tube voltage range of 60kVp to 90kVp and X-Ray Tube Current range of 01 mA to 16 mA. OPG system shall have Voltage and Current modulation.	Compatible computer system CompatiblePrinter Compatible UPS with minimum 30 mns back-up	Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY and AERB setup and certificate to be provided by the vendor"	1
		2.3. Focal Spot should be approximately 0.5mm. 2.4. Total Filtration should be more than 2.5mm Al Equivalent. 2.5. The OPG unit should have sensors based on CCD / CMOS / CSI technology. 2D cephalometric and 2D Panoramic sensor should be separate and should work without removing any one of the two. 3. OPG Scanning Specifications: 3.1. It should allow, patient positioning in Seated or standing with / without wheelchair with temporal support and positioning guides. 3.2. Magnification offered by the OPG unit should be not more than 1.29 for panoramic. 3.3. OPG programs should be full panoramic, segmented panoramic, maxillary sinus, LA TMJ x 2, LATMJ x 4, Cephalometric mode, Lat. Skull, PA views, PA Ceph, Lat. Ceph, PNS view, SubmentoVertex View. 3.4. Media format for OPG Output file shall be Medical Format; JPEG, DICOM Image Archive orDICOM Removable Media. 3.5. Should be DICOM compatible. CD / DVD, optical disc archiving. 3.6. The sensor should be always ready for taking an image. The images should be automatically stored into the selected patient's file. No interaction with the pc, keyboard or mouse need be required during image acquisition. 4. OPG Software Specifications: 4.1. OPG shall come with fully loaded software with a number of image processing capabilities. Software version must be specified. From contrast and brightness to magnification and full-screendisplay, giving a full range of tools to help maximize diagnostic findings. 4.2. The Software should be provided with timely updates of software free of cost for the duration of 10 years. 4.3. Should come with multiple licenses (at least 15) that can be utilized for maintaining a uniform and well-defined database.				
		4.4. Network connectivity - It should have Integrated DICOM interface that supports printing from DICOM modalities:10/100/1000 Base-T Ethernet connection to imager. 5. OPG Accessories and Attachments Specification: 5.1 The vendor should provide lead aprons (minimum 2), apron hangers, apron stands, and thyroid shields. 6. OPG additional specifications and Room construction requirements: 6.1. The bidder shall ensure complete compliance with AERB norm and shall provide, at his own cost,site registration and approval from AERB for duration of 10 Years from the date of commissioning of the machine. QA reports have to have generated every 02 Years and submitted with GDC for theentire duration of Warranty and CMC. 6.2. The bidder on their own behalf shall arrange for a demonstration of the unit and the software either at an existing installation location or at GDC premises. GDC Raipur reserves the right to technically disqualify a machine if it is not at par with the quality output or build quality..				0

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
36	Intraoral x ray machine (Floor mounted – movable on wheels) with compatible voltage stabilizer	SPECIFICATION FOR DENTAL X-RAY: 1. Operation should be conventional as well as automatic. 2. Completely micro controller based digital timer assuring the accuracy of the exposure time selected. 3. Ease of operation as all the functions can be selected from the remote control as well as timer. 4. Feather touch keypad and length of exposure cable should be 5 to 6 meters. 5. Digital timer with the accuracy of 0.01 Sec (0.01 Sec to 4.00 Sec). 6. Patient selection Switches (Thin, Normal and Obese) 7. Film Speed selection switches (3 Speeds) 8. RVG mode for RVG sensor. 9. An excellent output of 65 kV to 70 kV, 7mAs to 10 mAs. 10. Audible and Visual indication of “X-Ray On” (Radiation indications). 11. Should provide compatible voltage stabilizer (Built in/External). 12. Collimating device should be 20 cm in length and parallel/square in a lead shield should also be provided. 13. Excellent, Mechanical maneuverability, long reach scissor arm Power requirements 230V, AC, 50 Hz, 15 Amps, Should not be Battery operated Protection Suitable stabilizer to be provided. High voltage protection for X-ray tube.	Compatible stabilizer	Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY and AERB setup and certificate to be provided by the vendor"	6
		1. CBCT Requirement and Specifications: 1.1. The machine should have Cone Beam technology for complete dental 2D and 3D volumetric imaging as well as cephalometric 2D imaging for radiographic examination with high accuracy of all clinical situations in dentistry to aid in diagnosis and treatment planning. 1.2. The CBCT unit should have both the options of Wall Mounted / Floor Mounted. 1.3. The CBCT unit should have an extended Cephalometric Arm with separate sensors for 2D Panoramic Imaging and 2D Cephalometric Imaging. The Cephalometric modality should come with Auto Tracing and have the facility of both customized or create templates.				1
		2. CBCT X Ray Tube Head Specifications: 2.1. X Ray Generator for the CBCT unit should be of high frequency 50 to 160 kHz and microprocessor controlled. 2.2. X Ray Generator for the CBCT should have an X-Ray tube voltage range of 50kVp to 120kVp and X- Ray Tube Current range of 01mA to 16mA. CBCT should have Voltage and Current modulation. 2.3. Focal Spot in the range of 0.7 mm or less. The equipment should have less radiation and better actual resolution with minimum focal spot size. 2.4. Total Filtration should be more than 2.5mm Al Equivalent 3. CBCT Scanning Specifications: 3.1. Patient Positioning in the CBCT should be Seated (on an integrated stool) or standing with / without wheelchair with temporal support or patient’s head immobilization devices and positioning. The Unit should have an advanced patient positing mechanism that can be controlled through the operation console.				0

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
37	Cone Beam Computed Tomography (CBCT)	3.2. Scan time shall range from 5.5 Seconds to 40 Seconds. Scout Scan facility should be available for Panoramic (the Software should automatically calculate the correct exposure settings and trajectory) and CBCT Scanning (precisely define the field-of-view position on a low dose scout view). 3.3. Programs offered by the CBCT should be with Multiple Anatomical Settings, Cephalography, Lateral, PA/ AP, Axial, Sagittal, Coronal, PAN with 3D Slice Navigation Segmented PAN, 3D Model, Implant Alignment, Sinus Reconstruction, TMJ views, Multiple Panoramic Modes: - (Adult, Child, Segmented), Extraoral bitewing, TMJ Open - Close -True Profile, Cephalometric (PA, AP, Both in the Asymmetric, & Symmetric projections,) Oblique, Submento Vertex and Carpus - Hand Wrist, Middle ear analysis, 3D Model Scanning, Endodontic mode for visualization of the finest details of the root canal morphology 3D face scan and Airway Analysis. Measure bone density to plan implants and abutments simultaneously within a 3D volume or a panoramic view.	compatible computer system CompatiblePrinter Compatible UPS with minimum 30 mns back-up	Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY and AERB setup and certificate to be provided by the vendor"	0
		3.4. Magnification offered by the CBCT unit should be not more than 1.29 for panoramic & 1.15 for Cephalometric and 1:1 for CBCT images. 3.5. Voxel Size for the offered CBCT shall be in the range of 50 µm - 400 µm. The unit should offer wide choice of selection of voxel sizes to suit different resolution modes or better. High resolution scan with highest resolution should not be limited to smaller FOV scans / sectional scans but should be achievable in larger FOV scans also. (Machine should be capable to take highest resolution scans at larger FOV also. 3.6. The CBCT Machine should come with a wide range of selectable FOV and should follow ALARA (as low as reasonably achievable) principle to allow for good image quality at the lowest reasonable radiation dose. The range of FOV provided should cater to all diagnostic and radiological exams. Smallest possible FOV shall be 4 x 4 or smaller, and largest possible FOV shall be 16 x 16 or higher upto 20 x 20. 3.7. Media format for CBCT Output file should be Medical Format; JPEG, DICOM Image Archive or DICOM Removable Media, Stored in Solid Server Database.				0
		3.8. Should be DICOM compatible and provision for incorporating with RIS (radiology information system) and PACS (Picture Archival and Communication System). CD / DVD, optical disc archiving. 3.9. The Unit should continuously monitor all functions and error messages should be displayed in case of abnormal functioning including (but not limited to) real-time tracking of the machine, assistance for selection of correct accessory for a particular exam, recall of preset parameters for patient by accessing history records. 3.10. The CBCT should have Face camera and should be able to capture and cover patient from different directions to cover maximum face area with all necessary software, which should capture face photo, which will help in many applications such as orthodontic and cephalometric estimations. Integrated face camera shall be preferred over external module. 3.11. Should have with Object / Model Scanning and should openly integrate with third Party software for surgical guide planning, orthognathic surgery planning, etc. 3.12. The Sensor Be Always Ready for taking An Image. The images should be automatically stored into the selected patient's file. No interaction with the pc, keyboard or mouse need be required during the image taking, no matter whether single images or multiple image sequences are taken.				0

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
		4. CBCT Software Specifications: 4.1. CBCT should come with fully loaded software with a number of image processing capabilities. Software version must be specified. From contrast and brightness to magnification and full-screen display, giving a full range of tools to help maximize diagnostic findings. 4.2. The Software should be provided with timely updates of software free of cost. Should provide minimum of 15 Software licenses. 4.3. Dedicated Endodontic mode should allow visualization of the finest anatomic details, Teeth Segmentation, slicing mode should have curved/ oblique / custom / orthogonal features. Complete implant workflow. 4.4. Should enable live comparison of images with and without MAR (Metal Artifact Reduction) filter to optimize diagnostic findings. Metal artefact reduction with the lowest loss of data will be preferred. 4.5. Advanced algorithm that reduces noise while preserving image details; Improves perception of cortical bone edge, ligament space,				0
		soft tissues, and small details. ANR with lowest loss of data will be preferred. 4.6. Should be possible to do prosthetically driven implant planning in the same software. A customizable report layout with the facility to store different report formats should be possible. It should have detailed implant library (regularly updated) 4.7. Should have the facility to convert DICOM into STL format in the same or separate software and have DICOM print facility in the software. Viewer file generation should be possible in the software for patients’ use. 4.8. Should have volumetric measurement for airway analysis. 4.9. Should have presets for 3D multiple Rendering. 3D clipping/ cropping facility from all views (Front-back/ Right- Left /Top- Bottom) Facility to see volume in 2D-CORONAL, sagittal and axial view in single window. Text, drawing, distance & angle measurement overlay facility on images should be available. 4.10. Should come with a Free, full functionality software with implant planning and extensive implant library, report generation, film composer and printing capability and should not be limited to only viewing. It should be shareable with referring doctors / reporting radiologists.				0
		5. 5.1. The vendor should provide lead aprons (minimum 2), apron hangers, apron stands, thyroid shields.5.2. Reporting room should be provided with LED film viewer with adjustable brightness and capable of holding 14” x 17” size film.				0
		6. CBCT additional specifications and Room construction requirements 6.1. The bidder shall ensure complete compliance with AERB norm and shall provide, at his own cost, site registration and approval from AERB for duration of 10 Years from the date of commissioning of the machine with QA test done every 02 years. 6.2. The bidder on their own behalf shall arrange for a demonstration of the unit and the software either at an existing installation location or at GDC premises. GDC Raipur reserves the right to technically disqualify a machine if it is not at par with the quality output or build quality.				0

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
		6.3. The Bidder on their own behalf shall provide a detailed drawing, with a floor plan presented to GDC Raipur for approval, to construct a dedicated room for housing the CBCT unit and its accessories, as well as OPG unit and Intraoral x-ray machine units and its accessories. This room shall be in accordance with the latest AERB norms. Bidder shall visit and inspect the site on their own behalf to understand the scope of works (Civil, electrical, networking, etc) as per the requirements of GDC, Raipur. GDC, Raipur shall be ultimate authority to decide the scope of works to be executed in the CBCT room and the Bidder has to abide by that while preparing the floor plan. Any deviation from the requirements mentioned by GDC, Raipur shall lead to disqualification from the tendering process. 6.4. The bidder, in accordance with approvals from GDC Raipur, should include construction of brick wall, if necessary, as per approved plan layout, plastering, painting, radiation shielding for doors and windows, lead glass partition if required, electrical work, plumbing, air conditioning (for CBCT room, console room and reporting room), Interior furnishing, false ceiling, firefighting equipment and flooring. A dedicated earthing has to be provided for the offered CBCT unit.				0
38	Intraoral x-ray unit (portable) hand held	It should have: 1. Generator - full wave rectified or constant potential. Introral DC x-Ray. 2. Generator Voltage kVp Range- 65kVp to 70kVp and Generator Current mA range 2-3.5mA. 3. Focal spot should be 0.4mm IEC. 4. Must have exposure time range from 0.01 s – 1.0 s. 5. Should either be Battery based or Capacitor Based DC portable X-Ray unit. 6. X-Ray shall be compatible with Digital Sensor, PSP Film and IOPA film. 7. Should Come with a circular collimator. 8. Total Filtration should be more than 1.5mm Al equivalent. 9. Must come with a Lead Apron of 0.5mm Lead Equivalent thickness and a thyroid collar. 10. Supplier should provide Escalation Matrix for Service of provided equipment.		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY and AERB setup and certificate to be provided by the vendor"	6
39	Dental Phantom Head Table	It should be attached with a phantom head simulating the model of a human head and mouth compatible with all type of available jaw sets. This table should be assembled and designed by using excellent quality components. Table Should be made up of Stainless Steel material with multiple storage drawers. It should consist of halogen light with adjustable arm, airrotor attachments with hand piece, micro motor attachments,control box, cord, hand piece along with foot controller.		Minimum 01 years	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	6
		Single vendor should provide the entire system. The technical requirement and scope of work is given below. 1. Description of Function: 1.1 Dental Laser should be provided with state-of-the-art equipment and should have been equipped with high quality equipment and consumables. 2. Equipment: 2.1. Laser type: Diode, Semiconductor Wavelength: 450nm, 650nm or 976nm. Wavelength and power: □ 976± 20nm : 0.2W-4W(CW), Peak Power 5W chopped 1Hz to 20KHz. □ chopped mode :5us – 0.9S □ 650± 20nm : 25mW-200mW (CW) □ 450± 20nm : 0.2W-3W(CW) chopped 1Hz to 20KHz. □ chopped mode: 5us – 0.9S 2.2. Laser classification: □ 976 nm: Class 4 □ 650 nm: Class 2 □ 450 nm: Class 4 2.3. Aiming beam: 650 ± 20 nm /Pmax< 5mW (CW) 2.4. Operation Mode: Continuous or pulsed. 2.5. Divergence:12.7° (According to IEC 60825-1: 2014) and diverging half angle: 0.22 mad/ 12.6 °				2

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
40	Dental Diode laser system with accessories	2.6. Optic Fiber delivery system with multiple tips of varying lengths and bio stimulation tips (both extraoral and intraoral) 2.7. Beam Emission Initiation: Only through Footswitch (Should be a wireless footswitch) 2.8. Display: High Quality Color Display with touch panel Medical approved 2.9. Cooling: Silent cooling fan, multi-channel copper tube electronic cooling		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	0
		2.10. Sleeve of Handpiece should be removable and autoclavable with 3 additional sleeves (metal) provided with the unit. 2.11. Up to 6 or more custom profiles – Can be used by more doctors, storing their preferences. 2.12. Preset programs: For every treatment, there is already a preset program. 2.13. Each present program shall display the correct tip to be used, adjustable laser parameters, and details regarding the procedures as to how it should be done. 2.14. Rechargeable battery 2.15. It should be easy to carry with a handle to move the laser from one place to another.> It should have a remote interlocking option so that if any personnel enter the room without safety goggles it should immediately cut off the laser operation.				
		3. Accessories: 3.1. A total of 50 Tips with varying lengths and thickness. These tips should be compatible with the inbuilt programs in the laser unit. 3.2. 6 Nos of protective eye wear glasses which can prevent all three wavelengths to harm the eye of the user. 3.3. 6 Nos of Extra Oral Bio stimulation Tips of varying sizes, 2 Nos of Intra Oral Desensitization Tips and 2 Nos of Teeth Whitening Tips. 3.4. Articulating paper set – 200 Nos 4. Training 4.1. Adequate training has to be provided for using the laser unit and its accessories, including handling and maintenance to the doctors and assistants at Government Dental College, Raipur				
41	Automatic Intraoral film processor	The automatic film processor should be exclusive for processing intraoral x-ray films It should have following features: 1. Voltage – 230 Volts 2. Frequency – 50 Hz 3. Current consumption – upto 8 Amp 4. Power – 300-500 W 5. Duty cycle – 100% 2. Bath temperature – 24OC 3. Processing time – 5-8 mins 4. Number of films per shift – 5-10 films 5. Equipment: 1. Automatic film processor 2. Daylight loading hood 3. Automatic bath heating 4. Film drying facility 5. Film size – 31 x 41 mm 6. Electrical safety conforms to standards for electrical safety		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
42	Dental Operating Microscope	It should be, Floor mounted dental operating microscope with working life of over 20,000 hours of continuous use with LED illumination, should have 5 step magnification changer, high definition camera attachment with the optional high definition interface that should allow remote viewing, wide choice of options regarding eyepiece magnifications, 0 - 180 Degree Inclined Binocular, unit should consist of Foot Pedal Control, Yellow and green filter and LCD Monitor		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
43	Dental Clinical Micro motor	The unit should have: Speed: 1000 rpm to 40,000 rpmTorque: 2.9 Ncm: Power: 45WUnit, should consist of 1 Micro motor engine, 1 Control Box, 1Foot Pedal, 1 user manual and 1 hand piece stand. The unit should have high torque even at slow speed for tooth structure cutting. The speed should be adjustable from 1000 rpm to 40,000 rpm. It should have the provision of being attached with dental chair or used without dental chair with the help of control box.		Minimum 01 years	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	200
44	Dental magnifying loupes with LED headlight	It should have: Magnification: 2.5x, 3x, 4x, 5x & 6x,Working distance: 370 mm or 420 mm or 460 mm or 500 mm, Field of vision: 120-55 mm, Resolution: 12-20		Minimum 01 years	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	8
45	Glass bead sterilizer	The well size should be (mm)- 59*58, Wattage 250 W, Temp control, Digital temp indicator with micro controller based, PID control, Temp range 220-240V, 50 Hz, Portable size, rapid warm up facility, high efficiency and highly accurate temperature, digital micro control beed, Autosleep mode			US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
46	Pulp Tester	Unit should consist of 2 Test Electrode, Cable Socket, Light Indicator and Battery. It should examine the vitality of dental pulp using the electric stimulation. It should works in an efficient way. Preset speed mode (high-mid- low speed).		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	4
47	Injectable GP condenser	Not in EEL Four preset temperature Wt. 80 gms Wt of charging base 195 gms Preset temp 150 degree C, 180 degree C,200,230 degree Celsius Warm up time in 0.2 secs Power input AC100V-240V, 50/60 hz 0.8 mpower output DC 15 V-1.6A Battery capacity 2000mAh Charging time 2.5 hours Power adaptor AC or DC		Minimum 01 years warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
48	Tissue floatation water bath	Not in EEL thermostatically controlled Inner tank (Single piece) made of black anodized aluminum. Housing made of powder coated aluminum. Front panel consist of temperature regulator knob, mains switch and pilot lamp. Inside black and outer surface finished in powder paint, with wide enough rim for drying the wet slides. Thermal loss is prevented by filling mineral wool (Insulation) in space between the outer body & inner tank. Heating elements are positioned in the floor and side walls & a natural mixing of the medium arises, providing better temperature distribution in the bath. 0 OT emperature is controlled by imported capillary type thermostat, from ambient to 70 C with an accuracy of +2 C. The low 110mm depth of the bath allows comfortable and safe working. Supplied with Dust Guard Lid made of black anodized aluminum		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
49	Biochemistry analyser	Not in EEL OPTICAL SYSTEM : 0.000 to 3.000 absorbance units (A) : ± 2 % or 0.007 whichever is higher, from 0 to 1.5 A, ± 3 % from 1.5 A to 3.0 A : Photodiode : Tungsten halogen Linear measurement range Photometric accuracy Optical measurement Light Source FILTERS : 340, 405, 510, 546, 578, 630 nm & two optional Type of filter Wave length : High quality narrow band interference FLOW CELL : upto 1000 µl : 18 µl flow cell 10 mm square cuvette Round tube for coagulation Sipping volume Triple cuvette system ANALYSIS MODES : Absorbance, End oint, Fixed Time, Kinetic, Differential & Ratio : PT, INR & APTT Biochemistry & Turbidimetry Coagulation TOUCH PANEL / KEYPAD 5” Color Graphics LCD Human machine interface PRINTER : On board thermal 28 columns, maintenance free INCUBATOR 0 : 15 Positions at 37 C MEMORY : 250 tests : 2500 patient results : 30 QC results/test Open tests Results QC POWER : 100 VA : 115 – 230 Volts ± 10%, 50 – 60 Hz Wattage Voltage POWER : 100 VA : 115 – 230 Volts ± 10%, 50 – 60 Hz		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	2
50	Smart dentin grinder for hospital, 300W-500W	Not in EEL Wattage 300-500 W Hospital usage Operation mode – electricity and manual Portable Voltage – 250 v single phase Frequency 50 Hz Display – digital Automatic – semi automatic		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1
51	Tracheostomy kit	Not in EEL Tracheostomy tube of the same size as insitu (with introducer/obturator if applicable), One tracheostomy tube one size smaller (with introducer if applicable), Spare inner tubes for double lumen trache tubes (if applicable), Spare ties (cotton and/or Velcro), Scissors, Resuscitation bag and mask (appropriate size for patient), One way valve (community use only - for resuscitation), Wall or portable suction equipment, Appropriate size suction catheters, One Heat Moisture Exchanger filter (HME) or tracheostomy bib, Water based lubricant for tube changes, Mucous trap with suction catheter for emergency suction		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
52	Cleft Surgery kit	Not in EEL Dingman Mouth gag with 3 Blades - Adults - 1 Set, Cleft Palate Raspatory French Pattern - Left & Right - 1 Pair, Cleft Palate Raspatory Barsky Type Double Ended - 1 Pcs, Cleft Palate Raspatory - Curved Upward - 1 Pcs, Cleft Palate Raspatory - Curved Downward - 1 Pcs, McIndoe Cleft Palate Raspatory - Medium - 1 Pcs, Mitchell Trimmer Double Ended - 1 Pcs, Cleft Palate Knife - Triangular - 1 Pcs, Cleft Palate Hooks Right Angle & Down Bent - 2 Pcs, Cronin Cleft Palate Elevator - Medium Right - 1 Pcs, Palate Dissector - Left & Right - 1 Pair, Spoon Shape Elevator Double Ended - 1 Pcs, Miller Bone File Double Ended - Curved - 1 Pcs, Cleft Palate Elevator Double Ended - Angled To Left & Right - 1 Pcs, Cleft Palate Elevator Double Ended -Angled to Side & Flat - 1 Pcs, Howarth Elevator Double Ended - 1 Pcs, Kilner Skin Retractor - 1 Pcs, Aluminium Anodized Box - Small - 1 Pcs		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1
53	Trauma kit with bone plating kit	Not in EEL Wire Cutter - Regular -01 No Wire Twisting Forceps - Fine 01 No Wire Twisting Forceps - Thick 01 No Aluminum Instruments Box with Silicon Mat & Lock -01 No Self holding screw driver of different sizes Screw driver of different sizes Plate bender Plate cutter Bone holding forceps		Minimum 03 years warranty and 05 years AMC and CMC after warranty	US FDA OR EUROPEAN CE APPROVAL MANDATORY. ISO CERTIFICATION MANDATORY	1
54	Pedo extraction forceps sets	It should be Madeup of high grade stainless steel material and it should have seven sets of instruments.			ISO certificate or CE certificate	4
55	Rubber dam kit for pedo patients	Kit should consist of Dental Dam sheets, rubber dam punch , Clamps, forecep, frame, template and other accessories.			ISO certificate or CE certificate	2
56	Formalin chamber	It should be madeup of acrylic fiber, should have min 5mm of thickness, with a transparent and corrosion- proof surface. Size should be approx 20" x 8" x 8", 23" x 10" x 8", and 14" x 8" x 8". Chambers should have have three sliding trays to hold instruments, facilitating easy handling and organization.			ISO certificate or CE certificate	2
57	Typhodonts	Pedo Jaw set should be equipped with Metal Articulator and it should have Set of 24 teeth and Screw-in type with soft Gingival			ISO certificate or CE certificate	3
58	Adjustable mouth prop set A . infant mouth gag B. Child mouth gag C. Adult mouth	It should be madeup of High strength stainless Steel material and it should be reusable.			ISO certificate or CE certificate	3

S.No .	Equipment Name	SPECIFICATION	Accessories	Warranty and Maintenance	License and Certification	Required Quantity
59	Double beak pliers anterior and posterior	It should be madeup of High quality Stailless steel material with double beak pliers for forming bands in patient’s mouth			ISO certificate or CE certificate	2
60	Bracket Holding Pliers	It should be madeup of high-garde stainless steel material.			ISO certificate or CE certificate	3
61	Straight Hoe Pliers	It should be versatile utility plier with serrations, it should accommodated wires upto 0.030’’ (0.76mm) . Length should be approx 14cm and weight should be approx 35gms.			ISO certificate or CE certificate	3
62	Curved Hoe Pliers	It Should be 4mm curved and 14cm long Utility plier with round and serrated tips with 45 degree angle.			ISO certificate or CE certificate	2
63	Papoose Board for stabilization of childrens	It should be available in Small, Regular, Large size and Extra large size.			ISO certificate or CE certificate	2
64	Articulators – Semi adjustable	It should be equipped With Face-bow Bennett and Condylar Guidance, Incisal Guide and Orbitale Indicator HANAU Spring-Bow should be auto-centered detachable transfer clamp assembly and built-in reference pointer Arcon-type Bennett Angle should be 0 to 30 degrees Condylar Guidance Angle should be -20 to +60 degrees			ISO certificate or CE certificate	1
65	Airotor Handpieces	Handpiece should be equipped with LED technology providing powerful and consistent illumination during procedures. Should be Compatible with all electrically operated dental chairs. Air Rotor hand piece clean head with a speed of 350000 RPM. Supplies with Titanium/ SS Air rotor torque hand piece. Ultra push type non retraction valve. Airotor Head Type-Standard Number of Holes-2 Hole or more Power-17 W or above Head Size- 011.2 x H 12.6 mm Auto		Minimum 01 years warranty	ISO certificate or CE certificate	50
66	Straight Handpiece for clinical micromotor	It should be autoclavable. With • Stainless steel body • Handle convenience • Max speed of 0 - 40,000 rpm.		Minimum 01 years warranty	ISO certificate or CE certificate	20
67	Supra gingival scalers	It should have Triangular blade in cross-section with cutting edges on all three points.			ISO certificate or CE certificate	5
68	Sub gingival scaler	It should be madeup of High-quality stainless steel and feature a slim profile, sharp, pointed tips and should be available in variety of shapes and sizes.			ISO certificate or CE certificate	5
69	Surgery Trolleys	Dimensions should be around 1070 mm L x 610 mm W x 900 mm H.			ISO certificate or CE certificate	8