

	Item Name: Fowler Beds (Electric)
S. No.	Description
1.	Buffer to buffer (2160mm L x 1000mm W x 600mm H) $\pm 2\%$ Engineering tolerance (without mattress)
2.	Bed frame size (2080mm L x 900mm W), Mattress Platform size (1960mm L x 900mm W) $\pm 2\%$ Engineering tolerance.
3.	Bed frame should be made from 50mm x 30mm x 1.6mm (16G) thick ERW tube with proper support.
4.	Four sections of bed are fitted with CRCA sheet duly pre-treated and powder coated with uniform capsule perforation which allows easy breathing of mattress.
5.	Bed frame mounted on round tube of minimum 31.70mm diameter and 16G thickness, fitted with 125mm castors, 2 with brake and having rustproof body.
6.	Bed should be provided with a Patient Hand Remote.
7.	Functions Operated via Hand Remote: Backrest Elevation & Knee Rest Elevation of at least 70° & 40° respectively.
8.	The Bed should have polymer moulded Tuck Away Railing on both side assisted by gas spring should have angle indicator on side railings (inside as well as outside).
9.	The bed should have ABS/Polymer moulded head and foot panel with MS pipe inserted on both side of the panel, minimum dia. 32 x 1.2mm. The head and foot panel should be easily removable without use of any tool.
10.	Should have hooks on bed frame on both side for holding urine/drainage bags.
11.	Four Corner Buffers of 125mm dia. each
12.	All the mild steel components should be pre-treated for de greasing, de-rusting and phosphating in 7 tanks process. After proper pre-treatment, the mild steel components are to be epoxy powder coated and oven baked at temperature of 180-200 degree.
13.	Certifications
14.	The manufacturer must have ISO 13485, ISO 9001 from NABCB Accredited Certification Body. (To be enclosed with Technical Submitting Documents)
15.	Test Reports
16.	Impact Test and Bend Test from NABL Accredited Laboratory.
17.	Powder Coating Thickness of 60 microns or above from NABL Accredited Laboratory.
18.	Stainless Steel Material of 304 Grade from NABL Accredited Laboratory.
19.	Salt-Spray Test from NABL Accredited Laboratory.

	Item Name: Bedside Locker (MS)
S. No.	Description
1.	Overall Size: 410L×410W×840H mm.
2.	Drawer having inbuilt sliding channel connected from bottom (no ball bearing channels allowed). Should be provided with recess to serve as handle
3.	Locker box made from machine pressed 0.8mm CRCA M.S. Sheet. Should be provided with recess to serve as handle
4.	Fitted with single piece die pressed, Seamless, Stainless Steel top with raised edges on three sides.
5.	Open Storage Space should be provided between the Drawer and Cabinet
6.	Legs frame made of 25mm square × 1.2mm thick CRC tube fitted with 2 pcs PVC stumps/ Rubber shoe in front and 2 pcs wheel at rear.
7.	Locker should be provided with two buffers at the back side.
8.	Finishing & workmanship in the medical furniture is of prime importance and must be of high standard. All corners shall be rounded off so that there shall be no sharp corners and holes should be burr free.
9.	All Mild Steel parts should be In-house, pre-treated and Epoxy powder coated with coating thickness 50 to 100 microns.
10.	Certifications:
11.	Test Reports of the material should be from NABCB Approved Lab.
12.	The manufacturer must have ISO 13485, ISO 9001 from NABCB Accredited Certification Body. (To be enclosed with Technical Submitting Documents)
13.	Test Reports
14.	Impact Test and Bend Test from NABL Accredited Laboratory.
15.	Powder Coating Thickness of 60 microns or above from NABL Accredited Laboratory.
16.	Stainless Steel Material of 304 Grade from NABL Accredited Laboratory.
17.	Salt-Spray Test from NABL Accredited Laboratory.

	Item Name: Overbed Table/Cardiac Table (Gear Operated)
S. No.	Description
1.	Overall dimensions should be 1000mm L x 435mm W x 825mm – 1180mm H.
2.	The top should be made of ABS having size of 880L x 435W x 28H mm.
3.	Aluminium extruded telescopic section fitted with smooth gear mechanism for height adjustment
4.	It should have MS base frame covered with durable high impact ABS Base cover.
5.	For ease in mobility, it should have castors of 50mm dia.
6.	Finishing & workmanship in the medical furniture is of prime importance and must be of high standard. All corners shall be rounded off so that there shall be no sharp corners and holes should be burr free.
7.	Top: MDF top with membrane press, should give anti-scratch Property with good surface finish
8.	M.S tubular parts, linkages, flats are to be In-house, pre-treated and Epoxy powder coated with coating thickness 50 to 100 microns
9.	Certifications:
10.	The manufacturer must have ISO 13485, ISO 9001 from NABCB Accredited Certification Body. (To be enclosed with Technical Submitting Documents)
11.	Test Reports
12.	Impact Test and Bend Test from NABL Accredited Laboratory.
13.	Powder Coating Thickness of 60 microns or above from NABL Accredited Laboratory.
14.	Stainless Steel Material of 304 Grade from NABL Accredited Laboratory.
15.	Salt-Spray Test from NABL Accredited Laboratory.

	Item Name: Attendant Stool (Fixed)
S. No.	Description
1.	The Four legs made of 25mm x 1.8G CRCA Tube and should be fitted with PVC Stumps/Shoes.
2.	The Height should be fixed at 490mm.
3.	The Footrest support should made of MS Tube.
4.	The top of the stool should be made of Mild Steel CRCA Sheet with double bent.
5.	M.S. tubular parts, linkages, flats are to be pre-treated and Epoxy powder coated with coating thickness 50 to 100 microns.
6.	The Overall Size (Approx.) should be 350L x 350W x 490H mm.
7.	Certifications
8.	The manufacturer must have ISO 13485, ISO 9001 from NABCB Accredited Certification Body. (To be enclosed with Technical Submitting Documents)
9.	Test Reports
10.	Impact Test and Bend Test from NABL Accredited Laboratory.
11.	Powder Coating Thickness of 60 microns or above from NABL Accredited Laboratory.
12.	Stainless Steel Material of 304 Grade from NABL Accredited Laboratory.
13.	Salt-Spray Test from NABL Accredited Laboratory.

	Item Name: Foot Step
S. No.	Description
1.	The Frame should be made of 1" x 18G CRCA Tube fitted with PVC Stumps/Shoes.
2.	The Step should be made of M.S. Top fitted with Aluminium strips
3.	M.S. tubular parts, linkages, flats are to be pre-treated and Epoxy powder coated with coating thickness 50 to 100 microns.
4.	The Overall Size (Approx.) should be 600L x 505W x 450H mm.
5.	The height of First step should be 230 mm and second step should be 220mm.
6.	Certifications
7.	The manufacturer must have ISO 13485, ISO 9001 from NABCB Accredited Certification Body. (To be enclosed with Technical Submitting Documents)
8.	Test Reports
9.	Impact Test and Bend Test from NABL Accredited Laboratory.
10.	Powder Coating Thickness of 60 microns or above from NABL Accredited Laboratory.
11.	Salt-Spray Test from NABL Accredited Laboratory.

Minimum warranty on all offered products should be 3 years

Technical Specifications for Gammow/HAPO bags

1. It should be operated with foot pump and light weight automation unit that work with built in rechargeable battery.
2. It should have minimum ONE year warranty.
3. It should have carry rods for HAPO bag for mobility.
4. It should have a pre-set leakage valve or PRV.
5. It should have repair kit.
6. The supplier should have Technology Absorption Certificate issued by Ministry Of Defence-DRDO, DIITM-New Delhi.
7. The TOT license certificate should be attached while submitting the tender.
8. The supplier should be either a direct OEM for Gammow/HAPO bags or should be an authorized dealer duly certified by the OEM partner, who in turn is certified by DRDO.
9. The HAPO bag & HAPO automation unit should be supplied as per LATEST DEBEL-DRDO specification no. DEBEL/DP-26/14-2024 Issue 'A' issued in July 2024.

Technical Specifications of LED Screen:

Details:

Outdoor LED screen is required to Display Information	
Total Screen Size: Width 4m x Height 3m.	
Outdoor Screen should have Provision of Both Front as well as Back Service, Installation Design to be decided as per Site requirement. (Design approval to be taken before installation)	
Outdoor Screen should have IP65 Front as well as IP65 Back Protection from Dust and Rain. Whenever required IP65 Test Certificate for Model needs to be submitted.	
Cabinets used should be Light Weight Aluminum or Magnesium Cabinets with Weight 20Kg per sq. meter or less to put minimum load on Structure.	
LED screen should have Good Viewing angle of Horizontal: 160 x Vertical 120 Degree to cover Maximum Area in open.	
Product should have High Refresh rate of 3840 Hz, Contrast of 6000:1 and Bit Depth 14 bit or higher for Proper colors reproduction.	
Screen should have Brightness of Minimum 5000 nits with Auto Brightness feature to change Brightness as per Light Condition.	
LED screen should be supplied with Original OEM UHD/4K LED controller with inbuilt Digital Signage Media Player and Software to Design and Play Multiple Contents.	
Signage Software should have provision to design Content on Screen and play through Network. Division of Screen in multiple parts to play different videos and image contents, URL Inputs, Streaming etc.	
LED Cabinet, UHD LED Controller, Digital Signage software and Configuration Software should be of same brand.	
OEM Should have experience of installing same Model in Same or larger Size Outdoor LED Screens in India in Government Department. Necessary Details of Customer to be provided for all Technical Main Specifications should be mentioned on Website of OEM to Verify Product.	
OEM Should have Direct Office and Service Center in India.	
The Bidder Should have Registered Office in UT Ladakh from last 5 Years.	

Outdoor Data Wall – Width: 4000mm / 4m x Height: 3000mm / 3m		
Sr. no.	Features	Specifications Required
1	Pixel pitch	P6.9 Category i.e. 6.94mm or less
2	Pixel Configuration	SMD 3 in 1 LED Package
3	Unit Case material	Magnesium or Aluminum Cabinets (to prevent from rust)
4	Brightness Minimum After Calibration	Typ. 5000 nit or more
5	Brightness Uniformity	97% or more
6	Contrast Ratio	6,000 :1 or more
7	Viewing Angle – Horizontal	160° or more
8	Viewing Angle – Vertical	120° or more
9	Color Temperature	3,500 ~ 9,000
10	Bit Depth	Minimum 14bit or more
11	Power Consumption – Max	700 (W/m2) or less
12	Refresh Rate	3,840 Hz or more
13	Working Temperature	0°C or less
14		+50°C or more

15	Working Humidity	10%RH
16		90%RH
17	IP Rating Front	IP65
18	IP Rating Back	IP65
19	Cabinet Quality Standard	Factory Made Size:1000mm x 1000mm Weight: 20 Kg per sqm or less
20	Controller Inputs Ports	HDMI x1, DPx1, OPS Slot
21	LED Controller of Same Make as LED Screen	Required
22	Resolution Supported	3840x2160 @ 60HZ or better
23	Output Pixel	8.2 Million or more
24	Control Ports	RS232C Loop and LAN
25	LAN Output	Minimum 16 or As per OEM
26	Temperature Sensor	Required
27	Chrome Adjustment	Required
28	Original LED Configuration Software of LED Screen OEM	Required
29	Brightness Adjustment	Required
30	Scaling	Required
31	HDR	Required
32	Controller Quality Standard	IEC 62368-1:2018, UL 60950-1:2007, BIS
33	Inbuilt Digital Signage Media player & Software for designing and playing content on Screen	Required
34	Operating System	WebOS/ Tizen/ Windows
35	Installation Options	Pole Mount