

Technical Specifications/Compliance Document

Dressing and Plasters

S No.	Items	Specifications	Make and Model and Warranty	Technical Specifications Offered by the vendor	Compliance Yes/No	Remarks If Any (Deviations)
1.	Zig-Zag Cotton Roll - 500 GMS	1. Product Description: Absorbent Cotton Wool. 100% Medical Grade Cotton. 2. Purpose: For medical use. 3. Conformity to specification for absorbent cotton IS 16468 (Latest Revision). 4. Conformity to Pharmacopoeia or Standard I.P. 5. Type of Absorbent cotton wool: Zig Zag. 6. Sterility: Sterile. 7. Acidity/Alkalinity: Must pass pH testing to ensure it is neutral to skin. 8. Net Weight: 500 Grams. 9. Performance Requirements for Absorbent Cotton shall be as per IS 16468(Latest) Compliance to Medical Device Rules (MDR) 2017 as amended till date 10. Availability of valid medical device license/Registration number for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date 11. ISO:13485 (Latest) Certification 12. Additional voluntary certification available ISO, CDSO Availability of Test Report for each supplied batch/product as per Medical Device Rules .(MDR) 2017 as amended till date 13. Submission of all necessary				

		certifications, licenses and test reports to the buyer at the time of bid submission and/or along with supplies as per buyer requirement 14. Shelf life in months from the date of manufacture: 36 months. 15. Minimum shelf life of the product at the time of delivery to the consignee: 3/4 th of Total Shelf Life. 16. Agree to provide advance sample of the product for buyer's approval before commencement of supply in case of bidding: Yes. 17. Labelling: Must indicate manufacturer name, weight, sterile/non-sterile, and manufacturing/expiry date				
2.	Paraffin gauze (10cm x 30cm)	1. Should consist of open weave fabric of cotton or cotton-viscose blend, uniformly impregnated with white or yellow soft paraffin, intended for medical use 2. Material – Base fabric: cotton or cotton-viscose blend - Fabric Type: Leno Weave With Two Picks In Each Shed Material Of Warp And Weft Threads: Cotton -Impregnation: Medical-grade soft paraffin -Dressing shall be free from toxic or harmful substances 3. Construction – open weave structure - Uniform mesh distribution - Non-adherent surface 4. product Specification Paraffin Gauze Dressings Shall Conform The Requirements Specified In Table 1 Of IS 16291 (Latest Revision) • Threads Per Dm, Warp: ≥ 74 • Threads Per Dm, Weft: ≥ 80				

		- Weight Per Square Meter (G/M2) After Extracting Ether Soluble Substances: ≥ 39 Paraffin content - Normal: 175–300 g/m². - Sterility - sterile packing ready to use in sterile surgical field 5. Workmanship & Finish - Clean and free from foreign matter - Uniform impregnation of paraffin - Free from linting and frayed edges - Manufactured under hygienic conditions 6. Sterility-Dressing should be sterile -Sterilization shall comply with relevant BIS standards 7. Sizes-10cm*30cm -Tolerance +/- 5% 8. Packaging- packed in sterile packs that contain 10 sheets of paraffin guaze piece layered in between tissue paper to avoid sticking of guaze packs - Packaging shall maintain sterility and integrity during storage and transport				
3.	Sterile Guaze Pad (10 cm x 10cm)	1. Soft & Hygienic. Sterile ETO Gamjeepads have the absorbent core made of 100% absorbent cotton placed between the layers of chlorine free bleached cotton gauze, Sterile 10 cm x 10 cm Gamjee Pad 2. Purpose : Gamjeepads are medical dressings used for a variety of purposes, primarily related to wound care 3. Steility : Sterile 4. Usage : Single-Use 5. Material : Woven with cotton 6. Width ($\pm 5\%$) 10 cm, Length ($\pm 5\%$) 10 cm (Pad) Thickness 20 mm				

		7. Packaging Type of Packing : Moisture Proof Packing 8. Shelf Life Shelf life from the date of manufacture in months : 36 month 9. Labelling: Must indicate manufacturer name, weight, sterile/non-sterile, and manufacturing/expiry date				
4.	Sterile Guaze Pad (20 cm x 10cm)	1. Soft & Hygienic. Sterile ETO Gamjeepad have the absorbent core made of 100% absorbent cotton placed between the layers of chlorine free bleached cotton gauze, Sterile 20 Cm x10 cm Gamjee Pad 2. Purpose : Gamjee pads are medical dressings used for a variety of purposes, primarily related to wound care 3. Steility : Sterile 4. Usage : Single-Use 5. Material : Woven with cotton 6. Width (±5%) 20 cm, Length (±5%) 10 cm (Pad)Thickness 20 mm 7. Packaging Type of Packing : Moisture Proof Packing 8. Shelf Life Shelf life from the date of manufacture in months : 36 month 9. Labelling: Must indicate manufacturer name, weight, sterile/non-sterile, and manufacturing/expiry date				
5.	Sterile Gauze piece (10 x10 cm)	1. Scope - Supply of X-ray detectable gauze swabs for hospital use - Product must comply with IS 10829:1993 standards				

		2. Material • Absorbent cotton gauze 100% • Plain or leno weave • Bleached white, odourless, free from impurities, non-toxic • Peroxide bleached (Chlorine-free) 3. Construction • 12-ply folded gauze swab • No loose threads or frayed edges • Edges properly folded inside 4. X-ray Detectable Element • Integrated radiopaque thread • Contains $\geq 55\%$ barium sulphate • Non-toxic and securely fixed 5. Size • 10×10 cm • Number of plies 12 6. Physical Requirements • Ends per dm ≥ 104 • Picks per dm ≥ 80 • Weight ≥ 22 g/m² • Absorbency time ≤ 10 seconds 7. Chemical Requirements • Water soluble substances $\leq 0.5\%$ • Foreign matter $\leq 1\%$ • Minimal fluorescence • Residual starch absent 8. Sterility • Should be Sterile • Sterile products must meet IS 10150 standards 9. Packaging • Pack of 10 • Sterile packs must be sealed and tamper-proof • Packing & marking shall be done as per the provisions mentioned in IS 16950 (Latest Revision) 10. Labelling				
--	--	--	--	--	--	--

		• Product name and size • Number of plies • Batch number • Manufacturing and expiry date • Manufacturer details • Sterility label if applicable 11. Quality Assurance • Manufacturer must provide certificate of compliance • Compliance to Medical Rules (MDR) 2017 as amended till date 12. Shelf Life • Minimum 3–5 years for sterile products • Must be clearly mentioned on packaging				
Additional Features/ Specifications Offered by the Bidder				(1) (2) (3)		