

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	17-07-2026 20:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	17-07-2026 20:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Chhattisgarh
विभाग का नाम/Department Name	Health And Family Welfare Department Chhattisgarh
संगठन का नाम/Organisation Name	Chhattisgarh Medical Services Corporation Limited
कार्यालय का नाम/Office Name	Chhattisgarh Medical Services Corporation Limited
शिकायत निवारण के संपर्क विवरण/ Contact details of Grievance redressal	HOD Email id :md.cgmsc@gov.in Buyer Email id: bme01cgmscl@gmail.com
कुल मात्रा/Total Quantity	125
वस्तु श्रेणी /Item Category	Refrigerated Centrifuge For General And Research Purpose (Q3) , Gel Documentation System (Q2) , water baths (Q3) , Biosafety Cabinet (Q2) , Medical Centrifuge (Q3) , Deep freezer (Q2) , Autoclave-1 , Autoclave-2 , BOD Incubator-1 , BOD Incubator-2 , Data Logger -80C, -20C , Data Logger - 80c, -20c, plus 4c 2no , Elisa plate reader-1 , ELISA Plate Reader-2 , Elisa Plate Reader-3 , Elisa plate washer-1 , ELISA Plate Washer-2 , Elisa Plate Washer-3 , Floater Rack , Gel Electrophoresis , Gel electrophoresis and documentation unit , Gel Electrophoresis andDocumentation Unit , Laminar Air Vertical Flow System , Ice Buckets , Ice flaker machine , Label Printer , Label Printer Brady Printer , Laboratory Incubator , Laboratory Rocker and Shaker , Laminar Air Flow Horizontal , Laminar air vertical flow 2 feet , Magnetic Stirrer-1 , Magnetic stirrer-2 , Magnetic Stirrer-3 , Micro pipette 0.5 - 10 ul , Micro pipette 10 - 100 ul , Micro pipette 100 - 1000 ul , Micro pipette 2 - 20 ul , Micro pipette 20 - 200 ul , Micropipettes , Microwave Oven For Agarose gel , Microwave Oven For agarose gels , Multi Channerl 30 - 300 ul , Multifunctional copier or printer or scannerFor office work , Pan electronic weighing balance-1 , Pan Electronic Weighing Balance-2 , PCR Workstation , PH Meter , Quick Spin , Quick Spinner-1 , Quick Spinner-2 , Quickspin for 1.5, 2ml Tubes. PCR Tubes and Plates , Refrigerated Micro centrifuge , Refrigerated Microfuge , Refrigerator , Refrigerator sample receiving , Themo Cycler-3 , Themocycler-2 , Thermo cycler-1 , Timer-1 , Timer-2 , Ultrapure Water System , Unichannel and Multichannel Pipettes , UV Dosimeter-1 , UV Dosimeter-2 , UV Dosimeter- 3 , Vortex Mixer-1 , Vortex Mixer-2

बिड विवरण/Bid Details	
GeMARPTS में खोजी गई स्ट्रिंग्स / Searched Strings used in GeMARPTS	Auloclave-1, Autoclave-2, BOD Incubator-1, BOD Incubator-2, Data Logger (-80C, -20C), Data Logger (-80c, -20c, +4c) (2no), Elisa plate reader-1, ELISA Plate Reader-2, Elisa Plate Reader-3, Elisa plate washer-1, ELISA Plate Washer-2, Elisa Plate Washer-3, Floater Rack, Gel Electrophoresis, Gel electrophoresis & documentation unit , Gel Electrophoresis & Documentation Unit, Laminar Air Vertical Flow System, Ice Buckets, Ice flaker machine, Label Printer, Label Printer (Brady Printer), Laboratory Incubator, Laboratory Rocker and Shaker, Laminar Air Flow (Horizontal), Laminar air vertical flow 2 feet, Magnetic Stirrer-1, Magnetic stirrer-2, Magnetic Stirrer-3, Micro pipette 0.5 - 10 ul, Micro pipette 10 - 100 ul, Micro pipette 100 - 1000 ul, Micro pipette 2 - 20 ul, Micro pipette 20 - 200 ul, Micropipettes , Microwave Oven (For Agarose gel), Microwave Oven (For agarose gels), Multi Channerl 30 - 300 ul , Multifunctional copier/printer/scanner(For office work), Pan electronic weighing balance-1, Pan Electronic Weighing Balance-2, PCR Workstation, PH Meter, Quick Spin, Quick Spinner-1, Quick Spinner-2, Quickspin for 1.5, 2ml Tubes. PCR Tubes and Plates, Refrigerated Micro centrifuge, Refrigerated Microfuge, Refrigerator, Refrigerator, Themo Cycler-3, Themocycler-2, Thermo cycler-1, Timer-1, Timer-2, Ultrapure Water System, Unichannel & Multichannel Pipettes, UV Dosimeter-1, UV Dosimeter-2, UV Dosimeter-3, Vortex Mixer-1, Vortex Mixer-2
	Searched String: Auloclave-1 1, 1, 3 - Triethoxybutane, Nitrile Coated/Nitrile Hand Gloves (V2), BCAA 2:1:1 Powder, BCAA 8:1:1 Powder, Multipara Monitor - Low End, 1-Bromohexadecane, Cypermethrin 1% Chalk, BOD Incubator, Cypermethrin 1% Dust, High End Desktop Computer Searched String: Autoclave-2 Dental Autoclave, Laminar air flow cabinets or stations, Vertical Autoclave, Autoclave (ICMR Specs), Autoclave (Horizontal), Sofa (V3), Horizontal Autoclave, Vertical Autoclave (FSSAI), Table Top Autoclave, Hydrothermal Autoclave Reactor Searched String: BOD Incubator-1 BOD Incubator, Nitrile Coated/Nitrile Hand Gloves (V2), BOD Incubator (V2), BOD Standard Solution, Laboratory Incubators, Egg Incubator, Incubator (FSSAI), Multipara Monitor - Low End, CO2 Incubator, High End Desktop Computer Searched String: BOD Incubator-2 BOD Incubator, Laboratory Incubators, BOD Incubator (V2), BOD Standard Solution, Laminar air flow cabinets or stations, Neonatal Incubator, Sofa (V3), Egg Incubator, XLPE Cable for Working Voltage up to and Including 1.1 KV Marked To IS 7098 (Part 1), CO2 Incubator Searched String: Data Logger (-80C, -20C) Digital Data Logger, Muffle/Box Furnace (V2), Data Logger Unit Storage, SHIRO VASTI CAP, Stainless Steel Dining Table

बिड विवरण / Bid Details	With Fixed Chair, Cabinet Dryer (V2), PVC Insulated Unsheathed Cable with Flexible Conductor ISI Marked To IS 694, V - Belts - Endless V - Belt for Industrial Purposes as per IS 2494, Projector Screen (V2), Open Jaw Wrenches (Spanners) as per IS 2028 Searched String: Data Logger (-80c, -20c, +4c) (2no) Muffle/Box Furnace (V2), Data Logger Unit Storage, Stainless Steel Dining Table With Fixed Chair, PVC Insulated Unsheathed Cable with Flexible Conductor ISI Marked To IS 694, V - Belts - Endless V - Belt for Industrial Purposes as per IS 2494, Projector Screen (V2), Open Jaw Wrenches (Spanners) as per IS 2028, C Arm Fluoroscope X - Ray Machine (V2), Intraocular Lenses (IOL), Correspondence Envelopes (V3) Conforming to IS 13892 Searched String: Elisa plate reader-1 ELISA Reader (V2), Ratchet Handle as per IS 7975, Elisa Reader (ICMR), ELISA reader and washer, ELISA Reader and Washer - Semi Automated, ELISA Plate Washer (V2), Dengue ELISA Test Kit, L 1 Fingerprint Registered Device Scanners for Bio - Metric Authentication (UIDAI) (V2), Plate Loaded T-Bar Row Machine, Mild Steel Sheet Searched String: ELISA Plate Reader-2 ELISA Reader (V2), Locking Compression Plates (V2), Elisa Reader (ICMR), ELISA reader and washer, Rubberized Weight Plates, ELISA Reader and Washer - Semi Automated, Non-Locking Bone Plate, ELISA Plate Washer (V2), Lat Pull Down (Version 2), Mild Steel Sheet Searched String: Elisa Plate Reader-3 ELISA Reader (V2), Locking Compression Plates (V2), Elisa Reader (ICMR), ELISA reader and washer, Chapati Plate with Puffer (V2), ELISA Reader and Washer - Semi Automated, ELISA Plate Washer (V2), Vitreous China Sanitary Appliances - Squatting Pans conforming to IS 2556 (Part 3), Non-Locking Bone Plate, Ratchet Handle as per IS 7975 Searched String: Elisa plate washer-1 ELISA Plate Washer (V2), High Pressure Washer, ELISA reader and washer, ELISA Reader and Washer - Semi Automated, Ratchet Handle as per IS 7975, Mild Steel Sheet, Dengue ELISA Test Kit, Plate Loaded T-Bar Row Machine, HIV ELISA Test Kits, Leaf Plate Making Machine Searched String: ELISA Plate Washer-2 ELISA Plate Washer (V2), Locking Compression Plates (V2), ELISA reader and washer, ELISA Reader and Washer - Semi Automated, Rubberized Weight Plates, Lat Pull Down (Version 2), Mild Steel Sheet, Non-Locking Bone Plate, Ratchet Handle as per IS 7975, Chapati Plate with Puffer (V2) Searched String: Elisa Plate Washer-3 ELISA Plate Washer (V2), Locking Compression Plates (V2), ELISA reader and washer, ELISA Reader and Washer - Semi Automated, Chapati Plate with Puffer (V2), Vitreous China Sanitary Appliances - Squatting Pans conforming to IS 2556 (Part 3), Non-Locking Bone Plate, Ratchet Handle as per IS
-------------------------	---

बिड विवरण / Bid Details	7975 Corrugated Boxes, High Pressure Washer
	Searched String: Floater Rack Dumbbell Rack, Metal Shelving Racks (Adjustable Type) conforming To Is 1883 (V2), Bathroom Rack, Networking / Server Rack, DRUM STACKING RACK, Wooden Shelf / Case / Rack / Credenza / Modular Storage (V2), Dumbbell Rack (V3), Servers (V2), Power Rack (V2), Heavy Duty Storage Racks Searched String: Gel Electrophoresis Horizontal Electrophoresis System, Blood Collection Tube, Vertical Electrophoresis System, Horizontal Electrophoresis System (V2), Vertical Electrophoresis System (V2), Air Freshener Solid or Gel (V2), Cockroach Gel, Ultrasound Gel, Cockroach Gel Fipronil, Gel Documentation System Searched String: Gel electrophoresis & documentation unit Gel Documentation System Searched String: Gel Electrophoresis & Documentation Unit Gel Documentation System Searched String: Laminar Air Vertical Flow System Laminar Air Flow (Horizontal) (FSSAI), Laminar air flow cabinets or stations, Laminar air flow cabinets or stations (V2), Water Purification System with UV (Ultra-Violet Process) or UV (Ultra-Violet Process)+UF (Ultra-Filtration Process) (V2), Vertical Electrophoresis System, Vertical Electrophoresis System (V2), Electronic PD Flow Metering System, Portable Electric Indoor Air Purifier Conforming To IS 17531, Vertical Garden Stands, Air Walker / Stroller - Outdoor Gym Equipment Searched String: Ice Buckets ice packs, Bucket Mop Wringer Trolley, Fire Buckets Stand (MRPL), Plastic Ice Box (V2), Ice Cube Machine, Ice Lined Refrigerator, Ice Piton'F' (Defence), Ice Hockey Skates, Ice Flake Making Machine, Ice Piton'G'(Defence) Searched String: Ice flaker machine Ice Cube Machine, Ice Flake Making Machine, ice packs, Sanitary Napkins Vending Machines, Ice Lined Refrigerator, High Vacuum Suction Machine (V2), Ice Hockey Skates, Sanitary Napkins Incinerator Machine with Smoke Control Unit, Ice Piton'G'(Defence), Roti Making Machine Searched String: Label Printer Label Printer / Barcode Printer, A3 Size Multifunction Printer (MFP), RFID Label Printer with Encoder, Thermal Ribbons for Label/Barcode Printer, Plotter Paper Roll (V2), Handmade paper Label, Ferrule Printer, Labels Stickers (V2), emblem label of govt of india, Multifunction Printer (Print, Scan, Copy) Searched String: Label Printer (Brady Printer) Label Printer / Barcode Printer, A3 Size Multifunction Printer

बिड विवरण / Bid Details	(MFP), RFID Label Printer with Encoder, Thermal Ribbons for Label/Barcode Printer, Plotter Paper Roll (V2), Ferrule
	Printer, Multifunction Printer (Print, Scan, Copy), Receipt Printer (V3), Computer Printer (V2), Industrial composite color 3D Printer Searched String: Laboratory Incubator Laboratory Incubators, Egg Incubator, Laboratory Steel Furniture, BOD Incubator, Ultra low temperature laboratory deep freezer, Incubator (FSSAI), CO2 Incubator, Shaker Incubator, Laboratory Graduated Cylinders as per IS 878, Transport Incubator Searched String: Laboratory Rocker and Shaker Sieve Shaker, Bar and gauge apparatus for laboratory, Plastic Rocker (Preschool Kit), Laboratory Glassware Washers, Laboratory Glassware - Burettes (V2) as per IS 1997, Test Sieves as per IS 460 (Part 1 and Part 2), Shaker Incubator, Glass Rods and Tubing for Laboratory Glassware confirming to IS 7374 (V3), Laboratory Incubators, Glass Rods and Tubing for Laboratory Glassware confirming to IS 7374 (V2) Searched String: Laminar Air Flow (Horizontal) Laminar Air Flow (Horizontal) (FSSAI), Laminar air flow cabinets or stations, Laminar air flow cabinets or stations (V2), Electronic PD Flow Metering System, Multiple Split System Air Conditioners and Air to Air Heat Pumps (VRF Air Conditioners) Conforming to IS 18728 Searched String: Laminar air vertical flow 2 feet Laminar Air Flow (Horizontal) (FSSAI), Laminar air flow cabinets or stations, Laminar air flow cabinets or stations (V2) Searched String: Magnetic Stirrer-1 magnetic stirrers, Nitrile Coated/Nitrile Hand Gloves (V2), Laboratory Stirrer, Magnetic Stand, High speed stirrer, Multipara Monitor - Low End, Thermal Conductivity Apparatus for liquid, Magnetic Compass, BOD Incubator, Magnetic Stirring Bar Searched String: Magnetic stirrer-2 magnetic stirrers, Laminar air flow cabinets or stations, Laboratory Stirrer, High speed stirrer, Dry Erase Writing Boards (V3), Magnetic Compass, Sofa (V3), Magnetic Stand, Permanent Magnetic Lifter, XLPE Cable for Working Voltage up to and Including 1.1 KV Marked To IS 7098 (Part 1) Searched String: Magnetic Stirrer-3 magnetic stirrers, High End Desktop Computer, Laboratory Stirrer, High speed stirrer, Entry and Mid Level Desktop Computer, 3 T MRI Machine, Interactive Panels with CPU, Magnetic Stand, XLPE Cable for Working Voltage up to and Including 1.1 KV Marked To IS 7098 (Part 1), Magnetic Compass Searched String: Micro pipette 0.5 - 10 ul Micro/Semi-Micro/Analytical weighing balance, Whistle (V2),

बिड विवरण / Bid Details	
GeMARPTS में खोजा गया परिणाम / Searched Result generated in GeMARPTS	Foreign body (metal) detector for veterinary use, Line Interactive UPS with AVR (V2), Respirable Dust Sampler as per IS 5182, Analog-Digital Circuits Development Platform, Mechanically Woven, Double - Twisted, Hexagonal Wire Mesh Gabions, Revet Mattresses and Rock Fall Netting as per IS 16014, Automatic Fire Ball for Fire - Extinguishing, PP Plastic Woven Non - Laminated Bags (for Seed Packaging) (V2)
	Searched String: Micro pipette 10 - 100 ul Universal Pipette Tips (V2), Online UPS (V2), Manual Single Channel Air Displacement Pipettors Or Micropipette (V2), Whistle (V2), PT 100 DUPLEX RTD (Air Circuit) (BHEL), Respirable Dust Sampler as per IS 5182, Heavy duty diesel ultra oil, Micro/Semi-Micro/Analytical weighing balance, Anti TB Drugs - Clofazimine 100 mg Capsules, Mechanically Woven, Double - Twisted, Hexagonal Wire Mesh Gabions, Revet Mattresses and Rock Fall Netting as per IS 16014
	Searched String: Micro pipette 100 - 1000 ul Universal Pipette Tips (V2), Hot air oven, Manual Single Channel Air Displacement Pipettors Or Micropipette (V2), Whistle (V2), Porcelain Insulators For Overhead Power Lines with Nominal Voltage Greater Than 1000 V Conforming To IS 731, Motorized Pipette Controller, Porcelain Insulators For Overhead Power Lines with Nominal Voltage up to and Including 1000 V Conforming To IS 1445, Light Portable Fire Monitor (V2), Alphacypermethrin 10.00% Suspension Concentrate (SC), Salt Spray Chamber
	Searched String: Micro pipette 2 - 20 ul Universal Pipette Tips (V2), Bucket Mop Wringer Trolley, Manual Single Channel Air Displacement Pipettors Or Micropipette (V2), Whistle (V2), Ammonium Sulphate, Flying Insect Control Traps - Fly Catcher, Combination therapy with 2 channel electrotherapy, vaccum and dual frequency ultrasound, Micro/Semi-Micro/Analytical weighing balance, TERA Package 2 - Bionic Inspired Flower (ASoSE), Online UPS (V2)
	Searched String: Micro pipette 20 - 200 ul Universal Pipette Tips (V2), Digital pH Meter, Manual Single Channel Air Displacement Pipettors Or Micropipette (V2), Whistle (V2), X Ray Machine (200 mA), Online UPS (V2), Distribution Transformer, 3 Phase, upto and Including 200 KVA ISI Marked To IS 1180 (Part 1), Distribution Transformer, 3 Phase, 6.6 kV, >200 KVA upto and Including 2500 KVA ISI Marked To IS 1180 (Part 1), Micro/Semi-Micro/Analytical weighing balance, Distribution Transformer, 3 Phase, 33 KV, >200 KVA upto and Including 2500 KVA ISI Marked To IS 1180 (Part 1) Searched String: Micropipettes Manual Single Channel Air Displacement Pipettors Or Micropipette (V2) Searched String: Microwave Oven (For Agarose gel) Microwave Ovens For Household And Similar Purposes As Per Is 11676 (V2), Atulya Microwave Sterilizer for (COVID - 19), Carry Oven for Welding Electrodes, Hot Air Oven (FSSAI), Digital Electronic Spirit Level (Digi level or Digital

बिड विवरण/Bid Details	inclinometer), Ice Gel Pack for Medical use, Fire safety first aid Water Gel Blankets, Microwave Disinfectant system for disposal of Hospital waste (Bio-medical waste) - (MHA Spec)
	Searched String: Microwave Oven (For agarose gels) Microwave Ovens For Household And Similar Purposes As Per Is 11676 (V2), Atulya Microwave Sterilizer for (COVID - 19), Carry Oven for Welding Electrodes, Hot Air Oven (FSSAI), Digital Electronic Spirit Level (Digi level or Digital inclinometer), Microwave Disinfectant system for disposal of Hospital waste (Bio-medical waste) - (MHA Spec), Fire safety first aid Water Gel Blankets Searched String: Multi Channerl 30 - 300 ul Anti TB Drugs - Isoniazid 300 mg Tablet, Online UPS (V2), 300mA X - Ray Machine (V2), X-Ray Detectable Absorbent Cotton Gauze Conforming to IS 16950, Topramezone 10g/l +Atrazine 300g/l SC (Suspension Concentrate), Hand Held Brush Cutter (V2), Topramezone 10g/l + Atrazine 300g/l SC (Suspension Concentrate), X-Ray Detectable Gauze Swab/Laparotomy Sponge (V2), Azoxystrobin 11.5% + Mancozeb 30.0% WP (Wettable Powder), Wind sock Searched String: Multifunctional copier/printer/scanner(For office work) Photocopier Trolley, Multipurpose Counter Machine Stickers for Post Office, A4 and Legal Size Multifunction Printer (MFP), Steel Filing Cabinets for General Office Purpose conforming to IS 3313 (V2), A3 Size Multifunction Printer (MFP), Chair for General Purpose, Multifunction Printer (Print, Scan, Copy) Searched String: Pan electronic weighing balance-1 Electronic Weighing Scales, Micro/Semi-Micro/Analytical weighing balance, Electronic Baby Weighing Scale - RBSK, Colony counter electronic balance, Two Pan Laboratory Balance Analytical Chemical Physical as per IS 9440, Electronic Lectern, Package No. 1 - ATAL TINKERING LAB OF NITI AAYOG ELECTRONICS DEVELOPMENT, ROBOTICS, INTERNET OF THINGS AND SENSORS, Analytical Semi - Micro Balances, Electronic Weigh Bridge (Without Civil Work), Electronic dosimeter Searched String: Pan Electronic Weighing Balance-2 Electronic Weighing Scales, Micro/Semi-Micro/Analytical weighing balance, Electronic Baby Weighing Scale - RBSK, Colony counter electronic balance, Electronic Weigh Bridge (Without Civil Work), Two Pan Laboratory Balance Analytical Chemical Physical as per IS 9440, Low Voltage MCCB with Thermal Magnetic/Electronic Trip Unit Conforming To IS/IEC 60947 (Part 2), Surveillance Drone / Unmanned Aerial Vehicle (V2) (Defence Forces - Navy, Army and Air Force), Moulded Case Circuit Breakers (MCCB) Conforming To IS/IEC 60947 (Part 2), Electronic component cleaner (co2) Searched String: PCR Workstation HEPA Workstation, Fixed Computer Workstation, PCR Tubes, PCR Mini Cooler, Anaesthesia Workstation (V2), Mobile Computer Workstation, Automated Real Time Micro PCR Machine, Anaesthesia Workstation (V3), PCR Master Mix, Digital Forensic Workstation

बिड विवरण/Bid Details	Searched String: PH Meter
	pH Meter, Digital pH Meter, Ph Meter Rod, Thermal Ribbon, pH Paper, true rms clamp meter, pH Electrode, digital conductivity meter, Multiparameter pH or mV or ISE or Temp Meter, Water Quality Meters / Analyzers Searched String: Quick Spin Spin Bike, ELEMENT SPIN ON OIL, Quick Lime Burnt, Domestic Wiper/Wiper Mopper (V3), Burnt Quick Lime as per NALCO specifications, HYDRAULIC FLAT FACE QUICK COUPLINGS, Minifuge, Exercise Bike, Spinal Needle Conforming to IS 7350, Stationary Exercise Bike (V2) Searched String: Quick Spinner-1 Quick Lime For Chemical Industries (V2) Conforming to IS 1540 Part 1, Nitrile Coated/Nitrile Hand Gloves (V2), Quick Lime Burnt, Burnt Quick Lime as per NALCO specifications, Multipara Monitor - Low End, HYDRAULIC FLAT FACE QUICK COUPLINGS, Hose connector, 1, 1, 3 - Triethoxybutane, BOD Incubator, BCAA 2:1:1 Powder Searched String: Quick Spinner-2 Quick Lime Burnt, Laminar air flow cabinets or stations, Burnt Quick Lime as per NALCO specifications, HYDRAULIC FLAT FACE QUICK COUPLINGS, Sofa (V3), 2-2-BIPYRIDYL, Blanket for AC Coach Passengers (V2) Indian Railways, Quick Lime For Chemical Industries (V2) Conforming to IS 1540 Part 1, XLPE Cable for Working Voltage up to and Including 1.1 KV Marked To IS 7098 (Part 1), 2-Phenoxyethanol Searched String: Quickspin for 1.5, 2ml Tubes. PCR Tubes and Plates Wrought Aluminium and Aluminium Alloy Bars, Rods,Tubes, Sections, Plates and Sheets for Electrical Application - IS 5082, XLPE Cable for Working Voltage up to and Including 1.1 KV Marked To IS 7098 (Part 1), PCR Mini Cooler, Leather Gauntlets and Mittens (V2) Conforming to IS 2573, Minifuge, Automotive Vehicles - Tubes for Pneumatic Tyres (V2) as per IS 13098, Ultra Violet (UV) Cabinet for Sterilization, Real Time PCR Kits (V2), Welding Electrode - Covered Arc Welding Electrode - Rutile Type (BEML) Searched String: Refrigerated Micro centrifuge Refrigerated Micro - Centrifuge, Refrigerated Centrifuge For General And Research Purpose, Refrigerated Centrifuge (FSSAI), Non Refrigerated Centrifuge for General Purpose and research, Refrigerated Blood Bag Component Centrifuge, Medical Centrifuge, Ice Lined Refrigerator, Laboratory Refrigerator (V2), Centrifuge Tubes (V2), Analytical Semi - Micro Balances Searched String: Refrigerated Microfuge Refrigerated Micro - Centrifuge, Refrigerated Centrifuge For General And Research Purpose, Refrigerated Centrifuge (FSSAI), Non Refrigerated Centrifuge for General Purpose and research, Refrigerated Blood Bag Component Centrifuge, Ice Lined Refrigerator, microcentrifuge tubes, Laboratory Refrigerator (V2), Laboratory Refrigerator, Centrifuge Tubes (V2)

बिड विवरण/Bid Details	Searched String: Refrigerator
	Laboratory Refrigerator, Refrigerated Centrifuge For General And Research Purpose, Laboratory Refrigerator (V3), Ice Lined Refrigerator, Non Refrigerated Centrifuge for General Purpose and research, Laboratory Refrigerator (V2), Ice Lined Refrigerator (V2), Refrigerator Truck for Vaccine Transport, Frost Free Refrigerator Training Jig(DVET), Reverse Osmosis based Water Treatment System above 50 LPH Capacity (V2) Searched String: Refrigerator Laboratory Refrigerator, Refrigerated Centrifuge For General And Research Purpose, Laboratory Refrigerator (V3), Ice Lined Refrigerator, Non Refrigerated Centrifuge for General Purpose and research, Laboratory Refrigerator (V2), Ice Lined Refrigerator (V2), Refrigerator Truck for Vaccine Transport, Frost Free Refrigerator Training Jig(DVET), Reverse Osmosis based Water Treatment System above 50 LPH Capacity (V2) Searched String: Themo Cyclor-3 Qualitative PCR Machine (Thermal Cyclor), High End Desktop Computer, Omega 3 Capsules, (3-Aminopropyl) triethoxysilane, Entry and Mid Level Desktop Computer, Layer - 3 Distribution Switch, Layer 3 - Access Switch, Interactive Panels with CPU, Layer - 3 Core Switch, XLPE Cable for Working Voltage up to and Including 1.1 KV Marked To IS 7098 (Part 1) Searched String: Themocycler-2 2-2-BIPYRIDYL, Laminar air flow cabinets or stations, 2-Phenoxyethanol, 2 - Nitrobenzaldehyde, Sofa (V3), 2-Naphthol, 2-Chlorobenzaldehyde, XLPE Cable for Working Voltage up to and Including 1.1 KV Marked To IS 7098 (Part 1), 5-5 Dithiobis -2-2 Nitrobenzoic Acid, Compostable Carry Bags Searched String: Thermo cyclor-1 Digital Thermo Hygrometer, Nitrile Coated/Nitrile Hand Gloves (V2), Qualitative PCR Machine (Thermal Cyclor), Radiation survey meter, 1, 1, 3 - Triethoxybutane, Multipara Monitor - Low End, BCAA 2:1:1 Powder, Electronic dosimeter, BCAA 8:1:1 Powder, BOD Incubator Searched String: Timer-1 Digital Timer, Nitrile Coated/Nitrile Hand Gloves (V2), Stopwatch (V2), Auto Drain Trap Valve- Solenoid operated (To drain Water) for Oxygen Concentrator, Aluminium Alloy Timer Pulley to Suit the Timer Belt (Defence), Multipara Monitor - Low End, Astronomical Timer for Street Light Control (Timer Controls) as per IS 9000, 1, 1, 3 - Triethoxybutane, BOD Incubator, Reciprocal Shakers - Mechanical Shaker Machine with Timer Searched String: Timer-2 Digital Timer, Laminar air flow cabinets or stations, Stopwatch (V2), Astronomical Timer for Street Light Control (Timer Controls) as per IS 9000, Aluminium Alloy Timer Pulley to Suit the Timer Belt (Defence), Sofa (V3), XLPE Cable for Working Voltage up to and Including 1.1 KV

बिड विवरण / Bid Details	Marked To IS 7098 (Part 1), 2-2-BIPYRIDYL, Desk and Bench Set for Classroom/Training Area, Reciprocal Shakers - Mechanical Shaker Machine with Timer
	Searched String: Ultrapure Water System Smart Water Management Control System, water baths, High quality water purification system, Reagent Grade Distilled water, Solar Water Heating System as per IS 12976:1990, Reverse Osmosis based Water Treatment System above 50 LPH Capacity (V2), Potable Water Ionization Machine, Water Purification System with UV (Ultra-Violet Process) or UV (Ultra-Violet Process)+UF (Ultra-Filtration Process) (V2), Environmental Chamber, Solar Photovoltaic Water Pumping Systems (V2) as per MNRE Specification Searched String: Unichannel & Multichannel Pipettes Serological Pipettes and Graduated Pipettes as per IS 4364, Volumetric Pipettes as per IS 1117, Digital Physiograph System, Manual Single Channel Air Displacement Pipettors Or Micropipette (V2), Motorized Pipette Controller, Erythrocyte Sedimentation Rate (ESR) Pipette Searched String: UV Dosimeter-1 Electronic dosimeter, Nitrile Coated/Nitrile Hand Gloves (V2), transilluminators uv, Chemical Splash Proof Goggle, UV Light, Multipara Monitor - Low End, PC Based Thermoluminescence Dosimeter (TLD) Reader, Ultra Violet (UV) Cabinet for Sterilization, UV - Visible Spectrophotometer (FSSAI), BOD Incubator Searched String: UV Dosimeter-2 Electronic dosimeter, Laminar air flow cabinets or stations, transilluminators uv, Ultra Violet (UV) Cabinet for Sterilization, UV Light, Sofa (V3), PC Based Thermoluminescence Dosimeter (TLD) Reader, XLPE Cable for Working Voltage up to and Including 1.1 KV Marked To IS 7098 (Part 1), UV - Visible Spectrophotometer (FSSAI), Desk and Bench Set for Classroom/Training Area Searched String: UV Dosimeter-3 Electronic dosimeter, High End Desktop Computer, transilluminators uv, Ultra Violet (UV) Cabinet for Sterilization, UV Light, Entry and Mid Level Desktop Computer, PC Based Thermoluminescence Dosimeter (TLD) Reader, UV - Visible Spectrophotometer (FSSAI), Interactive Panels with CPU, UV - C Disinfection System (V2) Searched String: Vortex Mixer-1 vortex mixers, Nitrile Coated/Nitrile Hand Gloves (V2), Vortex Flow meter, Ribbon Blender Mixer, Multipara Monitor - Low End, Audio Mixer (V2), Spotting Scope, Self Propelled Auto Batch Mixer, BOD Incubator, Lab Multi Sample Mixer Searched String: Vortex Mixer-2 vortex mixers, Laminar air flow cabinets or stations, Vortex Flow meter, High Capacity Portable Pump For Fire Fighting Conforming To IS 12717, Audio Mixer (V2), Sofa (V3), Ribbon Blender Mixer, Spotting Scope, Self Propelled Auto Batch Mixer, XLPE Cable for Working Voltage up to and

बिड विवरण / Bid Details Including 1.1 KV Marked To IS 7098 (Part 1)	
अधिसूचना के लिए चयनित प्रासंगिक श्रेणियाँ / Relevant Categories selected for notification	• Refrigerated Centrifuge For General And Research Purpose • Gel Documentation System • water baths
बीओक्यू शीर्षक / BOQ Title	VRDL MERGED RETENDER
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) / Minimum Average Annual Turnover of the bidder (For 3 Years)	96 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का) / OEM Average Turnover (Last 3 Years)	192 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष / Years of Past Experience Required for same/similar service	3 Year (s)
वर्षों के अनुभव एवं टर्नओवर से एमएसई को छूट प्राप्त है / MSE Relaxation for Years Of Experience and Turnover	Yes Complete
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Relaxation for Years Of Experience and Turnover	Yes Complete
विक्रेता से मांगे गए दस्तावेज़ / Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC),Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेन् है/ Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	1
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	3
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
विगत प्रदर्शन / Past Performance	40 %

बिड विवरण/Bid Details	
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	No
बिड का प्रकार/Type of Bid	Two Packet Bid
प्राथमिक उत्पाद श्रेणी/Primary product category	Biosafety Cabinet
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Payment Timelines	Payments shall be made to the Seller within 45 days of issue of consignee receipt-cum-acceptance certificate (CRAC) and on-line submission of bills (This is in supersession of 10 days time as provided in clause 12 of GeM GTC)
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	HDFC Bank
ईएमडी राशि/EMD Amount	192113

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	HDFC Bank
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	3.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	42

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने है। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

(c).ईएमडी और संपादन जमानत राशि लाभार्थी के पक्ष में होनी चाहिए। / Earnest Money Deposit (EMD) shall also be accepted by the buyer in the form of a surety bond.

लाभार्थी /Beneficiary :

Managing Director
Chhattisgarh Medical Services Corporation Limited, Health and Family Welfare Department Chhattisgarh,
Chhattisgarh Medical Services Corporation Limited,
(Md Cgmsc)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No
--	----

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं/सेवा प्रदाता को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ / Purchase Preference to MSE OEMs/ Service Provider available upto price within $L1+X\%$	15
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माता/सेवा प्रदाता को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Percentage of Bid quantity/amount for MSE OEMs/ Service Provider Purchase preference	25

1. If the bidder is a Micro or Small Enterprise as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Experience Criteria" as defined above subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
2. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover, shall upload the supporting documents to prove his eligibility for Relaxation.
3. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Experience Criteria" as defined above subject to their meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
4. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Bidder Turnover" as defined above subject to their meeting of quality and technical specifications. If the bidder is DPIIT Registered OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover shall upload the supporting documents to prove his eligibility for Relaxation.
5. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
6. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value

should meet this criterion.

7. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

8. **Purchase preference to Micro and Small Enterprises (MSEs):** Purchase preference will be given to MSEs from the State of Bid Inviting Authority having valid Udyam Certificate and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry/ State of Bid Inviting Authority. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service, and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017 and its subsequent Orders/Notifications issued by concerned Ministry. Benefits of MSE will be allowed only if seller/service provider is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

9. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

10. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 40% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Pre Bid Detail(s)

मूल्य भिन्नता खंड दस्तावेज़/Pre-Bid Date and Time	प्री-बिड स्थान/Pre-Bid Venue
07-07-2026 12:00:00	Chhattisgarh Medical Services Corporation Limited 3rd Floor, South East Corner, Vyawshayk Parisar, Sector 27, Near Mini Market, Atal Nagar, Nawa Raipur - 492015, Chhattisgarh.

Refrigerated Centrifuge For General And Research Purpose (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Performance Parameters	Type of Motor	Brushless(without carbons), With Carbon Brushless

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Type of lid lock	Both Manual and Electronic, Electronic lid locking with automatic lid opening at end of run, click seal biocontainment lid
	No of acceleration and deceleration profiles in built for rotor speed while automatic braking	"Programmable from 50 to 200 Seconds in Interval of 5 Seconds", Acceleration & Deceleration time to be less than 15 sec, Not available, Acceleration and Deceleration Profile-9/10, Option to accelerate and decelerate speed at various rates, Time Set Range 1 min – 99 min; 1 min increments
	Configuration of Rotor supplied with centrifuge machine	Swing Out, Angle rotor, swing out and angle rotor
	Size of rotor supplied with centrifuge in ml	"4x100 ml(for angle rotor) RPM 13000", "6x50 mL(for angle rotor) RPM 13000", "8x25 ml(for angle rotor) RPM 13000", "12x15ml(for angle rotor) RPM 13000", "4 Micro Strips(for angle rotor) RPM 13000", "12x2ml(for angle rotor) RPM 20000", "24x1.5ml(for angle rotor) RPM 16000", "8x5ml(for angle rotor) RPM 16000", "24X2 ml(for angle rotor) RPM 16000", "8X15 ml(for angle rotor) RPM 13000", "24 x 0.5 ml(for angle rotor) RPM 20000", "16X15 ml(for angle rotor) RPM 6000", "24x15ml(for angle rotor) RPM 6000", 16x15ml((For swing out type rotor RPM 5000), 24x15ml(FOR swing out type rotor RPM 5000), 32x15ml(FOR swing out type rotor RPM 5000), 4x50ml(FOR swing out type rotor RPM 5000), 2 Micro Plates of 96 wells Each(FOR swing out type RPM 3000), 2 micro plates of 96 wells each (for swing out type or angle rotor RPM 3500), PCR strip rotor with 1500 rpm/18000G which can hold upto 4xPCR strips RPM15000, 6x100 with 15000rpm, 8x50, 14500rpm, 30x15ml, 5650rpm, 30x2ml, 15200 rpm, 30x2ml, 14000rpm, 18x5ml, 15000rpm, Adapter for 4x50ml 5000rpm, 24 Microplates 4200rpm, 16 Microplates 4700rpm, Fixed Angle Rotor accommodating atleast 24 x 1.5/2 ml tubes (go upto a speed of approximately 14000 RPM)
	Warranty period for equipment & Rotor (on site and comprehensive warranty) in years	1, 2, 3, 4, 5
Additional Parameters	Condensed Water Drainage System	Yes, No
	The Equipment shall be installed and demonstrated before Acceptance	Yes, No(No installation and demonstration required and consignee can install)
	Servo Controller Voltage Stabilizer supplied with Item	Yes, No
	Servo stabilizer ratings in KVA	1, 2, Not supplied, 5 KVA
	Warranty of stabilizer in years	3, 4, 5, NA(If no stabilizer provided)

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Gel Documentation System (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Performance Parameters	Camera imager for imaging applications	DNA gels, Protein gels, Fluorescence, Colorimetric detection, Chemi, 3 colour detection, Infrared fluorescence
	Type of Camera	CMOS, CCD
	System should enable detection of picogram levels of sample with chemiluminescence	Yes, No Or higher
	Display	LED, LCD, Touch screen, NA (If used with workstation)
	Machine should have onboard capture and analyse software for ease of use and automation, with PC connectivity and USB connection	Yes, No Or higher
	Camera with fixed focus motorized lens	Yes, No Or higher
	Cooling Temperature available for the camera in °C	-25, -35, -57, NA (if Chemi is not available)
	Camera resolution in Mega Pixel (MP)	6 Mega Pixel, 5 Mega Pixel, 4 Mega Pixel, 3 Mega Pixel, 7 Mega Pixel, >= 8 Mega Pixel
	Compact dark room should include	UV Transilluminater, UV Transilluminator with White Light source, UV transilluminator with White and Blue Light sources

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	UV transillumination wavelength (nm)	254, 302, 365
	Imaging area in cm x cm with tolerance of $\pm 10\%$	15x17, 20x20, 25x30, 20 x 24
	Machine should include positions motorised filter wheel with suitable filter for EtBr, SYBR Green, Coomassie Blue, Silver Stain	1, 5, 7
	Sample placement slot should be adjustable as per user requirement and application specific trays for UV, White and Chemiluminescent Blot samples should be provided	Yes, No Or higher
	Modes of image capturing	Automatic, Semi automatic, Manual
	Analysis software	1D, 1D & 2D
Miscellaneous Parameters	Comprehensive Warranty in years	2
	Number of installations of the Machine machines in Central /State/PSU Govt Hospitals(Note: Seller should supply a performance certificate of the device to the buyer if demanded after placement of order)	No installations, Upto 3, More than 3

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Water Baths (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Water Bath Chamber Dimensions, Capacity and Construction	Water bath Chamber Capacity	2 to 4 litre, > over 4 to 6 litre, >over 8 to 10 litre, >over 10 to 14 litre, >over 14 to 18 litre, >over 18 to 22 litre, >over 22 to 26 litre, >over 26 to 30 litre, over 6 to 8 litre, over 30 to 34 Litre
	Chamber Construction:	single walled rectangular chamber, Double walled rectangular chamber, Double walled rectangular chamber with mineral wool/solid insulation material filling in between
	Material of Outer Chamber Construction:	Epoxy coated/Powder coated mild steel, Epoxy / Powder coated Stainless Steel SS304
	Number of racks provided with Water bath unit	0, 2, 4, 6, 8, 10, 12, 14
	Number of 75 mm diameter Holes provided:	0, 6, 8, 12, 15, 24
Heating Arrangement Parameters	Heating Capacity (>=):	800, 1000, 1500, 2000, 2400, 2500, 3000, 200, 1200 (Watt)
Control Panel and Display Unit	Control Panel has:	ON/Off Switch, Heating & Temperature (Set & Actual) Indicators, Power mains Indicator, Temperature controller, Circulation Controller (if Water bath has circulation facility), Refrigeration Controller (if Water bath has chamber refrigeration facility), Shaking Controller (if Water bath has shaking facility)
	Adjustable shaking frequency in rpm	20-200, NA
	Type of Temperature controller	Thermostat based control, Digital controller cum Indicator PID Based control Or higher
	Water bath Display type:	Analog Display, Digital (numerical/application specific) backlit LCD display, LED
	Integrated circulation pump for temperature homogeneity	Yes, No Or higher
	Removable bottom plate and shaking inserts	Yes, No, N/A for shaking insert if the equipment provided without circulation feature Or higher
Water bath Temperature	Temperature range of bath chamber:	5°C above ambient to 99.9°C, 5°C above ambient to 90°C, 5°C above ambient to 95°C, 20°C to 99.9°C, 10°C to 100°C, Ambient to 100°C, 10°C to 150°C
	Temperature control sensitivity:	±0.5°C or better, ±0.2°C or better, ±0.1°C or better Or higher
	Least temperature increment Step available	5 degree C, 1 degree C, 0.1 degree C, 0.01 degree C Or higher
	Temperature Accuracy:	±5°C or better, ±1°C or better, ±0.5°C or better, ±0.2°C or better, ±0.1°C or better Or higher
	Safety Features:	Low fluid Alarm, High limit Alarm, Over temperature Alarm & cutoff, Redundant safety thermostat

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Power Requirement, Certifications and Warranty	Power requirement:	220 V 50 Hz Single Phase AC, 440 V 50Hz three phase AC
	Availability of Test Report from Central Govt/NABL/ILAC accredited lab to prove conformity to specification	Yes, No
	Equipment has CE certification	Yes, No
	Warranty Period:	3, 5 Or higher (year)
Timer Unit	Timer Provided	Yes, No
	Timer Setting (minimum increment step)	0 - 300 (second)
Miscellaneous Features	Drain plug in Chamber provided	Yes, No
	Data Logging facility through (RS 232/USB/HDMI/Bluetooth etc) to computer available:	Yes, No
	Drain screw for emptying of water bath	Yes, No Or higher
	Seamless and waterproof keypad	Yes, No Or higher

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Water Baths (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Water Bath Chamber Dimensions, Capacity and Construction	Water bath Chamber Capacity	2 to 4 litre, > over 4 to 6 litre, >over 8 to 10 litre, >over 10 to 14 litre, >over 14 to 18 litre, >over 18 to 22 litre, >over 22 to 26 litre, >over 26 to 30 litre, over 6 to 8 litre, over 30 to 34 Litre
	Chamber Construction:	single walled rectangular chamber, Double walled rectangular chamber, Double walled rectangular chamber with mineral wool/solid insulation material filling in between
	Material of Outer Chamber Construction:	Epoxy coated/Powder coated mild steel, Epoxy / Powder coated Stainless Steel SS304
	Number of racks provided with Water bath unit	0, 2, 4, 6, 8, 10, 12, 14
	Number of 75 mm diameter Holes provided:	0, 6, 8, 12, 15, 24
Heating Arrangement Parameters	Heating Capacity (>=):	800, 1000, 1500, 2000, 2400, 2500, 3000, 200, 1200 (Watt)
Control Panel and Display Unit	Control Panel has:	ON/Off Switch, Heating & Temperature (Set & Actual) Indicators, Power mains Indicator, Temperature controller, Circulation Controller (if Water bath has circulation facility), Refrigeration Controller (if Water bath has chamber refrigeration facility), Shaking Controller (if Water bath has shaking facility)
	Adjustable shaking frequency in rpm	20-200, NA
	Type of Temperature controller	Thermostat based control, Digital controller cum Indicator PID Based control Or higher
	Water bath Display type:	Analog Display, Digital (numerical/application specific) backlit LCD display, LED
	Integrated circulation pump for temperature homogeneity	Yes, No Or higher
	Removable bottom plate and shaking inserts	Yes, No, N/A for shaking insert if the equipment provided without circulation feature Or higher
Water bath Temperature	Temperature range of bath chamber:	5°C above ambient to 99.9°C, 5°C above ambient to 90°C, 5°C above ambient to 95°C, 20°C to 99.9°C, 10°C to 100°C, Ambient to 100C, 10°C to 150°C
	Temperature control sensitivity:	±0.5°C or better, ±0.2°C or better, ±0.1°C or better Or higher
	Least temperature increment Step available	5 degree C, 1 degree C, 0.1 degree C, 0.01 degree C Or higher
	Temperature Accuracy:	±5°C or better, ±1°C or better, ±0.5°C or better, ±0.2°C or better, ±0.1°C or better Or higher
	Safety Features:	Low fluid Alarm, High limit Alarm, Over temperature Alarm & cutoff, Redundant safety thermostat

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Power Requirement, Certifications and Warranty	Power requirement:	220 V 50 Hz Single Phase AC, 440 V 50Hz three phase AC
	Availability of Test Report from Central Govt/NABL/ILAC accredited lab to prove conformity to specification	Yes, No
	Equipment has CE certification	Yes, No
	Warranty Period:	3, 5 Or higher (year)
Timer Unit	Timer Provided	Yes, No
	Timer Setting (minimum increment step)	0 - 300 (second)
Miscellaneous Features	Drain plug in Chamber provided	Yes, No
	Data Logging facility through (RS 232/USB/HDMI/Bluetooth etc) to computer available:	Yes, No
	Drain screw for emptying of water bath	Yes, No Or higher
	Seamless and waterproof keypad	Yes, No Or higher

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Water Baths (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Water Bath Chamber Dimensions, Capacity and Construction	Water bath Chamber Capacity	2 to 4 litre, > over 4 to 6 litre, >over 8 to 10 litre, >over 10 to 14 litre, >over 14 to 18 litre, >over 18 to 22 litre, >over 22 to 26 litre, >over 26 to 30 litre, over 6 to 8 litre, over 30 to 34 Litre
	Chamber Construction:	single walled rectangular chamber, Double walled rectangular chamber, Double walled rectangular chamber with mineral wool/solid insulation material filling in between
	Material of Outer Chamber Construction:	Epoxy coated/Powder coated mild steel, Epoxy / Powder coated Stainless Steel SS304
	Number of racks provided with Water bath unit	0, 2, 4, 6, 8, 10, 12, 14
	Number of 75 mm diameter Holes provided:	0, 6, 8, 12, 15, 24
Heating Arrangement Parameters	Heating Capacity (>=):	800, 1000, 1500, 2000, 2400, 2500, 3000, 200, 1200 (Watt)
Control Panel and Display Unit	Control Panel has:	ON/Off Switch, Heating & Temperature (Set & Actual) Indicators, Power mains Indicator, Temperature controller, Circulation Controller (if Water bath has circulation facility), Refrigeration Controller (if Water bath has chamber refrigeration facility), Shaking Controller (if Water bath has shaking facility)
	Adjustable shaking frequency in rpm	20-200, NA
	Type of Temperature controller	Thermostat based control, Digital controller cum Indicator PID Based control Or higher
	Water bath Display type:	Analog Display, Digital (numerical/application specific) backlit LCD display, LED
	Integrated circulation pump for temperature homogeneity	Yes, No Or higher
	Removable bottom plate and shaking inserts	Yes, No, N/A for shaking insert if the equipment provided without circulation feature Or higher
Water bath Temperature	Temperature range of bath chamber:	5°C above ambient to 99.9°C, 5°C above ambient to 90°C, 5°C above ambient to 95°C, 20°C to 99.9°C, 10°C to 100°C, Ambient to 100C, 10°C to 150°C
	Temperature control sensitivity:	±0.5°C or better, ±0.2°C or better, ±0.1°C or better Or higher
	Least temperature increment Step available	5 degree C, 1 degree C, 0.1 degree C, 0.01 degree C Or higher
	Temperature Accuracy:	±5°C or better, ±1°C or better, ±0.5°C or better, ±0.2°C or better, ±0.1°C or better Or higher
	Safety Features:	Low fluid Alarm, High limit Alarm, Over temperature Alarm & cutoff, Redundant safety thermostat

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Power Requirement, Certifications and Warranty	Power requirement:	220 V 50 Hz Single Phase AC, 440 V 50Hz three phase AC
	Availability of Test Report from Central Govt/NABL/ILAC accredited lab to prove conformity to specification	Yes, No
	Equipment has CE certification	Yes, No
	Warranty Period:	3, 5 Or higher (year)
Timer Unit	Timer Provided	Yes, No
	Timer Setting (minimum increment step)	0 - 300 (second)
Miscellaneous Features	Drain plug in Chamber provided	Yes, No
	Data Logging facility through (RS 232/USB/HDMI/Bluetooth etc) to computer available:	Yes, No
	Drain screw for emptying of water bath	Yes, No Or higher
	Seamless and waterproof keypad	Yes, No Or higher

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

Biosafety Cabinet (2 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Performance Parameters	Type of Configuration	Freestanding, Benchtop

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Material used for working area	Stainless steel 304 grade, Stainless steel 316 grade
	Intensity of light in lux	? 600, ? 900
	Filtration efficiency of ULPA filter	99.999% for 0.1 micron particles, ?99.9995%, @0.12micron particles, NA
	Width of the working area in feet	3ft and more, 4ft and more
	Depth of the working area in feet	2ft and more, 3ft and more, 4ft and more
	Height of the working area in feet	2ft and more, 3ft and more, 4ft and more
	Bio saftey cabimet should have Provision for fitting taps	Taps for vaccum, water and non combustible gas, Not provided
Additional Accessories	Magnehelic Gauge provided	1, NA
	Hepafilter	1, NA
	Gas burner	1, NA
Miscellaneous Parameters	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	5, 1, 2, 3, 4

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	2	60

Biosafety Cabinet (2 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Performance Parameters	Type of Configuration	Freestanding, Benchtop
	Material used for working area	Stainless steel 304 grade, Stainless steel 316 grade
	Intensity of light in lux	? 600, ? 900
	Filtration efficiency of ULPA filter	99.999% for 0.1 micron particles, ?99.9995%, @0.12micron particles, NA
	Width of the working area in feet	3ft and more, 4ft and more
	Depth of the working area in feet	2ft and more, 3ft and more, 4ft and more
	Height of the working area in feet	2ft and more, 3ft and more, 4ft and more
	Bio safety cabinet should have Provision for fitting taps	Taps for vacuum, water and non combustible gas, Not provided
Additional Accessories	Magnehelic Gauge provided	1, NA
	Hepafilter	1, NA
	Gas burner	1, NA
Miscellaneous Parameters	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	5, 1, 2, 3, 4

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009, Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	2	60

Refrigerated Centrifuge For General And Research Purpose (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Performance Parameters	Type of Motor	Brushless(without carbons), With Carbon Brushless
	Type of lid lock	Both Manual and Electronic, Electronic lid locking with automatic lid opening at end of run, click seal biocontainment lid
	No of acceleration and deceleration profiles in built for rotor speed while automatic braking	"Programmable from 50 to 200 Seconds in Interval of 5 Seconds", Acceleration & Deceleration time to be less than 15 sec, Not available, Acceleration and Deceleration Profile-9/10, Option to accelerate and decelerate speed at various rates, Time Set Range 1 min – 99 min; 1 min increments
	Configuration of Rotor supplied with centrifuge machine	Swing Out, Angle rotor, swing out and angle rotor
	Size of rotor supplied with centrifuge in ml	"4x100 ml(for angle rotor) RPM 13000", "6x50 mL(for angle rotor) RPM 13000", "8x25 ml(for angle rotor) RPM 13000", "12x15ml(for angle rotor) RPM 13000", "4 Micro Strips(for angle rotor) RPM 13000", "12x2ml(for angle rotor) RPM 20000", "24x1.5ml(for angle rotor) RPM 16000", "8x5ml(for angle rotor) RPM 16000", "24X2 ml(for angle rotor) RPM 16000", "8X15 ml(for angle rotor) RPM 13000", "24 x 0.5 ml(for angle rotor) RPM 20000", "16X15 ml(for angle rotor) RPM 6000", "24x15ml(for angle rotor) RPM 6000", 16x15ml((For swing out type rotor RPM 5000), 24x15ml(FOR swing out type rotor RPM 5000), 32x15ml(FOR swing out type rotor RPM 5000), 4x50ml(FOR swing out type rotor RPM 5000), 2 Micro Plates of 96 wells Each(FOR swing out type RPM 3000), 2 micro plates of 96 wells each (for swing out type or angle rotor RPM 3500), PCR strip rotor with 1500 rpm/18000G which can hold upto 4xPCR strips RPM15000, 6x100 with 15000rpm, 8x50, 14500rpm, 30x15ml, 5650rpm, 30x2ml, 15200 rpm, 30x2ml, 14000rpm, 18x5ml, 15000rpm, Adapter for 4x50ml 5000rpm, 24 Microplates 4200rpm, 16 Microplates 4700rpm, Fixed Angle Rotor accommodating atleast 24 x 1.5/2 ml tubes (go upto a speed of approximately 14000 RPM)
	Warranty period for equipment & Rotor (on site and comprehensive warranty) in years	1, 2, 3, 4, 5
Additional Parameters	Condensed Water Drainage System	Yes, No
	The Equipment shall be installed and demonstrated before Acceptance	Yes, No(No installation and demonstration required and consignee can install)
	Servo Controller Voltage Stabilizer supplied with Item	Yes, No
	Servo stabilizer ratings in KVA	1, 2, Not supplied, 5 KVA

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Warranty of stabilizer in years	3, 4, 5, NA(If no stabilizer provided)

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Medical Centrifuge (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Standards	Conformity to standards	European CE (Notified Body)With 4 Digit Notified Body No., BIS, CE
Performance Parameters	Rotor configuration	Swing Out Rotor, Angle Rotor
	Maximum Volume of the tube testable	Vacutainers, 10 ml centrifuges tubes, 15 ml centrifuges tubes
	Rotor Size	8x15ml,, 12x15ml,, 16x15ml,, 24x15ml, 32x15ml, 36x15ml, 16 Vacuutainers, 24 Vacuutainers, 32 Vacuutainers, 48 Vacuutainers, 64 Vacuutainers, NA
	Warranty in Years	1, 2, 3, 4, 5
Reports	ISO certification for Manufacturer	Yes, No

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Deep Freezer (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Standards	Conformity to Standards	CE (with 4 digit notified number), USFDA, BIS (As per medical directive)
Performance Parameters	Capacity in liters	250-300, 310-350, 360-400, 401-500, 501-600, 105 - 150, 155-195, 200-245, < 200, 301-500, 501-700, 701-900, 901-1100, 1101-1200, 200-250
	Type of cabinet	Chest, Horizontal, Vertical, Upright, Upright double door
	Castors	Heavy duty lockable, NA (if no castors provided)
	Number of corrosion resistant racks	2, 3, 4, 5, 6, 7, 8, 9, >10
	Number of inner storage compartments with insulated doors and	3, 4, 5, 1, 2
	Warranty of Freezer in Years (from the date of installation)	2, 3, 4, 5, 1
Additional Parameters	CO₂ cylinder should be supplied with freezer for backup	Yes, No
	Temperature Chart recorder to be provided	Yes, No

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

Deep Freezer (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Standards	Conformity to Standards	CE (with 4 digit notified number), USFDA, BIS (As per medical directive)
Performance Parameters	Capacity in liters	250-300, 310-350, 360-400, 401-500, 501-600, 105 - 150, 155-195, 200-245, < 200, 301-500, 501-700, 701-900, 901-1100, 1101-1200, 200-250
	Type of cabinet	Chest, Horizontal, Vertical, Upright, Upright double door
	Castors	Heavy duty lockable, NA (if no castors provided)
	Number of corrosion resistant racks	2, 3, 4, 5, 6, 7, 8, 9, >10
	Number of inner stroage compartments with insulated doors and	3, 4, 5, 1, 2
	Warranty of Freezer in Years (from the date of installation)	2, 3, 4, 5, 1
Additional Parameters	CO₂ cylinder should be supplied with freezer for backup	Yes, No
	Tempearture Chart recorder to be provided	Yes, No

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

Auloclave-1

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Autoclave-2

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	2	60

BOD Incubator-1

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

BOD Incubator-2

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

Data Logger -80C, -20C

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	4	60

Data Logger -80c, -20c, Plus 4c 2no**तकनीकी विशिष्टियाँ /Technical Specifications**

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	4	60

Elisa Plate Reader-1**तकनीकी विशिष्टियाँ /Technical Specifications**

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

ELISA Plate Reader-2

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Elisa Plate Reader-3

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

Elisa Plate Washer-1

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

ELISA Plate Washer-2

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
-------------------------------	---------------------------

BOQ Detail Document[View File](#)

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Elisa Plate Washer-3**तकनीकी विशिष्टियाँ /Technical Specifications****Specification Document**[View File](#)**BOQ Detail Document**[View File](#)

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

Floater Rack**तकनीकी विशिष्टियाँ /Technical Specifications****Specification Document**[View File](#)**BOQ Detail Document**[View File](#)

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	20	60

Gel Electrophoresis

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Gel Electrophoresis And Documentation Unit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Gel Electrophoresis AndDocumentation Unit

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

Laminar Air Vertical Flow System

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	2	60

Ice Buckets

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Ice Flaker Machine

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Label Printer

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Label Printer Brady Printer

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Laboratory Incubator**तकनीकी विशिष्टियाँ /Technical Specifications**

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Laboratory Rocker And Shaker**तकनीकी विशिष्टियाँ /Technical Specifications**

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Laminar Air Flow Horizontal**तकनीकी विशिष्टियाँ /Technical Specifications**

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Laminar Air Vertical Flow 2 Feet**तकनीकी विशिष्टियाँ /Technical Specifications**

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	2	60

Magnetic Stirrer-1**तकनीकी विशिष्टियाँ /Technical Specifications**

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Magnetic Stirrer-2**तकनीकी विशिष्टियाँ /Technical Specifications**

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Magnetic Stirrer-3

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	2	60

Micro Pipette 0.5 - 10 UI

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	2	60

Micro Pipette 10 - 100 UI

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Micro Pipette 100 - 1000 UI

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	3	60

Micro Pipette 2 - 20 UI

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	2	60

Micro Pipette 20 - 200 UI

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	5	60

Micropipettes

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	2	60

Microwave Oven For Agarose Gel

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Microwave Oven For Agarose Gels

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Multi Channerl 30 - 300 UI

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Multifunctional Copier Or Printer Or ScannerFor Office Work

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Pan Electronic Weighing Balance-1

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Pan Electronic Weighing Balance-2

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

PCR Workstation

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

PH Meter

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	1	60

Quick Spin

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	5	60

Quick Spinner-1

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Quick Spinner-2

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	2	60

Quickspin For 1.5, 2ml Tubes. PCR Tubes And Plates

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Refrigerated Micro Centrifuge

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Refrigerated Microfuge

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Refrigerator

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	2	60

Refrigerator Sample Receiving

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	2	60

Themo Cyclor-3

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	2	60

Themocycler-2

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Thermo Cycler-1

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

Timer-1

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	2	60

Timer-2

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	2	60

Ultrapure Water System

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Unichannel And Multichannel Pipettes

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

UV Dosimeter-1

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

UV Dosimeter-2

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Erumalla Naveen	494334,Late Smt. Indira Gandhi Memorial Government Medical College,Nandanmara, Administrative Building, kanker, Chhattisgrah	1	60

UV Dosimeter-3

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	2	60

Vortex Mixer-1

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Nahrel	495009,Chhattisgarh Institute of Medical Sciences (CIMS), Sardar Vallabh Bhai Patel Hospital Bilaspur (Chhattisgarh) - 495009, India	1	60

Vortex Mixer-2

तकनीकी विशिष्टियाँ /Technical Specifications

Specification Document	View File
BOQ Detail Document	View File

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Anil Kumar Baghel	492001,Jail Road, Pt. J.N.M Medical College, Raipur	2	60

Special terms and conditions-Version:3 effective from 11-11-2022 for category Refrigerated Centrifuge For General And Research Purpose

1.

- All Provisions of Drugs and Cosmetics Act, 1940 and Rules made there under as amended till date will always be applicable. This will include all notifications issued by *Central Drugs Standard Control Organisation(CDSCO)*, Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
- The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
- The price offered by the seller shall not, in any case exceed the DPCO controlled price, if any, fixed by the Central/State Government, the Maximum Retail Price (MRP) and the selling price. The seller must reduce the prices if there is any reduction in DPCO ceiling price, if any.
- Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty

period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

5. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
6. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
7. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
8. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.
OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
9. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
10. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
11. **Software:** All software updates should be provided free of cost during warranty period.
12. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

Special terms and conditions-Version:2 effective from 24-11-2023 for category Gel Documentation System

1.
 - o **Comprehensive warranty**
Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares,. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

- **Service centres**
Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address ,telephone numbers, e mails etc at time of making the supplies .It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled .Details of toll free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
- **Source of supply**
It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
- **Packing and Marking**
Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take in to consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed .Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date , brief description of goods including quantity ,. Packing list reference number , country of origin of goods and any other relevant details.
- **Spare Parts**
Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM . It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies .In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.
OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
- **Installation, Training, Manuals**
Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and Supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations , training and manuals the same shall also be applicable.
- **Electrical safety checking**
Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
- Software All software updates should be provided free of cost during warranty period.

Special terms and conditions-Version:2 effective from 06-01-2023 for category water baths

1. Comprehensive warranty:

Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares,. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables .Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

2. Service centres:

Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address ,telephone numbers, e mails etc at time of making the supplies .It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where

supplies are made within reasonable distance from where the service calls can be handled .Details of toll free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.

3. Source of supply:

It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them .

4. Packing and Marking: Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take in to consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed .Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date , brief description of goods including quantity ,. Packing list reference number , country of origin of goods and any other relevant details.

5. Spare Parts:

Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM . It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies .In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.

OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied

6. Installation, Training, Manuals:

Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and Supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations , training and manuals the same shall also be applicable.

7. Electrical safety checking:

Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.

8. Software:

All software updates should be provided free of cost during warranty period.

Special terms and conditions-Version:2 effective from 06-01-2023 for category Biosafety Cabinet

1.
 - **Comprehensive warranty**
Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares,. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables .Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
 - **Service centres**
Details of Service outlets in India to render services for equipment to be furnished to

buyer/consignees with complete address ,telephone numbers, e mails etc at time of making the supplies .It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled .Details of toll free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.

- **Source of supply**

It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.

- **Packing and Marking**

Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take in to consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed .Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date , brief description of goods including quantity ,. Packing list reference number , country of origin of goods and any other relevant details.

- **Spare Parts**

Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM . It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies .In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.

OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.

- **Installation, Training, Manuals**

Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignees's representatives for operating and maintaining the equipment and Supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations , training and manuals the same shall also be applicable.

- **Electrical safety checking**

Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.

- **Software** All software updates should be provided free of cost during warranty period.

Special terms and conditions-Version:2 effective from 11-02-2025 for category Medical Centrifuge

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard. (As Applicable)
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc. (As Applicable)
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any

- reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
 6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
 7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
 8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
 9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
 10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
 11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
 12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
 13. **Software:** All software updates should be provided free of cost during warranty period.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

अस्वीकरण/Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer, is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs and including the modification of technical specifications and / or terms and conditions governing the bid. All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM. If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.

However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।/In terms

of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---