

Kaolin based EMS 4X4 Dressing

1. The active ingredient should be Kaolin
2. The dimensions of the gauze should be 10 cm X 10 cm (tolerance of +/- 5%)
3. The gauze should be suitable for field injuries, minor cuts and lacerations resulting in traumatic bleeding
4. The gauze should be a soft, white, nonwoven, hydrophilic gauze made of Rayon & Polyester and impregnated with kaolin
5. The gauze should be heat free and have no any exothermic reactions and should not contain any animal or human proteins
6. The shelf life of the gauze should be 3 years
7. The gauze should be packaged in an easy-tear sterile pouch
8. The gauze should have CE certification & US FDA Approval

Chitosan Haemostatic Dressing 8x8cm

1. Active ingredients 100% Chitosan with a net positive charge.
2. In the form of dressing pad of size 8x8cms and active on all sides
3. To be supplied in a moisture proof easy to use pack.
4. Should be capable of stopping bleeding without causing any exothermic, skin irritation reaction and should be biocompatible.
5. The haemostatic dressing should have a pH between 5 to 7.
6. The haemostatic dressing should not release any particle in water or saline.
7. Should be gamma sterile with NABL Test reports & Certification.
8. It should have preferably 05 years shelf life printed on pack.
9. Should be ISO13485 and BIS 23485
10. OEM should provide clinical evidence and market standing for at least 5 years.
11. Manufactured in India - Indian MDR, CE from notified body / USFDA.

TWIN PACK CHEST SEAL WITH VENT

1. The Chest Seal with vent should be meant for the prevention, management and treatment of an open and/or tension pneumothorax potentially caused by a penetrating chest trauma.
2. The 2 chest seals in the twin pack should be individually packed to be used for entry and exit wound
3. This chest seal device should have 3-vented channels designed to
 - a. prevent airflow into the chest cavity during inhalation while allowing air to escape through the vent channels during exhalation.
 - b. allow blood to escape and also provide a backup fail-safe system even if two of the three channels become obstructed

4. The adhesion in the vent chest seal should be formulated using Advanced hydrogel technology to provides superior adhesion where the casualty may be covered in blood, sweat or body hair.
5. Each chest seal should be packaged in a compact, easy-to-open foil pouch.
6. Each chest seal should also include a gauze pad to wipe the wound surface prior to application
7. The Vent Chest Seal should have a large pull tab for single-step peel-and-apply application
8. Each chest seal should have transparent backing for easy placement over the wound area.
9. The chest seal should have a shelf life of 5 years
10. The chest seal should be CE certified or US FDA approved
11. The product should be NATO approves and carry a NSN (NATO Stock Number)
12. Each chest seal dimensions should be 5 inches X 5 inches (excluding pulling tab)
13. The weight of the twin pack chest seal should be 71 gms (tolerance +/- 5%)

Burn Dressing 10cm X 10cm

1. The burn dressing should be effective and 10cm x 10cm size
2. The dressing should be sterile and made with 96% purified water and 0.1% Hyaluronic Acid (HA)
3. The dressing should provide quick relief, cool the injured area, and serve as a barrier from the environment, reducing the potential for further contamination.

SURVIVAL BLANKET LARGE

1. The survival blanket is a high performing insulating blanket for protection from cold, wind and rain as well as an effective aid in the prevention and treatment of hypothermia.
2. It should be made of Reflexcell: expanding 3-layer metallised OPP film (Oriented Polypropylene) and of Green Colour.
3. It should be a large blanket with adhesive re-usable single closure at front to allow maximum mobility whilst maintaining body temperature.
4. Its size when packed should be ~22cm x 11cm x 5cm and when open should be approx. Length 2.30 metres, width 2.2 metres.
5. Its weight should be less than 500 gm.
6. It should have a TOG warmth (Thermal Overall Grade) value of 7.5
7. It should have CE Certification and NATO NSN Number or have an Indian Certification (BIS or ISO13485 for the particular product) by CDSCO nominated organization