

Messprotokoll – EC-2025-000123

Erstellt: 2025-10-17 16:41:52

Allgemeine Metadaten

barcode	EC-2025-000123
manufacturer	ACME Components GmbH
orig_name	C-10u-16V
category	Kondensator
status	Kalibriert
operator	C. Kessler
location	Labor A-2
comment	Standardmessung nach Sichtprüfung.

OpenBIS


project	ELEC-CHECK
experiment	EC-EXP-2025-10
sample	EC-SAMPLE-000987
dataset	DS-2025-10-17-01
space	ELECTRONICS
user	ckessler
session_token	SESSION-PLACEHOLDER
link	https://openbis.example.org/dss/webapp/#!/datastore_browser/DS-2025-10-17-01

Notizen

Messung mit Level 0.3V und 0.6V. Keine Auffälligkeiten.

Messergebnisse

Messung



Messwerte

Zeit	Frequenz	Level (mV)	equiv	Z (Ω)	theta_deg ($^{\circ}$)	u_primary	u_secondary	range_name
14:11:33	100 Hz	300	SER	0.995265	-0.0841516	0.0199076	1.1	4 Ω
14:11:36	100 Hz	600	SER	0.998709	-0.0720117	0.0199764	1.1	4 Ω
14:11:45	120 Hz	300	SER	0.991341	-0.0826439	0.0198291	1.1	4 Ω
14:11:49	120 Hz	600	SER	0.996272	-0.0439692	0.0199277	1.1	4 Ω
14:11:58	400 Hz	300	SER	0.993363	0.0125467	0.0198695	1.1	4 Ω
14:12:01	400 Hz	600	SER	0.99661	-0.0122938	0.0199345	1.1	4 Ω
14:12:10	1 kHz	300	SER	0.995217	0.0361351	0.0199066	1.1	4 Ω
14:12:13	1 kHz	600	SER	0.998208	0.00658938	0.0199664	1.1	4 Ω
14:12:22	4 kHz	300	SER	0.9937	0.148866	0.0198763	1.1	4 Ω
14:12:26	4 kHz	600	SER	0.998132	0.143842	0.0199649	1.1	4 Ω
14:12:35	10 kHz	300	SER	0.997865	0.334748	0.0199596	1.1	4 Ω
14:12:38	10 kHz	600	SER	1.00181	0.357486	0.0200384	1.1	4 Ω
14:12:48	40 kHz	300	SER	0.992001	1.1588	0.0198491	1.1	4 Ω
14:12:51	40 kHz	600	SER	1.00184	0.684741	0.0200458	1.1	4 Ω
14:13:00	50 kHz	300	SER	0.994981	1.61826	0.0199087	1.1	4 Ω
14:13:03	50 kHz	600	SER	1.00367	1.00107	0.0200824	1.1	4 Ω
14:13:12	75 kHz	300	SER	0.996898	1.84821	0.0299237	1.7	4 Ω
14:13:15	75 kHz	600	SER	1.00621	1.44152	0.0302029	1.7	4 Ω
14:13:24	100 kHz	300	SER	0.997823	1.81876	0.0299514	1.7	4 Ω
14:13:28	100 kHz	600	SER	1.00789	1.75542	0.0302532	1.7	4 Ω