


Dynamics Tutorial (Alpha)



Introduction

This tutorial will take you through the analysis of some dynamics data in CcpNmr Analysis Version 3.1 using the alpha-version of the Relaxation Data Analysis Module. We assume that you are already familiar with the basics of how the program works, e.g. by doing our Beginner Tutorial.

You can download the example data to go with this tutorial from the tutorials page of our website (<https://ccpn.ac.uk/support/tutorials/>). We are grateful to Dr Fred Muskett for making the spectra of GB1 available to us for use in this tutorial. You will be analysing T1 and T2 relaxation and heteronuclear NOE data, and then going on to do reduced spectral density mapping and Hsqc ETA(η) analysis. We have not included any theoretical background to these experiments. Please see other publications for this information.

The tutorial is divided into sections, each of them has a set of simple actions: you will see a descriptive image on top and a full description below. (Note that images are representative, and that there may be small differences between your setup and that shown in the tutorial.)

Contents:


1. Loading Data
2. T1 and T2 Data
3. Heteronuclear NOE Data
4. Combined Analyses
5. Reduced Spectral Density Mapping
6. Hsqc- η analysis
7. Exporting Graphs
8. Examining the Graphs

Start CcpNmr Analysis V3

Apple/Linux users by using the terminal command *bin/assign* in the *ccpnmr* directory

Windows users by double-clicking on the *assign.bat* file in the bin directory of the *ccpnmr* directory

Loading data



1A Open project and load spectra


- Find the **GB1_Assigned.ccpn** project folder in the Dynamics Tutorial data directory. This project contains an assigned HSQC spectrum of GB1.
- Select the folder in your file browser and drag it onto the **Sidebar** or **Drop Area**.
- Find the **spectra** directory in the Dynamics Tutorial data directory
- Select and drag the following three folders into the Sidebar or Drop Area:

GB1_T1

GB1_T2


GB1_HetNOE

If asked whether you want to continue loading multiple items, click **Yes**.



If you drop the folders into the Drop Area they will open directly. If you drop them on the Sidebar they will simply be visible in the **Spectra** section of the sidebar and can be opened in a SpectrumDisplay module at a later stage.

Loading data




1B Create the T1 SpectrumGroup and Series

- Select all the T1 spectra in the sidebar, right-click and select Make SpectrumGroup from Selected

OR

- Expand SpectrumGroups in the sidebar and double-click on <New SpectrumGroup>
- In the Edit SpectrumGroup pop-up, drag all the T1 spectra from the left hand side the right hand side of the pop-up
- Give your SpectrumGroup a new Name, e.g. T1

Loading data



1c Create the T1 Series

- Go to the **Series** tab of the Edit SpectrumGroup pop-up
- Enter **s** (seconds) as your unit (either type it or select from the drop-down menu).
- Enter the relaxation delay (in seconds) for each spectrum. You can derive these from the spectrum names (note that these are given in ms, not s). Note that some delay times are repeated in order to give an indication of the measurement error.
- If your spectra are not ordered in ascending order, then click the **Ascending** button to do so!
- Finally click on **Save**.

Loading data

T2Data

The screenshot shows the T2Data interface. On the left, a sidebar lists spectra: SP:T1_9_728ms, SP:T2_1_3-6ms, SP:T2_2_7-21, SP:T2_3_21, SP:T2_4_21, SP:T2_5_36ms, SP:T2_6_54ms, SP:T2_7_54ms, SP:T2_8_75ms, SP:T2_9_101ms, and SP:saturated. A context menu is open over the first few spectra, with 'Make SpectrumGroup From Selected' highlighted. To the right is a 'New SpectrumGroup' dialog with a 'Series' tab selected. It shows a table with 'Series Type' set to 'Float'. The 'Series Units' dropdown is set to 's'. The table lists the spectra from the sidebar with their corresponding values: SP:T2_1_3-6ms (0.0036), SP:T2_2_7-21ms (0.0072), SP:T2_3_21-6ms (0.0216), SP:T2_4_21-6ms (0.0216), SP:T2_5_36ms (0.036), SP:T2_6_54ms (0.054), SP:T2_7_54ms (0.054), SP:T2_8_75-6ms (0.756), and SP:T2_9_101ms (0.101).

HetNOE

The screenshot shows the HetNOE interface. On the left, a sidebar lists spectra: SP:T2_9_101ms, SP:saturated_noe, SP:unsaturated_noe, and a 'SpectrumGroup' section containing '<New Spec>'. A context menu is open over the spectra in the sidebar, with 'Make SpectrumGroup From Selected' highlighted. To the right is a 'New SpectrumGroup' dialog with a 'Series' tab selected. It shows a table with 'Series Type' set to 'Float'. The 'Series Units' dropdown is set to 'AU'. The table lists the spectra from the sidebar with their corresponding values: SP:saturated_noe (1.0) and SP:unsaturated_noe (0.0).

1D Create T2 and Heteronuclear NOE SpectrumGroups and Series


- Repeat the procedure for creating a SpectrumGroup and setting up the Series for the T2 data

Create a SpectrumGroup and Series for the Heteronuclear NOE data:

- Select the **saturated_noe** and **unsaturated_noe** spectra in the sidebar, right-click and select **Make SpectrumGroup from Selected**.
- Give the SpectrumGroup a **Name** (e.g. **HetNOE**)
- In the Series tab, set your units to **AU** (arbitrary units) and enter the following values:

saturated_noe	1.0
unsaturated_noe	0.0

Loading data



N	spectrumGroupName	spectrumPath	spectrumName	series	seriesUnit
2	1 T1	GB1_T1/T1_1_8ms.ucsf	T1_1_8ms	0.008	s
3	2 T1	GB1_T1/T1_2_48ms.ucsf	T1_2_48ms	0.048	s
4	3 T1	GB1_T1/T1_3_120ms.ucsf	T1_3_120ms	0.12	s
5	4 T1	GB1_T1/T1_4_120ms.ucsf	T1_4_120ms	0.12	s
6	5 T1	GB1_T1/T1_5_224ms.ucsf	T1_5_224ms	0.224	s
7	6 T1	GB1_T1/T1_6_360ms.ucsf	T1_6_360ms	0.36	s
8	7 T1	GB1_T1/T1_7_528ms.ucsf	T1_7_528ms	0.528	s
9	8 T1	GB1_T1/T1_8_528ms.ucsf	T1_8_528ms	0.528	s
10	9 T1	GB1_T1/T1_9_728ms.ucsf	T1_9_728ms	0.728	s
11	10 T1	GB1_T1/T1_10_960ms.ucsf	T1_10_960ms	0.96	s
12	1 T2	GB1_T2/T2_1_3-6ms.ucsf	T2_1_3-6ms	0.0036	s
13	2 T2	GB1_T2/T2_2_7-2ms.ucsf	T2_2_7-2ms	0.0072	s
	3 T2	GB1_T2/T2_3_21-6ms.ucsf	T2_3_21-6ms	0.0216	s


1E Import from Excel Lookup File (optional)

As an alternative to loading the spectra and setting up the SpectrumGroups and Series manually, you can also enter all the information into an Excel file and drag this into the program. This will automatically load the spectra and create the SpectrumGroups and Series for you.

The figure above shows the column headings required for the Excel file to load correctly. SpectrumPaths are relative to the location of the Excel file.

- Find the **GB1_Assigned ccpn** project folder in the Dynamics Tutorial data directory and drag it onto the **Sidebar** or **Drop Area**.
- Select **Yes**, if asked whether you want to open a new project or not
- Find the **RelaxLookup.xls** Excel file in the tutorial data **spectra** folder and drag it into the Sidebar or Drop Area.
- Expand the **Spectra** and **SpectrumGroups** sections of the sidebar to see that the data has been imported.


Loading data




1F Check alignment to HSQC

- Close any open SpectrumDisplays (click on the in the top right corner).
- Drag the GB1_HSQC spectrum from the sidebar into the DropArea.
- Drag the T1_1_8ms spectrum from the sidebar on top of the other spectrum.

These should overlay well, meaning that you will be able copy the peaks from the GB1_HSQC spectrum to the T1 spectra without having to change their positions.



T1 and T2 Data



2A Open Relaxation Module

At this point you can either continue from Section 1 or start using our [Section1_completed ccpn](#) project.


- Go to Main Menu → Macro → Run CCPN Macros → RelaxationModule(Alpha).

This will open the alpha-version Relaxation Module.

- Open the Settings Panel by clicking the gear icon in the top left of the module window.

This is where you will set up your data and run all the calculations required before inspecting it in the main module.

T1 and T2 Data



2B Copy Peaks

In the **General Setup** section you will need to select a **SpectrumGroup** (which specifies which spectra you want to use in your analysis) and a **Collection** (which is a collection of peak groups across the spectra).

We have already created our SpectrumGroups:

- Select the **T1 SpectrumGroup** from the drop-down menu.

You will need to set up the Collection, as we don't have this yet:

- Click on the **Setup...** button to set up your Collection
- In the Series Peak Collection pop-up set:

Select SpectrumGroup: SG:T1

Source PeakList: PL:GB1_HSQC.2

CollectionName: T1Peaks

Select Refit Peaks At Position

And keep the remaining default options, as shown above.

- Click **Okay** which will start the creation of the Peak Collections. This will probably take a minute or so.

The screenshot shows the 'Input data' tab selected in the top navigation bar. The interface is divided into several sections:

- General Setup:**
 - Create SpectrumGroup: Create ... (button)
 - SpectrumGroup: > select-to-add < (dropdown menu) currently set to SG:T1
- Experiment:**
 - Experiment Name: Radio buttons for N/A, T1 (selected), T2, and HetNoe.
- Input DataTables:**

Input DataTable Name	T1Input
Create Input DataTable	Create
Input DataTable(s)	> select-to-add < (dropdown menu) currently set to None
- Results DataTable:**

Results DataTable Name	SeriesAnalysisResultsData
Fit and Fetch Results Data	Fit
Results DataTable	> Select < (dropdown menu)

2C Create Input DataTable

You will now create your Input DataTable, a table containing all the information needed to fit your data.

In the **Experiment** Section:

- Select **T1**

In the **Input DataTables** Section:

- Enter a **Input DataTable Name**, e.g. **T1Input**
- Click on the **Create** button to create your Input DataTable
- Check that this new Input Data Table is the only one shown in the **Input Data Table(s)** box.

The screenshot shows the 'Input DataTables' section of the configuration interface. The 'Create Input DataTable' section is highlighted with a blue box. The 'Input DataTable(s)' dropdown menu now contains the entry 'DT:T1Input'.

T1 and T2 Data

Input data	Calculation	Fitting	Appearance
Peak Property	height		
Calculation Options <input checked="" type="radio"/> Blank <input type="radio"/> HetNoe <input type="radio"/> R2/R1 <input type="radio"/> Reduced_Spectral_Density_Mapping			
Input data	Calculation	Fitting	Appearance
Optimiser Options Optimiser Method: leastsq			
Input data	Calculation	Fitting	Appearance
Fitting Options Fitting Model: OnePhaseDecay			
Input data	Calculation	Fitting	Appearance
Input DataTables Input DataTable Name: T1Input			
Input data	Calculation	Fitting	Appearance
Create Input DataTable			
Input data	Calculation	Fitting	Appearance
Input DataTable(s): > select-to-add <			
Input data	Calculation	Fitting	Appearance
DT:T1Input			
Input data	Calculation	Fitting	Appearance
Results DataTable Name: T1Results	Fit and Fetch Results Data	Fit	

2D Fitting

Move to the **Calculation** tab:

- Check that **Calculation Options** is set to **Blank**

Now move to the **Fitting** tab:

- Make sure the **Fitting Model** is set to **OnePhaseDecay**. Hover over this option to see information on the equation used).

A model to describe the rate of a decay.
 Model:

$$Y = \text{amplitude} \cdot \exp(-\text{rate} \cdot X)$$

 X: the various time values
 amplitude: the Y value when X (time) is zero. Same units as Y
 rate: the rate constant, expressed in reciprocal of the X axis time units, e.g.: Second-1.


Move back to the **Input data** tab:

- Enter a **Results DataTable Name**, e.g. **T1Results**
- Click on the **Fit** button to run the fitting routine.

Your **Results DataTable** will then be automatically set to your new **T1Results** table. The table and graphs in the main module will be filled with this data.

Results DataTable	
Results DataTable	DT:T1Results

T1 and T2 Data



2E Inspect Results

Now move to the Appearance tab:

- Make sure the Y Axis Data is set to **rate** from the drop-down menu.
- Close the **Settings** tab by clicking on the gear icon in the top left again.
- If the Refresh button has gone orange , then press this to refresh the graphs and table.
- If the data aren't shown in full, then press the auto-zoom icon above the bar chart to auto-scale it.
- If you click on a row in the Table or on a bar in the Bar Chart, then the matching bar/row will be selected and in the Fitting Plot you will see a graph showing your data points (in spectrum colours) and the fit.

The table contains both the data points and fitting parameters.

If you have a SpectrumDisplay open this will navigate to the peaks corresponding to the residue selected.

T1 and T2 Data

Input data **Calculation** **Fitting** **Appearance**

General Setup

Create SpectrumGroup **SpectrumGroup** SG:T1

Experiment

Experiment Name: T2

Input DataTables

Input DataTable Name: T2Input

Results DataTable Name: T2Results

Fit and Fetch Results Data

Results DataTable: DT:T2Results

#	collectionPid	exper
1	CO:A.18.LYS.H,N_1	T2
2	CO:A.19.LEU.H,N_1	T2
3	CO:A.20.ILE.H,N_1	T2
4	CO:A.21.LEU.H,N_1	T2
5	CO:A.22.ASN.H,N_1	T2
6	CO:A.23.GLY.H,N_1	T2
7	CO:A.24.LYS.H,N_1	T2
8	CO:A.25.THR.H,N_1	T2
9	CO:A.26.LEU.H,N_1	T2
10	CO:A.27.LYS.H,N_1	T2
11	CO:A.28.GLY.H,N_1	T2
12	CO:A.29.GLU.H,N_1	T2
13	CO:A.30.THR.H,N_1	T2
14	CO:A.31.THR.H,N_1	T2
15	CO:A.32.THR.H,N_1	T2
16	CO:A.33.GLU.H,N_1	T2


rate

nmrResidueCode

2F T2 Data

- Open the **Settings** panel with the gear icon.
- In the **Input data** tab:
- In the **General Setup** section, right-click on your **SG:T1** SpectrumGroup and select **Remove** or **Remove all**.
 - Repeat sections **2B-2E** for the T2 data. Use the same options adjusting T1 to T2 in each instance and selecting **T2** as your **Experiment Name**.

Heteronuclear NOE



3A Heteronuclear NOE Calculation Setup

- Open the **Settings** panel with the gear icon.

In the **Input data** tab:

- In the **General Setup** section, right-click on your **SG:T1 SpectrumGroup** and select **Remove** or **Remove all**.
- Select **SG:HetNOE** as your **SpectrumGroup** from the drop-down menu
- Click on **Setup ...** to set up your Input Collection

In the Series Peak Collection pop-up:

- Select the **SG:HetNOE** SpectrumGroup
- Give your Collection a Name, e.g. **HetNOEPeaks**
- Select to **Refit Peaks At Position**
- Click **OK**

Heteronuclear NOE

Experiment

Experiment Name	<input type="radio"/> N/A <input type="radio"/> T1 <input type="radio"/> T2 <input checked="" type="radio"/> HetNoe
-----------------	--

Input DataTables

Input DataTable Name	HetNOEInput
Create Input DataTable	Create

Input data Calculation Fitting Appearance

Peak Property: height

Calculation Options

Calculation Options	<input type="radio"/> Blank <input checked="" type="radio"/> HetNoe <input type="radio"/> R2/R1 <input type="radio"/> Reduced_Spectral_Density_Mapping
---------------------	---

Input data Calculation Fitting Appearance

Optimiser Options

Optimiser Method	leastsq
Fitting Error Method	Default

Fitting Options

Fitting Model	<input checked="" type="radio"/> Blank <input type="radio"/> OnePhaseDecay <input type="radio"/> ExponentialDecay <input type="radio"/> InversionRecovery
---------------	--

Results DataTable Name: HetNOEResults

Fit and Fetch Results Data: **Fit**

Results DataTable: DT:HetNOEResults

3B Heteronuclear NOE Calculation

In the Settings **Input data** tab:

- Set the **Experiment Name** to **HetNoe**
- Give your Input DataTable a name, e.g. **HetNOEInput**
- Click on **Create**

In the Settings **Calculation** tab:

- Select **HetNoe**

In the Settings **Fitting** tab:


- Select **Blank**

In the **Input data** tab:

- Give your Results DataTable a name, e.g. **HetNOEResults**
- Click on the **Fit** button to run the calculation.

Your **Results DataTable** will then be automatically set to your new **HetNOEResults** table. The table and graphs in the main module will be filled with this data.

Heteronuclear NOE



3B Inspecting the Heteronuclear NOE data

The bar chart should automatically show the Heteronuclear NOE on the Y-axis.

- Close the **Settings** tab by clicking on the gear icon.
- If necessary, press the auto-zoom icon above the bar chart to auto-scale it.
- If you click on a row in the Table or on a bar in the Bar Chart, then the matching bar/row will be selected.
If you have a SpectrumDisplay open this will navigate to the peaks corresponding to the residue selected.
The Fitting Plot area in the top right hand corner of the Relaxation module, will remain empty, as no fitting is required when analysing the Heteronuclear NOE data.

Combined Data Analyses

Input DataTables	
Input DataTable Name	<default>
Create Input DataTable	Create
Input DataTable(s)	> select-to-add < DT:T1Results DT:T2Results
Results DataTable Name	R2R1Results
Fit and Fetch Results Data	Fit

Input data		Calculation	Fitting	Appearance
Peak Property	height			
Calculation Options				
Calculation Options	<input type="radio"/> Blank <input type="radio"/> HetNoe <input checked="" type="radio"/> R2/R1 <input type="radio"/> Reduced_Spectral_Density_Mapping			


Input data		Calculation	Fitting	Appearance
Optimiser Options				
Optimiser Method	leastsq			
Fitting Error Method	Default			
Fitting Options				
Fitting Model	<input checked="" type="radio"/> Blank <input type="radio"/> OnePhaseDecay <input type="radio"/> ExponentialDecay <input type="radio"/> InversionRecovery			

4A R2/R1 Calculation Setup

At this point you can either continue from Section 3 or start using our **Section3_completed ccpn** project.

- Open the **Relaxation Module Settings** panel with the gear icon.
 - In the **Input data** tab, select the **Input DataTables** to be
 - DT:T1Results**
 - DT:T2Results**
 - If need be, remove the **DT:HetNOEResults** DataTable by right-clicking on it and selecting **Remove**.
 - Provide a **Results DataTable Name**, e.g. **R2R1Results**
- In the **Fitting** tab:
- Set your **Calculation Options** to be **R2/R1**.
- In the **Fitting** tab:
- Make sure the **Fitting Model** is set to **Blank**.

Combined Data Analyses



4B R2/R1 Calculation

In the **Settings Input data** tab:

- Click on the **Fit** button.
- Close the Settings tab with the gear icon and inspect your results in the main part of the module.
- If necessary, press the auto-zoom icon above the bar chart to auto-scale it.
- If you click on a row in the Table or on a bar in the Bar Chart, then the matching bar/row will be selected.

If you have a SpectrumDisplay open this will navigate to the peaks corresponding to the residue selected.

The Fitting Plot area in the top right hand corner of the Relaxation module, will remain empty, as no fitting was required for this calculation.

Reduced Spectral Density Mapping

Input DataTables	
Input DataTable Name	<default>
Create Input DataTable	<input type="button" value="Create"/>
Input DataTable(s)	<input type="button" value="> select-to-add <"/> DT:T1Results DT:T2Results DT:HetNOEResults
Results DataTable Name	RDSMResults
Fit and Fetch Results Data	<input type="button" value="Fit"/>

Input data		Calculation	Fitting	Appearance
Peak Property	<input type="text" value="height"/> Calculation Options			
Calculation Options	<input type="radio"/> Blank <input type="radio"/> HetNoe <input type="radio"/> R2/R1 <input checked="" type="radio"/> Reduced_Spectral_Density_Mapping			

Input data		Calculation	Fitting	Appearance
Optimiser Options				
Optimiser Method	<input type="text" value="leastsq"/>			
Fitting Error Method	<input type="text" value="Default"/>			
Fitting Options				
Fitting Model	<input checked="" type="radio"/> Blank <input type="radio"/> OnePhaseDecay <input type="radio"/> ExponentialDecay <input type="radio"/> InversionRecovery			

5A Reduced Spectral Density Mapping Setup

At this point you can either continue from Section 4 or start using our **Section4_completed ccpn** project.

- Open the Relaxation Module Settings panel with the gear icon.
 - In the **Input data** tab, select the **Input DataTables** to be
 - DT:T1Results**
 - DT:T2Results**
 - DT:HetNOEResults**
 - Provide a **Results DataTable Name**, e.g. **RSDMResults**
- In the **Calculation** tab:
- Set your **Calculation Options** to be **Reduced_Spectral_Density_Mapping**
- In the **Fitting** tab:
- Make sure the **Fitting Model** is set to **Blank**.

Reduced Spectral Density Mapping

Input DataTables	
Input DataTable Name	<default>
Create Input DataTable	<input type="button" value="Create"/>
Input DataTable(s)	<input type="button" value="> select-to-add <"/> DT:T1Results DT:T2Results DT:HetNOEResults
Results DataTable Name	RDSMResults
Fit and Fetch Results Data	<input type="button" value="Fit"/>

Input data	Calculation	Fitting	Appearance
SpectrumDisplay			
Select SpectrumDisplays	<input type="button" value="> select-to-add <"/> <input type="button" value="<Use active>"/>		
Navigate trigger	<input checked="" type="radio"/> Single Click <input type="radio"/> Double Click <input type="radio"/> Disabled		
BarGraph			
X Axis Data	nmrResidueCode		
Y Axis Data	<input type="button" value="J0"/> <input checked="" type="button" value="J0"/> JwX JwH argA argB Rsqr chisqr redchi aic bic		
Threshold Value Calculation			
Threshold Factor			
Threshold Value			
Above Threshold Colour			
Below Threshold Colour			
Intraceable Observation Col	 CCPNpurple		

5_B Reduced Spectral Density Mapping Calculation

In the **Settings Input data** tab:

- Click on the Fit button.

In the **Appearance** tab:

- From the **Y Axis Data** drop-down menu select J0, JwX and JwH in turn to view the spectral densities at 0, ω_N and ω_H , respectively.
- If necessary, press the orange Update button  or auto-zoom button  to view the data on the bar chart.


The Table, Bar Chart and Spectrum Display are dynamically linked as usual and no fits are present.

6 Relaxation Exchange via $\eta_{xy/z}$

Relaxation Exchange rates determination via δ_{NH} and $\eta_{xy/z}$ analysis

- Spectra
 - ▷ SP:T1_1
 - ▷ SP:T1_2
 - ▷ SP:T1_3
 - ▷ SP:T1_4
 - ▷ SP:T1_5
 - ▷ SP:T1_6
 - ▷ SP:T1_7
 - ▷ SP:T1_8
 - ▷ SP:T1_9
 - ▷ SP:T1_10
 - ▷ SP:T2_1
 - ▷ SP:T2_2
 - ▷ SP:T2_3
 - ▷ SP:T2_4
 - ▷ SP:T2_5
 - ▷ SP:T2_6
 - ▷ SP:T2_7
 - ▷ SP:T2_8
 - ▷ SP:T2_9
 - ▷ SP:unsat
 - ▷ SP:sat
 - ▷ SP:GB1_HSQC
 - ▷ SP:ETAz_AP_1
 - ▷ SP:ETAz_AP_2
 - ▷ SP:ETAz_AP_3
 - ▷ SP:ETAz_AP_4
 - ▷ SP:ETAz_AP_5
 - ▷ SP:ETAz_IP_1
 - ▷ SP:ETAz_IP_2
 - ▷ SP:ETAz_IP_3
 - ▷ SP:ETAz_IP_4
 - ▷ SP:ETAz_IP_5
 - ▷ SP:ETAxz_AP_1
 - ▷ SP:ETAxz_AP_2
 - ▷ SP:ETAxz_AP_3
 - ▷ SP:ETAxz_AP_4
 - ▷ SP:ETAxz_AP_5
 - ▷ SP:ETAxz_AP_6
 - ▷ SP:ETAxz_IP_1
 - ▷ SP:ETAxz_IP_2
 - ▷ SP:ETAxz_IP_3
 - ▷ SP:ETAxz_IP_4
 - ▷ SP:ETAxz_IP_5
 - ▷ SP:ETAxz_IP_6
- SpectrumGroups
 - <New SpectrumGr...
 - ▷ SG:T1
 - ▷ SG:T2
 - ▷ SG:HetNoe
 - ▷ SG:ETAz_AP
 - ▷ SG:ETAz_IP
 - ▷ SG:ETAxz_AP
 - ▷ SG:ETAxz_IP

In this section is described how to setup the analysis for obtaining the $\eta_{xy/z}$ ratios from in-phase and anti-phase HSQC spectra



6A η_{xy} and η_z analysis setup


- If not continuing from step 5 of this tutorial, open the **Section5_completed ccpn** project.
- Load the Excel file **ETAs_Lookup.xls** from the spectra/GB1_ETAs directory. Two new spectrumGroups and their spectra will appear on the sidebar
- Go to **Main Menu → Macro → Run CCPN Macros → SetupETAsInputData**.
Allow a couple of minutes for the macro to run.

This macro will produce 2 new output/result DataTables containing the η_{xy} and η_z values. (For more details on what the macro does, open it with the macroEditor from the main menu.)

- Open the RelaxationAnalysis module (Go to **Main Menu → Macro → Run CCPN Macros → RelaxationModule**) and select the the new result dataTables to inspect the data.

As part of the following step, you will extract the Relaxation Exchange rates.

Exporting Graphs



* DSSP secondary structure coding:

<i>H</i> : <i>a</i> -helix	<i>G</i> : β_{10} helix
<i>I</i> : π -helix	<i>E</i> : extended strand
<i>B</i> : residue in isolated β -bridge	<i>S</i> : bend
<i>T</i> : <i>H</i> -bonded turn	<i>C</i> : no secondary structure

7A Export predefined plots in pdf format


At this point you can either continue from Section 6 or start using our [Section6_completed ccpn](#) project.

- Go to Main Menu → Macro → Run CCPN Macros → **Relaxation Plots Popup**
- As your **DataTable**, select your Reduced Spectral Density Mapping Results table (**DT:RSDMResults**)
- Select a Directory where you want to save your plots
The Sequence and Secondary Structure (in DSSP code*) for GB1 are already provided in the pop-up.
- Keep all **Reports** selected
- Click on **Generate**

7B Calculate and Export the Relaxation Exchange rates

- Go to Main Menu → Macro → Run CCPN Macros → **Calculate_RelaxationExchange_via_ETAs**

Exporting Graphs




7C Exporting data for use in external graphing programs

- Right-click on the table and select either **Export Visible Table** or **Export All Columns**. The former is sensitive to any filtering you might have done on the table while the latter is not.


7D Exporting individual charts

- Right-click on the bar chart or fitting graph and select **Export**.

You will be presented with a small pop-up offering you several options, including the ability to export in an image or a scalable vector graphics format.



Examining the Graphs



8 Examine the graphs

The graphs are also available in the **graphs** folder of the tutorial data directory.

- Examine the various graph that you have produced
- **For each graph**, summarise what kind of data they display and what you can learn from them.
- What is your overall assessment of the dynamics of GB1? Is it adequately described by a small sphere tumbling in solution?

Contact Us

Website:

www ccpn ac uk

Suggestions and comments:

support@ccpn.ac.uk

Issues and bug report:

<https://forum.ccpn.ac.uk/>

Cite Us

Skinner, S. P. et al. CcpNmr AnalysisAssign: a flexible platform for integrated NMR analysis. *J. Biomol. NMR* (2016). doi:10.1007/s10858-016-0060-y