


S25 vs S26 — STRESS-STRAIN REFERENCE

Every recorded data point of both frame-capture fracture runs, for line-by-line comparison against extensometer software



HEADLINE NUMBERS

	S25 · VC2	S26 · VC3
UTS	46.16 MPa	47.17 MPa
at strain	2.19 %	2.44 %
σ_y (0.2 %)	41.68 MPa	46.47 MPa
at strain	1.50 %	1.92 %
E (fixed window)	2.976 GPa	2.508 GPa
ϵ_f	4.18 %	5.74 %
σ at fracture	41.59 MPa	42.54 MPa
Toughness	1564 kJ/m ³	2309 kJ/m ³
Force anchor	280 N	308 N
Load samples	1648	1421
Captured frames	1445	1682

WHAT IS IN THIS DOCUMENT


- p2 Why the reported E differs between the two runs — and why σ_y follows it
- p3 COMPARISON TABLE — both runs resampled onto one common strain axis (0.10 % steps), the form to read across
- p4-6 FULL RESOLUTION · S25 — all 768 raw samples in acquisition order
- p7-10 FULL RESOLUTION · S26 — all 903 raw samples in acquisition order

Stress is ENGINEERING stress (force ÷ 80 mm² nominal), anchor-corrected: the preload tared away at the start of the test is added back, which is why these values exceed the tared reading in the CSV header. Strain is DIC gauge strain over an 80 mm gauge, zeroed at Px0. Both are exactly what the app's own report button computes — `utm_analysis.analyze()`, no separate maths.

Highlighted rows: ■ Yield (0.2 % offset) ■ UTS (peak stress) ■ Fracture

WHY THE TWO RUNS REPORT A DIFFERENT E

The fit window is fixed; the toe it lands on is not



analyze() fits E over a FIXED 0.05–0.40 % strain window. A specimen that seats more slowly has a longer toe, so that window lands further down the curve and returns a softer modulus. Searching instead for each specimen's own straightest run:

S25 fixed window 2.976 GPa → straightest run 3.123 GPa over 0.13–0.65 % (R^2 0.9998)

S26 fixed window 2.508 GPa → straightest run 3.464 GPa over 0.33–0.85 % (R^2 0.9996)

σ_y is a CONSEQUENCE of this. The 0.2 % offset line has slope E, so a softer E tilts it down, meets the curve later, and pushes σ_y toward UTS: S25 reports σ_y at 90.3 % of its UTS, S26 at 98.5 %. The σ_y spread (41.7 vs 46.5 MPa) is largely the E spread re-expressed.

UTS needs no fit at all and agrees to 2.2 % — it is the number to quote when comparing against the extensometer.

COMPARISON TABLE — COMMON STRAIN AXIS (1/1)

Both runs resampled to 0.10 % strain steps · 64 rows · exact yield / UTS / fracture samples inserted at their true strain

ϵ (%)	S25 σ	S26 σ	Δ	landmark
0.00	—	—	—	
0.10	6.20	5.84	-0.37	
0.20	8.86	8.28	-0.58	
0.30	12.05	10.64	-1.41	
0.40	15.11	13.64	-1.48	
0.50	18.34	17.12	-1.22	
0.60	21.31	20.59	-0.71	
0.70	24.14	24.27	+0.14	
0.80	26.90	27.56	+0.66	
0.90	29.45	30.60	+1.15	
1.00	31.81	33.47	+1.67	
1.10	34.11	35.89	+1.79	
1.20	36.23	38.12	+1.89	
1.30	38.25	40.16	+1.92	
1.40	40.10	41.78	+1.68	
1.50	41.67	43.13	+1.47	
1.50	41.68	43.15	+1.47	S25 yield
1.60	43.05	44.27	+1.22	
1.70	44.13	45.18	+1.05	
1.80	44.97	45.88	+0.92	
1.90	45.57	46.39	+0.82	
1.92	45.67	46.47	+0.80	S26 yield

ϵ (%)	S25 σ	S26 σ	Δ	landmark
2.00	45.94	46.73	+0.79	
2.10	46.11	46.95	+0.84	
2.19	46.16	47.06	+0.91	S25 UTS
2.20	46.16	47.08	+0.92	
2.30	46.07	47.13	+1.06	
2.40	45.94	47.16	+1.22	
2.44	45.87	47.17	+1.30	S26 UTS
2.50	45.75	47.15	+1.39	
2.60	45.54	47.11	+1.57	
2.70	45.31	47.04	+1.73	
2.80	45.10	46.96	+1.86	
2.90	44.87	46.86	+1.99	
3.00	44.66	46.74	+2.08	
3.10	44.45	46.61	+2.17	
3.20	44.23	46.47	+2.24	
3.30	44.05	46.32	+2.28	
3.40	43.85	46.17	+2.32	
3.50	43.67	46.03	+2.36	
3.60	43.47	45.89	+2.42	
3.70	43.27	45.74	+2.47	
3.80	43.06	45.60	+2.54	
3.90	42.83	45.46	+2.63	

ϵ (%)	S25 σ	S26 σ	Δ	landmark
4.00	42.54	45.32	+2.77	
4.10	42.17	45.17	+3.00	
4.18	41.59	45.07	+3.48	S25 fracture
4.20	—	45.03	—	
4.30	—	44.90	—	
4.40	—	44.79	—	
4.50	—	44.66	—	
4.60	—	44.54	—	
4.70	—	44.41	—	
4.80	—	44.29	—	
4.90	—	44.17	—	
5.00	—	44.02	—	
5.10	—	43.89	—	
5.20	—	43.76	—	
5.30	—	43.62	—	
5.40	—	43.47	—	
5.50	—	43.28	—	
5.60	—	43.08	—	
5.70	—	42.76	—	
5.75	—	42.54	—	S26 fracture

σ in MPa. Δ = S26 — S25. “—” means that strain is past that specimen's fracture, so it has no stress there.

Highlighted rows: Yield (0.2 % offset) UTS (peak stress) Fracture

FULL RESOLUTION — S25 · VC2 (1/3)

All 768 recorded samples in acquisition order · UTS 46.16 MPa · σ_y 41.68 MPa · E 2.976 GPa · ϵ_f 4.18 % · 1445 frames captured

#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)
1	0.0024	3.525	54	0.1094	6.420	107	0.2360	9.917	160	0.3637	14.039	213	0.5069	18.493
2	0.0040	3.633	55	0.1139	6.480	108	0.2381	10.063	161	0.3660	14.118	214	0.5099	18.580
3	0.0041	3.551	56	0.1177	6.541	109	0.2410	10.138	162	0.3696	14.361	215	0.5104	18.667
4	0.0053	3.589	57	0.1202	6.605	110	0.2429	10.213	163	0.3698	14.201	216	0.5128	18.750
5	0.0070	3.681	58	0.1240	6.671	111	0.2473	10.289	164	0.3709	14.280	217	0.5165	18.840
6	0.0090	3.960	59	0.1273	6.733	112	0.2479	10.359	165	0.3754	14.443	218	0.5200	18.927
7	0.0095	3.743	60	0.1286	6.793	113	0.2536	10.432	166	0.3805	14.524	219	0.5268	19.011
8	0.0130	3.822	61	0.1347	6.853	114	0.2551	10.506	167	0.3824	14.685	220	0.5310	19.267
9	0.0136	3.945	62	0.1367	6.913	115	0.2572	10.580	168	0.3837	14.607	221	0.5316	19.179
10	0.0148	3.907	63	0.1377	6.976	116	0.2588	10.655	169	0.3896	14.765	222	0.5322	19.095
11	0.0168	3.988	64	0.1382	7.042	117	0.2621	10.727	170	0.3936	14.849	223	0.5329	19.359
12	0.0214	4.048	65	0.1427	7.111	118	0.2632	10.871	171	0.3943	14.935	224	0.5389	19.448
13	0.0249	4.055	66	0.1432	7.171	119	0.2654	10.799	172	0.3978	15.019	225	0.5404	19.530
14	0.0259	4.161	67	0.1440	7.298	120	0.2695	10.950	173	0.3991	15.103	226	0.5418	19.702
15	0.0282	4.089	68	0.1442	7.231	121	0.2695	11.024	174	0.4060	15.184	227	0.5436	19.613
16	0.0296	4.225	69	0.1482	7.365	122	0.2720	11.097	175	0.4070	15.268	228	0.5452	19.793
17	0.0311	4.276	70	0.1506	7.427	123	0.2752	11.173	176	0.4100	15.432	229	0.5536	19.880
18	0.0323	4.386	71	0.1532	7.492	124	0.2760	11.252	177	0.4115	15.352	230	0.5554	19.967
19	0.0328	4.327	72	0.1542	7.558	125	0.2808	11.331	178	0.4128	15.595	231	0.5570	20.056
20	0.0342	4.494	73	0.1604	7.624	126	0.2811	11.401	179	0.4131	15.515	232	0.5593	20.140
21	0.0353	4.438	74	0.1604	7.692	127	0.2847	11.476	180	0.4165	15.678	233	0.5640	20.226
22	0.0407	4.607	75	0.1663	7.763	128	0.2863	11.552	181	0.4183	15.768	234	0.5668	20.316
23	0.0417	4.553	76	0.1679	7.832	129	0.2909	11.708	182	0.4216	15.850	235	0.5694	20.489
24	0.0438	4.663	77	0.1701	7.899	130	0.2916	11.630	183	0.4239	15.936	236	0.5700	20.403
25	0.0458	4.717	78	0.1729	7.965	131	0.2956	11.780	184	0.4245	16.017	237	0.5754	20.578
26	0.0501	4.774	79	0.1747	8.028	132	0.2961	11.855	185	0.4301	16.099	238	0.5776	20.671
27	0.0520	4.834	80	0.1751	8.091	133	0.2998	12.086	186	0.4354	16.181	239	0.5802	20.755
28	0.0540	4.890	81	0.1781	8.233	134	0.3008	11.929	187	0.4385	16.267	240	0.5860	20.928
29	0.0554	5.000	82	0.1802	8.160	135	0.3024	12.162	188	0.4388	16.356	241	0.5861	21.019
30	0.0564	4.944	83	0.1806	8.304	136	0.3028	12.008	189	0.4418	16.442	242	0.5861	20.837
31	0.0584	5.061	84	0.1840	8.373	137	0.3040	12.241	190	0.4447	16.525	243	0.5899	21.104
32	0.0644	5.119	85	0.1851	8.445	138	0.3091	12.317	191	0.4466	16.608	244	0.5949	21.273
33	0.0656	5.288	86	0.1899	8.511	139	0.3119	12.392	192	0.4508	16.691	245	0.5952	21.187
34	0.0661	5.232	87	0.1926	8.714	140	0.3122	12.466	193	0.4541	16.778	246	0.6024	21.365
35	0.0677	5.176	88	0.1935	8.575	141	0.3202	12.546	194	0.4550	16.864	247	0.6053	21.453
36	0.0681	5.344	89	0.1939	8.642	142	0.3216	12.621	195	0.4596	16.949	248	0.6064	21.539
37	0.0705	5.403	90	0.1978	8.784	143	0.3216	12.696	196	0.4613	17.123	249	0.6120	21.624
38	0.0721	5.465	91	0.1996	8.851	144	0.3253	12.779	197	0.4613	17.035	250	0.6137	21.711
39	0.0735	5.522	92	0.2021	8.920	145	0.3259	12.937	198	0.4646	17.288	251	0.6145	21.800
40	0.0740	5.580	93	0.2075	8.987	146	0.3266	12.860	199	0.4649	17.204	252	0.6168	21.887
41	0.0806	5.639	94	0.2082	9.061	147	0.3306	13.012	200	0.4683	17.380	253	0.6206	21.981
42	0.0861	5.698	95	0.2096	9.133	148	0.3329	13.083	201	0.4707	17.465	254	0.6247	22.068
43	0.0871	5.817	96	0.2110	9.202	149	0.3368	13.164	202	0.4733	17.548	255	0.6269	22.155
44	0.0877	5.759	97	0.2157	9.270	150	0.3395	13.244	203	0.4799	17.633	256	0.6309	22.244
45	0.0884	5.875	98	0.2162	9.343	151	0.3400	13.322	204	0.4827	17.719	257	0.6329	22.329
46	0.0905	5.935	99	0.2175	9.413	152	0.3429	13.481	205	0.4842	17.803	258	0.6380	22.418
47	0.0940	6.117	100	0.2193	9.487	153	0.3441	13.401	206	0.4884	17.892	259	0.6412	22.512
48	0.0950	5.995	101	0.2231	9.555	154	0.3490	13.561	207	0.4905	17.978	260	0.6418	22.598
49	0.0974	6.058	102	0.2258	9.627	155	0.3506	13.639	208	0.4931	18.062	261	0.6439	22.682
50	0.0980	6.175	103	0.2307	9.701	156	0.3579	13.719	209	0.4937	18.147	262	0.6477	22.771
51	0.1022	6.237	104	0.2334	9.845	157	0.3608	13.799	210	0.4951	18.234	263	0.6512	22.856
52	0.1050	6.297	105	0.2356	9.772	158	0.3628	13.958	211	0.4987	18.323	264	0.6524	22.946
53	0.1074	6.356	106	0.2357	9.991	159	0.3631	13.880	212	0.5057	18.410	265	0.6608	23.036

Highlighted rows:  Yield (0.2 % offset)

 UTS (peak stress)

 Fracture

Y = yield U = UTS F = fracture

FULL RESOLUTION — S25 · VC2 (2/3)

All 768 recorded samples in acquisition order · UTS 46.16 MPa · σ_y 41.68 MPa · E 2.976 GPa · ϵ_f 4.18 % · 1445 frames captured

#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)
266	0.6644	23.124	319	0.8344	27.758	372	1.0175	32.360	425	1.2220	36.772	478	1.4438	40.841
267	0.6658	23.297	320	0.8384	27.848	373	1.0204	32.443	426	1.2285	36.853	479	1.4468	40.914
268	0.6669	23.211	321	0.8414	27.939	374	1.0269	32.526	427	1.2344	37.014	480	1.4531	40.981
269	0.6694	23.386	322	0.8452	28.033	375	1.0296	32.613	428	1.2347	36.933	481	1.4561	41.053
270	0.6705	23.474	323	0.8512	28.204	376	1.0371	32.696	429	1.2364	37.094	482	1.4663	41.122
271	0.6737	23.557	324	0.8514	28.118	377	1.0388	32.781	430	1.2464	37.257	483	1.4663	41.196
272	0.6792	23.641	325	0.8540	28.291	378	1.0435	32.868	431	1.2486	37.173	484	1.4700	41.266
273	0.6807	23.825	326	0.8555	28.380	379	1.0471	32.955	432	1.2588	37.413	485	1.4770	41.333
274	0.6812	23.735	327	0.8588	28.472	380	1.0508	33.041	433	1.2589	37.338	486	1.4799	41.402
275	0.6886	23.909	328	0.8654	28.559	381	1.0545	33.126	434	1.2613	37.491	487	1.4816	41.473
276	0.6930	23.998	329	0.8678	28.643	382	1.0594	33.209	435	1.2701	37.576	488	1.4839	41.542
277	0.6996	24.255	330	0.8746	28.728	383	1.0638	33.296	436	1.2716	37.654	489	1.4945	41.609
278	0.6997	24.086	331	0.8748	28.817	384	1.0706	33.382	437	1.2716	37.729	490	1.5014	41.681 Y
279	0.7002	24.170	332	0.8804	28.909	385	1.0744	33.468	438	1.2823	37.808	491	1.5033	41.753
280	0.7047	24.339	333	0.8839	29.168	386	1.0781	33.548	439	1.2853	37.892	492	1.5077	41.817
281	0.7083	24.429	334	0.8842	28.995	387	1.0791	33.629	440	1.2886	37.972	493	1.5113	41.883
282	0.7133	24.521	335	0.8862	29.081	388	1.0794	33.717	441	1.2900	38.050	494	1.5171	41.954
283	0.7155	24.609	336	0.8932	29.250	389	1.0834	33.800	442	1.2928	38.123	495	1.5189	42.024
284	0.7171	24.693	337	0.8970	29.339	390	1.0888	33.883	443	1.2979	38.199	496	1.5265	42.088
285	0.7202	24.782	338	0.8993	29.425	391	1.0911	33.968	444	1.3018	38.285	497	1.5299	42.153
286	0.7237	24.868	339	0.9014	29.512	392	1.0937	34.046	445	1.3035	38.366	498	1.5363	42.221
287	0.7247	24.955	340	0.9082	29.598	393	1.1022	34.129	446	1.3054	38.443	499	1.5401	42.286
288	0.7282	25.041	341	0.9099	29.689	394	1.1030	34.217	447	1.3092	38.516	500	1.5467	42.352
289	0.7316	25.124	342	0.9156	29.774	395	1.1069	34.299	448	1.3154	38.593	501	1.5492	42.419
290	0.7341	25.212	343	0.9171	29.863	396	1.1080	34.383	449	1.3201	38.667	502	1.5521	42.484
291	0.7390	25.304	344	0.9192	29.950	397	1.1140	34.468	450	1.3260	38.744	503	1.5603	42.552
292	0.7424	25.392	345	0.9217	30.037	398	1.1176	34.558	451	1.3306	38.827	504	1.5631	42.621
293	0.7499	25.480	346	0.9270	30.120	399	1.1236	34.641	452	1.3338	38.904	505	1.5740	42.682
294	0.7508	25.569	347	0.9343	30.207	400	1.1282	34.722	453	1.3377	38.981	506	1.5782	42.740
295	0.7548	25.657	348	0.9351	30.292	401	1.1337	34.808	454	1.3439	39.056	507	1.5832	42.800
296	0.7570	25.744	349	0.9418	30.378	402	1.1379	34.891	455	1.3484	39.135	508	1.5883	42.862
297	0.7629	25.832	350	0.9429	30.469	403	1.1414	34.971	456	1.3504	39.211	509	1.5911	42.925
298	0.7687	26.002	351	0.9494	30.557	404	1.1428	35.140	457	1.3572	39.293	510	1.5923	42.992
299	0.7689	25.920	352	0.9499	30.644	405	1.1452	35.054	458	1.3591	39.367	511	1.6014	43.056
300	0.7738	26.090	353	0.9553	30.734	406	1.1507	35.221	459	1.3643	39.440	512	1.6064	43.119
301	0.7770	26.179	354	0.9553	30.818	407	1.1553	35.302	460	1.3681	39.513	513	1.6066	43.178
302	0.7782	26.267	355	0.9580	30.901	408	1.1579	35.383	461	1.3700	39.590	514	1.6135	43.241
303	0.7815	26.357	356	0.9633	30.986	409	1.1623	35.469	462	1.3734	39.668	515	1.6178	43.304
304	0.7839	26.444	357	0.9654	31.072	410	1.1670	35.551	463	1.3755	39.746	516	1.6266	43.360
305	0.7889	26.536	358	0.9732	31.156	411	1.1683	35.632	464	1.3834	39.818	517	1.6321	43.419
306	0.7920	26.625	359	0.9758	31.244	412	1.1695	35.715	465	1.3850	39.894	518	1.6343	43.475
307	0.7967	26.798	360	0.9768	31.330	413	1.1745	35.795	466	1.3921	39.970	519	1.6393	43.531
308	0.7973	26.712	361	0.9802	31.421	414	1.1808	35.875	467	1.3951	40.042	520	1.6439	43.593
309	0.7997	26.886	362	0.9854	31.512	415	1.1830	35.963	468	1.4008	40.113	521	1.6535	43.652
310	0.8023	26.970	363	0.9897	31.674	416	1.1871	36.043	469	1.4011	40.186	522	1.6581	43.710
311	0.8082	27.053	364	0.9911	31.593	417	1.1924	36.124	470	1.4094	40.259	523	1.6637	43.769
312	0.8112	27.143	365	0.9960	31.759	418	1.1988	36.208	471	1.4130	40.331	524	1.6685	43.825
313	0.8166	27.235	366	1.0028	31.843	419	1.2028	36.290	472	1.4159	40.403	525	1.6738	43.875
314	0.8194	27.323	367	1.0035	31.926	420	1.2080	36.372	473	1.4242	40.476	526	1.6846	43.933
315	0.8210	27.413	368	1.0096	32.018	421	1.2080	36.453	474	1.4303	40.553	527	1.6890	43.989
316	0.8262	27.501	369	1.0136	32.106	422	1.2119	36.530	475	1.4331	40.696	528	1.6948	44.044
317	0.8295	27.670	370	1.0136	32.194	423	1.2198	36.689	476	1.4375	40.624	529	1.6948	44.100
318	0.8300	27.588	371	1.0145	32.277	424	1.2218	36.607	477	1.4387	40.769	530	1.7034	44.155

Highlighted rows:  Yield (0.2 % offset)

 UTS (peak stress)

 Fracture

Y = yield U = UTS F = fracture

FULL RESOLUTION — S25 · VC2 (3/3)

All 768 recorded samples in acquisition order · UTS 46.16 MPa · σ_y 41.68 MPa · E 2.976 GPa · ϵ_f 4.18 % · 1445 frames captured

#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)
531	1.7102	44.205	584	2.0798	46.093	637	2.6268	45.476	690	3.2287	44.178	743	3.8489	42.951
532	1.7142	44.258	585	2.0893	46.107	638	2.6367	45.454	691	3.2394	44.159	744	3.8639	42.920
533	1.7182	44.315	586	2.0995	46.112	639	2.6473	45.433	692	3.2553	44.139	745	3.8662	42.894
534	1.7220	44.366	587	2.1057	46.119	640	2.6614	45.405	693	3.2652	44.120	746	3.8870	42.868
535	1.7309	44.417	588	2.1195	46.132	641	2.6771	45.376	694	3.2810	44.098	747	3.8938	42.836
536	1.7362	44.466	589	2.1280	46.140	642	2.6846	45.350	695	3.2867	44.071	748	3.9130	42.803
537	1.7428	44.517	590	2.1280	46.134	643	2.6943	45.330	696	3.3010	44.045	749	3.9156	42.769
538	1.7449	44.571	591	2.1399	46.143	644	2.7021	45.305	697	3.3115	44.021	750	3.9301	42.743
539	1.7532	44.619	592	2.1490	46.143	645	2.7168	45.275	698	3.3192	43.997	751	3.9410	42.715
540	1.7571	44.667	593	2.1629	46.149	646	2.7287	45.251	699	3.3319	43.975	752	3.9524	42.680
541	1.7698	44.716	594	2.1708	46.150	647	2.7407	45.224	700	3.3413	43.951	753	3.9706	42.644
542	1.7740	44.766	595	2.1834	46.151	648	2.7530	45.200	701	3.3605	43.934	754	3.9806	42.607
543	1.7825	44.816	596	2.1894	46.156	649	2.7607	45.175	702	3.3707	43.908	755	3.9911	42.567
544	1.7857	44.860	597	2.2041	46.156	650	2.7779	45.146	703	3.3745	43.887	756	4.0048	42.528
545	1.7921	44.905	598	2.2093	46.146	651	2.7847	45.124	704	3.3897	43.870	757	4.0145	42.491
546	1.7975	44.954	599	2.2238	46.142	652	2.7994	45.100	705	3.4047	43.847	758	4.0285	42.450
547	1.8063	44.996	600	2.2276	46.138	653	2.8106	45.078	706	3.4132	43.821	759	4.0439	42.352
548	1.8084	45.041	601	2.2379	46.129	654	2.8159	45.052	707	3.4264	43.797	760	4.0439	42.402
549	1.8130	45.081	602	2.2462	46.125	655	2.8269	45.025	708	3.4447	43.777	761	4.0648	42.297
550	1.8221	45.125	603	2.2592	46.124	656	2.8424	44.995	709	3.4521	43.760	762	4.0807	42.242
551	1.8338	45.169	604	2.2758	46.116	657	2.8478	44.972	710	3.4537	43.738	763	4.0990	42.174
552	1.8407	45.213	605	2.2880	46.102	658	2.8629	44.950	711	3.4700	43.712	764	4.1121	42.097
553	1.8418	45.252	606	2.2918	46.089	659	2.8744	44.924	712	3.4883	43.690	765	4.1176	42.012
554	1.8484	45.293	607	2.2954	46.076	660	2.8859	44.898	713	3.5029	43.671	766	4.1419	41.906
555	1.8589	45.333	608	2.3031	46.069	661	2.9000	44.870	714	3.5029	43.649	767	4.1601	41.773
556	1.8623	45.374	609	2.3250	46.054	662	2.9145	44.847	715	3.5220	43.624	768	4.1756	41.593
557	1.8712	45.412	610	2.3370	46.039	663	2.9233	44.823	716	3.5300	43.598	F		
558	1.8784	45.447	611	2.3435	46.032	664	2.9352	44.797	717	3.5488	43.575			
559	1.8829	45.487	612	2.3496	46.016	665	2.9443	44.773	718	3.5544	43.558			
560	1.8873	45.523	613	2.3612	45.999	666	2.9587	44.749	719	3.5629	43.536			
561	1.8975	45.555	614	2.3735	45.982	667	2.9651	44.723	720	3.5761	43.512			
562	1.9046	45.588	615	2.3843	45.969	668	2.9815	44.699	721	3.5880	43.491			
563	1.9121	45.619	616	2.3945	45.950	669	2.9962	44.676	722	3.6002	43.468			
564	1.9210	45.650	617	2.4069	45.931	670	3.0006	44.654	723	3.6151	43.444			
565	1.9275	45.687	618	2.4153	45.912	671	3.0198	44.632	724	3.6264	43.421			
566	1.9328	45.720	619	2.4314	45.889	672	3.0236	44.607	725	3.6347	43.399			
567	1.9441	45.780	620	2.4405	45.868	673	3.0298	44.575	726	3.6475	43.373			
568	1.9443	45.750	621	2.4496	45.847	674	3.0496	44.554	727	3.6582	43.350			
569	1.9591	45.807	622	2.4595	45.828	675	3.0580	44.534	728	3.6759	43.330			
570	1.9662	45.832	623	2.4737	45.807	676	3.0701	44.507	729	3.6802	43.305			
571	1.9691	45.858	624	2.4849	45.784	677	3.0814	44.479	730	3.6948	43.279			
572	1.9807	45.878	625	2.4977	45.763	678	3.0958	44.456	731	3.7096	43.255			
573	1.9884	45.903	626	2.5036	45.743	679	3.1053	44.431	732	3.7131	43.230			
574	1.9950	45.927	627	2.5212	45.695	680	3.1157	44.408	733	3.7344	43.204			
575	2.0020	45.947	628	2.5212	45.716	681	3.1291	44.389	734	3.7378	43.181			
576	2.0090	45.964	629	2.5400	45.673	682	3.1395	44.363	735	3.7448	43.158			
577	2.0173	45.983	630	2.5465	45.650	683	3.1466	44.340	736	3.7623	43.134			
578	2.0320	46.003	631	2.5632	45.625	684	3.1611	44.317	737	3.7828	43.107			
579	2.0377	46.019	632	2.5686	45.601	685	3.1763	44.293	738	3.7876	43.083			
580	2.0438	46.040	633	2.5800	45.575	686	3.1882	44.270	739	3.8005	43.059			
581	2.0552	46.055	634	2.5956	45.551	687	3.1913	44.246	740	3.8117	43.030			
582	2.0604	46.068	635	2.6041	45.523	688	3.2060	44.225	741	3.8190	42.998			
583	2.0743	46.078	636	2.6188	45.498	689	3.2167	44.200	742	3.8350	42.974			

Highlighted rows: Yield (0.2 % offset)

 UTS (peak stress)

 Fracture

Y = yield U = UTS F = fracture

FULL RESOLUTION — S26 · VC3 (1/4)

All 903 recorded samples in acquisition order · UTS 47.17 MPa · σ_y 46.47 MPa · E 2.508 GPa · ϵ_f 5.74 % · 1682 frames captured

#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)
1	0.0044	3.866	54	0.1898	8.017	107	0.3702	12.753	160	0.5123	17.649	213	0.6508	22.490
2	0.0055	3.893	55	0.1965	8.102	108	0.3744	12.848	161	0.5202	17.742	214	0.6534	22.682
3	0.0089	3.925	56	0.1980	8.185	109	0.3794	12.943	162	0.5214	17.833	215	0.6544	22.770
4	0.0107	3.964	57	0.2000	8.278	110	0.3847	13.035	163	0.5221	17.926	216	0.6604	22.862
5	0.0149	4.010	58	0.2095	8.368	111	0.3877	13.120	164	0.5262	18.017	217	0.6624	22.954
6	0.0196	4.069	59	0.2115	8.455	112	0.3882	13.304	165	0.5282	18.110	218	0.6650	23.048
7	0.0258	4.141	60	0.2185	8.629	113	0.3885	13.211	166	0.5288	18.198	219	0.6702	23.140
8	0.0322	4.240	61	0.2193	8.538	114	0.3890	13.395	167	0.5302	18.293	220	0.6717	23.230
9	0.0342	4.437	62	0.2258	8.723	115	0.3918	13.487	168	0.5321	18.389	221	0.6748	23.322
10	0.0342	4.347	63	0.2276	8.807	116	0.3970	13.581	169	0.5334	18.485	222	0.6790	23.413
11	0.0357	4.479	64	0.2295	8.889	117	0.4020	13.672	170	0.5409	18.577	223	0.6822	23.504
12	0.0448	4.507	65	0.2329	8.979	118	0.4058	13.858	171	0.5439	18.667	224	0.6857	23.692
13	0.0451	4.573	66	0.2348	9.068	119	0.4071	13.767	172	0.5448	18.760	225	0.6857	23.596
14	0.0487	4.669	67	0.2439	9.155	120	0.4086	13.948	173	0.5474	18.857	226	0.6890	23.781
15	0.0513	4.739	68	0.2448	9.247	121	0.4094	14.042	174	0.5525	18.947	227	0.6916	23.968
16	0.0573	4.792	69	0.2489	9.340	122	0.4111	14.135	175	0.5557	19.039	228	0.6927	23.871
17	0.0586	4.878	70	0.2495	9.429	123	0.4151	14.223	176	0.5561	19.134	229	0.6932	24.062
18	0.0638	4.968	71	0.2551	9.515	124	0.4189	14.315	177	0.5577	19.227	230	0.6961	24.156
19	0.0640	5.057	72	0.2569	9.601	125	0.4211	14.408	178	0.5600	19.321	231	0.6990	24.246
20	0.0658	5.134	73	0.2588	9.694	126	0.4271	14.500	179	0.5644	19.411	232	0.7021	24.334
21	0.0722	5.218	74	0.2663	9.780	127	0.4283	14.598	180	0.5651	19.507	233	0.7071	24.426
22	0.0782	5.306	75	0.2665	9.872	128	0.4305	14.689	181	0.5660	19.604	234	0.7094	24.518
23	0.0790	5.391	76	0.2697	9.964	129	0.4354	14.782	182	0.5711	19.883	235	0.7132	24.614
24	0.0844	5.472	77	0.2774	10.051	130	0.4410	14.876	183	0.5712	19.698	236	0.7159	24.708
25	0.0913	5.557	78	0.2791	10.139	131	0.4411	15.056	184	0.5717	19.791	237	0.7216	24.798
26	0.0946	5.637	79	0.2827	10.226	132	0.4413	14.963	185	0.5772	19.974	238	0.7231	24.890
27	0.0965	5.717	80	0.2859	10.317	133	0.4465	15.155	186	0.5777	20.071	239	0.7268	24.983
28	0.0991	5.798	81	0.2919	10.405	134	0.4470	15.248	187	0.5857	20.164	240	0.7288	25.074
29	0.1009	5.875	82	0.2953	10.495	135	0.4481	15.333	188	0.5866	20.253	241	0.7314	25.166
30	0.1055	5.956	83	0.2953	10.588	136	0.4514	15.426	189	0.5896	20.348	242	0.7333	25.258
31	0.1075	6.040	84	0.3033	10.677	137	0.4538	15.525	190	0.5943	20.443	243	0.7347	25.438
32	0.1110	6.128	85	0.3062	10.767	138	0.4570	15.706	191	0.5965	20.538	244	0.7356	25.348
33	0.1145	6.213	86	0.3084	10.858	139	0.4590	15.618	192	0.6025	20.634	245	0.7385	25.623
34	0.1197	6.296	87	0.3091	10.949	140	0.4595	15.796	193	0.6057	20.727	246	0.7386	25.532
35	0.1211	6.384	88	0.3116	11.038	141	0.4629	15.889	194	0.6067	20.822	247	0.7426	25.714
36	0.1219	6.468	89	0.3149	11.125	142	0.4680	15.981	195	0.6076	20.913	248	0.7458	25.808
37	0.1259	6.553	90	0.3171	11.214	143	0.4693	16.074	196	0.6084	21.003	249	0.7486	25.897
38	0.1325	6.638	91	0.3211	11.306	144	0.4736	16.169	197	0.6089	21.098	250	0.7518	25.989
39	0.1354	6.719	92	0.3254	11.401	145	0.4791	16.265	198	0.6140	21.189	251	0.7543	26.081
40	0.1442	6.803	93	0.3320	11.487	146	0.4824	16.352	199	0.6151	21.281	252	0.7593	26.168
41	0.1444	6.892	94	0.3334	11.569	147	0.4826	16.627	200	0.6181	21.376	253	0.7634	26.258
42	0.1490	6.978	95	0.3373	11.659	148	0.4838	16.532	201	0.6227	21.472	254	0.7670	26.350
43	0.1521	7.063	96	0.3389	11.751	149	0.4847	16.721	202	0.6243	21.569	255	0.7680	26.442
44	0.1566	7.243	97	0.3434	11.845	150	0.4849	16.441	203	0.6287	21.662	256	0.7721	26.537
45	0.1568	7.153	98	0.3455	11.936	151	0.4854	16.816	204	0.6303	21.753	257	0.7756	26.626
46	0.1600	7.327	99	0.3476	12.024	152	0.4890	16.911	205	0.6317	21.842	258	0.7770	26.717
47	0.1635	7.409	100	0.3509	12.116	153	0.4951	16.999	206	0.6361	21.934	259	0.7793	26.805
48	0.1744	7.498	101	0.3562	12.212	154	0.4996	17.092	207	0.6412	22.027	260	0.7813	26.896
49	0.1758	7.587	102	0.3598	12.300	155	0.5012	17.188	208	0.6450	22.121	261	0.7871	26.990
50	0.1797	7.673	103	0.3624	12.482	156	0.5067	17.283	209	0.6463	22.311	262	0.7874	27.084
51	0.1826	7.758	104	0.3630	12.389	157	0.5080	17.370	210	0.6471	22.215	263	0.7875	27.172
52	0.1835	7.845	105	0.3647	12.576	158	0.5081	17.457	211	0.6491	22.400	264	0.7911	27.351
53	0.1881	7.930	106	0.3693	12.664	159	0.5116	17.555	212	0.6501	22.587	265	0.7924	27.259

Highlighted rows: Yield (0.2 % offset)

UTS (peak stress)

Fracture

Y = yield U = UTS F = fracture

FULL RESOLUTION — S26 · VC3 (2/4)

All 903 recorded samples in acquisition order · UTS 47.17 MPa · σ_y 46.47 MPa · E 2.508 GPa · ϵ_f 5.74 % · 1682 frames captured

#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)
266	0.7964	27.445	319	0.9505	32.132	372	1.1234	36.625	425	1.3269	40.570	478	1.5653	43.970
267	0.7987	27.539	320	0.9528	32.213	373	1.1247	36.545	426	1.3314	40.644	479	1.5739	44.019
268	0.8049	27.622	321	0.9572	32.296	374	1.1296	36.704	427	1.3339	40.717	480	1.5792	44.071
269	0.8073	27.711	322	0.9586	32.385	375	1.1319	36.781	428	1.3368	40.790	481	1.5843	44.130
270	0.8117	27.891	323	0.9644	32.475	376	1.1349	36.863	429	1.3394	40.859	482	1.5914	44.188
271	0.8117	27.803	324	0.9683	32.563	377	1.1428	36.946	430	1.3440	40.923	483	1.5970	44.239
272	0.8166	27.981	325	0.9708	32.648	378	1.1446	37.022	431	1.3495	41.064	484	1.6020	44.289
273	0.8166	28.070	326	0.9719	32.733	379	1.1463	37.103	432	1.3502	40.995	485	1.6104	44.342
274	0.8185	28.157	327	0.9735	32.818	380	1.1492	37.182	433	1.3543	41.131	486	1.6153	44.399
275	0.8222	28.250	328	0.9758	32.903	381	1.1530	37.262	434	1.3580	41.206	487	1.6187	44.453
276	0.8244	28.344	329	0.9833	32.988	382	1.1590	37.345	435	1.3624	41.278	488	1.6238	44.502
277	0.8297	28.433	330	0.9853	33.071	383	1.1630	37.505	436	1.3685	41.408	489	1.6261	44.553
278	0.8312	28.521	331	0.9888	33.158	384	1.1647	37.425	437	1.3699	41.341	490	1.6320	44.605
279	0.8360	28.608	332	0.9900	33.243	385	1.1655	37.580	438	1.3811	41.478	491	1.6405	44.651
280	0.8368	28.698	333	0.9931	33.326	386	1.1713	37.650	439	1.3835	41.549	492	1.6422	44.703
281	0.8391	28.788	334	0.9951	33.410	387	1.1787	37.731	440	1.3877	41.618	493	1.6464	44.756
282	0.8396	28.876	335	1.0017	33.496	388	1.1814	37.809	441	1.3911	41.681	494	1.6552	44.805
283	0.8423	28.966	336	1.0039	33.660	389	1.1841	37.884	442	1.3976	41.745	495	1.6616	44.856
284	0.8477	29.051	337	1.0053	33.579	390	1.1912	37.962	443	1.4024	41.812	496	1.6670	44.902
285	0.8501	29.145	338	1.0088	33.748	391	1.1943	38.046	444	1.4063	41.881	497	1.6718	44.948
286	0.8547	29.230	339	1.0141	33.918	392	1.2001	38.124	445	1.4120	41.944	498	1.6788	44.994
287	0.8585	29.321	340	1.0160	33.831	393	1.2018	38.198	446	1.4149	42.012	499	1.6849	45.086
288	0.8603	29.417	341	1.0187	34.003	394	1.2044	38.272	447	1.4221	42.083	500	1.6850	45.043
289	0.8626	29.501	342	1.0194	34.089	395	1.2069	38.348	448	1.4245	42.148	501	1.6968	45.132
290	0.8651	29.593	343	1.0239	34.172	396	1.2099	38.427	449	1.4289	42.208	502	1.6998	45.180
291	0.8662	29.764	344	1.0257	34.257	397	1.2149	38.504	450	1.4361	42.273	503	1.7078	45.227
292	0.8682	29.680	345	1.0302	34.339	398	1.2169	38.582	451	1.4387	42.339	504	1.7154	45.266
293	0.8722	29.854	346	1.0330	34.422	399	1.2187	38.657	452	1.4429	42.405	505	1.7207	45.315
294	0.8768	29.946	347	1.0374	34.504	400	1.2251	38.735	453	1.4501	42.477	506	1.7272	45.360
295	0.8803	30.034	348	1.0428	34.586	401	1.2295	38.807	454	1.4525	42.540	507	1.7272	45.403
296	0.8811	30.120	349	1.0496	34.754	402	1.2318	38.881	455	1.4551	42.602	508	1.7390	45.448
297	0.8853	30.295	350	1.0502	34.669	403	1.2367	38.960	456	1.4583	42.667	509	1.7409	45.489
298	0.8858	30.205	351	1.0560	34.923	404	1.2416	39.038	457	1.4641	42.727	510	1.7499	45.527
299	0.8892	30.385	352	1.0589	34.841	405	1.2473	39.109	458	1.4692	42.788	511	1.7534	45.568
300	0.8921	30.471	353	1.0626	35.004	406	1.2514	39.182	459	1.4732	42.852	512	1.7578	45.607
301	0.8963	30.640	354	1.0626	35.087	407	1.2550	39.260	460	1.4787	42.911	513	1.7632	45.650
302	0.8987	30.556	355	1.0654	35.169	408	1.2598	39.338	461	1.4870	42.975	514	1.7726	45.688
303	0.9035	30.730	356	1.0680	35.253	409	1.2601	39.409	462	1.4879	43.038	515	1.7776	45.729
304	0.9056	30.819	357	1.0727	35.414	410	1.2633	39.483	463	1.4961	43.097	516	1.7837	45.768
305	0.9100	30.911	358	1.0745	35.333	411	1.2694	39.557	464	1.5023	43.155	517	1.7897	45.806
306	0.9114	30.998	359	1.0769	35.495	412	1.2720	39.629	465	1.5093	43.218	518	1.7949	45.848
307	0.9128	31.085	360	1.0797	35.576	413	1.2745	39.704	466	1.5165	43.276	519	1.8002	45.885
308	0.9164	31.173	361	1.0838	35.660	414	1.2780	39.780	467	1.5174	43.337	520	1.8060	45.920
309	0.9183	31.258	362	1.0881	35.743	415	1.2835	39.925	468	1.5240	43.401	521	1.8096	45.953
310	0.9230	31.343	363	1.0940	35.819	416	1.2844	39.852	469	1.5304	43.457	522	1.8183	45.994
311	0.9254	31.436	364	1.1005	35.901	417	1.2886	39.997	470	1.5333	43.512	523	1.8253	46.032
312	0.9267	31.524	365	1.1049	35.982	418	1.2956	40.072	471	1.5355	43.573	524	1.8346	46.065
313	0.9306	31.612	366	1.1054	36.137	419	1.2980	40.143	472	1.5392	43.624	525	1.8393	46.101
314	0.9364	31.698	367	1.1071	36.060	420	1.3050	40.212	473	1.5404	43.683	526	1.8507	46.136
315	0.9372	31.784	368	1.1115	36.222	421	1.3099	40.285	474	1.5468	43.744	527	1.8526	46.168
316	0.9433	31.958	369	1.1130	36.303	422	1.3127	40.358	475	1.5540	43.800	528	1.8573	46.202
317	0.9433	31.867	370	1.1150	36.386	423	1.3169	40.428	476	1.5562	43.857	529	1.8679	46.236
318	0.9476	32.047	371	1.1212	36.466	424	1.3228	40.501	477	1.5644	43.915	530	1.8738	46.267

Highlighted rows:  Yield (0.2 % offset)

 UTS (peak stress)

 Fracture

Y = yield U = UTS F = fracture

FULL RESOLUTION — S26 · VC3 (3/4)

All 903 recorded samples in acquisition order · UTS 47.17 MPa · σ_y 46.47 MPa · E 2.508 GPa · ϵ_f 5.74 % · 1682 frames captured

#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)
531	1.8823	46.298	584	2.2954	47.131	637	2.8134	46.954	690	3.3849	46.201	743	3.9626	45.365
532	1.8887	46.327	585	2.3047	47.134	638	2.8205	46.944	691	3.3923	46.185	744	3.9761	45.347
533	1.8976	46.357	586	2.3188	47.144	639	2.8357	46.933	692	3.4070	46.165	745	3.9941	45.332
534	1.8992	46.389	587	2.3284	47.148	640	2.8357	46.921	693	3.4170	46.151	746	3.9994	45.317
535	1.9096	46.417	588	2.3326	47.148	641	2.8508	46.910	694	3.4316	46.135	747	4.0067	45.299
536	1.9158	46.448	589	2.3389	47.149	642	2.8629	46.901	695	3.4416	46.121	748	4.0265	45.281
537	1.9244	46.472	590	2.3475	47.150	643	2.8739	46.888	696	3.4484	46.105	749	4.0327	45.269
538	1.9294	46.499	591	2.3624	47.152	644	2.8857	46.874	697	3.4606	46.088	750	4.0429	45.252
539	1.9407	46.524	592	2.3672	47.151	645	2.8921	46.863	698	3.4654	46.070	751	4.0572	45.236
540	1.9462	46.551	593	2.3823	47.155	646	2.9044	46.853	699	3.4885	46.054	752	4.0679	45.222
541	1.9524	46.582	594	2.3823	47.157	647	2.9178	46.842	700	3.4965	46.039	753	4.0748	45.208
542	1.9622	46.609	595	2.3921	47.156	648	2.9357	46.829	701	3.5019	46.026	754	4.0873	45.188
543	1.9669	46.629	596	2.4018	47.157	649	2.9357	46.817	702	3.5080	46.008	755	4.0986	45.170
544	1.9696	46.648	597	2.4091	47.161	650	2.9445	46.804	703	3.5242	45.993	756	4.1148	45.159
545	1.9802	46.675	598	2.4197	47.160	651	2.9580	46.788	704	3.5371	45.974	757	4.1198	45.145
546	1.9863	46.702	599	2.4278	47.163	652	2.9715	46.776	705	3.5503	45.953	758	4.1336	45.132
547	1.9977	46.723	600	2.4405	47.165	653	2.9830	46.760	706	3.5586	45.942	759	4.1450	45.117
548	2.0021	46.739	601	2.4460	47.163	654	2.9892	46.749	707	3.5675	45.930	760	4.1587	45.097
549	2.0088	46.759	602	2.4591	47.159	655	3.0018	46.736	708	3.5802	45.912	761	4.1701	45.083
550	2.0128	46.781	603	2.4662	47.157	656	3.0084	46.721	709	3.5959	45.896	762	4.1772	45.069
551	2.0225	46.802	604	2.4817	47.154	657	3.0213	46.706	710	3.6067	45.881	763	4.1902	45.052
552	2.0321	46.820	605	2.4921	47.150	658	3.0347	46.695	711	3.6187	45.861	764	4.1941	45.033
553	2.0336	46.839	606	2.5011	47.149	659	3.0427	46.682	712	3.6244	45.845	765	4.2130	45.018
554	2.0439	46.856	607	2.5117	47.143	660	3.0550	46.662	713	3.6346	45.831	766	4.2184	45.005
555	2.0556	46.866	608	2.5221	47.143	661	3.0616	46.651	714	3.6513	45.814	767	4.2300	44.997
556	2.0650	46.880	609	2.5273	47.141	662	3.0811	46.637	715	3.6648	45.797	768	4.2423	44.983
557	2.0695	46.901	610	2.5398	47.141	663	3.0861	46.625	716	3.6730	45.784	769	4.2505	44.967
558	2.0752	46.920	611	2.5473	47.135	664	3.1003	46.613	717	3.6798	45.768	770	4.2678	44.955
559	2.0863	46.933	612	2.5574	47.131	665	3.1051	46.595	718	3.6881	45.750	771	4.2737	44.942
560	2.0952	46.942	613	2.5642	47.127	666	3.1195	46.580	719	3.6994	45.736	772	4.2803	44.929
561	2.1056	46.960	614	2.5748	47.120	667	3.1318	46.566	720	3.7215	45.719	773	4.2917	44.910
562	2.1105	46.975	615	2.5805	47.118	668	3.1439	46.551	721	3.7269	45.702	774	4.3069	44.896
563	2.1222	46.987	616	2.5962	47.116	669	3.1579	46.532	722	3.7371	45.688	775	4.3267	44.880
564	2.1288	47.000	617	2.6037	47.104	670	3.1620	46.516	723	3.7421	45.672	776	4.3305	44.870
565	2.1386	47.012	618	2.6120	47.097	671	3.1750	46.500	724	3.7583	45.654	777	4.3381	44.857
566	2.1483	47.022	619	2.6209	47.091	672	3.1892	46.483	725	3.7715	45.640	778	4.3481	44.842
567	2.1510	47.027	620	2.6337	47.084	673	3.1991	46.472	726	3.7842	45.626	779	4.3634	44.826
568	2.1647	47.033	621	2.6435	47.079	674	3.2138	46.459	727	3.7886	45.608	780	4.3746	44.814
569	2.1686	47.050	622	2.6545	47.073	675	3.2224	46.443	728	3.8016	45.595	781	4.3857	44.803
570	2.1820	47.062	623	2.6642	47.068	676	3.2320	46.425	729	3.8122	45.579	782	4.3988	44.790
571	2.1912	47.064	624	2.6709	47.062	677	3.2381	46.409	730	3.8247	45.564	783	4.4058	44.775
572	2.1956	47.073	625	2.6853	47.052	678	3.2565	46.394	731	3.8416	45.549	784	4.4190	44.760
573	2.2010	47.082	626	2.6960	47.043	679	3.2622	46.377	732	3.8476	45.534	785	4.4295	44.750
574	2.2102	47.090	627	2.7063	47.042	680	3.2757	46.360	733	3.8563	45.518	786	4.4424	44.732
575	2.2185	47.101	628	2.7166	47.035	681	3.2806	46.345	734	3.8701	45.500	787	4.4482	44.722
576	2.2288	47.104	629	2.7228	47.025	682	3.2933	46.328	735	3.8826	45.485	788	4.4584	44.709
577	2.2386	47.105	630	2.7359	47.018	683	3.3131	46.311	736	3.8959	45.465	789	4.4720	44.694
578	2.2470	47.112	631	2.7437	47.011	684	3.3148	46.296	737	3.9046	45.452	790	4.4851	44.678
579	2.2567	47.117	632	2.7578	47.000	685	3.3349	46.265	738	3.9120	45.439	791	4.4942	44.665
580	2.2661	47.123	633	2.7735	46.991	686	3.3349	46.280	739	3.9181	45.424	792	4.5036	44.652
581	2.2732	47.125	634	2.7758	46.984	687	3.3491	46.249	740	3.9320	45.411	793	4.5102	44.636
582	2.2818	47.125	635	2.7845	46.972	688	3.3608	46.233	741	3.9465	45.394	794	4.5240	44.621
583	2.2842	47.128	636	2.7968	46.961	689	3.3733	46.218	742	3.9591	45.379	795	4.5406	44.603

Highlighted rows: Yield (0.2 % offset) UTS (peak stress) Fracture Y = yield U = UTS F = fracture

FULL RESOLUTION — S26 · VC3 (4/4)

All 903 recorded samples in acquisition order · UTS 47.17 MPa · σ_y 46.47 MPa · E 2.508 GPa · ϵ_f 5.74 % · 1682 frames captured

#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)	#	ϵ (%)	σ (MPa)
796	4.5486	44.586	849	5.1188	43.867	902	5.7359	42.622
797	4.5546	44.580	850	5.1315	43.850	903	5.7446	42.537 F
798	4.5742	44.565	851	5.1453	43.836			
799	4.5781	44.553	852	5.1614	43.823			
800	4.5982	44.542	853	5.1739	43.805			
801	4.6036	44.532	854	5.1763	43.788			
802	4.6103	44.514	855	5.1833	43.778			
803	4.6270	44.497	856	5.2011	43.764			
804	4.6340	44.487	857	5.2126	43.747			
805	4.6469	44.477	858	5.2210	43.729			
806	4.6624	44.461	859	5.2289	43.720			
807	4.6678	44.451	860	5.2400	43.701			
808	4.6806	44.435	861	5.2510	43.686			
809	4.6913	44.421	862	5.2644	43.671			
810	4.7050	44.409	863	5.2670	43.651			
811	4.7093	44.397	864	5.2831	43.633			
812	4.7308	44.382	865	5.3015	43.622			
813	4.7347	44.365	866	5.3096	43.606			
814	4.7393	44.352	867	5.3257	43.591			
815	4.7592	44.338	868	5.3323	43.570			
816	4.7666	44.325	869	5.3381	43.552			
817	4.7759	44.315	870	5.3520	43.541			
818	4.7838	44.302	871	5.3634	43.525			
819	4.8023	44.286	872	5.3747	43.503			
820	4.8087	44.271	873	5.3863	43.489			
821	4.8265	44.262	874	5.3938	43.476			
822	4.8337	44.248	875	5.4069	43.457			
823	4.8401	44.230	876	5.4125	43.435			
824	4.8546	44.216	877	5.4347	43.410			
825	4.8661	44.203	878	5.4386	43.393			
826	4.8794	44.189	879	5.4594	43.357			
827	4.8919	44.176	880	5.4594	43.376			
828	4.9051	44.163	881	5.4701	43.335			
829	4.9090	44.152	882	5.4839	43.316			
830	4.9219	44.137	883	5.4940	43.299			
831	4.9344	44.123	884	5.5007	43.278			
832	4.9469	44.110	885	5.5194	43.255			
833	4.9531	44.092	886	5.5288	43.225			
834	4.9589	44.079	887	5.5421	43.203			
835	4.9732	44.064	888	5.5606	43.179			
836	4.9790	44.054	889	5.5676	43.153			
837	4.9888	44.038	890	5.5762	43.128			
838	4.9999	44.022	891	5.5918	43.102			
839	5.0035	44.006	892	5.6051	43.073			
840	5.0165	43.995	893	5.6170	43.042			
841	5.0295	43.982	894	5.6323	43.008			
842	5.0443	43.969	895	5.6374	42.972			
843	5.0500	43.959	896	5.6510	42.931			
844	5.0641	43.941	897	5.6585	42.888			
845	5.0748	43.926	898	5.6713	42.843			
846	5.0850	43.908	899	5.6868	42.799			
847	5.0995	43.895	900	5.7033	42.749			
848	5.1071	43.884	901	5.7118	42.692			

Highlighted rows:  Yield (0.2 % offset)

 UTS (peak stress)

 Fracture

Y = yield U = UTS F = fracture