

### SARS-CoV-2 IgG (COV2G)

For Use Under Emergency Use Authorization Only.

For *in vitro* diagnostic use.

For Prescription Use Only.

The results of this semiquantitative test should not be interpreted as an indication or degree of immunity or protection from infection.

Current Revision and Date <sup>a</sup>	Rev. 01, 2021-01	
Product Name	Dimension Vista SARS-CoV-2 IgG (COV2G)	<b>REF</b> K7771 11417771
Abbreviated Product Name	Dimension Vista COV2G	
Test Name/ID	COV2G	
Systems	Dimension Vista System using software version 3.9.3 or higher	
Materials Required but Not Provided	SARS-CoV-2 IgG calibrator (COV2G CAL/CV2G CAL)	<b>REF</b> KC872 11417772
	SARS-CoV-2 IgG Quality Control (COV2G QC/CV2G QC)	<b>REF</b> KC873 11417773
	Dimension Vista LOCI Reaction Vessels	<b>REF</b> KS855 10445212
Specimen Types	Serum, dipotassium EDTA plasma, lithium heparin plasma	
Sample Volume	2 µL	
Measuring Interval	600–140,000 QUAL Units	

<sup>a</sup> A vertical bar in the page margin indicates technical content that differs from the previous version.

#### For USA

- Use of this test is limited to laboratories certified under Clinical Laboratory Improvement Amendments of 1988 (CLIA) that meet requirements to perform moderate- or high-complexity testing.
- Negative results do not preclude acute SARS-CoV-2 infection. If acute infection is suspected, direct testing for SARS-CoV-2 is necessary.
- Results from antibody testing should not be used to diagnose or exclude acute SARS-CoV-2 infection.
- Positive results may be due to past or present infection with non-SARS-CoV-2 coronavirus strains, such as coronavirus HKU1, NL63, OC43, or 229E.

## Intended Use

The Dimension Vista® SARS-CoV-2 IgG (COV2G) assay is a chemiluminescent immunoassay intended for qualitative and semiquantitative detection of IgG antibodies to SARS-CoV-2 in human serum and plasma (dipotassium EDTA and lithium heparin) using the Dimension® Vista® System. The Dimension Vista® SARS-CoV-2 IgG (COV2G) assay is intended for use as an aid in identifying individuals with an adaptive immune response to SARS-CoV-2, indicating recent or prior infection. At this time, it is unknown for how long antibodies persist following infection and if the presence of antibodies confers protective immunity. The Dimension Vista® SARS-CoV-2 IgG (COV2G) assay should not be used to diagnose acute SARS-CoV-2 infection.

Testing is limited to laboratories certified under the Clinical Laboratory Improvement Amendments (CLIA) of 1988, 42 U.S.C 263a, that meet requirements to perform moderate or high complexity tests. Results are for the detection of SARS-CoV-2 IgG antibodies. IgG antibodies to SARS-CoV-2 are generally detectable in blood several days after initial infection, although the duration of time antibodies are present post-infection is not well characterized. Individuals may have detectable virus present for several weeks following seroconversion. Laboratories within the United States and its territories are required to report all results to the appropriate public health authorities.

The sensitivity of the Dimension Vista® SARS-CoV-2 IgG (COV2G) assay early after infection is unknown. Negative results do not preclude acute SARS-CoV-2 infection. If acute infection is suspected, direct testing for SARS-CoV-2 is necessary.

False positive results for the Dimension Vista® SARS-CoV-2 IgG (COV2G) assay may occur due to cross-reactivity from pre-existing antibodies or other possible causes.

The Dimension Vista® SARS-CoV-2 IgG (COV2G) assay is only for use under the Food and Drug Administration's Emergency Use Authorization.

## Summary and Explanation

COVID-19 (coronavirus disease 2019) is the illness resulting from infection with SARS-CoV-2 (severe acute respiratory syndrome coronavirus 2) virus.<sup>1-5</sup> The virus spreads readily from person-to-person or possibly from environmental exposure.<sup>6</sup> Evidence supports spread by both asymptomatic and symptomatic individuals.<sup>7</sup> Antibodies appear approximately 1–3 weeks post-symptom onset in most patients and are produced in both symptomatic and asymptomatic infection.<sup>8,9</sup> Unlike typical seroconversion profiles, near-simultaneous production of both IgM and IgG has been observed in symptomatic patients with confirmed SARS-CoV-2. Titer of antibody may be higher in symptomatic disease, though additional data is needed to confirm this.<sup>10,11</sup> Antibodies produced to structural proteins of the virus include spike antibody and nucleocapsid antibody. Data show both IgM and IgG antibodies for these structural proteins appear with seroconversion. IgM eventually disappears, but IgG remains detectable in most patients. Spike is a transmembrane glycoprotein comprised of two regions: S1 and S2. S1 mediates recognition and binding of the Angiotensin-converting enzyme 2 receptor (ACE2) on host cells, and S2 facilitates viral fusion and entry.<sup>12,13</sup> The majority of S1 is comprised of the receptor binding domain (RBD) that binds directly to ACE2 and is highly immunogenic. The S1 RBD in SARS-CoV-2 contains both unique and conserved sequences compared to other beta-coronaviruses. Multiple vaccines in development target or include the S1 RBD, as initial data indicate antibodies to this region can be neutralizing.<sup>14-23</sup> The ability to identify specific antibodies associated with neutralization will be an important adjunct to the detection of an immune response to the SARS-CoV-2 virus.

## Principles of the Procedure

The Dimension Vista COV2G assay is a homogeneous, sandwich chemiluminescent immunoassay based on Luminescent Oxygen Channeling Immunoassay (LOCI) technology. The LOCI reagents include two synthetic bead reagents and a biotinylated anti-Human IgG antibody. The first bead reagent (Sensibeads) is coated with streptavidin and contains photosensitizer dye. The second bead reagent (Chemibeads) is coated with anti-FITC (Fluorescein isothiocyanate) antibody and contains chemiluminescent dye. For this assay the anti-FITC antibody coated-Chemibeads are predecorated with fluoresceinated S1 RBD antigen. The patient sample is incubated with Chemibeads. After 1 minute the biotinylated antibody is added to form bead-SARS-CoV-2 antigen-biotinylated antibody sandwiches. After incubation, Sensibeads are added and bind to the biotin to form bead-pair immunocomplexes. Illumination of the complex at 680 nm generates singlet oxygen from Sensibeads which diffuses into the Chemibeads, triggering a chemiluminescent reaction. The resulting signal is measured at 612 nm and is a direct function of the amount of SARS-CoV-2 IgG antibody present in the sample.<sup>24-26</sup>

## Reagents

Material Description	Storage	Stability <sup>a</sup>
<b>Dimension Vista COV2G</b>	Unopened at 2–8°C	Until expiration date on product
Wells 1–2 <sup>c</sup>		
Liquid Chemibeads (200 µg/mL, recombinant S1 RBD antigen); bovine serum albumin; mouse IgG	Onboard <sup>b</sup> Open well	30 days 3 days
Wells 3–4		
Liquid biotinylated antibody (40 µg/mL, monoclonal mouse anti-human IgG); mouse IgG; bovine serum albumin; bovine gamma globulin; goat serum <sup>d</sup>		
Wells 5–6		
Empty		
Wells 7–8		
Liquid Sensibeads (1500 µg/mL); bovine serum albumin; human gamma globulin		
Wells 9–12		
Assay buffer; bovine serum albumin		

<sup>a</sup> Refer to *Storage and Stability*.

<sup>b</sup> Refer to *Onboard Stability*.

<sup>c</sup> Wells are numbered consecutively from the wide end of the cartridge.

<sup>d</sup> Goat serum contains sodium azide (0.00062%).

## Warnings and Precautions

For *in vitro* diagnostic use.

For Prescription Use Only.

For Use Under Emergency Use Authorization Only.

This test has not been FDA cleared or approved, but has been authorized for emergency use by FDA under an Emergency Use Authorization (EUA) for use by laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C 263a, that meet requirements to perform moderate or high complexity tests.

This test has been authorized only for detecting the presence of IgG antibodies against SARS-CoV-2, not for any other viruses or pathogens.

The emergency use of this test is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of in vitro diagnostic tests for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug, and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.

Safety data sheets (SDS) available on [siemens-healthineers.com](https://www.siemens-healthineers.com).



**H317, H412**  
**P280, P273,**  
**P302+P352,**  
**P333+P313,**  
**P362+P364, P501**

#### **Warning!**

May cause an allergic skin reaction. Harmful to aquatic life with long lasting effects.

Wear protective gloves/protective clothing/eye protection/face protection. Avoid release to the environment. IF ON SKIN: Wash with plenty of soap and water. If skin irritation or rash occurs: Get medical advice/attention. Take off contaminated clothing and wash it before reuse. Dispose of contents and container in accordance with all local, regional, and national regulations.

**Contains:** reaction mass of 5-chloro-2-methyl-2H-isothiazol-3-one and 2-methyl-2H-isothiazol-3-one (3:1); (Dimension Vista COV2G)



#### **Warning! Potential Biohazard**

Contains human source material.

**Caution:** All products containing human source material should be treated as potentially infectious. Source material from human blood from which this product was derived was found negative when tested in accordance with current FDA required tests described in 21 CFR 610.40(a) and (b). No known test methods can offer assurance that products derived from human sources will not transmit infectious agents; this material should be handled using good laboratory practices and universal precautions.

#### **CAUTION**

This device contains material of animal origin and should be handled as a potential carrier and transmitter of disease.

Contains sodium azide as a preservative. Sodium azide can react with copper or lead plumbing to form explosive metal azides. On disposal, flush reagents with a large volume of water to prevent buildup of azides. Disposal into drain systems must be in compliance with prevailing regulatory requirements.

Dispose of hazardous or biologically contaminated materials according to the practices of your institution. Discard all materials in a safe and acceptable manner and in compliance with prevailing regulatory requirements.

## **Storage and Stability**

Store unopened product at 2–8°C. Do not use products beyond the expiration date printed on the product labeling.

For details about product onboard stability, refer to *Reagents*.

## **Onboard Stability**

Discard products at the end of the onboard stability interval.

For details about product onboard stability, refer to *Reagents*.

Do not use products beyond the expiration date printed on the product labeling.

## Specimen Collection and Handling

Serum and plasma (dipotassium EDTA and lithium heparin) are the recommended specimen types for this assay. Do not use heat-inactivated specimens.

The handling and storage information provided here is based on data or references maintained by the manufacturer. It is the responsibility of the individual laboratory to use all available references and/or its own studies when establishing alternate stability criteria to meet specific needs.

### Collecting the Specimen

- Observe universal precautions when collecting specimens. Handle all specimens as if they are capable of transmitting disease.<sup>27</sup>
- Follow recommended procedures for collection of diagnostic blood specimens by venipuncture.<sup>28</sup>
- Follow the instructions provided with your specimen collection device for use and processing.<sup>29</sup>
- Specimens with high turbidity or particulates should be centrifuged before analysis.
- Allow blood specimens to clot completely before centrifugation.<sup>30</sup>
- Keep tubes capped at all times.<sup>30</sup>

### Storing the Specimen

Test specimens as soon as possible after collecting. Store specimens at 2–8°C if not tested immediately within 16 hours.<sup>31</sup>

Separated specimens in the primary collection device are stable for up to 7 days at 2–8°C.<sup>31</sup>

Separated specimens may be frozen for up to 1 month at  $\leq -20^{\circ}\text{C}$ . The samples may be frozen once. Do not store in a frost-free freezer. Thoroughly mix thawed specimens and centrifuge before using.

### Transporting the Specimen

Package and label specimens for shipment in compliance with applicable federal and international regulations covering the transport of clinical specimens and etiological agents.

Store samples capped and upright at 2–8°C upon arrival. If shipment is expected to exceed 7 days, ship specimens frozen.

### Preparing the Samples

This assay requires 2  $\mu\text{L}$  of serum or plasma for a single determination. This volume does not include the unusable volume in the sample container or the additional volume required when performing duplicates or other tests on the same sample. For information about determining the minimum required volume, refer to the Dimension Vista Operator's Guide and Online Help.

Do not use samples with apparent contamination.

Remove particulates by centrifugation according to CLSI guidance and the collection device manufacturer's recommendations.<sup>30</sup>

For a complete list of appropriate sample containers, refer to the Dimension Vista Operator's Guide and Online Help.

Before placing samples on the system, ensure that samples are free of:

- Bubbles or foam.
- Fibrin or other particulate matter.

## Procedure

### Materials Provided

The following materials are provided:

REF	Contents	Number of Tests
K7771 11417771	Dimension Vista COV2G Flex Reagent Cartridge	4 x 60

### Materials Required but Not Provided

The following materials are required to perform this assay, but are not provided:

REF	Description	
	Dimension Vista System	
KC872 11417772	SARS-CoV-2 IgG calibrator (COV2G CAL/CV2G CAL)	2 x 1.0 mL calibrator level 1/A 2 x 1.0 mL calibrator level 2/B 2 x 1.0 mL calibrator level 3/C 2 x 1.0 mL calibrator level 4/D 2 x 1.0 mL calibrator level 5/E
KC873 11417773	SARS-CoV-2 IgG Quality Control (COV2G QC/CV2G QC)	6 x 1.0 mL positive quality control 6 x 1.0 mL negative quality control
KS855 10445212	Dimension Vista LOCI Reaction Vessels	

### Assay Procedure

For information about loading samples and ordering tests, refer to the Dimension Vista Operator's Guide and Online Help. For information about running in semiquantitative mode, refer to the refer to the Dimension Vista Operator's Guide and Online Help.

The system automatically performs the following steps:

1. For serum or plasma, delivers 20 µL of antigen-coated Chemibeads into the reaction vessel.
2. Adds 2 µL of sample into the reaction vessel.
3. Dispenses 20 µL of the biotinylated antibody reagent into the reaction vessel.
4. Briefly mixes and incubates for 12 minutes at 37°C.
5. Dispenses 20 µL of the Sensibeads into the reaction vessel.
6. Incubates the mixture at 37°C.
7. Dispenses 100 µL of the Assay Buffer into the reaction vessel.
8. Illuminates the complex at 680 nm. This generates singlet oxygen from the Sensibeads which diffuses into the Chemibeads, triggering a chemiluminescent reaction.
9. Measures emission at 612 nm and is a direct function of the total SARS-CoV-2 antibody concentration in the sample.
10. Reports "POS", "NEG" results (qualitative) and QUAL units (semiquantitative) for "POS" results within the acceptable measuring interval.

Test Duration: approximately 16 minutes

## Preparing the Reagents

All reagents are liquid and ready to use.

## Preparing the System

For information about loading reagents, refer to the Dimension Vista Operator's Guide and Online Help.

## Performing Calibration

For calibration of the Dimension Vista COV2G assay, use SARS-CoV-2 IgG calibrator (COV2G CAL/CV2G CAL). Use the calibrators in accordance with the calibrator instructions for use.

## Calibration Procedure

The Dimension Vista COV2G assay uses a five-level semiquantitative calibration that is similar to the Urine Drugs of Abuse assays on the Dimension Vista System. Refer to *Urine Drugs of Abuse and Serum Toxicology Calibration* in the Dimension Vista Operator's Guide.

Follow the steps in the *Entering Calibrator IFU Information* section described in *Calibration and Quality Control* in the Dimension Vista Operator's Guide.

Calibration may be performed by loading levels 1/A through 5/E onto the Dimension Vista System prior to calibration or by performing the calibration from cups. Refer to *Calibration and Quality Control* in the Dimension Vista Operator's Guide.

## Calibration Frequency

Calibrate the assay every 14 days.

In addition, perform a calibration:

- At the end of the calibration interval.
- When changing lot numbers of reagents.
- When indicated by quality control results.
- After major maintenance or service.

Follow government regulations or accreditation requirements for calibration frequency. Individual laboratory quality control programs and procedures may require more frequent calibration.

## Performing Quality Control

For quality control of the Dimension Vista COV2G, use the SARS-CoV-2 IgG Quality Control (COV2G QC/CV2G QC) or an equivalent product at least once during each day that samples are analyzed. Additional quality control material can be used at the discretion of the laboratory. Use the quality control material in accordance with the quality control instructions for use.

In addition, perform quality control:

- Following a valid calibration.
- With use of a new lot of reagent.
- When troubleshooting test results that do not match clinical conditions or symptoms.

Follow government regulations or accreditation requirements for quality control frequency.

Acceptable performance is achieved when the analyte values obtained are within the expected control interval for the system, as indicated by the manufacturer of the control material or within the interval determined by an internal laboratory quality control procedure.

Individual laboratory quality control programs and procedures may require more frequent quality control testing. For information about entering quality control definitions, refer to the Dimension Vista Operator's Guide and Online Help.

## Taking Corrective Action

If the quality control results do not fall within the expected control interval, do not report results. Perform corrective actions in accordance with established laboratory protocol. For suggested protocol, refer to the Dimension Vista Operator's Guide and Online Help.

## Results

### Calculation of Results

The system determines the result using the calculation scheme described in the Dimension Vista Operator's Guide and Online Help. The system reports the results as "POS" or "NEG" relative to the assay cutoff and a numeric value in QUAL units for "POS" results within the acceptable measuring interval. The Dimension Vista COV2G assay cutoff analyte value is 1000 QUAL and is used to distinguish between positive and negative. Refer to *Interpretation of Results*.

### Interpretation of Results

#### Interpretation of Qualitative Results

Analyte Value	Result	Test Result Interpretation
< 1000 QUAL	NEG	IgG antibodies to SARS-CoV-2 are not detected
≥ 1000 QUAL	POS	IgG antibodies to SARS-CoV-2 are detected

#### Interpretation of Semiquantitative Results

Analyte Value	Interpretation	Description
< 1000 QUAL	NEG	IgG antibodies to SARS-CoV-2 are not detected
≥ 1000 QUAL to 140,000 QUAL	POS; Numerical value is reported to the end user	IgG antibodies to SARS-CoV-2 are detected
> 140,000 QUAL	POS; Value above ULMI is not reported <sup>a</sup>	IgG antibodies to SARS-CoV-2 are detected

<sup>a</sup> ULMI- upper limit of measuring interval.

Results of this assay should always be interpreted in conjunction with the patient's medical history, clinical presentation, and other findings.

## Limitations

The following information pertains to limitations of the assay:

- The clinical applicability of a quantitative or semiquantitative result is currently unknown and cannot be interpreted as an indication or degree of immunity nor protection from infection, nor compared to other SARS-CoV-2 antibody assays.
- This device should not be used to diagnose or exclude acute SARS-CoV-2 infection. Direct testing for SARS-CoV-2 with a molecular assay should be performed to evaluate acute infection in symptomatic individuals.
- Performance characteristics have not been established for the assay used in conjunction with other manufacturer's assays for specific SARS-CoV-2 serological markers.



- Performance of the assay has not been established with cord blood, neonatal specimens, cadaver specimens, or body fluids other than serum or plasma.
- Performance has only been established with the specimen types listed in the Intended Use. Other specimen types have not been evaluated and should not be used with this assay.
- Results obtained with the assay may not be used interchangeably with values obtained with different manufacturer's test methods.
- Results are not intended to be used as the sole basis for patient management decisions. Test results should be interpreted in conjunction with clinical observations, patient history, epidemiological information, and other laboratory findings.
- It is currently unknown how long SARS-CoV-2 antibodies persist following infection and if the presence of antibodies confers protective immunity.
- A positive test result may not indicate previous SARS-CoV-2 infection. Consider other information, including clinical history and local disease prevalence, in assessing the need for a second, but different, serology test to confirm an immune response.
- A positive test result does not exclude past or present infection by other coronaviruses, such as SARS-CoV-1, MERS-CoV, HKU1, 229E, NL63, or OC43, or due to cross-reactivity from pre-existing antibodies or other possible causes.
- Due to the risk of false positive results, confirmation of positive results should be considered using a second, different IgG assay.
- A negative test result for an individual subject indicates absence of detectable anti-SARS-CoV-2 antibodies. Negative results do not preclude SARS-CoV-2 infection and should not be used as the sole basis for patient management decisions. A negative result can occur if the quantity of the anti-SARS-CoV-2 antibodies present in the specimen is below the detection limits of the assay, or the antibodies that are detected are not present during the stage of disease in which a sample is collected.
- A negative test result does not exclude the possibility of exposure to or infection with SARS-CoV-2. SARS-CoV-2 antibodies may not be detectable in patients with recent infections (7–10 days or less) or in samples collected from patients less than 7 days from a positive polymerase chain reaction (PCR) result. Patient specimens may be negative if collected during the early (preseroconversion) phase of illness or due to a decline in titer over time. In addition, the immune response may be depressed in elderly, immunocompromised, or immunosuppressed patients.
- The test may have lower sensitivity for IgG detection in symptomatic individuals prior to 8 days since symptom onset.
- It is not known at this time if the presence of antibodies to SARS-CoV-2 confers immunity to infection.
- This test should not be used for donor screening.

## Conditions of Authorization for the Laboratory

The Dimension Vista COV2G assay Letter of Authorization, along with the authorized Fact Sheet for Healthcare Providers, the authorized Fact Sheet for Patients, and authorized labeling are available on the FDA website:

<https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/vitro-diagnostics-euas>

Authorized laboratories using the Dimension Vista COV2G assay ("your product" in the conditions below) must adhere to the Conditions of Authorization indicated in the Letter of Authorization as listed below:

- Authorized laboratories<sup>a</sup> using your product must include with test result reports, all authorized Fact Sheets. Under exigent circumstances, other appropriate methods for disseminating these Fact Sheets may be used, which may include mass media.
- Authorized laboratories using your product must use the product as outlined in the Instructions for Use. Deviations from the authorized procedures, including the authorized instruments, authorized clinical specimen types, authorized control materials, authorized other ancillary reagents and authorized materials required to use the Dimension Vista COV2G assay are not permitted.
- Authorized laboratories that receive your product must notify the relevant public health authorities of their intent to run the assay prior to initiating testing.
- Authorized laboratories using your product must have a process in place for reporting test results to healthcare providers and relevant public health authorities, as appropriate.
- Authorized laboratories must collect information on the performance of your product and report to DMD/OHT7-OIR/OPEQ/ CDRH (via email: CDRH EUA Reporting@fda.hhs.gov) and Siemens Healthineers Technical Support (tel: +1 (800) 441-9250) any suspected occurrence of false positive or false negative results and significant deviations from the established performance characteristics of the assay of which they become aware.
- All laboratory personnel using your product must be appropriately trained in automated immunoassay techniques and use appropriate laboratory and personal protective equipment when handling this kit, and use your product in accordance with the authorized labeling. All laboratory personnel using the assay must also be trained in and be familiar with the interpretation of results of your product.
- Siemens Healthineers, authorized distributors, and authorized laboratories using the Dimension Vista COV2G assay must ensure that any records associated with this EUA are maintained until otherwise notified by FDA. Such records will be made available to FDA for inspection upon request.

<sup>a</sup> The letter of authorization refers to, "Laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet requirements to perform moderate or high complexity tests as "authorized laboratories".

## Performance Characteristics

### Measuring Interval

The measuring interval for the Dimension Vista COV2G assay is from 600–140,000 QUAL. Results are reported as positive ( $\geq 1000$  QUAL) or negative ( $< 1000$  QUAL). Positive results are reported with a semiquantitative numeric value in QUAL units within the measuring interval.

**Note** Positive results indicated above the upper limit of the assay measuring interval are reported as positive and numeric value is reported as ">140,000 QUAL".

### Detection Capability

Detection capability was determined in accordance with CLSI Document EP17-A2.<sup>32</sup>

Assay results obtained at individual laboratories may vary from the data presented.

The LoB corresponds to the highest measurement result that is likely to be observed for a blank sample with a probability of 95%. The LoB is 550 QUAL.

The LoD corresponds to the lowest concentration of IgG antibodies to SARS-CoV-2 that can be detected with a probability of 95%. The LoD is 600 QUAL.

The LoQ corresponds to the lowest concentration of IgG antibodies to SARS-CoV-2 in a sample at which the within-laboratory %CV is  $\leq 20\%$ . The LoQ is 600 QUAL.

The lower limit of the measuring interval is defined by the LoQ (600 QUAL); however, report negative patient results as  $< 1000$  QUAL.

## Seroconversion Sensitivity

Up to 17 serial draws from 12 SARS-CoV-2 PCR positive patients from the United States were collected and evaluated using the Dimension Vista COV2G assay. All samples were tested in singlicate. The seroconversion study demonstrates the change in COV2G results over time. At this time, it is unknown for how long antibodies persist following infection and if the presence of antibodies confers protective immunity. The following results were obtained.

Panel ID	Bleed Number	Days Post PCR Positive	QUAL	Result Interpretation
Panel 1	1	4	897	Negative
	2	7	17,790	Positive
	3	24	136,739	Positive
	4	25	>140,000	Positive
	5	26	>140,000	Positive
	6	27	>140,000	Positive
	7	28	>140,000	Positive
	8	29	129,754	Positive
	9	30	125,732	Positive
	10	31	134,961	Positive
Panel 2	1	7	3205	Positive
	2	10	9836	Positive
	3	21	26,026	Positive
	4	23	39,134	Positive
Panel 3	1	4	2457	Positive
	2	5	6030	Positive
	3	6	12,513	Positive
	4	7	17,900	Positive
Panel 4	1	1	1787	Positive
	2	3	13,168	Positive
	3	4	20,120	Positive
Panel 5	1	2	13,077	Positive
	2	3	32,980	Positive

Panel ID	Bleed Number	Days Post PCR Positive	QUAL	Result Interpretation
Panel 6	1	9	42,245	Positive
	2	10	43,586	Positive
	3	11	47,824	Positive
	4	12	48,792	Positive
	5	13	52,834	Positive
	6	14	48,967	Positive
	7	15	51,158	Positive
Panel 7	1	8	129,737	Positive
	2	9	>140,000	Positive
	3	9	127,177	Positive
	4	10	139,079	Positive
	5	11	>140,000	Positive
	6	12	131,533	Positive
	7	13	127,966	Positive
	8	14	115,635	Positive
	9	15	112,802	Positive
	10	16	103,413	Positive
	11	17	97,500	Positive
	12	18	93,199	Positive
	13	19	76,020	Positive
	14	20	75,352	Positive
	15	21	69,365	Positive
	16	22	72,934	Positive
	17	23	74,815	Positive
Panel 8	1	22	60,907	Positive
	2	23	57,543	Positive
	3	24	55,638	Positive
	4	25	56,022	Positive
	5	26	56,504	Positive
	6	26	30,070	Positive
	7	27	32,893	Positive

Panel ID	Bleed Number	Days Post PCR Positive	QUAL	Result Interpretation
Panel 9	1	7	15,730	Positive
	2	10	57,286	Positive
	3	11	69,861	Positive
	4	15	93,201	Positive
	5	16	94,548	Positive
	6	17	95,945	Positive
	7	18	97,521	Positive
	8	19	101,439	Positive
Panel 10	1	0	1687	Positive
	2	1	2009	Positive
Panel 11	1	5	700	Negative
	2	6	773	Negative
	3	7	994	Negative
	4	8	1334	Positive
	5	11	2854	Positive
	6	12	3703	Positive
	7	13	4694	Positive
Panel 12	1	0	11,675	Positive
	2	1	17,903	Positive

Assay results obtained at individual laboratories may vary from the data presented.

## Clinical Agreement

Positive percent agreement and negative percent agreement were determined in accordance with CLSI Document EP12-A2.<sup>33</sup> The performance of the Dimension Vista COV2G assay was determined by testing a total of 1754 samples using the Dimension Vista System.

## Positive Percent Agreement

Positive percent agreement was determined by testing 245 samples from individuals with a clinical diagnosis of COVID-19 based on a positive polymerase chain reaction (PCR) method and presented by days post-symptom onset. The results are shown in the table below:

Days Post Symptom Onset	N <sup>a</sup>	Positive	Negative	Positive Percent Agreement (95% CI) <sup>b</sup>
≤ 7	113	69	44	61.1% (51.8%–69.5%)
8–14	42	39	3	92.9% (81.0%–97.5%)
≥ 15	90	90	0	100.0% (95.9%–100.0%)

<sup>a</sup> Number of samples tested.

<sup>b</sup> Confidence Interval.

## Negative Percent Agreement

Negative percent agreement was determined by testing 1509 samples collected prior to the COVID-19 outbreak (before mid-December 2019) from apparently healthy individuals and apparently healthy pregnant women in the United States. The results are shown in the table below:

Group	N <sup>a</sup>	Negative	Positive	Negative Percent Agreement (95% CI) <sup>b</sup>
Healthy Individuals	1442	1442	0	100.0% (99.7%–100.0%)
Healthy, Pregnant Women	67	67	0	100.0% (94.6%–100.0%)
<b>Total</b>	<b>1509</b>	<b>1509</b>	<b>0</b>	<b>100.0%</b> <b>(99.7%–100.0%)</b>

<sup>a</sup> Number of samples tested.

<sup>b</sup> Confidence Interval.

## Single Site Precision Study

The assay was designed to have the following precision.

Concentration Interval	Precision	
QUAL	Repeatability (% CV)	Within-Laboratory (% CV)
800–2000	≤ 10.0% CV	≤ 12.0% CV
> 2000	≤ 12.0% CV	≤ 15.0% CV

Precision was determined in accordance with CLSI Document EP05-A3.<sup>34</sup> Samples were assayed on the Dimension Vista System in duplicate in 2 runs per day for 20 days.

The following results were obtained:

#### Semiquantitative Analysis

Specimen Type	N <sup>a</sup>	Mean (QUAL)	Repeatability		Within-Laboratory	
			SD <sup>b</sup> (QUAL)	CV <sup>c</sup> (%)	SD (QUAL)	CV (%)
Serum A	80	779	12.0	1.54	18.0	2.31
Serum C	80	911	15.2	1.66	21.9	2.41
Serum D	80	1410	17.3	1.22	37.9	2.68
Serum B	80	2925	58.1	1.99	94.3	3.22
Serum E	80	102,139	2151	2.11	4421	4.33
Control 1 <sup>d</sup>	80	523	13.5	2.58	16.6	3.18
Control 2 <sup>e</sup>	80	2193	33.6	1.53	62.6	2.85

<sup>a</sup> Number of results.

<sup>b</sup> Standard deviation.

<sup>c</sup> Coefficient of variation.

<sup>d</sup> Negative.

<sup>e</sup> Positive.

Assay results obtained at individual laboratories may vary from the data presented.

## Precision Study for Evaluation of Instrument-to-Instrument and Lot-to-Lot Variability (Reproducibility)

Instrument and lot variability of the Dimension Vista COV2G assay was evaluated on 2 Dimension Vista instruments using 2 reagent lots. Samples and assay controls (a negative control and a positive control) were assayed in duplicate in 2 runs per day for 3 days. The data were analyzed to calculate the following components of precision: repeatability, between-run, between-day, between-lot, between-instrument, and reproducibility (total). Results of the Dimension Vista COV2G assay are presented in the following table:

The following results were obtained:

Specimen Type	N <sup>a</sup>	Mean QUAL	Repeatability		Between–Run		Between–Day		Between–Lot		Between Instrument/Site		Total Precision	
			SD <sup>b</sup> QUAL	CV <sup>c</sup> (%)	SD QUAL	CV (%)	SD QUAL	CV (%)	SD QUAL	CV (%)	SD QUAL	CV (%)	SD QUAL	CV (%)
Serum A	48	782	9.7	1.2	7.8	1.0	3.0	0.4	24.4	3.1	0.0	0.0	27.6	3.5
Serum C	48	918	14.7	1.6	9.2	1.0	1.9	0.2	21.3	2.3	10.6	1.2	29.5	3.2
Serum D	48	1446	19.8	1.4	15.6	1.1	11.9	0.8	37.4	2.6	0.0	0.0	46.6	3.2
Serum B	48	3057	19.8	0.6	15.6	0.5	11.9	0.4	37.4	1.2	0.0	0.0	46.6	1.5
Serum E	48	104,903	2992.0	2.9	0.0	0.0	1702.6	1.6	3557.5	3.4	4727.3	4.5	6845.0	6.5
Control 1 <sup>d</sup>	48	515	14.1	2.7	11.2	2.2	0.0	0.0	33.2	6.4	0.0	0.0	37.7	7.3
Control 2 <sup>e</sup>	48	2396	49.8	2.1	44.3	1.8	0.0	0.0	223.4	9.3	0.0	0.0	233.1	9.7

<sup>a</sup> Number of results.

<sup>b</sup> Standard deviation.

<sup>c</sup> Coefficient of variation.

<sup>d</sup> Negative.

<sup>e</sup> Positive.



## Specimen Equivalency

Specimen equivalency was determined using the weighted Deming linear regression model of matched sample sets (serum, dipotassium EDTA plasma, and lithium heparin plasma) from the same donors on the Dimension Vista System in accordance with CLSI Document EP35.<sup>35</sup> Slope, intercept and correlation coefficient were calculated. Results from plasma samples were compared to serum results. The following results were obtained:

Specimen (y)	N <sup>a</sup>	Sample Interval (QUAL)	y-intercept	Slope	r <sup>b</sup>
Dipotassium EDTA plasma	45	1542–108,569	-1.30	0.991	0.999
Lithium Heparin plasma	45	1542–108,569	-18.81	0.989	0.999

<sup>a</sup> Number of samples tested.

<sup>b</sup> Correlation coefficient.

Assay results obtained at individual laboratories may vary from the data presented.

## Interferences

Interfering testing was performed on endogenous and potentially interfering substances commonly found in serum and plasma specimens in accordance with CLSI Document EP07-ED3.<sup>36</sup> Bias is the difference in the results between the control sample (does not contain the interferent) and the test sample (contains the interferent) expressed in percent. The Dimension Vista COV2G assay has ≤ 10% interference from hemoglobin, bilirubin, intralipid, and biotin at 1200 ng/mL. The Dimension Vista COV2G assay showed > 10% interference from biotin at 3500 ng/mL.

## Hemolysis, Icterus, and Lipemia (HIL)

The following results were obtained:

Substance	Substance Concentration Conventional Units (SI Units)	Analyte Concentration (QUAL)	Bias %
Hemoglobin	1000 mg/dL (0.625 mmol/L)	563 1168	-1.5 -0.3
Bilirubin, conjugated	40 mg/dL (475 mmol/L)	549 1152	0.9 -0.1
Bilirubin, unconjugated	40 mg/dL (684 mmol/L)	557 1137	-1.9 1.2
Lipemia (Intralipid <sup>®</sup> ) <sup>a</sup>	1500 mg/dL (16.95 mmol/L)	549 1155	2.0 1.4

<sup>a</sup> SI units calculated as triolein.

Assay results obtained at individual laboratories may vary from the data presented.

## Interfering Substances

The following results were obtained:

Substance	Substance Concentration Conventional Units (SI Units)	Analyte Concentration (QUAL)	Bias %
Biotin	1200 ng/mL	552 1133	-0.6 -2.6

Assay results obtained at individual laboratories may vary from the data presented.

## Cross-Reactivity

Cross-reactivity was determined in accordance with CLSI Document EP07-ED3.<sup>36</sup> The assay was evaluated for potential cross-reactivity in specimens with other viral and microbial antibodies and other disease states. No false positives were observed with the potential cross-reactants tested below.

Clinical Category	Number Tested	Number Positive with Dimension Vista COV2G Assay
Adenovirus	5	0
Anti-Influenza A virus IgG and IgM	7	0
Anti-Influenza B virus IgG and IgM	10	0
Anti-hepatitis B virus IgG and IgM	10	0
Anti-Respiratory Syncytial Virus IgG	9	0
Antinuclear antibody (ANA)	10	0
Bordetella pertussis IgG	10	0
Cytomegalovirus IgG	5	0
Epstein Barr Virus	10	0
Hepatitis B core antigen (anti-HBc) IgM	16	0
Hepatitis B surface antigen (HBs Ag)	10	0
Hepatitis C virus (HCV) IgG and IgM	10	0
Human immunodeficiency virus (HIV) total antibody	10	0
Haemophilus influenzae IgG	10	0
Multiple myeloma	10	0
Streptococcus pneumoniae IgG	8	0
Treponema pallidum	10	0
Toxoplasma gondii IgG	9	0
<b>Total</b>	<b>169</b>	<b>0</b>

Assay results obtained at individual laboratories may vary from the data presented.

## Linearity

Linearity testing was performed in accordance with CLSI Document EP06-A.<sup>37</sup> Patient pools containing high levels of SARS-CoV-2 IgG (1 serum, 1 dipotassium EDTA plasma, and 1 lithium heparin plasma) were diluted with negative basepool to prepare a dilution series comprised of nine test levels. Each level was tested in 3 replicates using 2 reagent lots in each specimen matrix using the Dimension Vista System. Linearity was evaluated for each type of specimen using 2 lots. Linearity was demonstrated for the analytical measuring interval of 600–140,000 QUAL with deviations from linearity within 15%. Taking into consideration the estimates of LoB, LoD, LoQ, precision, and linearity, the analytical measuring interval of the Dimension Vista COV2G assay is 600–140,000 QUAL.

## Standardization

Values assigned to the SARS-CoV-2 IgG calibrator (COV2G CAL/CV2G CAL) are traceable to the established SARS-CoV-2 cutoff through the Dimension Vista System patient sample concordance testing.

Currently no reference standard is available for this assay.

## Technical Assistance

For customer support, contact your local technical support provider or distributor.

[siemens.com/healthineers](https://www.siemens.com/healthineers)

## References
























1. Lai CC, Shih TP, Ko WC, Tang HJ, Hsueh PR.. Severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) and coronavirus disease-2019 (COVID-19): The epidemic and the challenges. *Int J Antimicrob Agents*. 2020;55(3):105924.
2. Guo YR, Cao QD, Hong ZS, Tan YY, Chen SD, Jin HJ, *et al*. The origin, transmission and clinical therapies on coronavirus disease 2019 (COVID-19) outbreak - an update on the status. *Mil Med Res*. 2020;7(1):11.
3. Li H, Liu SM, Yu XH, Tang SL, Tang CK. Coronavirus disease 2019 (COVID-19): current status and future perspectives. *Int J Antimicrob Agents*. 2020;29:105951.
4. Wu Z and McGoogan JM. Characteristics of and Important Lessons From the Coronavirus Disease 2019 (COVID-19) Outbreak in China: Summary of a Report of 72314 Cases From the Chinese Center for Disease Control and Prevention . *JAMA* 2020;323(13):2648.
5. Pan A, Liu L, Wang C, Guo H, Hao X, Wang Q, *et al*. Association of Public Health Interventions With the Epidemiology of the COVID-19 Outbreak in Wuhan, China. *JAMA*. 2020;6130.
6. Rothan HA and Byrareddy SN. The epidemiology and pathogenesis of coronavirus disease (COVID-19) outbreak. *J Autoimmun*. 2020;109:102433.
7. Lai CC, Liu YH, Wang CY, Wang YH, Hsueh SC, Yen MY, *et al*. Asymptomatic carrier state, acute respiratory disease, and pneumonia due to severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2): Facts and myths. *J Microbiol Immunol Infect*. 2020;pii: S1684-1182;(20):30040-2.
8. Sethuraman N, Jeremiah SS, Ryo A. Interpreting Diagnostic Tests for SARS-CoV-2. *JAMA*. 2020; doi: 10.1001/jama.2020.8259.
9. Long Q, Liu B, Deng H, *et al*. Antibody responses to SARS-CoV-2 in patients with COVID-19. *Nat Med*. 2020; doi: 10.1038/s41591-020-0897-1.
10. Zhao J, Yuan Q, Wang H, Liu W, Liao X, Su Y, *et al*. Antibody responses to SARS-CoV-2 in patients of novel coronavirus disease 2019. *Clin Infect Dis*. 2020;pii:344.
11. Tay, MZ, Poh, CM, Rénia L, MacAry PA, Ng LFP.. The trinity of COVID-19: immunity, inflammation and intervention. *Nat Rev Immunol*. 2020;20(6):363–374. doi: 10.1038/s41577-020-0311-8.
12. Ou X, Liu Y, Lei X, *et al*. Characterization of spike glycoprotein of SARS-CoV-2 on virus entry and its immune cross-reactivity with SARS-CoV. *Nat Commun*. 2020;11(1):1620. doi: 10.1038/s41467-020-15562-9.
13. Walls AC, Park JY, Tortorici MA, Wall A, McGuire AT, Veesler D. Structure, Function, and Antigenicity of the SARS-CoV-2 Spike Glycoprotein. *Cell*. 2020;181(2):281–292 doi: 10.1016/j.cell.2020.02.058.
14. Ni F, Ye F, Cheng M, *et al*. Detection of SARS-CoV-2-Specific Humoral and Cellular Immunity in COVID-19 Convalescent Individuals. *Immunity*. 2020:S1074-7613: doi: 10.1016/j.immuni.2020.04.023.





















15. Li K, Wu M, Huang B, et al. The Dynamic Changes of Antibodies against SARS-CoV-2 during the Infection and Recovery of COVID-19. <https://doi.org/10.1101/2020.05.18.20105155>. Accessed June 9, 2020.
16. Wu F, Wang A, Liu M, et al. Neutralizing antibody responses to SARS-CoV-2 in a COVID-19 recovered patient cohort and their implications. <https://doi.org/10.1101/2020.03.30.20047365>. Accessed June 9, 2020.
17. Ju B, Zhang Q, Ge J, et al. Human neutralizing antibodies elicited by SARS-CoV-2 infection. *Nature*. 2020; doi: 10.1038/s41586-020-2380-z.
18. Zhou G, Zhao Q. Perspectives on therapeutic neutralizing antibodies against the Novel Coronavirus SARS-CoV-2. *Int. J. Biol. Sci.* 2020;16(10):1718–1723. doi: 10.7150/ijbs.45123.
19. Jiang S, Hillyer C, Lanying D. Neutralizing Antibodies against SARS-CoV-2 and Other Human Coronaviruses. *Trends Immunol.* 2020;41(5):355–359. doi:10.1016/j.it.2020.03.007.
20. Kai-Wang To K, Tsang OT-Y, Leung W-S, et al. Temporal profiles of viral load in posterior oropharyngeal saliva samples and serum antibody responses during infection by SARS-CoV-2: an observational cohort study. *Lancet Infect Dis.* 2020;20(5):565–574. doi: 10.1016/S1473-3099(20)30196-1.
21. Zhu F-C, Li Y-H, Guan X-H, et al. Safety, tolerability, and immunogenicity of a recombinant adenovirus type-5 vectored COVID-19 vaccine: a dose-escalation, open-label, non-randomized, first-in-human trial. *Lancet*. 2020;S0140–6736. doi: 10.1016/S0140-6736(20)31208-3.
22. Chen WH, Strych U, Hotez PJ, Bottazzi ME. The SARS-CoV-2 Vaccine Pipeline: an Overview. *Curr Trop Med Rep.* 2020;1-4. doi:10.1007/s40475-020-00201-6.
23. Amanat F, Krammer F. SARS-CoV-2 Vaccines: Status Report. *Immunity.* 2020;52(4):583–589. doi: 10.1016/j.immuni.2020.03.007.
24. Ullman EF, Kirakossian H, Switchenko AC, Ishkanian J, Ericson M, Wartchow CA, et al. Luminescent oxygen channeling assay (LOCI®): sensitive, broadly applicable homogenous immunoassay method. *Clin Chem* 1996;42(9):1518–1526.
25. Ullman EF, Kirakossian H, Singh S, Wu ZP, Irvin BR, Pease JS, et al. Luminescent oxygen channeling immunoassay: Measurement of particle binding kinetics by chemiluminescence. *Proc. Natl. Acad. Sci. USA*, 1994;91:5426-5430.
26. Ullman EF, Homogenous Immunoassays. In: Wild D, ed. *The Immunoassay Handbook*, 2nd ed. London, UK: Nature Publishing Group 2001:192-194.
27. Clinical and Laboratory Standards Institute. *Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline—Fourth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2014. CLSI Document M29-A4.
28. Clinical and Laboratory Standards Institute. *Procedures for the Collection of Diagnostic Blood Specimens by Venipuncture; Approved Standard—Sixth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2007. CLSI Document GP41-A6.
29. Clinical and Laboratory Standards Institute. *Tubes and Additives for Venous and Capillary Blood Specimen Collection; Approved Standard—Sixth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2010. CLSI Document GP39-A6.
30. Clinical and Laboratory Standards Institute. *Procedures for the Handling and Processing of Blood Specimens for Common Laboratory Tests; Approved Guideline—Fourth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2010. CLSI Document GP44-A4.
31. Data on file at Siemens Healthcare Diagnostics.
32. Clinical and Laboratory Standards Institute. *Evaluation of Detection Capability for Clinical Laboratory Measurement Procedures; Approved Guideline—Second Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2012. CLSI Document EP17-A2.

33. Clinical and Laboratory Standards Institute. *User Protocol for Evaluation of Qualitative Test Performance; Approved Guideline—Second Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2008. CLSI Document EP12-A2.
34. Clinical and Laboratory Standards Institute. *Evaluation of Precision of Quantitative Measurement Procedures; Approved Guideline—Third Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2014. CLSI Document EP05-A3.
35. Clinical and Laboratory Standards Institute. *Assessment of Equivalence or Suitability of Specimen Types for Medical Laboratory Measurement Procedures—First Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2019. CLSI Document EP35.
36. Clinical and Laboratory Standards Institute. *Interference Testing in Clinical Chemistry; Approved Guideline—Third Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2018. CLSI Document EP07-ED3.
37. Clinical and Laboratory Standards Institute. *Evaluation of the Linearity of Quantitative Measurement Procedures: A Statistical Approach; Approved Guideline*. Wayne, PA: Clinical and Laboratory Standards Institute; 2003. CLSI Document EP06-A.

## Definition of Symbols

The following symbols may appear on the product labeling:


Symbol	Symbol Title	Symbol	Symbol Title
	Legal Manufacturer		Authorized Representative in the European Community
	Use-by Date		Batch Code
	Catalog Number		Contains sufficient for <n> tests
	Consult Instruction for Use		Version of Instruction for Use
	Internet URL address to access the electronic instructions for use		Revision
	<i>In vitro</i> diagnostic medical device		Unique Device Identification (UDI) barcode
<b>RxOnly</b>	Prescription device (US only)		CE Marking
	CE Marking with Notified Body		Keep away from sunlight and/or heat
	Temperature Limit		Lower Limit of Temperature
	Upper Limit of Temperature		Do not freeze
	Do not re-use		This way up
	Recycle		Caution / Warning

Symbol	Symbol Title	Symbol	Symbol Title
	Biological Risk		Explosive (GHS)
	Flammable (GHS)		Oxidizing (GHS)
	Corrosive (GHS)		Toxic (GHS)
	Irritant (GHS)		Respiratory / Internal Health (GHS)
	Environmental (GHS)		Common Units
	International System of Units		Date format (year-month-day)
	Date format (year-month)		Printed with soy ink
	Mixing of substances		Non-sterile
	Contents		Reconstitution volume
	Level		Scalers

## Legal Information

Dimension, Dimension Vista, EXL, and LOCI are trademarks of Siemens Healthineers.

© 2021 Siemens Healthineers. All rights reserved.

 Siemens Healthcare Diagnostics Inc.  
500 GBC Drive  
Newark, DE 19714  
USA  
siemens-healthineers.com

**Siemens Healthineers Headquarters**  
Siemens Healthcare GmbH  
Henkestr. 127  
91052 Erlangen  
Germany  
Phone: +49 9131 84-0  
siemens-healthineers.com

## SARS-CoV-2 IgG Quality Control (COV2G QC/ CV2G QC)

Current Revision and Date <sup>a</sup>		Rev. 01, 2021-01
Product Name	Dimension Vista/Dimension EXL SARS-CoV-2 IgG Quality Control (COV2G QC/CV2G QC)	<b>REF</b> KC873 11417773
Abbreviated Product Name	Dimension Vista COV2G QC Dimension EXL CV2G QC	
	6 x 1.0 mL positive quality control 6 x 1.0 mL negative quality control	
Systems	Dimension Vista System Dimension EXL with LOCI Module Dimension EXL 200	

<sup>a</sup> A vertical bar in the page margin indicates technical content that differs from the previous version.



For USA

For Use Under Emergency Use Authorization Only.

For *in vitro* diagnostic use.

For Prescription Use Only.

### Intended Use

The SARS-CoV-2 IgG Quality Control (COV2G QC/CV2G QC) material is an *in vitro* diagnostic product to monitor the accuracy of the SARS-CoV-2 IgG antibody assay (COV2G QC/CV2G QC) on the Dimension Vista® System and the Dimension® EXL™ integrated chemistry system with LOCI® module.

## Material Description

Material Description	Storage	Stability <sup>a</sup>
Dimension Vista/Dimension EXL SARS-CoV-2 IgG Quality Control (COV2G QC/CV2G QC) 1.0 mL/vial	Unopened at -25 – -15°C	Until expiration date on product
Neg (-): human plasma base; Pos (+): bovine serum albumin containing SARS-CoV-2 Spike S1 Antibody	Opened at 2–8°C	10 days after opening product
	Punctured at 2–8°C <sup>b</sup>	10 days after puncturing product
	Unopened at 2–8°C	10 days after thaw

<sup>a</sup> Refer to *Storage and Stability*.

<sup>b</sup> Dimension Vista System.

**Note** Once the cap is removed, assigned values are stable for 10 days when re-capped immediately after use and stored at 2–8°C. Do not use re-capped vials on board the instrument.

## Warnings and Precautions

For Use Under Emergency Use Authorization Only.

For *in vitro* diagnostic use.

For Prescription Use Only.

This product has not been FDA cleared or approved, but has been authorized for emergency use by FDA under an EUA for use by laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet requirements to perform moderate or high complexity tests.

This product has been authorized only for detecting the presence of IgG antibodies against SARS-CoV-2, not for any other viruses or pathogens.

The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of *in vitro* diagnostic tests for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug and Cosmetic Act, 21 U.S.C. § 360bbb3(b)(1), unless the declaration is terminated or authorization is revoked sooner.

### CAUTION

Federal (USA) law restricts this device to sale by or on the order of a licensed healthcare professional.

Safety data sheets (SDS) available on [siemens-healthineers.com](https://www.siemens-healthineers.com).





H317, H412  
P280, P273,  
P302+P352,  
P333+P313,  
P362+P364, P501

#### Warning!

May cause an allergic skin reaction. Harmful to aquatic life with long lasting effects.

Wear protective gloves/protective clothing/eye protection/face protection. Avoid release to the environment. IF ON SKIN: Wash with plenty of soap and water. If skin irritation or rash occurs: Get medical advice/attention. Take off contaminated clothing and wash it before reuse. Dispose of contents and container in accordance with all local, regional, and national regulations.

**Contains:** reaction mass of 5-chloro-2-methyl-2H-isothiazol-3-one and 2-methyl-2H-isothiazol-3-one (3:1); (Dimension Vista/Dimension EXL COV2G QC/CV2G QC Pos and Neg)



#### Caution Potential Biohazard

Contains human source material.

**Caution:** Contains human source material. Each donation of human blood or blood component was tested by FDA-approved methods for the presence of antibodies to human immunodeficiency virus type 1 (HIV-1) and type 2 (HIV-2), as well as for hepatitis B surface antigen (HBsAg) and antibody to hepatitis C virus (HCV). The test results were negative (not repeatedly reactive). No test offers complete assurance that these or other infectious agents are absent; this material should be handled using good laboratory practices and universal precautions.<sup>1-3</sup>

#### CAUTION

This device contains material of animal origin and should be handled as a potential carrier and transmitter of disease.

Dispose of hazardous or biologically contaminated materials according to the practices of your institution. Discard all materials in a safe and acceptable manner and in compliance with prevailing regulatory requirements.

**Note** For information about quality control material preparation, refer to *Preparing the Quality Control Materials*.

## Storage and Stability

Store unopened product at -25 to -15°C. Do not use products beyond the expiration date printed on the product labeling.

Do not refreeze.

For information about product storage and stability, refer to *Material Description*.

## Performing Quality Control

For quality control of the Dimension Vista COV2G and Dimension EXL CV2G assays, use the Dimension Vista/Dimension EXL COV2G QC/CV2G QC at least once during each day that samples are analyzed.

In addition, perform quality control:

- Following a valid calibration.
- With use of a new lot of reagent.
- When troubleshooting test results that do not match clinical conditions or symptoms.

Additional quality control material can be used at the discretion of the laboratory.

### Qualitative

The Dimension Vista/Dimension EXL COV2G QC/CV2G QC should be used in accordance with the Dimension Vista/Dimension EXL COV2G QC/CV2G QC IFU. The Dimension Vista/Dimension EXL COV2G QC/CV2G QC is a separate qualitative (POS/NEG) quality control material. The system reports quality control results as "POS" or "NEG" relative to the assay cutoff.

### Semiquantitative

The negative quality control will generate an "Assay Range" flag due to the assigned value being below the LoQ of 610 Ind units. This flag generation is expected. Troubleshooting is recommended when quality control recovery is outside the lot-specific range listed in the lot-specific value sheet or laboratory generated quality control ranges. Dilution studies are not necessary with a known standard or quality control with the Dimension EXL CV2G assay when this flag is generated for the negative quality control.

Use the following lot-specific materials to perform quality control:

- For the assigned values and quality control definitions, refer to the quality control lot-specific value sheet 

CONTROL	LOT	VAL
---------	-----	-----

 provided. The assigned values are traceable to the standardization of the assay.

For instructions about how to perform the quality control procedure, refer to the Dimension EXL Operator's Guide and Online Help or Dimension Vista Operator's Guide and Online Help.

Follow your laboratory's quality control procedures if the results obtained do not fall within the acceptable limits. For information about entering quality control definitions, refer to the Dimension EXL Operator's Guide and Online Help or Dimension Vista Operator's Guide and Online Help.

Follow government regulations or accreditation requirements for quality control frequency. Individual laboratory quality control programs and procedures may require more frequent quality control testing.

## Preparing the Quality Control Materials

Quality controls are frozen. Allow to equilibrate to room temperature (18–30°C) for one hour (no more than two hours), then gently mix and invert the vials to ensure homogeneity of the material. The required volume for testing depends on the number of replicates. For information about sample volume requirements, refer to the Dimension EXL Operator's Guide and Online Help or Dimension Vista Operator's Guide and Online Help.

**Note** Use quality control material within the stability limits specified in *Storage and Stability* and discard any remaining material.

## Quality Control Procedure

The required volume for testing depends on the number of replicates. For information about sample volume requirements, refer to the Dimension EXL Operator's Guide and Online Help or Dimension Vista Operator's Guide and Online Help.

Use the following lot-specific materials to perform quality control:

- For the assigned values and quality control definitions, refer to the quality control lot-specific value sheet 

CONTROL	LOT	VAL
---------	-----	-----

 provided. The assigned values are traceable to the standardization of the assay.
- Generate lot-specific barcode labels to use with the quality control samples.

For instructions about how to perform the quality control procedure, refer to the Dimension EXL Operator's Guide and Online Help or Dimension Vista Operator's Guide and Online Help.

## Taking Corrective Action

Acceptable performance is achieved when the analyte values obtained are within the expected control interval for the system, as indicated by the manufacturer of the control material or within the interval determined by an appropriate internal laboratory quality control scheme.

If the quality control results do not fall within the expected control interval, do not report results. Perform corrective actions in accordance with established laboratory protocol. If the results continue to fall outside your laboratory's protocol, recalibrate the assay and run the controls again.

If necessary, contact your local technical support provided or distributor for assistance.

## Limitations

The results obtained using quality control material depend on several factors. Erroneous results can occur from causes such as improper storage, inadequate mixing, or sample handling errors associated with system or assay procedures.

The Dimension Vista/Dimension EXL SARS-CoV-2 IgG Quality Control (COV2G QC/CV2G QC) is only for use with the Dimension Vista/Dimension EXL SARS-CoV-2 IgG antibody assay (COV2G/CV2G). Assay values have not been established for assays other than SARS-CoV-2 IgG Quality Control (COV2G QC/CV2G QC).

Each laboratory should establish corrective measures if individual values fall outside the interval. Follow the applicable government regulations and local guidelines for quality control.

## Technical Assistance

For customer support, contact your local technical support provider or distributor.







siemens-healthineers.com
































## References





- Centers for Disease Control. Perspectives in disease prevention and health promotion update: Universal precautions for prevention of transmission of human immunodeficiency virus, hepatitis B virus and other bloodborne pathogens in healthcare settings. *MMWR*. 1988;37(24):377–382, 387–388.
- Clinical and Laboratory Standards Institute. *Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline—Fourth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2014. CLSI Document M29-A4.
- Clinical and Laboratory Standards Institute. *Procedures for the Handling and Processing of Blood Specimens for Common Laboratory Tests; Approved Guideline—Fourth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2010. CLSI Document GP44-A4.

## Definition of Symbols

The following symbols may appear on the product labeling:

Symbol	Symbol Title	Symbol	Symbol Title
	Legal Manufacturer		Authorized Representative in the European Community
	Use-by Date		Batch Code
	Catalog Number		Contains sufficient for <n> tests

Symbol	Symbol Title	Symbol	Symbol Title
	Consult Instruction for Use		Version of Instruction for Use
	Internet URL address to access the electronic instructions for use	<b>Rev.</b> 	Revision
	<i>In vitro</i> diagnostic medical device		Unique Device Identification (UDI) barcode
<b>RxOnly</b>	Prescription device (US only)		CE Marking
	CE Marking with Notified Body		Keep away from sunlight and/or heat
	Temperature Limit		Lower Limit of Temperature
	Upper Limit of Temperature		Do not freeze
	Do not re-use		This way up
	Recycle		Caution / Warning
	Biological Risk		Explosive (GHS)
	Flammable (GHS)		Oxidizing (GHS)
	Corrosive (GHS)		Toxic (GHS)
	Irritant (GHS)		Respiratory / Internal Health (GHS)
	Environmental (GHS)		Common Units
	International System of Units	<b>YYYY-MM-DD</b>	Date format (year-month-day)
<b>YYYY-MM</b>	Date format (year-month)		Printed with soy ink
	Mixing of substances		Non-sterile

Symbol	Symbol Title	Symbol	Symbol Title
	Contents		Reconstitution volume
	Level		Scalars

## Legal Information

Dimension, Dimension Vista, EXL, and LOCI are trademarks of Siemens Healthineers.

© 2021 Siemens Healthineers. All rights reserved.



Siemens Healthcare Diagnostics Inc.  
500 GBC Drive  
Newark, DE 19714  
USA  
[siemens-healthineers.com](http://siemens-healthineers.com)

### Siemens Healthineers Headquarters

Siemens Healthcare GmbH  
Henkestr. 127  
91052 Erlangen  
Germany  
Phone: +49 9131 84-0  
[siemens-healthineers.com](http://siemens-healthineers.com)



## SARS-CoV-2 IgG calibrator (COV2G CAL/ CV2G CAL)

<b>Current Revision and Date<sup>a</sup></b>	Rev. 01, 2021-01
<b>Product Name</b>	Dimension Vista/Dimension EXL SARS-CoV-2 IgG calibrator (COV2G CAL/ CV2G CAL) <span style="float: right;"><b>REF</b> KC872 11417772</span>
<b>Abbreviated Product Name</b>	Dimension Vista COV2G CAL Dimension EXL CV2G CAL
	2 x 1.0 mL calibrator level 1/A 2 x 1.0 mL calibrator level 2/B 2 x 1.0 mL calibrator level 3/C 2 x 1.0 mL calibrator level 4/D 2 x 1.0 mL calibrator level 5/E
<b>Systems</b>	Dimension Vista System Dimension EXL with LOCI Module Dimension EXL 200

<sup>a</sup> A vertical bar in the page margin indicates technical content that differs from the previous version.



For USA

For Use Under Emergency Use Authorization Only.

For *in vitro* diagnostic use.

For Prescription Use Only.

### Intended Use

The SARS-CoV-2 IgG calibrator (COV2G CAL/CV2G CAL) is an *in vitro* diagnostic product for calibration of the SARS-CoV-2 IgG antibody assays (COV2G/CV2G) on the Dimension Vista® System and the Dimension® EXL™ integrated chemistry system with LOCI® module.

## Material Description

Material Description	Storage	Stability <sup>a</sup>
Dimension Vista/Dimension EXL SARS-CoV-2 IgG calibrator (COV2G CAL/ CV2G CAL) 1.0 mL/vial	Unopened at -25 – -15°C	Until expiration date on product
Level 1/A: human plasma base; Level 2/B: human plasma base containing SARS-CoV-2 Spike S1 Antibody; Levels 3/C, 4/D, and 5/E: bovine serum albumin containing SARS-CoV-2 Spike S1 Antibody	Opened at 2–8°C	10 days after opening product
	Punctured at 2–8°C <sup>b</sup>	10 days after puncturing product
	Unopened at 2–8°C	10 days after thaw

<sup>a</sup> Refer to *Storage and Stability*.

<sup>b</sup> Dimension Vista System.

**Note** Once the cap is removed, assigned values are stable for 10 days when re-capped immediately after use and stored at 2–8°C. Do not use re-capped vials on board the instrument.

## Warnings and Precautions

For Use Under Emergency Use Authorization Only.

For *in vitro* diagnostic use.

For Prescription Use Only.

This product has not been FDA cleared or approved, but has been authorized for emergency use by FDA under an EUA for use by laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet requirements to perform moderate or high complexity tests.

This product has been authorized only for detecting the presence of IgG antibodies against SARS-CoV-2, not for any other viruses or pathogens.

The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of *in vitro* diagnostic tests for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug and Cosmetic Act, 21 U.S.C. § 360bbb3(b)(1), unless the declaration is terminated or authorization is revoked sooner.

### CAUTION

Federal (USA) law restricts this device to sale by or on the order of a licensed healthcare professional.

Safety data sheets (SDS) available on [siemens-healthineers.com](https://www.siemens-healthineers.com).





H317, H412  
P280, P273,  
P302+P352,  
P333+P313,  
P362+P364, P501

#### Warning!

May cause an allergic skin reaction. Harmful to aquatic life with long lasting effects.

Wear protective gloves/protective clothing/eye protection/face protection. Avoid release to the environment. IF ON SKIN: Wash with plenty of soap and water. If skin irritation or rash occurs: Get medical advice/attention. Take off contaminated clothing and wash it before reuse. Dispose of contents and container in accordance with all local, regional, and national regulations.

**Contains:** reaction mass of 5-chloro-2-methyl-2H-isothiazol-3-one and 2-methyl-2H-isothiazol-3-one (3:1); (Dimension Vista/Dimension EXL COV2G CAL/CV2G CAL Levels 1/A, 2/B, 3/C, 4/D, and 5/E)



#### Caution Potential Biohazard

Contains human source material.

**Caution:** Contains human source material. Each donation of human blood or blood component was tested by FDA-approved methods for the presence of antibodies to human immunodeficiency virus type 1 (HIV-1) and type 2 (HIV-2), as well as for hepatitis B surface antigen (HBsAg) and antibody to hepatitis C virus (HCV). The test results were negative (not repeatedly reactive). No test offers complete assurance that these or other infectious agents are absent; this material should be handled using good laboratory practices and universal precautions.<sup>1-3</sup>

#### CAUTION

This device contains material of animal origin and should be handled as a potential carrier and transmitter of disease.

Dispose of hazardous or biologically contaminated materials according to the practices of your institution. Discard all materials in a safe and acceptable manner and in compliance with prevailing regulatory requirements.

**Note** For information about calibrator preparation, refer to *Preparing the Calibrators*.

## Storage and Stability

Store unopened product at -25 to -15°C. Do not use products beyond the expiration date printed on the product labeling.

Do not refreeze.

For information about product storage and stability, refer to *Material Description*.

## Performing Calibration

For calibration of the Dimension Vista/Dimension EXL SARS-CoV-2 IgG antibody assay (COV2G/ CV2G), use SARS-CoV-2 IgG calibrator (COV2G CAL/CV2G CAL).

Use the following lot-specific materials to perform calibration:

- For the assigned values and calibrator definitions, refer to the calibrator lot-specific value sheet 

CAL	LOT	VAL
-----	-----	-----

 provided. The assigned values are traceable to the standardization of the assay.

Perform a calibration:

- At the end of the calibration interval.
- When changing lot numbers of reagents.

- When indicated by quality control results.
- After major maintenance or service.

Follow government regulations or accreditation requirements for calibration frequency. Individual laboratory quality control programs and procedures may require more frequent calibration.

For instructions about how to perform the calibration procedure, refer to the Dimension EXL Operator's Guide and Online Help or Dimension Vista Operator's Guide and Online Help.

For information about calibration, refer to the assay instructions for use.

## Preparing the Calibrators

Calibrators are frozen. Allow to equilibrate to room temperature (18-30°C) for one hour (no more than two hours), then gently mix and invert the vials to ensure homogeneity of the material.

## Calibration Procedure

Use the following lot-specific materials to perform calibration:

- For the assigned values and calibrator definitions, refer to the calibrator lot-specific value sheet 

CAL	LOT	VAL
-----	-----	-----

 provided. The assigned values are traceable to the standardization of the assay.
- Generate lot-specific barcode labels to use with the calibrator samples, if necessary.

For instructions about how to perform the calibration procedure, refer to the Dimension EXL Operator's Guide and Online Help or Dimension Vista Operator's Guide and Online Help.

## Technical Assistance

For customer support, contact your local technical support provider or distributor.






























[siemens-healthineers.com](http://siemens-healthineers.com)









## References

1. Centers for Disease Control. Perspectives in disease prevention and health promotion update: Universal precautions for prevention of transmission of human immunodeficiency virus, hepatitis B virus and other bloodborne pathogens in healthcare settings. *MMWR*. 1988;37(24):377–382, 387–388.
2. Clinical and Laboratory Standards Institute. *Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline—Fourth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2014. CLSI Document M29-A4.
3. Clinical and Laboratory Standards Institute. *Procedures for the Handling and Processing of Blood Specimens for Common Laboratory Tests; Approved Guideline—Fourth Edition*. Wayne, PA: Clinical and Laboratory Standards Institute; 2010. CLSI Document GP44-A4.

## Definition of Symbols

The following symbols may appear on the product labeling:


Symbol	Symbol Title	Symbol	Symbol Title
	Legal Manufacturer		Authorized Representative in the European Community
	Use-by Date		Batch Code
	Catalog Number		Contains sufficient for <n> tests
	Consult Instruction for Use		Version of Instruction for Use
 <a href="https://www.siemens.com/eifu">siemens.com/eifu</a>	Internet URL address to access the electronic instructions for use		Revision
	<i>In vitro</i> diagnostic medical device		Unique Device Identification (UDI) barcode
<b>RxOnly</b>	Prescription device (US only)		CE Marking
	CE Marking with Notified Body		Keep away from sunlight and/or heat
	Temperature Limit		Lower Limit of Temperature
	Upper Limit of Temperature		Do not freeze
	Do not re-use		This way up
	Recycle		Caution / Warning
	Biological Risk		Explosive (GHS)
	Flammable (GHS)		Oxidizing (GHS)
	Corrosive (GHS)		Toxic (GHS)
	Irritant (GHS)		Respiratory / Internal Health (GHS)
	Environmental (GHS)		Common Units

Symbol	Symbol Title	Symbol	Symbol Title
	International System of Units	YYYY-MM-DD	Date format (year-month-day)
YYYY-MM	Date format (year-month)		Printed with soy ink
	Mixing of substances		Non-sterile
	Contents		Reconstitution volume
	Level		Scalers

## Legal Information

Dimension, Dimension Vista, EXL, and LOCI are trademarks of Siemens Healthineers.

© 2021 Siemens Healthineers. All rights reserved.

 Siemens Healthcare Diagnostics Inc.  
500 GBC Drive  
Newark, DE 19714  
USA  
siemens-healthineers.com

### Siemens Healthineers Headquarters

Siemens Healthcare GmbH  
Henkestr. 127  
91052 Erlangen  
Germany  
Phone: +49 9131 84-0  
siemens-healthineers.com