

SARS-CoV-2/Flu for BD MAX™ System

For Emergency Use Authorization (EUA) only
For *In Vitro* Diagnostic Use
For use with the BD MAX™ System

REF

445011
P0256(01)
2021-02
English



IVD

R_x Only

INTENDED USE

BD SARS-CoV-2/Flu for BD MAX™ System is an automated multiplexed real-time RT-PCR test intended for the simultaneous qualitative detection and differentiation of nucleic acid from SARS-CoV-2, influenza A and/or influenza B in nasopharyngeal and anterior nasal swabs collected from individuals suspected of respiratory viral infection consistent with COVID-19 by their healthcare provider. Clinical signs and symptoms of respiratory viral infection due to SARS-CoV-2 and influenza can be similar. Testing is limited to laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet the requirements to perform moderate or high complexity tests.

BD SARS-CoV-2/Flu for BD MAX System is intended for use in simultaneous detection and differentiation of SARS-CoV-2, influenza A, and/or influenza B nucleic acid in clinical specimens, and is not intended to detect influenza C. The SARS-CoV-2, influenza A, and/or influenza B RNA is generally detectable in upper respiratory samples during the acute phase of infection.

Positive results are indicative of active infection but do not rule out bacterial infection or co-infection with other pathogens not detected by the test. Clinical correlation with patient history and other diagnostic information is necessary to determine patient infection status. The agent detected may not be the definite cause of disease.

Laboratories within the United States and its territories are required to report all SARS-CoV-2 results to the appropriate public health authorities.

Negative results do not preclude SARS-CoV-2, influenza A, and/or influenza B infection and should not be used as the sole basis for treatment, or other patient management decisions. Negative results must be combined with clinical observations, patient history, and/or epidemiological information.

BD SARS-CoV-2/Flu for BD MAX System is intended for use by qualified clinical laboratory personnel specifically instructed and trained on the use of the BD MAX System and in vitro diagnostic procedures. BD SARS-CoV-2/Flu for BD MAX System is only for use under the Food and Drug Administration's Emergency Use Authorization.

EXPLANATION OF THE TEST

Total nucleic acid (TNA) is isolated and purified using BD SARS-CoV-2/Flu for BD MAX System from nasopharyngeal or anterior nasal swabs collected in BD Universal Viral Transport System (UVT) or Copan Universal Transport Media System (UTM). Patient sample is transferred to the BD Molecular Respiratory Sample Buffer Tube provided with the BD SARS-CoV-2/Flu for BD MAX System kit and placed in the BD MAX System. The BD Respiratory Unitized Reagent Strip contains a combination of lytic and extraction reagents designed to perform cell lysis and TNA extraction. Eluted TNA is transferred to the BD SARS-CoV-2/Flu master mix. The final rehydrated master mix is transferred to a PCR cartridge for rRT-PCR.

The BD SARS-CoV-2/Flu for BD MAX System utilizes multiplexed primers and probes targeting RNA from the nucleocapsid phosphoprotein gene (N1 and N2 regions) of the SARS-CoV-2 coronavirus, a conserved region of the matrix protein M1 gene for influenza A, conserved regions of the matrix protein M1 gene and hemagglutinin (HA) gene for influenza B, and the human RNase P gene. The primer and probe sets for SARS-CoV-2 are based on the United States Centers for Disease Control and Prevention (US CDC) assay for specific detection of SARS-CoV-2 by amplifying two unique regions of the N gene (i.e., N1 and N2). SARS-CoV-2 targets, N1 and N2, are indistinguishable as they are detected in the same optical channel. Influenza B targets, M1 and HA, are also indistinguishable and are detected in the same optical channel.

An internal control targeting the human RNase P gene will be co-amplified along with SARS-CoV-2, influenza A, and influenza B gene targets (if present) and will serve as an endogenous nucleic acid extraction control present in all properly collected patient samples. This control serves as both an extraction control and an internal amplification control.

PRINCIPLES OF THE PROCEDURE

A combination of lytic and extraction reagents is used to perform cell lysis and TNA extraction. Nucleic acids released from the target organisms are captured on magnetic affinity beads. The beads, together with the bound nucleic acids, are washed and the nucleic acids are eluted by a combination of heat and pH variation. Eluted TNA is added to neutralization buffer, mixed, and transferred to BD SARS-CoV-2/Flu master mix for rehydration. After reconstitution, the BD MAX System dispenses a fixed volume of rRT-PCR-ready solution containing extracted nucleic acids into the PCR Cartridge. Microvalves on the cartridge are sealed by the system prior to initiating PCR in order to contain the amplification mixture and thus prevent evaporation and contamination.

The amplified cDNA targets are detected using hydrolysis (TaqMan®) probes, labeled at one end with a fluorescent reporter dye (fluorophore), and at the other end, with a quencher moiety. Probes labeled with different fluorophores are used to detect the targets in different optical channels of the BD MAX System. When the probes are in their native state, the fluorescence of the fluorophore is quenched due to its proximity to the quencher. However, in the presence of target cDNA, the probes hybridize to their complementary sequences and are hydrolyzed by the 5'-3' exonuclease activity of the DNA polymerase as it synthesizes the nascent strand along the cDNA template. As a result, the fluorophores are separated from the quencher molecules and fluorescence is emitted. The amount of fluorescence detected in the optical channels is directly proportional to the quantity of the corresponding probe that is hydrolyzed. The BD MAX System monitors these signals at each cycle of the PCR and interprets the data at the end of the reaction to provide qualitative test results for each analyte.

REAGENTS AND MATERIALS

REF	Contents	Quantity
445011	BD SARS-CoV-2/Flu for BD MAX System Master Mix (D9) <i>Dried PCR Master Mix containing nucleotides and specific molecular probes (0.005% w/v) and primers (0.009% w/v) along with PCR enzyme (0.004% w/v).</i>	24 tests (2 x 12 tubes)
	BD Respiratory for BD MAX System Extraction Tube (D4) <i>Dried extraction reagent containing DNA/RNA magnetic affinity beads (6.41% w/v) and Proteinase K (6.7% w/v).</i>	24 tests (2 x 12 tubes)
	BD Respiratory for BD MAX System Unitized Reagent Strip <i>Unitized Reagent Strip containing wash buffer with 0.004% v/v Tween® 20 (0.75 mL), elution buffer with 0.004% v/v Tween 20 (0.75 mL), and neutralization buffer with 0.004% v/v Tween 20 (0.75 mL) reagents and disposable pipette tips necessary for sample processing and TNA extraction.</i>	24 tests
	BD Molecular Respiratory Sample Buffer Tube <i>(with 2% v/v of Triton® X-100)</i>	24 (2 x 12 tubes)
	Septum Caps	25

EQUIPMENT AND MATERIALS REQUIRED BUT NOT PROVIDED

- BD MAX System (BD Catalog No. 441916)
- BD MAX Sample Rack (BD Catalog No. 441935, 443550, 443551, 444807, or 444808)
- BD MAX PCR Cartridges (BD Catalog No. 437519)
- Copan UTM Collection Kit
- BD UVT Collection Kit
- Vortex Genie 2 (VWR Cat. No. 58815-235 or equivalent)
- Multi-Tube Vortex Mixer (VWR Catalog No. 58816-115 or equivalent)
- Rack compatible with a multi-tube vortexer (e.g., Cryogenic Vial Holder or equivalent)
- Variable Volume Calibrated Pipettor (750 µL volume capable)
- Aerosol resistant micropipette tips
- Disposable gloves, powderless

WARNINGS AND PRECAUTIONS

Danger



H312 Harmful in contact with skin.

H315 Causes skin irritation.

H317 May cause an allergic skin reaction.

H319 Causes serious eye irritation.



H334 May cause allergy or asthma symptoms or breathing difficulties if inhaled.

H335 May cause respiratory irritation.

H350 May cause cancer.

H360 May damage fertility or the unborn child.



H402 Harmful to aquatic life.

H411 Toxic to aquatic life with long lasting effects.

P201 Obtain special instructions before use.

P202 Do not handle until all safety precautions have been read and understood.

P233 Keep container tightly closed.

P261 Avoid breathing dust/fume/gas/mist/vapors/spray.

P264 Wash thoroughly after handling.

P271 Use only outdoors or in a well-ventilated area.

P272 Contaminated work clothing should not be allowed out of the workplace.

P273 Avoid release to the environment.

P280 Wear protective gloves/protective clothing/eye protection/face protection.

P281 Use personal protective equipment as required.

P284 [In case of inadequate ventilation] wear respiratory protection.

P308+P313 IF exposed or concerned: Get medical advice/attention.

P332+P313 IF skin irritation occurs: Get medical advice/attention.

P333+P313 IF skin irritation or rash occurs: Get medical advice/attention.

P342+P311 If experiencing respiratory symptoms: Call a POISON CENTER/doctor.

P337+P313 IF eye irritation persists: Get medical advice/attention.

P305+P351+P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P302+P352 IF ON SKIN: Wash with plenty of water.

P312 Call a POISON CENTER/doctor if you feel unwell.

P321 Specific treatment.

P342+P311 If experiencing respiratory symptoms: Call a POISON CENTER/doctor/...

P304+P340 IF INHALED: Remove person to fresh air and keep comfortable for breathing.

P362+P364 Take off contaminated clothing and wash it before reuse.

P363 Wash contaminated clothing before reuse.

P391 Collect spillage.

P403 Store in a well-ventilated place.

P405 Store locked up.

P501 Dispose of contents/container to an appropriate treatment and disposal facility in accordance with applicable laws and regulations, and product characteristics at time of disposal.

- For *in vitro* diagnostic use under Emergency Use Authorization only.
- For Prescription Use only.
- This product has not been FDA cleared or approved, but has been authorized for emergency use by FDA under an EUA for use by authorized laboratories.
- This product has been authorized for use by laboratories certified under the Clinical Laboratory Improvement Amendments (CLIA) of 1988, 42 U.S.C. 263a, that meet requirements to perform moderate or high complexity tests.
- This product has been authorized only for the detection of nucleic acid from SARS-CoV-2, influenza A, and influenza B, not for any other viruses or pathogens.
- The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of *in vitro* diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug and Cosmetic Act, 21 U.S.C. §360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.
- Positive results are indicative of the presence of SARS-CoV-2, influenza A, and/or influenza B RNA.
- Laboratories within the United States and its territories are required to report all SARS-CoV-2 results to the appropriate public health authorities.

- All patient samples should be handled as if infectious, using good laboratory procedures as outlined in the CLSI Document M29-A4¹ and in Biosafety in Microbiological and Biomedical Laboratories.² Only personnel proficient in handling infectious materials and the use of BD SARS-CoV-2/Flu and BD MAX System should perform this procedure.
- All human-sourced materials should be considered potentially infectious and should be handled with universal precautions. If spillage occurs, follow appropriate site procedures.
- Closely follow procedures and guidelines provided to ensure that the test is performed correctly. Any deviation from the procedures and guidelines may affect optimal test performance.
- Do not use expired reagents and/or materials.
- Do not use the kit if the label that seals the outer box is broken upon arrival.
- Do not use reagents if the protective pouches are open or broken upon arrival.
- Do not use reagents if desiccant is not present or broken inside reagent pouches.
- Do not remove desiccant from reagent pouches.
- Close protective pouches of reagents promptly with the zip seal after each use. Remove any excess air in the pouches prior to sealing.
- Protect reagents against heat and humidity. Prolonged exposure to humidity may affect product performance.
- Do not use reagents if the foil has been broken or damaged.
- Do not mix reagents from different pouches and/or kits and/or lots.
- Do not interchange or re-use caps, as contamination may occur and compromise test results.
- Check Unitized Reagent Strips for proper liquid fills (ensure that the liquids are at the bottom of the tubes).
- Check Unitized Reagent Strips to ensure that all pipette tips are present.
- Proceed with caution when using chemical solutions, as Extraction Tube barcode readability may be altered.
- Good laboratory technique is essential to the proper performance of this assay. Extreme care should be taken to preserve the purity of all materials and reagents.
- In cases where other PCR tests are conducted in the same general area of the laboratory, care must be taken to ensure that the BD SARS-CoV-2/Flu components, any additional reagents required for testing, and the BD MAX System are not contaminated. Avoid microbial and ribonuclease (RNase)/deoxyribonuclease (DNase) contamination of reagents at all times. The use of sterile RNase/DNase-free disposable aerosol resistant or positive displacement pipette tips is recommended. Use a new tip for each specimen. Gloves must be changed before manipulating reagents and cartridges.
- To avoid contamination of the environment by amplicons, do not break apart the BD MAX PCR Cartridge after use. The seals of the BD MAX PCR Cartridges are designed to prevent contamination.
- The laboratory should routinely perform environmental monitoring to minimize the risk of cross-contamination.
- Wear protective clothing and disposable gloves while handling all reagents.
- Wash hands thoroughly after performing the test.
- Do not pipette by mouth.
- Do not smoke, drink, chew or eat in areas where specimens or kit reagents are being handled.
- Dispose of unused reagents and waste in accordance with local, state, provincial and/or federal regulations.
- Consult the BD MAX System User's Manual³ for additional warnings, precautions and procedures.

STORAGE

- BD SARS-CoV-2/Flu for BD MAX System components are stable at 2–25 °C through the stated expiration date. Do not use expired components.
- The BD SARS-CoV-2/Flu Master Mix and Extraction Tubes are provided in sealed pouches. To protect from humidity, immediately re-seal after opening.
- Reagent tubes are stable for up to 14 days at 2–25 °C after initial opening and re-sealing of the pouch.

INSTRUCTIONS FOR USE

Swab Specimen Collection/Transport in Universal Viral Transport (UVT) or Universal Transport Medium (UTM)

Note: Wear gloves when handling specimens. If gloves come in contact with the specimen, immediately change them to prevent contamination of other specimens.

1. Nasopharyngeal or anterior nasal swab specimens should be collected and expressed directly into the BD Universal Viral Transport System or the Copan Universal Transport Media System according to their respective package insert instructions.
2. After collection, specimens can be stored for up to 24 hours at 2–25 °C.
3. If delivery and processing of samples exceeds specified time period, specimens should be transported in dry ice and once in laboratory frozen at -70 °C or colder.

BD Molecular Respiratory Sample Buffer Tube Preparation for use with nasopharyngeal or anterior nasal swab specimens in Universal Viral Transport (UVT) or Universal Transport Media (UTM)

Note: Wear gloves when handling specimens. If gloves come in contact with the specimen, immediately change them to prevent contamination of other specimens.

Note: If frozen, allow Universal Viral Transport (UVT) or Universal Transport Media (UTM) specimen to come to room temperature before proceeding.

1. Uncap the BD Molecular Respiratory Sample Buffer Tube and transfer (using a calibrated, variable pipette) 750 µL from the UVT/UTM specimen directly into the BD Molecular Respiratory Sample Buffer Tube.
2. Recap the tube with a blue septum cap and vortex or mix by inversion 8-10 times.
3. Label the BD Molecular Respiratory Sample Buffer Tube with patient information.

Note: Do not obscure the barcodes on the tube. Obscuring the barcode may result in BD MAX System catalog failure and inability to test the sample.

4. Repeat Steps 1 to 3 for each UVT/UTM sample that will be tested on the BD MAX System.
5. Proceed directly with the BD MAX System Operation.

BD MAX System Operation

Note: Refer to the BD MAX System User's Manual³ for detailed instructions (Operation section).

1. Power on the BD MAX System (if not already done) and log in by entering <user name> and <password>.
2. Gloves must be changed before manipulating reagents and cartridges.
3. Remove the required number of Unitized Reagent Strips from the BD SARS-CoV-2/Flu kit. Gently tap each Unitized Reagent Strip onto a hard surface to ensure that all the liquids are at the bottom of the tubes.
4. Remove from the protective pouches the required number of Extraction Tube(s) and Master Mix Tube(s) from the BD SARS-CoV-2/Flu kit.
5. Remove excess air, and close pouches with the zip seal.
6. For each sample to be tested, place one (1) Unitized Reagent Strip on the BD MAX System Rack, starting with Position 1 of Rack A.
7. Snap one (1) Extraction Tube (D4) (white foil) into each Unitized Reagent Strip in Position 1 as shown in Figure 1.
8. Snap one (1) BD SARS-CoV-2/Flu Master Mix tube (D9) (green foil) into each Unitized Reagent Strip in Position 2 as shown in Figure 1.




Figure 1: Snap Extraction Tubes and Master Mix Tube into Unitized Reagent Strips

9. Click on the Run Tab and then the Inventory subtab. Enter the kit lot number for the BD SARS-CoV-2/Flu kit (for lot traceability) by either scanning the barcode with the scanner or by manual entry.

NOTE: Repeat step 9 each time a new kit lot is used.

10. Navigate to the Worklist. Using the pull down menu select <BD SARS CoV2 Flu 74>.
11. Select the appropriate kit lot number (found on the outer box of the BD SARS-CoV-2/Flu kit) from the pull down menu.
12. Enter the BD Molecular Respiratory Sample Buffer Tube ID, Patient ID and Accession Number (if applicable) into the Worklist, either by scanning the barcode with the scanner or by manual entry.
13. Repeat step 12 for all remaining Sample Buffer Tubes.
14. Place the Sample Buffer Tubes in the BD MAX System Rack(s) corresponding to the Unitized Reagent Strips assembled in steps 6 to 8.

15. Place the required number of BD MAX PCR Cartridge(s) into the BD MAX System (refer to Figure 2):
 - Each BD MAX PCR Cartridge accommodates 1 run of up to 12 samples for a total of 12 samples.
 - The BD MAX System will automatically select the position and row on the BD MAX PCR Cartridge for each run.
 - BD MAX PCR Cartridges are used on a per-run AND rack basis (1 run per cartridge and 1 cartridge per rack).
 - To maximize use of BD MAX PCR Cartridges, using 2000 Sample Mode, select Run Wizard under the Worklist tab for lane assignments.
 - Consult the BD MAX System User's Manual³ for more details.



Figure 2: Load BD MAX PCR Cartridges

16. Load rack(s) into the BD MAX System (refer to Figure 3).



Figure 3: Load Rack(s) into the BD MAX System

17. Close the BD MAX System lid and click the **<Start>** to begin processing.

NOTE: When an Indeterminate (IND), Unresolved (UNR), or Incomplete (INC) result is obtained, or when an External Control failure occurs, repeat test from the primary sample (refer to Repeat Test Procedure section).

QUALITY CONTROL

Quality control procedures monitor the performance of the assay. Laboratories must establish the number, type and frequency of testing control materials according to guidelines or requirements of local, provincial, state, federal and/or country regulations or accreditation organizations in order to monitor the effectiveness of the entire analytical process. For general Quality Control guidance, the user may wish to refer to Clinical Laboratory Standards Institute documents MM3⁴ and EP12.⁵

1. External Control materials are not provided by BD. External Positive and Negative Controls are not used by the BD MAX System software for the purpose of sample test result interpretation. External Controls are treated as if they were patient samples. (Refer to Table 2 for the interpretation of External Control assay results.)
2. One (1) External Positive Control and one (1) External Negative Control should be run at least daily until adequate process validation is achieved on the BD MAX System in each laboratory setting. Reduced frequency of control testing should be in accordance with applicable regulations.
3. The External Positive Control is intended to monitor for substantial reagent failure. The External Negative Control is intended to detect reagent or environmental contamination (or carry-over) by target nucleic acids.
4. Various types of External Controls are recommended to allow the user to select the most appropriate for their laboratory quality control program.

- a. External Negative Control (ENC): Commercially available control material, such as Microbiology[®] Helix Elite™ Inactivated Standard Negative Cellularity Control (refer to Table 1), or a previously characterized sample known to be negative. BD recommends that the External Negative Control be prepared prior to the External Positive Control in order to reduce the potential for contamination as a result of control preparation.
- b. External Positive Control (EPC): Commercially available control materials, such as the Microbiology[®] Helix Elite™ standards listed below (refer to Table 1), or previously characterized samples known to be positive.

Table 1: Commercially Available Standards for External Controls

Commercially Available Standards	Part Number
Microbiology [®] Helix Elite™ Synthetic Standard SARS-CoV-2 Synthetic RNA (N gene Targets)	HE0060S
Microbiology [®] Helix Elite™ Inactivated Standard Inactivated Influenza A/B and Respiratory Syncytial Virus	HE0044N
Microbiology [®] Helix Elite™ Inactivated Standard Negative Cellularity Control	HE0058N

5. Suggested procedure for preparing an EPC or ENC using Microbiology[®] Helix Elite™ Standards (see below) has been verified by BD. However, the choice of EPC and ENC for the BD SARS-CoV-2/Flu for BD MAX System is ultimately the decision of the laboratory, in accordance with applicable local, state, and/or federal regulations, accreditation requirements, and the laboratory's standard Quality Control (QC) procedures.
6. Preparation of Microbiology[®] Helix Elite™ Inactivated Standard Negative Cellularity Control as an External Negative Control:
 - a. Add 750 µL of nuclease free water into a BD Molecular Respiratory Sample Buffer Tube.
 - b. Rehydrate the Microbiology[®] Negative Cellularity Control Standard with 100 µL of nuclease free water.
 - c. Dilute the rehydrated standard 1:10 in nuclease free water (10 µL standard to 90 µL nuclease free water).
 - d. Spike 75 µL of the diluted standard into the Sample Buffer Tube.
 - e. Cap the External Negative Control Sample Buffer Tube and vortex for 10-30 seconds or invert 8-10 times. Process on the BD MAX System.
7. Preparation of Microbiology[®] Helix Elite™ Standards as an External Positive Control:
 - a. Add 750 µL of nuclease free water into a BD Molecular Respiratory Sample Buffer Tube.
 - b. Rehydrate the Microbiology[®] Negative Cellularity Control Standard, Microbiology[®] SARS-CoV-2 Synthetic RNA Standard, and Microbiology[®] Inactivated Influenza A/B and Respiratory Syncytial Virus Standard each with 100 µL of nuclease free water.
 - c. Dilute the Microbiology[®] Negative Cellularity Control Standard and Microbiology[®] Inactivated Influenza A/B and Respiratory Syncytial Virus Standard 1:10 in nuclease free water (10 µL standard to 90 µL nuclease free water).
 - d. Dilute the Microbiology[®] SARS-CoV-2 Synthetic RNA Standard 1:100 in nuclease free water (10 µL standard to 990 µL nuclease free water).
 - e. Spike 75 µL of the diluted Microbiology[®] Negative Cellularity Control Standard, 50 µL of the diluted Microbiology[®] SARS-CoV-2 Synthetic RNA Standard, and 50 µL of the diluted Microbiology[®] Inactivated Influenza A/B and Respiratory Syncytial Virus Standard into the Sample Buffer Tube.
 - f. Cap the External Positive Control Sample Buffer Tube and vortex for 10-30 seconds or invert 8-10 times. Process on the BD MAX System.
8. Preparation of previously characterized nasopharyngeal specimen in UVT/UTM as an External Positive or Negative Control:
 - a. Transfer 750 µL of the specimen to a BD Molecular Respiratory Sample Buffer Tube.
 - b. Cap the External Control Sample Buffer Tube and vortex for 10-30 seconds or invert 8-10 times. Process on the BD MAX System.
9. All External Controls should yield the expected results (Table 2) with no failed external controls (Unresolved, Indeterminate, Incomplete results).

Table 2: BD SARS-CoV-2/Flu External Control Expected Results

Control Type	Control	Used to Monitor	Expected Result		
			CoV2 (SARS-CoV-2)	Flu A	Flu B
Negative External Control	Known Negative Specimen	Reagent and/or environmental contamination	NEG	NEG	NEG
	Microbiology Negative External Control				
Positive External Control	Known Positive Specimen*	Substantial reagent failure including primer and probe integrity	POS/NEG	POS/NEG	POS/NEG
	Microbiology Positive External Control				

*Known Positive Specimens are expected to be positive only for the virus(es) present in the specimen.

10. An External Negative Control that yields a positive test result is indicative of a specimen handling and/or contamination event. Review the specimen handling technique to avoid mix-up and/or contamination. An External Positive Control that yields a negative result is indicative of a specimen handling/preparation problem. Review the specimen handling/preparation technique.
11. An External Control that yields an Unresolved, Indeterminate, or Incomplete test result is indicative of a reagent or a BD MAX System failure. Check the BD MAX System monitor for any error messages. Refer to the Troubleshooting section of the BD MAX System User's Manual³ for interpretation of warning and error codes. If the problem persists, use reagents from an unopened pouch or use a new BD SARS-CoV-2/Flu for BD MAX System kit.
12. The RNase P gene serves as both an Extraction and Internal Amplification Control. In the event that SARS-CoV-2, influenza A, and influenza B are negative, an RNase P result must be positive for the SARS-CoV-2, influenza A, and influenza B results to be valid negative results. When either SARS-CoV-2, influenza A, and/or influenza B target results are positive, RNase P result is ignored. An Unresolved (UNR) result is indicative of specimen-associated inhibition or reagent failure. Repeat any sample reported as Unresolved according to the "Repeat Test Procedure" section below.

RESULT INTERPRETATION

Results are available on the **<Results>** tab in the **<Results>** window on the BD MAX System monitor. The BD MAX System software automatically interprets test results. Results are reported for each of the analytes. A test result may be called as NEG (negative), POS (positive) or UNR (unresolved) based on the amplification status of the target and the Extraction and Internal Amplification Control, RNase P. IND (Indeterminate) or INC (Incomplete) results are due to BD MAX System failure. BD SARS-CoV-2/Flu for BD MAX System results interpretation is described below in Table 3.

Table 3: BD SARS-CoV-2/Flu for BD MAX System Result Interpretation

CoV2 (SARS-CoV-2)	FluA	FluB	Result Displayed ^a	Actions
Pos	Neg	Neg	CoV2 POS FluA NEG FluB NEG	Report as: SARS-CoV-2 Detected Influenza A Not Detected Influenza B Not Detected
Neg	Pos	Neg	CoV2 NEG FluA POS FluB NEG	Report as: SARS-CoV-2 Not Detected Influenza A Detected Influenza B Not Detected
Neg	Neg	Pos	CoV2 NEG FluA NEG FluB POS	Report as: SARS-CoV-2 Not Detected Influenza A Not Detected Influenza B Detected
Pos	Pos	Neg	CoV2 POS FluA POS FluB NEG	Report as: SARS-CoV-2 Detected Influenza A Detected Influenza B Not Detected
Pos	Neg	Pos	CoV2 POS FluA NEG FluB POS	Report as: SARS-CoV-2 Detected Influenza A Not Detected Influenza B Detected
Neg	Pos	Pos	CoV2 NEG FluA POS FluB POS	Report as: SARS-CoV-2 Not Detected Influenza A Detected Influenza B Detected
Pos	Pos	Pos	CoV2 POS FluA POS FluB POS	Report as: SARS-CoV-2 Detected Influenza A Detected Influenza B Detected
Neg	Neg	Neg	CoV2 NEG FluA NEG FluB NEG	Report as: SARS-CoV-2 Not Detected Influenza A Not Detected Influenza B Not Detected
			UNR ^b	Repeat Test ^c
			IND ^d (with Warning or Error Codes ^e)	Repeat Test ^c
			INC ^f (with Warning or Error Codes ^e)	Repeat Test ^c

^a Laboratories should report their diagnostic result as appropriate and in compliance with their specific reporting system. Laboratories within the United States and its territories are required to report all SARS-CoV-2 results to the appropriate public health authorities.

^b Unresolved

^c Repeat Test by preparing a fresh sample buffer tube from the primary UVT/UTM sample.

^d Indeterminate

^e Refer to Troubleshooting section of the BD MAX System User's Manual³ for interpretation of warning and error codes.

^f Incomplete

UNRESOLVED, INDETERMINATE, AND INCOMPLETE RESULTS

When an Indeterminate (IND), Unresolved (UNR), or Incomplete (INC) result is obtained, a repeat test from the primary sample must be performed. If an External Control fails, repeat testing of all specimens conducted on the same day using freshly prepared External Controls (see Quality Control).

Unresolved Result

Unresolved results may be obtained in the event that specimen-associated inhibition or reagent failure prevents proper target or RNase P amplification. Sample(s) can be repeated from the primary sample. Uncap a new BD Molecular Respiratory Sample Buffer Tube and transfer (using a calibrated, variable pipette) 750 µL from the UVT/UTM specimen directly into the BD Molecular Respiratory Sample Buffer Tube. Restart from the BD MAX System Operation section.

Indeterminate Result

Indeterminate results may be obtained in the event that a System failure occurs. Sample(s) can be repeated from the primary sample. Uncap a new BD Molecular Respiratory Sample Buffer Tube and transfer (using a calibrated, variable pipette) 750 µL from the UVT/UTM specimen directly into the BD Molecular Respiratory Sample Buffer Tube. Restart from the BD MAX System Operation section.

Incomplete Result

Incomplete results may be obtained in the event that Specimen Preparation or the PCR did not reach its expected time points. Sample(s) can be repeated from the primary sample. Uncap a new BD Molecular Respiratory Sample Buffer Tube and transfer (using a calibrated, variable pipette) 750 µL from the UVT/UTM specimen directly into the BD Molecular Respiratory Sample Buffer Tube. Restart from the BD MAX System Operation section.

External Control Failure

External Controls should yield expected results when tested. If samples have to be repeated due to an incorrect External Control result, the samples should be repeated from the primary sample along with freshly prepared External Controls. Restart from the BD MAX System Operation section.

LIMITATIONS OF THE PROCEDURE

- BD SARS-CoV-2/Flu for BD MAX System has been evaluated only for use on the BD MAX System.
- Reliable results depend on proper sample collection, storage, and handling procedures.
- Performance of BD SARS-CoV-2/Flu for BD MAX System has only been established in nasopharyngeal swab specimens. Anterior nasal swabs are considered an acceptable specimen type for use with the BD SARS-CoV-2/Flu for BD MAX System, but performance with this specimen type has not been established.
- Use of BD SARS-CoV-2/Flu for BD MAX System with other specimen types has not been assessed and performance characteristics are unknown.
- The clinical performance has not been established in all circulating variants but is anticipated to be reflective of the prevalent variants in circulation at the time and location of the clinical evaluation. Performance at the time of testing may vary depending on the variants circulating, including newly emerging strains of SARS-CoV-2 and their prevalence, which change over time.
- Detection of SARS-CoV-2, influenza A, and/or influenza B RNA may be affected by sample collection methods, patient factors (e.g., presence of symptoms), and/or stage of infection.
- As with any molecular test, mutations within the target regions of the BD SARS-CoV-2/Flu for BD MAX System test could affect primer and/or probe binding resulting in failure to detect the presence of virus.
- Due to inherent differences between technologies, it is recommended that, prior to switching from one technology to the next, users perform method correlation studies in their laboratory to qualify technology differences. One hundred percent agreement between the results should not be expected due to aforementioned differences between technologies. Users should follow their own specific policies/procedures.
- False negative or invalid results may occur due to interference. The RNase P endogenous control is included to help identify the specimens containing substances that may interfere with nucleic acid isolation and PCR amplification.
- Good laboratory practices and careful adherence to the procedures specified in this Instructions For Use document are necessary to avoid contamination of reagents.
- The effect of interfering substances has only been evaluated for those listed in this labeling. Potential interference has not been evaluated for substances other than those described in the Interfering Substances section below. Interference by substances other than those described in the Interfering Substances section below could lead to erroneous results.
- Human blood, Flonase, Zicam and tobramycin were found to interfere with BD SARS-CoV-2/Flu for BD MAX System at concentrations greater than 0.2% v/v, 1.7% v/v, 0.5% v/v and 0.4 µg/mL in UVT, respectively.
- BD SARS-CoV-2/Flu for BD MAX System has not been evaluated for patients receiving intranasally administered influenza vaccine.
- The performance of this device has not been assessed in a population vaccinated against COVID-19.
- Enterovirus C (Coxsackievirus A17) was shown to interfere with BD SARS-CoV-2/Flu for BD MAX System at a concentration above 1.00E+04 TCID₅₀/mL in UVT.
- Results from analytical studies with contrived co-infected samples showed potential for competitive interference of influenza B at low concentrations (~2x LoD) when SARS-CoV-2 concentration is ≥1.00E+06 genomic copies/mL.
- The test is not intended to differentiate influenza A subtypes or influenza B lineages. If differentiation of specific influenza subtypes and lineages is needed, additional testing, in consultation with state or local public health departments, is required.

CONDITIONS OF AUTHORIZATION FOR THE LABORATORY

The BD SARS-CoV-2/Flu for BD MAX System Letter of Authorization, along with the authorized Fact Sheet for Healthcare Providers, the authorized Fact Sheet for Patients, and authorized labeling are available on the FDA website: <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/vitro-diagnostics-euas>. However, to assist clinical laboratories using the BD SARS-CoV-2/Flu for BD MAX System, the relevant Conditions of Authorization are listed below.

- Authorized laboratories¹ using the BD SARS-CoV-2/Flu for BD MAX System must include with result reports all authorized Fact Sheets. Under exigent circumstances, other appropriate methods for disseminating these Fact Sheets may be used, which may include mass media.
- Authorized laboratories using BD SARS-CoV-2/Flu for BD MAX System must perform BD SARS-CoV-2/Flu for BD MAX System as outlined in the authorized labeling. Deviations from the authorized procedures, including the authorized instruments, authorized extraction methods, authorized clinical specimen types, authorized control materials, authorized ancillary reagents, and authorized material required to perform the BD SARS-CoV-2/Flu for BD MAX System test are not permitted.
- Authorized laboratories that receive the BD SARS-CoV-2/Flu for BD MAX System test must notify the relevant public health authorities of their intent to run the test prior to initiating testing.
- Authorized laboratories using the BD SARS-CoV-2/Flu for BD MAX System test must have a process in place for reporting test results to healthcare providers and relevant public health authorities, as appropriate.
- Authorized laboratories must collect information on the performance of the test and report to DMD/OHT7-OIR/OPEQ/CDRH (via email: CDRH-EUA-Reporting@fda.hhs.gov) and to Becton, Dickinson and Company Customer Technical Support 1.800.638.8663 any suspected occurrence of false positive or false negative results and significant deviations from the established performance characteristics of the BD SARS-CoV-2/Flu for BD MAX System test of which they become aware.
- All laboratory personnel using the BD SARS-CoV-2/Flu for BD MAX System test must be appropriately trained in RT-PCR techniques, use appropriate laboratory, and personal protective equipment when handling this kit and use the test in accordance with the authorized labeling.
- Becton, Dickinson and Company, its authorized distributors, and authorized laboratories using the BD SARS-CoV-2/Flu for BD MAX System must ensure that any records associated with this EUA are maintained until otherwise notified by FDA. Such records will be made available to FDA for inspection upon request.

¹The letter of authorization refers to, "Laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet requirements to perform moderate or high complexity tests" as "authorized laboratories".

NON-CLINICAL PERFORMANCE EVALUATION

Limit of Detection (LoD)

The analytical sensitivity of the BD SARS-CoV-2/Flu for BD MAX System was assessed with limiting dilution for seven respiratory viruses (SARS-CoV-2, two strains of influenza A H1N1, two strains of influenza A H3N2, and two strains of influenza B). To confirm co-spiking of analytes does not impact analytical sensitivity, LoD was determined with co-spiked analytes, one strain per analyte. LoD studies determine the lowest detectable concentration of virus at which at least 95% of all (true positive) replicates test positive.

To determine the LoD, quantified inactivated SARS-CoV-2 or quantified influenza culture fluids were serially diluted into simulated nasopharyngeal matrix for a total of 5 concentrations levels with 2-fold serial dilutions between each level. Confirmation of the estimated LoD was performed with one reagent lot in replicates of 20 prepared in simulated nasopharyngeal matrix. The LoD is defined as the lowest concentration at which ≥95% of all replicates are expected to test positive. The verified LoD values for each virus tested are summarized in Table 4.

Table 4. BD SARS-CoV-2/Flu for BD MAX System Limit of Detection

Strain ID	LoD Concentration (in UVT)
SARS-CoV-2/USA-WA1/2020 ^a	700 GC/mL
Influenza A H1N1 Brisbane/59/07 ^b	0.025 TCID ₅₀ /mL
Influenza A H1N1 Idaho/07/2018 ^b	0.20 TCID ₅₀ /mL
Influenza A H3N2 Switzerland/9715293/13 ^b	0.10 TCID ₅₀ /mL
Influenza A H3N2 Kansas/14/2017 ^a	4.8 TCID ₅₀ /mL
Influenza B Colorado/06/17 ^a	0.05 TCID ₅₀ /mL
Influenza B Phuket/3073/13 ^b	0.06 TCID ₅₀ /mL

^a Co-spiked analyte LoD

^b Individual analyte LoD

Reactivity/Inclusivity

The N1 and N2 primers and probes utilized for SARS-CoV-2 detection within the BD SARS-CoV-2/Flu for BD MAX System are identical in sequence to those reported in the CDC 2019-Novel Coronavirus (2019-nCoV) Real-Time RT-PCR Diagnostic Panel. An *in silico* comparison of the N1 and N2 primer sets was performed using all available high quality SARS-CoV-2 sequences submitted to the GISAID EpiCoV database⁶ by January 13, 2021 (n=329,434). Alignments against the N gene showed that both N1 and N2 primer/probe sets are a perfect match to 93.8% of sequences in the database, 96.8% of the sequences were a perfect match to the N1 primer set region, and 97.0% were a perfect match to the N2 primer set region. In total, 99.9% are a perfect match to either the N1 or the N2 region primer sets.

An *in silico* comparison of the influenza A primer set was performed using all available high quality Influenza A M1 (matrix protein) gene sequences submitted to the GISAID EpiCoV database⁶ between May 1, 2008 and October 21, 2020 (n=87,051). Multiple alignment of the matrix gene showed that 90.2% of sequences are a perfect match to the primer/probe set while an additional 7.8% of sequences have a single base mismatch in the 5' end of a single primer. Multiple mismatches to the primers and probe occurred in only 0.25% of sequences.

An *in silico* comparison of the influenza B primer sets was performed using all available high quality Influenza B M1 gene and HA gene sequences submitted to the GISAID EpiCoV database⁶ between May 1, 2008 and October 21, 2020. A total of 23,972 matrix and 49,852 HA sequences were used in this analysis. Multiple alignment of the M1 gene showed that 97.2% of sequences are a perfect match to the primer/probe set and 74.8% of HA sequences had one or fewer base pair mismatches.

BD SARS-CoV-2/Flu for BD MAX System was evaluated against multiple strains of influenza A H1N1 and H3N2 and influenza B strains including both the Yamagata and Victoria lineages. A total of 20 influenza A and 5 influenza B strains were evaluated at levels near the analytical LoD. Three replicates were tested for each strain. All strains were detected at 3x LoD except for one influenza A H1N1 strain (A/Wisconsin/505/2018 pdm09) and one influenza A H3N2 strain (A/Texas/71/2017), which were detected at 6x LoD.

Table 5. Analytical Reactivity/Inclusivity for BD MAX SARS-CoV-2/Flu for BD MAX System

Virus	Strain	Type	Virus Concentration in UVT	Virus Concentration Relative to LoD	CoV2 (SARS-CoV-2) Result	Flu A Result	Flu B Result
Influenza A	H1N1	A/Maryland/08/2013 (H1N1) pdm09 Antiviral Resistance	1158.93 TCID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Bangladesh/3002/2015 (H1N1) pdm09	2313.38 CEID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Iowa/53/2015 (H1N1) pdm09	71.33 TCID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Michigan/272/2017 (H1N1) pdm09	39.7 TCID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Wisconsin/505/2018 (H1N1) pdm09	42.5 TCID ₅₀ /mL	6x LoD	NEG	POS	NEG
		A/St. Petersburg/61/2015 (H1N1) pdm09	2600.9 CEID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Michigan/45/2015 (H1N1) pdm09	5748.7 CEID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Louisiana/08/2013 (H1N1) pdm09 Antiviral Resistance	220.0 TCID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/North Carolina/4/2014 (H1N1) pdm09 Antiviral Resistance	2002.3 CEID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/New York/18/2009 (H1N1) pdm09 Antiviral Resistance	14.5 TCID ₅₀ /mL	3x LoD	NEG	POS	NEG
	H3N2	A/California/02/2014 (H3N2)	3.1 TCID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Alaska/232/2015 (H3N2)	1650.9 CEID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Singapore/INFIMH-16-0019/2016 (H3N2)	42.2 FFU ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Texas/71/2017 (H3N2)	60.9 FFU ₅₀ /mL	6x LoD	NEG	POS	NEG
		A/Arizona/45/2018 (H3N2)	173.5 FFU ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Hong Kong/4801/14 (H3N2)	0.8 TCID ₅₀ /mL	3x LoD	NEG	POS	NEG
		A/Norway/466/14 (H3N2)	0.1 U/mL	3x LoD	NEG	POS	NEG
		A/South Australia/55/14 (H3N2)	0.2 U/mL	3x LoD	NEG	POS	NEG
		A/Stockholm/6/14 (H3N2)	0.1 U/mL	3x LoD	NEG	POS	NEG
		A/Wisconsin/04/2018 (H3N2)	2793.9 CEID ₅₀ /mL	3x LoD	NEG	POS	NEG
Influenza B	Victoria	B/Maryland/15/2016	1.34 TCID ₅₀ /mL	3x LoD	NEG	NEG	POS
		B/Hong Kong/286/2017	0.63 TCID ₅₀ /mL	3x LoD	NEG	NEG	POS
		B/Hawaii/01/2018 (NA D197N)	236.1 TCID ₅₀ /mL	3x LoD	NEG	NEG	POS
	Yamagata	B/Guangdong-Liwan/1133/2014	1208.8 CEID ₅₀ /mL	3x LoD	NEG	NEG	POS
		B/Oklahoma/10/2018 (NA D197N)	491.5 TCID ₅₀ /mL	3x LoD	NEG	NEG	POS

In addition, the BD SARS-CoV-2/Flu was shown to be inclusive for the CDC Human Influenza Virus Panel (2020). The lowest concentration where at least one out of five replicates are positive from a 5-fold dilution scheme is reported as the minimum reactive concentration.

Table 6. CDC Human Influenza Virus Panel (2020) Results

Influenza Virus (Type/Subtype)	Virus strain name	Minimum Reactive Concentration (EID ₅₀ /mL)
Influenza A/H3N2	A/Perth/16/2009	1.02E+03
	A/Hong Kong 2671 2019	4.05E+02
Influenza A/H1N1	A/Christ Church/16/2010	8.11E+03
	A/Guangdongmaonan/1536/2019	6.45E+02
Influenza B/Victoria lineage	B/Michigan/09/2011	8.13E-01
	B/Washington/02/2019	1.62E+02
Influenza B/Yamagata lineage	B/Texas/81/2016	1.02E+02
	B/Phuket 3073/2013	8.13E+02

Cross-Reactivity

An *in silico* analysis was performed to evaluate the potential for all primers and probes contained within the BD SARS-CoV-2/Flu for BD MAX System master mix to amplify and detect unintended organisms. Each primer was 'BLAST' against the full nt database and alignments were kept if there were no more than three (3) base pair mismatches across the length of the primer, the 3' end of the primer matched the subject sequence, and no gaps were introduced to "force" an alignment. The plus/minus orientation between the primer (query) and the subject (database sequence) was determined, and all two-primer combinations (including each primer with itself) were identified where one primer matched the plus strand and the other matched the minus, representing potential amplicons. Amplicons were kept if the minus strand primer was downstream of the plus strand primer and the resulting amplicons were less than or equal to 3,000 base pairs long.

Influenza A: No relevant cross-reactivity was discovered.

Influenza B: No relevant cross-reactivity was discovered.

SARS-CoV-2: All identified hits are either SARS-CoV-2 or a closely related coronavirus from non-human species. No relevant cross-reactivity was discovered.

Additionally, 46 organisms and 1 nasopharyngeal pool were evaluated for cross-reactivity with the BD SARS-CoV-2/Flu for BD MAX System. The bacterial cells, yeasts, and viruses were tested in the BD Molecular Respiratory Sample Buffer Tube. All organisms tested produced negative results when tested at the concentrations in Table 7.

Table 7. Cross-Reactivity Testing Results

Organism	Concentration of Organism in Sample Buffer Tube	Negative Results (Negative/Total)
Adenovirus - type 1	1.00E+05 TCID ₅₀ /mL	3/3
Adenovirus - type 4	1.00E+05 TCID ₅₀ /mL	3/3
Adenovirus - type 7	1.00E+05 TCID ₅₀ /mL	3/3
<i>Bordetella pertussis</i>	1.00E+06 CFU/mL	3/3
<i>Candida albicans</i>	1.00E+06 CFU/mL	3/3
<i>Chlamydia pneumonia</i>	1.00E+06 IFU/mL	3/3
<i>Corynebacterium diphtheriae</i>	1.00E+06 CFU/mL	3/3
Cytomegalovirus	4.17E+04 U/mL	3/3
Enterovirus B (Echovirus 6)	1.00E+05 U/mL	3/3
Enterovirus C (Coxsackievirus A17)	1.00E+05 TCID ₅₀ /mL	3/3
Enterovirus D	1.00E+05 U/mL	3/3
Epstein Barr virus	1.00E+05 copies/mL	3/3
<i>Escherichia coli</i>	1.00E+06 CFU/mL	3/3
<i>Haemophilus influenzae</i>	1.00E+06 CFU/mL	3/3
Herpes simplex virus Type 1	1.41E+04 U/mL	3/3
Herpes simplex virus Type 2	1.41E+04 U/mL	3/3
Human coronavirus 229E	1.00E+05 U/mL	3/3
Human coronavirus HKU1 ^a	1.00E+06 GC/mL	3/3
Human coronavirus NL63	1.41E+04 TCID ₅₀ /mL	3/3
Human coronavirus OC43	1.00E+05 TCID ₅₀ /mL	3/3
Human Metapneumovirus (hMPV)	1.00E+05 TCID ₅₀ /mL	3/3
<i>Lactobacillus acidophilus</i>	1.00E+06 CFU/mL	3/3
<i>Legionella pneumophila</i>	1.00E+06 CFU/mL	3/3
Measles	1.00E+05 U/mL	3/3

Organism	Concentration of Organism in Sample Buffer Tube	Negative Results (Negative/Total)
MERS-coronavirus ^a	1.00E+05 copies/mL	3/3
<i>Moraxella catarrhalis</i>	1.00E+06 CFU/mL	3/3
Mumps	1.00E+05 U/mL	3/3
<i>Mycobacterium tuberculosis</i> ^a	1.00E+06 copies/mL	3/3
<i>Mycoplasma pneumoniae</i>	1.00E+06 CFU/mL	3/3
<i>Neisseria meningitidis</i>	5.00E+03 CFU/mL	3/3
<i>Neisseria gonorrhoeae</i>	1.00E+06 CFU/mL	3/3
Parainfluenza virus 1	1.00E+05 TCID ₅₀ /mL	3/3
Parainfluenza virus 2	1.00E+05 U/mL	3/3
Parainfluenza virus 3	1.00E+05 TCID ₅₀ /mL	3/3
Parainfluenza virus 4	1.00E+05 TCID ₅₀ /mL	3/3
<i>Pneumocystis jirovecii</i> (PJP)	1.00E+05 nuclei/mL	3/3
Pooled human expressed nasopharyngeal swab matrix	n/a	3/3
<i>Pseudomonas aeruginosa</i>	1.00E+06 CFU/mL	3/3
Respiratory syncytial virus	1.00E+05 U/mL	3/3
Rhinovirus	1.00E+05 TCID ₅₀ /mL	3/3
SARS-coronavirus ^a	1.00E+05 GE/mL	3/3
<i>Staphylococcus aureus</i>	1.00E+06 CFU/mL	3/3
<i>Staphylococcus epidermis</i>	1.00E+06 CFU/mL	3/3
<i>Streptococcus pneumoniae</i>	1.00E+06 CFU/mL	3/3
<i>Streptococcus pyogenes</i>	1.00E+06 CFU/mL	3/3
<i>Streptococcus salivarius</i>	5.00E+03 CFU/mL	3/3
Varicella-zoster virus	1.00E+04 U/mL	3/3

^a Genomic DNA or RNA tested

Microbial Interference

Forty-six (46) organisms and one nasopharyngeal pool were evaluated for potential interference with the BD SARS-CoV-2/Flu for BD MAX System. Organisms were tested at high concentration ($\geq 10^5$ CFU/mL, cells or genome equivalents/mL, $\geq 10^5$ IFU/mL or TCID₅₀/mL, or highest concentration available) in the presence of assay analytes (SARS-CoV-2, influenza A and influenza B) co-spiked at 3x LoD. Enterovirus C (Coxsackievirus A17) was shown to interfere in the detection of influenza A and influenza B at concentrations above 1.00E+04 TCID₅₀/mL.

Table 8. Microbial Interference Testing Results

Organism	Concentration of Organism in Sample Buffer Tube	CoV2 (SARS-CoV-2)	FluA	Flu B
		Positive/Total		
Adenovirus - type 1	1.00E+05 TCID ₅₀ /mL	3/3	3/3	3/3
Adenovirus - type 4	1.00E+05 TCID ₅₀ /mL	3/3	3/3	3/3
Adenovirus - type 7	1.00E+05 TCID ₅₀ /mL	3/3	3/3	3/3
<i>Bordetella pertussis</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
<i>Candida albicans</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
<i>Chlamydia pneumoniae</i>	1.00E+06 IFU/mL	3/3	3/3	3/3
<i>Corynebacterium diphtheriae</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
Cytomegalovirus	4.17E+04 U/mL	3/3	3/3	3/3
Enterovirus B (Echovirus 6)	1.00E+05 U/mL	3/3	3/3	3/3
Enterovirus C (Coxsackievirus A17)	1.00E+05 TCID ₅₀ /mL	3/3	1/3	0/3
	1.00E+04 TCID ₅₀ /mL	3/3	3/3	3/3
Enterovirus D	1.00E+05 U/mL	3/3	3/3	3/3
Epstein Barr virus	1.00E+05 copies/mL	3/3	3/3	3/3
<i>Escherichia coli</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
<i>Haemophilus influenzae</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
Herpes simplex virus Type 1	1.41E+04 U/mL	3/3	3/3	3/3
Herpes simplex virus Type 2	1.41E+04 U/mL	3/3	3/3	3/3
Human coronavirus 229E	1.00E+05 U/mL	3/3	3/3	3/3
Human coronavirus HKU1 ^a	1.00E+06 GC/mL	3/3	3/3	3/3
Human coronavirus NL63	1.41E+04 TCID ₅₀ /mL	3/3	3/3	3/3
Human coronavirus OC43	1.00E+05 TCID ₅₀ /mL	3/3	3/3	3/3
Human Metapneumovirus (hMPV)	1.00E+05 TCID ₅₀ /mL	3/3	3/3	3/3
<i>Lactobacillus acidophilus</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
<i>Legionella pneumophila</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
Measles	1.00E+05 U/mL	3/3	3/3	3/3
MERS-coronavirus ^a	1.00E+05 copies/mL	3/3	3/3	3/3
<i>Moraxella catarrhalis</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
Mumps	1.00E+05 U/mL	3/3	3/3	3/3

Organism	Concentration of Organism in Sample Buffer Tube	CoV2 (SARS-CoV-2)	FluA	Flu B
		Positive/Total		
<i>Mycobacterium tuberculosis</i> ^a	1.00E+06 copies/mL	3/3	3/3	3/3
<i>Mycoplasma pneumoniae</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
<i>Neisseria meningitidis</i>	5.00E+03 CFU/mL	3/3	3/3	3/3
<i>Neisseria gonorrhoeae</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
Parainfluenza virus 1	1.00E+05 TCID ₅₀ /mL	3/3	3/3	3/3
Parainfluenza virus 2	1.00E+05 U/mL	3/3	3/3	3/3
Parainfluenza virus 3	1.00E+05 TCID ₅₀ /mL	3/3	3/3	3/3
Parainfluenza virus 4	1.00E+05 TCID ₅₀ /mL	3/3	3/3	3/3
<i>Pneumocystis jirovecii</i> (PJP)	1.00E+05 nuclei/mL	3/3	3/3	3/3
Pooled human expressed nasopharyngeal swab matrix	n/a	3/3	3/3	3/3
<i>Pseudomonas aeruginosa</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
Respiratory syncytial virus	1.00E+05 U/mL	3/3	3/3	3/3
Rhinovirus	1.00E+05 TCID ₅₀ /mL	3/3	3/3	3/3
SARS-coronavirus ^a	1.00E+05 GE/mL	3/3	3/3	3/3
<i>Staphylococcus aureus</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
<i>Staphylococcus epidermis</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
<i>Streptococcus pneumoniae</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
<i>Streptococcus pyogenes</i>	1.00E+06 CFU/mL	3/3	3/3	3/3
<i>Streptococcus salivarius</i>	5.00E+03 CFU/mL	3/3	3/3	3/3
Varicella-zoster virus	1.00E+04 U/mL	3/3	3/3	3/3

^a Genomic DNA or RNA tested

Mixed Infection / Competitive Interference

To access potential competitive interference between SARS-CoV-2, influenza A, and influenza B, samples were tested in replicates of five (5) where low (approximately 2x their respective LoD) concentration of two analytes were mixed with high (approximately 1.00E+06 TCID₅₀/mL in UVT or 1.00E+06 GC/mL in UVT) concentration of the other analyte in simulated nasopharyngeal matrix.

Results of the study demonstrated that SARS-CoV-2 at 1.00E+06 GC/mL in UVT inhibits influenza B detection when present at low concentration (~2x LoD) in a sample. Inhibition was not observed when SARS-CoV-2 concentration was diluted to 1E+05 GC/mL in UVT.

Table 9: Mixed Infection / Competitive Interference Results

Co-Infection Condition	Virus 1 (High)		Virus 2		Virus 3		Positive Results		
	Description	Concentration	Description	Concentration	Description	Concentration	CoV2 (SARS-CoV-2)	Flu A	Flu B
1	SARS-CoV-2 (USA-WA1/2020)	1.00E+06 GC/mL	Influenza A (Kansas/14/17)	9.6 TCID ₅₀ /mL	Influenza B (Colorado/06/17)	0.10 TCID ₅₀ /mL	5/5	5/5	4/5
2	Influenza A (Michigan/45/2015 (H1N1) pdm09)	1.00E+06 TCID ₅₀ /mL	SARS-CoV-2 (USA-WA1/2020)	1400 GC/mL	Influenza B (Colorado/06/17)	0.10 TCID ₅₀ /mL	5/5	5/5	5/5
3	Influenza B (Guangdong-Liwan/1133/2014)	1.00E+06 TCID ₅₀ /mL	SARS-CoV-2 (USA-WA1/2020)	1400 GC/mL	Influenza A (Kansas/14/17)	9.6 TCID ₅₀ /mL	5/5	5/5	5/5
4	SARS-CoV-2 (USA-WA1/2020)	1.00E+05 GC/mL	Influenza B (Colorado/06/17)	0.10 TCID ₅₀ /mL	N/A		5/5	N/A	5/5

Interfering Substances

Nine (9) biological and chemical substances that may be present in nasopharyngeal swab specimens were evaluated for potential interference with the BD SARS-CoV-2/Flu for BD MAX System in the absence and presence of assay analytes (SARS-CoV-2, influenza A and influenza B). Whole human blood was found to interfere at levels above 0.2% volume/volume. Flonase was found to interfere at levels above 1.7% volume/volume. Tobramycin was found to interfere at levels above 0.4 µg/mL. Zicam was found to interfere at levels above 0.5% volume/volume. Results demonstrated no reportable interference with any other substance tested (refer to Table 8).

Table 10: Endogenous and Commercial Exogenous Substances Tested with BD SARS-CoV-2/Flu for BD MAX System

Brand Name or Description	Active Ingredient	Concentration Tested	Positive Testing (Positive/Total)			Negative Testing (Negative/Total)	Result
			CoV2 (SARS-CoV-2)	FluA	FluB		
Mucin	Purified Mucin	60 ug/mL	3/3	3/3	3/3	3/3	NI
Whole Human Blood	N/A	2% v/v	2/3 ^a	2/3 ^a	2/3 ^a	3/3	I
		0.2% v/v	3/3	3/3	3/3	3/3	NI
Nasal corticosteroids - Flonase	Fluticasone	17% v/v	1/3	0/3	1/3	3/3	I
		1.7% v/v	3/3	3/3	3/3	3/3	NI
Nasal gel – Zicam	Galphimia glauca, luffa operculata, sabadilla	5% v/v	2/3 ^b	2/3 ^b	2/3 ^b	3/3	I
		0.5% v/v	3/3	3/3	3/3	3/3	NI
Homeopathic allergy relief medicine – Afrin	Oxy metazoline hydrochloride	8% v/v	3/3	3/3	3/3	3/3	NI
Throat lozenges, oral anesthetic and analgesic - Cepacol	Benzocaine, Menthol	0.8 mg/mL	3/3	3/3	3/3	3/3	NI
Anti-viral drugs (Relenza)	Zanamivir	3.3 mg/mL	3/3	3/3	3/3	3/3	NI
Antibiotic, nasal ointment (Mupirocin)	Mupirocin	10 mg/mL	3/3	3/3	3/3	3/3	NI
Antibacterial, systemic (Tobramycin)	Tobramycin	4 ug/mL	3/3	3/3	3/3	2/3 ^b	I
		0.4 ug/mL	3/3	3/3	3/3	3/3	NI

^a: Indeterminate (IND) result

^b: Unresolved (UNR) result

I: Reportable Interference with the BD SARS-CoV-2/Flu for BD MAX System at high concentrations.

NI: No reportable interference with the BD SARS-CoV-2/Flu for BD MAX System.

CLINICAL EVALUATION

Clinical performance characteristics of the BD SARS-CoV-2/Flu for BD MAX System were determined from a total of 232 frozen retrospective nasopharyngeal swabs in UVT/UTM obtained from two external sources with historical positive or negative results for either SARS-CoV-2, influenza A, or influenza B. The specimens were collected as part of routine patient care between 30 November 2019 and 3 September 2020 from 116 males and 116 females ranging in age from 5 months to over 89 years old. All the specimens were tested in a blinded and randomized fashion with the BD MAX SARS-CoV-2/Flu for BD MAX System and reference methods (RM). The RM for SARS-CoV-2 was an EUA authorized high sensitivity RT-PCR assay. The RM for influenza A and influenza B was an FDA-cleared RT-PCR assay.

Tables 11 through 13 describe the performance characteristics of the BD SARS-CoV-2/Flu for BD MAX System that were observed during the clinical evaluation.

Table 11: SARS-CoV-2 Clinical Performance

	Reference Method			
	SARS-CoV-2	POSITIVE	NEGATIVE	Total
BD SARS-CoV-2/Flu for BD MAX System	POSITIVE	50	0	50
	NEGATIVE	2 ^a	30 ^b	32
	Total	52	30	82
SARS-CoV-2 PPA: 96.2% (50/52) (95% CI: 87.0% - 98.9%)				
SARS-CoV-2 NPA: 100% (30/30) (95% CI: 88.7% - 100%)				

^a2/2 specimens were tested with discrepant method and yielded negative results for SARS-CoV-2. One (1) historical result was positive and the other negative.

^bOne unresolved (UNR) result was obtained during initial testing with BD SARS-CoV-2/Flu for BD MAX System and was excluded from analysis.

Table 12: Influenza A Clinical Performance

	Reference Method			
	Influenza A	POSITIVE	NEGATIVE	Total
BD SARS-CoV-2/Flu for BD MAX System	POSITIVE	59 ^a	1 ^b	60
	NEGATIVE	0	90	90
	Total	59	91	150
Influenza A PPA: 100% (59/59) (95% CI: 93.9% - 100%)				
Influenza A NPA: 98.9% (90/91) (95% CI: 94.0% - 99.8%)				

^aOne indeterminate (IND) result was obtained during initial testing with BD SARS-CoV-2/Flu for BD MAX System and was excluded from analysis

^bThe specimen was tested with the discrepant method and yielded a positive result for influenza A. Historical result was positive.

Table 13: Influenza B Clinical Performance

BD SARS-CoV-2/Flu for BD MAX System	Reference Method			
	Influenza B	POSITIVE	NEGATIVE	Total
	POSITIVE	59	0	59
	NEGATIVE	1 ^a	90 ^b	91
	Total	60	90	150

Influenza B PPA: 98.3% (59/60) (95% CI: 91.1% – 99.7%)
Influenza B NPA: 100% (90/90) (95% CI: 95.9% - 100%)

^a The specimen was tested with the discrepant method and yielded a positive result for influenza B. Historical result was positive.

^b One indeterminate (IND) result was obtained during initial testing with BD SARS-CoV-2/Flu for BD MAX System and was excluded from analysis.

REFERENCES

1. Clinical and Laboratory Standards Institute. Protection of laboratory workers from occupationally acquired infections; Approved Guideline. Document M29 (Refer to the latest edition).
2. Centers for Disease Control and Prevention, and National Institutes of Health. Biosafety in microbiological and biomedical laboratories. Chosewood L.C. and Wilson D.E. (eds) (2009). HHS Publication No. (CDC) 21–1112.
3. BD MAX System User's Manual (refer to the latest revision) BD Life Sciences, Sparks, MD 21152 USA.
4. Clinical and Laboratory Standards Institute. Molecular Diagnostic Methods for Infectious Diseases; Approved Guideline, document MM3 (Refer to the latest edition).
5. Clinical and Laboratory Standards Institute. User Protocol for Evaluation of Qualitative Test performance; Approved Guideline, Document EP12 (Refer to the latest edition).
6. Shu, Y., McCauley, J. (2017) GISAID: Global initiative on sharing all influenza data – from vision to reality. EuroSurveillance, 22(13) DOI:10.2807/1560-7917.ES.2017.22.13.30494 PMID: PMC5388101

Change History

Revision	Date	Change Summary
(01)	2021-02	Initial release.

US Customers only: For symbol glossary, refer to bd.com/symbols-glossary



Manufacturer / Производител / Výrobce / Fabrikant / Hersteller / Κατασκευαστής / Fabricante / Tootja / Fabricant / Proizvođač / Gyártó / Fabbicante / Аткарушы / 제조업체 / Gamintojas / Ražotājs / Tilraker / Producent / Producător / Производитель / Výrobca / Proizvođač / Tilverkare / Üretici / Виробник / 生产厂商



Use by / Исползвайте до / Spottebuje do / Brug før / Verwendbar bis / Χρήση έως / Usar antes de / Kasutada enne / Date de péremption / 사용 기한 / Upotrijebiti do / Felhasználhatóság dátuma / Usare entro / Δείν πайдалануға / Naudokite iki / Izlijet lidz / Houdbaar tot / Brukes for / Stosować do / Prazo de validade / A se utiliza până la / Исползовать до / Použite do / Upotrebiti do / Använd före / Son kulanma tarihi / Використати до/line / 使用截止日期

YYYY-MM-DD / YYYY-MM (MM = end of month)
ГГГГ-ММ-ДД / ГГГГ-ММ (ММ = края на месеца)
RRRR-MM-DD / RRRR-MM (MM = konec měsíce)
AAAA-MM-DD / AAAA-MM (MM = slutning af måned)
JJJJ-MM-TT / JJJJ-MM (MM = Monatsende)
EEEE-MM-HH / EEEE-MM (MM = τέλος του μήνα)
AAAA-MM-DD / AAAA-MM (MM = fin del mes)
AAAA-KK-PP / AAAA-KK (KK = kuu lõpp)
AAAA-MM-JJ / AAAA-MM (MM = fin du mois)
GGGG-MM-DD / GGGG-MM (MM = kraj mjeseca)
EEEE-HH-NN / EEEE-HH (HH = hónap utolsó napja)
AAAA-MM-GG / AAAA-MM (MM = fine mese)
ЖЖЖЖ-АА-КК / ЖЖЖЖ-АА (АА = айдың соңы)
YYYY-MM-DD/YYYY-MM (MM = 월말)
MMMM-MM-DD / MMMM-MM (MM = mėnesio pabaiga)
GGGG-MM-DD/GGGG-MM (MM = mėneša beigas)
JJJJ-MM-DD / JJJJ-MM (MM = einde maand)
AAAA-MM-DD / AAAA-MM (MM = sluten av månaden)
RRRR-MM-DD / RRRR-MM (MM = koniec miesiąca)
AAAA-MM-DD / AAAA-MM (MM = fim do mês)
AAAA-LL-ZZ / AAAA-LL (LL = sfârșitul lunii)
ГГГГ-ММ-ДД / ГГГГ-ММ (ММ = конец месяца)
RRRR-MM-DD / RRRR-MM (MM = koniec miesiąca)
GGGG-MM-DD / GGGG-MM (MM = kraj meseca)
AAAA-MM-DD / AAAA-MM (MM = slutet av månaden)
YYYY-AA-GG / YYYY-AA (AA = айна sonu)
PPPP-MM-DD / PPPP-MM (MM = кінець місяця)
YYYY-MM-DD / YYYY-MM (MM = 月末)



Catalog number / Каталоген номер / Katalognové číslo / Katalognummer / Αριθμός καταλόγου / Número de catálogo / Katalooginumber / Numéro catalogue / Kataloški broj / Katalogsszám / Numero di catalogo / Каталог номер / 카탈로그 번호 / Katalog / numeris / Kataloga numurs / Catalogus nummer / Numer katalogowy / Număr de catalog / Номер по каталогу / Katalognové číslo / Kataloški broj / Katalog numarası / Номер за каталогом / 目录号



Authorized Representative in the European Community / Оторизирани представител в Европейската общност / Autorizovaný zástupce pro Evropském společenství / Autoriseret repræsentant i De Europæiske Fællesskaber / Autorisierter Vertreter in der Europäischen Gemeinschaft / Εξουσιοδοτημένος αντιπρόσωπος στην Ευρωπαϊκή Κοινότητα / Representante autorizado en la Comunidad Europea / Volitatud esindaja Euroopa Nõukogus / Représentant autorisé pour la Communauté européenne / Autorizirani predstavnik u Europskoj uniji / Meghatalmazott képviselő az Európai Közösségekben / Rappresentante autorizzato nella Comunità Europea / Европа қауымдастығындағы уәкілетті екіп / 유럽 공동체의 위임 대표 / Įgaliojasis atstovas Europos Bendrijoje / Pārvietotais pārstāvis Eiropas Kopienā / Bevoegde vertegenwoordiger in de Europese Gemeenschap / Autoriseret representant i EU / Autoryzowane przedstawicielstwo we Wspólnocie Europejskiej / Representante autorizado na Comunidade Europeia / Rezentantul autorizat pentru Comunitatea Europeană / Уполномоченный представитель в Европейском сообществе / Autorizovaný zástupca v Európskom spoločenstve / Autorizovano predstavništvo u Evropskoj uniji / Auktoriserad representant i Europeiska gemenskapen / Avvura Toptulugu Yetkili Temsilcisi / Уповноважений представник у країнах ЄС / 歐洲共同體授權代表



In Vitro Diagnostic Medical Device / Медицински уред за диагностика ин витро / Lékařské zařízení určené pro diagnostiku in vitro / In vitro diagnostisk medinsk anordning / Medizinisches In-vitro-Diagnostikum / In vitro διαγνωστική κτηρή συσκευή / Dispositivo médico para diagnóstico in vitro / In vitro diagnostisk meditsiniaparatuur / Dispositif médical de diagnostic in vitro / Medicinska pomagala za In Vitro Dijagnostiku / In vitro diagnostikai orvosi eszköz / Dispositivo medicale per diagnostic in vitro / Жасанды жағдайда жүргізілетін медициналық диагностика аспабы / In Vitro Diagnostik 의료 기기 / In vitro diagnostikos prietais / Medicīnas ierīces, ko lieto in vitro diagnostikā / Medisch hulpmiddel voor in-vitro diagnostiek / In vitro diagnostisk medisinsk utstyr / Urządzenie medyczne do diagnostyki in vitro / Dispositivo médico para diagnóstico in vitro / Dispozitiv medical pentru diagnostic in vitro / Медицинский прибор для диагностики ин витро / Medicínska pomôcka na diagnostiku in vitro / Medicinski uređaj za in vitro dijagnostiku / Medicinteknisk produkt för in vitro diagnostik / In Vitro Diagnostik Tibbi Cihaz / Медицинский прибор для диагностики in vitro / 体外诊断医疗设备



Temperature limitation / Температурни ограничения / Teplotní omezení / Temperaturbegrænsning / Temperaturbegrenzung / Περιορισμοί θερμοκρασίας / Limitación de temperatura / Temperaturi pīrang / Limites de température / Dozvoljena temperatura / Hőmérsékleti határ / Limiti di temperatura / Температураны шектеу / 온도 제한 / Laikymo temperatūra / Temperatūras ierobežojumi / Temperaturlimit / Temperaturbegrensning / Ograniczenie temperatury / Limites de temperatura / Limite de temperatură / Ограничення температури / Ограничение температуры / Ograničenje teploty / Ograničenje temperature / Temperaturgräns / Sıcaklık sınırlaması / Обмеження температури / 温度限制



Batch Code (Lot) / Код на партидата / Kód (číslo) šarže / Batch-kode (lot) / Batch-Code (Charge) / Κωδικός παρτίδας (παρτίδα) / Código del lote (lote) / Partii kood / Numéro de lot / Lot (kod) / Tétel száma (Lot) / Codice batch (lotto) / Топтама коды / 배치 코드(로트) / Partijos numeris (LOT) / Partijas kods (laidens) / Lot nummer / Batch-kode (parti) / Kod partii (serie) / Código do lote / Cod de serie (Lot) / Код партии (лот) / Kód série (šarža) / Kod serije / Partinummer (Lot) / Parti Kodu (Lot) / Код партии / 批号 (亚批)



Contains sufficient for <n> tests / Съдържащието е достатъчно за <n> теста / Dostatečné množství pro <n> testů / Indeholder tilstrækkeligt til <n> tests / Ausreichend für <n> Tests / Περιέχει επαρκή ποσότητα για <n> εξετάσεις / Contenido suficiente para <n> pruebas / Kifüladare <n> testide jaoks / Contenu suffisant pour <n> tests / Szárazza <n> testov / <n> teszthez elegendő / Contenido suficiente per <n> test / <n> тесттери үшін жеткілікті / <n> 테스트가 충분히 포함됨 / Pākanamas kiekis atlikti <n> testų / Satur pietiekami <n> pārbaudēm / Inhoud voldoende voor <n> testen / Inholder tilstrækkelig til <n> tester / Zawiera ilość wystarczającą do <n> testów / Conteúdo suficiente para <n> testes / Conținut suficient pentru <n> teste / Достаточна для <n> тестов(a) / Obsah vystačí na <n> testov / Szárazza dovoljan za <n> testova / Innehåller tillräckligt för <n> analyser / <n> test için yeterli mabzeme içerir / Вистачить для аналізів <n> / 足夠進行 <n> 次检测



Consult Instructions for Use / Направете справка в инструкциите за употреба / Prostudujte pokyny k použití / Se brugsanvisningen / Gebrauchsanweisung beachten / Συμβουλευτείτε τις οδηγίες χρήσης / Consultar las instrucciones de uso / Lugeda kasutusjuhendit / Consulter la notice d'emploi / Koristi upute za upotrebu / Olvassa el a használati utasítást / Consultare le istruzioni per l'uso / Пайдалану ыңауаулығымен танысып алыңыз / 사용 지침 참조 / Skaitykite naudojimo instrukcijas / Skatīt lietošanas pamācību / Raadpleeg de gebruiksaanwijzing / Se i brugsanvisningen / Zobacz instrukcja użytkowania / Consultar as instruções de utilização / Consultati instrucțiunile de utilizare / См. руководство по эксплуатации / Pozri Pokyny na používanie / Pogledajte uputstvo za upotrebu / Se brugsanvisningen / Kullanim Talimatları'na başvurun / Διαβ. ηστρουκλιών z використання / 请参阅使用说明



Do not reuse / Не използвайте отново / Nepoužívejte opakovaně / Ikke til genbrug / Nicht wiederverwenden / Μην επαναχρησιμοποιείτε / No reutilizar / Mitte kasutada korduvalt / Ne pas réutiliser / Ne koristiti ponovo / Egyszer használatos / Non riutilizzare / Пайдаланбаңыз / 재사용 금지 / Tik vienkartiniam naudojimui / Nelietot atkārtoti / Niet opnieuw gebruiken / Kun til engangsbruk / Nie stosować powtórnie / Não reutilize / Nu refolosiți / Не использовать повторно / Nepoužívejte opakovane / Ne utreibjauvte porogo / Får ej återanvändas / Tekrar kulanmayın / Не використовувати повторно / 请勿重复使用



Serial number / Серийен номер / Sériové číslo / Seriennummer / Seriennummer / Σειριακός αριθμός / N° de série / Seerianumber / Numéro de série / Serijski broj / Sorozatszám / Numero di serie / Топтамалық номер / 일련 번호 / Serijos numeris / Sérijas numurs / Serie nummer / Numer serijny / Número de série / Număr de serie / Серийный номер / Ser numarası / Номер серії / 序列号



For IVD Performance evaluation only / Само за оценка качеството на работа на IVD / Pouze pro vyhodnocení výkonu IVD / Kun til evaluering af IVD ydelse / Nur für IVD-Leistungsbewertungszwecke / Μόνο για αξιολόγηση απόδοσης IVD / Sólo para la evaluación del rendimiento en diagnóstico in vitro / Ainult IVD seadme hindamiseks / Réserve à l'évaluation des performances IVD / Samo u znanstvene svrhe za In Vitro Dijagnostiku / Kizárolag in vitro diagnosztikához / Solo per valutazione delle prestazioni IVD / Жасанды жағдайда «пробирка шінде» диагностикада тек жұмысты бағалау үшін / IVD 성능 평가에 대해서만 사용 / Tik IVD prietais veikimo charakteristikoms tikrinti / Vienigi IVD darbi bas novērtēšanai / Uitsluitend voor doeltreffendheidsonderzoek / Kun for evaluering af IVD-ydelse / Tyko do oceny wydajności IVD / Uso esclusivo para avaliação de IVD / Numai pentru evaluarea performanței IVD / Только для оценки качества диагностики ин витро / Samo za procenu učinka u in vitro dijagnostici / Endast för utvärdering av diagnostisk användning in vitro / Yalnızca IVD Performans değerlendirmesi için / Тільки для оцінювання якості діагностики in vitro / 仅限 IVD 性能评估

For US: "For Investigational Use Only"



Lower limit of temperature / Долой лимит на температурата / Dolní hranice teploty / Nedre temperaturgrænse / Temperaturuntergrenze / Κατώτερο όριο θερμοκρασίας / Limite inferior de temperatura / Alumine temperatuuripiir / Limite inférieure de température / Najnižja dovoljena temperatura / Alsó hőmérsékleti határ / Limite inferiore di temperatura / Температураның төменгі рускат шегі / 하한 온도 / Žemiausia laikymo temperatūra / Temperatūras zemākā robeža / Laagste temperatuurlimit / Limite temperaturgrense / Dolna granica temperatury / Limite mínimo de temperatura / Limită minimă de temperatură / Нижний предел температуры / Spodná hranica teploty / Donja granica temperature / Nedre temperaturgräns / Sıcaklık alt sınırı / Мінімальна температура / 温度下限

	Control / Контроль / Kontrola / Kontrol / Kontrolle / Μάρτυρας / Kontroll / Contrôle / Controllo / Бақылау / 컨트롤 / Kontrolé / Kontrolle / Controle / Controllo / Контроль / kontroll / Контроль / 对照
	Positive control / Положительный контроль / Pozitívny kontrola / Positiv kontrol / Positive Kontrolle / Θετικός μάρτυρας / Control positivo / Positiivne kontroll / Contrôle positif / Pozitívna kontrola / Pozitívny kontroll / Controllo positivo / Оң бақылау / 양성 컨트롤 / Teigiana kontrolė / Pozitívna kontrolė / Positiive controle / Kontrola dodatná / Controllo positivo / Control pozitiv / Положительный контроль / Pozitíf kontrol / Позитивний контроль / 阳性对照试剂
	Negative control / Отрицательный контроль / Negatívny kontrola / Negatív kontrol / Negative Kontrolle / Αρνητικός μάρτυρας / Control negativo / Negatiivne kontroll / Contrôle négatif / Negatívna kontrola / Negatívny kontroll / Controllo negativo / Негативный контроль / Negatívna kontrolė / Negatívna kontrolė / Negatiive controle / Kontrola ujemna / Controllo negativo / Control negativ / Отрицательный контроль / Negatív kontrol / Негативний контроль / 阴性对照试剂
	Method of sterilization: ethylene oxide / Метод на стерилизация: етиленов оксид / Způsob sterilizace: etylenoxid / Steriliseringmetode: ethylenoxid / Sterilisationsmethode: Ethylenoxid / Μέθοδος αποστείρωσης: αιθυλοξείδιο / Método de esterilización: óxido de etileno / Steriliseerimismetod: etüleenoksiid / Méthode de stérilisation: oxyde d'éthylène / Metoda sterilizacji: etilen oksid / Sterilizálás módszere: etilén-oxid / Metodo di sterilizzazione: ossido di etilene / Стерилизация әдісі: этилен тотығы / 소독 방법: 에틸렌옥사이드 / Sterilizavimo būdas: etileno oksidas / Sterilizācijas metode: etilēnoksīds / Gesteriliseerd met behulp van ethyleenoxide / Steriliseringmetode: etylenoxid / Metoda sterylizacji: tlenek etylu / Método de esterilização: óxido de etileno / Metodă de sterilizare: oxid de etilenă / Метод стерилизации: этиленоксид / Metóda sterilizácie: etýlénoxid / Metoda sterilizacije: etilen oksid / Steriliseringsmetod: etenoxid / Sterilizasyon yöntemi: etilen oksit / Метод стерилизації: етиленоксидом / 灭菌方法: 环氧乙烷
	Method of sterilization: irradiation / Метод на стерилизация: иррадиация / Způsob sterilizace: záření / Steriliseringmetode: bestrålning / Sterilisationsmethode: Bestrahlung / Μέθοδος αποστείρωσης: ακτινοβολία / Método de esterilización: irradiación / Steriliseerimismetod: kiirgus / Méthode de stérilisation: irradiation / Metoda sterilizacije: zračenje / Sterilizálás módszere: besugárzás / Metodo di sterilizzazione: irradiazione / Стерилизация әдісі: сәуле түсірі / 소독 방법: 방사 / Sterilizavimo būdas: radiacija / Sterilizācijas metode: apstarošana / Gesteriliseerd met behulp van bestraling / Steriliseringmetode: bestrålning / Metoda sterylizacji: napromienianie / Método de esterilização: irradiação / Metodă de sterilizare: iradiere / Метод стерилизации: облучение / Metóda sterilizácie: ožarovanie / Metoda sterilizacije: zračevanje / Steriliseringsmetod: strålning / Sterilizasyon yöntemi: ışınlama / Метод стерилизації: опромінення / 灭菌方法: 辐射
	Biological Risks / Биологични рискове / Biologická rizika / Biologisk fare / Biogefährdung / Βιολογικά κίνδυνοι / Riesgos biológicos / Bidoogised riskid / Risques biologiques / Biološki rizik / Bídógialag veszélyes / Rischio biologico / Биологическое опасение / 생물학적 위험 / Biologinis pavojus / Biologiské risiko / Biologisk risiko / Zagrożenia biologiczne / Perigo biológico / Riscuri biologice / Биологическая опасность / Biologické riziko / Biološki rizici / Biologisk risk / Biyolojik Riskler / Біологічна небезпека / 生物学风险
	Caution, consult accompanying documents / Внимание, направьте справка в придружащите документи / Pozor! Prostudujte si příloženou dokumentaci! / Forsigtig, se ledsagende dokumenter / Achtung, Begleitdokumente beachten / Προσοχή, συμβουλευτείτε τα συνοδευτικά έγγραφα / Precaución, consultar la documentación adjunta / Ettevaatus! Luge da kaasnevat dokumentatsiooni / Attention, consulter les documents joints / Upozorenje, koristi prateću dokumentaciju / Figyelem! Olvassa el a mellékelt tájékoztatót / Attenzione: consultare la documentazione allegata / Абайлаңыз, тілсі құжаттармен танысыңыз / 주의, 동봉된 설명서 참조 / Dămesio, žiūrėkite priedamus dokumentus / Piesardzība, skatīt pavaddokumentus / Voorzichtig, raadpleeg bijgevoegde documenten / Forsiktig, se vedlagt dokumentasjon / Należy zapoznać się z dołączonymi dokumentami / Cuidado, consulte a documentação bmeccida / Atenție, consultați documentele însoțitoare / Внимание: см. прилагаемую документацию / Výstraha, poz srievočné dokumenty / Pažnja! Pogledajte priložena dokumenta / Obs! Se medföljande dokumentation / Dikkat, birlikte verilen belgeleri başurun / Увага: див. супутню документацію / 小心, 请参阅附带文档。
	Upper limit of temperature / Горен лимит на температурата / Horní hranice teploty / Øvre temperaturgrænse / Temperaturobergrenze / Ανώτερο όριο θερμοκρασίας / Limite superior de temperatura / Ülemine temperatuuripiir / Limite supérieure de température / Gornja dozvoljena temperatura / Felső hőmérsékleti határ / Limite superiore de temperatura / Температураның жусат етілен жоғарғы шеі / 상한 온도 / Aukščiausia laikymo temperatūra / Augšējā temperatūras robeža / Hoogste temperatuurimiet / Øvre temperaturgrense / Gorna granica temperatury / Limite máximo de temperatura / Limită maximă de temperatură / Верхний предел температуры / Horná hranica teploty / Gornja granica temperature / Øvre temperaturgräns / Sıcaklık üst sınırı / Максимальна температура / 温度上限
	Keep dry / Пазете сухо / Skladujte v suchém prostredí / Opbevaes tort / Trocklagem / Φυλάξτε το στεγνό / Mantener seco / Hoida kuivas / Conserver au sec / Držati na suhom / Száraz helyen tartandó / Tenere all'asciutto / Құрғақ күйінде ұста / 건조 상태 유지 / Laikykite sausai / Uzglabāt sausu / Droog houden / Hoides tort / Przechowywać w stanie suchym / Manter seco / A se feri de umezală / Не допускать попадания влаги / Uchovávejte v suchu / Držite na suwom miejscu / Fönvaras tort / Kuru bir şekilde muhafaza edin / Берегти від вологи / 请保持干燥
	Collection time / Време на събиране / Čas odběru / Opsamlingsstidspunkt / Entnahmeuhrzeit / Ωρα συλλογής / Hora de recogida / Kogumisaeg / Heure de prélèvement / Sati prikupljanja / Mntavétel időpontja / Ora di raccolta / Жынау уақыты / 수집 시간 / Paėmimo laikas / Savākšanas laiks / Verzamelstid / Tid provetaking / Godzina pobrania / Hora de colecta / Ora colectării / Время сбора / Doba odboru / Vreme prikupljanja / Uppsamlingsstid / Toplama zamanı / Час забору / 采集时间
	Peel / Обенете / Otevíte zde / Abn / Abziehen / Αποκόλληστε / Desprender / Koorida / Décoller / Ολοκίτι skin / Húzza le / Staccare / Üstüñi қабатын алып таста / 벗기기 / Pięśń cia / Atłmēt / Schlen / T reik av / Oderwać / Destacar / Se dezlipęse / Отклеить / Odtrhně / Oljušiti / Dra isär / Аутма / Відклеїти / 撕下
	Perforation / Перфорация / Perforace / Perforering / Δάτρηση / Perforación / Perforatsioon / Perforacija / Perforálás / Perforazione / Tęcki тесу / 절취선 / Perforacija / Perforacija / Perforatie / Perforacja / Perfuração / Perforare / Перфорация / Perforada / Perforasyon / Перфорация / 穿孔
	Do not use if package damaged / Не използвайте, ако опаковката е повредена / Ne používejte, je-li obal poškozený / Må ikke anvendes hvis emballagen er beskadiget / Inhal beschädigter Packung nicht verwenden / Μη χρησιμοποιείτε εάν η συσκευασία έχει υποστεί ζημιά / No usar si el paquete está dañado / Mitte kasutada, kui pakend on kahjustatud / Ne pas l'utiliser si l'emballage est endommagé / Ne koristiti ako je oštećeno pakiranje / Ne használni, ha a csomagolás sérült / Non usare se la confezione è danneggiata / Егер пакет бұзылған болса, пайдаланба / पैकिङ کا 손상된 경우 사용 금지 / Jei pakuoatė pažeista, nenaudoti / Nelietoti, ja iepakojums bojāts / Niet gebruiken indien de verpakking beschadigd is / Må ikke brukes hvis pakke er skadet / Nie używać, jeśli opakowanie jest uszkodzone / Não usar se a embalagem estiver danificada / A nu se folosi dacă pachetul este deteriorat / Не использовать при повреждении упаковки / Ne pouzivat, ak je obal poškozený / Ne koristiti ako je pakovanje oštećeno / Använd ej om förpackningen är skadad / Ambalaj hasar götmüşse kullanmayın / Не використовувати за пошкодженої упаковки / 如果包装破损, 请勿使用
	Keep away from heat / Пазете от топлина / Nəystavujte priľišnému teplu / Må ikke udsættes for varme / Vor Wärme schützen / Κρατήστε το μακριά από τη θερμότητα / Mantener alejado de fuentes de calor / Hoida eemal valgusest / Protéger de la chaleur / Držati dalje od izvora topline / Óvja a melegtől / Tenere lontano dal calore / Limite superiore de / 열을 피해야 함 / Laikyti atokiau nuo šilumos šaltinių / Sargătino karstuma / Beschermen tegen warmte / Må ikke utsættes for varme / Przechowywać z dala od źródeł ciepła / Manter ao abrigo do calor / A se feri de căldură / Не нагревать / Uchovávejte mimo zdroja tepla / Držite dalje od toplote / Får ej utsättas för värme / Isidan uzak tutun / Берегти від тепла / 请远离热源
	Cut / Срежете / Odstřihněte / Klip / Schneiden / Κόψτε / Cortar / Lõigata / Découper / Reži / Vágja ki / Tagliare / Kecişiz / 잘라내기 / Kirpti / Nogriez / Krippen / Kutt / Odciać / Cortar / Descupați / Отрезать / Odstrhně / Iseći / Klipp / Kesme / Pozriati / 剪下
	Collection date / Дата на събиране / Datum odběru / Opsamlingsdato / Entnahmedatum / Ημερομηνία συλλογής / Fecha de recogida / Kogumiskupäev / Date de prélèvement / Dani prikupljanja / Mntavétel dátuma / Data di raccolta / Жынаған тзбекүні / 수집 날짜 / Paėmimo data / Savākšanas datums / Verzamelatum / Dato provetaking / Data pobrania / Data de colecta / Ora colectării / Дата сбора / Doba odboru / Datum prikupljanja / Uppsamlingsstid / Toplama tarihi / Дата забору / 采集日期
	µL/test / µL/тест / µL/Test / µL/έξταση / µL/pueba / µL/teszt / µL/테스트 / мкл/тест / µL/tyrimas / µL/pārbaude / µL/teste / мкл/анализ / µL/检测
	Keep away from light / Пазете от светлина / Nəystavujte svetlu / Må ikke udsættes for lys / Vor Licht schützen / Κρατήστε το μακριά από το φως / Mantener alejado de la luz / Hoida eemal valgusest / Conserver à l'abri de la lumière / Držati dalje od svetla / Fény nem érheti / Tenere al riparo dalla luce / Қазақылған жерде ұста / 빛을 피해야 함 / Laikyti atokiau nuo šilumos šaltinių / Sargătino gaismas / Niet blootstellen aan zonlicht / Må ikke utsættes for lys / Przechowywać z dala od źródeł światła / Manter ao abrigo da luz / Feriti de lumină / Хранить в темноте / Uchovávejte mimo dosahu svetla / Držite dalje od svetlosti / Får ej utsättas för ljus / Isikdan uzak tutun / Берегти від світла / 请远离光线
	Hydrogen gas generated / Образуван е водород газ / Možnost úniku plynného vodíku / Frembringer hydrogengas / Wasserstoffgas erzeugt / Δημιουργία αερίου υδρογόνου / Producción de gas de hidrógeno / Vesinikgaasi tekitatud / Produit de l'hydrogène gazeux / Sadrži hydrogen vodik / Hidrogén gázt fejeleszt / Produzione di gas idrogeno / Газтөктес сыерти пайда болды / 수소 가스 생성됨 / Išiskiria vandenilio dujas / Rodas udegnis / Waterstoff gas generert / Hydrogengas generert / Powoduje powstawanie wodoru / Produção de gás de hidrogénio / Generare gaz de hidrogen / Выделение водорода / Vytobené použitím vodíka / Oslobada se vodonik / Genererad välgas / Ağıda çıkan hidrojen gazı / Реакция з виділенням водню / 会产生氢气
	Patient ID number / ИД номер на пациента / ID pacienā / Patientens ID-nummer / Patienten-ID / Αριθμός αναγνώρισης ασθενούς / Número de ID del paciente / Patsiendi ID / No d'identification du patient / Identifikációs broj pacienā / Beteg azonosító száma / Numero ID paziente / Пациенттин идентификациялык нөмрү / 환자 ID 번호 / Paciento identifikavimo numeris / Pacienā ID numurs / Identificatienummer van de patiënt / Pasientens ID-nummer / Numer ID pacienā / Número da ID do doente / Număr ID pacient / Идентификационный номер пациента / Identifikačné číslo pacienā / ID broj pacienā / Patientnummer / Hasta kimlik numarası / Идентификатор пациента / 患者标识号



Fragile, Handle with Care / Чупливо, Работете с необходимото внимание. / Křehké, Při manipulaci postupujte opatrně. / Forsiktig, kan gå i stykker. / Zerbrechlich, vorsichtig handhaben. / Ευθραusto. Χειρτάτε το με προσοχή. / Frágil, Manipular con cuidado. / Öm, käsitlege ettevaatlikult. / Fragile, Manipuler avec précaution. / Lomjivo, rukujte pažljivo. / Törékeny! Óvatosan kezelendő. / Fragile, maneggiare con cura. / Сыныш, абайлап пайдаланыңыз. / 조심 깨지기 쉬운 처리 / Trapu, elkités atsargiai / Trausis; rickotis uzmanig / Breekbaar, voorzichtig behandelen. / Ömtälg, händler försiktig. / Kucha zawartość, przenosić ostrożnie. / Frágil, Manuseie com Cuidado. / Frágil, manipulați cu atenție. / Хрупкое! Обращаться с осторожностью. / Křehké, vyžaduje sa opatrná manipulácia. / Lomjivo - rukujte pažljivo. / Bräckligt. Hantera försiktigt. / Kolay Kırılır, Dikkatli Taşıyın. / Тендітпа, звертатися з обережністю / 易碎, 小心轻放

Rx Only

This only applies to US: "Caution: Federal Law restricts this device to sale by or on the order of a licensed practitioner." / S'applique uniquement aux États-Unis: "Caution: Federal Law restricts this device to sale by or on the order of a licensed practitioner." / Vale solo per gli Stati Uniti: "Caution: Federal Law restricts this device to sale by or on the order of a licensed practitioner." / Gilt nur für die USA: "Caution: Federal Law restricts this device to sale by or on the order of a licensed practitioner." / Sólo se aplica a los EE.UU.: "Caution: Federal Law restricts this device to sale by or on the order of a licensed practitioner."



bd.com/e-labeling



Technical Information: In the United States contact BD Technical Service and Support at 1.800.638.8663 or bd.com.



Becton, Dickinson and Company
7 Loveton Circle
Sparks, Maryland 21152 USA

BD, the BD Logo, and MAX are trademarks of Becton, Dickinson and Company or its affiliates. All other trademarks are the property of their respective owners. © 2021 BD. All rights reserved.

SARS-CoV-2/Flu for BD MAX™ System

IVD Rx Only

REF 445011

USA – For Use Under Emergency Use Authorization (EUA) Only

USA:



bd.com/e-labeling

Please contact BD Technical Service and Support at 1.800.638.8663 or bd.com for questions or if you require a printed copy free of charge or need technical support to access the Instructions for Use.

- This product has not been FDA cleared or approved, but has been authorized for emergency use by FDA under an EUA for use by authorized laboratories;
- This product has been authorized only for the detection of nucleic acid from SARS CoV-2, influenza A virus and influenza B virus and not for any other viruses or pathogens; and
- The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of in vitro diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.

Non - USA:



bd.com/e-labeling

Please contact your local BD representative or bd.com if you require a printed copy free of charge or need technical support to access the Instructions for Use. / Veuillez contacter le représentant local BD ou consulter le site bd.com si vous avez besoin d'une copie imprimée gratuite ou d'une assistance technique pour accéder à la notice d'utilisation. / Contattare il rappresentante BD locale o bd.com se si necessita di una copia stampata gratuita o si ha bisogno di supporto tecnico per accedere alle Istruzioni per l'uso. / Setzen Sie sich mit Ihrer zuständigen BD-Vertretung in Verbindung oder besuchen Sie bd.com, wenn Sie ein kostenloses gedrucktes Exemplar oder technische Unterstützung beim Zugriff auf die Gebrauchsanweisung benötigen. / Póngase en contacto con el representante local de BD o bd.com si necesita una copia impresa gratuita o ayuda técnica para acceder a las Instrucciones de uso. / Potřebujete-li tištěnou kopii návodu k použití bez poplatku, nebo technickou podporu pro přístup k návodu k použití, obraťte se prosím na místního zástupce společnosti BD nebo na webovou stránku bd.com. / Kontakt den lokale BD-repræsentant eller bd.com, hvis du har brug for en gratis papirversion eller teknisk support for at få adgang til brugsanvisningen. / Παρακαλούμε επικοινωνήστε με τον τοπικό σας αντιπρόσωπο της BD ή επισκεφθείτε τη διεύθυνση bd.com εάν επιθυμείτε ένα δωρεάν εκτυπωμένο αντίγραφο ή χρειάζεστε τεχνική υποστήριξη για την πρόσβαση στις Οδηγίες Χρήσης. / Obratite se lokalnom predstavniku tvrtke BD ili posjetite bd.com ako vam je potreban besplatni tiskani primjerak ili tehnička podrška za pristup Uputama za uporabu. / Ta kontakt med den lokale BD-representanten eller bd.com hvis du trenger gratis utskrift eller har behov for teknisk støtte for å få tilgang til bruksanvisningen. / Contacte o representante local da BD ou bd.com se necessitar de uma cópia impressa gratuita ou precisar de assistência técnica para aceder às Instruções de utilização. / Contate o representante local da BD ou visite bd.com se desejar solicitar uma cópia impressa gratuita ou necessitar de suporte técnico para acessar as Instruções de uso. / Kontakta din lokale BD-representant eller gå till bd.com om du behöver en tryckt kostnadsfri kopia eller teknisk support för att komma åt bruksanvisningen. / Kullanım Talimatlarının ücretsiz basılı bir kopyası için veya Kullanım Talimatlarına erişmek üzere teknik desteğe ihtiyacınız varsa, lütfen yerel BD temsilcinizle iletişim kurun veya bd.com adresini ziyaret edin. / Зв'яжіться з місцевим представником компанії BD або зверніться за посиланням bd.com, якщо вам потрібна безкоштовна друкована копія або технічна підтримка, щоб отримати доступ до інструкцій із використання.

Please visit go.bd.com/BDMAX to learn more!



BD, the BD Logo, and MAX are trademarks of Becton, Dickinson and Company or its affiliates. © 2021 BD. All rights reserved.