

**SAFETY DATA SHEET**

according to the Global Harmonized System (and with all of the information required by the HPR)

Revision Date 06/17/2018

Version 1.4

**SECTION 1. Identification****Product identifier**

Product number	632805
Millipore Ref.	4200-0140
Product name	Guava Instrument Cleaning Fluid (ICF) (100 mL) for all Guava Flow Cytometers

**Relevant identified uses of the substance or mixture and uses advised against**

Identified uses	Laboratory cleaning
-----------------	---------------------

**Details of the supplier of the safety data sheet**

Company	Millipore (Canada) Ltd   109 Woodbine Downs Blvd. Unit 5   Etobicoke   Ontario M9W 6Y1   Canada   General Inquiries: +1 800-645-5476   Monday to Friday, 9:00 AM to 4:00 PM Eastern Time (GMT-5) MilliporeSigma is a business of Merck KGaA, Darmstadt, Germany.
---------	---

Emergency telephone	800-424-9300 CHEMTREC (USA) +1-703-527-3887 CHEMTREC (International) 24 Hours/day; 7 Days/week
---------------------	--

**SECTION 2. Hazards identification****GHS Classification**

Skin irritation, Category 2, H315  
Eye irritation, Category 2A, H319

For the full text of the H-Statements mentioned in this Section, see Section 16.

**GHS-Labeling***Hazard pictograms*

*Signal Word*  
Warning

*Hazard Statements*  
H315 Causes skin irritation.  
H319 Causes serious eye irritation.

*Precautionary Statements*

# SAFETY DATA SHEET

according to the Global Harmonized System (and with all of the information required by the HPR)

Product number	632805	Version 1.4
Product name	Guava Instrument Cleaning Fluid (ICF) (100 mL) for all Guava Flow Cytometers	

---

P264 Wash skin thoroughly after handling.  
P280 Wear protective gloves/ eye protection/ face protection.  
P302 + P352 IF ON SKIN: Wash with plenty of soap and water.  
P305 + P351 + P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.  
P321 Specific treatment (see supplemental first aid instructions on this label).  
P332 + P313 If skin irritation occurs: Get medical advice/ attention.  
P337 + P313 If eye irritation persists: Get medical advice/ attention.  
P362 Take off contaminated clothing and wash before reuse.

## Other hazards

None known.

---

## SECTION 3. Composition/information on ingredients

Chemical nature	Aqueous solution
Remarks	WHMIS hazardous composition: No ingredients are hazardous according to the CPR criteria.

---

## SECTION 4. First aid measures

### Description of first-aid measures

#### *Inhalation*

After inhalation: fresh air.

#### *Skin contact*

In case of skin contact: Take off immediately all contaminated clothing. Rinse skin with water/ shower.

#### *Eye contact*

After eye contact: rinse out with plenty of water. Call in ophthalmologist. Remove contact lenses.

#### *Ingestion*

After swallowing: make victim drink water (two glasses at most). Consult doctor if feeling unwell.

Never give anything by mouth to an unconscious person.

### Most important symptoms and effects, both acute and delayed

irritant effects

Cough, Shortness of breath

### Indication of any immediate medical attention and special treatment needed

No information available.

---

## SECTION 5. Fire-fighting measures

### Extinguishing media

#### *Suitable extinguishing media*

Use extinguishing measures that are appropriate to local circumstances and the surrounding environment.

# SAFETY DATA SHEET

according to the Global Harmonized System (and with all of the information required by the HPR)

Product number	632805	Version 1.4
Product name	Guava Instrument Cleaning Fluid (ICF) (100 mL) for all Guava Flow Cytometers	

---

## *Unsuitable extinguishing media*

For this substance/mixture no limitations of extinguishing agents are given.

## **Special hazards arising from the substance or mixture**

Not combustible.

Ambient fire may liberate hazardous vapors.

## **Advice for firefighters**

### *Special protective equipment for fire-fighters*

Stay in danger area only with self-contained breathing apparatus. Prevent skin contact by keeping a safe distance or by wearing suitable protective clothing.

### *Further information*

none

---

## **SECTION 6. Accidental release measures**

### **Personal precautions, protective equipment and emergency procedures**

Advice for non-emergency personnel: Do not breathe vapors, aerosols. Avoid substance contact. Evacuate the danger area, observe emergency procedures, consult an expert.

Advice for emergency responders: Protective equipment see section 8.

### **Environmental precautions**

No special precautionary measures necessary.

### **Methods and materials for containment and cleaning up**

Observe possible material restrictions (see sections 7 and 10).

Take up with liquid-absorbent and neutralizing material (e.g. Chemizorb® OH<sup>-</sup>, Art. No. 101596).

Dispose of properly. Clean up affected area.

No special precautionary measures necessary.

---

## **SECTION 7. Handling and storage**

### **Precautions for safe handling**

Observe label precautions.

### **Conditions for safe storage, including any incompatibilities**

Requirements for storage areas and containers

No aluminum, tin, or zinc containers.

Tightly closed.

Store at +10°C to +30°C (+50°F to +86°F).

---

## **SECTION 8. Exposure controls/personal protection**

### **Exposure limit(s)**

Contains no substances with occupational exposure limit values. Contains no substances with occupational exposure limit values.

### **Engineering measures**

Technical measures and appropriate working operations should be given priority over the use of personal protective equipment.

---

# SAFETY DATA SHEET

according to the Global Harmonized System (and with all of the information required by the HPR)

Product number	632805	Version 1.4
Product name	Guava Instrument Cleaning Fluid (ICF) (100 mL) for all Guava Flow Cytometers	

---

## Individual protection measures

Protective clothing should be selected specifically for the workplace, depending on concentration and quantity of the hazardous substances handled. The chemical resistance of the protective equipment should be inquired at the respective supplier.

### *Hygiene measures*

Immediately change contaminated clothing. Apply skin- protective barrier cream. Wash hands and face after working with substance.

### *Eye/face protection*

Safety glasses

### *Hand protection*

full contact:

Glove material:	Nitrile rubber
Glove thickness:	0.11 mm
Break through time:	> 480 min

splash contact:

Glove material:	Nitrile rubber
Glove thickness:	0.11 mm
Break through time:	> 480 min

The protective gloves to be used must comply with the specifications of EC Directive 89/686/EEC and the related standard EN374, for example KCL 741 Dermatril® L (full contact), KCL 741 Dermatril® L (splash contact).

The breakthrough times stated above were determined by KCL in laboratory tests acc. to EN374 with samples of the recommended glove types.

This recommendation applies only to the product stated in the safety data sheet and supplied by us as well as to the purpose specified by us. When dissolving in or mixing with other substances and under conditions deviating from those stated in EN374 please contact the supplier of CE-approved gloves (e.g. KCL GmbH, D-36124 Eichenzell, Internet: [www.kcl.de](http://www.kcl.de)).

### *Other protective equipment:*

protective clothing

### *Respiratory protection*

required

Recommended Filter type: Filter P 2 (acc. to DIN 3181) for solid and liquid particles of harmful substances

The entrepreneur has to ensure that maintenance, cleaning and testing of respiratory protective devices are performed according to the instructions of the producer. These measures have to be properly documented.

---

## SECTION 9. Physical and chemical properties

Physical state	liquid
Color	clear colorless
Odor	No strong odor known.
Odor Threshold	No information available.

## SAFETY DATA SHEET

according to the Global Harmonized System (and with all of the information required by the HPR)

Product number	632805	Version 1.4
Product name	Guava Instrument Cleaning Fluid (ICF) (100 mL) for all Guava Flow Cytometers	

---

pH	12.7 - 13 at 68 °F (20 °C)
Melting point	No information available.
Boiling point	No information available.
Flash point	Not applicable
Evaporation rate	No information available.
Flammability (solid, gas)	No information available.
Lower explosion limit	No information available.
Upper explosion limit	No information available.
Vapor pressure	No information available.
Relative vapor density	1.06
Density	No information available.
Relative density	No information available.
Water solubility	No information available.
Partition coefficient: n-octanol/water	No information available.
Autoignition temperature	No information available.
Decomposition temperature	No information available.
Viscosity, dynamic	No information available.
Explosive properties	Not classified as explosive.
Oxidizing properties	none

---

### SECTION 10. Stability and reactivity

#### Reactivity

See below

#### Chemical stability

The product is chemically stable under standard ambient conditions (room temperature) .

#### Possibility of hazardous reactions

Gives off hydrogen by reaction with metals.

Exothermic reaction with:

## SAFETY DATA SHEET

according to the Global Harmonized System (and with all of the information required by the HPR)

Product number	632805	Version 1.4
Product name	Guava Instrument Cleaning Fluid (ICF) (100 mL) for all Guava Flow Cytometers	

---

Strong acids and oxidizing agents, Halogenated compounds, anhydrides, Alcohols, and their compounds

Violent reactions possible with:

The generally known reaction partners of water.

### Conditions to avoid

no information available

### Incompatible materials

Metals, animal/vegetable tissues, various plastics, glass

Hydrogen may form upon contact with light metals (danger of explosion!).

### Hazardous decomposition products

no information available

---

## SECTION 11. Toxicological information

### Information on toxicological effects

#### *Likely route of exposure*

Eye contact, Skin contact

#### *Acute oral toxicity*

Symptoms: Irritations of mucous membranes in the mouth, pharynx, oesophagus and gastrointestinal tract.

#### *Acute inhalation toxicity*

Symptoms: Cough, Shortness of breath

Symptoms: Possible damages:, mucosal irritations

#### *Skin irritation*

Mixture causes skin irritation.

#### *Eye irritation*

Mixture causes serious eye irritation.

#### *Specific target organ systemic toxicity - single exposure*

The substance or mixture is not classified as specific target organ toxicant, single exposure.

#### *Specific target organ systemic toxicity - repeated exposure*

The substance or mixture is not classified as specific target organ toxicant, repeated exposure.

#### *Aspiration hazard*

Regarding the available data the classification criteria are not fulfilled.

### Carcinogenicity

IARC

No ingredient of this product present at levels greater than or equal to 0.1% is identified as probable, possible or confirmed human carcinogen by IARC.

OSHA

No component of this product present at levels greater than or equal to 0.1% is on OSHA's list of regulated carcinogens.

## SAFETY DATA SHEET

according to the Global Harmonized System (and with all of the information required by the HPR)

Product number	632805	Version 1.4
Product name	Guava Instrument Cleaning Fluid (ICF) (100 mL) for all Guava Flow Cytometers	

---

NTP	No ingredient of this product present at levels greater than or equal to 0.1% is identified as a known or anticipated carcinogen by NTP.
ACGIH	No ingredient of this product present at levels greater than or equal to 0.1% is identified as a carcinogen or potential carcinogen by ACGIH.

### Further information

After uptake:

pain, oedema, Vomiting, shock, death

Other dangerous properties can not be excluded.

Handle in accordance with good industrial hygiene and safety practice.

---

## SECTION 12. Ecological information

### Ecotoxicity

No information available.

### Persistence and degradability

#### *Biodegradability*

The methods for determining the biological degradability are not applicable to inorganic substances.

### Bioaccumulative potential

No information available.

### Mobility in soil

No information available.

#### *Additional ecological information*

Biological effects:

Harmful effect due to pH shift.

No ecological problems are to be expected when the product is handled and used with due care and attention.

---

## SECTION 13. Disposal considerations

The information presented only applies to the material as supplied. The identification based on characteristic(s) or listing may not apply if the material has been used or otherwise contaminated. It is the responsibility of the waste generator to determine the toxicity and physical properties of the material generated to determine the proper waste identification and disposal methods in compliance with applicable regulations. Disposal should be in accordance with applicable regional, national and local laws and regulations.

# SAFETY DATA SHEET

according to the Global Harmonized System (and with all of the information required by the HPR)

Product number	632805	Version 1.4
Product name	Guava Instrument Cleaning Fluid (ICF) (100 mL) for all Guava Flow Cytometers	

---

---

## SECTION 14. Transport information

### Land transport (DOT)

UN number	UN 1814
Proper shipping name	POTASSIUM HYDROXIDE SOLUTION
Class	8
Packing group	III
Environmentally hazardous	--

### Air transport (IATA)

UN number	UN 1814
Proper shipping name	POTASSIUM HYDROXIDE SOLUTION
Class	8
Packing group	III
Environmentally hazardous	--
Special precautions for user	no

### Sea transport (IMDG)

UN number	UN 1814
Proper shipping name	POTASSIUM HYDROXIDE SOLUTION
Class	8
Packing group	III
Environmentally hazardous	--
Special precautions for user	yes
EmS	F-A S-B

---

## SECTION 15. Regulatory information

### United States of America

### Canada

This product has been classified in accordance with the hazard criteria of the Hazardous Products Regulations (HPR) and the SDS contains all the information required by the HPR.

### Notification status

TSCA: All components of the product are listed in the TSCA-inventory.

DSL: All components of this product are on the Canadian DSL

---

## SECTION 16. Other information

### Training advice

Provide adequate information, instruction and training for operators.

---



## SAFETY DATA SHEET

according to the Global Harmonized System (and with all of the information required by the HPR)

Product number

632805

Version 1.4

Product name

Guava Instrument Cleaning Fluid (ICF) (100 mL) for all Guava Flow Cytometers

---

### Labeling

*Hazard pictograms*



*Signal Word*

Warning

*Hazard Statements*

H315 Causes skin irritation.

H319 Causes serious eye irritation.

*Precautionary Statements*

Response

P302 + P352 IF ON SKIN: Wash with plenty of soap and water.

P305 + P351 + P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P313 Get medical advice/ attention.

### Key or legend to abbreviations and acronyms used in the safety data sheet

Used abbreviations and acronyms can be looked up at [www.wikipedia.org](http://www.wikipedia.org).

Revision Date 06/17/2018

---

The information contained herein is based on the present state of our knowledge. It characterizes the product with regard to appropriate safety precautions. It does not represent a warranty of any product properties and we assume no liability for any loss or injury which may result from the use of this information. Users should conduct their own investigations to determine the suitability of the information.

***All rights reserved. Millipore and the "M" Mark are registered trademarks of Merck KGaA, Darmstadt, Germany.***