Training for 150 epochs: training accuracy reaches 100 %, new-patient accuracy stays near 52 %

ECGFounder MLP head on the frozen ECGFounder embedding; patient-grouped 5-fold cross-validation (repeat 0, 837 ECGs, original labels).
Lines: mean of the 5 folds; shaded: lowest to highest fold. Accuracy measured after each epoch with dropout off.