SeizeLUT: A Self-Testing, Multiplier-Free LUT Network for Multimodal Wearable Seizure Detection on an FPGA

[Authors]
Department of Electronics and Communication Engineering, National Institute of Technology Warangal, India

***Abstract*—Wearable seizure detectors must run for days on a small battery and, being medical devices, must not fail silently. We present SeizeLUT, a seizure detector whose classifier is a Differentiable Weightless Neural Network (DWN) of 50 six-input lookup tables (LUTs) that map one-to-one onto FPGA LUTs, fed by a bit-exact, multiplier-free integer front end for behind-the-ear EEG, ECG, EMG and motion. On the official split of the SeizeIT2 dataset (125 patients, 11,626 hours), the final network reaches an event-F1 of 0.248 ± 0.016 with 0.8 false alarms per day on 29 held-out patients, above a TreeLUT model that needs 6.7 times more LUTs, a logic-gate network and the ChronoNet baseline, in 398 LUTs and no DSP or block RAM after Vivado implementation (6.7 and 13.2 times fewer LUTs than TreeLUT and a logic-gate network, and the fastest of the three at 468 MHz). A pre-registered two-stage variant that confirms each alarm against the post-ictal period with a second LUT network detects 73.3 % of seizures at 10.9 false alarms per hour (published SVM: 71.1 % at 11.0), and the whole cascade fits in about 1,200 LUTs. Because the classifier is a set of truth tables, every single hardware fault can be enumerated: of 3,592 single faults, 2,890 change at least one clinical decision on 8,714 hours of training data, a built-in self-test fed with 1,024 consecutive recorded windows detects only 50.7 % of them, and 255 test vectors selected for coverage detect all of them; the on-chip self-test takes 10.3 µs. On a Kria KV260 the classifier runs in the fabric behind an AXI4-Lite interface with fault injection, and the integer front end runs on the Arm cores; the design is verified bit-exact on 431,960 windows at two operating points and on every injected fault.**

I. Introduction

Patients are often unaware of their seizures, so diaries undercount them [1], and wearable detection is clinically recommended today only for tonic-clonic seizures [2]. Behind-the-ear (bte) EEG is discreet but sees two channels; adding ECG, EMG and motion helps [3], [4], but the detectors that fuse them are floating-point software or multiply-accumulate networks on microcontrollers [5]. Two requirements of a wearable medical device are rarely addressed together: a classifier cheap enough for a tiny always-on fabric, and a way to know that the hardware still computes what was validated, since a silent fault turns into missed seizures or false alarms. Soft errors in FPGA configuration memory are a known risk for neural-network accelerators [6]–[8], and generic scrubbing [9] says nothing about whether a fault matters clinically.

LUT-native networks replace weighted sums by lookup tables: each neuron is a Boolean function of a few input bits that maps onto one FPGA LUT. DWNs [10], logic-gate networks [11] and LUT networks [12] train such networks by gradient descent, and tree ensembles map to LUTs in the same spirit [13]. They have been applied to ECG arrhythmia [14] and EEG classification [15], but not to seizure detection, with its extreme class imbalance (0.12 % seizure time), patient variability, missing sensors and event-level clinical scoring. Their small, explicit truth tables also make a property possible that weighted networks lack: the complete set of single hardware faults can be enumerated and judged by its clinical effect.

Contributions. (i) The first LUT-native seizure detector and the first benchmark of LUT-native models (DWN, logic-gate network, TreeLUT) against standard machine learning on the official SeizeIT2 split with SzCORE event scoring [16], with distillation from a tree ensemble to cut false alarms and modality dropout for missing sensors. (ii) A bit-exact, multiplier-free integer front end for bte-EEG, ECG, EMG and motion. (iii) A fully LUT-native post-ictal verification cascade that detects more seizures than the dataset paper's SVM at a slightly lower false-alarm rate, in about 1,200 LUTs. (iv) *Self-testing*: an exhaustive single-fault analysis of the deployed classifier on 8,714 hours of data, a coverage-driven choice of on-chip test vectors that detects every decision-relevant fault with 255 vectors, and fault-injection hardware that checks this on the device. (v) A verified Kria KV260 implementation, and an independent reproduction of the published ChronoNet baseline, whose released code selects its model on the test patients.

II. Related Work

Table I compares hardware seizure detectors. ASICs reach nJ–µJ per classification with patient-specific SVMs and tree ensembles on implanted or scalp EEG [17]–[19]; an SNN on a small iCE40 FPGA detects seizures on four scalp channels [20]; transformers and fusion networks run on microcontrollers [5], [21]. Most use EEG only and epoch-level metrics; to our knowledge none reports what a hardware fault does to its decisions. On SeizeIT2, the dataset paper reports SVM and ChronoNet [22] baselines [23]; tree ensembles won the ICASSP 2023 wearable-EEG challenge [24], [25]. The only weightless seizure work is a WiSARD study on four patients [26]; for DWNs the thermometer encoder can dominate the cost [27]. Resilience studies of neural accelerators inject faults statistically [7], [8] or harden the most vulnerable layers [6]; with LUT networks the fault space is small enough to enumerate completely.

TABLE I

HARDWARE SEIZURE DETECTORS AS REPORTED BY EACH WORK. PS: PATIENT-SPECIFIC TRAINING OR CALIBRATION; PI: PATIENT-INDEPENDENT, SCORED ON HELD-OUT PATIENTS. DETECTION METRICS DIFFER BETWEEN WORKS (EVENT-, EPOCH- OR SAMPLE-BASED) AND ARE NOT DIRECTLY COMPARABLE. ‡FROM THE COMPARISON TABLE OF [19]; N.R.: NOT REPORTED; *VIVADO VECTORLESS ESTIMATE (17 MW DYNAMIC AT 400 MHZ, ONE WINDOW PER CLOCK); FAULTS: SHARE OF DECISION-RELEVANT SINGLE FAULTS DETECTED BY THE SELF-TEST.

Work	Platform	Data (patients), eval.	Inputs	Detection (as reported)	Cost	Energy/inf.	Faults
Altaf'15 [17]	ASIC 180 nm	CHB-MIT (14)‡, PS	EEG, 16 ch	sens 95.7 %, spec 98 %‡	1.56 mm^2/ch‡	2.73 µJ‡	n.r.
Shoaran'18 [18]	ASIC 65 nm	iEEG (26), PS	iEEG	F1 87.9 % (block split)	1.0 mm^2	41.2 nJ	n.r.
NeuralTree'22 [19]	ASIC 65 nm	CHB-MIT (24)+iEEG (6), PS	EEG/iEEG, $\leq$64 ch	sens 95.6 %, spec 96.8 %	3.48 mm^2	0.227 µJ	n.r.
Vittimberga'25 [28]	Artix-7 (synth.)	CHB-MIT (5), PS	EEG, 4 ch	sens >98 % (samples)	192 LUT, 3 DSP	n.r.	n.r.
Busia'25 [20]	iCE40UP5K	CHB-MIT (9), PS	EEG, 4 ch	100 % events, 0.3 FA/h	3.3k LC, 2 DSP, 14 BRAM	4.55 nJ (core)	n.r.
EEGformer'24 [21]	GAP9 MCU	CHB-MIT (8), PS	EEG, 4 ch	73 % events, 0.15 FA/h	150 kB, 14.7 MOPS	0.31 mJ	n.r.
BrainFuseNet'24 [5]	GAP9 MCU	PEDESITE (6), PS	EEG 4 ch, PPG, ACC	95 % events, 0.58 FA/day	2.29 M MAC	0.11 mJ	n.r.
SeizeIT2 [23]	software	SeizeIT2 (29 test), PI	bte-EEG, 2 ch	SVM 71.1 % at 11.0 FA/h	–	–	–
SeizeLUT (ours)	XCK26 (KV260)	SeizeIT2 (29 test), PI	bte-EEG 2 ch, ECG, EMG, ACC, GYR	event-F1 0.248, 0.8 FA/day; two-stage 73.3 % at 10.9 FA/h	398 LUT, 0 DSP/BRAM (two-stage ≈1,200 LUT)	≈43 pJ*	**100 %**

III. DATASET AND PROTOCOL

Data. SeizeIT2 (OpenNeuro ds005873 v1.1.0) [23]: 125 patients with focal epilepsy from five European monitoring units, 2,850 recordings, 11,626 hours of two-channel bte-EEG, single-lead ECG and EMG at 256 Hz and accelerometer and gyroscope at 25 Hz. Our audit found 883 seizures (702 in the 96 training patients, 181 in the 29 test patients). *Split and tuning.* Official split: training sub-001–096, test sub-097–125. Every choice is made by five-fold patient-wise cross-validation inside the training patients; each final model is trained on all 96 and scored once per seed on the test patients. *Scoring.* SzCORE event scoring [16] (any overlap, 30 s pre- and 60 s post-ictal tolerance, events closer than 90 s merged), averaged over patients, with false alarms per 24 h. A decision is made every second; an alarm is raised when at least k of the last 10 decisions are positive. Per-patient paired Wilcoxon tests; an extension is kept only if $p < 0.05$ in its favour.

IV. METHOD

A. Integer multimodal front end

Samples arrive as int16. Every filter is a second-order section with coefficients rounded to at most five signed powers of two and second-order error feedback. Every second, 2-s windows give per-channel and cross-channel EEG magnitude, line length, mobility, zero crossings, extrema and band magnitudes; ECG heart rate and RR variability in the log domain; EMG magnitude, line length, zero crossings and a band ratio; accelerometer and gyroscope magnitude, jerk, activity and band magnitude; presence bits; and, for most features, the difference from a causal 64-s average: 117 features without a multiplier or divider.

B. Thermometer encoding and DWN

Each of the 40 selected features (gain ranking of a LightGBM model inside each training set) is compared with 8 thresholds at training quantiles, $b_{f,i} = [x_f > t_{f,i}]$. A layer of 50 LUT6 neurons reads these bits through a learnable mapping and is trained with extended finite differences [10]; the score is the popcount of the seizure group minus that of the background group. Of seven families of extensions tested on the training folds (stateful networks, encoders, label-free adaptation, distillation, robust losses, modality handling, interpretation), two were adopted: *distillation* from a LightGBM teacher (soft targets, weight 0.5, temperature 2), and *modality dropout*, which removes ECG, EMG and motion independently with probability 0.2 per training window.

C. Post-ictal verification

Stage 1 is the final DWN at a sensitive threshold (-7, $k = 10$). Every candidate alarm is checked 96 s later by a second 50-LUT DWN on the 24 most informative of 101 statistics: 16-s ictal and 64-s post-ictal means of 45 features, their changes from the preceding 128 s, the stage-1 score and the candidate's duration. All windows are powers of two, so every mean is a sum and a shift. The settings were chosen on the training folds and pre-registered before a single test evaluation.

D. Self-testing

Fault model. The fabric classifier is 146 comparators, 50 LUT6 and an adder tree. We enumerate every single fault in the part that the network adds: each of the $50 \times 64 = 3{,}200$ truth-table bits flipped (an upset in configuration memory), and each comparator output and LUT output stuck at 0 or 1 ($292 + 100$), 3,592 faults in all. *Relevance.* Each fault is applied to all 31.4 million windows (8,714 h) of the 96 training patients. A truth-table flip at LUT j, entry a changes the score by ± 1 exactly on the windows that address (j, a), so its effect is computed in one pass; stuck-at faults are recomputed per window. A fault is *decision-relevant* if it changes at least one decision at the operating threshold, and *alarm-relevant* if it changes an alarm. *Vector selection.* A test vector detects a fault when the faulty score differs from the golden one. Since every window addresses exactly one entry per LUT, at least 64 vectors are needed when a LUT uses all its entries; we cover the addressed entries rarest first, each time taking, among the training windows that address the entry, the one that covers most uncovered entries, then add windows for stuck-at faults still undetected. *Hardware.* The built-in self-test (BIST) streams the vectors from block RAM at one per clock and compares score, decision and alarm in the fabric; fault-injection registers force any comparator or LUT output

TABLE II
FIVE-FOLD CROSS-VALIDATION ON THE 96 TRAINING PATIENTS (SZCORE, MEAN OVER PATIENTS; BEST-F1 OPERATING POINT; ESTIMATED HARDWARE COST).

Model	AUROC	F1	Sens	FA/24h	Cost
LightGBM	0.870	0.292	0.327	3.6	≈30.6k LUTs
CatBoost	0.863	0.286	0.337	3.0	≈27.7k LUTs
TreeLUT (5-bit leaves)	0.857	0.266	0.329	4.2	≈6.4k LUTs
DWN, 50 LUT6	0.840	0.211	0.315	2.4	≈960 LUTs
DWN + distillation (final)	0.851	0.229	0.267	0.9	≈870 LUTs
DWN + modality dropout	0.849	0.188	0.275	1.9	≈945 LUTs
Random forest	0.852	0.198	0.216	0.5	≈323k LUTs
Logic-gate network	0.843	0.171	0.330	4.9	≈7.5k LUTs
INT8 MLP	0.830	0.147	0.225	7.8	9,600 mult.

or flip any truth-table entry, so the coverage can be measured on the device.

V. RESULTS

A. Baseline reproduction

On the 29 test patients our port of ChronoNet with the released protocol detects 74.5 % of seizures at 22.6 false alarms per hour (published: 84.2 % at 100.5) and matches the published per-recording AUPR (0.109 against 0.104). The released code selects the training epoch on the test patients; a clean run (last epoch) gives 67.1 % at 11.9 false alarms per hour, and an event-F1 of 0.027 at 106 false alarms per day. The clean run is our ChronoNet reference.

B. Model comparison and missing sensors

Table II compares the model families on cross-validation. The DWN is the most compact model, with one of the lowest false-alarm rates, and beats the INT8 MLP on every metric; it remains below gradient-boosted trees ($p = 0.004$ against LightGBM). Distillation cut false alarms from 2.4 to 0.9 per day ($p = 3 \times 10^{-7}$) at a slightly higher event-F1 (0.229, $p = 0.26$) and became the final model. The base loses most of its accuracy without ECG (event-F1 0.092); modality dropout keeps it working (0.172, $p = 0.005$; without motion 0.192 against 0.136, $p = 0.021$). Out of fold, the base detects 88 % of focal to bilateral tonic-clonic seizures but 17 % of focal impaired-awareness ones.

C. Held-out test patients

The design was frozen before any of our models was scored on a test patient (Table III). The final DWN reaches an event-F1 of 0.248 ± 0.016 with 0.8 false alarms per day, above its cross-validation value: the gain from distillation holds on unseen patients (base 0.172). It is above TreeLUT (0.233) at a seventh of the LUTs, above the logic-gate network (0.191) and the INT8 MLP (0.155), level with random forest and CatBoost, and below LightGBM (0.308, $p = 0.039$). Against clean ChronoNet it is better per patient ($p = 0.027$): ChronoNet detects more seizures (79 %) but raises over a hundred times more false alarms.

TABLE III
HELD-OUT TEST PATIENTS (29): MEAN ± STD OVER 5 SEEDS, SZCORE EVENT METRICS AT THE BEST-F1 OPERATING POINT CHOSEN ON THE TRAINING FOLDS; p: PAIRED WILCOXON ON PER-PATIENT EVENT-F1 AGAINST CLEAN CHRONONET.

Model	F1	Sens	FA/24h	p vs ChronoNet	Cost
DWN + distillation (final)	0.248±0.016	0.303	0.8	0.03	895 LUTs
DWN base	0.172±0.044	0.340	2.7	0.01	980 LUTs
DWN + modality dropout	0.242±0.030	0.272	0.4	0.03	951 LUTs
LightGBM	0.308±0.013	0.331	0.6	0.03	30,750 LUTs
XGBoost	0.294±0.014	0.301	0.4	0.06	48,295 LUTs
CatBoost	0.254±0.009	0.330	1.2	0.03	27,720 LUTs
Random forest	0.246±0.003	0.245	0.2	0.3	346,576 LUTs
Logistic regression	0.123±0.002	0.255	5.1	0.6	117 mult.
TreeLUT (5-bit leaves)	0.233±0.044	0.292	1.0	0.02	6,330 LUTs
INT8 MLP	0.155±0.008	0.249	1.7	0.02	9,600 mult.
Logic-gate network	0.191±0.021	0.291	1.4	0.03	7,452 LUTs
ChronoNet (clean)	0.027	0.790	106.3	–	CNN+GRU




Fig. 1. Seizures detected against false alarms per hour on the 29 test patients (dataset paper's metrics). Curves: threshold sweeps of the frozen models (mean of 5 seeds); dots: operating points chosen on the training folds; stars: two-stage detector with post-ictal verification (pre-registered); open markers: published values.

D. Comparison with the dataset paper

Fig. 1 uses the dataset paper's metrics (any-overlap sensitivity and false alarms per hour). At a matched 1.2 false alarms per hour the final DWN detects 43.2 % of seizures against 34.8 % for ChronoNet, and its window AUROC is 0.883 against 0.746. The pre-registered two-stage detector reaches 73.3 ± 1.4 % at 10.9 ± 0.6 false alarms per hour over five seeds (training folds: 81.9 % at 9.6), against 71.1 % at 11.0 for the published SVM; verification adds 3.6–6.5 points over the DWN alone at the same rate. This analysis was added after the final test and is a second look at the same patients; it suits seizure logging (96-s decision delay) rather than instant alarms.

E. Hardware cost

Table IV gives the cost of the three LUT-native classifiers, each generated from its frozen final model and checked bit-exact against it in simulation, implemented out of context by

TABLE IV
COST OF THE LUT-NATIVE CLASSIFIERS (XCK26; INPUT REGISTERS, COMPARATORS, NETWORK, SCORE, DECISION AND ALARM LOGIC). VIVADO 2026.1 OUT OF CONTEXT, 2.5-NS CONSTRAINT, VECTORLESS POWER AT 400 MHZ; YOSYS `SYNTH_XILINX`.

	Vivado				Yosys		
Model	LUTs	FFs	Fmax	Dyn. P	LUTs	DSP	Test F1
DWN + distillation (ours)	**398**	454	**468 MHz**	**17 mW**	405	0	0.248
TreeLUT (5-bit leaves)	2,677	1,461	294 MHz	88 mW	2,721	0	0.233
Logic-gate network	5,249	2,208	154 MHz	523 mW	5,758	0	0.191
Post-ictal verifier (stage 2)	–	–	–	–	799	0	–

TABLE V
SINGLE FAULTS OF THE FABRIC CLASSIFIER ON THE 96 TRAINING PATIENTS (8,714 H; FROZEN THRESHOLD) AND THE DECISION-RELEVANT FAULTS DETECTED (%) BY THREE BIST VECTOR SETS: 1,024 CONSECUTIVE WINDOWS OF ONE RECORDING, 1,024 RANDOM TRAINING WINDOWS AND THE 255 COVERAGE-SELECTED WINDOWS.

Fault	All	Obs.	Dec.	Alarm	Consec.	Random	**Selected**
truth-table bit flipped	3,200	2,750	2,498	1,674	43.6	79.4	**100.0**
comparator output stuck at 0/1	292	292	292	289	94.5	100.0	**100.0**
LUT output stuck at 0/1	100	100	100	100	100.0	100.0	**100.0**
all	3,592	3,142	2,890	2,063	50.7	82.2	**100.0**

Vivado 2026.1 on the XCK26 under the same 2.5-ns clock constraint, and synthesised by Yosys (`synth_xilinx`) as a second, open flow. The DWN needs 398 LUTs and 454 flip-flops with no DSP or block RAM, 6.7 times fewer LUTs than TreeLUT and 13.2 times fewer than the logic-gate network, and it is also the fastest: 468 MHz against 294 and 154 MHz. With default switching activity Vivado estimates 17 mW of dynamic power at 400 MHz for it, against 88 and 523 mW. The post-ictal verifier, with 24-bit comparators on window sums, needs 799 LUTs (Yosys); integer and floating-point verifiers agree on 99.998 % of the 118,975 training candidates, so the complete two-stage detector costs about 1,200 LUTs and no multiplier. In the KV260 design the whole core (AXI4-Lite interface, BIST with its vector ROM in 9 block RAMs, fault injection, debug taps, demo I/O) uses 1,908 LUTs and the full design 2,345 LUTs, 2.0 % of the device; timing closes at 100 MHz with 3.3 ns of setup slack. The fault-injection and debug logic, needed only in test builds, adds 298 LUTs (Yosys).

F. Faults and self-test

Of the 3,592 single faults, 3,142 change at least one score on the training data; 450 truth-table entries (14 %) are never addressed in 8,714 hours, so a flip there is harmless on such data. At the frozen threshold 2,890 faults are decision-relevant and 2,063 alarm-relevant (3,101 decision-relevant at the sensitive threshold). Single faults are mostly mild but not negligible: the worst truth-table flip changes 0.12 % of all decisions at the frozen threshold and the worst stuck-at fault 0.69 %; at the sensitive threshold a single stuck-at fault changes 16.3 %.

Table V and Fig. 2 show how many decision-relevant faults a BIST detects. Stuck-at faults on wires are easy: almost any




Fig. 2. Decision-relevant single faults (frozen threshold, training data) detected by the first m BIST vectors.




Fig. 3. SeizeLUT on the Kria KV260. The fault-injection registers and the self-test let the device check that the fabric still computes the validated classifier.

data exercises them. The hard faults are flips of truth-table entries that the data rarely address, exactly the configuration-memory upsets of an FPGA. 1,024 consecutive windows of one recording around a seizure, the natural choice of a golden test, detect 50.7 % (58.9 % of the alarm-relevant ones, and 43.6 % of the decision-relevant truth-table flips): consecutive windows revisit the same entries. 1,024 random training windows reach 82.2 %. The 255 coverage-selected windows detect all observable faults, decision- and alarm-relevant at every operating point. They now open the BIST ROM, which runs 1,024 vectors in 1,028 cycles (10.3 µs at 100 MHz), so the device can test itself after every one-second window at a duty cycle of 0.001 %. In simulation, each of the 3,592 faults injected through the fault registers gives exactly the BIST error count and first failing vector that the software model predicts. [board: the same campaign on the KV260: faults injected, detected, hardware/software mismatches, time per fault.]

G. KV260 implementation

On a Kria KV260 (Zynq UltraScale+ XCK26, Fig. 3) the integer front end runs in C on the Arm Cortex-A53 cores and the classifier runs in the fabric behind an AXI4-Lite interface

at 100 MHz, with a frozen and a programmable operating point, read-back of the 146 comparator bits and 50 LUT outputs, the BIST and fault injection. Verification before the board: the C front end, batch and second by second, reproduces all 117 features and both artefact flags of all 431,960 windows of the 10 exported test recordings exactly; the core, driven through its registers as the board software drives it, reproduces score, decision, alarm, comparator bits and LUT outputs on all 431,960 windows at two operating points (863,920 window checks, Verilator), with a latency of 3 cycles. [board: on-board replay of the 431,960 windows, measured latency and throughput, front-end time per hour of signal on the A53, SOM power from the INA260 monitor.]

VI. Discussion and Limitations

Accuracy per LUT favours the DWN by a wide margin, and the thermometer encoder, not the network, dominates its cost, as also measured by [27]; it also dominates the verifier's cost through its wide comparators. Self-testing is cheap for LUT networks because their fault space is small and their behaviour is exactly computable, which a weighted network with thousands of multipliers does not offer. The analysis covers single faults in the network part (comparator outputs, truth tables, LUT outputs), not the adder tree, the AXI logic or the comparator constants, and relevance is measured on training data. Focal seizures remain mostly undetected, consistent with the clinical evidence [2]. Results come from retrospective recordings of one dataset.

VII. Conclusion

A 50-LUT weightless network on integer multimodal features detects seizures across patients without multipliers in 398 LUTs, reaching an event-F1 of 0.248 with 0.8 false alarms per day on 29 held-out patients, and a two-stage version in about 1,200 LUTs detects more seizures than the dataset paper's SVM. Its truth tables make complete fault analysis possible: 255 selected self-test vectors detect every single fault that changes a decision, and the device can run them after every window.

Reproducibility

Code, configurations, decision log, export packages, the Verilog generator, the KV260 kit and the fault list are released with the paper. Dataset: SeizeIT2, DOI 10.18112/openneuro.ds005873.v1.1.0.

References

[1] C. E. Elger and C. Hoppe, "Diagnostic challenges in epilepsy: seizure under-reporting and seizure detection," *The Lancet Neurology*, vol. 17, no. 3, pp. 279–288, 2018.

[2] S. Beniczky, S. Wiebe, J. Jeppesen, W. O. Tatum, M. Brázdil, Y. Wang, S. T. Herman, and P. Ryvlin, "Automated seizure detection using wearable devices: A clinical practice guideline of the International League Against Epilepsy and the International Federation of Clinical Neurophysiology," *Epilepsia*, vol. 62, no. 3, pp. 632–646, 2021.

[3] K. Vandecasteele, T. De Cooman, C. Chatzichristos, E. Cleeren, L. Swinnen, J. Macea Ortiz, S. Van Huffel, M. Dümpelmann, A. Schulze-Bonhage, M. De Vos, W. Van Paesschen, and B. Hunyadi, "The power of ECG in multimodal patient-specific seizure monitoring," *Epilepsia*, vol. 62, no. 10, pp. 2333–2343, 2021.

[4] J. Zhang, L. Swinnen, C. Chatzichristos, V. Broux, R. Proost, K. Jansen, B. Mahler, N. Zabler, N. Epitashvilli, M. Dümpelmann, A. Schulze-Bonhage, E. Schriewer, U. Ermis, S. Wolking, F. Linke, Y. Weber, M. Symmonds, A. Sen, A. Biondi, M. P. Richardson, A. Sulaiman I, A. I. Silva, F. Sales, G. Vértes, W. Van Paesschen, and M. De Vos, "Multimodal wearable EEG, EMG and accelerometry measurements improve the accuracy of tonic-clonic seizure detection," *Physiological Measurement*, vol. 45, no. 6, 2024.

[5] T. M. Ingolfsson, X. Wang, U. Chakraborty, S. Benatti, A. Bernini, P. Ducouret, P. Ryvlin, S. Beniczky, L. Benini, and A. Cossettini, "BrainFuseNet: Enhancing wearable seizure detection through EEG-PPG-accelerometer sensor fusion and efficient edge deployment," *IEEE Transactions on Biomedical Circuits and Systems*, vol. 18, no. 4, pp. 720–733, 2024.

[6] F. Libano, B. Wilson, J. Anderson, M. J. Wirthlin, C. Cazzaniga, C. Frost, and P. Rech, "Selective hardening for neural networks in FPGAs," *IEEE Transactions on Nuclear Science*, vol. 66, no. 1, pp. 216–222, 2019.

[7] G. Li, S. K. S. Hari, M. Sullivan, T. Tsai, K. Pattabiraman, J. Emer, and S. W. Keckler, "Understanding error propagation in deep learning neural network (DNN) accelerators and applications," in *Proceedings of the International Conference for High Performance Computing, Networking, Storage and Analysis (SC)*, 2017.

[8] B. Reagen, U. Gupta, L. Pentecost, P. Whatmough, S. K. Lee, N. Mulholland, D. Brooks, and G.-Y. Wei, "Ares: A framework for quantifying the resilience of deep neural networks," in *Proceedings of the 55th Annual Design Automation Conference (DAC)*, 2018, pp. 17:1–17:6.

[9] *UltraScale Architecture Soft Error Mitigation Controller, LogiCORE IP Product Guide (PG187), v3.1*, AMD, 2025.

[10] A. T. L. Bacellar, Z. Susskind, M. Breternitz Jr., E. John, L. K. John, P. M. V. Lima, and F. M. G. França, "Differentiable weightless neural networks," in *Proceedings of the 41st International Conference on Machine Learning (ICML)*, ser. Proceedings of Machine Learning Research, vol. 235. PMLR, 2024, pp. 2277–2295. [Online]. Available: https://proceedings.mlr.press/v235/bacellar24a.html

[11] F. Petersen, C. Borgelt, H. Kuehne, and O. Deussen, "Deep differentiable logic gate networks," in *Advances in Neural Information Processing Systems (NeurIPS)*, vol. 35, 2022. [Online]. Available: https://openreview.net/forum?id=vF3WefcoePW

[12] Y. Umuroglu, Y. Akhauri, N. J. Fraser, and M. Blott, "LogicNets: Co-designed neural networks and circuits for extreme-throughput applications," in *30th International Conference on Field-Programmable Logic and Applications (FPL)*, 2020, pp. 291–297.

[13] A. Khataei and K. Bazargan, "TreeLUT: An efficient alternative to deep neural networks for inference acceleration using gradient boosted decision trees," in *Proceedings of the 2025 ACM/SIGDA International Symposium on Field Programmable Gate Arrays (FPGA '25)*. ACM, 2025.

[14] W. Mommen, L. Keuninckx, P. Detterer, A. Colpaert, and P. Wambacq, "Inter-patient ECG arrhythmia classification with LGNs and LUTNs," 2026.

[15] S. Dharia, S. Smith, and C. Valderrama, "Differentiable logic gate networks for low-latency EEG classification on edge devices," in *Proceedings of the 39th Canadian Conference on Artificial Intelligence*, ser. Proceedings of Machine Learning Research, vol. 318. PMLR, 2026, pp. 377–391. [Online]. Available: https://proceedings.mlr.press/v318/dharia26a.html

[16] J. Dan, U. Pale, A. Amirshahi, W. Cappelletti, T. M. Ingolfsson, X. Wang, A. Cossettini, A. Bernini, L. Benini, S. Beniczky, D. Atienza, and P. Ryvlin, "SzCORE: Seizure community open-source research evaluation framework for the validation of electroencephalography-based automated seizure detection algorithms," *Epilepsia*, 2024.

[17] M. A. Bin Altaf, C. Zhang, and J. Yoo, "A 16-channel patient-specific seizure onset and termination detection SoC with impedance-adaptive transcranial electrical stimulator," *IEEE Journal of Solid-State Circuits*, vol. 50, no. 11, pp. 2728–2740, 2015.

[18] M. Shoaran, B. Allahgholizadeh Haghi, M. Taghavi, M. Farivar, and A. Emami-Neyestanak, "Energy-efficient classification for resource-constrained biomedical applications," *IEEE Journal on Emerging and Selected Topics in Circuits and Systems*, vol. 8, no. 4, pp. 693–707, 2018.

[19] U. Shin, C. Ding, B. Zhu, Y. Vyza, A. Trouillet, E. Revol, S. Lacour, and M. Shoaran, "NeuralTree: A 256-channel 0.227-μJ/class versatile neural activity classification and closed-loop neuromodulation SoC,"

IEEE Journal of Solid-State Circuits, vol. 57, no. 11, pp. 3243–3257, 2022.

[20] P. Busia, G. Leone, A. Matticola, L. Raffo, and P. Meloni, "Wearable epilepsy seizure detection on FPGA with spiking neural networks," *IEEE Transactions on Biomedical Circuits and Systems*, vol. 19, no. 6, 2025.

[21] P. Busia, A. Cossettini, T. M. Ingolfsson, S. Benatti, A. Burrello, V. J. B. Jung, M. Scherer, M. A. Scrugli, A. Bernini, P. Ducouret, P. Ryvlin, P. Meloni, and L. Benini, "Reducing false alarms in wearable seizure detection with EEGformer: A compact transformer model for MCUs," *IEEE Transactions on Biomedical Circuits and Systems*, vol. 18, no. 3, pp. 608–621, 2024.

[22] S. Roy, I. Kiral-Kornek, and S. Harrer, "ChronoNet: A deep recurrent neural network for abnormal EEG identification," in *Artificial Intelligence in Medicine (AIME 2019)*, ser. Lecture Notes in Computer Science, vol. 11526. Springer, 2019.

[23] M. Bhagubai, C. Chatzichristos, L. Swinnen, J. Macea, J. Zhang, L. Lagae, K. Jansen, A. Schulze-Bonhage, F. Sales, B. Mahler, Y. Weber, W. Van Paesschen, and M. De Vos, "SeizeIT2: Wearable dataset of patients with focal epilepsy," *Scientific Data*, vol. 12, p. 1228, 2025.

[24] M. Bhagubai, L. Swinnen, E. Cleeren, W. Van Paesschen, M. De Vos, and C. Chatzichristos, "Towards automated seizure detection with wearable EEG – grand challenge," *IEEE Open Journal of Signal Processing*, vol. 5, pp. 717–724, 2024.

[25] I. Al-Hussaini and C. S. Mitchell, "SeizFt: Interpretable machine learning for seizure detection using wearables," *Bioengineering*, vol. 10, no. 8, p. 918, 2023.

[26] K. de Aguiar, F. M. G. França, V. C. Barbosa, and C. A. D. Teixeira, "Early detection of epilepsy seizures based on a weightless neural network," in *37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, 2015, pp. 4470–4474.

[27] M. Mecik and M. Kumm, "Implementation and analysis of thermometer encoding in DWN FPGA accelerators," in *Asilomar Conference on Signals, Systems, and Computers*, 2025.

[28] A. Vittimberga, G. Nicolini, and G. Scotti, "Real-time multi-channel epileptic seizure detection exploiting an ultra-low-complexity algorithm–hardware co-design approach," *Sensors*, vol. 25, no. 22, p. 6889, 2025.