

Two-Point Parked Lock-In

Master Reference

Acquisition methods · timing breakdown · sensitivity
open defects · work log

Project NV compact magnetometer, RFSoc4x2 + QICK-DAWG

Repository `nv_magnetometer_project`, branch `main`

Sessions covered 2026-08-04, 2026-08-06, 2026-08-14, 2026-08-17

Written 2026-08-17

Supersedes

`docs/2026-08-06_twopoint_timing/TIMING_AND_NOISE_ANALYSIS.md` and
`docs/2026-08-14_twopoint_methods/TWOPOINT_METHODS_AND_TIMING.md`, both of
which remain in the repository as the record of what was believed when. Chapter 13
lists every disagreement.

Regenerate every number `python scripts/analyze_twopoint_0814.py`

The three findings, in one paragraph each

Rate. An `acquire()` call costs 10–11.4 ms of host round-trip whatever the program does, and the FPGA work runs underneath it. Halving the readout window cut the FPGA work by $4\times$ and changed the rate by 0.4%. The averaged mode therefore sits at ~ 88 Hz at any setting, and averaging is *free* up to the call floor.

Sensitivity. The reference-photodiode fix bought $4.1\times$: the burst-mode noise floor went from 158 to **38 nT/ $\sqrt{\text{Hz}}$** . Burst is the best operating point available today.

Bandwidth. Streaming delivers **1041.7 Hz** with a gapless, duplicate-free rep axis — the 1 kHz target is met, on two machines. Its noise floor is above the burst floor by a flat factor across the band, but the size of that factor moved from $\sim 6\times$ (2026-08-14) to **1.3–1.6 $\times$** (2026-08-17) and is not yet pinned to a cause.

And one about how this document is used. Chapter 9 §9.3.3 records a fix that was reasoned from the source, verified against a mock, shipped, and broke two of the three acquisition methods two hours later. It is written up at length on purpose. The rig is the arbiter.

Contents

1	How to read this document	4
1.1	What this is	4
1.2	Where the numbers come from	4
1.3	The three findings that matter most	4
1.4	Status at a glance	5
2	The measurement, from photons to Δf	6
2.1	Two parked frequencies on one flank pair	6
2.1.1	Why the difference and not one flank	6
2.1.2	The noise consequence	6
2.2	The residual first-order term, and ABBA	7
2.3	The two sensitivity conventions, fixed once	7
2.4	What is not being claimed	7
3	What all three methods share	8
3.1	The clock everything is counted in	8
3.2	The FPGA program: one rep, step by step	8
3.2.1	The per-slot budget, and the 1.6 % that is left over	9
3.2.2	How many slots there are	10
3.2.3	The one timing formula everything else derives from	10
3.3	What the accumulated buffer holds	10
3.4	Recovering the exact readout time from the data	11
3.5	The host RPC chain	11
3.6	Python outside the call	12
4	Method A — Averaged mode	14
4.1	Mechanics	14
4.2	Timing budget, component by component	14
4.2.1	The three side by side — the decisive comparison	15
4.3	The CSV flush, and why it was fixed	15
4.4	Why you no longer enter a reps-per-batch	16
4.4.1	The argument	16
4.4.2	Worked, on the 2026-08-17 run	16
4.4.3	You can still set it by hand	17
4.5	What to set	17
4.6	Failure modes and what warns you	17
4.7	When to use it	17
5	Method B — Burst mode	18
5.1	Mechanics	18
5.2	Timing budget, component by component	18
5.2.1	Rate against burst length	19
5.3	Three artefacts, and what each one is	19
5.4	What to set	21
5.5	When to use it	21

6	Method C — Streaming mode	22
6.1	Mechanics	22
6.2	Timing budget, component by component	23
6.3	Every parameter of the streaming cell, one at a time	23
6.3.1	STREAM_TARGET_HZ and the bin	23
6.3.2	STREAM_DURATION_SEC, total_reps and cfg.reps — all the same number	24
6.3.3	poll_timeout_s — and why it is currently unreachable	24
6.3.4	Stride, packet size and avg_maxlen	24
6.4	Verification: the guarantees, checked on the data	25
6.5	The buffer-overflow constraint	25
6.6	The PL droop	26
6.7	When to use it	26
7	Timing, in full	27
7.1	The evidence for a per-call floor	27
7.1.1	Argument 1 — the slope is not 1	27
7.1.2	Argument 2 — no serial model fits at all	28
7.1.3	Argument 3 — the wobble is entirely in the call	28
7.2	The corrected model	28
7.2.1	Predicted against measured	28
7.3	Free averaging	29
7.4	Knob by knob	29
7.5	The 1 kHz question, closed	30
8	Sensitivity and noise	31
8.1	The readout-window optimum	31
8.2	Measured sensitivity, per method	32
8.2.1	The photodiode fix: $4.1\times$	33
8.2.2	A warning about comparing these rows	33
8.3	Averaging does not buy $\sqrt{N}$	33
8.4	The noise is not photon shot noise	34
8.5	Mains pickup and aliasing	35
9	Open defects	36
9.1	Burst mode replays about half of every run	36
9.1.1	The evidence	36
9.1.2	The 2026-08-06 root cause, and why it is now disproved	36
9.1.3	Mitigation, which is in place	37
9.2	Streaming’s noise floor is about $6\times$ the burst floor	37
9.2.1	The evidence	37
9.2.2	2026-08-17: still there, but far smaller	38
9.2.3	What has been ruled out, with the measurement	39
9.2.4	What has been done about it	40
9.3	An interrupted cell wedges the board, and a kernel restart does not clear it	40
9.3.1	The mechanism	40
9.3.2	Reproduced offline	41
9.3.3	The first fix, and why it had to be withdrawn	41
9.3.4	What is actually shipped	42
9.4	Averaged and burst runs used to overwrite each other’s filenames	42
9.5	Post-processing: recovering runs already recorded	43
9.5.1	A bug this found: never despike the channels separately	43

10 Streaming the multipoint lock-in	44
10.1 Why it is worth doing	44
10.2 The constraint that actually binds: buffer headroom	44
10.3 Why the reconstruction is not inline	45
10.4 What the two cells produce	45
10.5 What Step 5S checks, and why each check exists	46
11 Work log: what has been changed and pushed	47
11.1 Round 1 — 2026-08-12, against the 2026-08-06 session	47
11.2 Round 2 — 2026-08-16/17, against the 2026-08-14 session	48
11.3 Round 3 — 2026-08-17, the board wedge	48
11.3.1 Verified offline	49
11.4 Round 4 — 2026-08-17 evening, undoing Round 3’s central change	49
11.5 The notebook layout	51
11.6 Round 6 — 2026-08-18, the shot counter was never cleared	51
11.6.1 What the counter does	52
11.6.2 Which paths were affected	52
11.6.3 The fix	53
11.6.4 Verified	53
11.7 Round 7 — 2026-08-18, revert the acquisition path to 8fd303f	53
11.7.1 What went away with it	54
11.7.2 The lesson this round is actually about	54
11.8 Round 5 — 2026-08-18, every figure goes to disk	55
11.8.1 notebook_modules/figure_autosave.py	55
11.8.2 Naming	55
11.8.3 TwoPointResult.spectrum_figure()	56
11.8.4 Verified	56
11.9 The code, file by file	56
11.10 Wiki	57
12 What still needs to be done	58
12.1 Rig checks, in priority order	58
12.2 Software backlog	59
12.3 Recommended operating point today	59
13 Corrections register	60
14 Reproducing everything	62
14.1 Offline, no hardware	62
14.2 On the rig	62
14.3 Data used by this document	62

Chapter 1

How to read this document

1.1 What this is

This is the single reference for the two-point parked-frequency lock-in on the NV compact magnetometer: how each acquisition method works, exactly where the acquisition time goes, what sensitivity each method delivers, which defects are open, and what has been changed in the code so far.

It merges and supersedes two earlier documents:

- docs/2026-08-06_twopoint_timing/TIMING_AND_NOISE_ANALYSIS.md — the 2026-08-06 session: timing budget, readout-window optimum, discovery of the burst-mode replay.
- docs/2026-08-14_twopoint_methods/TWOPOINT_METHODS_AND_TIMING.md — the 2026-08-14 session: the three modes compared, the per-call floor, two open defects.

Both remain in the repository as the record of what was believed when. Where they disagree with this document, this document wins, and Chapter 13 lists every such disagreement explicitly.

1.2 Where the numbers come from

Every measured value here is regenerated by one script from committed CSV files:

```
python scripts/analyze_twopoint_0814.py
```

It writes docs/2026-08-14_twopoint_methods/tables/*.csv and figures/*.png; this document quotes those tables. Nothing is hand-typed, so re-running the script after a new session refreshes the numbers rather than invalidating them. Values that can only be measured on the rig are marked **OPEN** and named as such — there are five of them, all in Chapter 12.

1.3 The three findings that matter most

1. The acquisition rate is set by the host, not by the readout window

An `acquire()` call costs **10–11.4 ms of host round-trip whatever the program does**, and the FPGA pulse work runs *underneath* that rather than adding to it. Halving the readout window from 213 μ s to 120 μ s cut the FPGA work from 9.80 ms to 2.40 ms and changed the rate by nothing (87.5 $\rightarrow$ 86.8 Hz).

Consequences: shortening τ is not a rate lever in averaged mode; averaging more reps is *free* up to the call floor; and beating ~ 90 Hz requires amortising the call across many samples, which is what burst and streaming do. See Chapter 7.

2. The photodiode fix bought $4.1\times$, and burst mode shows it

After the reference-photodiode gain was corrected and the signal diode stopped saturating, the burst-mode noise floor went from **158 to 38 nT/ $\sqrt{\text{Hz}}$** — a factor of 4.1 at the same two parked points. Burst is the best operating point available today by a wide margin. See Chapter 8.

3. Streaming reaches 1 kHz, but its noise floor is $\sim 6\times$ burst

Streaming delivered **1041.7 Hz** with a gapless, duplicate-free rep axis across 125 000 reps — the 1 kHz target is met. Its noise floor, however, sits a flat $\sim 6\times$ above burst across 10–500 Hz for reasons not yet identified. Chapter 9 lists what has been ruled out and the one-minute rig test that decides it.

1.4 Status at a glance

Item	State	Number	Where
Averaged mode rate	DONE	86.8–88.8 Hz	§4
Burst mode rate	DONE	4167 Hz in-burst	§5
Streaming rate	DONE	1041.7 Hz	§6
Best measured noise floor	DONE	38 nT/ $\sqrt{\text{Hz}}$ (burst)	§8.2
Readout window chosen	DONE	120 μs	§8.1
Timing model	DONE	max(host, FPGA)	§7.2
Free averaging implemented	DONE	42 reps fit, 10 used	§7.3
Per-mode post-processing	DONE	3 pipelines	§9.5
Notebook reorganised	DONE	4A/4B/4C + 5A/5B/5C	§11.5
Burst replay root cause	OPEN	49.9% of batches	§9.1
Streaming noise excess	OPEN	$\sim 6\times$ burst	§9.2
Per-acquire noise term	OPEN	does not average down	§8.4
ABBA ordering benefit	OPEN	never tested on hardware	§2.2
Calibrated system sensitivity	OPEN	no field reference yet	§2.4

Chapter 2

The measurement, from photons to Δf

Before any timing discussion, this is what a “sample” actually is. All three acquisition methods produce the same quantity by the same arithmetic; they differ only in how many FPGA repetitions go into one sample and how the host collects them.

2.1 Two parked frequencies on one flank pair

A single ODMR resonance is fitted (Lorentzian, from the Step 1 reference sweep), giving its centre f_0 , half-width γ , baseline B , and the two maximum-slope points f_- and f_+ on either flank. Rather than sweeping, the microwave source is *parked* alternately at f_- and f_+ and the photoluminescence is read at each.

$$z_{\pm} = \frac{\text{counts at } f_{\pm}}{\text{normalisation}} \quad \text{normalised PL at each parked point} \quad (2.1)$$

$$L = z_+ - z_- \quad \text{the lock-in signal} \quad (2.2)$$

$$\Delta f = \frac{L - (b_+ - b_-)}{m_- - m_+} \quad \text{shift of the resonance, in MHz} \quad (2.3)$$

where $b_{\pm}$ are the calibration baselines at the parked points and $m_{\pm}$ the flank slopes, both measured off the reference sweep. Field is then $\Delta B = \Delta f / \gamma_{\text{NV}}$ with $\gamma_{\text{NV}} = 28.024 \text{ kHz } \mu\text{T}^{-1}$ along the tracked NV axis.

2.1.1 Why the difference and not one flank

If the resonance shifts by δ , z_- moves down the negative-slope flank and z_+ moves up the positive-slope one: the shift appears with opposite sign in the two channels and *adds* in L . A change in laser power or collection efficiency scales both channels the same way and *cancels* in L to first order.

This works, and by a large factor. In the 2026-08-06 heat-gun/motor run, z_- and z_+ swung **13.4% peak-to-peak together** while the difference stayed flat — a common-mode rejection ratio of **32×**.

2.1.2 The noise consequence

Because Δf is linear in L ,

$$\sigma(\Delta f) = \frac{\sigma(z_+ - z_-)}{|m_- - m_+|}. \quad (2.4)$$

Two things follow that recur throughout this document:

1. **Collapsed dip depth costs sensitivity directly.** The slopes come from the calibration. If the resonance moves off the parked pair, or the laser/MW level drifts, the real slopes flatten while the assumed ones do not, and $\sigma(\Delta f)$ grows in proportion. Every acquisition cell now prints the measured dip depth at both parked points against the calibrated value.
2. **Post-processing must act on the difference, not on the channels.** Filtering z_- and z_+ independently breaks the cancellation. This is not hypothetical: see §9.5.1.

2.2 The residual first-order term, and ABBA

The cancellation is not perfect, because the two points are not read at the same instant. In the default "forward" ordering the rep visits f_- then f_+ , one readout window apart, *always in the same direction*. A common-mode drift d per unit time therefore contributes $+d\tau$ to L every rep: the estimator is differentiating the common mode.

"abba" ordering emits f_-, f_+, f_+, f_- and folds slot 0 with slot 3 and slot 1 with slot 2. A linear drift contributes $+d, +2d, +3d, +4d$ across the four slots, and the folding cancels it exactly. It doubles the readouts per rep, so one "abba" rep costs and averages exactly two "forward" reps — the cancellation is paid for in rep count, not in time.

OPEN — ABBA has never been tested on hardware

Implemented and verified in simulation; free at equal averaging. Given the $32\times$ rejection already measured and the 13.4% common-mode swing, the residual first-order term is plausibly significant, but no A/B has been run. Pass criterion and protocol: §12.1, check S7.

2.3 The two sensitivity conventions, fixed once

Two numbers are quoted throughout, and they are not the same thing.

$\eta = \sigma_B \sqrt{2\Delta t}$ — the “implied” sensitivity

The white-noise amplitude spectral density that a measured standard deviation σ_B at sample spacing Δt would correspond to. Cheap to compute from any trace. *Only meaningful if the trace is actually white*: drift inflates σ without raising the noise density.

Measured floor — the median ASD in the flat band

Read directly off a Welch periodogram between 10 Hz and $0.4f_s$. Makes no whiteness assumption.

Both are reported side by side, and **when they disagree the floor is the one to trust**; the ratio is a direct readout of how drift-dominated the trace is. For example the 2026-08-14 burst run gives $\eta = 56$ but a floor of $38 \text{ nT}/\sqrt{\text{Hz}}$, a ratio of 1.5 — the trace is not white.

The convention is fixed in code, in `notebook_modules/twopoint_spectra.py`, whose PSD normalisation is verified against Parseval’s theorem (the integral of the PSD reproduces the variance to 0.02%) and against a synthetic tone of known amplitude.

Correction — an earlier $\sqrt{2}$ and an earlier $\sqrt{\text{reps}}$

The 2026-08-06 document used $\eta = \sigma_B \sqrt{\Delta t}$, smaller than the convention above by $\sqrt{2}$. Its figures are restated here in the $\sqrt{2\Delta t}$ convention where they are compared with 2026-08-14 numbers.

Separately, the first draft of that document paired a *reps-averaged* sweep-point σ with a *single-rep* time, understating η by $\sqrt{\text{reps}} \approx 10$. That was corrected on 2026-08-12 by cross-calibrating against the burst runs; see §8.1.

2.4 What is not being claimed

Every sensitivity figure in this document is a **readout-chain** figure: it comes from the measured scatter of Δf and the ODMR-derived slopes. None of them is a calibrated system sensitivity, because that requires a known applied field, which has not been done. The distinction matters for the project’s sensitivity claims and is the reason `NV Project/wiki/concepts/sensitivity.md` keeps these numbers separate from the roadmap target.

Chapter 3

What all three methods share

The three acquisition methods run the *same* FPGA program and read the *same* accumulated buffer. They differ only in how many repetitions the host takes per call and how it drains the buffer. Everything in this chapter is therefore common to all three, and the per-method chapters that follow only have to describe the differences.

3.1 The clock everything is counted in

Every duration in the program is stored as an integer number of cycles — the `_treg` suffix on every `NVConfiguration` field. The conversion is done by `soccfg.us2cycles()` and the clock is

$$f_{\text{tProc}} = 307.2 \text{ MHz} \quad \implies \quad 1 \text{ treg} = \frac{1}{307.2 \text{ MHz}} = 3.2552 \text{ ns.} \quad (3.1)$$

That number is not taken from the firmware printout; it is recovered from the data, by the argument in §3.4. The recovered divisor for the 213 μs runs is $D = 65434$, and $65434/307.2 \times 10^6 = 213.001 \mu\text{s}$ — and `us2cycles(213)` rounds $213 \times 307.2 = 65433.6$ to exactly 65 434. The two agree to the last cycle, which is what makes it safe to quote the rest of this chapter in nanoseconds.

Field	Value	treg	Exact duration
<code>readout_integration_tus = 120</code> (current)	120 μs	36 864	120.0000 μs
<code>readout_integration_tus = 213</code> (08-06)	213 μs	65 434	213.0013 μs
<code>relax_delay_treg</code> (set in the run cells)	—	1000	3.2552 μs
<code>relax_delay_tns</code> (default_config)	500 000 ns	153 600	500.0000 μs
<code>synci(100)</code> in <code>initialize()</code>	—	100	325.52 ns

`default_config.relax_delay_tns = 500000` is *overridden* by every acquisition cell, each of which sets `relax_delay_treg = 1000`. The 500 μs value never reaches the FPGA in any of the three methods; it would double the rep time if it did.

3.2 The FPGA program: one rep, step by step

`MultipointLockinODMR.body()` (`notebook_modules/multipoint_lockin_program.py`) emits the following assembly *per parked frequency*, i.e. *per emission slot*. The last column is what each step contributes to the tProc time cursor — which is not the same as how long the step lasts, and the difference is the whole content of this table.

#	ASM step	What it does	Advances the cursor by
1	<code>mw_frequency_register.set_to(f)</code>	Retunes the MW generator at runtime. This is the trick that lets 2 (or 16) non-uniform frequencies live in one program instead of one upload each.	0. Register writes are tProc instructions, not scheduled events.
2	<code>pulse(ch=mw_channel, t=0)</code>	Constant-amplitude MW pulse. Its length is <code>readout_integration_treg</code> — the MW is on for exactly the readout window, not longer.	Sets the <i>generator</i> timestamp to $0 + \tau = 120.0000\ \mu\text{s}$
3	<code>trigger_no_off(adcs, pins=[laser_gate], width=readout_integration_treg, adc_trig_offset=0, t=0)</code>	Opens the laser gate <i>and</i> arms the ADC accumulator, both at $t = 0$. This step is the rep.	Sets the <i>readout</i> timestamp to $0 + \tau = 120.0000\ \mu\text{s}$
4	<code>trigger(pins=[laser_gate], width=relax_delay_treg, t=readout_integration_treg)</code>	Closing trigger: drives the laser gate low at the end of the slot. Pins only — no adcs.	0. See the box below.
5	<code>wait_all()</code>	Blocks the tProc until every timestamp is reached.	0 beyond what 2 and 3 already set
6	<code>sync_all(relax_delay_treg)</code>	New cursor = max timestamp + 1000 treg.	3.2552 μs

Why step 4's 3.2552 μs costs nothing, from the source

`trigger()` only writes a timestamp inside `if any([adcs, ddr4, mr]):` (`quickdawg/nvpulsing/nvaverageprogram.py`). Step 4 passes `pins` and nothing else, so that branch is skipped entirely: the call emits two `seti` output-set instructions and touches no channel's timestamp. Its width is a property of the laser gate waveform, not of the schedule. `trigger_no_off()` is the same function with the closing `seti` commented out — which is what the name means. It *does* set the readout timestamp, to `t_start + ro_length` converted into tProc cycles, and that is step 3's entry above.

3.2.1 The per-slot budget, and the 1.6 % that is left over

Summing the last column: one slot advances the cursor by $\tau + 3.2552\ \mu\text{s}$, so the programmed rep time is

$$t_{\text{rep}}^{\text{programmed}} = n_{\text{slots}} \times (n_{\text{readouts}} \tau + 3.2552\ \mu\text{s}). \quad (3.2)$$

At the two operating points that have been measured:

	$n_{\text{slots}}\tau$	Eq. 3.2	measured cadence	Eq. 3.2 – measured
$\tau = 213\ \mu\text{s}$, 2 slots	426.0 μs	432.5 μs	423.8–425.4 μs	+7 to 9 μs (1.6–2.0%)
$\tau = 120\ \mu\text{s}$, 2 slots	240.0 μs	246.5 μs	240.0 $\mu\text{s}^\dagger$	6.5 μs (2.7%)

[†]recovered exactly from the value quantisation (§3.4); the 213 μs figures are wall-clock cadences across four 2026-08-06 burst runs.

The measurement sits *below* even $n_{\text{slots}}\tau$ at $\tau = 213\ \mu\text{s}$, so the two `sync_all` terms are not merely small — they are not visible at all. The model used everywhere in this document therefore drops them; §3.2.3 states it and `time_per_rep()` implements it.

3.2.2 How many slots there are

multipoint_pair_order	multipoint_skip_reference	Slots/rep	Readouts/rep
"forward"	True (<i>current default</i>)	2	2
"forward"	False	2	4
"abba"	True	4	4
"abba"	False	4	8

`multipoint_skip_reference = True` drops the per-slot “MW-off” reference readout. It is on because the MW cannot actually be gated off on this setup, so the reference was a contaminated copy of the signal rather than a clean normaliser — it halved the FPGA time and removed a systematic. With it on, z is normalised against a stored constant (`ref_norm_counts` from the calibration JSON) instead of a per-shot reference.

3.2.3 The one timing formula everything else derives from

$$t_{\text{rep}} = n_{\text{slots}} \times n_{\text{readouts per slot}} \times \tau \quad (3.3)$$

Worked through for every configuration this project runs, at $\tau = 120 \mu\text{s}$:

Configuration	slots	reads/slot	t_{rep}	cadence	reads/shot
two-point, forward, skip ref (<i>default</i>)	2	1	240 μs	4167 Hz	2
two-point, forward, with ref	2	2	480 μs	2083 Hz	4
two-point, ABBA, skip ref	4	1	480 μs	2083 Hz	4
16-point, forward, skip ref	16	1	1.92 ms	521 Hz	16
16-point, forward, with ref (<i>Step 4S</i>)	16	2	3.84 ms	260 Hz	32

The last column is not decoration: it is what the accumulated buffer budget is spent on, and §3.3 and the streaming chapters both come back to it.

The relax windows do not appear in Eq. 3.3, and that is measured

An earlier model added the relax windows, giving $n_{\text{slots}} \times (\tau + 2 \times 3.2552 \mu\text{s})$. Reading the source narrows that to *one* `sync_all` per slot rather than two (the closing `trigger` sets no timestamp — see the box in §3.2), and the measurement removes even that one: across four 2026-08-06 burst runs the cadence is 423.8–425.4 μs at $\tau = 213 \mu\text{s}$, against 426.0 μs for $2 \times \tau$ and 432.5 μs for the sync-inclusive sum. The measurement is *below* 2τ , so the sync term is not resolvable. `time_per_rep()` implements Eq. 3.3; `include_relax=True` restores the old sum as a conservative upper bound if you want headroom rather than an estimate.

3.3 What the accumulated buffer holds

The ADC does not hand back samples; it hands back a *sum*. For each readout trigger, the firmware accumulates `readout_integration_treg` raw ADC samples into one 64-bit integer entry, and writes one such entry per readout per rep. So after a run of `reps` repetitions with R readouts per rep, the buffer holds `reps` $\times$ R integers, laid out as (`reps`, `nsweep`, R) — with `nsweep` = 1 here.

`analyze_results()` then does two divisions:

$$\text{stored value} = \underbrace{\frac{\text{integer accumulator sum}}{\text{readout_integration_treg}}}_{\text{always}} \times \underbrace{\text{reps}}_{\text{averaged mode only}} \quad (3.4)$$

Two consequences, both used heavily later:

1. **The stored counts are τ -independent in the mean** (they are a mean ADC level, not a total), which is why the counts do not grow when τ is raised. Their *variance* does fall as $1/\tau$, which is the sensitivity benefit of a longer window.
2. **The exact FPGA readout time is recoverable from the recorded numbers.** This is the measurement that settled the timing question; it gets its own section.

3.4 Recovering the exact readout time from the data

Because the numerator of Eq. 3.4 is an integer, every stored count is an integer multiple of

$$\frac{1}{D}, \quad D \equiv \text{treg} \times \text{reps}.$$

Searching for the smallest divisor D that makes *all* of a run's counts integral therefore recovers D exactly, with no reliance on what the notebook claims it configured. The readout time follows:

$$t_{\text{ADC per call}} = n_{\text{slots}} \times \frac{D}{f_{\text{clk}}}, \quad f_{\text{clk}} = 307.2 \text{ MHz} \quad (3.5)$$

The clock is pinned independently by two runs: $213 \mu\text{s} \leftrightarrow \text{treg} = 65434$ and $120 \mu\text{s} \leftrightarrow \text{treg} = 36864$, both giving 307.2 MHz.

Only the product is identifiable. $D = 36864$ is $(\text{treg} = 36864) \times (\text{reps} = 1)$ and $(\text{treg} = 768) \times (\text{reps} = 48)$ equally well. That is sufficient, because Eq. 3.5 depends only on D . In *burst* mode there is no rep averaging, so D is treg and the readout window follows directly — which is how the 2026-08-14 burst run was confirmed to be at $\tau = 120.0 \mu\text{s}$, exactly as configured.

Implemented as `twopoint_postprocess.recover_readout_quantum()` and `readout_seconds()`.

Why this matters practically

It replaced a 12% error. The burst cadence had been estimated as `acq_seconds / n_samples` over the real batches, which reads $269 \mu\text{s}$ against a true $240 \mu\text{s}$ because it includes the host share of each acquire. That 12% stretch went straight into the time axis, the quoted rate, and the frequency scale of every spectrum.

3.5 The host RPC chain

`NVAveragerProgram.acquire()` (`quickdawg/nvpulsing/nvaverageprogram.py`, lines 163–304) issues this sequence of Pyro remote calls to the board on **every** call:

Stage	Call	Notes
1	<code>config_all(soc, load_envelopes, reset, load_mem)</code>	Uploads the program. <code>load_mem=True</code> , so the binary and data memory are re-sent on every call. Until 2026-08-12 this raised <code>TypeError</code> on every call (wrong keyword) and fell through to defaults.
1a	<code>→ soc.start_src("internal")</code>	inside <code>config_all</code>
1b	<code>→ soc.stop_tproc(lazy=not reset)</code>	documented no-op on tProc v1 unless <code>reset=True</code>
1c	<code>→ load_envelopes, load_bin_program, load_mem</code>	re-uploaded each call
2	<code>soc.start_src(start_src)</code>	again, from <code>acquire</code>
3	<code>config_bufs(soc, enable_avg=True, enable_buf=False)</code>	arms the accumulator
4	<code>soc.reload_mem()</code>	no-op unless tProc v2
5	<code>soc.start_readout(total_ count, ...)</code>	starts the board-side worker <i>and</i> the tProc
6	<code>soc.poll_data() × N</code>	blocks here while the FPGA runs
7	<code>Pyro deserialise → analyze_results() reshape</code>	host-side

The critical structural fact: stage 6 is where the FPGA time goes

The host does not wait for the FPGA *after* finishing its own work — it waits *inside* stage 6, which is part of the chain. So the FPGA time is **concurrent with** the host chain, not additive to it, up to the point where the FPGA run outlasts the rest of the chain. That single fact is the whole content of Chapter 7, and it is what the 2026-08-06 analysis got wrong.

OPEN — the per-stage split has never been measured

The stage list above is read off the source. The *time* each stage takes can only be measured on the rig, by wrapping the soc proxy so every remote call is timed. `scripts/profile_twopoint_acquire.py` does exactly that and has never been run, because the RFSoc is not reachable from the analysis machine. Until it is, the chain is known only in total (10–11.4 ms) and not stage by stage.

This is the single largest gap in this document. Run:

```
python scripts/profile_twopoint_acquire.py -ip 192.168.0.103
```

3.6 Python outside the call

Measured directly as (loop period – `acq_seconds`), per batch:

Run	Python outside <code>acquire()</code>	Share of period
2026-08-14 $\tau=120$, 10 reps	0.067 ms	0.58%
2026-08-14 $\tau=120$, 23 reps	0.056 ms	0.49%
2026-08-06 $\tau=213$, 23 reps	0.058 ms	0.51%

Row-dict construction, the conversion arithmetic and the history append together cost well under a tenth of a millisecond. **The Python side is not a bottleneck in any mode** and the earlier vectorisation of the per-sample conversion did its job. It is listed here so that it can be ruled out rather than suspected.

The exception is the CSV write, which is not per-sample and is covered per mode.

Chapter 4

Method A — Averaged mode

Notebook cell	Step 4A (analysis: Step 5A)
Code path	<code>prog.acquire(progress=False)</code>
One call returns	the <i>mean</i> over <code>reps</code> reps: one sample
Measured rate	86.8–88.8 Hz , at every τ and rep count tested
Rate limited by	the host call (~ 11.4 ms), <i>not</i> the readout window
Duty cycle	20.9% at 10 reps, 48.3% at 23 reps, 85.7% at $\tau=213/23$ reps
Measured floor	274–501 nT/ $\sqrt{\text{Hz}}$ (varies $1.8\times$ between runs)
Open defects	none

4.1 Mechanics

The program runs `reps` repetitions. `analyze_results()` averages over the rep axis (Eq. 3.4 divides by `reps`) and returns one number per parked frequency. The host then converts that pair into one Δf and one CSV row, and immediately calls `acquire()` again.

The cost of this design. The rep axis is *discarded*. Those `reps` repetitions were already independent time samples spaced by t_{rep} ; averaging them away throws that time resolution out. Methods B and C exist to keep it.

4.2 Timing budget, component by component

Three operating points, all measured. Read the **Adds?** column first: it is what distinguishes this mode from the other two.

Operating point 1: $\tau = 120\ \mu\text{s}$, 10 reps — what actually ran on 2026-08-14

Component	Time	Share	Adds?	How obtained
ADC integration ($10 \times 240\ \mu\text{s}$)	2.400 ms	20.9%	no	exact, Eq. 3.5, $D=368640$
of which MW pulse	0 ms	0%	no	concurrent with the readout
of which relax/sync	~ 0 ms	$\sim 0\%$	no	absorbed; measured, §3.2
<code>acquire()</code> call, fastest observed	10.302 ms	89.8%	—	p05 of <code>acq_seconds</code>
<code>acquire()</code> call, median	11.408 ms	99.4%	yes	median of <code>acq_seconds</code>
<code>acquire()</code> jitter (p95–p05)	1.624 ms	14.2%	yes	host/network scheduling
Python outside <code>acquire()</code>	0.067 ms	0.58%	yes	period – <code>acq_seconds</code>
CSV flush stall (worst)	215.8 ms	—	yes	2 events $> 3\times$ median period
TOTAL per sample	11.475 ms	100%		$\rightarrow$ 87.1 Hz

Read this table as a maximum, not a sum

$2.400 + 11.408 = 13.8$ ms, but the measured period is 11.475 ms. The two do not add, because

the ADC integration happens *inside* stage 6 of the host chain (§3.5). The period is

$$\begin{aligned}\text{period} &= \max(\text{host chain}, \text{reps} \times t_{\text{rep}}) + \text{Python outside} \\ &= \max(11.408, 2.400) + 0.067 = 11.475 \text{ ms.}\end{aligned}$$

Only 2.400 ms of every 11.475 ms is spent collecting photons. The remaining 79% is the host talking to the board.

Operating point 2: $\tau = 120 \mu\text{s}$, 23 reps

Component	Time	Share	Note
ADC integration: $23 \times 240 \mu\text{s}$	5.520 ms	48.3%	$D=847872$, exact
<code>acquire()</code> call, fastest	10.266 ms	89.7%	
<code>acquire()</code> call, median	11.384 ms	99.5%	
<code>acquire()</code> jitter	1.819 ms	15.9%	
Python outside	0.056 ms	0.49%	
CSV flush stalls	0 ms	—	none in this run
TOTAL per sample	11.440 ms	100%	→ 87.4 Hz

Operating point 3: $\tau = 213 \mu\text{s}$, 23 reps — the 2026-08-06 default

Component	Time	Share	Note
ADC integration: $23 \times 426 \mu\text{s}$	9.798 ms	85.7%	$D=1504982$, exact
<code>acquire()</code> call, fastest	10.218 ms	89.4%	
<code>acquire()</code> call, median	11.369 ms	99.5%	
<code>acquire()</code> jitter	1.877 ms	16.4%	
Python outside	0.058 ms	0.51%	
CSV flush stall (worst)	219.5 ms	—	4 events
TOTAL per sample	11.428 ms	100%	→ 87.5 Hz

4.2.1 The three side by side — the decisive comparison

	$\tau=120, 10 \text{ reps}$	$\tau=120, 23 \text{ reps}$	$\tau=213, 23 \text{ reps}$	spread
ADC integration	2.400 ms	5.520 ms	9.798 ms	4.1×
Period	11.475 ms	11.440 ms	11.428 ms	1.004×
Rate	87.1 Hz	87.4 Hz	87.5 Hz	1.005×
Duty cycle	20.9%	48.3%	85.7%	4.1×

This one row is the whole finding

The ADC integration time varies by **4.1×** across these three operating points. The period varies by **0.4%**. There is no readout-window or rep-count setting that makes this mode faster than $\sim 88 \text{ Hz}$.

4.3 The CSV flush, and why it was fixed

The worst single stall in these runs is 219.5 ms — nineteen missed samples. The cause was the old save routine calling `pd.DataFrame(rows).to_csv(...)` every 5 seconds, rewriting *every row collected so far*: $O(N^2)$ over a run. Of the 94 outlier batches across the 2026-08-06 session, **46 sat within 150 ms of a 5-second boundary**.

Replaced by `twopoint_runner.IncrementalCsvWriter`, which appends only the new rows (header once). On a 40 000-row run this is $10.2\times$ cheaper and the stall is flat with run length instead of growing.

4.4 Why you no longer enter a reps-per-batch

Every version of this notebook before 2026-08-16 asked you for `REPS_PER_BATCH`, and every value anyone entered was wrong, because the right value is not a preference — it is a measurement of the host, and the host is the only thing that sets it.

4.4.1 The argument

An `acquire()` call costs 8.5–11.4 ms of host round-trip *whatever the program does*, and the FPGA work runs underneath that rather than adding to it (Ch. 7 is the evidence). So the period is

$$T = \max \left(\underbrace{T_{\text{host}}}_{\text{fixed}}, \underbrace{\text{reps} \times t_{\text{rep}}}_{\text{yours}} \right), \quad (4.1)$$

which is *flat* in reps until the second term overtakes the first. Every rep below the crossing is free: the call takes the same wall-clock time either way, so an under-filled batch is not a faster measurement, it is the same measurement with photons thrown away. Setting reps by hand can only do one of two things — leave sensitivity on the table, or cross the floor and cost rate.

`AVG_REPS_PER_BATCH = None` therefore replaces the question with the arithmetic:

$$\text{reps_that_fit}() = \left\lfloor \frac{T_{\text{floor}}}{t_{\text{rep}}} \right\rfloor. \quad (4.2)$$

Two details in there are deliberate. It uses the *floor* — the fastest call observed, 5th percentile — rather than the typical call, so that filling the budget does not make the quick calls FPGA-bound and drag the rate down. And T_{floor} comes from `measure_host_floor()` run at the top of the cell (`AVG_CALIBRATE_FLOOR = True`, ~ 0.3 s) rather than from the stored `HOST_FLOOR_S = 10.1 ms`, because the floor is rig- and session-dependent: it has moved by 1.6 ms across the four sessions on record.

4.4.2 Worked, on the 2026-08-17 run

Quantity	Value	Where it came from
measured floor T_{floor}	8.52 ms	<code>measure_host_floor()</code> , 5th percentile of 20 calls
median call	8.71 ms	same measurement
t_{rep}	240 μs	Eq. 3.3, 2 slots $\times$ 120 μs
<code>reps_that_fit()</code>	35	$\lfloor 8.52/0.240 \rfloor$
FPGA work at 35 reps	8.40 ms	$35 \times 240 \mu\text{s}$ — 99% of the floor used
achieved rate	104.5 Hz	1067 samples in 10.2 s

The cell then reports how much was left over, and on that run it reported that 36 reps would still have fitted — a 1.4% improvement it declined to take automatically, because the floor is measured with jitter and rounding up would sometimes cross it.

4.4.3 You can still set it by hand

`AVG_REPS_PER_BATCH = <int>` overrides Eq. 4.2 and the cell prints (`set by hand`) instead of (`auto: filled the floor`). Reasons to do it: reproducing an old run exactly, or deliberately going *past* the floor to trade rate for averaging — at $\tau = 120\ \mu\text{s}$, 100 reps gives 24 ms of FPGA work and therefore 42 Hz, with $\sqrt{100/35} = 1.7\times$ the averaging. Below 35 there is no reason at all.

4.5 What to set

Parameter	Recommended	Why
<code>AVG_REPS_PER_BATCH</code>	None	Fills the measured call floor automatically via <code>reps_that_fit()</code> — 35 reps at $\tau=120$ on 2026-08-17, against the 10 used on 08-14. See §4.4 and §7.3.
<code>AVG_CALIBRATE_FLOOR</code>	True	Measures the floor on this rig at startup; it already differs by 1.2 ms between the two sessions. Costs ~ 0.3 s.
<code>readout_integration_tus</code>	120	Inside the flat sensitivity band (§8.1). Not a rate lever here.
<code>AVG_PAIR_ORDER</code>	"forward"	"abba" is free at equal averaging but untested (§2.2).
<code>AVG_SKIP_REFERENCE</code>	True	The MW-off reference is contaminated on this setup.
<code>AVG_SHOW_PLOT</code>	False	A live 3-panel redraw costs 20–100 ms, which at an 11 ms period would dominate the loop.

4.6 Failure modes and what warns you

Failure	Symptom	Guard
Resonance drifts off the parked pair	Dip depth collapses, $\sigma(\Delta f)$ grows as $1/\text{depth}$ (Eq. 2.4)	<code>report_calibration_validity()</code> prints measured vs calibrated depth and depth/noise; warns below 50% or ratio < 3
60 Hz mains	Appears at 26.8–28.8 Hz, indistinguishable from signal	<code>alias_report()</code> in Step 5A; cannot be removed (§8.5)
Averaging does not help	σ gets <i>worse</i> with more reps	OPEN — see §8.4
Peaking transient	Per-batch transient from an MW polarisation pulse at each batch boundary	<code>pre_init=False</code> on the live program, enforced by <code>assert</code> ; primed once with a throwaway <code>pre_init=True</code> acquire
Stale buffer	Would show as a bit-identical repeat	Freshness guard; zero occurrences in 20 365 averaged batches

4.7 When to use it

When ~ 88 Hz is enough. It is the only one of the three methods with no open defect, its timestamps are honest, and it has never been observed to replay a buffer. For drone sensing and MCG, where bandwidth is the point, it is the wrong choice — go to Method B or C.

Chapter 5

Method B — Burst mode

Notebook cell	Step 4B (analysis: Step 5B)
Code path	<code>prog.acquire(per_rep=True, reset_tproc=True)</code>
One call returns	<i>every</i> rep as its own sample: reps samples
Measured rate	4167 Hz in-burst, 3089 Hz net
Rate limited by	the FPGA cadence; the host cost is amortised over the burst
Duty cycle	74.1%
Measured floor	38 nT/$\sqrt{\text{Hz}}$ — best of the three
Open defects	OPEN 49.9% of bursts are replays (§9.1)

5.1 Mechanics

The FPGA program is identical to Method A’s. The only change is that `analyze_results(per_rep=True)` *keeps* the rep axis: the second division in Eq. 3.4 is skipped, and the call returns a $(\text{reps} \times n_{\text{freqs}})$ array. Every row is an independent time sample spaced by t_{rep} .

The ~ 11 ms host chain is therefore paid once per *burst* of 500 samples instead of once per sample — a $500\times$ amortisation, which is where the rate comes from.

The structural difference from Method A. Here the FPGA time and the host time *do* add. Stage 6 of the chain (§3.5) cannot return until the tProc has finished all 500 reps, which takes 120 ms — far longer than the rest of the chain. The mode is genuinely FPGA-bound, which is exactly what Method A is not.

5.2 Timing budget, component by component

Operating point: $\tau = 120\ \mu\text{s}$, 500 reps per burst (2026-08-14)

Component	Time per burst	Share	How obtained
<i>Productive time</i>			
ADC integration: $500 \times 240\ \mu\text{s}$	120.000 ms	81.7%	exact, Eq. 3.5
<i>Overhead, per real burst</i>			
Host RPC chain (adds here)	14.697 ms	10.0%	measured wall – exact FPGA
<code>acquire()</code> total, real burst	134.697 ms	91.7%	median of <code>acq_seconds</code> , real batches
p05 / p95	118.0 / 136.3 ms		spread is host jitter
<i>Pure waste — the open defect</i>			
Stale replay acquire	12.533 ms	8.5%	$\times 185$ batches
total run time wasted	2.32 s of 30.11 s	7.7%	produces no usable data
<i>Result</i>			
Total per usable sample	0.324 ms		30.11 s / 92 814 samples
Net rate	3089 Hz		
In-burst rate	4167 Hz		$1/t_{\text{rep}}$, exact
Duty cycle	74.1%		22.32 s of FPGA in 30.11 s

Where the 26% of non-productive time goes

Where the 30.11 s run went	Time	Share
ADC integration (productive)	22.32 s	74.1%
Host chain, 186 real bursts	2.73 s	9.1%
Stale replay acquires (waste)	2.32 s	7.7%
Flushes, row-0 drops, residual	2.74 s	9.1%

Fixing the replay defect would recover the 7.7% directly and would also remove the wasted host chain on those calls — roughly a 15% rate gain, taking the net rate from 3089 Hz toward ~3600 Hz.

Operating point: $\tau = 213 \mu\text{s}$, 500 reps per burst (2026-08-06)

Component	Time per burst	Share	Note
ADC integration: $500 \times 426 \mu\text{s}$	213.001 ms	100.3%	exact
Host RPC chain	<2.13 ms	<1%	below the resolution of timing a 213 ms call
<code>acquire()</code> total, real burst	212.267 ms	100%	measured
Stale replay acquire	12.648 ms	6.0%	$\times 106$ batches = 1.34 s wasted
Total per usable sample	0.514 ms		
Net rate	1946 Hz		
In-burst rate	2347 Hz		
Duty cycle	82.9%		

Note that the host share is a *smaller* fraction here because the FPGA run is longer — the amortisation improves with burst length. That is the argument for large BURST_REPS, bounded only by how long a gap in the time series is acceptable and by the accumulated-buffer size.

5.2.1 Rate against burst length

With host chain $H \approx 14.7$ ms and stale overhead $S \approx 12.5$ ms per real burst, the net rate is

$$f_{\text{net}} = \frac{N}{N t_{\text{rep}} + H + S} \quad (5.1)$$

which at $\tau = 120 \mu\text{s}$ gives:

BURST_REPS	FPGA per burst	Net rate	Duty cycle
100	24.0 ms	1961 Hz	47%
250	60.0 ms	2874 Hz	69%
500	120.0 ms	3413 Hz	82%
1000	240.0 ms	3771 Hz	90%
2000	480.0 ms	3963 Hz	95%

(The 500-rep row predicts 3413 Hz against 3089 Hz measured; the difference is the CSV flushes and the dropped row-0 samples.) The returns flatten past ~1000 reps, and the gaps grow: a 2000-rep burst leaves a 27 ms hole in the time series every 480 ms.

5.3 Three artefacts, and what each one is

The 2026-08-06 `Burst Mode.png` showed four visually distinct problems. All four are now explained, and three are fixed.

What you see	What it is	Status
Large spike, strictly every ~ 0.45 s	Row 0 of each <i>stale</i> batch: $+1908 \text{ kHz} \approx +68 \mu\text{T}$ mean. Real batches read -72 kHz at row 0, so this is the buffer read, not physics.	Fixed by dropping stale batches
Smaller, less regular spikes	Genuine single-rep glitches, recorded twice , because the replay reproduces them. E.g. -1443 kHz at sample 824 appears in both batch 11 and its replay batch 12.	De-duplicated; the glitches themselves are real and unexplained
Dense/sparse striping	The timestamp model . Samples were stamped across the <i>measured</i> acquire window: 13 ms for a stale batch, 425 ms for a real one — a $33\times$ alternating compression of the time axis, even though the true cadence is a constant $424.8 \mu\text{s}$.	Fixed: timestamps now come from the cadence
Break every ~ 5 s	The <code>SAVE_EVERY_SEC = 5.0</code> full-DataFrame rewrite. 267 ms in one run.	Fixed by the append-only writer
Row 0 runs 1.4% high in <i>real</i> batches too	A genuine post-idle transient : 1210.6 against 1193.4 ADC. Distinct from the replay, and present even in clean batches.	Dropped at acquisition and in post-processing




Figure 5.1: Burst-mode stale reads on 2026-08-06. The stale replays (red) compress into narrow vertical stripes because their 1000 samples are stamped across a 13 ms window while the real samples (blue) spread over 425 ms. Circled points are the row-0 transients — one every ~ 0.45 s, exactly the periodicity seen in the original plot.

5.4 What to set

Parameter	Recommended	Why
BURST_ON_STALE	"drop"	Re-acquires, then <i>skips</i> the batch rather than writing a replay to the CSV. Costs duty cycle, not correctness.
BURST_STALE_RETRIES	2	The race is transient; a repeat call usually lands on a completed run.
BURST_REPS	500–1000	See the rate table above; returns flatten past 1000 and the time-series gaps grow.
BURST_DROP_FIRST_REP	True	Row 0 runs 1.4% high — a real transient.
BURST_RESET_TPROC	True	Kept available and harmless, but measured not to fix the replay (§9.1).

5.5 When to use it

Today, for best sensitivity: this is the mode. $38 \text{ nT}/\sqrt{\text{Hz}}$ at an effective 3.1 kHz beats streaming by $\sim 6\times$ and averaged mode by $7\text{--}13\times$. The price is a 26% duty cycle loss and a $\sim 15 \text{ ms}$ hole in the time series every 135 ms, which matters for a continuous MCG trace and does not matter for a noise-floor measurement.

If the replay defect is ever fixed, this mode gets $\sim 15\%$ faster for free. If the streaming noise defect is fixed, Method C becomes strictly better for continuous recording.

Chapter 6

Method C — Streaming mode

Notebook cell	Step 4C (analysis: Step 5C)
Code path	<code>prog.stream(total_reps=N)</code> — a generator
One packet returns	38–500 reps, continuously, as they arrive
Measured rate	1041.7 Hz (or 4167 Hz with no host binning)
Rate limited by	the FPGA cadence alone
Duty cycle	100.0%
Measured floor	240 nT/ $\sqrt{\text{Hz}}$
Open defects	OPEN floor is $\sim 6 \times \text{burst}$ (§9.2); 25% PL droop (§6.6)

6.1 Mechanics

One `config_all`, one `start_readout(total_reps)`, then nothing but draining the streamer queue with `poll_data()` in a loop. The board-side worker thread transfers completed shots out of the accumulated buffer autonomously while the tProc keeps running.

Because nothing is re-armed mid-run:

- the replay race of Method B **cannot occur** — there is no second configuration to race against;
- there is **no inter-burst gap** — the FPGA never stops;
- timestamps follow the FPGA cadence **exactly**, rather than being reconstructed from host arrival times, which only record when the host happened to drain the queue.

cfg.reps is the run length. The tProc loops `reps` times and ends, so the rep count must be set before compilation and cannot change mid-run. `stream()` raises if `cfg.reps != total_reps` rather than silently truncating.

Host-side binning. `STREAM_TARGET_HZ` sets an integer number of reps to average on the host per output sample. Whole reps only, so the achieved rate is the cadence divided by an integer rather than exactly the target: at $\tau = 120 \mu\text{s}$, a 1000 Hz target gives `bin = 4` and therefore 1041.7 Hz. Leftover reps carry into the next packet, so no rep is dropped and no bin straddles a packet boundary.

6.2 Timing budget, component by component

Operating point: $\tau = 120\ \mu\text{s}$, $\text{bin} = 4$ (2026-08-14)

Component	Per output sample	Share	How obtained
ADC integration ($4 \times 240\ \mu\text{s}$)	960.0 μs	100.0%	exact, Eq. 3.5, $D=147456$
Host RPC chain	0 μs	0%	paid once for the whole run the FPGA does not stop between packets
Inter-packet gap	0 μs	0%	
Timestamp jitter	0.0 μs	0%	cadence-derived; min = max = 960.0 μs
TOTAL per output sample	960.0 μs	100%	measured spacing
Rate	1041.67 Hz		predicted 1041.67 Hz — <i>exact</i>
Duty cycle	100.0%		125 000 reps $\times$ 240 μs in 29.999 s

This is what a clean timing budget looks like

Every microsecond of the 960 μs sample period is ADC integration. The measured sample spacing has **min = max = 960.0 μs** — zero jitter, because the timestamps come from the cadence rather than from arrival times. Predicted and measured rate agree to the digit. Compare Method A, where 79% of the period is host overhead, and Method B, where 26% is overhead and waste.

6.3 Every parameter of the streaming cell, one at a time

Step 4C has six knobs and three of them behave in ways that are not obvious from their names. This section is the reference for all of them.

6.3.1 STREAM_TARGET_HZ and the bin

You ask for a rate; the cell converts it into an integer number of reps to average per output sample, and that integer — not your request — sets the rate:

$$\text{bin} = \max\left(1, \text{round}\left(\frac{1}{\text{STREAM_TARGET_HZ} \times t_{\text{rep}}}\right)\right), \quad f_{\text{achieved}} = \frac{1}{\text{bin} \times t_{\text{rep}}}. \quad (6.1)$$

At $t_{\text{rep}} = 240\ \mu\text{s}$, asking for 1000 Hz gives $\text{bin} = \text{round}(4.167) = 4$ and therefore **1041.7 Hz**, not 1000 Hz. That is why every streamed file in this project reports 1041.7 Hz: it is the nearest achievable rate, and reps cannot be split.

Four things follow from the bin being a *mean* of consecutive reps rather than a decimation, and all four matter:

1. **It averages, so it buys $\sqrt{\text{bin}}$ in σ .** Dropping 3 of every 4 reps would not. At $\text{bin} = 4$ that is a factor of 2 — the same free averaging Method A gets from filling the call floor (§7.3), obtained here without giving up any rate you were going to use.
2. **It anti-aliases, at $1/(2\text{bin}t_{\text{rep}})$.** A boxcar of bin reps has its first null at f_{achieved} and rolls off as $|\text{sinc}|$ before it, so noise between $f_{\text{achieved}}/2$ and the raw 4167 Hz is suppressed rather than folded back in. This is the reason to bin in the acquisition cell rather than to record raw and decimate afterwards.

3. **Bins never straddle a packet boundary.** Reps left over at the end of a packet are carried into the next one (`_pending` in the cell), so a bin is always `bin` consecutive reps of the same continuous run. Packet boundaries are an artefact of when the host drained the queue and mean nothing physical; a bin that straddled one would still be correct, but the carry makes the sample count exactly $\lfloor N_{\text{reps}}/\text{bin} \rfloor$ regardless of how the packets happened to land.
4. **The timestamp is the bin centre.** $(\text{first_rep} + (i + 0.5) * \text{bin}) * \text{cadence}$. Using the bin's first rep would bias every sample early by $\text{bin } t_{\text{rep}}/2 = 480 \mu\text{s}$ at `bin = 4`.

bin	Per output sample	Rate	Note
1	240 μs	4166.7 Hz	STREAM_TARGET_HZ = None; every rep kept what ran on 2026-08-14 and 2026-08-17
2	480 μs	2083.3 Hz	
4	960 μs	1041.7 Hz	
8	1920 μs	520.8 Hz	
42	10.08 ms	99.2 Hz	
			equivalent averaging to Method A, at 100% duty

The last row is worth noting: streaming with `bin = 42` produces the same averaging as the recommended Method A operating point, at the same rate, but with 100% duty cycle instead of 21% — i.e. *if the streaming noise excess were resolved*, streaming would dominate Method A at every rate.

6.3.2 STREAM_DURATION_SEC, total_reps and cfg.reps — all the same number

The `tProc` loops `cfg.reps` times and then ends. The rep count *is* the run length, it is compiled into the program, and it cannot be changed once the program is uploaded. So the cell computes

$$\text{total_reps} = \text{round}\left(\frac{\text{STREAM_DURATION_SEC}}{t_{\text{rep}}}\right)$$

and builds the program with `cfg.reps` set to it. `stream()` raises `ValueError` if the two disagree rather than quietly streaming a different length than you asked for. At 30 s and $t_{\text{rep}} = 240 \mu\text{s}$ that is 125 000 reps, giving $125\,000/4 = 31\,250$ samples — which is exactly the row count of every streamed CSV in the 2026-08-17 session.

There is no way to stop early and keep a valid file beyond interrupting the cell: Ctrl-C ends the generator, the `finally` leaves the board idle, and `stream_qc()["aborted"]` records that the run was short.

6.3.3 poll_timeout_s — and why it is currently unreachable

`stream(poll_timeout_s=10.0)` is a client-side guard: if the board sends nothing for this long, raise instead of waiting forever. It only does anything when the poll is bounded, and as of the 2026-08-17 fix the poll is blocking by default (`POLL_TIMEOUT_S = None`), so on the healthy path this guard is never reached. That is deliberate, and §9.3 is the argument for it. What protects an interrupted run is the generator's `finally`, not this timeout.

To arm it for a diagnostic run, set the class attribute for the session:

```
MultipointLockinODMR.POLL_TIMEOUT_S = 5.0
```

6.3.4 Stride, packet size and avg_maxlen

None of these are host-side settings; they are consequences of the firmware and the board worker, and they are what `stream_headroom()` reports.

Quantity	What it is
<code>avg_maxlen</code>	Length of the accumulated buffer, in entries, from the firmware. One entry per readout per rep.
<code>reads_per_shot</code>	$n_{\text{slots}} \times n_{\text{readouts per slot}} - 2$ for the two-point default, 32 for a 16-point run with references.
<code>shots_that_fit</code>	<code>avg_maxlen // reads_per_shot</code> . The buffer budget is shared across every readout in a shot, so a 16-point run has sixteen times less room than a two-point one.
stride	<code>max(1, int(0.1 * shots_that_fit))</code> . The board worker transfers this many shots at a time — 10% of the buffer, chosen by qick, not by us.
packet size	Whatever the worker had ready when the host polled: on 2026-08-14, 38–500 reps, median 205. It varies, and <i>it does not matter</i> : the rep axis is carried in <code>first_rep/n_reps</code> , so the timestamps are exact regardless of how the data was parcelled up. <code>stream_qc()</code> proves that on the data.

The failure mode is overflow: if unread samples ever reach `avg_maxlen`, the worker raises and the run dies. Run `stream_headroom()` before a long stream and check that `shots_that_fit` is several strides’ worth. This has never bound on the two-point path and is a live constraint on the multipoint one (Ch. 10).

`stream_headroom()` silently failed for the whole 2026-08-17 session

It read `qd.soc['readouts'][ch]['avg_maxlen']`, and `qd.soc` is a Pyro Proxy: a proxy forwards *method calls* but not `__getitem__`, so this raised ‘Proxy’ object is not subscriptable and the cell printed “could not read avg_maxlen” instead of a number. The subscriptable object is `qd.soccfg`, the local QickConfig that qick’s `make_proxy()` builds from `soc.get_cfg()`. Fixed 2026-08-17; the check now runs.

6.4 Verification: the guarantees, checked on the data

Streaming’s value is its exactness, so Step 5C checks it rather than assuming it. On the 2026-08-14 run:

Check	Result	Meaning
<code>rep_index</code> gaps	0	no reps dropped; cadence timestamps valid throughout
<code>rep_index</code> step	4, uniformly	host binning consistent, no straddled bins
Duplicate rows	0 of 31 250	the replay race really is absent
Reps collected	125 000 / 125 000	nothing lost
Packets	152	38–500 reps each, median 205
Duty cycle	100.0%	the FPGA never idled

`stream()` also now refuses to reshape a packet whose payload size does not match its own rep count. Such a packet would silently misalign every subsequent sample and mix the two parked frequencies together — which would look exactly like a flat broadband noise excess. No occurrence has been observed, but it would previously have been invisible.

6.5 The buffer-overflow constraint

The board worker transfers in strides of roughly 10% of the accumulated buffer and raises if unread samples reach `avg_maxlen`. `stream_headroom()` reports the actual margin at runtime rather than assuming it:

```
{avg_maxlen, reads_per_shot, shots_that_fit, seconds_that_fit, worker_stride_shots}
```

At $\tau = 120\ \mu\text{s}$ with one read per shot the cadence is $\sim 4\ \text{kHz}$, so even a modest buffer allows a poll interval of order a second. No overflow has been observed, and the 500-sample first packet is no noisier than the 205-sample ones, which argues against overflow being behind the noise excess.

6.6 The PL droop

Real, and specific to this mode

With no gap between acquires the sample heats. On the 2026-08-14 run the PL fell **25% over 30 s**, smoothly, from 993 to 744 ADC counts.

It is mostly common-mode and largely cancels in $z_+ - z_-$: it is 85% of the *common-mode* variance but only **2.8%** of the Δf variance. But it walks the operating point down the flank, so the slopes the conversion assumes drift out of validity over the run.

Step 5C detrends it (rolling median, corner $\sim f_s/\text{window}$) and reports the magnitude. The proper fix is a periodic re-zero during the run, which is not implemented.

6.7 When to use it

For anything that needs a *continuous* record above a few hundred Hz — MCG, drone transients — streaming is the only method that provides it: no gaps, exact timestamps, 100% duty cycle. Its rate is proven.

Its noise floor is currently $\sim 6\times$ worse than burst, which for absolute sensitivity work makes burst the better choice today. **Resolving §9.2 is the single highest-value open item in this document**, because it would make streaming better than both other methods at every operating point.

Chapter 7

Timing, in full

Chapters 4–6 gave each method’s budget. This chapter gives the evidence that the model behind them is right, and the practical consequences.

7.1 The evidence for a per-call floor

Applying the exact-recovery technique of §3.4 to all 21 usable averaged runs across three sessions (tables/timing.csv):

Session	Config	ADC integration	Fastest call	Median period	Rate
2026-08-04	$D=654340$	4.26 ms	9.35–9.53 ms	10.48–10.62 ms	94–95 Hz
2026-08-04	$D=706560$	4.60 ms	9.29 ms	10.84 ms	92 Hz
2026-08-04/06	$D=1504982$	9.80 ms	9.61–10.29 ms	10.93–11.54 ms	87–91 Hz
2026-08-14	$D=368640$	2.40 ms	10.14–10.30 ms	11.26–11.52 ms	87–89 Hz
2026-08-14	$D=847872$	5.52 ms	10.27 ms	11.39 ms	87.8 Hz

Two runs were excluded from the fit below because their loop period exceeds $1.5\times$ their acquire time — most of it was spent outside `acquire()` (live plotting on, or operator interaction), so they say nothing about the call.




Figure 7.1: Left: measured batch period against FPGA readout work. The dashed line is the serial model the data does not follow; the red dotted line is the per-call floor. Right: the same data as rate — a $4\times$ change in FPGA work leaves the rate at ~ 88 Hz.

7.1.1 Argument 1 — the slope is not 1

Fitted *within* a session, because comparing across sessions would confound the FPGA change with the floor itself, which moved ~ 1 ms between 2026-08-04/06 and 2026-08-14:

Session	FPGA work	Period	$d(\text{period})/d(\text{FPGA})$
2026-08-04/06	4.26 $\rightarrow$ 9.80 ms ($2.3\times$)	10.55 $\rightarrow$ 11.43 ms	+0.140
2026-08-14	2.40 $\rightarrow$ 5.52 ms ($2.3\times$)	11.38 $\rightarrow$ 11.39 ms	+0.003

A serial model period = host + FPGA requires slope = 1.000. Neither session comes close.

7.1.2 Argument 2 — no serial model fits at all

The 2026-08-14 runs used 2.40 ms of FPGA work, and their *fastest* call was 10.14 ms. Under a serial model that demands a host term of 7.74 ms, against the 1.60 ms the 2026-08-06 analysis reported for the same code on the same rig eight days earlier. There is no constant, non-negative host term that fits both sessions.

Under the max-model both are consistent: the host chain is ~ 10 –11.4 ms in both, and the FPGA work is simply hidden inside it.

7.1.3 Argument 3 — the wobble is entirely in the call

The instantaneous rate spans 81.4–100.1 Hz, which is what is seen live as “85–97 Hz”. Decomposing it:

- The FPGA term is **deterministic** — 23 reps of identical pulse work vary by well under a microsecond, and the burst runs confirm the cadence is stable to 0.4% across runs.
- `acq_seconds` spans 10.1–12.1 ms (jitter 1.6–1.9 ms p95–p05).
- Python outside the call stays at 0.06 ms.

So **100% of the wobble is host/network scheduling inside `acquire()`** — Windows scheduling plus TCP/Pyro latency on stages 1–6.

7.2 The corrected model

$$\text{period} = \max\left(\underbrace{H}_{\text{host call}}, \underbrace{\text{reps} \times t_{\text{rep}}}_{\text{FPGA}}\right) + \underbrace{P}_{\text{Python, } \sim 0.06 \text{ ms}} \quad (7.1)$$

with, on this rig,

$$H_{\text{typical}} = 11.4 \text{ ms}, \quad H_{\text{floor}} = 10.1 \text{ ms}.$$

Two constants rather than one, because the two uses want different ends of the distribution: H_{typical} predicts the rate, H_{floor} sizes the free averaging (fill past the fastest call and even the quick calls become FPGA-bound).

7.2.1 Predicted against measured

τ	reps	FPGA	Old model	Eq. 7.1	Measured
213 μs	23	9.80 ms	160 Hz	87.7 Hz	87.5 Hz
120 μs	23	5.52 ms	400 Hz	87.7 Hz	87.8 Hz
120 μs	10	2.40 ms	180 Hz	87.7 Hz	86.8 Hz

The model giving the *same* answer for all three is the point.

`measure_host_floor()` calibrates H on the rig in about a second, using a `reps=1` probe whose FPGA work is 240 μs — so whatever the call still costs is the host chain, measured rather than inferred. Step 4A calls it at startup. Given that H already differs by 1.2 ms between the two sessions, the defaults above should be treated as a starting point.

7.3 Free averaging

The most actionable consequence in this document

If the FPGA work hides under the call, then **running fewer reps than fit is pure loss**. The call takes the same wall-clock time either way, so the unused milliseconds are photons that were never collected.

At $\tau = 120\ \mu\text{s}$, $t_{\text{rep}} = 240\ \mu\text{s}$, so $H_{\text{floor}}/t_{\text{rep}} = 10.1/0.240 = \mathbf{42\ reps}$ fit inside the floor. The 2026-08-14 runs used **10**.

reps	FPGA	Duty cycle	Rate	Note
10	2.40 ms	21%	87.7 Hz	what ran
23	5.52 ms	48%	87.7 Hz	
42	10.08 ms	88%	87.7 Hz	the fill point
50	12.00 ms	100%	83.3 Hz	now FPGA-bound; rate starts to fall
100	24.00 ms	100%	41.7 Hz	

Implemented as `reps_that_fit()`, used automatically when `AVG_REPS_PER_BATCH = None`. `describe_timing()` prints the headroom:

```
2 freqs, order=forward (2 slots/rep), tau=120.0 us, 10 reps -> 240 us/rep, 2.40
ms FPGA work
averaged mode: 2.40 ms of FPGA work hides under the 10.1 ms per-call floor (24%
used) -> 87.7 Hz
FREE AVERAGING: 42 reps also fit inside the floor. Raising 10 -> 42 costs no rate
and buys up to 2.0x in sigma if the noise is white.
```

The caveat is in that last clause

“if the noise is white” — and §8.4 shows it currently is not. The averaged mode’s noise does not average down at all; it gets *worse* with more reps. Filling the budget is still free, so it is still worth doing, but do not expect the full $\sqrt{4.2} = 2.0\times$ until the per-acquire noise term is understood.

7.4 Knob by knob

Knob	Effect on rate	Effect on sensitivity	Verdict
readout_integration_tus (τ)	None in averaged mode below the floor; <i>is</i> the cadence in burst/stream	Flat for $\tau \leq 140\mu\text{s}$, worse above (§8.1)	Not a rate lever in Method A. Keep at $120\mu\text{s}$
reps	None below the floor; linear above	Should be $1/\sqrt{N}$; measured far less (§8.4)	Free averaging up to the floor
Acquisition method	12–47 $\times$	See Chapter 8	The only real rate lever
multipoint_skip_reference	2 $\times$ (already on)	Removes a contaminated normaliser	Keep on
multipoint_pair_order	2 $\times$ per rep for "abba"	Cancels linear common-mode drift; free at equal averaging	OPEN , untested
n_freqs	Linear	2 is the minimum for a two-point estimator	Fixed at 2
BURST_REPS	Sub-linear; flattens past ~ 1000	None directly	500–1000
STREAM_TARGET_HZ	Sets the bin divisor exactly	$1/\sqrt{\text{bin}}$ if white	Free choice
LIVE_SHOW_PLOT	20–100 ms per redraw	None	Keep False
SAVE_EVERY_SEC rewrite	25–220 ms every 5 s	None	Fixed: append-only

7.5 The 1 kHz question, closed

The 2026-08-06 analysis concluded that 1 kHz was unreachable with one call per sample. That conclusion **survives the retraction and in fact gets stronger**: the per-call cost is ~ 11 ms, not 1.6 ms, so a 1 ms period is unreachable by a factor of eleven rather than by 1.6.

The route it identified — one `start_readout` with continuous polling — was built, and on 2026-08-14 it delivered **1041.7 Hz** exactly as designed (§6.2). The rate target is met. Only its noise is open.

Chapter 8

Sensitivity and noise

8.1 The readout-window optimum

From 92 ODMR sweeps at 17 readout windows on 2026-08-06 (docs/2026-08-06_twopoint_timing/tables/integration_time.csv). Lorentzian fits over 2840–2900 MHz; noise taken from the point-to-point scatter of the off-resonance wings, which is immune to slow drift across the sweep.

τ (μs)	contrast	σ_z	$\sigma_z\sqrt{\tau}$	$\sigma(\Delta f)$ kHz	t_{rep} μs	η nT/ $\sqrt{\text{Hz}}$
50	0.541	3.01e-3	0.0213	51.3	109	186 ± 14
90	0.517	2.13e-3	0.0202	37.0	189	177 ± 14
110	0.515	1.90e-3	0.0199	33.7	229	178 ± 8
120	0.514	1.98e-3	0.0217	34.9	249	192 ± 17
130	0.507	1.70e-3	0.0194	30.6	269	175 ± 15
140	0.506	1.71e-3	0.0202	30.6	289	181 ± 7
170	0.513	2.01e-3	0.0262	35.6	349	232 ± 5
190	0.495	2.09e-3	0.0288	38.4	389	264 ± 22
213	0.531	1.52e-3	0.0222	25.9	435	188 ± 14

Three results:

1. **Contrast is flat** (0.49–0.54) across the whole range. A longer window buys no signal.
2. $\sigma_z\sqrt{\tau}$ is **flat below $\sim 140 \mu\text{s}$ (0.0207) and rises above it (0.0252)**. Below $140 \mu\text{s}$ the readout averages down as $\sqrt{\text{time}}$; above it, the extra window length collects drift instead of statistics. *This is the physical reason an optimum exists.*
3. η is **flat at $185 \pm 7 \text{ nT}/\sqrt{\text{Hz}}$ for $\tau \leq 140 \mu\text{s}$** , degrading to 232 ± 20 over 150 – $200 \mu\text{s}$. Individual windows inside the flat band are not distinguishable — the sweep-to-sweep scatter at fixed τ is ± 14 , larger than the spread between them.




Figure 8.1: Contrast, noise scaling and sensitivity against readout window, 2026-08-06.

Decision: $\tau = 120 \mu\text{s}$. Inside the flat band, and $1.75\times$ faster *per rep* than $213 \mu\text{s}$. Note carefully that this does **not** make Method A faster (Chapter 7) — it makes burst and streaming faster, and it lets more reps fit under Method A’s call floor.

Two caveats that remain

The 213 μs point is an outlier. It reads $188 \text{ nT}/\sqrt{\text{Hz}}$, well below the $150\text{--}200 \mu\text{s}$ trend it should continue, on an unusually low σ_z ($1.52\text{e-}3$). It was the configured default, so those sweeps were probably taken under different conditions. The $\sigma_z\sqrt{\tau}$ trend across $150\text{--}200 \mu\text{s}$ is the more reliable signal.

The shape assumes reps was constant across the scan. That is the natural reading of a one-variable scan, and it is what flat $\sigma_z\sqrt{\tau}$ implies, but `reps` is not recorded in the sweep CSVs and cannot be recovered from file timestamps (those gaps are dominated by operator cadence: median 4s while t_{rep} varies $4\times$). A `reps` change part-way through could mimic the rise above $140 \mu\text{s}$.

8.2 Measured sensitivity, per method

From `tables/modes.csv`. Both conventions of §2.3 are given; when they disagree, trust the floor.

Session	Method	Run	Rate	$\sigma(\Delta f)$	η	Floor
08-14	averaged	144602	87.8 Hz	56.4 kHz	304	278
08-14	averaged	144655	88.8 Hz	98.1 kHz	526	501
08-14	averaged	145010	87.9 Hz	92.1 kHz	496	486
08-14	averaged	145321	86.8 Hz	54.6 kHz	296	274
08-14	stream	145442	1041.7 Hz	160.8 kHz	251	240
08-14	burst	150458	4166.7 Hz (3089 net)	71.4 kHz	56	38
08-06	averaged	144936	86.7 Hz	95.0 kHz	515	499
08-06	burst	145051	2347 Hz (2108 net)	256.1 kHz	267	158

(η and floor in $\text{nT}/\sqrt{\text{Hz}}$.)




Figure 8.2: Rate against measured noise floor by method, 2026-08-14. The runs were taken minutes apart rather than interleaved — see the warning below.

8.2.1 The photodiode fix: $4.1\times$

The headline sensitivity result

Burst-mode noise floor, same two parked points, same measurement:

$$2026-08-06: 158 \text{ nT}/\sqrt{\text{Hz}} \longrightarrow 2026-08-14: \mathbf{38 \text{ nT}/\sqrt{\text{Hz}}} \quad (4.1\times)$$

attributable to the reference-photodiode gain correction and the removal of signal-diode saturation. Burst mode is the only method that currently shows the full benefit.

8.2.2 A warning about comparing these rows

These runs are not strictly comparable

They were taken minutes apart, not interleaved, and §8.4 shows the per-rep noise varied $3.6\times$ between two averaged runs **three minutes apart** with identical settings (145010 vs 145321: 5.97% vs 2.45%).

Any A/B comparison across separate runs on this rig is unreliable. Only interleaved measurements settle a method comparison, which is why the rig test in §12.1 interleaves. Treat the method ranking above as indicative.

8.3 Averaging does not buy $\sqrt{N}$

Block-averaging the real burst samples within each burst, 2026-08-06:

reps averaged	$\sigma(\Delta f)$ kHz	ideal $1/\sqrt{N}$	excess
1	248.6	248.6	1.00
4	141.0	124.3	1.13
8	110.8	87.9	1.26
23	86.7	51.8	1.67
64	76.5	31.1	2.46
128	72.3	22.0	3.29

The predicted 86.7 kHz at $N = 23$ matches the averaged runs, which came in at 73–110 kHz for 11 of 12 runs (median 94 kHz) — an independent consistency check, since the two are measured in completely different modes. But the excess grows with N : past ~ 30 reps, averaging is nearly free of benefit.




Figure 8.3: Noise spectrum and block-averaging behaviour, 2026-08-06. Above ~ 10 Hz the spectrum is white at ~ 165 nT/ $\sqrt{\text{Hz}}$; the excess is all below 10 Hz.

8.4 The noise is not photon shot noise

This is the sharpest version of the same finding, and the most important open item in the document. Per-rep normalised-PL noise, scaled back from whatever averaging each run used (tables/per_rep_noise.csv):

Session	Method	reps averaged	per-rep $\sigma(z)$	relative
2026-08-06	burst	1	0.0076	0.67%
2026-08-06	burst	1	0.0091	0.83%
2026-08-06	averaged	23	0.0114	1.11%
2026-08-06	averaged	23	0.0161	1.38%
2026-08-14	burst	1	0.0059	0.61%
2026-08-14	averaged	10	0.0214	2.45%
2026-08-14	averaged	10	0.0513	5.97%
2026-08-14	averaged	23	0.0755	8.95%
2026-08-14	stream	4	0.0522	7.57%

Two observations:

1. **Burst is reproducibly 0.61–0.83% in both sessions; everything else is 1.1–9%.** Burst measures the per-rep noise *directly*; the other rows infer it by scaling up a reps-averaged σ . The gap means the inference is wrong — the noise in those modes is not a per-rep term.
2. **Averaged mode gets *worse* with more reps** (10 reps $\rightarrow$ 2.45%, 23 reps $\rightarrow$ 8.95%). Shot noise cannot do that.

OPEN — the per-acquire noise term

The dominant noise term is **per-acquire, not per-rep**, so it survives averaging entirely. **If it were removed, Method A would inherit Method B's $4\times$ better noise at no cost in rate**, and free averaging (§7.3) would deliver its full $\sqrt{N}$.

Candidate mechanisms, none tested: a start-of-batch transient whose amplitude varies with the (jittery) idle gap between calls; ADC offset drift re-sampled per call; laser/MW state settling after the gap. The row-0 transient measured in burst mode (+1.4%) is direct evidence that *something* happens at a batch boundary.

Protocol: §12.1, check S6.

8.5 Mains pickup and aliasing

From `tables/spectra.csv`:

Method	Rate	Strongest line	Excess over floor	60 Hz appears at
averaged	86.8–88.8 Hz	—	—	26.8–28.8 Hz (aliased)
stream	1041.7 Hz	61.7 Hz	10.2×	60 Hz
burst	4166.7 Hz	58.5 Hz	3.0×	60 Hz

The mains line is real and strong — **10×** **the noise floor** in the streamed data, with odd harmonics at 185, 308 and 465 Hz.

At 88 Hz, 60 Hz folds to ~27 Hz and becomes unfixable

Nyquist at 87.9 Hz is 43.9 Hz, so a 60 Hz tone aliases to $|60 - 87.9| = 27.9$ Hz — the middle of the measurement band. Once folded there it is indistinguishable from a real 27.9 Hz signal, and notching it would remove real signal along with it.

This is a structural argument for running at ≥ 1 kHz for drone and MCG work, *independent of sensitivity*: only there can the pickup be identified for what it is and removed. `twopoint_spectra.alias_report()` raises this automatically and Step 5A prints it on every averaged run.

Note that the 2026-08-06 analysis reported “no 60 Hz line”. That was correct for what it could see: at 87.5 Hz the line is not *at* 60 Hz, and the burst runs of that session, which could have resolved it, were half replays with a badly compressed time axis. The line was there all along.

Chapter 9

Open defects

9.1 Burst mode replays about half of every run

9.1.1 The evidence

Every burst run in both sessions shows the same signature: acquires strictly alternate short and long, and the short ones return a *bit-identical* copy of the previous batch.

Run	Batches	Stale	Fraction	Pattern	Recorded	True
2026-08-06 _142358	211	106	50.2%	13 / 425 ms	3911 Hz	1946 Hz
2026-08-06 _142443	66	33	50.0%	13 / 424 ms	4182 Hz	2091 Hz
2026-08-06 _145051	44	23	52.3%	13 / 425 ms	4421 Hz	2110 Hz
2026-08-06 _145144	14	7	50.0%	13 / 425 ms	4455 Hz	2227 Hz
2026-08-14 _150458	371	185	49.9%	12 / 120 ms	6161 Hz	3089 Hz

The separation is clean with nothing in between: the duplicate fraction is **0.957 in the fast batches** and 0.069 in the real ones. Within a stale batch the only rows that differ from the previous batch are row 0 and a block around rows 770–818 — and 819 is exactly the streamer’s transfer stride, $0.1 \times \text{avg_maxlen}/\text{reads_per_shot}$.

Detection uses two independent tests, either of which condemns a batch:

1. $\geq 90\%$ of rows identical to the previous batch;
2. wall-clock duration under half the FPGA work the batch claims to contain.

Both are needed. On _145051, batch 24 is 82% duplicate — under the fraction threshold, because the run’s one CSV-flush stall shifted the alignment — but returned in 12.9 ms against 425 ms of claimed pulse work.

9.1.2 The 2026-08-06 root cause, and why it is now disproved

DISPROVED — `config_all` / `reset_tproc` was not the mechanism

The 2026-08-06 analysis found a real bug. `quickdawg` called `config_all(..., load_pulses=...)`, but `quick` 0.2.386 names that parameter `load_envelopes`. **Every acquire raised `TypeError`**, was silently caught, and fell back to `config_all(soc)` with `reset=False` — and `reset=False` means `stop_tproc(lazy=True)`, which `quick` documents as a no-op on `tProc` v1. **The `tProc` was never stopped between runs**, so the accumulated buffer was re-armed under a possibly-still-running program.

That was fixed on 2026-08-12 (commit 942d9e8) and a `reset_tproc=` argument was threaded through `acquire()`.

The 2026-08-14 run used the fix and the replay fraction did not change: 49.9% (185/371). The rig was definitely running the new code — $\tau = 120 \mu\text{s}$ took effect and the new streaming cell produced output.

So the keyword bug was real and worth fixing, but it was *not* the cause of the replay. **The root cause is unknown.**

What the elimination does tell us: the mechanism is not the `tProc` failing to stop, because forcing the stop changes nothing. The remaining candidates involve the accumulated buffer or the shot counter rather than the program:

- `config_bufs()` re-arming the accumulator without clearing the counter, so `poll_data` returns the previous run’s completed shots as if they were new;
- the board-side worker thread from the previous `start_readout` not being joined, so its queue still holds the old transfer;
- `start_readout`’s shot counter being read before the `tProc` has reset it.

None has been tested. A single instrumented burst run on the rig, printing the shot counter and the queue depth before and after each `start_readout`, would discriminate.

9.1.3 Mitigation, which is in place

`cfg.multipoint_on_stale = "drop"` makes the freshness guard re-acquire (`multipoint_stale_retries`, default 2) and, if every attempt still returns the same buffer, return `None` so the caller **skips the batch**. A replay is never written to the CSV as though it were data.

This costs duty cycle, not correctness — the recorded rate becomes honest: 3089 Hz rather than the 6161 Hz the raw file claims. Simulated against a board that replays every second call, all 20 requested batches were recovered as real data.

The guard is a byte comparison (`blake2b`) against the previous raw buffer, so it costs microseconds and catches the failure *whatever* its mechanism — which is why it survived the root cause being wrong.

9.2 Streaming’s noise floor is about $6\times$ the burst floor

9.2.1 The evidence

Streaming and burst ran ten minutes apart at the same τ , on the same hardware, with nothing changed on the bench (confirmed with the operator). Their amplitude spectral densities differ by a **flat factor of ~ 6 across the whole band** (`tables/read_path.csv`):

Band	Stream	Burst (cleaned)	Ratio
10–30 Hz	234.8	38.6	6.08
30–60 Hz	234.1	44.1	5.31
60–100 Hz	240.8	39.8	6.06
100–200 Hz	231.1	38.7	5.97
200–350 Hz	242.4	38.8	6.25
350–500 Hz	250.2	39.7	6.30

$(nT/\sqrt{Hz.})$



Figure 9.1: Streaming against burst amplitude spectral density. Both floors are flat; the ratio is constant across two and a half decades.

9.2.2 2026-08-17: still there, but far smaller

The 2026-08-17 morning session gives an independent look at the same comparison, after the photodiode gain change and on a different computer. It is a weaker measurement than the 08-14 one — the runs are ten minutes apart rather than interleaved, and the whole session was a magnet test with the magnet moved by hand — so each run is quoted twice: over the full 30 s, and over the first 5 s before the magnet was brought in at all. From `tables/0817_read_paths.csv`, medians over 4 stream and 6 de-staled burst runs:

Band	full 30 s			first 5 s, magnet away		
	Stream	Burst	Ratio	Stream	Burst	Ratio
10–30 Hz	113.3	78.4	1.45	127.6	95.9	1.33
100–200 Hz	105.5	77.2	1.37	131.2	86.2	1.52
350–500 Hz	105.6	78.0	1.35	137.3	87.8	1.56

$(nT/\sqrt{Hz.})$




Figure 9.2: 2026-08-17 morning. Points are individual runs, lines the median. The run-to-run scatter within each mode is a factor of ~ 2 , which is the main reason this session cannot replace the interleaved A/B.

The excess is real but it is $1.3\text{--}1.6\times$, not $6\times$

Streaming is still worse than burst, still by a flat factor across the band — the signature has not changed. But the size has: **$1.3\text{--}1.6\times$** here against **$5.3\text{--}6.3\times$** on 08-14. Both windows agree, so it is not the magnet.

Three things differ between the sessions and none has been isolated: the photodiode gain change, `mw_gain` (this session ran at 10 500; the current notebook default is 15 000), and the `stream()` configuration sequence, which was aligned with `acquire()`'s on 2026-08-16 — *after* the 08-14 data and *before* this session. That last one is the leading candidate precisely because it was the only difference in the read path, and this is the first data taken with it.

Do not treat the $6\times$ figure as current. Do not treat $1.4\times$ as settled either: `profile_twopoint_acquire.py -compare-read-paths` interleaves the two modes on the same program and is still the measurement that closes this.

9.2.3 What has been ruled out, with the measurement

Candidate	Measurement	Verdict
Mains pickup	The 60 Hz comb (8 harmonics) accounts for 4.5% of the stream's Δf variance	ruled out
Slow drift / the PL droop	Below ~ 10 Hz is 2.8% of the Δf variance (it is 85% of the <i>common-mode</i> variance, but that cancels)	ruled out
Dropped reps	0 <code>rep_index</code> gaps in 125 000 reps	ruled out
Duplicated samples	0 duplicate rows in 31 250 samples	ruled out
Packet-boundary artefacts	σ at the first 5 rows of a packet is $0.96\times$ the mid-packet σ (30.7 vs 31.8 ADC)	ruled out
Buffer overflow as packets fill	The 500-sample first packet is <i>no noisier</i> than the 205-sample ones	ruled out
Torn reads (partial accumulations)	The residual is symmetric, skew $+0.12$, kurtosis 1.24. Torn reads would skew strongly low	ruled out

A flat, structureless, frequency-independent multiplicative excess is what a **read-path** difference looks like, not what a physical disturbance looks like. Both paths pull from the same accumulated buffer; they differ in how they configure and drain it.

9.2.4 What has been done about it

`stream()`'s configuration sequence now mirrors `NVAveragerProgram.acquire()` step for step. It did not before:

	<code>acquire()</code>	<code>stream()</code> , before 2026-08-16
<code>config_all</code>	yes	yes
<code>start_src</code>	yes	yes
<code>config_bufs</code>	yes	yes, <i>different order</i>
<code>reload_mem</code>	yes	missing entirely
<code>start_readout</code>	yes	yes

That is the cheapest candidate to eliminate and it costs nothing if irrelevant. `stream()` also now refuses to reshape a packet whose payload size does not match its own rep count, which would misalign every subsequent sample and mix the two parked frequencies — a failure that would produce exactly this signature and that was previously invisible.

OPEN — neither change is verified to fix it

Highest-priority open item. The test is designed and takes about a minute:

```
python scripts/profile_twopoint_acquire.py -compare-read-paths -tau 120 -reps 500
```

It alternates `acquire(per_rep=True)` and `stream()` on the *same program object*, five times, within seconds. Identical FPGA work; only the read path differs. It compares raw normalised counts rather than kHz, so it does not depend on the ODMR calibration still being valid.

- ratio $\approx 1.0 \Rightarrow$ the $\sim 6\times$ was the environment after all, and streaming needs no caveat;
- ratio $\gg 1 \Rightarrow$ streaming really is noisier on identical work, and it is a software defect. Follow with `-stream-scan` to see whether it grows with buffer fill.

9.3 An interrupted cell wedges the board, and a kernel restart does not clear it

Found 2026-08-17, when Step 4C hung at its auto-zero `acquire()` and Ctrl-C produced a traceback ending in `sock.recv(size, socket.MSG_WAITALL)`. Fixed the same day. It is recorded here because it silently contaminates data rather than only costing time, and because the recovery is not obvious.

9.3.1 The mechanism

`qick`'s `poll_data()` takes a `timeout` argument that defaults to `None`, and its board-side body is

```
length, data = streamer.data_queue.get(block=True, timeout=timeout)
```

so with the default it blocks *forever*. `qickdawg`'s drain loop called `qd.soc.poll_data()` with no arguments, so every `acquire` inherited that default. The queue is fed by the board's readout worker, which only transfers once the `tProc` shot counter advances past a stride. If the counter never advances — a program that did not start, a `tProc` left running by an earlier cell, an abandoned `stream()` generator — nothing is ever queued and the RPC never replies.

The client then sits in `sock.recv(..., MSG_WAITALL)` waiting for a reply that is not coming. That is the traceback above, and the four consequences are:

1. **The wait is unbounded.** No timeout, anywhere in the path.

2. **Ctrl-C only unblocks the client.** The board-side Pyro worker thread stays blocked in `queue.get()` for the life of the daemon.
3. **That thread is still a consumer.** When the next run finally queues a packet, the abandoned `get()` can take it instead of the live caller — so a subsequent run silently loses its first packet.
4. **Restarting the notebook kernel does not help**, because the wedged thread is on the board, not in the client.

9.3.2 Reproduced offline

Against the real `qick.streamer.DataStreamer` driven by a fake board whose shot counter never advances (`scripts/` is not needed; the check is eight lines):

call	argument	result
<code>poll_data()</code>	<code>timeout=None</code> (the default)	never returned; killed at 3 s
<code>poll_data(totaltime=0.1, timeout=1.0)</code>	<code>finite</code>	returned <code>[]</code> after 1.01 s

9.3.3 The first fix, and why it had to be withdrawn

title=Bounding the poll broke two of the three methods. Do not reinstate it.

The obvious fix — pass a finite timeout so the block becomes a `queue.Empty` — shipped on 2026-08-17 at 14:23 (commit 9832e07) with `POLL_TIMEOUT_S = 1.0`. It was verified against a hand-written mock, never against the board. Two hours later, on the rig:

Step	Mode	reps	Result
4A	averaged	35	worked, 104.5 Hz
4B	burst	2000	<code>TimeoutError: acquire() read 0 of 2000 shots and then saw nothing for 10.0 s</code>
4C	stream	2000	hung <i>anyway</i> , in <code>Pyro4 socketutil.recvall</code>

Both failures are at the same call — the dedicated large-reps auto-zero acquire — and 4A never makes one, which is why it survived.

The reason a bounded poll is not a safe drop-in is in `poll_data`'s body (`qick/qick.py`):

```
time_end = time.time() + totaltime
while streamer.count < streamer.total_count and time.time() < time_end:
    length, data = streamer.data_queue.get(block=True, timeout=timeout)
    if streamer.stop_flag.is_set() or data is None:
        break          # <-- the packet is DISCARDED, not counted
    streamer.count += length
    new_data.append((length, data))
```

With `timeout=None` the `get` blocks until a packet exists, so the RPC always makes forward progress and a slow first stride is simply a wait. With a finite timeout it returns `[]` instead, and the client loop turns that into a hard failure — a transient board-side condition becomes a dead run. The `stop_flag` branch makes it worse: it drops the packet it just took, so a bounded poll can consume data and report none.

9.3.4 What is actually shipped

title=Blocking poll, bounded cleanup, explicit recovery

1. **The poll blocks, as qick intends.** `POLL_TIMEOUT_S = None` is the default in `NVaveragerProgram`, and both `_drain_readout()` and `stream()` read it from there so the two paths cannot drift apart. Set it to a float for a diagnostic run and the bounded behaviour — including `stream()`'s `poll_timeout_s` — comes back.
2. **An empty return is an error, not a spin.** Under a blocking poll, `[]` cannot mean “the board is slow”; it means the readout ended, or somebody else drained the queue. Both loops raise, naming how many shots they got. The original code spun on the same condition, hammering the board with RPCs; a naive fix would have short-read zeros into the tail of `d_buf` instead.
3. **Interrupts still clean up** — this is the half of 9832e07 worth keeping, and it works fine with a blocking poll, because Pyro raises `KeyboardInterrupt` inside `sock.recv`. The drain loop catches `BaseException` and `stream()` has a `finally`, which for a generator also runs on `GeneratorExit`. Both call `_abort_readout()`: stop the `tProc`, then stop the readout.
4. **Recovery is explicit and staged.** `recover()` first stops the `tProc`, stops the readout and *drains the data queue* — the draining is what keeps the next run from picking up leftovers, and it is enough after an ordinary `Ctrl-C`. Only if the board is still not idle does it escalate to `DataStreamer.start_worker()`, which installs fresh queues so anything blocked on the old one is orphaned, at the cost of leaking a board-side thread. `recover(force=True)` goes straight there.
5. **ensure_idle() reports; it does not act.** It used to call `recover()` automatically when it found the board busy. Replacing the streamer's queues implicitly, at the top of every acquisition cell, possibly on a false positive from the probe, is a larger risk than the condition it guards. It now prints and returns `False`, and Steps 4A/4B/4C raise and point at Step 0.

title=Still open: does the wedge explain the burst replay? Probably not.

Consequence 3 above is a mechanism by which one run's packet is delivered to a *different* run's consumer, which fits the alternating short/long signature.

The 2026-08-17 morning session argues against it. Those runs were taken on a second computer at commit 8fd303f — before any of this — and their stale fraction is **50.0–50.8% across six burst files** (`tables/0817_burst_audit.csv`), the same as 08-06 and 08-14. A wedge is an incident; it should not reproduce at exactly one batch in two, on two machines, across three sessions. Whatever causes the replay is structural.

Still worth measuring, and free: run a burst now that the poll is blocking again and read `BURST_STALE_FRACTION`.

9.4 Averaged and burst runs used to overwrite each other's file-names

Step 4A and Step 4B both wrote `twopoint_lockin_live_<stamp>.csv`. Nothing was lost — the timestamps differ — but an averaged run and a burst run were indistinguishable by name, which is why identifying which recorded file was which mode has repeatedly needed column inspection rather than a glance at the directory.

Fixed 2026-08-17: Step 4A writes `twopoint_lockin_avg_*`, Step 4B writes `twopoint_lockin_*`

`burst_*`, Step 4C already wrote `twopoint_lockin_stream_*.find_live_csvs()` in `analyze_twopoint_session.py` accepts all three prefixes so previously recorded sessions still load unchanged. Files already on disk keep their `_live_` names and are still found; only the mode has to be read from the columns, which `detect_mode()` does anyway.

9.5 Post-processing: recovering runs already recorded

Independently of any hardware fix, existing burst files can be made usable. The pipeline is `process_burst()` in `twopoint_postprocess.py`, reachable from the command line as `scripts/clean_burst_lockin.py`. It applies, in order:

1. **Stale-batch removal**, by the two tests above. Reports the fraction.
2. **Row-0 transient removal** from every batch.
3. **Re-timing on the exact cadence** from §3.4, with each surviving burst marked as its own segment so the inter-burst dead time is an explicit gap. Nothing is interpolated across it and spectra are computed per segment.
4. **Differential despiking** — a causal Hampel filter on `peak_shift_kHz`.

Order matters. Staleness is judged on the *unfiltered* frame, because the batch table reconstructs each batch’s start time from its first row; dropping row 0 first would shift every start by half a sample.

What it cannot do. The samples lost to the dead time between bursts are not there, and the genuine single-rep glitches that were double-counted are *de*-duplicated, not explained. It recovers a correct 3.1 kHz record from a 6.2 kHz one; it does not recover 6.2 kHz.

9.5.1 A bug this found: never despiking the channels separately

Per-channel despiking makes the data worse

The two parked channels are **0.84–0.97 correlated** through the common-mode PL. A Hampel filter run on them independently flagged **30 and 23** samples with only **4 in common** — so most replacements are one-sided, and each one-sided replacement breaks the cancellation the estimator depends on and *injects* differential noise.

Measured effect:

	raw	per-channel	differential
averaged 145321	54.6 kHz	55.8 kHz	53.6 kHz
stream 145442	160.8 kHz	175.5 kHz	155.2 kHz

The default is now `despike="shift"`, which filters the differential. Per-channel remains available as `despike="channels"` for diagnosing raw-channel glitches, and prints a warning saying it will usually raise σ .

The live-loop despiker is a separate case and legitimately per-channel: there, the raw channels must be cleaned *before* the conversion exists. It is off by default (`AVG_DESPIKE_ENABLED = False`).

Chapter 10

Streaming the multipoint lock-in

Everything so far has been the two-point method: one pair of parked frequencies on one flank pair, giving the field projection along a single NV axis. The multipoint notebook (`Modules/Lockin_module.ipynb`) parks 16 frequencies instead, which is what makes a full three-axis reconstruction possible — and until 2026-08-17 it had only the slowest of the three acquisition paths.

This chapter is what changes when `stream()` is pointed at it. Nothing in the FPGA program needs to change: `emission_order()`, `_fold_slots()` and `stream()` are all written in terms of n_{slots} , so 16 parked points work exactly as 2 do. What changes is the arithmetic, and one constraint that has never bound on the two-point path.

10.1 Why it is worth doing

`acquire()` costs 8.5–11.4 ms of host round-trip whatever the program does (Ch. 7), and Steps 4/4B/4BN call it once per sample. With 16 points and a reference the FPGA work is 3.84 ms — comfortably *under* the floor — so the live loop is host-bound at

$$\frac{1}{\max(10 \text{ ms}, 3.84 \text{ ms})} \approx 26 \text{ Hz},$$

and no FPGA setting moves it. `stream()` pays the host cost once for the whole run, so the rate becomes the cadence itself:

Path	slots	reads/slot	t_{rep}	Rate	Note
Step 4 / 4B / 4BN (live loop)	16	2	3.84 ms	26 Hz	host-bound, not FPGA-bound
Step 4S , reference kept	16	2	3.84 ms	260 Hz	the default; reconstructable
Step 4S, <code>skip_reference</code>	16	1	1.92 ms	521 Hz	projections only — see below

A factor of ten, and it comes with the two structural properties from Ch. 6: nothing is re-armed mid-run, so the replay race that costs burst mode half its batches cannot occur; and the samples are consecutive reps of one program, so the time axis is the cadence exactly.

The reference readout stays on, and that is not a rate compromise

It is tempting to set `multipoint_skip_reference = True` and double the rate, as the two-point path does. Do not, for any run you intend to reconstruct.

The vector inversion consumes `peak_NN / peak_NN_ref` — each parked point normalised by its own off-resonance reference. That division is what puts the values on the calibration’s scale, because the calibration itself is built from `mw_on / median(mw_on)` inside `_load_operator_full_scan`. Feed it unnormalised counts and it returns field values around $1 \times 10^5 \mu\text{T}$, which is not a subtle failure but is an easy one to mistake for a geometry problem.

The two-point path can skip the reference because it normalises against a stored constant from the calibration JSON instead (§3.2). The multipoint path has no equivalent.

10.2 The constraint that actually binds: buffer headroom

The board’s accumulated buffer holds `avg_maxlen` entries, one per readout per rep, and the board worker raises if unread entries ever reach that. The budget is shared across *every* readout in a shot, so what matters is

$$\text{shots_that_fit} = \left\lfloor \frac{\text{avg_maxlen}}{n_{\text{slots}} \times n_{\text{readouts per slot}}} \right\rfloor.$$

Configuration	reads/shot	relative headroom	
two-point, skip reference (<i>all data so far</i>)	2	1×	has never come close
two-point, with reference	4	1/2×	
16-point, skip reference	16	1/8×	
16-point, with reference (Step 4S)	32	1/16×	check before every long run

Sixteen times less room, and the stride the worker transfers in is 10% of it, so the margin between “fine” and “the run dies at 20 seconds” is much narrower than anything the two-point path has explored. Step 4S calls `stream_headroom()` before starting and warns if fewer than four strides fit.

The headroom number has never actually been read on this rig

`stream_headroom()` was broken for the whole 2026-08-17 session — it subscribed the Pyro proxy instead of `qd.soccfg` (§6.3.4) — so every run printed “could not read avg_maxlen” and proceeded blind. It is fixed, but that means the first Step 4S run is also the first time this rig’s `avg_maxlen` will be known. If it turns out to be tight, the levers in order of preference are: shorten τ (the buffer is counted in *entries*, so this does not help — it only makes the entries arrive faster, which is worse), reduce the parked-point count, or drop the reference and give up reconstruction.

10.3 Why the reconstruction is not inline

Step 4BN reconstructs the field vector inside its acquisition loop, once per batch. Step 4S does not, and the difference is not stylistic.

At 26 Hz there is 38 ms between samples and the fit costs a few milliseconds; there is room. At 260 Hz there is 3.84 ms between samples and there is not. More importantly, the consequence of falling behind is different in kind: the live loop simply runs slower, because nothing is happening on the board between its `acquire()` calls. A stream has the `tProc` running continuously, so a consumer that cannot keep up lets unread shots accumulate in the buffer of §10.2 until it overflows and *the run dies*. Slow post-processing inside a stream loop is not a performance question, it is a correctness one.

So Step 4S writes files and does nothing else, and Step 5S does the work afterwards, once, over the whole run — through `nv_toolkit.tui._compute_live_snapshot`, the same entry point the existing offline reconstruction cell uses. Same calibration, same inversion, so a streamed run and a live run are directly comparable.

`S5_DECIMATE` is there for the case where even the offline fit is too slow to be comfortable: it reconstructs every N th sample. Note that this is a genuine decimation, unlike the acquisition-side bin of §6.3.1 — it neither averages nor anti-aliases. Bin in Step 4S; decimate in Step 5S only to make a first look fast.

10.4 What the two cells produce

File	Contents
<code>multipoint_lockin_stream_<stamp>.csv</code>	One row per output sample: <code>packet</code> , <code>rep_index</code> , <code>reps_averaged</code> , <code>time_s</code> , <code>timestamp_epoch_s</code> , then <code>peak_NN</code> , <code>peak_NN_freq_mhz</code> and <code>peak_NN_ref</code> for each parked point.
<code>..._wide.csv</code>	<code>time_s</code> plus <code>peak_NN</code> normalised by its reference — the format the inversion consumes. Rebuilt by Step 5S rather than trusted, so a run copied in from another machine works.
<code>..._projection_rows.csv</code>	Per-block Δf and new peak frequency.
<code>..._vector_rows.csv</code>	ΔB_x , ΔB_y , ΔB_z per sample.

The naming matches the live-loop convention (`multipoint_lockin_live_*` and its `_wide` / `_vector_rows` / `_projection_rows` companions), so the existing post-processing and the session folders under `data/results/` stay consistent.

10.5 What Step 5S checks, and why each check exists

Check	What a failure means
<code>rep_index</code> gaps	A dropped rep. Every timestamp <i>after</i> the first gap is wrong, because the axis is built by counting reps rather than by measuring time. This is the one failure that silently corrupts the frequency scale of every spectrum downstream.
duplicate <code>rep_index</code>	Would be the burst replay appearing where it should be impossible. Has never been seen in a stream.
timestamp spacing	Should be identical to four decimal places. Spread here means the same thing as a rep gap.
PL droop	A continuous run has no relaxation gap between batches, so the common-mode level walks over the run — $\sim 25\%$ over 30 s on the two-point path. It is common to every parked point, which is what the reference normalisation removes.
rank-deficient samples	A geometry problem with the parked plan, not a streaming one: the same plan fails the same way in Step 4B.

Not yet run on hardware

Everything in this chapter is derived from the timing model and from `stream()`'s behaviour on the two-point path, which is well characterised. The multipoint streaming cells have been checked for syntax and for name resolution against the notebook's own globals, and nothing more. The first rig run should compare a Step 4S run against a Step 4BN run taken back to back: the reconstructed ΔB should agree within noise, and Step 4S should report ten times the rate with `reps_missing` = 0.

Chapter 11

Work log: what has been changed and pushed

Repository: github.com/chalach49266/nv-compact-magnetometer, branch `main`. All commits below are pushed. The two sessions are separated because the second one partly retracts the first.

title=Read Round 7 first — the acquisition path was reverted

On 2026-08-18 the acquisition *modules* were reset to revision `8fd303f`, the one the second machine takes good data with. Everything Rounds 3, 4 and 6 changed on that path is **no longer in the code**. The notebooks were kept. Rounds 1, 2 and 5 stand: measurement, analysis and figure saving, none of which touch acquisition.

11.1 Round 1 — 2026-08-12, against the 2026-08-06 session

Commit	Scope	What it did
5ea75ab	data:	2026-08-06 session snapshot (drone sweeps, burst runs, 17-point integration-time scan).
50fe514	docs:	The timing & noise analysis, plus <code>analyze_twopoint_session.py</code> and <code>profile_twopoint_acquire.py</code> .
942d9e8	perf:	Corrected the <code>config_all</code> keyword; τ 213 $\rightarrow$ 120 μ s; added ABBA ordering; append-only CSV writer.
6765284	fix:	<code>MultipointLockinODMR.stream()</code> — the continuous streaming path.
9eed2e1	feat:	<code>clean_burst_lockin.py</code> , to salvage burst runs already recorded.
16fd9ba	docs:	Changelog, rig checklist, cross-references.
f700223	fix:	Sensitivity omitted the $\sqrt{\text{reps}}$ factor, understating η by $\sim 10\times$.

Scorecard on round 1. Of the four changes intended to improve things:

Change	Outcome	Verdict
τ 213 $\rightarrow$ 120 μ s	PARTIAL	Took effect and is right for sensitivity, but bought no rate — the premise that τ set the rate was wrong.
<code>config_all</code> fix + <code>reset_tproc</code>	OPEN	Real bug, correctly fixed, but did not stop the replay .
<code>stream()</code>	DONE	Works: 1041.7 Hz , gapless. Its noise is a new open defect.
Append-only CSV writer	DONE	Removed the 25–220 ms periodic stalls.
ABBA ordering	OPEN	Implemented, never tested on hardware.

11.2 Round 2 — 2026-08-16/17, against the 2026-08-14 session

Commit	Scope	What it did
a98fd2f	data:	2026-08-14 session (101 MB): four averaged runs, one stream run, one burst run all at $\tau=120\mu\text{s}$; a 12-point integration-time sweep (1–213 μs , 5 repeats each); an 8-peak reference sweep; a multipoint run. Committed <i>before</i> any code change, so the analysis has a fixed baseline.
6eb7eb5	docs:	Retraction of the “rate is FPGA-bound” conclusion in the 2026-08-06 document, with the evidence, plus marking §4.3’s root cause disproved.
524dbb4	feat:	<code>twopoint_spectra.py</code> (365 lines) and <code>twopoint_postprocess.py</code> (790 lines); <code>clean_burst_lockin.py</code> rewritten to delegate. Found and fixed the per-channel despiking bug and the 12% burst-cadence error.
6893bcb	perf:	Corrected rate model (<code>HOST_CALL_S/HOST_FLOOR_S</code> replacing <code>HOST_OVERHEAD_S</code>); <code>reps_that_fit()</code> ; <code>measure_host_floor()</code> ; <code>multipoint_on_stale="drop"</code> ; <code>stream()</code> sequence aligned with <code>acquire()</code> and packet-size validation.
8fd303f	refactor:	Notebook split into Step 4A/5A, 4B/5B, 4C/5C, Step 6. New <code>twopoint_runner.py</code> (236 lines) for the shared plumbing. 1497 lines of notebook changed.
68e95a9	feat:	Three rig tests: <code>-compare-read-paths</code> , <code>-stream-scan</code> , <code>-floor</code> .
f1bd0eb	docs:	The 2026-08-14 methods document, <code>analyze_twopoint_0814.py</code> , five generated tables and three figures; pointer fixes in <code>LOCKIN_ACQUIRE.md</code> and <code>twopoint_lockin_notes.md</code> .

11.3 Round 3 — 2026-08-17, the board wedge

Triggered by a live failure at the rig rather than by analysis: Step 4C hung at its auto-zero `acquire()`, and `Ctrl-C` would not release it. Diagnosed to `poll_data()`’s `timeout=None` default and fixed the same day. Full account in §9.3; the naming collision in §9.4 was found while tracing which file each mode had written.

title=Read Round 4 before acting on anything in this section

The central change below — bounding every poll — was **withdrawn the same day**. It broke burst and streaming on first contact with the board. §9.3.3 is the account; §11.4 is what replaced it. The interrupt cleanup and the naming split survive unchanged.

Scope	What it did
<code>fix(qickdawg):</code>	<code>acquire()</code> 's drain loop extracted to <code>_drain_readout()</code> and bounded: an explicit <code>poll_data</code> timeout, a no-progress deadline that raises <code>TimeoutError</code> , and a <code>BaseException</code> handler that stops the <code>tProc</code> and the readout before re-raising. Adds <code>POLL_TOTALTIME_S</code> , <code>POLL_TIMEOUT_S</code> , <code>ACQUIRE_IDLE_S</code> , <code>_acquire_timeout_s()</code> and <code>_abort_readout()</code> .
<code>fix(twopoint):</code>	<code>stream()</code> passes the same explicit timeout — which is what finally makes its own <code>poll_timeout_s</code> reachable — and wraps its loop in <code>try/finally</code> , so an interrupted or abandoned generator leaves the board idle. Records <code>stream_qc()["aborted"]</code> .
<code>feat(twopoint):</code>	<code>MultipointLockinODMR.recover()</code> and <code>ensure_idle()</code> . The pre-flight is wired into Steps 4A, 4B and 4C.
<code>fix(notebook):</code>	Step 4A writes <code>twopoint_lockin_avg_*</code> and Step 4B <code>twopoint_lockin_burst_*</code> instead of both writing <code>_live_</code> ; <code>find_live_csvs()</code> accepts all three prefixes.

11.3.1 Verified offline

No board was available, so all three paths were exercised against fakes built from the real `qick` sources:

Check	What it drives	Result
Root cause	real <code>DataStreamer</code> , counter frozen; <code>poll_data()</code> vs <code>poll_data(timeout=1.0)</code>	hangs vs returns in 1.01 s
Healthy drain	<code>_drain_readout()</code> from the edited file, 50 shots in 10-shot packets	50/50, buffer bit-exact
Stalled <code>tProc</code>	same, board returns nothing	<code>TimeoutError</code> at the budget, board cleaned
Interrupt	<code>KeyboardInterrupt</code> mid-drain	re-raised, board cleaned
Stream abort	real <code>stream()</code> source, consumer breaks after 10 reps	<code>finally</code> ran, <code>aborted=True</code>
Stream completion	same, run to 20/20 reps	cleaned, <code>aborted=False</code>

11.4 Round 4 — 2026-08-17 evening, undoing Round 3's central change

Round 3 shipped at 14:23 and the rig session at 14:25–14:29 showed it had broken two of the three acquisition methods. The evidence is the notebook's own saved outputs:

Cell	Step	exec	reps in the failing call	Result
18	4A averaged	21	35	worked — 1067 samples, 104.5 Hz
22	4B burst	25	2000	<code>TimeoutError: read 0 of 2000 shots in 10 s</code>
26	4C stream	27	2000	hung in <code>Pyro4 socketutil.recvall</code> ; released by <code>Ctrl-C</code>
24	5B analysis	47	—	<code>TypeError: ... not 'NoneType' (collateral)</code>

Both real failures are the dedicated large-reps auto-zero `acquire()`, which 4A never makes. The comparison that settled it was the 2026-08-17 *morning* session, taken on a second computer at `8fd303f` — before Round 3 — which streamed 31 250 samples at 1041.7 Hz with zero missing reps, four times over. Same rig, same board, same day; the only difference was the code.

Scope	What it did
<code>revert(qickdawg):</code>	<code>POLL_TIMEOUT_S</code> back to <code>None</code> , restoring <code>qick</code> 's blocking <code>poll_data()</code> . The no-progress deadline is now consulted only when the poll is bounded, which it no longer is by default; set the class attribute to a float for a diagnostic run. §9.3.3 carries the reason at length so this is not re-attempted.
<code>fix(qickdawg):</code>	An empty return under a blocking poll now raises, naming how many shots were read. Previously that condition spun on the board with RPCs; the naive alternative would short-read zeros into the tail of <code>d_buf</code> .
<code>fix(twopoint):</code>	<code>ensure_idle()</code> reports and no longer calls <code>recover()</code> implicitly. <code>recover()</code> is staged: stop, drain the data queue, and only escalate to <code>start_worker()</code> if the board is still not idle (<code>force=True</code> skips ahead). Steps 4A/4B/4C raise and point at the new Step 0 instead of continuing into a hang.
<code>fix(twopoint):</code>	<code>stream_headroom()</code> read <code>qd.soc['readouts']</code> , and <code>qd.soc</code> is a Pyro Proxy — not subscriptable — so the check printed an error for the whole 2026-08-17 session instead of a number. It now uses <code>qd.soccfg</code> . This is the check the multipoint streaming path actually depends on.
<code>fix(notebook):</code>	<code>twopoint_postprocess.process(None)</code> raises a named error saying which run cell to re-run; Steps 5A/5B/5C fall back to the newest matching CSV on disk. <code>default_config.mw_gain</code> hoisted to a named <code>MW_GAIN</code> constant with its provenance — it went 10 500→15 000 inside a commit whose subject was “normalise JSON encoding”.
<code>feat(notebook):</code>	Steps 4S and 5S in <code>Modules/Lockin_module.ipynb</code> : streaming for the multipoint lock-in with its own post-processing and vector reconstruction (Ch. 10).
<code>test(scripts):</code>	<code>scripts/test_drain_readout.py</code> , which exercises the drain loop against <code>qick</code> 's <i>real</i> <code>poll_data</code> rather than a mock — the gap that let Round 3 ship. 17 checks, no board needed.
<code>docs(scripts):</code>	<code>scripts/analyze_twopoint_0817.py</code> regenerates every 2026-08-17 number in this document.

The process lesson, which is the expensive one

Round 3's commit message says “Verified offline (no board)”. It was — against a mock written alongside the fix, which modelled `poll_data` as “returns packets, or `[]` on timeout”. The real `poll_data` also discards a packet when `stop_flag` is set, and returns `[]` the moment its 100 ms `totaltime` expires. Neither behaviour was in the mock, and both matter.

`scripts/test_drain_readout.py` now extracts `poll_data`'s source from the installed `qick` and executes it, and fails loudly if that extraction stops matching. A mock of the component you are debugging tests your belief about it, not it.

title=Two more traps found while tracing this, both still live

- **qickdawg resolves to the wrong checkout unless the project root is first on `sys.path`.** It is pip-installed editable from `~/Documents/PhD/NV Compact Magnetometer/qick-dawg/src`, which carries none of this project's patches. The notebooks are safe — cell 0 inserts `PROJECT_ROOT` at position 0 — but a bare `python scripts/foo.py` is not. `test_drain_readout.py` asserts which copy it imported; other scripts do not.
- **The burst replay reproduces on a second machine.** 50.0–50.8% across six 2026-08-17 morning burst files, at 8fd303f. It is not an artefact of this laptop, and it is not

the wedge (§9.1).

11.5 The notebook layout

Modules/Two-point_Lockin_module.ipynb, 34 cells. Step 4 used to be a single 26 kB cell with a LIVE_BURST_MODE switch, and Step 5 a single generic analysis cell that treated all three methods alike — which is wrong in a different way for each of them.

Cells	Section	Contents
0–5	setup	project root, MW_GAIN and default_config (τ set here), DSA attenuation
6–7	Step 0	board recovery. Skip it on a normal day; run it after any interrupted acquisition (§9.3)
8–10	ODMR sweep	the reference sweep
11–14	Step 1a / 1b	pick the peak to track; place the parked pair and build TWOPOINT_CALIB (this is where the slopes and baselines come from)
15–18	Step 2 / 3	one batch at the two parked frequencies; apply the conversion
19–20	Step 4A	averaged acquisition. Measures the host floor, sizes reps to fill it, prints the alias warning. Writes twopoint_lockin_avg_*
21–22	Step 5A	process_averaged + ASD, ~15 lines
23–24	Step 4B	burst acquisition, on_stale="drop", cadence-based timestamps, reports BURST_STALE_FRACTION. Writes twopoint_lockin_burst_*
25–26	Step 5B	process_burst + per-segment ASD
27–28	Step 4C	streaming acquisition, stream_qc() assertions. Writes twopoint_lockin_stream_*
29–30	Step 5C	process_stream + ASD, optional mains notch
31–32	Step 6	post-process any saved run; mode detected from the columns; side-by-side comparison table
33	Caveats	what the method does and does not measure

Each Step 4x opens with a one-RPC ensure_idle() that raises and points at Step 0 rather than starting into a board that is not free. Each Step 5x falls back to the newest run of its own mode on disk when its partner's *_LAST_CSV is unset, so a failed acquisition produces a sentence rather than a TypeError from inside pathlib.

Each Step 4x is self-contained — it builds its own program, primes, auto-zeroes and writes its own CSV — with the parameters that change run to run inline, where they are edited. Each Step 5x reads the run its partner just wrote. The shared mechanical plumbing (append-only CSV writer, row building, the timing/freshness/calibration reports, the standard figure) lives in notebook_modules/twopoint_runner.py so that 4A and 4B do not carry 150 duplicated lines each.

11.6 Round 6 — 2026-08-18, the shot counter was never cleared

One missing line, and it was never in this file

acquire() and stream() arm the board-side readout without first zeroing the tProc shot counter. qick's own AcquireProgram.start_round() does it one line after reload_mem():

```
soc.reload_mem()
# make sure count variable is reset to 0
```

```
soc.clear_tproc_counter(addr=self.counter_addr)
```

The vendored `qickdawg/nvpulsing/nvaverageprogram.py` reproduces every other step of that sequence and drops this one.

11.6.1 What the counter does

It is `tProc` *data memory* at `counter_addr`, and nothing in the acquire path resets it: the program only ever increments it, `stop_tproc(lazy=True)` is a documented no-op on `tProc` v1 (§3.5, stage 1b), and `reload_mem()` reloads the program’s own data, not this address. So it carries the previous run’s final value into the next run.

The board-side worker (`qick/streamer.py`) then does:

```
shots = self.soc.get_tproc_counter(addr=counter_addr)
if shots >= min(last_shots+stride, total_shots):
    newshots = shots - last_shots # last_shots starts at 0
    data = self.soc.get_accumulated(ch=ch, address=addr,
    length=newshots*reads_per_count)
```

With a stale counter that test passes on the worker’s *first* look, before the new program has written anything, and it transfers whatever is sitting in the accumulated buffer. Two outcomes, depending only on how the stale value compares to the new run’s length:

Stale counter	What happens
<code>> total_count</code>	One oversized packet. This is the 2026-08-18 crash: a 100-rep priming <code>acquire()</code> inherited 2000 from the preceding auto-zero and raised <code>ValueError: could not broadcast input array from shape (4000,2) into shape (200,2)</code> — 2000×2 readouts into a 100×2 array.
<code>= total_count</code>	No error at all. The call returns in milliseconds with the <i>previous</i> run’s contents.
<code>< total_count</code>	A short head of stale data, then the run proceeds normally.

11.6.2 Which paths were affected

`reset_tproc=True` routes `config_all` through `stop_tproc(lazy=False)`, which on `tProc` v1 is `tproc.reset() + reload_program()` — and a reset plausibly zeroes data memory as a side effect. That is why the damage is uneven:

Call site	<code>reset_tproc</code>	Exposure
Step 4A, every sample	default <code>False</code>	every call after the first inherited the last one’s count
Step 4B, the burst <code>acquire</code>	<code>True</code>	probably protected by the reset
Step 4B/4C priming and auto-zero	default <code>False</code>	where the crash happened
Step 4C <code>stream()</code>	n/a	armed <code>start_readout</code> directly, never cleared
ODMR sweep, Steps 1–3	default <code>False</code>	one <code>acquire</code> per sweep point

Do not yet credit this with the burst replay

It is the obvious suspect for §9.1 — “the call returns in 13 ms instead of 425 ms with 95.7% of its rows bit-identical to the batch before” is exactly the middle row of the table above. But burst mode is the one path that passes `reset_tproc=True`, so if the v1 reset does zero the

counter, burst was already protected and its replay has another cause. The rig settles it in one line, run straight after a burst batch:

```
qd.soc.get_tproc_counter(addr=1)
```

Zero means the reset clears it and the replay is something else. Non-zero means this was the cause all along, and §9.1 can be closed. Step 0's `recover()` now prints the value for exactly this reason.

11.6.3 The fix

Where	Change
<code>acquire()</code>	<code>clear_tproc_counter(addr=self.counter_addr)</code> between <code>reload_mem()</code> and <code>start_readout()</code> , matching qick. One RPC, ~1 ms, on a call that already costs 10–11 ms.
<code>stream()</code>	the same, in the same position.
<code>_drain_readout()</code>	an oversized packet now raises a named <code>RuntimeError</code> that says the counter did not start at zero, instead of letting numpy raise <code>could not broadcast</code> , which names neither cause nor fix.
<code>_abort_readout()</code>	clears it on the way out, so an interrupted cell does not hand a poisoned counter to whatever runs next.
<code>recover()</code>	reads the counter, prints it, clears it.

No notebook cell changed. Every acquisition in both notebooks goes through `acquire()` or `stream()`; the vendored qickdawg has exactly one `start_readout` call site and `stream()` is the other, and every LockinODMR-style subclass delegates to `NVAveragerProgram.acquire` via `super()`.

11.6.4 Verified

`scripts/test_shot_counter.py`, 21 checks. It drives qick's *real* `DataStream._run_readout` against a fake board whose counter persists between runs and whose program ramps the counter rather than completing instantly — the arithmetic under test lives in that worker, so a mock of the worker would prove nothing. It reproduces the exact reported shapes, (4000,2) into (200,2), and shows the clear removes them; shows the equal-counter case completing silently with data read before the program produced it; and confirms the call is present in the shipped `acquire()`, `stream()`, `_abort_readout()` and `recover()`, in the right order relative to `reload_mem` and `start_readout`.

11.7 Round 7 — 2026-08-18, revert the acquisition path to 8fd303f

Three rounds of changes to the acquisition path each fixed something real and each left the rig worse than it found it. On 2026-08-18 the operator reported that streaming still did not run and that averaged mode had fallen to 48 Hz, against the 87–105 Hz it had held all week, and asked for the path to be made identical to the revision on the second machine.

File	State	Why
quickdawg/nvpulsing/nvaverageprogram.py	byte-identical to 8fd303f	the acquire loop and the poll
notebook_modules/multipoint_lockin_program.py	byte-identical to 8fd303f	stream(), the timing model
all 35 quickdawg/*.py	identical	nothing else on the path drifted
Modules/*.ipynb	kept	the revert is the acquisition <i>ma</i> the notebooks keep Step 0, the <i>_</i> /_burst/_stream_names, mw_g 15000, the newest-run fallbacks, 4S/5S and figure saving
notebook_modules/twopoint_postprocess.py	kept	post-processing only; never on t quire path

The only notebook edit the revert forced: Step 0 and the three pre-flight guards called `recover()` and `ensure_idle()`, which live in the acquisition module and went away with it. Both are now written inline against the raw QickSoc proxy — `stop_tproc`, `streamer.stop_readout`, `poll_data(totaltime=-1)`, `streamer.readout_running` — all plain remote methods qick itself exposes. So the notebook keeps the escape hatch and is independent of which module revision is loaded. The guard also downgraded from fatal to a warning on probe failure: a board-state probe must never be able to stop a run.

11.7.1 What went away with it

Gone from the modules	Consequence
<code>POLL_TIMEOUT_S</code> , <code>_drain_readout</code>	the poll is qick's plain <code>poll_data()</code> again — blocking, which is the behaviour that worked.
<code>_abort_readout()</code>	an interrupted cell no longer stops the tProc on the way out. After a Ctrl-C mid-readout, run Step 0; if it still reports busy, restart the board's Pyro server. A kernel restart will not clear it.
<code>ensure_idle()</code> , <code>recover()</code>	reimplemented in the notebook instead, as above.
<code>clear_tproc_counter()</code>	the stale-counter behaviour of §11.6 is live again — see the box below.

The hardware operating point is *unchanged*: `mw_gain = 15000`, $\tau = 120\ \mu\text{s}$, `AVG_REPS_PER_BATCH = None`, `BURST_REPS = 500`, `STREAM_TARGET_HZ = 1000`. Those live in the notebook, not in the reverted modules.

The stale shot counter is real and is now unfixed

§11.6 is not retracted — the mechanism was reproduced against qick's own `DataStream._run_readout`, and 8fd303f is missing the same call qick makes. It is an open defect of the reverted revision, and it is the leading candidate for the burst replay of §9.1. `scripts/test_shot_counter.py` is kept for that reason, scoped down to the mechanism: 13 checks, all passing, none of them asserting anything about the shipped code. If the fix is revisited, that test is the check it needs to pass.

11.7.2 The lesson this round is actually about

Every one of Rounds 3, 4 and 6 was reasoned from the source, verified offline against a test, and shipped — and each one changed the rate or broke a mode on first contact with the rig. Round 3 was verified against a mock and broke burst and streaming. Round 6 was verified against qick's real worker thread and, whatever else it did, coincided with averaged mode falling to 48 Hz.

Offline verification says the code does what you think. It says nothing about whether what you think is worth doing. **On this rig, the acquisition path changes one thing at a time, with a before-and-after run on the bench, or it does not change.**

Session	Averaged rate	Path revision
2026-08-14	86.8–88.8 Hz	8fd303f lineage
2026-08-17 morning (<i>other machine</i>)	—	8fd303f
2026-08-17 afternoon	104.5 Hz	9832e07 (burst + stream broken)
2026-08-18	48 Hz	Round 6
2026-08-18, after this revert	<i>to be measured</i>	8fd303f

The last row is the first thing to fill in at the rig.

11.8 Round 5 — 2026-08-18, every figure goes to disk

Neither notebook wrote a single PNG. Every plot lived only in cell output, which "Restart & Clear Outputs" discards and which cannot be opened next to the CSV it describes. The `*_result.png/*_fft.png` pairs already sitting in `data/results/` were produced by a separate tool on the acquisition machine, not by these notebooks — nothing in this repository wrote them.

11.8.1 `notebook_modules/figure_autosave.py`

Saving is automatic rather than a `savefig` line per plot, because there are twenty-three plotting sites across the two notebooks and adding a line to each guarantees the twenty-fourth is forgotten.

The module hooks the two moments at which the inline backend destroys a figure:

Hook	What it catches
<code>post_execute</code> callback	Figures a cell drew but never explicitly showed. It must run <i>before</i> matplotlib-inline's <code>flush_figures</code> , which displays and then closes everything; IPython fires <code>post_execute</code> in registration order, so <code>enable()</code> registers its own callback and then re-registers <code>flush_figures</code> to move it to the back of the queue.
wrapper around <code>pyplot.show</code>	Figures closed mid-cell. <code>show()</code> on this backend closes what it displays, and <code>twopoint_runner.plot_run</code> , <code>TwoPointResult.figure(show=True)</code> and the live loops in <code>Lockin_module.ipynb</code> all call it.

Both funnel into `capture()`, and the “already written” marker lives on the figure object, so the two hooks cannot duplicate each other's work.

11.8.2 Naming

`set_run(csv, tag=...)` points the module at a run: the PNGs take the CSV's stem and land in the CSV's own folder. The numbering restarts on every `set_run`, so **re-running a cell overwrites that cell's own PNGs** instead of accumulating copies.

Cell	Writes
Step 4A/4B/4C <code>set_run(AVG_CSV, tag="run")</code>	<code>twopoint_lockin_avg_<ts>_run_result.png</code>
Step 5A/5B/5C <code>set_run(A5_CSV)</code>	<code>twopoint_lockin_avg_<ts>_result.png</code>
	<code>twopoint_lockin_avg_<ts>_fft.png</code>

A figure is classified `_fft` only when *all* of its axes are log-log spectra. That matters here: the three-panel figure from `TwoPointResult.figure()` carries an ASD as its last panel, and must not claim the `_fft` name from the standalone spectrum figure of the same run. A cell can override the classification entirely with `fig.set_label("droop")`, giving `<stem>_droop.png`.

11.8.3 `TwoPointResult.spectrum_figure()`

Added so that every two-point run gets a real `_fft.png` rather than only the ASD panel buried at a third height inside `figure()`. Full width, with the white floor, η from σ , and every line above $2\times$ the floor annotated with its excess. Steps 5A/5B/5C call it after `figure()`.

11.8.4 Verified

- `scripts/test_figure_autosave.py` — 18 checks, driven through a real `InteractiveShell` with `%matplotlib inline` rather than a mock, because hook ordering against `flush_figures` is the whole mechanism. Confirms the ordering explicitly, and covers the unshown-figure path, the `plt.show()` path, the mixed-figure classification, tags, overwrite-on-rerun, empty figures, `disable()/re-enable()` and the label override.
- `scripts/test_notebook_figure_saving.py` — runs Step 5B and Step 5C *out of the .ipynb files* against scratch copies of the 2026-08-17 runs, and checks both PNGs appear beside the CSV and that a second run overwrites them. 6 checks.
- A static pass over both notebooks asserts every cell that draws calls `set_run` *before* its first plot. That caught one cell where the call had landed after the plots and would have filed them under the previous run's name.

The live loops write one figure, not a movie

`Lockin_module.ipynb` Steps 4 and 4B redraw a single figure in place while the run proceeds. What reaches disk is that figure's *final* state, which is the whole run — not a frame per update. Step 4BN and Step 4S are headless and draw once at the end, so they are unaffected.

There is also no `_fft.png` for the multipoint live cells: their time base is the host loop period, not a cadence, so an ASD computed on it would put a frequency scale on jitter. The streamed Step 5S has a real cadence and does produce one.

11.9 The code, file by file

File	State	Role
<i>Acquisition</i>		
<code>notebook_modules/multipoint_lockin_program.py</code>	modified	The FPGA program, all three read paths, the timing model, the freshness guard.
<code>quickdawg/nvpulsing/nvaverageprogram.py</code>	modified	Vendored; the <code>config_all</code> fix and <code>reset_tproc</code> plumbing.
<code>Modules/ Twopoint_Lockin_module.ipynb</code>	restructured	32 cells: setup, ODMR, Steps 1–3, then 4A/5A, 4B/5B, 4C/5C, 6.
<i>Analysis (new)</i>		
<code>notebook_modules/ twopoint_spectra.py</code>	new	Every FFT. One-sided PSD, $\eta = \sigma\sqrt{2\Delta t}$, mains-line finder, alias report, comb notch. Verified against Parseval.
<code>notebook_modules/ twopoint_postprocess.py</code>	new	One pipeline per mode + mode detection + exact readout-quantum recovery + CLI.

File	State	Role
notebook_modules/ twopoint_-runner.py	new	Shared acquisition plumbing: append-only CSV, row building, the three reports, the standard figure.
notebook_modules/burst_-qc.py	reused	Batch table, stale masks, retiming, segment PSD, block averaging.
notebook_modules/ spike_-rejection.py	reused	HampelDespiker.
<i>Scripts</i>		
scripts/ analyze_twopoint_-0814.py	new	Regenerates every number and figure in this document.
scripts/ profile_twopoint_-acquire.py	modified	Static signature check + per-RPC timing + the three rig tests.
scripts/ clean_burst_-lockin.py	rewritten	Burst CLI, now delegating to twopoint_-postprocess.
scripts/ analyze_twopoint_-session.py	unchanged	The 2026-08-06 analysis; still regenerates that document.
<i>Documents</i>		
docs/twopoint_master_-reference/	new	This document.
docs/ 2026-08-14_twopoint_-methods/	new	Markdown source + generated tables/figures.
docs/ 2026-08-06_twopoint_-timing/	annotated	The earlier analysis, with corrections marked in place.
docs/LOCKIN_ACQUIRE.md	pointer fixed	Repeated the retracted 1.6 ms figure.
docs/ twopoint_lockin_-notes.md	pointer fixed	Same.

11.10 Wiki

The synthesis is filed in the Obsidian vault at NV Project/.

- **New page:** wiki/sources/twopoint_three_modes_2026_08_14.md
- **Correction banner:** wiki/sources/twopoint_timing_and_burst_defect.md
- **Updated:** wiki/experiments/lockin_tracking.md (new “August 14–17” section, and the August 6–12 one partly retracted); wiki/concepts/sensitivity.md (three new rows, plus a warning that the two-point noise does not average down); wiki/questions/open_questions.md (three questions answered, four new ones with pass criteria); wiki/overview.md (status table and the “Where We Stand” list); index.md (routing).
- **Log entry:** log.md, dated 2026-08-17.

Not committed to git: the vault sits inside the home-directory repository, which contains unrelated and sensitive untracked files. The files are on disk and sync through OneDrive.

Chapter 12

What still needs to be done

12.1 Rig checks, in priority order

Everything here needs the RFSoc at 192.168.0.103, reachable only from the Windows rig machine. Each has a pass criterion.

#	Check	Command	Pass criterion
S0	Does the wedge fix close the burst replay? Now the cheapest decisive test, because §9.3 gives §9.1 its first mechanism since <code>reset_tproc</code> was eliminated. <i>Do this first.</i> ~1 min.	Run Step 4B for 30 s and read <code>BURST_STALE_FRACTION</code> .	0% $\Rightarrow$ the replay was the wedge and both defects close. Still ~50% $\Rightarrow$ they are independent; S1 next.
S1	Burst vs stream, interleaved. Decides whether the ~6 $\times$ streaming excess is the read path or the room. <i>Do this first.</i> ~1 min.	<code>profile_twopoint_-acquire.py -compare-read-paths -tau 120 -reps 500</code>	ASD ratio within 1.25$\times$ $\Rightarrow$ environment. Well above $\Rightarrow$ software defect.
S2	Does the streaming excess grow with run length (buffer fill)?	<code>profile_twopoint_-acquire.py -stream-scan</code>	Floor at 125 k reps within 1.3$\times$ of the floor at 2 k reps.
S3	Confirm the per-call floor and the flat rate.	<code>profile_twopoint_-acquire.py -floor -tau 120 -reps 1 4 10 23 42 100</code>	Fastest call 9–11.5 ms with a reps=1 probe; period flat to 15% across reps whose FPGA work fits under it.
S4	Fill in §3.5 — the per-RPC split, the one structural gap in this document.	<code>profile_twopoint_-acquire.py -ip 192.168.0.103</code>	A time for each of stages 1–7. No pass/fail; this is a measurement that has never been made.
S5	Free averaging. 30 s Step 4A with <code>AVG_REPS_PER_BATCH = None</code> .	Notebook Step 4A	Still ~88 Hz and $\sigma(\Delta f) \leq$ 54.6 kHz . Given §8.4, expect less than the naive 2 $\times$.
S6	The per-acquire noise term. Interleave averaged runs at reps = 10 / 23 / 42 within one sitting.	Notebook Step 4A, three configs, alternating	Does σ fall as $1/\sqrt{\text{reps}}$? If not, the term is per-call — then test whether it tracks the idle gap by varying <code>SAVE_EVERY_SEC</code> or inserting a deliberate sleep.
S7	ABBA A/B at fixed τ and equal averaging ("forward" 46 reps vs "abba" 23 reps), interleaved.	Notebook Step 4A, two configs	Report the change in $\sigma(\Delta f)$. Any reduction is free.
S8	Honest burst rate with the drop policy. 10 s Step 4B.	Notebook Step 4B, <code>BURST_ON_STALE="drop"</code>	Step 5B reports n_stale_batches = 0 on the file Step 4B just wrote, and <code>BURST_STALE_FRACTION</code> is printed.

#	Check	Command	Pass criterion
S9	Instrument the replay. Print the shot counter and queue depth before/after each <code>start_readout</code> during a burst run.	needs a small patch to <code>acquire()</code>	Discriminates the three candidate mechanisms in §9.1.

12.2 Software backlog

Item	Priority	Why
Periodic re-zero during a streaming run	high	The 25% PL droop (§6.6) walks the operating point off the calibrated flank over 30 s. Currently only detrended in post.
Root-cause the burst replay	high	Would recover $\sim 15\%$ of burst rate and remove the need for the drop policy.
Real-time mains notch in the acquisition path	medium	The current notch is offline, non-causal, and assumes stationarity. At ≥ 1 kHz a causal notch is feasible and the line is $10\times$ the floor.
Fold the multipoint notebook onto the same modules	medium	<code>twopoint_spectra.py</code> and <code>twopoint_postprocess.py</code> are mode-generic; the multipoint path still has its own analysis code.
Calibrated field reference	medium	Every η here is a readout-chain figure (§2.4). A known applied field would turn them into system claims.
<code>remove_offset</code> is documented but not implemented	low	<code>NVAveragerProgram.acquire()</code> accepts the argument and never uses it. Harmless today (both paths are consistent) but a latent trap.
Allan deviation on the long runs	low	Needed for the drift/averaging-time story; the June time-integration data exists but has not been analysed drift-aware.

12.3 Recommended operating point today

Until S1 resolves the streaming question

For best sensitivity — Method B, burst.

$\tau = 120\ \mu\text{s}$, `BURST_REPS` = 500, `BURST_ON_STALE` = "drop". $38\ \text{nT}/\sqrt{\text{Hz}}$ at an effective 3.1 kHz. Accept the ~ 15 ms gap every 135 ms.

For a continuous trace — Method C, streaming.

$\tau = 120\ \mu\text{s}$, `STREAM_TARGET_HZ` = 1000. 1041.7 Hz, gapless, exact timestamps, but $\sim 6\times$ the noise until §9.2 is closed.

For slow work and sanity checks — Method A, averaged.

$\tau = 120\ \mu\text{s}$, `AVG_REPS_PER_BATCH` = None so the call floor is filled. ~ 88 Hz, no open defects — but remember 60 Hz is aliased into the band here.

Chapter 13

Corrections register

Every claim from an earlier document that this one contradicts, in one place.

Source	Earlier claim	Current position	Evidence
2026-08-06 doc §1, §2	“The rate is FPGA-bound, not host-bound”; host round-trip ~ 1.6 ms	Retracted. Host-bound, ~ 10 – 11.4 ms per call; the FPGA runs underneath it	§7.1
2026-08-06 doc §2.6	τ is “the main lever”; reps trades against rate one for one	Both backwards. Neither affects the rate below the floor; reps is a free <i>sensitivity</i> lever	§7.4
2026-08-06 doc §2.4	“Host round-trip 1.60 ms (14.1%)” as a budget row	That was a <i>residual</i> under an assumed serial model, not a measurement	§7.1
2026-08-06 doc §4.3	config_all / never-stopped tProc is the root cause of the burst replay	Disproved. The fix was applied and 49.9% still replayed	§9.1
2026-08-06 doc §3.2	“No 60 Hz line, no discrete tones”	The line is there and is $10\times$ the floor; at 87.5 Hz it is <i>aliased</i> to ~ 28 Hz and the burst runs that could have resolved it were half replays	§8.5
2026-08-06 doc, 1st draft	$\eta \approx 19$ nT/ $\sqrt{\text{Hz}}$ per rep	185 nT/ $\sqrt{\text{Hz}}$. The draft paired a reps-averaged σ with a single-rep time, omitting $\sqrt{\text{reps}} \approx 10$	§2.3
2026-08-06 doc, η convention	$\eta = \sigma_B \sqrt{\Delta t}$	$\eta = \sigma_B \sqrt{2\Delta t}$, matching what the notebook prints; a $\sqrt{2}$ difference	§2.3
2026-08-14 doc, 1st draft	Streaming excess is “ $4.2\times$ ”	5.3 – $6.3\times$. The 4.2 used the <i>raw</i> burst floor; the cleaned floor is 38.6 rather than 55 nT/ $\sqrt{\text{Hz}}$	§9.2
Notebook Step 4 comment, pre-2026-08-12	“ ~ 10 ms of fixed host overhead, FPGA work that fits inside it is FREE”	Substantially correct , and wrongly “corrected” on 2026-08-12. Restored	§7.2
time_per_rep() , pre-2026-08-12	$n_{\text{slots}} \times (\tau + 2 t_{\text{relax}})$	$n_{\text{slots}} \times \tau$; the relax windows are absorbed	§3.2
Burst cadence, pre-2026-08-16	From acq_seconds/n_samples : 269 μs	240.0 μs exactly, from the read-out quantum. The old value included the host share	§3.4
Despiking, pre-2026-08-16	Filter the two parked channels independently	Filter the <i>differential</i> . Per-channel raises σ	§9.5.1

The methodological lesson

Four of the twelve rows above come from the same mistake: **fitting a model with n free parameters to data taken at one operating point**. The 2026-08-06 session ran at a single (τ, reps) , and every timing conclusion drawn from it was under-determined — the serial-versus-overlapped question simply cannot be answered from one point, however carefully the residuals are computed.

The 2026-08-14 session settled it in an afternoon purely by varying two knobs. Any future timing claim should span at least three operating points before it is written down.

Chapter 14

Reproducing everything

14.1 Offline, no hardware

```
# every number and figure in this document
python scripts/analyze_twopoint_0814.py
# post-process any run; the mode is detected from the columns
python -m twopoint_postprocess <csv> [--outdir DIR] [--notch-mains]
# burst-specific CLI with the original flags
python scripts/clean_burst_lockin.py <live_csv> --outdir analysis/
# the 2026-08-06 analysis, unchanged
python scripts/analyze_twopoint_session.py
# static check of the config_all signature
python scripts/profile_twopoint_acquire.py --check-only
# rebuild this document
cd docs/twopoint_master_reference && latexmk -pdf TWOPOINT_MASTER_REFERENCE.tex
```

14.2 On the rig

```
python scripts/profile_twopoint_acquire.py --floor --tau 120
python scripts/profile_twopoint_acquire.py --compare-read-paths --tau 120 --reps
500
python scripts/profile_twopoint_acquire.py --stream-scan
python scripts/profile_twopoint_acquire.py --ip 192.168.0.103
# the full per-RPC profile (fills section 3.5)
```

14.3 Data used by this document

Path	Contents
data/results/ 081426 (Sensitivity increase update)/	2026-08-14: 4 averaged + 1 stream + 1 burst two-point run, all at $\tau=120\text{ }\mu\text{s}$; 12-point integration sweep; 8-peak reference sweep; multipoint run. 101 MB.
data/twopoint_lockin/08062026/	2026-08-06: drone sweeps, 4 burst runs, 12 averaged runs, 17-point integration sweep (92 ODMR sweeps).
docs/2026-08-14_twopoint_methods/tables/	Generated: <code>timing.csv</code> , <code>mode_budgets.csv</code> , <code>modes.csv</code> , <code>read_path.csv</code> , <code>per_rep_noise.csv</code> , <code>spectra.csv</code> .