

红外测温仪校准证书

设备编号：多8

校准时间：2025-04-08 10:25:22

校准温度(℃)	标准温度(℃)	测量值1(℃)	测量值2(℃)	测量值3(℃)
-25.00	-24.91	-24.98	-24.98	-24.99
-25.00	-21.67	-21.68	-21.78	-21.78
-20.00	-19.91	-20.14	-20.05	-20.13
-5.00	-4.74	-4.68	-4.79	-4.71
-5.00	-2.66	-2.65	-2.68	-2.67
0.00	0.10	0.48	0.41	0.40
5.00	5.37	5.61	5.54	5.55
5.00	6.78	6.43	6.47	6.47
25.00	25.31	24.59	24.42	24.51
25.00	25.78	25.01	25.08	25.07
30.00	30.42	29.31	29.12	29.24
40.00	40.21	40.30	40.29	40.29
40.00	39.37	39.40	39.45	39.42
60.00	60.30	60.05	60.03	60.10