

# 红外测温仪校准证书

设备编号: N26

校准时间: 2025-04-07 15:00:52

校准温度点	标准温度	测量温度
60.00	58.04	58.30
60.00	60.08	60.10
50.00	50.40	50.53
40.00	39.26	39.18
40.00	40.11	40.16
30.00	30.30	30.38
25.00	25.69	25.70
25.00	25.21	25.27
10.00	11.15	11.63
10.00	10.33	10.36
-5.00	-3.12	-2.94
-5.00	-4.76	-4.87
-20.00	-17.52	-17.62
-20.00	-19.99	-20.62
-30.00	-25.43	-27.14
-30.00	-29.92	-29.70