

红外测温仪校准证书

设备编号：多15

校准时间：2025-04-08 10:24:39

校准温度(℃)	标准温度(℃)	测量值1(℃)	测量值2(℃)	测量值3(℃)
70.00	70.50	69.95	69.94	70.03
60.00	58.10	58.49	58.47	58.33
50.00	50.40	50.97	50.99	50.90
50.00	48.53	49.07	49.26	49.02
40.00	39.45	40.37	40.60	40.27
40.00	40.36	39.54	41.63	41.41
30.00	30.34	10.62	10.62	10.62
30.00	30.67	10.62	10.62	10.62
30.00	29.92	10.62	10.62	10.62
30.00	30.31	10.62	10.62	10.62
25.00	25.22	10.62	10.62	10.62
25.00	25.19	10.62	10.62	10.62
20.00	20.44	10.62	10.62	10.62
20.00	20.44	10.62	10.62	10.62
15.00	16.03	10.62	10.62	10.62
15.00	15.76	10.62	10.62	10.62
10.00	11.22	10.62	10.62	10.62
5.00	6.40	10.62	10.62	10.62
5.00	5.63	10.62	10.62	10.62
0.00	1.65	1.64	1.74	1.88
0.00	0.15	0.02	-0.11	-0.08
0.00	0.56	0.51	0.39	0.50

校准温度(℃)	标准温度(℃)	测量值1(℃)	测量值2(℃)	测量值3(℃)
-5.00	-3.15	-3.13	-3.00	-3.01
-5.00	-4.52	-4.58	-4.70	-4.57
-10.00	-9.57	-9.60	-9.69	-9.61
-15.00	-12.71	-12.81	-12.81	-12.88
-20.00	-19.98	-20.12	-20.15	-20.27
-25.00	-22.35	-21.83	-22.31	-22.30
-25.00	-24.84	-24.80	-24.82	-24.92
-30.00	-27.16	-25.10	-26.55	-26.24
-30.00	-29.89	-29.97	-29.92	-30.10