This copy will not be updated. Verify the status of this version in DiaDoc or contact your ROCHE representative prior to use.
Diese Kopie wird nicht aktualisiert. Überprüfen Sie vor Gebrauch den Status dieser Version im DiaDoc oder kontaktieren Sie Ihren ROCHE Ansprechpartner.
Esta copia no va a ser actualizada. Verifique el “status” de esta versión en DiaDoc o contacte a su representante de ROCHE antes de utilizarla.

Document Information:

Document Author:	Smorenburg, Penny {DQME~CAPETOWN}
Business Area / Unit:	Logistics & All Sites
Confidentiality:	Confidential
Document Class:	Quality System Record
Document Type:	Quality System Record with Review
Document Creator:	Bhagat, Hitesh {DQOF~BRANCHBURG}
Document Lifecycle Status:	Effective
Valid From:	06-Mar-2026 00:21:00 (UTC)
Valid To:	
Document Title:	MQMS-PM-GSP-05-CC-02
Document Number:	0000000000001200000521954
Document Version:	49
Template:	No
Global Group:	Modular Quality Management Sys
Global SubGroup:	Global Standard Process
Local Group:	Changing Content
Language:	English
Site:	Global
Document Description:	PRI Screening Criteria for Erroneous Results
Description of change:	1. Added CT-1174 (Gram Stain Core Kit), and CT-1175 (Congo Red Stain Core Kit) under Ventana Special Stains 2. Added CT-1176 (SSR Solution) under Ventana Ancillary Reagents 3. Added CT-1177 (NFL [Elecsys NfL/ Neurofilament light chain]) under Specific Proteins (REQ01389R)
Justification for change:	1. Design Change to a existing product with a new name 2. Gram Staining Kit split into the Gram Stain Core Kit and the SSR Solution 3. New product launch (REQ01389R)

Electronic Signatures:

Signed By:	smorenbp (Penny Smorenburg {DQME})
Role:	Approver
Signature Differentiation:	Post Market Quality
Signed Date:	03-Mar-2026 16:47:33 (UTC)
Signed By:	neblg (Gabriele Nebl {DQOF})
Role:	Approver
Signed Date:	04-Mar-2026 16:02:43 (UTC)

Purpose: The purpose of this document is to provide criteria for identifying erroneous results that must be classified as PRI. This document is to be used in conjunction with the information provided in MQMS-PM-GSP-05 and MQMS-PM-GSP-05-CC-01.

PRI Screening Criteria for Erroneous Results
Changes since last version

Version	Summary of Changes
Anex 07a v 32 Transitioned to MQMS-PM-GSP-05-CC-02 Version 2	Added criteria for AFP-L3, Lp-PLA2, HCV Duo, Urine Testing of SEC, NEC, HYA, PAT, Sperm,Crystals Updated Anti-Tg, Cortisol, HCG, Kappa and Lamda FLC, Procalcitonin, S100 Updated PT critria to include units of measure "Seconds" to exclusion list For Thyroid parameters with exclusions, added clarifying informatino regarding reference values and clinical interpretation Updated Section Headers for Influenza A/B, Influenza A/B and RSV and Clostridium difficile to include cobas Liat Harmonized naming of Beta 2 Microglobuin tests Added vTAC to appropriate blood gas testing paramters Added Diagenode AD_hMPV_EV_RV and Diagenode Parainfluenza assays
Annex 07a v33 transitioned to MQMS-PM-GSP-05-CC-02 Version 3	Added criteria for Fentanyl Moved criteria for Lp-PLA2 from Cardiac Markers section to Enzyme section Corrected spelling for beta-2 Microglobuline CT-208 to beta-2-Microglobulin Added Criteria for Factors IX, XI, XII and XIII for Coagulation Testing by Professionals
MQMS-PM-GSP-05-CC-02 Version 3 to Version 4 (Annex 07a retired 1 Dec 2021)	Urine Strip and Sediment Testing Section Name updated to Urine Strip Testing and added a note for sediment evaluation for applicable tests Added new criteria for CoVid 19 At Home Test and SARS CoV-2 Rapid Antigen Test Pro Updated ssDW screening criteria name to Trisomy Screening Added Preeclampsia screening criteria Clarified the following statement for applicable thyroid tests: A PRI is not required when differing results lead to the same clinical interpretation. This applies when a Roche result is compared to Roche result or Roche compared to another manufacturer. For BS assays, removed the references to the different platforms (i.e., cobas s 201, cobas 6800/8800) Added criteria for ADCY9 for Genomics & Oncology Removed criteria for Navify Mutation Profiler and removed Navify Therapy Matcher from list of parameters as NAVIFY Therapy Matcher has been absorbed into NAVIFY Mutation Profiler as one product, referred to as NAVIFY® Mutation Profiler version 2.1. Updated Table F - BKV Monitoring (plasma) and Table G - BKV Monitoring (urine) Removed Table H as invalid/delayed results for SARS-CoV-2 tests is no longer PRI criteria Separated criteria for HIV-1/2 - Qualitative Tests- Diagnosis/Infant testing into separate entries. Removed Invalid/Delayed criteria as PRI for SARs-CoV-2 tests (cobas 6800/8800 and Liat) Added criteria for cobas SARS-CoV-2 nucleic acid test for use on the cobas Liat System Added Elastic Stain Core Kit Special Stains Van Gieson Counter Stain under Ventana Special Stains
Version 4 to Version 5	Minor Formatting Changes
Version 5 to Version 6	Clarified MetHb Criteria is for absolute values Clarified Naming of CoV-2 Antigen Testing Products Added criteria for CoV-2 FluA/B Rapid Antigen Test Clarified %Q criteria for PT testing Removed Troponin I for other Point of Care tests Clarified names of Troponin T testing for Other Point of Care tests Clarified SARS-CoV-2 - c6800/8800 System criteria is for qualitative and quantitative test
Version 6 to Version 7	Combined and updated Covid/Pilot 19 At Home Test (US Lay Use Test from SD Biosensors [OTC} and SARSC C0V-2 Rapid Antigen Test US Professional Use Test from SD Biosensor [Pro]. Clarified comparisons are for SD Biosensor to SD Biosensor within 24 hour time frame for US products only Added Criteria for Lipoprotein A - LPACX (Companion Diagnostic Only) Added Criteria for SARS IGRA Added Criteria for Inteferon Gamma Beta 2 Microglobulin - Corrected spelling and moved criteria to Specific Proteins to remove redundant criteria in different sections Added 5800 analyzer to appropriate parameters
Version 7 to Version 8	Added PRI Criteria for LightMix® Modular Monkeypox Virus Added PRI Criteria for GenMark products Corrected typo in spelling of Interferon Gamma and moved to new Section - Cytokines (Lymphokins) / Immunomodulators
Version 8 to Version 9	Added PRI Criteria for PRAME (EPR20330) Rbt Mono PAB Updated Name for Troponin T hs to Troponin T hs / Gen5 for clarification purposes only
Version 9 to Version 10	Added PRI criteria for VENTANA FOLR1(1-2.1)- RxDx Assay and cobas MPXV test (monkeypox) Removed criteria for LightMix Modular Monkeypox Virus (RUO) Simplified and harmonized criteria for PRI criteria for SD Biosensor Tests
Version 10 to Version 11	Replaced the term 'guidance' with the term 'document' in the 'Purpose' section Replaced 'DONT TOUCH THIS' seen under Unique Identifier with 'Please do not modify' Added note to explain how updated information will be captured GenMark PRI criteria were modified to incorporate medical expertise from CDMA and to align with other Roche molecular products, including (1) defining criteria based on hazardous situations for specific targets or groups of targets and (2) approaching equivalent targets similarly. In addition, PRI criteria now also include consideration of those targets associated with public health-reportable diseases. Updated PRI criteria for SD Biosensor test Removed Navify Mutation Profiler as this is a standalone system. Test/parameters for this standalone system are captured under oncology section
Version 11 to Version 12	Updated CT-338 (SD Biosensor) to clarify when a case is not considered a PRI. Updated GenMark criteria (CT-346 and CT-379) to update virus 3 to 4 and capture 2 targets.
Version 12 to Version 13	Added PRI Criteria for CD8 (SP239) Rabbit Monoclonal PAB test Updated a GenMark Unique Identifier ID Updated PRI criteria for Calcium (CT-007 and CT-297)
Version 13 to Version 14	Updated PRI criteria CT-037 (HBDH) Updated PRI criteria CT-130 (FSH) Updated CT-135 and CT-139 (PlGF) with a lowercase letter "L"

Version 14 to Version 15	Added criteria for Anti-HEV IgG Added criteria for Anti-HEV IgM
Version 15 to Version 16	Deleted Sections 'Ventana RUO' (all RUO products are captured on Parameter/Test 'RUO, LUO, ASR, IB, etc') Added criteria for Anti-AREG (SP272) Primary Ab Added criteria for Anti-EREG (SP326) Primary Ab Moved Section 'All other Ventana products not listed'
Version 16 to Version 17	Added criteria for aPTT HighS Added criteria for aPTT MedS Added criteria for aPTT LowS Updated PRI criteria CT-182 (Anti-HBc IgM) Updated PRI criteria CT-183 (Anti-Hbe) Updated PRI criteria CT-184 (Anti-HBs) Added criteria for CLDN18 (43-14A) RxDx Assay in Gastric Carcinoma
Version 17 to Version 18	Added criteria for Specific Proteins (AAGP)
Version 18 to Version 19	Added criteria for Ventana Antibodies - FGFR2b (FPR2-D) Assay 1 Added criteria for Therapeutic Drug Monitoring - MTX (ONLINE TDM Methotrexate) Added additional sample type to Cortisol criteria
Version 19 to Version 20	Updated Parameter/Test for CT-972 (TIB Assay HSV 1/2 & VSV) Added criteria for TIB UC hAdV PCR on x800 Added criteria for anti-CMV Added criteria for p40 (SP225) Rabbit Monoclonal PAB Update PRI criteria CT-342 (IGRA SARS-CoV-2) Retire all Reflotron criteria
Version 20 to Version 21	Added criteria for cobas® SARS-CoV-2, Influenza A/B & RSV nucleic acid test for use on the cobas Liat System Added criteria for p40 (BC28) Mouse Monoclonal Primary Antibody
Version 21 to Version 22	Added criteria for cobas Malaria for use on the 5800/6800/8800 Systems Updated criteria for Ventana - All other Ventana products not listed Fixed typo for TIB Assay HSV 1/2 & VSV (CT-972); VSV was fixed to VZV
Version 22 to Version 23	Added criteria for Special Proteins, Clinical Chemistry (HSCRP) Added criteria for Ventana Molecular Probes (VENTANA® HER2 Dual ISH DNA Probe Cocktail) Added criteria for Ventana Detection Kits (VENTANA® Red ISH DIG Detection Kit) Added criteria for Ventana Detection Kits (VENTANA® Silver ISH DNP Detection Kit) Under the tables, updated RMD term to Molecular Lab
Version 23 to Version 24	Added criteria for TIB Molbiol Removed several Ventana Parameters/Tests that fall under different sections or the product has been obsolesced Added a 'Note' under 'Criteria to classify as PRI' for 'All Cell Marque Complaints'
Version 24 to Version 25	Updated PRI criteria for CT-337, CT-968, CT-936 (under Section Virology) to remove 'False Positive' for SARS-CoV-2 & Influenza A/B Added 4 new criteria for Fertility / Hormones_iST1_MS Updated 'Section', 'Parameter/Test' and/or 'Criteria to classify as PRI' information for several products Removed several criteria that are no longer available or an equivalent criteria is captured
Version 25 to Version 26	Updated criteria for Coagulation Testing by Professionals (PT) (CT-248) Added 12 new criteria under Section: Virology, including adding Table H Added 3 new criteria for Ventana Antibodies and one new criteria for Ventana Special Stains
Version 26 to Version 27	Updated criteria for Consumer Tests performed by Professionals or Patients (PT) (CT-CT-286) Added 2 new criteria under Section: Bone Markers_iVitD_Mass Spec
Version 27 to Version 28	Removed 'False Positive' for CT-208 (under Section GenMark) Updated criteria for TIB Molbiol
Version 28 to Version 29	Added new criteria for Algorithm (KFRE) Updated several Blood Screening criteria Added new criteria for Continuous Glucose Monitoring device Split criteria for Consumer test performed by lay-user and professionals; creating a new criteria for professionals Added new criteria for Immunology (Pro-C3) Added 9 new criteria for LumiraDx tests Added new criteria for Ventana Antibodies (VENTANA MET (SP44) RxDx Assay) Update several Ventana criteria
Version 29 to Version 30	Updated criteria for Microbiology & HPV (Mycobacteria tuberculosis) Updated criteria for Consumer Tests performed by Professionals (Glucose)
Version 30 to Version 31	Added 'cobas Liat' to the sections for CT-356, CT-357, and CT-922 Updated criteria for CT-336, CT-375 and CT-376 under Virology
Version 31 to Version 32	Updated CT-988 to include 'False positve result' and 'Delayed Results' Added 7 new criteria under Section: Fertility / Hormones_iST2_MS (CT-1094 thru CT-1100)
Version 32 to Version 33	Added 4 new criteria under section Therapeutic Drug Monitoring_iISD_MS (CT-1101 through CT-1104) Adjusted criterion for CT-1085: Error > 1% absolute for values >/= 6.4% (e.g. 6.4% reported as 5.3%) Added new criterion CT-1105 for VENTANA Molecular Probes (EBER1 DIG Probe - ASR only)
Version 33 to Version 34	Added new criterion CT-1106 for Roche in vitro Companion Diagnostics (Elecsys Anti-AAVrh74)
Version 34 to Version 35	Added new criteria for Enzymes, PG I and PG II (CT-1107,CT-1108) Added new criterion CT-1109 for Drugs of Abuse (ETG) Added new criterion CT-1110 for Infectious Diseases (cobas® liat Bordetella panel nucleic acid test) Added new criteria CT-1111 to CT-1118 for US RBSS assays Updated CT-825 to add values for children Updated CT-380 to add criteria for blood screening Updated CT-1090: added criteria for no PRI Updated CT-965 and CT-975 (Virology): removed false positive criterion
Version 35 to Version 36	Added CT-1115 (HIV Duo) under US RBSS
Version 36 to Version 37	Updated CT-936 to include cobas® Respiratory 4-flex for use on the cobas 58/68/8800 systems Added new criteria CT-1119 to CT-1126 under section Therapeutic Drug Monitoring_iAED2_MS Removed CT-173 (Troponin T) from Cardiac Markers

Version 37 to Version 38	Updated CT-1111 to CT-1118: Changed US RBSS to RBSS (US), and adjusted criteria. Updated CT-181, CT-185, CT-186, CT-188, CT-189, CT-191, CT-192, CT-195, CT-205, CT-207, and CT-988 under Infectious Diseases to align RBSS criteria. Added new criteria CT-1129 to CT-1139 to new Section - RBSS (ROW), to align criteria with RBSS (US). Updated CT-336, CT-375, and CT-376 under Virology, to align with RBSS criteria. Updated CT-367, CT-370, CT-371, CT-823, CT-919, and CT-931 under Blood Screening, to align with RBSS criteria. Updated CT-884 under Blood Screening (US only), to align with RBSS criteria. Updated CT-184 under Infectious Diseases: removed blood donations stipulation, and added PSI to clinical allegations.
Version 38 to Version 39	Revised RBSS criteria from v38 to amend "discrepant positive results" to "discrepant results".
Version 39 to Version 40	Added new criterion CT-1128 for Ventana Antibodies (VENTANA HER2 (CAL27) Assay) Added new criteria CT-1140 to CT-1141 under Therapeutic Drug Monitoring_iAB1_MS Updated CT-1127 (Phospho-Tau (181P) Plasma) under Specific Proteins Updated CT-1107 and CT-1108 (PG I and PG II) under Enzymes
Version 40 to Version 41	Added new criteria CT-1142 (Kidney KlinRisk) and CT-1143 (LGI-Flag) under Algorithm Updated CT-174 (Troponin T hs / Gen5, under Cardiac Markers) to reference Gen6 in the parameter description Added new criteria CT-1144 to CT-1146 under Therapeutic Drug Monitoring_i3MV_MS
Version 41 to Version 42	Updated CT-369 (DPX – Parvo B19) under Blood Screening, to align with RBSS criteria. Added new criteria CT-1147 to CT-1150 under Therapeutic Drug Monitoring_iAED1_MS
Version 42 to Version 43	Added CT-1151 (GenMark Respiratory Panel 3 [RP3]) under GenMark Added CT-1152 to CT-1153 (PAX-8 [SP348] Rabbit Monoclonal Antibody, and TRPS1 [SP532] Rabbit Monoclonal Antibody) under Ventana Antibodies Added CT-1154 (Elecsys Anti-TPO II) under Thyroid
Version 43 to Version 44	Updated header in column D from "No evaluation for PRI is required" to "No escalation for PRI ER required" Updated header in column E from "Criteria to classify as PRI" to "Criteria to classify as PRI ER" Added CT-1155 (BV [Bacterial Vaginosis]) under Microbiology & HPV Added CT-1156 (Candida Vaginitis [CV]) under Microbiology & HPV / cobas Liat Updated CT-921: Section changed from "Microbiology & HPV" to "Microbiology & HPV / cobas Liat" Added CT-1157 and CT-1158 (VENTANA TROP2 [EPR20043] Rabbit Monoclonal Antibody; and VENTANA TROP2 [EPR20043] Rabbit Monoclonal Antibody RUO Assay) under Ventana Antibodies Added CT-1159 and CT-1160 (Meropenem [iMERO]; and Piperacillin [iPIPE]) under Therapeutic Drug Monitoring_iAB2_MS Added CT-1161 (Elecsys Dengue Ag) under Infectious Diseases
Version 44 to Version 45	Added CT-1162 (Beta-Amyloid [1-40] CSF) under Specific Proteins Updated CT-1127 (Phospho-Tau (181P) Plasma) under Specific Proteins Added criterion for CT-049 (Beta-Amyloid (1-42)), CT-916 (Phospho-Tau (181) CSF), and CT-917 (Total-Tau CSF), under Specific Proteins [previously no criteria specified] Updated CT-058 (HbA1c) under Specific Proteins Updated CT-273 (HbA1c) under Other Point of Care Tests Updated CT-1085 (HbA1c) under LumiraDx tests
Version 45 to Version 46	Further updated CT-058 (HbA1c) under Specific Proteins Further updated CT-273 (HbA1c) under Other Point of Care Tests Further updated CT-1085 (HbA1c) under LumiraDx tests
Version 46 to Version 47	Further updated CT-058 (HbA1c) under Specific Proteins Further updated CT-273 (HbA1c) under Other Point of Care Tests Further updated CT-1085 (HbA1c) under LumiraDx tests Added CT-1163 (Apolipoprotein E4 Plasma, Elecsys) under Specific Proteins Added CT-1164 (Elecsys Dengue IgM) under Infectious Diseases

Version 47 to Version 48	Added CT-1165 (DABI) under Therapeutic Drug Monitoring, for Partner Channel Technoclone Updated CT-048 (ASLO) under Specific Proteins Deleted CT-208 and combined targets with CT-345 under Genmark; also added Bordetella parapertussis Updated CT-1551 Parameter description and PRI criteria under Genmark Added CT-1166 (LightMix Newborn) under TIB MolBiol Added CT-1167 (Hematoxylin III) under Ventana Ancillary Reagents Added CT-1168 (VENTANA BAP1 (SP527) Rabbit Monoclonal Antibody) under Ventana Antibodies Updated CT-952 (6-Acetylmorphine) under Drugs of Abuse to "6-Acetylaminomorphine (6AM) ROW & US only", and added Partner Channel and sample type information Deleted CT-105 (Methamphetamine) and combined it with CT-096 under Drugs of Abuse Updated CT-096 (Amphetamines) under Drugs of Abuse to "Amphetamines (AMP2), Methamphetamine", and added PRI criteria and sample type Updated CT-097 (Barbiturates) under Drugs of Abuse to "Barbiturates (BARB) ROW", and added PRI criteria and sample type Updated CT-098 (Benzodiazepines) under Drugs of Abuse to "Benzodiazepines (BNZ2) ROW", and added sample type Updated CT-953 (Buprenophene) under Drugs of Abuse to "Buprenophene (BUP) ROW", and added Partner Channel and sample type Updated CT-099 (Cannabinoids) under Drugs of Abuse to "Cannabinoids (THC2) ROW", and added PRI criteria and sample type Updated CT-100 (Cocaine Metabolite) under Drugs of Abuse to "Cocaine Metabolite (COC2) ROW", and added PRI criteria and sample type Updated CT-101 (Ethanol (Alcohol)) under Drugs of Abuse to "Ethanol (alcohol, ETOH2", updated PRI criteria, and added sample type Updated CT-1109 (Ethyl glucuronide (ETG)) under Drugs of Abuse to add Partner Channel and sample type Updated CT-998 (Fentanyl) under Drugs of Abuse to "Fentanyl (FEN)", and added Partner Channel and sample type Updated CT-954 (Hydrocodone) under Drugs of Abuse to "Hydrocodone (HYD)", and added Partner Channel and sample type Updated CT-102 (LSD) under Drugs of Abuse to "Lysergic Acid Diethylamide (LSD)", and added PRI criteria and sample type Updated CT-103 (Methadone) under Drugs of Abuse to "Methadone (MDN2) ROW", and added PRI criteria and sample type Updated CT-104 (Methadone Metabolite (EDDP)) under Drugs of Abuse with Partner Channel and sample type Updated CT-106 (Methaqualone) under Drugs of Abuse to "Methaqualone (MTQL)", and added PRI criteria and sample type Updated CT-107 (Opiates) under Drugs of Abuse to "Opiates (OPI2)", and added PRI criteria and sample type Updated CT-108 (Oxycodone) under Drugs of Abuse to "Oxycodone (OXY)", and added Partner Channel and sample type Updated CT-109 (Phencyclidine (PCP)) and CT-110 (Propoxyphene (PPX)) under Drugs of Abuse with PRI criteria and sample type Added CT-1169 (Barbiturates (BARB) US only), CT-1170 (Benzodiazepines (BNZ2) US only), CT-1171 (Cannabinoids (THC2) US only), CT-1172 (Cocaine Metabolite (COC2) US only), and CT-1173 (Methadone (MDN2) US only) under Drugs of Abuse
Version 48 to Version 49	Added CT-1174 (Gram Stain Core Kit), and CT-1175 (Congo Red Stain Core Kit) under Ventana Special Stains Added CT-1176 (SSR Solution) under Ventana Ancillary Reagents Added CT-1177 (NFL [Elecsys NfL /Neurofilament light chain]) under Specific Proteins

PRI Screening Criteria for Erroneous Results (ER)
NOTE: Updated information are captured in red for easy visibility. Use the native Excel file when working with this document.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1090	Algorithm	KFRE		Clinically implausible results (alleged by a medical professional). Incorrect medical decisions (alleged by a medical professional). KFRE always/repeatedly returns the same output across multiple patients irrespective of input value. Numeric results <0% (negative results). Numeric results >100%. No PRI Evaluation is needed: - Only comparison of Roche assay to Roche assay result or Roche to competitor's result if not excluded by disclaimer is qualified to meet PRI criteria as given by the Roche assay. - If non-comparable assays are used, it is the customer lab's responsibility to ensure the values are correct. - If no Roche assay at all is used, it is the customer lab's responsibility to ensure the values are correct.
CT-1142	Algorithm	Kidney KlinRisk		Algorithm leading to discrepant score results, deviation >/= 30% Clinically implausible results (alleged by a medical professional). Incorrect medical decisions (alleged by a medical professional). Algorithm always/repeatedly returns the same output across multiple patients, irrespective of input value. Algorithm gives mathematically implausible results. No "PRI ER" Evaluation is needed: - Only comparison of Roche assay to Roche assay result or Roche to competitor's result if not excluded by disclaimer is qualified to meet PRI criteria as given by the Roche assay. - If non-comparable assays are used, it is the customer lab's responsibility to ensure the values are correct. - If no Roche assay at all is used, it is the customer lab's responsibility to ensure the values are correct. Note: even if no Roche Assay is used, a discrepant result of the algorithm may indicate a malfunction. Escalation to "PRI Malfunction" to be evaluated.
CT-1143	Algorithm	LGI-Flag		Algorithm leading to discrepant score results, deviation >/= 30% Algorithm result leading to an incorrect scoring. Mathematically implausible results (ER) may indicate a HW/SW malfunction (e.g. scoring below 0.1 or above 1.0, score repeatedly the same with different input data). If results of components of the calculation are deemed to be discrepant, the specific PRI ER criteria apply if a Roche test was used. ---------- PRI ER evaluation is not required: Age < 40 y and > 100 y (years inclusive). Error Message codes: - Non-Fatal Error: indicates a specific non-severe malfunction that did not affect the score calculation; In most cases, non-fatal errors should be referred to as FYI messages. - Fatal Error: indicates a malfunction that prevents the processing of the request, no calculation result will be provided; Fatal errors are sometimes expected when dealing with real world data in healthcare organizations.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-905	Anemia	Active B12		Error > 30% at values 30- 90 pmol/L
CT-175	Anemia	Ferritin		Error > 50%
CT-177	Anemia	Folate (RBC) (hemolysate sample)		Error > 30% at values > 170 ng/mL (385.1 nmol/L) hemolysate sample
CT-176	Anemia	Folate (serum, plasma)		Error > 30% at values > 2 ng/mL (4.5 nmol/L)
CT-178	Anemia	Vitamin B12		Error > 30% at values 140-400 pmol/L (190.4 - 544 pg/mL)
CT-291	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	BB	X	
CT-293	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	BEact	X	
CT-294	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	BEecf	X	
CT-295	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	BO2	X	
CT-296	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	BUN (Urea)		Error > 50% at values > 40 mg/dL (6.7 mmol/L)
CT-297	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Ca2+		Error > 20% at values < 1.14 mmol/L Error > 20% at values > 1.35 mmol/L
CT-299	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	cHCO3-st	X	
CT-300	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Cl-		Error > 10%
CT-302	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Creatinine		Error > 20% at values 0.9 - 1,1 mg/dL (79.6-97.2 µmol/L) Error > 30% at values > 1,1 mg/dL (97.2 µmol/L)
CT-303	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	ctCO2(B)	X	
CT-306	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	FIO2	X	
CT-307	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	FO2Hb	X	
CT-308	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Glucose		Error > 15 mg/dL (0.8 mmol/L) at values ≤ 50 mg/dL (2.8 mmol/L) Error > 30% at values > 50 mg/dL (2.8 mmol/L)

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-309	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	H+	X	
CT-310	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Hbf	X	
CT-311	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Hct		Error > 5% (absolute) at values < 25% (e.g. 6% reported as 12%) and > 45% (e.g 55% reported as 65%)
CT-312	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	HHb	X	
CT-313	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	K+		Error > 0.4 mmol/L at values 4 - 6 mmol/L Error > 0.8 mmol/L at values ≥ 6 mmol/L **<u>For children < 12 month</u>: Error > 0.6 mmol/L**
CT-314	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Lactate		Error > 30% at values < 4.5 mg/dL (0.5 mmol/L) Error > 30% at values > 15 mg/dL (1.7 mmol/L)
CT-315	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Li+		Error > 20% at values > 1.2 mmol/L Error > 20% at values < 0.9 mmol/L
CT-317	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Na+		Error > 5 mmol/L at values < 135 mmol/L Error > 5 mmol/L at values > 145 mmol/L
CT-318	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	nCa2+	X	
CT-319	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	O2Hb	X	
CT-320	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	p50		Error > 20%
CT-325	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	pHst	X	
CT-327	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Qs	X	
CT-328	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Qs/Qt	X	
CT-329	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Qt	X	

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-330	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	R	X	
CT-331	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	RI	X	
CT-333	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	SulfHb	X	
CT-335	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test	Whole Blood Bilirubin		In neonates*: Error > 30% at values < 6.0 mg/dL (100.0 µmol/L) Error > 20% at values ≥ 6.0 mg/dL (100.0 µmol/L) *Neonate is a child up to an age of 28 days.
CT-292	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test / V-TAC	BE		Error > 1 mmol/L
CT-298	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	cHCO3-		Error > 3 mmol/L
CT-301	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	COHb		Error > 5% abs. at values < 30%
CT-304	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	ctCO2(P)	X	
CT-305	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	ctO2	X	
CT-316	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	MetHb		Error > 4% abs at values of 0 - 10% Error > 40% rel at values > 10%
CT-321	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	pACO2	X	
CT-322	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	pAO2	X	
CT-323	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	pCO2		Error > 6 mmHg (0.8 kPa) at values < 75 mmHg (10 kPa) Error > 10% for values > 75 mmHg (10 kPa)
CT-324	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	pH		Error > 0.08
CT-326	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	pO2		Error > 12 mmHg (1.6 kPa) at values < 150 mmHg (20 kPa)

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-332	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	sO2		Error > 4% abs. (i.e. 94% instead of 99%)
CT-334	Blood Gases, Electrolytes, Metabolytes and CO-Oximetry Test/ V-TAC	tHb		Error > 2 g/dL (20 g/L)
CT-823	Blood Screening	Babesia		False negative. False positive. Discrepant results.
CT-373	Blood Screening	Dengue virus/Chikungunya virus (DenV/ChikV)		False Non-Reactive False Reactive if false reactive result led to rejection of an organ or a tissue for transplantation **NOTE:** When used as aid in diagnosis, no evaluation for PRI is required.
CT-370	Blood Screening	DPX – HAV		False negative. False positive. Discrepant results.
CT-369	Blood Screening	DPX – Parvo B19		False negative. False positive. Discrepant results.
CT-372	Blood Screening	Hepatitis E (HEV)		False Non-Reactive False Reactive if the false reactive result led to rejection of an organ or a tissue for transplantation **NOTE:** When used as aid in diagnosis, no evaluation for PRI is required.
CT-367	Blood Screening	MPX		False negative. False positive. Discrepant results.
CT-371	Blood Screening	West Nile Virus (WNV)		False negative. False positive. Discrepant results.
CT-919	Blood Screening	Zika		False negative. False positive. Discrepant results.
CT-931	Blood Screening	cobas Malaria for use on the 5800/6800/8800 Systems		False negative. False positive. Discrepant results.
CT-884	Blood Screening (US only)	anti-CMV		False negative. False positive. Discrepant results. Please note: Elecsys Anti-CMV and Elecsys CMV IgG are not on the same market. Elecsys anti-CMV is intended for US blood donor market.
CT-161	Bone Markers	Anti-CCP		Error > 30% at values > 17 U/mL Error > 30% at values ≤ 200 U/mL Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI
CT-162	Bone Markers	Beta-Crosslaps		Error > 30% at values > 0.45 ng/ml

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-163	Bone Markers	Osteocalcin		Error > 30%
CT-166	Bone Markers	P1NP		Error > 30% at values > 15 µg/L
CT-165	Bone Markers	PTH		Error > 4.5 pg/mL for results ≥6 pg/mL to ≤ 15 pg/mL Error > 30% for values > 15 pg/mL
CT-164	Bone Markers	PTH (1-84)		Error > 30%
CT-167	Bone Markers	Vitamin D		Error > 50% at values > 10 ng/mL
CT-1024	Bone Markers_iVitD_Mass Spec	25-Hydroxy Vitamin D total		1. Deviations of results > 30 % for values > 10.0 ng/mL and </= 30.0 ng/mL 2. Deviations of results > 50% for all values > 30.0 ng/mL, including results generated after dilution
CT-1025	Bone Markers_iVitD_Mass Spec	24,25-Dihydroxy Vitamin D total		Deviations of results > 50 % for values >/=1.00 ng/mL. The 24,25(OH)2-VitD values should be further evaluated in the 25(OH)VitD / 24,25(OH)2VitD ratio. • When all comparison measurement results lie < 1.00ng/mL, no PRI Criterion is applicable: it is strongly recommended to proceed with calculation of the 25(OH)-VitD /24,25(OH)2-VitD ratio for accurate interpretation in clinical settings. • When ratio 25(OH)VitD / 24,25(OH)2VitD is evaluated, all measurements must be released by the same method and on same system from the same sample (i.e. cobas i 601 MS).
CT-168	Cardiac Markers	CK-MB (mass)		Women: Error > 4.9 ng/mL Men: Error > 8.8 ng/mL
CT-169	Cardiac Markers	GDF-15	X	
CT-170	Cardiac Markers	Myoglobin		Error > 30% at values > 40 ng/mL Error > 50% at values > 500 ng/mL
CT-171	Cardiac Markers	proBNP		Error > 30% at values > 300 ng/L
CT-172	Cardiac Markers	Troponin I		Error >0.06 ng/mL for values ≤ 0.3ng/mL Error >20% for values > 0.3 ng/mL
CT-174	Cardiac Markers	Troponin T hs / Gen5 and Gen6		Error > 2.8 pg/mL for values ≥ 5 to < 14 pg/mL Error > 20% for values ≥ 14 pg/mL
CT-987	Cardiac Markers	Lp-PLA2		Error > 50% at values 225 U/L. Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI.
CT-908	Coagulation Testing by Professionals	Anti-Xa		Error > 30%
CT-243	Coagulation Testing by Professionals	aPTT		Error > 30%
CT-875	Coagulation Testing by Professionals	aPTT HighS		Error > 30%
CT-876	Coagulation Testing by Professionals	aPTT MedS		Error > 30%
CT-877	Coagulation Testing by Professionals	aPTT LowS		Error > 30%
CT-244	Coagulation Testing by Professionals	AT III		Error > 30% at values < 70%
CT-245	Coagulation Testing by Professionals	D-Dimer		Error > 30% at values > 0.5 µg FEU/mL

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-959	Coagulation Testing by Professionals	Factor II		Error > 30% at values < 60% Error > 30% at values > 115%
CT-960	Coagulation Testing by Professionals	Factor V		Error > 30% at values < 70% Error > 30% at values > 120%
CT-961	Coagulation Testing by Professionals	Factor VII		Error > 30% at values < 40% Error > 30% at values > 110%
CT-830	Coagulation Testing by Professionals	Factor VIII		Error > 30% at values < 60% Error > 30% at values > 150%
CT-999	Coagulation Testing by Professionals	Factor IX		Error >6.0% (absolute) for values < 20.0% Error >30% (relative) for values 20.0 - 55.0% Error > 30% (relative) for values > 160%
CT-938	Coagulation Testing by Professionals	Factor X		Error > 30% at values < 70%
CT-224	Coagulation Testing by Professionals	Factor XI		Error >6.0% (absolute) for values < 20.0% Error >30% (relative) for values 20.0 - 70.0% Error > 30% (relative) for values > 130%
CT-232	Coagulation Testing by Professionals	Factor XII		Error >6.0% (absolute) for values < 20.0% Error >30% (relative) for values 20.0 - 60.0% Error > 30% (relative) for values > 150%
CT-233	Coagulation Testing by Professionals	Factor XIII		Error >6.0% (absolute) for values < 20.0% Error >30% (relative) for values 20.0 - 60.0%
CT-246	Coagulation Testing by Professionals	Fibrinogen		Error > 30% at values < 200 mg/dL Error > 30% at values > 700 mg/dL
CT-939	Coagulation Testing by Professionals	Free Protein S		Error > 30% at values < 75%
CT-247	Coagulation Testing by Professionals	GDF-15		Error > 20%
CT-940	Coagulation Testing by Professionals	Lupus S, Lupus C		False negative and false positive result
CT-941	Coagulation Testing by Professionals	Protein C		Error > 50% at values < 70%

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-248	Coagulation Testing by Professionals	PT		Error ≥ 0.5 INR for all values ≤ 2.0 Error > 30% for all values > 2.0 INR For measurements using %Quick Error > 20% at values >30% Q Error > 30% at values ≤ 30%Q **If any of the following circumstances apply, a PRI is not required, regardless of the magnitude of the numeric differences:** -Elapsed time >12h (PT Rec) or >24h (PT Screen / PT Owren) between specimen drawing and measurement time (including repeats) can result in a significant variation value. NOTE: If the elapsed time is not known, the results should be evaluated for PRI.- - Discrepant results provided compare %Quick to INR - Discrepant results between Owren and Quick based PT assays - Discrepant results provided compare %Quick to INR - Patients on heparin above a certain threshold will be in a dynamic state beyond that recommended for near patient PT testing, resulting in a significant variation in INR values: Levels: 1. UFH: PT Rec 1.0 IU/mL, PT Owren 1.0 IU/mL, PT Screen NPP 0.65 IU/mL, INR=2.0 0.2 IU/mL 2. LMWH: PT Rec 1.5 IU/mL, PT Owren 1.5 IU/mL, PT Screen 1.2 IU/mL NOTE: If the dosage of heparin is not known, evaluation for PRI should be performed. - Certain anticoagulants will have a significant impact on the outcome of PT results. Discrepant results cannot be evaluated for patients under therapy with the following anticoagulants: Bivalirudin, Rivaroxaban, Apixaban, Edoxaban, Dabigatran, Argatroban, Fondaparinux - Discrepant results from patients on interfering drugs listed in the respective PT assay method sheet do not require escalation as PRI
CT-249	Coagulation Testing by Professionals	Thrombin Time		Error >30% **If any of the following circumstances apply, a PRI is not required, regardless of the magnitude of the numeric differences:** -Elapsed time longer than 4 hours between the drawing of specimens can result in a significant variation values. NOTE: If the elapsed time is not known, the results should be evaluated for PRI. -Elapsed time longer than 2 hours for heparinzed samples between the drawing of specimens can result in a significant variation values. NOTE: If the elapsed time is not known, the results should be evaluated for PRI. -Certain anticoagulants in the patient plasma will have a significant impact on the outcome of TT results. Discrepant results cannot be evaluated for patients under therapy with the following anticoagulants: Bivalirudin, Dabigatran or Argatroban -Discrepant results from patients on intefering drugs listed in the respective TT assay method sheet do not require escalation as PRI
CT-962	Coagulation Testing by Professionals	vWF Act		Error > 30% at values < 70% Error > 30% at values > 100%
CT-288	Consumer Tests (Others)	Cholesterol		Error > 50% at values 150 – 300 mg/dL (3.9 – 7.8 mmol/L)
CT-289	Consumer Tests (Others)	Lactate		Error > 30% at values > 15 mg/dL (1.7 mmol/L)
CT-290	Consumer Tests (Others)	Triglycerides		Error > 30% at values > 200 mg/dL (2.3 mmol/L)
CT-287	Consumer Tests performed by Professionals or Patients	aPTT		Error > 30%

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-285	Consumer Tests performed by Professionals	Glucose		Error > 15 mg/dL (0.8 mmol/L) at values ≤ 50 mg/dL (2.8 mmol/L) Error > 30% at values > 50 mg/dL (2.8 mmol/L) Discrepant results for neonates (i.e. newborn child up to 28 days) Error > 15 mg/dL (0.8 mmol/L) at values ≤ 75 mg/dL (4.2 mmol/L) Error > 15% at values > 75 mg/dL (4.2 mmol/L) **The following apply to all discrepant results:** • Discrepant results only qualify for PRI in instances where the blood samples used for the comparison have been collected within 15 minutes. • Allegation for discrepant results of any magnitude based on a potential interference.
CT-1091	Consumer test performed by lay-users (Accu-Chek)	Glucose		At least 2 glucose reading results from a valid comparison, obtained within 15 minutes of each other, fall outside A and B range of the Consensus Error Grid Tool.
CT-286	Consumer Tests performed by Professionals or Patients	PT		Error ≥ 0.5 INR for all values ≤ 2.0 Error > 30% for all values > 2.0 INR For measurements using %Quick Error > 20% Quick at values >30%Q Error > 30% at values ≤ 30%Q For measurements using Seconds (sec) Error ≥ 6 sec for all values ≤ 24 sec Error > 30% for all values > 24 sec <u>Complaints must be evaluated for PRI for any discrepancies under the following circumstances:</u> - Roche meter to Roche meter comparisons, including back to back with the same Roche meter <u>If any of the following circumstances apply, a PRI is not required for INR/ %Quick /Seconds, regardless of the magnitude of the numeric differences.</u> • Discrepant results involving a meter compared to a lab test where no adverse event is alleged. • Discrepant results to other POC meters; only Roche meter to Roche meter comparisons qualify for PRI. • Discrepant results compared to another unknown POC meter (the name of the POC meter that was used to compare results is not known) • Elapsed time longer than 4 hours between the drawing of specimens can result in a significant variation values. NOTE: If the elapsed time is not known, the results should be evaluated for PRI. • Coaguchek XS and Pro II models: Incorrect result is accompanied by a “c” • Vantus/INRange: Incorrect results accompanied by a blue control bottle • Testing strips used were expired • Strips used were compromised as noted by the indicator on the back of the strip • Discrepant results provided compare %Quick or Seconds to INR • Discrepant results were from the same puncture (fingerstick) or same finger
CT-1092	Continuous Glucose Monitoring (CGM) device sensor used by Consumer	Glucose		≥ 2 pairs of values, each pair with one Blood Glucose and one Continuous Glucose Measurement reading, fall outside zones A, B and C of the Consensus Error Grid Tool. Precondition: • Valid comparison method • Time interval 10 - 15 min for 2 samples

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-930	Custom Biotech	FFPET RNA Isolation Kit IVD ONLY		Incorrrect diagnostic results caused by the FFPET RNA Isolation Kit Product is adulterated and/or Product is misbranded
CT-343	Cytokines (Lymphokines) / Immunomodulators	Interferon Gamma (IFN-gamma)	X	
CT-952	Drugs of Abuse	6-Acetylaminomorphine (6AM) ROW & US only	X	Partner Channel Lin-Zhi International. Sample type: Urine.
CT-096	Drugs of Abuse	Amphetamines (AMP2), Methamphetamine		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.
CT-097	Drugs of Abuse	Barbiturates (BARB) ROW		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine, Serum, Plasma.
CT-1169	Drugs of Abuse	Barbiturates (BARB) US only		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.
CT-098	Drugs of Abuse	Benzodiazepines (BNZ2) ROW		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine, Serum, Plasma.
CT-1170	Drugs of Abuse	Benzodiazepines (BNZ2) US only		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.
CT-953	Drugs of Abuse	Buprenophene (BUP) ROW	X	Partner Channel Lin-Zhi International. Sample type: Urine.
CT-099	Drugs of Abuse	Cannabinoids (THC2) ROW		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine, Serum, Plasma.
CT-1171	Drugs of Abuse	Cannabinoids (THC2) US only		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.
CT-100	Drugs of Abuse	Cocaine Metabolite (COC2) ROW		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine, Serum, Plasma.
CT-1172	Drugs of Abuse	Cocaine Metabolite (COC2) US only		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.
CT-101	Drugs of Abuse	Ethanol (Alcohol, ETOH2)		Error >/= 30% Sample type: Urine, Serum, Plasma.
CT-1109	Drugs of Abuse	Ethyl glucuronide (ETG)	X	Partner Channel Lin-Zhi International. Sample type: Urine.
CT-998	Drugs of Abuse	Fentanyl (FEN)	X	Partner Channel Lin-Zhi International. Sample type: Urine.
CT-954	Drugs of Abuse	Hydrocodone (HYD)	X	Partner Channel Lin-Zhi International. Sample type: Urine.
CT-102	Drugs of Abuse	Lysergic Acid Diethylamide (LSD)		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.
CT-103	Drugs of Abuse	Methadone (MDN2) ROW		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine, Serum, Plasma.
CT-1173	Drugs of Abuse	Methadone (MDN2) US only		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-104	Drugs of Abuse	Methadone Metabolite (EDDP)	X	Partner Channel Lin-Zhi International. Sample type: Urine.
CT-106	Drugs of Abuse	Methaqualone (MTQL)		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.
CT-107	Drugs of Abuse	Opiates (OPI2)		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.
CT-108	Drugs of Abuse	Oxycodone (OXY)	X	Partner Channel Thermoscientific Sample type: Urine
CT-109	Drugs of Abuse	Phencyclidine (PCP)		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.
CT-110	Drugs of Abuse	Propoxyphene (PPX)		False negative or false positive when confirmed false positive or negative by GC/MS or LC/MS. Sample type: Urine.
CT-025	Enzymes	Acid Phosphatase	X	
CT-026	Enzymes	Aldolase	X	
CT-027	Enzymes	Alkaline Phosphatase		Error > 50% at values > 100 U/L
CT-028	Enzymes	ALT / GPT		Error > 50% at values > 20 U/L
CT-029	Enzymes	Amylase (pancreatic)		Error > 50% at values > 60 U/L
CT-030	Enzymes	Amylase (total)		Error > 50% at values > 100 U/L
CT-031	Enzymes	AST / GOT		Error > 50% at values > 20 U/L
CT-032	Enzymes	Cholinesterase		Error > 30% at values < 2000 U/L
CT-033	Enzymes	CK		Error > 30% at values > 50 U/L
CT-034	Enzymes	CK-MB (activity)		Error > 20% at values between 20 - 40 U/L Error > 30% at values > 41 U/L
CT-035	Enzymes	GGT		Error > 50% at values > 20 U/L
CT-036	Enzymes	GLDH		Error > 30% at values > 5 U/L
CT-037	Enzymes	HBDH		30% >/= 127 U/L
CT-038	Enzymes	LDH/LD (L or P mode)		Error > 30% at values > 200 U/L
CT-039	Enzymes	Lipase		Error > 50% at values > 70 U/L
CT-1107	Enzymes	PG I		1) Incorrect results (falsely low / falsely high) of PGI exceeding a relative deviation of 30%. 2) Incorrect relevant clinical decision based on PGI without actual harm
CT-1108	Enzymes	PG II		1) Incorrect results (falsely low / falsely high) of PGII exceeding a relative deviation of 30%. 2) Incorrect relevant clinical decision based on PGII without actual harm.
CT-122	Fertility / Hormones	ACTH		Error > 30% at values > 10 pmol/L
CT-123	Fertility / Hormones	AFP		Error > 30%. Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI.
CT-124	Fertility / Hormones	AMH		Error > 20% at values > 0.7 pmol/L (0.1 ng/mL)
CT-834	Fertility / Hormones	Androstenedione	X	
CT-127	Fertility / Hormones	Cortisol		Error >40% for serum and plasma samples Error > 50% for urine samples and saliva sample
CT-128	Fertility / Hormones	DHEA-S		Error > 30% at values > 5 µg/dL (0.14 µmol/L)

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-129	Fertility / Hormones	Estradiol		Error > 30% at values > 30 pg/mL
CT-126	Fertility / Hormones	Free beta-HCG		Error > 20% at values > 10 IU/L
CT-130	Fertility / Hormones	FSH		Error > 30%
CT-125	Fertility / Hormones	HCG		Error > 3.00 mIU/mL at values <= 10 mIU/mL Error > 30% at values > 10 mIU/mL
CT-951	Fertility / Hormones	HCG + beta		Error >1.5 mIU/mL at values ≤ 5 mIU/mL Error >30% at values > 5 mIU/mL
CT-132	Fertility / Hormones	hGH		Basic values and suppression: Error > 0.5 ng/mL at values > 1 ng/mL Stimulation tests: Error > 30% at values > 8 ng/mL
CT-906	Fertility / Hormones	IGF-1		Error > 30% at values < 100 ng/mL Error > 30% at values > 300 ng/mL
CT-942	Fertility / Hormones	IGFBP-3	X	
CT-131	Fertility / Hormones	Insulin		Error > 50%
CT-133	Fertility / Hormones	LH	X	
CT-134	Fertility / Hormones	PAPP-A		Error > 30% at values < 4000 mIU/L
CT-135	Fertility / Hormones	PlGF		Error > 20% at values > 20 pg/mL
CT-136	Fertility / Hormones	Progesterone	X	
CT-137	Fertility / Hormones	Prolactin		Error > 30% at values > 3 ng/mL (63.3 µIU/mL)
CT-138	Fertility / Hormones	sFLT1		Error > 20% at values > 500 pg/mL
CT-139	Fertility / Hormones	sFLT1/ PlGF ratio		Calculated ratio could lead to a wrong conclusion with respect to diagnosis and prediction of preeclampsia
CT-140	Fertility / Hormones	SHBG		Error > 30% at values > 5 nmol/L
CT-141	Fertility / Hormones	Testosterone		Error > 30% at values > 3 nmol/L
CT-340	Fertility / Hormones - Software	Preeclampsia Screening		Customer alleges a discrepant false negative assessment which results in serious injury or death
CT-142	Fertility / Hormones - Software	Trisomy Screening		Customer alleges a discrepant false high assessment which results in unnecessary interventions
CT-932	Fertility / Hormones_iST1_MS	Aldosterone		Deviation of results > 12.5 pg/mL for values >/= 25.0 pg/mL and < 125 pg/mL Deviation of results > 30% for values =/> 125 pg/mL, including diluted sample results. PRI Criterion validity: • only comparison between results released by mass spectrometry method. • only Roche results of Aldosterone iPack ST1 compared to Roche results Aldosterone iPack ST1 on cobas i 601 analyzers qualify for PRI. • Results from other system/method/reagent combinations should not be compared to Roche cobas i 601(cobas MS) results. • If the comparison method is not known, results should be evaluated for PRI. • Deviations of exactly 30% do not apply for PRI Criterion, as considered still acceptable.
CT-933	Fertility / Hormones_iST1_MS	Androstenedione	X	

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-934	Fertility / Hormones_iST1_MS	Estradiol		Deviation of results > 0.750 pg/mL for values >/= 1.50 pg/mL and < 7.50 pg/mL. Deviation of results > 30% for values >/= 7.50 pg/mL, including diluted sample results. PRI Criterion validity: • Only comparison between results released by mass spectrometry method. • Only Roche results of Estradiol of iPack ST1 compared to Roche results Estradiol iPack ST1 on cobas i 601 analyzers qualify for PRI. • Results from other system/method/reagent combinations should not be compared to Roche cobas i 601 (cobas MS) results. • If the comparison method is not known, results should be evaluated for PRI. Deviations of exactly 30% do not apply for PRI Criterion, as considered still acceptable.
CT-935	Fertility / Hormones_iST1_MS	Testosterone		Deviation of results > 0.00500 ng/mL for values >/= 0.0100 ng/mL and < 0.0500 ng/mL Deviation of results > 30% for values >/= 0.0500 ng/mL, including diluted sample results. PRI Criterion validity: • Only comparison between results released by mass spectrometry method. • Only Roche results of Testosterone of iPack ST1 compared to Roche results Testosterone iPack ST1 on cobas i 601 analyzers qualify for PRI. • Results from other system/method/reagent combinations should not be compared to Roche cobas i 601 (cobas MS) results. • If the comparison method is not known, results should be evaluated for PRI. • Deviations of exactly 30% do not apply for PRI Criterion, as considered still acceptable.
CT-1094	Fertility / Hormones iST2_MS	i17OHP (17-Hydroxyprogesterone)		Deviation of results > 50% for values above 0.125ng/mL and < 2.00ng/mL Deviation of results > 30% for values >/= 2.00ng/mL and </= 10.00ng/mL Deviation of results >50% for values > 10.00ng/mL, including diluted samples
CT-1095	Fertility / Hormones iST2_MS	iCORT (Cortisol total)		Deviation of results in Serum/Plasma samples > 30% for values =/> 5 ng/mL, including diluted samples
CT-1096	Fertility / Hormones iST2_MS	iDHT (Dihydrotestosterone)		Deviation of results > 30% for values =/> 0.100 ng/mL, including diluted samples
CT-1097	Fertility / Hormones iST2_MS	iCORN (Cortisone)	X	
CT-1098	Fertility / Hormones iST2_MS	iDHEA (Dehydroepiandrosterone)	X	
CT-1099	Fertility / Hormones iST2_MS	iDHEAS (Dehydroepiandrosterone-Sulfate	X	
CT-1100	Fertility / Hormones iST2_MS	iPROG (Progesterone)	X	
CT-345	GenMark	Respiratory Panel - any of the following targets: - Bordetella pertussis - Bordetella parapertussis - Chlamydia pneumoniae - Influenza (A, A H1, A H1-2009, A H3, B) - Legionella pneumophila - MERS-CoV - Mycoplasma pneumoniae - SARS-CoV-2		False Negative

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-346	GenMark	Respiratory Panel - any of the following targets: - Adenovirus - Coronaviruses (229E, NL63, OC43, HKU1) - Human Bocavirus - Human Metapneumovirus - Human Rhinovirus/Enterovirus - Parainfluenza Virus 1-4 - Respiratory Syncytial Virus (A, B)	X	
CT-1151	GenMark	Respiratory Panel - any of the following targets: - Bordetella pertussis - Legionella pneumophila		False Positive
CT-347	GenMark	Blood Culture Identification - Gram Negative - any of the following targets: - Haemophilus influenzae - Neisseria meningitidis - Salmonella - Stenotrophomonas maltophilia - IMP - KPC - NDM - OXA (OXA-23 and OXA-48) - VIM		False Negative False Positive
CT-379	GenMark	Blood Culture Identification - Gram Negative - any of the following targets: - Acinetobacter baumannii - Bacteroides fragilis - Citrobacter - Cronobacter sakazakii - Enterobacter (non-cloacae complex) - Enterobacter cloacae complex - Escherichia coli - Fusobacterium nucleatum - Fusobacterium necrophorum - Klebsiella oxytoca - Klebsiella pneumoniae group - Morganella morganii - Proteus - Proteus mirabilis - Pseudomonas aeruginosa - Serratia - Serratia marcescens - CTX-M - Pan Gram Positive - Pan Candida		False Negative

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-869	GenMark	Blood Culture Identification - Gram Positive - any of the following targets: - Listeria - Listeria monocytogenes - Streptococcus pneumoniae		False Negative False Positive
CT-362	GenMark	Blood Culture Identification - Gram Positive - any of the following targets: - Enterococcus - Enterococcus faecalis - Enterococcus faecium - Staphylococcus aureus - Staphylococcus lugdunensis - Streptococcus agalactiae - Streptococcus pyogenes - mecA - mecC - vanA - vanB - Pan Gram Negative - Pan Candida		False Negative
CT-866	GenMark	Blood Culture Identification - Gram Positive - any of the following targets: - Bacillus cereus group - Bacillus subtillis group - Corynebacterium - Cutibacterium acnes (Propionibacterium acnes) - Lactobacillus - Micrococcus - Staphylococcus - Staphylococcus epidermidis - Streptococcus - Streptococcus anginosus group	X	
CT-374	GenMark	Blood Culture Identification - Fungal Panel - for the following target: - Candida auris		False Negative False Positive

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-867	GenMark	Blood Culture Identification - Fungal Panel - for the following target: - Candida albicans - Candida dubliniensis - Candida famata - Candida glabrata - Candida guilliermondii - Candida kefyr - Candida krusei - Candida lusitaniae - Candida parapsilosis - Candida tropicalis - Cryptococcus gattii - Cryptococcus neoformans - Fusarium - Rhodotorula		False Negative
CT-348	Genomics & Oncology	BRAF V600 Mutation		False Mutation Detected False No Mutation Detected
CT-350	Genomics & Oncology	EGFR Mutation		False Mutation Detected False No Mutation Detected
CT-973	Genomics & Oncology	EZH2		False Mutation Detected False No Mutation Detected
CT-351	Genomics & Oncology	Factor II		False Mutation Detected False No Mutation Detected
CT-352	Genomics & Oncology	Factor V		False Mutation Detected False No Mutation Detected
CT-349	Genomics & Oncology	KRAS Mutation		False Mutation Detected False No Mutation Detected
CT-250	Genomics & Oncology	ADCY9		False Detected False Not Detected
CT-979	Genomics & Oncology	PIK3CA		False Negative False Positive

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1093	Immunology	Pro-C3		Numeric PRI Criterion: Error ≥ 30% Pro-C3 result (whole range) Clinical PRI Criterion: • Clinically implausible results (alleged by a medical professional) due to the Pro-C3 result as part of the ADAPT algorithm. • Incorrect medical decisions (alleged by a medical professional) due to the Pro-C3 result as part of the ADAPT algorithm. No PRI Evaluation is needed: • Not to be used in a primary care population without prior confirmatory tests. • Not intended for use in patients with other chronic liver disease including hepatitis B or C. • Not to be used in patients with excessive alcohol consumption. • One of the 4 input data for adapt has not been used, because all 4 ADAPT input values: platelet count, PRO-C3 assay result, diabetes status (yes/no), and patient age must be used. • ADAPT should be used in an enriched adult (≥ 18 years). • Error in input values platelet count, diabetes status, patient age do not require PRI evaluation. • Different than Roche method for Pro-C3 testing used. • Different ADAPT algorithm used than provided with the Pro-C3 method sheet. • Pro-C3 used outside ADAPT algorithm.
CT-179	Infectious Diseases	Anti-HAV		Error > + 50% at values > 20 IU/L
CT-826	Infectious Diseases	Anti-HAV II (qualitative)		False negative or false positive result
CT-180	Infectious Diseases	Anti-HAV IgM		False negative or false positive result
CT-181	Infectious Diseases	Anti-HBc		False negative. False positive. Discrepant results.
CT-182	Infectious Diseases	Anti-HBc IgM		False negative False positive
CT-183	Infectious Diseases	Anti-HBe		False negative or false positive result False non-reactive or false reactive False COI > 1.0 (non-reactive) or false COI </= 1.0 (reactive)
CT-184	Infectious Diseases	Anti-HBs		**Disease monitoring:** - False positive for anti-HBs - False reactive for anti-HBs - False high >/= 10 IU/L - False negative for anti-HBs - False non-reactive for anti-HBs - False low < 10 IU/L **Vaccination / Booster:** - False high >/= 100 IU/L - False low < 100 IU/L **Clinical allegations to be escalated to PRI/PSI:** - Incorrect result leading to missed vaccination / revaccination / booster - Incorrect result leading to unnecessary revaccination / booster - Incorrect assessment as non-responder to vaccination based on incorrectly low anti-HBs titers.
CT-185	Infectious Diseases	Anti-HCV		False negative. False positive. Discrepant results.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-967	Infectious Diseases	Anti-SARS-CoV-2		**False Positive:** Elecsys result positive and repeat testing on Elecsys is negative. Anti-SARS-C0V-2 results are not comparable to Anti-SARS-CoV-2 S results and should not be compared. OR Elecsys result positive and two different manufacturer products detecting SARS-CoV2 N (nucleocapsid antibodies) are negative. If it is not known if the other manufacturer products detect SARS CoV2 N, PRI evaluation is required. **False Negative:** Elecsys result negative and repeat testing on Elecsys is positive OR An Elecsys result is negative and the Patient is PCR positive and PCR result was 30 days or more prior to the Elecsys testing. If the date of PCR testing is not known, PRI is not required.
CT-974	Infectious Diseases	Anti-SARS-CoV-2 S		**False Positive:** Elecsys result positive and repeat testing on Elecsys is negative. Anti-SARS-C0V-2 S results are not comparable to Anti-SARS-CoV-2 results and should not be compared. OR Elecsys result positive and two different manufacturer products detecting SARS CoV2 S(spike antibodies) are negative. If it is not known if the other manufacturer products detect SARS CoV2 S, PRI evaluation is required. **False Negative:** Elecsys result negative and repeat testing on Elecsys is positive OR An Elecsys result is negative and the Patient is PCR positive and PCR result was 30 days or more prior to the Elecsys testing. If the date of PCR testing is not known, PRI is not required.
CT-338	Infectious Diseases	Covid / Pilot 19 At Home Test (US Lay Use Test from SD Biosensor) [OTC] SARS-CoV-2 Rapid Antigen Test US Professional Use Test from SD Biosensor [Pro] SARS-CoV-2 Rapid Antigen Tests (NP, Nasal or Nasal PST from SD BioSensor) for Rest of World (RoW) not for US		SD Biosensor test result compared to SD Biosensor test result: False positive or false negative result within 24 hours. SD Biosensor test result compared to PCR test: False positive or false negative result within 24 hours. Note: If the time between the comparison tests is unknown the case should NOT be considered as a PRI. For non-erroneous results issue (e.g. labeling): If any of the following circumstances apply, a PRI is not required: - If the instructions for use or quick reference guide is missing. - If the expiration date / lot number label on the outer KIT box is missing.
CT-969	Infectious Diseases	SARS-CoV-2 Rapid Antibody Test (SD Biosensor)		Erroneous results that led to Death, Serious Injury and/or Serious Harm

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-977	Infectious Diseases	Roche SARS-CoV-2 Antigen Test		**False Positive:** Elecsys result positive and repeat testing on Elecsys is negative OR Elecsys result positive and PCR is negative **False Negative:** Elecsys result negative and repeat testing on Elecsys is positive OR Elecsys result is negative and 2 different antigen assays are positive
CT-341	Infectious Diseases	SARS-CoV-2 / FluA/B Rapid Antigen (AG) test from SD Biosensor		**For SARS-CoV-2:** SD Biosensor test result compared to SD Biosensor test result: False positive or false negative result within 24 hours. SD Biosensor test result compared to PCR test: False positive or false negative result within 24 hours. **For Influenza A/B:** SD Biosensor test result compared to SD Biosensor test result: False positive or false negative result within 24 hours. SD Biosensor test result compared to PCR test: False positive or false negative result within 24 hours. For either test, If the time between the comparison tests is unknown the case should NOT be considered as a PRI.
CT-186	Infectious Diseases	Chagas		False negative. False positive. Discrepant results.
CT-870	Infectious Diseases	Anti-HEV IgG		False negative result False non-reactive result Result incorrectly < 0.15 U/mL
CT-872	Infectious Diseases	Anti-HEV IgM		False negative or false positive result False non-reactive or false reactive result Result incorrectly COI < 1.0 or result incorrectly >/= 1.0
CT-1161	Infectious Diseases	Elecsys Dengue Ag		False negative results False positive results Incorrect clinical decision without actual harmful event Delay of results should be covered by PRI malfunction
CT-1164	Infectious Diseases	Elecsys Dengue IgM		False Positive Results. False Negative Results. Incorrect clinical decision/diagnosis without actual harm. Delayed Results: no ER; shall be evaluated for PRI Malfunction. Actual harm is covered by PSI escalation.
CT-955	Infectious Diseases	EBV EBNA IgG		False negative or false positive result
CT-956	Infectious Diseases	EBV VCA IgG		False negative or false positive result
CT-957	Infectious Diseases	EBV IgM		False negative or false positive result

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-187	Infectious Diseases	HBeAg (Qual)	X	
CT-985	Infectious Diseases	HBeAg (Quant)	X	
CT-188	Infectious Diseases	HBsAg		False negative. False positive. Discrepant results.
CT-189	Infectious Diseases	HBsAg Confirmatory Test		False negative. False positive. Discrepant results.
CT-190	Infectious Diseases	HBsAg quant		Error > 30% at values > 100 IU/mL
CT-988	Infectious Diseases	HCV DUO		False negative. False positive. Discrepant results.
CT-193	Infectious Diseases	HIV Ag Confirmatory Test		False negative or false positive result
CT-191	Infectious Diseases	HIV combi PT		False negative. False positive. Discrepant results.
CT-192	Infectious Diseases	HIV DUO		False negative MAIN result. Sub results from this test are not evaluated for PRI. False positive. Discrepant results.
CT-194	Infectious Diseases	HIV p24 Ag		False negative result
CT-205	Infectious Diseases	HTLV –I/II		False negative. False positive. Discrepant results.
CT-342	Infectious Diseases	IGRA SARS-CoV-2		Erroneous reactive result. Erroneous non-reactive result. <u>PRI Criterion Validity:</u> - Only comparison between results released with IGRA Method on Roche Elecsys Systems designed for IGRA SARS application. - Only comparison between results released by re-testing of the sample from the same blood draw and measured with designed T cell stimulation tubes and Reagent/working solutions having IGRASARS label. - Result interpretation is only valid if the Ag Tube is used in conjunction with the control tubes (NC and C) provided and if the blood draw occurred at the same time. Assure use of NC, Ag and PC Tubes from the same manufacturing lot only. - PRI Criterion is not applicable for the partial measurements, from NC, PC and Ag tube, being the mandatory steps within IGRA Method.
CT-943	Infectious Diseases	RPR		False negative result
CT-207	Infectious Diseases	Syphilis		False negative. False positive. Discrepant results.
CT-206	Infectious Diseases	TPLA		False negative result
CT-195	Infectious Diseases	CMV IgG		False negative. False positive. Discrepant results.
CT-197	Infectious Diseases	CMV IgG Avidity		False low or false high avidity index

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-196	Infectious Diseases	CMV IgM		False negative result
CT-198	Infectious Diseases	HSV 1 IgG		False negative or false positive result
CT-199	Infectious Diseases	HSV 2 IgG		False negative or false positive result
CT-200	Infectious Diseases	Rubella IgG		False negative or false positive result.
CT-201	Infectious Diseases	Rubella IgM		False negative result.
CT-202	Infectious Diseases	Toxoplasma IgG		False negative or false positive result.
CT-204	Infectious Diseases	Toxoplasma IgG Avidity		False low or false high avidity index
CT-203	Infectious Diseases	Toxoplasma IgM		False negative or false positive result
CT-883	Infectious Diseases	TIB UC hAdV PCR on x800		False negative result
CT-920	Infectious Diseases	Zika IgG		False negative or false positive result
CT-1110	Infectious Diseases	cobas® liat Bordetella panel nucleic acid test		False Positive and False Negative results for *Bordetella pertussis*
CT-040	ISE	ISE-Chloride		Error > 10%
CT-043	ISE	ISE-Lithium		Error > 20% at values < 0.9 mmol/L Error > 20% at values > 1.2 mmol/L
CT-041	ISE	ISE-Potassium		Error > 0.4 mmol/L at values 4 - 6 mmol/L Error > 0.8 mmol/L at values > 6 mmol/L **For children < 12 month: Error > 0.6 mmol/L**
CT-042	ISE	ISE-Sodium		Error > 5 mmol/L at values < 135 mmol/L Error > 5 mmol/L at values > 145 mmol/L
CT-1081	LumiraDx tests	CRP		Error > 30% Regardless of the magnitude: • Incorrect clinical decision (e.g. inappropriate/missed therapy/diagnostic measures) due to an erroneous result and allegation not meeting PSI criteria. If any of the following circumstances apply, a PRI is not required, regardless of the magnitude of the numeric differences: • In individuals younger than 2 years old. • Not intended as risk marker of coronary heart disease.
CT-1082	LumiraDx tests	D-Dimer		Error > 30% Regardless of the magnitude: • Around the MDP (500 µg/L FEU 0.5/ mg/L FEU) discrepant values leading to incorrect classification (false low results: incorrectly VTE considered to be unlikely. False high results: Incorrectly VTE cannot be excluded). • Incorrect clinical decision (e.g. inappropriate/missed therapy/diagnostic measures) due to an erroneous result and allegation not meeting PSI criteria. If any of the following circumstances apply, a PRI is not required, regardless of the magnitude of the numeric differences: • In individuals younger than 18 years old. • Pregnant women.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1083	LumiraDx tests	INR		Error > 30% Error ≥ 0.5 INR for all values ≤ 2 Regardless of the magnitude: • Discrepant values leading to incorrect classification (false low results: incorrectly INR considered to be below TR / within TR. False high results: Incorrectly INR considered to be above TR / within TR. • Incorrect clinical decision (e.g. inappropriate/missed therapy/diagnostic measures) due to an erroneous result and allegation not meeting PSI criteria. If any of the following circumstances apply, a PRI is not required, regardless of the magnitude of the numeric differences: • In individuals younger than 18 years old. • Patients being treated with any direct thrombin inhibitors, including Hirudin, Lepirudin, Bivalirudin, and Argatroban. • Patient samples INR > 4.5 treated with Fondaparinux.
CT-1084	LumiraDx tests	NT-proBNP		Error > 30% Regardless of the magnitude: • Around the MDP (125 pg/mL (ng/L) discrepant values leading to incorrect classification (false low results: incorrectly probability of heart failure considered to be “very low”. False high results: Incorrectly diagnosis of heart failure established). • Incorrect clinical decision (e.g. inappropriate/missed therapy/diagnostic measures) due to an erroneous result and allegation not meeting PSI criteria. If any of the following circumstances apply, a PRI is not required, regardless of the magnitude of the numeric differences: • In individuals younger than 18 years old.
CT-1085	LumiraDx tests	HbA1c		Error =/>0.5% (absolute) for values of 5.2 - 6.4%* Error =/> 1% (absolute) for values =/> 6.5% Error =/> 5 mmol/mol for values 33 - 47 mmol/mol* Error =/>10 mmol/mol for values =/> 48 mmol/mol * endpoints included. Regardless of the magnitude: Incorrect clinical decision or classification (e.g. inappropriate/missed therapy/diagnostic measures) due to an erroneous result and allegation not meeting PSI criteria. If any of the following circumstances apply, a PRI is not required, regardless of the magnitude of the numeric differences: • In individuals younger than 2 years old. • In conditions associated with an altered relationship between HbA1c and glycemia and/or decreased red blood cell survival, the LumiraDx HbA1c test should not be used. Some examples include diseasestate/homozygous hemoglobinopathies (including sickle cell disease), hemolytic anemia, iron deficiency anemia, polycythemia, pregnancy (second and third trimesters and the postpartum period), glucose-6-phosphate dehydrogenase deficiency, malignancies, severe chronic hepatic and renal diseases, HIV, hemodialysis, recent blood loss, recent transfusion, or erythropoietin therapy. • Specimens containing HbF (>7.2%), including patients with Hereditary Persistence of Fetal Hemoglobin (HPFH), may result in lower than expected HbA1c values and should not be reported.
CT-1086	LumiraDx tests	SARS-CoV-2 Ag		False negative False positive

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1087	LumiraDx tests	SARS-CoV-2 Ag Ultra		False negative False positive
CT-1088	LumiraDx tests	SARS-CoV-2 Ag & Flu A/B		For SARS-CoV-2: - False negative - False positive For FLU A/B: - False negative - False positive
CT-1089	LumiraDx tests	SARS-CoV-2 Ag & RSV		For SARS-CoV-2: - False negative - False positive For RSV: - False negative - False positive
CT-1155	Microbiology & HPV	BV (Bacterial Vaginosis)	X	
CT-365	Microbiology & HPV	Group A Strep	X	
CT-361	Microbiology & HPV	Herpes Simplex Virus 1/2		False Negative
CT-366	Microbiology & HPV	Human Papilloma Virus	X	
CT-358	Microbiology & HPV	Methicillin-resistant *Staphylococcus aureus (MRSA)*	X	
CT-359	Microbiology & HPV	Methicillin-sensitive *Staphylococcus aureus (MSSA)*	X	
CT-355	Microbiology & HPV	*Mycobacterium intracellulare* in immunocompromised patients	X	
CT-354	Microbiology & HPV	*Mycobacterium tuberculosis*		False Positive False Negative
CT-824	Microbiology & HPV	MTB-RIF/INH		False Positive False Negative
CT-353	Microbiology & HPV	SeptiFast		False Negative
CT-970	Microbiology & HPV	ResistancePlus GC (Antibiotic resistance for *Neisseria gonorrhoeae*)	X	
CT-971	Microbiology & HPV	ResistancePlus MG (Antibiotic resistance for *Mycoplasma genitalium*)		False Negative
CT-356	Microbiology & HPV / cobas Liat	*Chlamydia trachomatis*	X	
CT-1156	Microbiology & HPV / cobas Liat	Candida Vaginitis (CV)	X	
CT-357	Microbiology & HPV / cobas Liat	*Neisseria gonorrhoeae*	X	
CT-921	Microbiology & HPV / cobas Liat	*Trichomonas vaginalis* (TV)		False Negative

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-922	Microbiology & HPV / cobas Liat	*Mycoplasma genitalium* (MG)		False Negative
CT-360	Microbiology & HPV / cobas Liat	*Clostridium difficile*	X	
CT-363	Microbiology & HPV / cobas Liat	Influenza A/B		False Negative
CT-364	Microbiology & HPV / cobas Liat	Influenza A/B and RSV		**For Influenza A/B target the following criteria apply:** False Negative **For RSV target the following criteria apply**: No evaluation for PRI is required
CT-880	Non-Invasive Prenatal Testing (NIPT)	Harmony CE-IVD		False Positive or False Negative results
CT-209	Other Immunochemistry Tests (Autoimmunity, Critical Care)	IL-6		Error >30% at values > 7 pg/mL For testing in the United States and for patients with ventilation problems and/or at risk of machine ventilation: Error > 30% at values < 35 pg/mL and Error > 30% at values > 7 pg/mL
CT-210	Other Immunochemistry Tests (Autoimmunity, Critical Care)	Procalcitonin		Error > 30% for values > 0.5 ng/mL to 10 ng/mL Error > 70% for values > 10 ng/mL
CT-211	Other Immunochemistry Tests (Autoimmunity, Critical Care)	S100		Error > 20% at values > 0.1 µg/L. For tumor monitoring testing should be performed with a single/same method.
CT-273	Other Point of Care Tests	HbA1c		Error =/>0.5% (absolute) for values of 5.2 - 6.4%* Error =/> 1% (absolute) for values =/> 6.5% Error =/> 5 mmol/mol for values 33 - 47 mmol/mol* Error =/>10 mmol/mol for values =/> 48 mmol/mol * endpoints included. Regardless of the magnitude: Incorrect clinical decision or classification (e.g. inappropriate/missed therapy/diagnostic measures) due to an erroneous result and allegation not meeting PSI criteria.
CT-277	Other Point of Care Tests	Troponin T Sensitive (Qualitative)		False negative result
CT-278	Other Point of Care Tests	CK-MB		Women: > 4.9 ng/ml Men: > 8.8 ng/mL
CT-910	Other Point of Care Tests	CRP		Error > 30%
CT-279	Other Point of Care Tests	D-Dimer		Error > 30% at values > 0.5 µg/mL
CT-280	Other Point of Care Tests	Myoglobin		Error > 30% at values > 50 ng/mL Error > 50% at values > 500 ng/mL
CT-281	Other Point of Care Tests	proBNP		False negative result
CT-283	Other Point of Care Tests	Troponin T		Difference > 30% at values > 0.1 ng/mL or results are inconsistent (e.g. second result shows different category)
CT-274	Other Point of Care Tests - Cobas Lipid Panel	Cholesterin total		Error > 50% at values 150 – 300 mg/dL (3.9 – 7.8 mmol/L)

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-275	Other Point of Care Tests - Cobas Lipid Panel	HDL Cholesterol		Error > 50% at values 35- 65 mg/dL (0.9-1.7 mmol/L)
CT-276	Other Point of Care Tests - Cobas Lipid Panel	Triglycerides		Error > 30% at values > 200 mg/dL (2.3 mmol/L)
CT-252	Platelet Testing	ADPtest		Aggregation (AUC) > 40 U in patients with a ADP receptor defect or Glanzman thrombasthenia confirmed by flow cytometry or genetic analysis. AUC = Area Under Curve; For all other patients: Hirudin sample: Error > 30% at values ≥ 52 U Error > 16U at values < 52 U Citrated sample: Error > 30% at values ≥ 40 U Error > 16U at values < 40 U
CT-251	Platelet Testing	ADPtest HS		Hirudin sample: Error > 30% at values ≥ 33 U Error > 16U at values < 33 U
CT-253	Platelet Testing	ASPItest		Aggregation (AUC) > 50 U in patients with a thromboxane receptor defect or Glanzman thrombasthenia confirmed by flow cytometry or genetic analysis AUC = Area Under Curve; For all other patients: Hirudin sample: Error > 30% at values ≥ 75 U Error > 20 U at values < 75 U Citrated sample: Error > 30% at values ≥ 40 U Error > 20 U at values < 40 U
CT-254	Platelet Testing	TRAPtest		Aggregation (AUC) > 50 U in patients with a Glanzman thrombasthenia confirmed by flow cytometry or genetic analysis AUC = Area Under Curve; For all other patients: Hirudin sample: Error > 30% at values ≥ 80 U Error > 24 U at values < 80 U Citrated sample: Error > 30% at values ≥ 80 U Error > 24U at values < 80 U
CT-1129	RBSS (ROW)	Anti-HBc		False negative. False positive. Discrepant results.
CT-1130	RBSS (ROW)	Anti-HBs		False negative. False positive. Discrepant results.
CT-1131	RBSS (ROW)	Anti-HCV		False negative. False positive. Discrepant results.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1132	RBSS (ROW)	Chagas		False negative. False positive. Discrepant results.
CT-1133	RBSS (ROW)	HBsAg		False negative. False positive. Discrepant results.
CT-1134	RBSS (ROW)	HBsAg Confirmatory Test		False negative. False positive. Discrepant results.
CT-1135	RBSS (ROW)	HIV combi PT		False negative. False positive. Discrepant results.
CT-1136	RBSS (ROW)	HIV duo		False negative MAIN result. Sub results from this test are not evaluated for PRI. False positive. Discrepant results.
CT-1137	RBSS (ROW)	HCV duo		False negative. False positive. Discrepant results.
CT-1138	RBSS (ROW)	HTLV		False negative. False positive. Discrepant results.
CT-1139	RBSS (ROW)	Syphilis		False negative. False positive. Discrepant results.
CT-1111	RBSS (US)	Anti-HBc		False negative. False positive. Discrepant results.
CT-1112	RBSS (US)	Anti-HCV		False negative. False positive. Discrepant results.
CT-1113	RBSS (US)	Chagas		False negative. False positive. Discrepant results.
CT-1114	RBSS (US)	HBsAg		False negative. False positive. Discrepant results.
CT-1118	RBSS (US)	HBsAg Confirmatory Test		False negative. False positive. Discrepant results.
CT-1115	RBSS (US)	HIV Duo		False negative MAIN result. Sub results from this test are not evaluated for PRI. False positive. Discrepant results.
CT-1116	RBSS (US)	HTLV		False negative. False positive. Discrepant results.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1117	RBSS (US)	Syphilis		False negative. False positive. Discrepant results.
CT-389	Research Use Only, Laboratory Use Only, Analyte Specific Reagents, Industrial Business or any additional non-In Vitro Diagnostic test commercially distributed by Roche Diagnostics	RUO, LUO, ASR, IB, etc.		Erroneous but believable results that led to Death, Serious Injury and/or Serious Harm
CT-390	RMD Tests not captured	RMD Tests not described		False Positive False Negative Miscall Over-quantitation if level is outside labeling claims Under-quantitation if level is outside labeling claims Invalid /delayed result when there is a delay in treatment that resulted in disease progression or further spread of disease
CT-1106	Roche in vitro Companion Diagnostics	Elecsys Anti-AAVrh74		False not elevated or false elevated result, OR False non-reactive or false reactive result OR Result incorrectly COI < 1.0 or result incorrectly >/= 1.0
CT-928	Sample Collection	AVENIO Millisect System Fill Station		Malfunction which already led or may potentially lead or contribute to death, serious injury and/or serious harm.
CT-929	Sample Collection	AVENIO Millisect System Software		Malfunction which already led or may potentially lead or contribute to death, serious injury and/or serious harm
CT-879	Specific Proteins	AAGP		Deviation of >/= 50%. Results incorrectly above normal range Results incorrectly below normal range Results incorrectly within normal range Incorrect increase of results (comparison measurements) within >/= 24 h by factor >/= 3 (verified by sample re-run). Incorrect clinical decision based on a reasonably incorrect AAGP result
CT-044	Specific Proteins	alpha-1-Acid Glycoprotein	X	
CT-045	Specific Proteins	alpha-1-Antitrypsin		Error > 30% at values > 2.0 g/L (36.8 µmol/L) Error > 0.3 g/L (5.5 µmol/L) at values < 1.1 g/L (16.6 µmol/L)
CT-926	Specific Proteins	Alpha2-Macroglobulin	X	
CT-819	Specific Proteins	Alpha 1-microglobulin	X	
CT-046	Specific Proteins	APO A1		Error > 0.3 g/L (10.7 µmol/L) at values < 1.0 g/L (35.7 µmol/L)
CT-047	Specific Proteins	APO B		Error > 30% at values > 1.2 g/l (2.34 µmol/L)
CT-1163	Specific Proteins	Apolipoprotein E4 Plasma, Elecsys		False negative results. Clinically false decisions based on result(s) of ApoE4 Plasma without actual patient harm.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-048	Specific Proteins	ASLO (applies to all tests for Antistreptolysin O, ASO, ASLOT, ASLO2)		Error >/= 30% at values >/= 100 IU/mL
CT-049	Specific Proteins	Beta-Amyloid (1-42) [= "Abeta 42"]		PRI ER Abeta 42 CSF: False negative results False positive results Clinically discrepant results or results discrepant to imaging Positron emission tomography (PET), PET-Scan <u>Standalone results:</u> please log PRI ER against the measured parameter <u>Ratio</u>: Elecsys β-Amyloid (1-42) CSF II shall be chosen to be the main product in case of allegation of an incorrect ratio (against β-Amyloid (1-40) CSF, pTau 181 CSF or tTau CSF).
CT-1162	Specific Proteins	Beta-Amyloid (1-40) CSF [= "Abeta 40"]		False negative results (ratio): falsely above/equal to the cut off ≥ 0.097 False positive results (ratio): falsely below the cut off < 0.097 Clinically discrepant results. Elecsys β-Amyloid (1-40) CSF assay is intended to be used in combination with the Elecsys β-Amyloid (1-42) CSF II assay as a ratio (no stand-alone product). Therefore only the results of the ratio apply for PRI ER. Important: Elecsys β-Amyloid (1-42) CSF II shall be chosen to be the main product in case of allegation of an incorrect ratio.
CT-144	Specific Proteins	Beta-2-Microglubulin		Error > 0.5 mg/L at values > 2.0 mg/L
CT-050	Specific Proteins	C3c		Error > 30% at values < 0.8 g/L
CT-051	Specific Proteins	C4		Error > 30% at values < 0.1 g/L
CT-820	Specific Proteins	Calprotectin	X	
CT-053	Specific Proteins	Ceruloplasmin		Error > 5 mg/dL (0.4 µmol/L) at values < 15 mg/dL (1.1 µmol/L)
CT-052	Specific Proteins	C-Peptide		Error > 0.3 ng/mL (0.10 nmol/L) at values < 1.1 ng/mL (0.37 nmol/L) Error > 30% at values > 4.4 ng/mL (1.47 nmol/L)
CT-054	Specific Proteins	CRP		Error > 30%
CT-056	Specific Proteins	CRPHS		Error > 30% at values > 1 mg/L
CT-887	Specific Proteins	HSCRP		Error > 30% at values ≥ 1 mg/L: applies to updated product under the name Tina-quant Cardiac high sensitivity CRP III
CT-055	Specific Proteins	Cystatin C		Error > 30% at values 0.8 - 2 mg/L
CT-057	Specific Proteins	Haptoglobin		Error > 0.1 g/L (1.0 µmol/L) at values ≤ 0.3 g/L (3.0 µmol/L)
CT-058	Specific Proteins	HbA1c		Error =/>0.5% (absolute) for values of 5.2 - 6.4%* Error =/> 1% (absolute) for values =/> 6.5% Error =/> 5 mmol/mol for values 33 - 47 mmol/mol* Error =/>10 mmol/mol for values =/> 48 mmol/mol * endpoints included. Regardless of the magnitude: Incorrect clinical decision or classification (e.g. inappropriate/missed therapy/diagnostic measures) due to an erroneous result and allegation not meeting PSI criteria.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-059	Specific Proteins	IgA		Error > 30%
CT-060	Specific Proteins	IgE		Error > 30% at values for - Neonates (i.e. newborn child up to 28 days) > 1.5 IU/mL (3.6 ng/mL) - Infants up to 1 yr > 15 IU/mL (36 ng/mL) - Children 1–5 years > 60 IU/mL (144 ng/mL) - Children > 6 years and Adults > 90 IU/mL (216 ng/mL)
CT-061	Specific Proteins	IgG		Error > 30%
CT-062	Specific Proteins	IgM		Error > 30%
CT-963	Specific Proteins	Kappa Free Light Chain		Error > 50% at values >20 mg/L. Only Diazyme results compared to Diazyme results qualify for PRI. Results from other manufacturers cannot be compared to Diazyme results. If the comparison method is not known, results should be evaluated for PRI.
CT-063	Specific Proteins	Kappa Light Chain	X	
CT-964	Specific Proteins	Lambda Free Light Chain		Error > 50% at values > 27 mg/L. Only Diazyme results compared to Diazyme results qualify for PRI. Results from other manufacturers cannot be compared to Diazyme results. If the comparison method is not known, results should be evaluated for PRI
CT-064	Specific Proteins	Lambda Light Chain	X	
CT-065	Specific Proteins	Lipoprotein (a)		Error > 50% at values > 75 nmol/L (300 mg/L)
CT-344	Specific Proteins	Lipoprotein A - LPACX (Companion Diagnostic only)		Error > 26.5% at values >75 nmol/L (300 mg/L)
CT-1177	Specific Proteins	NFL (Elecsys NfL /Neurofilament light chain)		- Deviation of results (relative) ≥ 40% for values > 1.5 pg/mL OR - Deviation of results (absolute) ≥ 0.4 pg/mL for values ≤ 1.5 pg/mL OR - Allegation of clinically discrepant test result, if confirmed by clinical manifestation, imaging with brain MRI, or other lab results. Validity of PRI ER criterion: - RUO use does not qualify for PRI ER. - IVD use qualifies for PRI ER criterion, as provided above. - Only results from comparison measurements within the same system/reagent/method combination for NfL qualify for PRI ER. Values determined in samples by different assay methods cannot be used interchangeably (Ref. Method Sheet 09880488500V1.0 English Elecsys NfL, 2025-11, V 1.0) - Percentage deviation of ≥ 40% applies for alleged -/+ deviation between results. - Absolute deviation ≥ 0.4 pg/mL applies for alleged -/+ deviation between results.
CT-815	Specific Proteins	NGAL		False low result < 250 ng/mL
CT-066	Specific Proteins	Percent CDT		Error > 50% at values >2.0% CDT Error > 30% at values >3.0% CDT
CT-916	Specific Proteins	Phospho-Tau (181) CSF		Phospho-Tau (181) CSF: False negative results False positive results Clinically discrepant results or results discrepant to imaging Positron emission tomography (PET), PET-Scan <u>Standalone results:</u> please log PRI ER against the measured parameter <u>Ratio</u>: Elecsys β-Amyloid (1-42) CSF II shall be chosen to be the main product in case of allegation of an incorrect ratio (against β-Amyloid (1-40) CSF, pTau 181 CSF or tTau CSF).

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1127	Specific Proteins	Phospho-Tau (181P) Plasma		Relative, numeric deviation >/= 30% False negative results (quantitative test results incorrectly below cut off, ≤ 0.934 pg/mL) False positive results (quantitative test results incorrectly above cut off, > 0.934 pg/mL) Incorrect clinical decision without actual harm. US only: Relative, numeric deviation >/= 30% False negative results (test results incorrectly below cut off, ≤ 0.722 pg/mL) False positive results (test results incorrectly above cut off, > 0.722 pg/mL) Incorrect clinical decision without actual harm.
CT-067	Specific Proteins	Prealbumin		Error > 5 mg/dL (0.91 µmol/L) at values < 20 mg/dL (3.64 µmol/L)
CT-068	Specific Proteins	Rheumatoid Factor		Error > 50% at values > 10 IU/mL
CT-069	Specific Proteins	Soluble Transferrin Receptor		Difference > 30% at values > 3.4 mg/L (52 nmol/L)
CT-917	Specific Proteins	Total-Tau CSF		Total-Tau CSF: False negative results False positive results Clinically discrepant results or results discrepant to imaging Positron emission tomography (PET), PET-Scan Standalone results: please log PRI ER against the measured parameter Ratio: Elecsys β-Amyloid (1-42) CSF II shall be chosen to be the main product in case of allegation of an incorrect ratio (against β-Amyloid (1-40) CSF, pTau 181 CSF or tTau CSF).
CT-070	Specific Proteins	Transferrin		Error > 0.5 g/L (6.3 µmol/L) at values < 2 g/L (25.2 µmol/L) Error > 30% at values > 3.6 g/L (45.4 µmol/L)
CT-836	Stago Products	Antiplasmin	X	
CT-837	Stago Products	APC resistance		Error > 25%
CT-838	Stago Products	aPTT		Error > 30%
CT-839	Stago Products	AT III		Error > 30% at values < 70%
CT-835	Stago Products	Beta-TG	X	
CT-840	Stago Products	D-Dimer		Error > 30% at values > 0.5 µg FEU/mL
CT-841	Stago Products	Factor II		Error > 25%
CT-842	Stago Products	Factor IV:AG	X	
CT-846	Stago Products	Factor IX		Error > 25%
CT-843	Stago Products	Factor V		Error > 25%
CT-844	Stago Products	Factor VII		Error > 25%
CT-845	Stago Products	Factor VIII		Error > 25%
CT-847	Stago Products	Factor X		Error > 25%
CT-848	Stago Products	Factor XI		Error > 25%
CT-849	Stago Products	Factor XII		Error > 25%
CT-850	Stago Products	Factor XIII		Error > 25%
CT-851	Stago Products	FDP TEST	X	
CT-852	Stago Products	Fibrinogen		Error > 30% at values < 200 mg/dL Error > 30% at values > 700 mg/dL

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-853	Stago Products	FM-Test	X	
CT-854	Stago Products	Hepato Quick		Error > 25%
CT-855	Stago Products	Neoplastin Plus		Error > 25%
CT-856	Stago Products	PAI	X	
CT-857	Stago Products	PF4	X	
CT-858	Stago Products	Plasminogen	X	
CT-859	Stago Products	Protein C		Error > 25%
CT-860	Stago Products	Protein S		Error > 25%
CT-861	Stago Products	PT		Error ≥ 0.5 INR for all values ≤ 2.0 INR Error > 30% for all values > 2.0 INR
CT-862	Stago Products	Reptilase	X	
CT-863	Stago Products	Thrombin Time		Error >30%
CT-864	Stago Products	t-PA	X	
CT-865	Stago Products	vWF (von Willebrand Factor)		Error > 25%
CT-002	Substrates	Albumin (CSF)		Error > 30% at values > 30 mg/dL
CT-001	Substrates	Albumin (serum, plasma)		Error > 0.5 g/dL at values < 3 g/dL Error > 30% at values > 5 g/dL
CT-003	Substrates	Ammonia		Error > 30% at values > 60 µg/dL (35.2 µmol/L)
CT-004	Substrates	Bicarbonate	X	
CT-005	Substrates	Bilirubin-Direct		Error > 0.3 mg/dL at values 0.4 - 1.0 mg/dL Error > 50% at values > 1.0 mg/dL
CT-006	Substrates	Bilirubin-Total		Error > 30% at values 0.9 - 3 mg/dL (15- 51 µmol/L) Error > 50% at values > 3 mg/dL (51 µmol/L) In term or near-term neonates:* Error > 30% at values < 6.0 mg/dL (100.0 µmol/L) Error > 20% at values ≥ 6.0 mg/dL (100.0 µmol/L) *Neonate is a child up to an age of 28 days.
CT-023	Substrates	BUN (Urea)		Error > 50% at values > 40 mg/dL (6.7 mmol/L)
CT-007	Substrates	Calcium		**Total Calcium (tCa)** Error > 20% at values < 2 mmol/L (8 mg/dL) Error > 20% at values > 2.5 mmol/L (10 mg/dL) **Ca2+** Error > 20% at values < 1.14 mmol/L Error > 20% at values > 1.35 mmol/L
CT-008	Substrates	Cholesterol		Error > 50% at values 150 – 300 mg/dL (3.9 – 7.8 mmol/L)
CT-009	Substrates	Creatinine		Error > 20% at values 0.9 - 1.1 mg/dL (79.6 - 97.2 µmol/L) Error > 30% at values > 1.1 mg/dL (97.2 µmol/L)
CT-010	Substrates	Fructosamine		Error > 30% at values > 200 µmol/L

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-011	Substrates	Glucose		Error > 15 mg/dL (0.8 mmol/L) at values ≤ 50 mg/dL (2.8 mmol/L) Error > 30% at values > 50 mg/dL (2.8 mmol/L)
CT-012	Substrates	HDL Cholesterol Direct		Error > 50% at values 35 - 65 mg/dL (0.9 – 1.7 mmol/L)
CT-013	Substrates	Homocysteine		Error > 30% at values > 10 µmol/L
CT-014	Substrates	Iron		Error > 50% at values < 40 µg/dL (7.2 µmol/L) Error > 50% at values > 160 µg/dL (28.7 µmol/L)
CT-015	Substrates	Lactate		Error > 30% at values < 4.5 mg/dL (0.5 mmol/L) Error > 30% at values > 15 mg/dL (1.7 mmol/L)
CT-016	Substrates	LDL Cholesterol Direct		Error > 50% at values 100 - 200 mg/dL (2.6 – 5.2 mmol/L)
CT-017	Substrates	Magnesium		Error > 0,25 mmol/L (0.6 mg/dL) at values ≤ 1 mmol/L (2.4 mg/dL) Error > 50% at values > 1 mmol/L (2.4 mg/dL)
CT-018	Substrates	Phosphorus		Error > 30% at values < 2.6 mg/dL (0.8 mmol/L) Error > 30% at values > 4.5 mg/dL (1.5 mmol/L)
CT-827	Substrates	sdLDL		False low and false high result with regard to cut-off 50 mg/dL
CT-021	Substrates	Total Protein (CSF)		Error > 50% at values > 300 mg/L
CT-020	Substrates	Total Protein (Serum/Plasma)		Error > 1 g/dL (10 g/L)
CT-019	Substrates	Triglycerides		Error > 30% at values > 200 mg/dL (2.3 mmol/L)
CT-022	Substrates	UIBC	X	
CT-024	Substrates	Uric Acid		Error > 30% at values < 3 mg/dL (180 µmol/L) Error > 30% at values > 5 mg/dL (297 µmol/L)
CT-071	Therapeutic Drug Monitoring	Acetaminophen		Error > 30% at values > 30 µg/mL (199 µmol/mL)
CT-072	Therapeutic Drug Monitoring	Amikacin		Error > 20% at values > 30 µg/mL (51.8 µmol/L) Error > 20% at values < 10 µg/mL (8.6 µmol/L)
CT-073	Therapeutic Drug Monitoring	Carbamazepine		Error > 20% at values >10 µg/mL (42.5 µmol/L Error > 20% at values < 4 µg/mL (17 µmol/L)
CT-074	Therapeutic Drug Monitoring	Cyclosporine		Error > 20% at values > 300 ng/mL (250 nmol/L) Error > 30% at values < 100 ng/mL (83 nmol/L)
CT-1165	Therapeutic Drug Monitoring	DABI	X	Partner Channel Technoclone
CT-075	Therapeutic Drug Monitoring	Digitoxin		Error >30% at values > 30 ng/mL (39.3 nmol/L) Error > 30% at values < 10 ng/mL (13.1 nmol/L)
CT-076	Therapeutic Drug Monitoring	Digoxin		Error > 30% at values > 1.2 ng/mL (1.5 nmol/L) Error > 30% at values < 0.6 ng/mL (0.8 nmol/L)
CT-077	Therapeutic Drug Monitoring	Everolimus		Error > 30% at values > 8 ng/mL (8.4 nmol/L) Error > 30% at values < 3 ng/mL (3.1 nmol/L)
CT-825	Therapeutic Drug Monitoring	Free Phenytoin		Error > 20% at values > 2.5 µg/mL (9.9 µmol/L) Error > 30% at values < 1.2 µg/mL (4.0 µmol/L) For children: Error > 20% at values > 0.5 µg/mL (2 µmol/L)

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-094	Therapeutic Drug Monitoring	Free Valproic Acid		Error > 20% at values > 15 µg/mL Error > 20% at values < 5 µg/mL Free valproic acid > total valproic acid
CT-078	Therapeutic Drug Monitoring	Gentamicin		Error > 20% at values > 10 µg/mL (21 µmol/L) Error > 1µg/mL (2.1 µmol/L) at values < 3 µg/mL (4.2 µmol/L)
CT-079	Therapeutic Drug Monitoring	Lidocaine		Error > 20% at values > 6 µg/mL Error > 20% at values < 1.5 µg/mL
CT-080	Therapeutic Drug Monitoring	Lithium		Error > 20% at values > 1,2 mmol/L Error > 20% at values < 0,9 mmol/L
CT-081	Therapeutic Drug Monitoring	Mycophenolic Acid (TMPA)		Error > 30% at values > 3,5 µg/mL (10.9 µmol/L) Error > 0.3 µg/mL (0.9 µmol/L) at values < 1.0 µg/mL (3.1 µmol/L)
CT-082	Therapeutic Drug Monitoring	NAPA		Error > 20% at values > 20 µg/mL (72 µmol/L)
CT-083	Therapeutic Drug Monitoring	Phenobarbital		Error > 20% at values > 30 µg/ml (129.3 µmol/L)
CT-084	Therapeutic Drug Monitoring	Phenytoin		Error > 20% at values > 20 µg/mL (79.2 µmol/L) Error > 30% at values < 10 µg/mL (39.6 µmol/L) For newborns: Error > 20% at values > 14 µg/mL (55.4 µmol/L) Error >30% at values < 6 µg/mL (23.8 µmol/L)
CT-085	Therapeutic Drug Monitoring	Primidone		Error > 20% at values > 12 µg/mL (55 µmol/L) Error > 20% at values < 5 µg/mL (23 µmol/L)
CT-086	Therapeutic Drug Monitoring	Procainamide		Error > 20% at values > 10 µg/mL (42.3 µmol/L) Error > 20% at values < 4 µg/mL (16.9 µmol/L)
CT-087	Therapeutic Drug Monitoring	Quinidine		Error > 20% at values > 5 µg/mL (15 µmol/L) Error > 20% at values < 2 µg/mL (6 µmol/L)
CT-088	Therapeutic Drug Monitoring	Salicylate		Error > 30% at values >200 µg/mL(1.45 mmol/L)
CT-089	Therapeutic Drug Monitoring	Sirolimus		Error > 30% at values > 20 ng/mL (21.9 nmol/L) Error > 30% at values < 5 ng/mL (5.5 nmol/L)
CT-090	Therapeutic Drug Monitoring	Tacrolimus		Error > 30% at values > 9 ng/mL (11.7 nmol/L) Error > 30% at values < 4 ng/mL (5.2 nmol/L)
CT-091	Therapeutic Drug Monitoring	Theophylline		Error > 20% at values > 20 µg/mL (111 µmol/L) Error > 20% at values < 6 µg/mL (33.3 µmol/L) Babies (up to 6 months including prematures [Babies born before 37 weeks of gestation]): Error > 20% at values > 6 µg/mL (33.3 µmol/L)
CT-092	Therapeutic Drug Monitoring	Tobramycin		Error > 20% at values > 8 µg/mL (17 µmol/L) Error > 1.0 µg/mL (2.3 µmol/L) at values < 3 µg/mL (6.3 µmol/L)
CT-093	Therapeutic Drug Monitoring	Valproic Acid		Error > 20% at values > 100 µg/mL (693 µmol/L) Error > 20% at values < 50 µg/mL (346.5 µmol/L) Free Valproic acid > Valproic acid
CT-095	Therapeutic Drug Monitoring	Vancomycine		Error > 20% at values > 30 µg/mL (20.6 µmol/L) Error > 2 µg/mL (1.4 µmol/L) at values < 10 µg/mL (7.0 µmol/L)
CT-882	Therapeutic Drug Monitoring	MTX (ONLINE TDM Methotrexate)		Error >/= 30% Incorrect clinical decision (e.g. inappropriate therapy) due to Erroneous Result

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1144	Therapeutic Drug Monitoring_i3MV_MS	Mycophenolic acid (iMPA)		Deviation >/= 0.230 µg/mL for values >/= 1.00 µg/mL and </= 1.5 µg/mL Deviation of results >22% for values > 1.5 µg/mL and for the whole measuring range
CT-1145	Therapeutic Drug Monitoring_i3MV_MS	Mycophenolic acid glucuronide (iMPAG)		Deviation of results >25% for values >/= 35 µg/mL
CT-1146	Therapeutic Drug Monitoring_i3MV_MS	Voriconazole (iVORI)		Deviation of results > 20% for values >/= 0.5 µg/mL and < 7.2 µg/mL
CT-1140	Therapeutic Drug Monitoring_iAB1_MS	Linezolid (iLINE)		Deviation of results >/= 24% for values >/= 1 µg/mL and throughout the measuring range
CT-1141	Therapeutic Drug Monitoring_iAB1_MS	Vancomycin (iVANC)		Deviation of results > 20% for values >/= 10 µg/mL and throughout the whole measuring range, including results for diluted samples.
CT-1159	Therapeutic Drug Monitoring_iAB2_MS	Meropenem (iMERO)		Deviation of results >/= 25% for values >/= 2.50 µg/mL for the whole measuring range
CT-1160	Therapeutic Drug Monitoring_iAB2_MS	Piperacillin (iPIPE)		Deviation of results >/= 25% for values >/= 8.00 µg/mL throughout the whole measuring range, including diluted samples.
CT-1147	Therapeutic Drug Monitoring_iAED1_MS	Phenytoin (iPHNY)		Deviation of results >/= 20% for values >/= 5.00 µg/mL and </= 40 µg/mL
CT-1148	Therapeutic Drug Monitoring_iAED1_MS	Phenytoin, free (iFPHY)		Deviation of results >/= 25% for values >/= 0.500µg/mL and </= 4.00µg/mL Free Phenytoin value > Phenytoin value ((FPHN > PHNY): measurements performed with the same system/reagent/method combination and from the same patient sample draw.
CT-1149	Therapeutic Drug Monitoring_iAED1_MS	Valproic Acid (iVPA)		Deviation of results >/= 20% for values >/= 40.00 µg/mL and </=120 µg/mL
CT-1150	Therapeutic Drug Monitoring_iAED1_MS	Valproic Acid, Free (iFVPA)		Absolute Deviation of results >/= 2.00 µg/mL for values >/= 4.00 µg/mL and </= 10.00 µg/mL Relative Deviation of results >/=25% for values > 10 µg/mL including results from diluted samples. Validity: all comparison values above 10 µg/mL Free Valproic Acid value > Valproic Acid value ((FVPA > VPA): measurements performed with the same system/reagent/method combination and from the same patient sample draw.
CT-1119	Therapeutic Drug Monitoring_iAED2_MS	Carbamazepine (iCARB)		Deviation of results > 20% for values > 4.00 µg/mL Deviation > 0.500 µg/mL for values >/= 1.00 µg/mL and < 4.00 µg/mL Carbamazepine value < Carbamazepine-10,11-epoxide value (CARB < CAEO) Validity of Criterion CARB < CAEO: measurements performed with the same system/reagent/method combination and from the same patient sample draw.
CT-1120	Therapeutic Drug Monitoring_iAED2_MS	Carbamazepine-10,11-epoxide (iCAEO)		Deviation of results > 20% for values >/= 0.400 µg/mL Carbamazepine-10,11-epoxide value > Carbamazepine (CAEO > CARB) Validity of Criterion CAEO > CARB: measurements performed with the same system/reagent/method combination and from the same patient sample draw.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1121	Therapeutic Drug Monitoring_iAED2_MS	Gabapentin (iGABA)		Deviation of results > 20% for values >/= 2.00 µg/mL
CT-1122	Therapeutic Drug Monitoring_iAED2_MS	Lamotrigine (iLAMO)		Deviation of results > 20% for values >/= 2.500 µg/mL Deviation of results > 0.400µg/mL for values < 2.500 µg/mL
CT-1123	Therapeutic Drug Monitoring_iAED2_MS	Phenobarbital (iPHNO)		Deviation of results > 20% for values >/= 7 µg/mL
CT-1124	Therapeutic Drug Monitoring_iAED2_MS	Primidone (iPRIM)		Deviation of results >20% for values >/= 4 µg/mL
CT-1125	Therapeutic Drug Monitoring_iAED2_MS	Topiramate (iTOPI)		Deviation of results >20% for values >/= 7.50 µg/mL Deviation > 0.750 µg/mL for values > 2.00 µg/mL and < 7.5 µg/mL
CT-1126	Therapeutic Drug Monitoring_iAED2_MS	Zonisamide (iZONI)		Deviation of results >20% for values >/= 8.00 µg/mL
CT-1101	Therapeutic Drug Monitoring_iISD_MS	Cyclospoprine		Deviation >/= 30% for values >/= 25 ng/mL including comparisons between diluted samples.
CT-1102	Therapeutic Drug Monitoring_iISD_MS	Everolimus		Deviation of results >/= 30% for values >/= 3.00 ng/mL, including comparisons between diluted samples.
CT-1103	Therapeutic Drug Monitoring_iISD_MS	Sirolimus		Deviation of results >/= 30% for values >/= 3.00 ng/mL including comparisons between diluted samples.
CT-1104	Therapeutic Drug Monitoring_iISD_MS	Tacrolimus		Deviation of results >/= 30% for values >/= 3.00 ng/mL including comparisons between diluted samples.
CT-111	Thyroid	Anti-TG		Error > 40% at values 25 - 50 IU/mL Errror > 30% at values 50 - 250 IU/mL
CT-112	Thyroid	Anti-TPO		Error > 50% at values > 35 IU/mL** **Due to the various standardization methods available, different assay setups and compositions like usage of different antibodies, it is possible to receive different results when using another manufacturer's test. Although the results are different, these could lead to the same clinical interpretation. It is essential to request the normal ranges in place when assessing for PRI and ask if the different results give different clinical interpretation. If both results are in the same range compared to the reference values (below - within - above), it is considered that the results are presenting the same clinical interpretation. In this situation PRI escalation of the case is not required. A PRI is not required when differing results lead to the same clinical interpretation. This applies when a Roche result is compared to Roche result or Roche compared to another manufacturer. If this information about clinical interpretation cannot be provided by the customer, evaluation for PRI is required.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1154	Thyroid	Elecsys Anti-TPO II		- Pregnant women: Deviation of results >/= 30% for values >/= 20 IU/mL - All other patient groups: Deviation of results >/= 50% for values >/= 20 IU/mL Validity: PRI Criterion is meant to be implemented for alleged - or + deviation of >/= 30% or >/= 50% Please note: Method Sheet: 08858454500V2.0 2025-08, V 2.0 English “The measured anti-TPO value of a patient’s sample can vary depending on the assay used. The laboratory finding must therefore always contain a statement on the anti-TPO assay method used. Due to methodological differences, such as the different epitopes targeted by the detection antibodies used in the Elecsys Anti-TPO II assay and other assays, numeric anti-TPO test results determined on patient samples by different assays cannot be directly compared with one another and could be the cause of erroneous medical interpretations.”
CT-113	Thyroid	Anti-TSHR		Error > 30% at values > 1.8 IU/L
CT-114	Thyroid	Calcitonin		Error > 30% at values > 6 pg/mL for female Error > 30% at values > 10 pg/mL for male
CT-115	Thyroid	FT3		Error > 20% and/or FT3 > T3** **Due to the various standardization methods available, different assay setups and compositions like usage of different antibodies, it is possible to receive different results when using another manufacturer's test. Although the results are different, these could lead to the same clinical interpretation. It is essential to request the normal ranges in place when assessing for PRI and ask if the different results give different clinical interpretation. A PRI is not required when differing results lead to the same clinical interpretation. This applies when a Roche result is compared to Roche result or Roche compared to another manufacturer. If this information about clinical interpretation cannot be provided by the customer, evaluation for PRI is required.
CT-116	Thyroid	FT4		Error > 20% and/or FT4 > T4** **Due to the various standardization methods available, different assay setups and compositions like usage of different antibodies, it is possible to receive different results when using another manufacturer's test. Although the results are different, these could lead to the same clinical interpretation. It is essential to request the normal ranges in place when assessing for PRI and ask if the different results give different clinical interpretation. If both results are in the same range compared to the reference values (below - within - above), it is considered that the results are presenting the same clinical interpretation. In this situation PRI escalation of the case is not required. A PRI is not required when differing results lead to the same clinical interpretation. This applies when a Roche result is compared to Roche result or Roche compared to another manufacturer. If this information about clinical interpretation cannot be provided by the customer, evaluation for PRI is required.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-117	Thyroid	T3		Error > 50%** **Due to the various standardization methods available, different assay setups and compositions like usage of different antibodies, it is possible to receive different results when using another manufacturer's test. Although the results are different, these could lead to the same clinical interpretation. It is essential to request the normal ranges in place when assessing for PRI and ask if the different results give different clinical interpretation. If both results are in the same range compared to the reference values (below - within - above), it is considered that the results are presenting the same clinical interpretation. In this situation PRI escalation of the case is not required. A PRI is not required when differing results lead to the same clinical interpretation. This applies when a Roche result is compared to Roche result or Roche compared to another manufacturer. If this information about clinical interpretation cannot be provided by the customer, evaluation for PRI is required.
CT-118	Thyroid	T4		Error > 30%** **Due to the various standardization methods available, different assay setups and compositions like usage of different antibodies, it is possible to receive different results when using another manufacturer's test. Although the results are different, these could lead to the same clinical interpretation. It is essential to request the normal ranges in place when assessing for PRI and ask if the different results give different clinical interpretation. If both results are in the same range compared to the reference values (below - within - above), it is considered that the results are presenting the same clinical interpretation. In this situation PRI escalation of the case is not required. A PRI is not required when differing results lead to the same clinical interpretation. This applies when a Roche result is compared to Roche result or Roche compared to another manufacturer. If this information about clinical interpretation cannot be provided by the customer, evaluation for PRI is required.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-119	Thyroid	TG		Error > 0.5 ng/mL at values > 0.4 and < 2 ng/mL** **Due to the various standardization methods available, different assay setups and compositions like usage of different antibodies, it is possible to receive different results when using another manufacturer's test. Although the results are different, these could lead to the same clinical interpretation. It is essential to request the normal ranges in place when assessing for PRI and ask if the different results give different clinical interpretation. If both results are in the same range compared to the reference values (below - within - above), it is considered that the results are presenting the same clinical interpretation. In this situation PRI escalation of the case is not required. A PRI is not required when differing results lead to the same clinical interpretation. This applies when a Roche result is compared to Roche result or Roche compared to another manufacturer. If this information about clinical interpretation cannot be provided by the customer, evaluation for PRI is required.
CT-120	Thyroid	TSH		Error > 20% ** **Due to the various standardization methods available, different assay setups and compositions like usage of different antibodies, it is possible to receive different results when using another manufacturer's test. Although the results are different, these could lead to the same clinical interpretation. It is essential to request the normal ranges in place when assessing for PRI and ask if the different results give different clinical interpretation. If both results are in the same range compared to the reference values (below - within - above), it is considered that the results are presenting the same clinical interpretation. In this situation PRI escalation of the case is not required. A PRI is not required when differing results lead to the same clinical interpretation. This applies when a Roche result is compared to Roche result or Roche compared to another manufacturer. If this information about clinical interpretation cannot be provided by the customer, evaluation for PRI is required.

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-121	Thyroid	T-Uptake (TBG, TBK)		Error > 50%** **Due to the various standardization methods available, different assay setups and compositions like usage of different antibodies, it is possible to receive different results when using another manufacturer's test. Although the results are different, these could lead to the same clinical interpretation. It is essential to request the normal ranges in place when assessing for PRI and ask if the different results give different clinical interpretation. If both results are in the same range compared to the reference values (below - within - above), it is considered that the results are presenting the same clinical interpretation. In this situation PRI escalation of the case is not required. A PRI is not required when differing results lead to the same clinical interpretation. This applies when a Roche result is compared to Roche result or Roche compared to another manufacturer. If this information about clinical interpretation cannot be provided by the customer, evaluation for PRI is required.
CT-891	TIB Molbiol	LightMix® AAT (PiS, PiZ)		False genotyping result (peaks in specific position)
CT-1026	TIB Molbiol	LightMix® ACE I/D		False genotyping result (peaks in specific position)
CT-1027	TIB Molbiol	LightMix® ApoB-100 R3500Q		False genotyping result (peaks in specific position)
CT-1028	TIB Molbiol	LightMix® Factor II Prothrombin		False genotyping result (peaks in specific position)
CT-1029	TIB Molbiol	LightMix® Factor V Leiden		False genotyping result (peaks in specific position)
CT-1030	TIB Molbiol	LightMix® Factor XIII		False genotyping result (peaks in specific position)
CT-1031	TIB Molbiol	LightMix® FGB [G-455A]		False genotyping result (peaks in specific position)
CT-1032	TIB Molbiol	LightMix® HFE H63D S65C C282Y		False genotyping result (peaks in specific position)
CT-1033	TIB Molbiol	LightMix® Lactose Intolerance		False genotyping result (peaks in specific position)
CT-1034	TIB Molbiol	LightMix® MTHFR A1298C		False genotyping result (peaks in specific position)
CT-1035	TIB Molbiol	LightMix® MTHFR C677T		False genotyping result (peaks in specific position)
CT-1036	TIB Molbiol	LightMix® Multi-SNiP DPYD		False genotyping result (peaks in specific position)

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1166	TIB Molbiol	LightMix Newborn (TREC, SMN1 or SMA, HBB)		Newborn screening for SCID: test TREC False negative result for screen SCID mutation False reactive result of TREC test False presence of detectable TRECs in sample Newborn screening for SMA: test SMN1 False negative result for screen SMA mutations False reactive result of SMN1 test (Signal for SMN1 present) Newborn screening for SCD: test HBB False negative result for screen SCD / HBB False One peak Tm ~67 °C of HBB test False homozygous HBB (wild type) result IMPORTANT note: The term “false negative result” is associated with the assessment of the screening result for mutation. A negative screen result means the absence of mutation in the survival motor neuron 1 (SMN1) gene, absence of profound impairment in T-cell development, or absence of hemoglobin mutations The term “reactive” is accordingly used to point to the direct test result, based on which the screening is concluded.
CT-1037	TIB Molbiol	LightMix® PAI-4G/5G		False genotyping result (peaks in specific position)
CT-1038	TIB Molbiol	LightMix® VKORC1		False genotyping result (peaks in specific position)
CT-892	TIB Molbiol	LightMix® TREC SMA HBB Newborn		False positive result, false negative result, false genotyping result (peaks in specific position)
CT-893	TIB Molbiol	LightMix® Modular GES (640)		False positive result, false negative result
CT-1039	TIB Molbiol	LightMix® Modular IMP (640)		False positive result, false negative result
CT-1040	TIB Molbiol	LightMix® Modular KPC (610)		False positive result, false negative result
CT-1041	TIB Molbiol	LightMix® Modular MCR-1 (640)		False positive result, false negative result
CT-1042	TIB Molbiol	LightMix® Modular NDM (530)		False positive result, false negative result
CT-1043	TIB Molbiol	LightMix® Modular OXA-23 (580)		False positive result, false negative result
CT-1044	TIB Molbiol	LightMix® Modular OXA-48 (580)		False positive result, false negative result
CT-1045	TIB Molbiol	LightMix® Modular VIM (500)		False positive result, false negative result
CT-894	TIB Molbiol	LightMix® Modular Adenovirus		False positive result, false negative result
CT-1046	TIB Molbiol	LightMix® Modular Astrovirus		False positive result, false negative result
CT-1047	TIB Molbiol	LightMix® Modular Norovirus GG1		False positive result, false negative result
CT-1048	TIB Molbiol	LightMix® Modular Norovirus GG1 (FAM)		False positive result, false negative result
CT-1049	TIB Molbiol	LightMix® Modular Norovirus GG2		False positive result, false negative result
CT-1050	TIB Molbiol	LightMix® Modular Rotavirus		False positive result, false negative result
CT-895	TIB Molbiol	LightMix® Modular Aeromonas		False positive result, false negative result
CT-1051	TIB Molbiol	LightMix® Modular Campylobacter		False positive result, false negative result
CT-1052	TIB Molbiol	LightMix® Modular Gastro panel Bacterial v 2.0		False positive result, false negative result
CT-1053	TIB Molbiol	LightMix® Modular Salmonella		False positive result, false negative result
CT-1054	TIB Molbiol	LightMix® Modular Shigella		False positive result, false negative result

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1055	TIB Molbiol	LightMix® Modular Yersinia		False positive result, false negative result
CT-896	TIB Molbiol	LightMix® Modular Blastocystis		False positive result, false negative result
CT-1056	TIB Molbiol	LightMix® Modular Cryptosporidium		False positive result, false negative result
CT-1057	TIB Molbiol	LightMix® Modular Dientamoeba		False positive result, false negative result
CT-1058	TIB Molbiol	LightMix® Modular Entamoeba histolytica		False positive result, false negative result
CT-1059	TIB Molbiol	LightMix® Modular Gastro panel Parasites v 2.0		False positive result, false negative result
CT-1060	TIB Molbiol	LightMix® Modular Giardia		False positive result, false negative result
CT-897	TIB Molbiol	LightMix® Modular Adenovirus (610)		False positive result, false negative result
CT-1061	TIB Molbiol	LightMix® Modular Bocavirus (530)		False positive result, false negative result
CT-1062	TIB Molbiol	LightMix® Modular Bocavirus (580)		False positive result, false negative result
CT-1063	TIB Molbiol	LightMix® Modular Enterovirus (500)		False positive result, false negative result
CT-1064	TIB Molbiol	LightMix® Modular Enterovirus (640)		False positive result, false negative result
CT-1065	TIB Molbiol	LightMix® Modular Metapneumovirus (hMPV) (580)		False positive result, false negative result
CT-1066	TIB Molbiol	LightMix® Modular Parechovirus (530)		False positive result, false negative result
CT-1067	TIB Molbiol	LightMix® Modular Parechovirus (640)		False positive result, false negative result
CT-1068	TIB Molbiol	LightMix® Modular Picornavirus HRV + EV (640)		False positive result, false negative result
CT-1069	TIB Molbiol	LightMix® Modular RSV A/B		False positive result, false negative result
CT-898	TIB Molbiol	LightMix® Modular Influenza A		False positive result, false negative result
CT-1070	TIB Molbiol	LightMix® Modular Influenza A H1 (H1N1)		False positive result, false negative result
CT-1071	TIB Molbiol	LightMix® Modular Influenza B		False positive result, false negative result
CT-899	TIB Molbiol	LightMix® Modular ParaInfluenza 1 (640)		False positive result, false negative result
CT-1072	TIB Molbiol	LightMix® Modular ParaInfluenza 2 (610)		False positive result, false negative result
CT-1073	TIB Molbiol	LightMix® Modular ParaInfluenza 3 (530)		False positive result, false negative result
CT-1074	TIB Molbiol	LightMix® Modular ParaInfluenza 4 (500)		False positive result, false negative result
CT-901	TIB Molbiol	LightMix® MERS-CoV RTPCR Kit		False positive result, false negative result
CT-1075	TIB Molbiol	LightMix® SARS-CoV-2 E+N UBC		False positive result, false negative result
CT-1076	TIB Molbiol	LightMix® Modular Sarbecovirus SARS-CoV-2		False positive result, false negative result
CT-902	TIB Molbiol	LightMix® Ureaplasmaurealyticum/parvum		False positive result, false negative result (also for melting)
CT-903	TIB Molbiol	LightMix® Modular Trichomonas vaginalis		False positive result, false negative result
CT-904	TIB Molbiol	LightMix® Modular Chikungunya Virus		False positive result, false negative result
CT-1077	TIB Molbiol	LightMix® Modular Dengue Virus		False positive result, false negative result
CT-1078	TIB Molbiol	LightMix® Modular Zika Virus (530)		False positive result, false negative result

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1079	TIB Molbiol	LightMix® Modular Zika Virus (640)		False positive result, false negative result
CT-911	TIB Molbiol	LightMix® Polyomavirus JC BK		False positive result, false negative result (also for melting)
CT-912	TIB Molbiol	UC-TIB-AdV		False positive result, false negative result
CT-913	TIB Molbiol	UC-TIB-HSV/VZV		False positive result, false negative result
CT-143	Tumor Markers / Oncology	AFP		Error > 30%. Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI
CT-989	Tumor Markers / Oncology	AFP-L3	X	
CT-958	Tumor Markers / Oncology	Anti-p53		False positive result at values > 0.10 µg/mL
CT-146	Tumor Markers / Oncology	CA 125		Error > 30% at values > 25 U/mL. Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI
CT-147	Tumor Markers / Oncology	CA 15-3		Error > 30% at values > 25 U/mL. Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI
CT-148	Tumor Markers / Oncology	CA 19-9		Error > 30% at values > 25 U/mL. Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI
CT-149	Tumor Markers / Oncology	CA 72-4		Error > 30% at values > 6 U/mL
CT-980	Tumor Markers / Oncology	CA 242		Error > 30% at values > 20 U/mL. Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI
CT-150	Tumor Markers / Oncology	Calcitonin		Error > 30% at values > 6 pg/mL for female Error > 30% at values > 10 pg/mL for male
CT-151	Tumor Markers / Oncology	CEA		Error > 50%. Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI
CT-152	Tumor Markers / Oncology	Cyfra 21-1		Error > 50% at values > 3 ng/mL
CT-153	Tumor Markers / Oncology	Ferritin		Error > 50%
CT-986	Tumor Markers / Oncology	GAAD / GALAD	X	
CT-145	Tumor Markers / Oncology	HCG + beta		Error > 1.5 mIU/mL at values ≤ 5 mIU/mL Error > 30% at values > 5 mIU/mL
CT-154	Tumor Markers / Oncology	HE4		Error > 30% at values > 40 pmol/L. Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI
CT-155	Tumor Markers / Oncology	NSE		Error > 50% at values > 17 ng/mL
CT-945	Tumor Markers / Oncology	PIVKA		Error > 30% at values > 20 ng/mL
CT-156	Tumor Markers / Oncology	ProGRP		Error > 30% at values > 80 pg/mL
CT-157	Tumor Markers / Oncology	PSA Free		Error > + 30% at values > 0.3 ng/mL and/or PSA Free > PSA Total

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-158	Tumor Markers / Oncology	PSA Total		Error > 30% at values > 3.0 ng/mL and/or PSA Free > PSA Total Error > 0.1 ng/mL at values < 0.2 ng/mL after confirmed prostatectomy or radiotherapy ** If values are <0.2 ng/mL and it is not known if prostatectomy or radiotherapy occurred, PRI evaluation is not required. Only Roche results compared to Roche results qualify for PRI. Results from other manufacturers cannot be compared to Roche results. If the comparison method is not known, results should be evaluated for PRI.
CT-159	Tumor Markers / Oncology	S 100		Error > 50% at values > 0.1 µg/L
CT-160	Tumor Markers / Oncology	SCC		Error > 30%
CT-223	Urine Testing - Clin. Chem. Analyzer	alpha-1-Microglobulin	X	
CT-927	Urine Testing - Clin. Chem. Analyzer	Alpha2-macroglobulin	X	
CT-226	Urine Testing - Clin. Chem. Analyzer	AMYL (total)	X	
CT-225	Urine Testing - Clin. Chem. Analyzer	AMY-P	X	
CT-227	Urine Testing - Clin. Chem. Analyzer	BUN (Urea)		Error > 50% at values > 800 mg/dL (134 mmol/L)
CT-946	Urine Testing - Clin. Chem. Analyzer	beta-2 Microglobulin	X	
CT-228	Urine Testing - Clin. Chem. Analyzer	Calcium	X	
CT-229	Urine Testing - Clin. Chem. Analyzer	Creatinine		Error > 30%
CT-990	Urine Testing - Clin. Chem. Analyzer	Crystals	X	
CT-230	Urine Testing - Clin. Chem. Analyzer	ETOH	X	
CT-231	Urine Testing - Clin. Chem. Analyzer	Glucose		Error > 30% at values > 20 mg/dL (1.1 mmol/L)
CT-991	Urine Testing - Clin. Chem. Analyzer	Hyaline casts (HYA)	X	
CT-234	Urine Testing - Clin. Chem. Analyzer	IgG		Error > 30%
CT-236	Urine Testing - Clin. Chem. Analyzer	Magnesium	X	
CT-235	Urine Testing - Clin. Chem. Analyzer	Microalbumin		Error > 30% at values > 20 mg/L
CT-816	Urine Testing - Clin. Chem. Analyzer	NGAL		False low result < 250 ng/mL

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-992	Urine Testing - Clin. Chem. Analyzer	Non-Squamous epithelial cells (NEC)	X	
CT-993	Urine Testing - Clin. Chem. Analyzer	Pathological casts (PAT)	X	
CT-237	Urine Testing - Clin. Chem. Analyzer	Phosphorus	X	
CT-994	Urine Testing - Clin. Chem. Analyzer	Sperm	X	
CT-995	Urine Testing - Clin. Chem. Analyzer	Squamous epithelial cells (SEC)	X	
CT-238	Urine Testing - Clin. Chem. Analyzer	Total Protein (Urine)		Error > 50%
CT-947	Urine Testing - Clin. Chem. Analyzer	TRSF2	X	
CT-239	Urine Testing - Clin. Chem. Analyzer	Uric Acid	X	
CT-240	Urine Testing - ISE Urine	ISE - Chloride	X	
CT-241	Urine Testing - ISE Urine	ISE - Potassium	X	
CT-242	Urine Testing - ISE Urine	ISE - Sodium	X	
CT-909	Urine Testing - Urine Strip Testing	Bacteria		False negative and false positive result
CT-212	Urine Testing - Urine Strip Testing	Bilirubin	X	
CT-213	Urine Testing - Urine Strip Testing	Erythrocytes		False negative result confirmed to be 50 Ery/µL or greater (more than two color blocks) **Note:** Criteria is related to particle concentration of native urine and not the sediment Native results to sediment results are not evaluated for PRI
CT-214	Urine Testing - Urine Strip Testing	Glucose		False negative result confirmed to be 250 mg/dL or greater (more than two color blocks)
CT-215	Urine Testing - Urine Strip Testing	Ketones		False negative result confirmed to be 50 mg/dL or greater (more than two color blocks)
CT-216	Urine Testing - Urine Strip Testing	Leucocytes		False negative result: sample contains > 100 leucocytes/µL **Note:** Criteria is related to particle concentration of native urine and not the sediment Native results to sediment results are not evaluated for PRI
CT-217	Urine Testing - Urine Strip Testing	Microalbumin (Micral)		False low result confirmed to be more than one color block in the range 50 to 100 mg/L
CT-981	Urine Testing - Urine Strip Testing	Mucus		False negative
CT-218	Urine Testing - Urine Strip Testing	Nitrite	X	
CT-219	Urine Testing - Urine Strip Testing	pH		Error > 2 color blocks
CT-220	Urine Testing - Urine Strip Testing	Protein		False negative result confirmed to be 100 mg/dL or greater (more than two color blocks)

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-221	Urine Testing - Urine Strip Testing	Specific Gravity	X	
CT-222	Urine Testing - Urine Strip Testing	Urobilinogen	X	
CT-918	Urine Testing - Urine Strip Testing	Yeast		False negative
CT-800	Ventana - All other Ventana products not listed	Ventana products Not listed		False Positive or False Negative released to Medical Record
CT-762	Ventana Ancillary Reagents	Amplification Kit	X	
CT-760	Ventana Ancillary Reagents	Antibody Diluent	X	
CT-776	Ventana Ancillary Reagents	Bluing	X	
CT-761	Ventana Ancillary Reagents	Endogenous Biotin Blocking	X	
CT-774	Ventana Ancillary Reagents	Hematoxylin	X	
CT-775	Ventana Ancillary Reagents	Hematoxylin II	X	
CT-1167	Ventana Ancillary Reagents	Hematoxylin III	X	
CT-786	Ventana Ancillary Reagents	HybReady	X	
CT-782	Ventana Ancillary Reagents	ISH Protease 1	X	
CT-783	Ventana Ancillary Reagents	ISH Protease 2	X	
CT-784	Ventana Ancillary Reagents	ISH Protease 3	X	
CT-763	Ventana Ancillary Reagents	OptiView Amplification kit	X	
CT-779	Ventana Ancillary Reagents	Protease 1	X	
CT-780	Ventana Ancillary Reagents	Protease 2	X	
CT-781	Ventana Ancillary Reagents	Protease 3	X	
CT-778	Ventana Ancillary Reagents	Red Counterstain II	X	
CT-1176	Ventana Ancillary Reagents	SSR Solution	X	
CT-409	Ventana Antibodies	Actin Muscle (HUC1-1)	X	
CT-410	Ventana Antibodies	ALK (D5F3)		False Positive or False Negative Released to Medical Record
CT-412	Ventana Antibodies	ALK1 (ALK01)		False Positive or False Negative Released to Medical Record
CT-907	Ventana Antibodies	Anti-p504s(SP116) Rabbit Monoclonal Pab	X	
CT-417	Ventana Antibodies	Basal Cell Cocktail (34bE12+p63)		False Positive or False Negative Released to Medical Record
CT-420	Ventana Antibodies	bcl-2 (124), CONFIRM		False Positive or False Negative Released to Medical Record
CT-421	Ventana Antibodies	bcl-2 (SP66)		False Positive or False Negative Released to Medical Record
CT-817	Ventana Antibodies	BRAF V600E (VE1)		False Positive or False Negative Released to Medical Record
CT-436	Ventana Antibodies	Calretinin (SP65), CONFIRM		False Positive or False Negative Released to Medical Record
CT-437	Ventana Antibodies	CD 15 (MMA), CONFIRM		False Positive or False Negative Released to Medical Record
CT-438	Ventana Antibodies	CD10 (SP67), VENTANA		False Positive or False Negative Released to Medical Record
CT-446	Ventana Antibodies	CD20 (L26), CONFIRM		False Positive or False Negative Released to Medical Record
CT-449	Ventana Antibodies	CD22 (SP104)	X	
CT-450	Ventana Antibodies	CD23 (SP23), CONFIRM		False Positive or False Negative Released to Medical Record
CT-452	Ventana Antibodies	CD3 (2GV6)		False Positive or False Negative Released to Medical Record

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-453	Ventana Antibodies	CD30 (Ber-H2), CONFIRM		False Positive or False Negative Released to Medical Record
CT-456	Ventana Antibodies	CD34 (QBEnd/10), CONFIRM		False Positive or False Negative Released to Medical Record
CT-458	Ventana Antibodies	CD4 (SP35)		False Positive or False Negative Released to Medical Record
CT-459	Ventana Antibodies	CD43 (L60)		False Positive or False Negative Released to Medical Record
CT-462	Ventana Antibodies	CD45 (LCA)(RP2/18), CONFIRM		False Positive or False Negative Released to Medical Record
CT-465	Ventana Antibodies	CD5 (SP19)	X	
CT-466	Ventana Antibodies	CD56 (123C3), CONFIRM		False Positive or False Negative Released to Medical Record
CT-471	Ventana Antibodies	CD68 (KP-1), CONFIRM	X	
CT-1008	Ventana Antibodies	CD68 (SP251v18) Rabbit Monoclonal Antibody		False Positive or False Negative Released to Medical Record
CT-472	Ventana Antibodies	CD7 (SP94)		False Positive or False Negative Released to Medical Record
CT-475	Ventana Antibodies	CD79a, CONFIRM	X	
CT-476	Ventana Antibodies	CD8 (SP57)		False Positive or False Negative Released to Medical Record
CT-477	Ventana Antibodies	CD99 (O13), CONFIRM		False Positive or False Negative Released to Medical Record
CT-900	Ventana Antibodies	CD117(EP10) Rabbit Monoclonal PAB		False Positive or False Negative Released to Medical Record
CT-481	Ventana Antibodies	Chromogranin A (LK2H10)		False Positive or False Negative Released to Medical Record
CT-966	Ventana Antibodies	CINtec PLUS Cytology		False Positive or False Negative Released to Medical Record
CT-482	Ventana Antibodies	c-kit (CD117), PATHWAY		False Positive or False Negative Released to Medical Record
CT-818	Ventana Antibodies	CLDN18 (43-14A) Analytic Assay	X	
CT-878	Ventana Antibodies	CLDN18 (43-14A) RxDx Assay in Gastric Carcinoma		False Positive or False Negative Released to Medical Record
CT-483	Ventana Antibodies	c-MET (SP44), CONFIRM		False Positive or False Negative Released to Medical Record
CT-485	Ventana Antibodies	c-MYC (Y69)		False Positive or False Negative Released to Medical Record
CT-488	Ventana Antibodies	Cyclin D1 (SP4-R)		False Positive or False Negative Released to Medical Record
CT-489	Ventana Antibodies	Cytokeratin (34bE12), CONFIRM		False Positive or False Negative Released to Medical Record
CT-490	Ventana Antibodies	Cytokeratin (AE1), CONFIRM		False Positive or False Negative Released to Medical Record
CT-491	Ventana Antibodies	Cytokeratin (AE3), CONFIRM		False Positive or False Negative Released to Medical Record
CT-492	Ventana Antibodies	Cytokeratin (CAM 5.2)		False Positive or False Negative Released to Medical Record
CT-493	Ventana Antibodies	Cytokeratin 10 (SP99)		False Positive or False Negative Released to Medical Record
CT-496	Ventana Antibodies	Cytokeratin 17 (SP95)		False Positive or False Negative Released to Medical Record
CT-498	Ventana Antibodies	Cytokeratin 20, CONFIRM		False Positive or False Negative Released to Medical Record
CT-500	Ventana Antibodies	Cytokeratin 5/6 (D5/16B4)		False Positive or False Negative Released to Medical Record
CT-501	Ventana Antibodies	Cytokeratin 7, CONFIRM		False Positive or False Negative Released to Medical Record
CT-503	Ventana Antibodies	Cytokeratin, Pan (AE1/AE3 & PCK26)		False Positive or False Negative Released to Medical Record
CT-505	Ventana Antibodies	Desmin (DE-R-11)		False Positive or False Negative Released to Medical Record
CT-923	Ventana Antibodies	DLL3 (SP347)	X	
CT-507	Ventana Antibodies	E Cadherin (36), VENTANA		False Positive or False Negative Released to Medical Record
CT-509	Ventana Antibodies	EGFR (3C6), CONFIRM		False Positive or False Negative Released to Medical Record
CT-510	Ventana Antibodies	EGFR (5B7), CONFIRM		False Positive or False Negative Released to Medical Record
CT-511	Ventana Antibodies	EGFR del (SP111)		False Positive or False Negative Released to Medical Record
CT-512	Ventana Antibodies	EGFR L858R (SP125)		False Positive or False Negative Released to Medical Record

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-513	Ventana Antibodies	EMA (E29)		False Positive or False Negative Released to Medical Record
CT-515	Ventana Antibodies	Epithelial Related Antigen (MOC-31)		False Positive or False Negative Released to Medical Record
CT-516	Ventana Antibodies	ERG (EPR3864)		False Positive or False Negative Released to Medical Record
CT-517	Ventana Antibodies	Estrogen Receptor (SP1), CONFIRM		False Positive or False Negative Released to Medical Record
CT-519	Ventana Antibodies	EZH2 (SP129)	X	
CT-534	Ventana Antibodies	Follicular Dendritic Cell (CAN.42)	X	
CT-387	Ventana Antibodies	FOLR1(1-2.1)- RxDx Assay		False Positive or False Negative Released to Medical Record
CT-544	Ventana Antibodies	Glypican 3 (GC33)		False Positive or False Negative Released to Medical Record
CT-550	Ventana Antibodies	Helicobacter pylori (SP48)		False Positive or False Negative Released to Medical Record
CT-553	Ventana Antibodies	HER-2/neu (4B5), PATHWAY		False Positive or False Negative Released to Medical Record
CT-1009	Ventana Antibodies	INSM1 (SP493) Rabbit Monoclonal Primary Antibody	X	
CT-1010	Ventana Antibodies	KAPPA LAMBDA		False Positive or False Negative Released to Medical Record
CT-566	Ventana Antibodies	Kappa (Polyclonal), CONFIRM		False Positive or False Negative Released to Medical Record
CT-567	Ventana Antibodies	Ki-67 (30-9), CONFIRM		False Positive or False Negative Released to Medical Record
CT-569	Ventana Antibodies	Lambda (Polyclonal)		False Positive or False Negative Released to Medical Record
CT-576	Ventana Antibodies	MART-1/melan A (A103), CONFIRM		False Positive or False Negative Released to Medical Record
CT-579	Ventana Antibodies	Melanoma Triple Cocktail (HMB45+A103+T311)		False Positive or False Negative Released to Medical Record
CT-580	Ventana Antibodies	Melanosome (HMB45), CONFIRM		False Positive or False Negative Released to Medical Record
CT-582	Ventana Antibodies	MITF C5/D5		False Positive or False Negative Released to Medical Record
CT-584	Ventana Antibodies	MLH-1 (M1)		False Positive or False Negative Released to Medical Record
CT-982	Ventana Antibodies	MSH2 (G219-1129)		False Positive or False Negative Released to Medical Record
CT-983	Ventana Antibodies	MSH6 (SP93)		False Positive or False Negative Released to Medical Record
CT-925	Ventana Antibodies	MSLN (SP74) Mesothelioma	X	
CT-587	Ventana Antibodies	MUC1 (H23)		False Positive or False Negative Released to Medical Record
CT-607	Ventana Antibodies	P120 Catenin (98)		False Positive or False Negative Released to Medical Record
CT-608	Ventana Antibodies	p16 - CINtec Histology Kit		False Positive or False Negative Released to Medical Record
CT-609	Ventana Antibodies	p16 - CINtec p16 Histology		False Positive or False Negative Released to Medical Record
CT-613	Ventana Antibodies	p53 (Bp53-11)		False Positive or False Negative Released to Medical Record
CT-614	Ventana Antibodies	p53 (DO-7)		False Positive or False Negative Released to Medical Record
CT-616	Ventana Antibodies	p63 (4A4)		False Positive or False Negative Released to Medical Record
CT-950	Ventana Antibodies	pan-TRK (EPR17341)		False Positive or False Negative Released to Medical Record
CT-617	Ventana Antibodies	PAX5 (SP34), CONFIRM	X	
CT-1152	Ventana Antibodies	PAX-8 (SP348) Rabbit Monoclonal Antibody		False Positive or False Negative Released to Medical Record
CT-915	Ventana Antibodies	PD-L1(SP263)		False Positive or False Negative Released to Medical Record
CT-949	Ventana Antibodies	PD-L1(SP142)		False Positive or False Negative Released to Medical Record
CT-984	Ventana Antibodies	PMS2 (A16-4)		False Positive or False Negative Released to Medical Record
CT-381	Ventana Antibodies	PRAME (EPR20330) Rbt Mono PAB		False Positive or False Negative Released to Medical Record

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-627	Ventana Antibodies	Progesterone Receptor (1E2), CONFIRM		False Positive or False Negative Released to Medical Record
CT-630	Ventana Antibodies	PSA (Polyclonal), CONFIRM		False Positive or False Negative Released to Medical Record
CT-821	Ventana Antibodies	ROS1 SP384 Rb Mono Primary Ab	X	
CT-635	Ventana Antibodies	S100 (4C4.9), CONFIRM		False Positive or False Negative Released to Medical Record
CT-636	Ventana Antibodies	S100 (Polyclonal), CONFIRM		False Positive or False Negative Released to Medical Record
CT-640	Ventana Antibodies	Smoothelin (R4A)	X	
CT-646	Ventana Antibodies	Synaptophysin (SP11), CONFIRM		False Positive or False Negative Released to Medical Record
CT-652	Ventana Antibodies	Thyroid Transcription Factor-1 (8G7G3/1)		False Positive or False Negative Released to Medical Record
CT-654	Ventana Antibodies	Topoisomerase II (JS5B4), CONFIRM		False Positive or False Negative Released to Medical Record
CT-659	Ventana Antibodies	TTF-1 (SP141)		False Positive or False Negative Released to Medical Record
CT-1153	Ventana Antibodies	TRPS1 (SP532) Rabbit Monoclonal Antibody		False Positive or False Negative Released to Medical Record
CT-660	Ventana Antibodies	Tyrosinase (T311), CONFIRM		False Positive or False Negative Released to Medical Record
CT-663	Ventana Antibodies	Vimentin (V9), CONFIRM	X	
CT-828	Ventana Antibodies	CD8 (SP239) Rabbit Monoclonal PAB test	X	
CT-873	Ventana Antibodies	Anti-AREG (SP272) Primary Ab	X	
CT-874	Ventana Antibodies	Anti-EREG (SP326) Primary Ab	X	
CT-881	Ventana Antibodies	FGFR2b (FPR2-D) Assay 1	X	
CT-885	Ventana Antibodies	p40 (SP225) Rabbit Monoclonal PAB		False Positive or False Negative Released to Medical Record
CT-886	Ventana Antibodies	p40 (BC28) Mouse Monoclonal Primary Antibody		False Positive or False Negative Released to Medical Record
CT-1168	Ventana Antibodies	VENTANA BAP1 (SP527) Rabbit Monoclonal Antibody		False Positive or False Negative Released to Medical Record
CT-1128	Ventana Antibodies	VENTANA HER2 (CAL27) Assay		False Positive Result or False Negative Result Reported to Physician
CT-1080	Ventana Antibodies	VENTANA MET (SP44) RxDx Assay		False Positive or False Negative Released to Medical Record
CT-1157	Ventana Antibodies	VENTANA TROP2 (EPR20043) Rabbit Monoclonal Antibody		False Positive or False Negative Released to Medical Record
CT-1158	Ventana Antibodies	VENTANA TROP2 (EPR20043) Rabbit Monoclonal Antibody RUO Assay	X	
CT-769	Ventana Bulk Reagents	10x SSC	X	
CT-771	Ventana Bulk Reagents	Cell Conditioning 2 (CC2)	X	
CT-767	Ventana Bulk Reagents	EZ Prep (10x)	X	
CT-765	Ventana Bulk Reagents	Liquid Coverslip (LCS) (Predilute)	X	
CT-768	Ventana Bulk Reagents	Reaction Buffer (10x)	X	
CT-772	Ventana Bulk Reagents	Ultra Cell Conditioning 1 (Ultra CC1)	X	
CT-773	Ventana Bulk Reagents	Ultra Cell Conditioning 2 (Ultra CC2)	X	
CT-764	Ventana Bulk Reagents	ULTRA LCS (Predilute)	X	
CT-785	Ventana Bulk Reagents	ultraView Silver Wash II	X	
CT-770	Ventana Bulk Reagents	Cell Conditioning 1 (CC1)	X	

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-798	Ventana Cell Marque Antibodies	All Cell Marque Complaints		Forward complaint to Cell Marque Cell Marque 6600 Sierra Collage Blvd. Rocklin, CA 95677 Direct: 1 (916) 746-8900 or 1 (800) 665-7284 Email: service@cellmarque.com Note: A global case is required whenever the corresponding Roche product(s) cannot be ruled out as causal OR in the event the case meets PSI criteria.
CT-724	Ventana Controls and Xenografts	EBER Positive Specimen Slide	X	
CT-714	Ventana Controls and Xenografts	HER2 Dual ISH 3 in 1 Xenograft	X	
CT-725	Ventana Controls and Xenografts	Kappa Positive Specimen Slide	X	
CT-726	Ventana Controls and Xenografts	Lambda Positive Specimen Slide	X	
CT-718	Ventana Controls and Xenografts	MET ISH Xenograft Control	X	
CT-720	Ventana Controls and Xenografts	Negative Control Reagents (IHC and ISH)	X	
CT-708	Ventana Controls and Xenografts	PATHWAY HER-2 4in1 Control slides		False Positive or False Negative Released to Medical Record
CT-717	Ventana Controls and Xenografts	TOP2A ISH Xenograft Control	X	
CT-704	Ventana Controls and Xenografts	Alu Positive Control Probe		False Positive or False Negative Released to Medical Record
CT-705	Ventana Controls and Xenografts	Alu Positive Control Probe II		False Positive or False Negative Released to Medical Record
CT-706	Ventana Controls and Xenografts	RNA Positive Control Probe		False Positive or False Negative Released to Medical Record
CT-703	Ventana Controls and Xenografts	U6	X	
CT-397	Ventana Detection Kits	ISH iVIEW Blue Detection Kit		False Positive or False Negative Released to Medical Record
CT-394	Ventana Detection Kits	ISH iVIEW Blue Plus Detection Kit		False Positive or False Negative Released to Medical Record
CT-404	Ventana Detection Kits	OptiView DAB IHC Detection Kit		False Positive or False Negative Released to Medical Record
CT-393	Ventana Detection Kits	UltraView Alkaline Phosphatase Red Detection Kit		False Positive or False Negative Released to Medical Record
CT-403	Ventana Detection Kits	ultraView Red ISH DIG Detection Kit		False Positive or False Negative Released to Medical Record
CT-399	Ventana Detection Kits	ultraView SISH DNP Detection Kit		False Positive or False Negative Released to Medical Record
CT-392	Ventana Detection Kits	Ultraview Universal DAB Detection Kit		False Positive or False Negative Released to Medical Record

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-888	Ventana Detection Kits	VENTANA® Red ISH DIG Detection Kit		False Positive or False Negative Released to Medical Record
CT-889	Ventana Detection Kits	VENTANA® Silver ISH DNP Detection Kit		False Positive or False Negative Released to Medical Record
CT-673	Ventana FITC Antibodies	FITC Anti-Albumin		False Positive or False Negative Released to Medical Record
CT-674	Ventana FITC Antibodies	FITC anti-C1Q		False Positive or False Negative Released to Medical Record
CT-676	Ventana FITC Antibodies	FITC Anti-C3		False Positive or False Negative Released to Medical Record
CT-671	Ventana FITC Antibodies	FITC Anti-C4		False Positive or False Negative Released to Medical Record
CT-668	Ventana FITC Antibodies	FITC Anti-Fibrinogen		False Positive or False Negative Released to Medical Record
CT-675	Ventana FITC Antibodies	FITC Anti-IgA		False Positive or False Negative Released to Medical Record
CT-667	Ventana FITC Antibodies	FITC Anti-IgG		False Positive or False Negative Released to Medical Record
CT-672	Ventana FITC Antibodies	FITC Anti-IgM		False Positive or False Negative Released to Medical Record
CT-669	Ventana FITC Antibodies	FITC Anti-Kappa		False Positive or False Negative Released to Medical Record
CT-670	Ventana FITC Antibodies	FITC anti-Lambda		False Positive or False Negative Released to Medical Record
CT-685	Ventana Molecular Probes	EBER		False Positive or False Negative Released to Medical Record
CT-890	Ventana Molecular Probes	VENTANA® HER2 Dual ISH DNA Probe Cocktail		False Positive or False Negative Released to Medical Record
CT-1105	Ventana Molecular Probes	EBER1 DIG Probe - ASR only	X	
CT-738	Ventana Special Stains	Alcian Blue		False Positive or False Negative Released to Medical Record
CT-742	Ventana Special Stains	Alcian Blue for PAS		False Positive or False Negative Released to Medical Record
CT-753	Ventana Special Stains	Congo Red		False Positive or False Negative Released to Medical Record
CT-1175	Ventana Special Stains	Congo Red Stain Core Kit		False Positive or False Negative Released to Medical Record
CT-747	Ventana Special Stains	Diastase		False Positive or False Negative Released to Medical Record
CT-733	Ventana Special Stains	Giemsa		False Positive or False Negative Released to Medical Record
CT-732	Ventana Special Stains	GMSII		False Positive or False Negative Released to Medical Record
CT-1174	Ventana Special Stains	Gram Stain Core Kit		False Positive or False Negative Released to Medical Record
CT-748	Ventana Special Stains	Green for Trichrome		False Positive or False Negative Released to Medical Record
CT-734	Ventana Special Stains	Iron		False Positive or False Negative Released to Medical Record
CT-750	Ventana Special Stains	Jones H&E Staining Kit		False Positive or False Negative Released to Medical Record
CT-740	Ventana Special Stains	Jones Light Green Staining Kit		False Positive or False Negative Released to Medical Record
CT-743	Ventana Special Stains	Light Green for PAS		False Positive or False Negative Released to Medical Record
CT-746	Ventana Special Stains	Mucicarmine Staining Kit		False Positive or False Negative Released to Medical Record
CT-739	Ventana Special Stains	PAS		False Positive or False Negative Released to Medical Record
CT-1011	Ventana Special Stains	PAS Stain Core Kit		False Positive or False Negative Released to Medical Record
CT-731	Ventana Special Stains	Reticulum II		False Positive or False Negative Released to Medical Record
CT-755	Ventana Special Stains	Steiner Cleaning kit		False Positive or False Negative Released to Medical Record
CT-751	Ventana Special Stains	Steiner II		False Positive or False Negative Released to Medical Record
CT-754	Ventana Special Stains	Trichrome Staining Kit		False Positive or False Negative Released to Medical Record
CT-255	Ventana Special Stains	Elastic Stain Core Kit Special Stains Van Gieson Counter Stain	X	
CT-1012	Virology	Adenovirus on the cobas Respiratory flex on cobas 5800/6800/8800 Systems	X	

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-831	Virology	BKV Monitoring - plasma		See Table F, BKV monitoring - plasma
CT-832	Virology	BKV Monitoring - urine		See Table G, BKV monitoring - urine
CT-1013	Virology	HSV 1 and 2 Test for use on the cobas® 4800 System		False Negative
CT-1014	Virology	common human coronavirus on the cobas Respiratory flex on cobas 5800/6800/8800 Systems	X	
CT-380	Virology	CMV monitoring / CMV blood screening		See Table B, Cytomegalovirus Monitoring For blood screening: False Non-Reactive and False Reactive (only if the false reactive result led to rejection of an organ or a tissue for transplantation).
CT-386	Virology	Ebola - *Diagnostic confirmation with the LightMix Ebola Zaire rRT-PCR Test (EUA)*		False Negative False Positive
CT-833	Virology	EBV Monitoring		See Table E, Epstein Barr (EBV) Monitoring
CT-388	Virology	Generic IVD, IUO Products and Lab Developed Tests	X	
CT-382	Virology	HBV monitoring		See Table D, Hepatitis B Baseline and/or Monitoring
CT-383	Virology	HCV - *Diagnostic Confirmation*		False Negative
CT-384	Virology	HCV baseline and monitoring		See Table C, Hepatitis C Virus Baseline and/or Monitoring
CT-1015	Virology	HCV diagnostic confirmation (cobas® HCV PSC claim extension)		False Negative
CT-1016	Virology	HCV baseline and monitoring (cobas® HCV PSC claim extension)		See Table H, Hepatitis C Virus Baseline and/or Monitoring (cobas® HCV PSC)
CT-385	Virology	HCV Genotyping	X	
CT-376	Virology	HIV Quantitative Tests – *Aid in HIV-1 Diagnosis*		False negative. False positive. Discrepant results.
CT-375	Virology	HIV-1/2 - Qualitative Tests- Diagnosis (non infants)		False negative. False positive. Discrepant results.
CT-336	Virology	HIV-1/2 - Qualitative Tests- Diagnosis (infants)		False negative. False positive. Discrepant results.
CT-377	Virology	HIV-1 baseline testing and monitoring		See Table A, Human Immunodeficiency Virus Baseline Testing and Monitoring
CT-1017	Virology	human metapneumovirus on the cobas Respiratory flex on cobas 5800/6800/8800 Systems	X	
CT-1018	Virology	human rhinovirus/enterovirus on the cobas Respiratory flex on cobas 5800/6800/8800 Systems	X	
CT-1019	Virology	influenza A on the cobas Respiratory flex on cobas 5800/6800/8800 Systems		False Negative
CT-1020	Virology	influenza B on the cobas Respiratory flex on cobas 5800/6800/8800 Systems		False Negative

Unique Identifier (Please do not modify)	Section	Parameter / Test	No escalation for PRI ER required	Criteria to classify as PRI ER
CT-1021	Virology	parainfluenza 1,2,3,4 on the cobas Respiratory flex on cobas 5800/6800/8800 Systems	X	
CT-1022	Virology	RSV on the cobas Respiratory flex on cobas 5800/6800/8800 Systems	X	
CT-1023	Virology	SARS-CoV-2 on the cobas Respiratory flex on cobas 5800/6800/8800 Systems		False Negative
CT-965	Virology	SARS-CoV-2 - cobas 5800/6800/8800 System (for qualitative and quantitative tests)		False Negative
CT-968	Virology	cobas SARS-CoV-2 & Influenza A/B nucleic acid test for use on the cobas Liat System		False Negative
CT-936	Virology	cobas® SARS-CoV-2, Influenza A/B & RSV nucleic acid test for use on the cobas Liat System AND cobas® Respiratory 4-flex for use on the cobas 58/68/8800 systems		**For SARS-CoV-2 target the following criteria apply:** False Negative **For Influenza A/B targets the following criteria apply:** False Negative **For RSV target the following criteria apply:** No evaluation for PRI is required
CT-972	Virology	TIB Assay HSV 1/2 & VZV		False Negative
CT-975	Virology	cobas SARS-CoV-2 & Influenza A/B Qualitative nucleic Acid test for use on the cobas 5800/ 6800/8800 System		False Negative
CT-996	Virology	Diagenode AD_hMPV_EV_RV (for cobas omni Utility Channel)	X	
CT-997	Virology	Diagenode Parainfluenza (for cobas omni Utility Channel)	X	
CT-337	Virology	cobas SARS-CoV-2 nucleic acid test for use on the cobas Liat System		False Negative
CT-700	Virology	cobas MPXV (monkeypox)	X	

PRI Screening Criteria for Erroneous Results

Table A

Human Immunodeficiency Virus Baseline Testing and Monitoring

HIV-1		Patient's Actual HIV Viral Load (copies/mL)				
		<LLoQ or TND	20 - <200	200 - <1,000	1,000 - <100,000	≥100,000
Reported Result	<LLoQ or TND		No	No	Yes	Yes
	20 - <200	No		No	No	Yes
	200 - <1,000	No	No		No	No
	1,000 - <100,000	No	No	No		No
	≥100,000	No	No	No	No	

Table B

Cytomegalovirus Monitoring

CMV		Patient's Actual CMV Viral Load (IU/mL)				
		TND	<LLoQ	LLoQ - ≤1820	1821 - 18,200	> 18,200
Reported Result	TND		No	No	Yes	Yes
	<LLoQ	No		No	Yes	Yes
	LLoQ - ≤1820	No	No		No	Yes
	1821 - 18,200	Yes	Yes	No		No
	> 18,200	Yes	Yes	No	No	

Table C

Hepatitis C Virus Baseline and/or Monitoring

HCV		Patient's Actual HCV Viral Load (IU/mL)				
		TND	<LLoQ	LLoQ - < 25	≥25 - ≤ 6,000,000	>6,000,000
Reported Result	TND		No	No	Yes	Yes
	<LLoQ	No		No	Yes	Yes
	LLoQ - <25	No	No		Yes	Yes
	≥25 - ≤ 6,000,000	No	No	No		No
	>6,000,000	No	No	No	No	

Table D

Hepatitis B Baseline and/or Monitoring

HBV		Patient's Actual HBV Viral Load (IU/mL)			
		<LLoQ or TND	LLoQ - ≤ 2000	2001 – 20,000	>20,000
Reported Result	<LLoQ or TND		No	Yes	Yes
	LLoQ - ≤ 2000	No		Yes	Yes
	2001 – 20,000	No	No		No
	>20,000	No	No	No	

Table E

Epstein Barr (EBV) Monitoring

EBV		Patient's Actual EBV Viral Load (IU/mL)				
		TND	<LLoQ	LLoQ - <1,000	1,000 - 10,000	>10,000
Reported Result	TND		No	No	Yes	Yes
	<LLoQ	No		No	Yes	Yes
	LLoQ - <1,000	No	No		No	Yes
	1,000 - 10,000	Yes	Yes	No		No
	>10,000	Yes	Yes	No	No	

Table F

BKV Monitoring (plasma)

BKV - plasma		Patient's Actual BKV Viral Load (IU/mL) - plasma			
		TND	<LLoQ	LLoQ - 1,000	>1,000
Reported Result	TND		No	No	Yes
	<LLoQ	No		No	Yes
	LLoQ - 1,000	No	No		No
	>1,000	Yes	Yes	No	

Table G

BKV Monitoring (urine)

BKV - urine		Patient's Actual BKV Viral Load (IU/mL) - urine			
		TND	<LLoQ	LLoQ - 10,000	>10,000
Reported Result	TND		No	No	Yes
	<LLoQ	No		No	Yes
	LLoQ - 10,000	No	No		No
	>10,000	Yes	Yes	No	

Table H

Hepatitis C Virus Baseline and/or Monitoring (cobas® HCV PSC)

HCV		Patient's Actual HCV Viral Load (IU/mL)				
		TND	<LLoQ	LLoQ - <1,000	≥1,000 to ≤ 6,000,000	>6,000,000
Reported Result	TND		No	No	Yes	Yes
	<LLoQ	No		No	Yes	Yes
	LLoQ - <1,000	No	No		Yes	Yes
	≥1,000 to ≤ 6,000,000	No	No	No		No
	>6,000,000	No	No	No	No	

Guidance for assessing Molecular Lab cases

If there is more than 1 result generated for a patient sample, and the true result is unknown, the case should be assessed using the worst-case scenario. See examples below:

• A patient sample generates a positive TV result during initial testing and negative TV result during repeat testing with the cobas® TV/MG Test for use on the cobas® 6800/8800 System; the true result is unknown. A "false positive" TV result for this specific test does not represent a PRI; however, a "false negative" TV result for this specific test does represent a PRI. As the true result is unknown, the case should be marked as a PRI as a "false negative" TV result represents a PRI.

• A Quality Control sample/panel generates a "Target Not Detected" result with the CAP/CTM HIV-1 Test on CAP/CTM system #1. On CAP/CTM system #2, the same QC panel generates a titer of "10^3 copies/mL". The expected result is unknown. If the "10^3 copies/mL" result was used as the "reported result" and the TND result was used as the "actual result" in the quantitative table for HIV-1, the case would not represent a PRI. However, if the "TND" result was used as the "reported result" and the "10^3 copies/mL" result was used the "actual result", the case would be marked as a PRI. Since the true result is unknown, the case should be marked using this "worst-case" scenario.

For systematic hardware/software/consumable malfunctions (failure to meet performance characteristics or otherwise perform as intended or designed) that could impact all assays utilized on the system (i.e., malfunction with the COBAS® AmpliPrep Instrument, which is used for blood-screening assays on the cobas s201 system and virology assays on the CAP/CTM system), all applicable assays utilized on the system must be evaluated for PRI determination. "Systematic" indicates that you would expect the same failure or malfunction to occur on the system, independent of which assay is being utilized.

For Molecular Lab screening lists, a result is considered delayed, in the context of PRI marking, if it requires a re-collection of the specimen from the patient.

If new information becomes available after the initial PRI assessment has been made, the case must be re-assessed based on the new information provided.