

Skeptic memo — Switching an In-Flight CSV Project to CSA

To: Skeptic

Re: Switching an In-Flight CSV Project to CSA

Basis: analysis-20261002T111530Z.md · Report · All files

Band (from §6): MEDIUM

Derived from §6 and the structured summary of analysis-20261002T111530Z.md; the six sections remain the source of truth.

In brief: The analysis rests its argument on regulatory text read at source and names the open project facts that would decide it (C1, C5).

It rests on the guidance being nonbinding and permitting alternative approaches (GT-1), validation being proportionate to risk (GT-3), the QMSR being in force (GT-4) and the record's required contents (GT-7).

Seven lines were set aside (§5). The lower-compliance-bar reading fails on GT-3 and GT-7 (C1), the one-method-per-package rule fails on GT-6 (C2), and starting now on FDA's endorsement fails because the guidance does not amend the firm's SOP (C3).

The weakest link is the near-tie between finishing under CSV and the boundary switch, both at 78, where a ± 1 change on any of five criteria moves the winner (C5).

An SOP that already permits risk-based assurance would open the gate in C3 (GT-9?). A remainder mostly not high process risk, with $R \times f \times r$ above B, would favour the boundary switch (C4, GT-10?).