

Skeptic view — Switching an In-Flight CSV Project to CSA

Derived from §6 and the structured summary of analysis-20261002T111530Z.md; the six sections remain the source of truth.

Band (from §6): MEDIUM

- **What does the argument rest on?** It rests on read regulatory text: the guidance is nonbinding and permits alternatives (GT-1), validation must be proportionate to risk (GT-3), and the record's contents are set (GT-7).
- **What alternatives were considered, and why were they set aside?** Seven lines were set aside: CSA as a lower bar, one method per package, starting now on FDA endorsement, an immediate full switch, a pause-then-switch, a single breakeven number, and a lifetime lock-in to CSV, each contradicted by a read ground truth or outscored (§5).
- **Where is the argument weakest?** The default is a near-tie: O1 and O3 both score 78, and a ± 1 change on any of five criteria moves the winner to O3 (C5).
- **What evidence would overturn it?** An SOP that already permits risk-based assurance would open the gate in C3 (GT-9?). A remaining scope mostly not high process risk, with $R \times f \times r$ above B, would favour the boundary switch (C4, GT-10?).