

Risk memo — Switching an In-Flight CSV Project to CSA

To: Risk

Re: Switching an In-Flight CSV Project to CSA

Basis: analysis-20261002T111530Z.md

Band (from §6): MEDIUM

Derived from §6 and the structured summary of analysis-20261002T111530Z.md; the six sections remain the source of truth.

In brief: The analysis separates the regulatory risk of CSA itself from the execution risk of changing method mid-project, and it carries a register of risks with a control for each (C7).

It retires the belief that CSA is a lower compliance bar, because the obligation is unchanged and only the evidence method differs (C1, A-2). It also retires the risk that executed evidence is lost under a switch, since scripted tests carry over with a risk mapping (C2, A-3).

Four ground truths are unverified: the SOP and plan state (GT-9?), the not-high-risk share of remaining scope (GT-10?), the ISO 13485 documented-procedure wording (GT-11?) and the system's scope (GT-13?).

If the system is drug/biologic GMP software or a device software function, the CSA guidance does not directly apply (A-5, GT-13?).

The §6 Pre-check caps the conclusion at MEDIUM, with MEDIUM as the lowest cited chain (§6). Any CSA activity run before the SOP and plan permit it is flagged as a self-inflicted nonconformance (C3, C7).