

Skeptic memo — Switching an In-Flight CSV Project to CSA

To: Skeptic

Re: Switching an In-Flight CSV Project to CSA

Basis: analysis-20261002T111530Z.md

Band (from §6): MEDIUM

Derived from §6 and the structured summary of analysis-20261002T111530Z.md; the six sections remain the source of truth.

In brief: The analysis rests its argument on read regulatory text and names the open project facts that would decide it (C1, C5).

It rests on the guidance being nonbinding and permitting alternative approaches (GT-1), validation being proportionate to risk (GT-3), the QMSR being in force (GT-4) and the record's required contents (GT-7).

Seven lines were set aside, among them CSA as a lower bar, one method per package, and starting now on FDA's endorsement (§5). An immediate full switch scored 53 and a pause-then-switch 67, against 78 for the default (C5).

The weakest link is the near-tie between finishing under CSV and the boundary switch, both at 78, where a ± 1 change on any of five criteria moves the winner (C5).

An SOP that already permits risk-based assurance would open the gate in C3 (GT-9?). A remainder mostly not high process risk, with $R \times f \times r$ above B, would favour the boundary switch (C4, GT-10?).