
G2: Lepto_minor



Sample Table

Well	DIN	Conc. [ng/µl]	Sample Description	Alert	Observations
Œ	-	48.8	Lepto_minor		Issue with ladder peak detection (too few peaks detected); Caution! Expired ScreenTape device

Peak Table

	Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	% Integrated Area	From [bp]	To [bp]	Peak Comment	Observations
Γ	100	8.50	8.50	-	39	164		Lower Marker
Γ	5769	31.3	-	81.00	3251	14643		
Γ	-	-	-	-	-	-		Sample Well

Generated: 09-Jul-2020