

## IMAGING DICOM GATEWAY IMPORTER USER MANUAL

**July 2010** 

Department of Veterans Affairs Office of Enterprise Development Health Provider Systems

### **Preface**

The purpose of this document is to help users understand the operation of the Importer application that resides on the DICOM Image Gateway and which is used to transfer DICOM objects from studies performed outside of the VA into VistA Imaging. The Importer can also be used to correct and import locally acquired DICOM objects that were rejected by VistA due to patient and/or study identification information mismatch.

#### **Revision Table**

Rev	Date	Notes	Notes
.1	05-15-2009	Document Creation.	Peter Kuzmak and D. Smith
.2	01-26-2010	Substantial update	Peter Kuzmak
.3	03-26-2010	Added Appendix B and Appendix C	Peter Kuzmak
.4	03-29-2010	Added Appendix D	Peter Kuzmak
.5	04-08-2010	Added section on deleting studies	Peter Kuzmak
.6	04-13-2010	Added Appendix E	Peter Kuzmak

### **Property of the US Government**

This is a controlled document. No changes to this document may be made without the express written consent of the VistA Imaging Office of Enterprise Development Office.

While every effort has been made to assure the accuracy of the information provided, this document may include technical inaccuracies and/or typographical errors. Changes are periodically made to the information herein and incorporated into new editions of this document.

Product names mentioned in this document may be trademarks or registered trademarks of their respective companies, and are hereby acknowledged.

VistA Imaging Office of Enterprise Development

Department of Veterans Affairs

Internet: <a href="http://www.va.gov/imaging">http://www.va.gov/imaging</a>
VA intranet: <a href="http://vaww.va.gov/imaging">http://vaww.va.gov/imaging</a>

1335 East-West Highway 3<sup>rd</sup> Floor

Silver Spring, MD 20910

(301) 734-0100

# <u>Caution</u>: Federal law restricts this device to use by or on the order of either a licensed practitioner or persons lawfully engaged in the manufacture or distribution of the product. Documentation Conventions

The following conventions are used in this manual.

- 1. Interactive user interface elements such as menus and button names are shown in **bold**
- 2. User input is also shown in **bold**
- 3. Keyboard key names are shown in **bold** and in brackets. For example: Press **<Enter>**
- 4. Sample output is shown in a monospace font.
- 5. A vertical bar is used to separate menu choices. For example, "Click File|Open" means: "Click the File menu, then click the Open option."

### **Table of Contents**

Prefac	ce iii	
Revisio	n Table	ii
Docum	entation Conventions	i
	ter 1 Introduction	
1.1	Background	
1.2	Importer Workflow Overview	
1.2.1	Importer Input	
1.2.2	Importer Workflow for Unordered Studies (Radiology Only)	
1.2.3	Importer Workflow for Ordered Studies	
Chapt	ter 2 Menu Overview and Parameter Setup	
2.1	Caché Terminal Setup	
2.2	Enabling the Importer.	
2.3	Importer Menus	
2.4	Importer Parameter Setup.	
2.5	Setting up Outside Imaging Locations for the DICOM Importer – See Appendix A	
2.6	Running the Importer from Remote Desktop – See Appendix B	
2.7	Handling Parent-Descendent Procedure Orders with the Importer – See Appendix C	
2.8	Problems Seen in Importing of DICOM Objects from Media – See Appendix D	
2.9	Using the BATCH_SEND_IMAGE.BAT utility script – See Appendix E	
Chant	ton 2 Importor Session Operation	10
<b>Chapt</b> 3.1	ter 3 Importer Session Operation Importer Session – Ordered Study Workflow	
3.1.1		
3.1.1	Importer Study Selection Initial Checking of DICOM Studies	
3.1.1.1		
3.1.1.2		
3.1.1.3	<u>.                                      </u>	
3.1.1.4	• • • • • • • • • • • • • • • • • • • •	
3.1.2	Associating the Selected Study with an Ordered Study	
3.1.3 3.1.4	Associating an Outside Contracted Study with the Order on VistA	
3.1.4	Importing the Ordered Study into VistA	
	Deleting (not Importing) Transmitted DICOM Studies	
3.2	Correcting Images with Erroneous Patient/Study Identification	
3.3		
3.3.1	Ordering Provider and Ordering Location	
3.3.2 3.3.3	Importer Study Selection	
	Identifying the VistA Patient	
3.3.4	Selecting the Corresponding VistA Radiology Procedure	
3.3.5	Importing the Unordered Radiology Study into VistA	
3.3.6	Display of an "Unordered Study" in CPRS	
3.4	Deleting (not Importing) Transmitted DICOM Studies	
3.5	Details of Study Checking.	
3.5.1	Flagging Non-supported DICOM Objects	
3.5.2	Identifying Studies that have Already Been Imported	
Chant	ter 4 Import Reconciliation Process	<b>4</b> 1

Chapter 5	Importer Session Log Files	43
Appendix A	· · · · · · · · · · · · · · · · · · ·	45
	side Imaging Locations for the DICOM Importer	
	fy Imaging Types and Divisions that need "No Credit" Imaging Locations	
	e the OUTSIDE STUDY entry in the CAMERA/EQUIP/RM file	
	ew Outside Imaging Locations for each Imaging Type to Divisions	
	e Outside Imaging Locations Parameters	
	SIDE STUDY" Camera/Equip/Room for all Radiology Imaging Location	
Step 6 – Popul	ate VistA Imaging Outside Imaging Location file	58
Appendix B	<b>.</b>	61
Running the In	nporter from a Remote Desktop	61
Microsoft Terr	ninal Server Installation	61
Licensing the T	Ferminal Server on Windows Server 2003	61
Installing Term	ninal Server	62
Setup the Rem	ote Desktop Client on a Windows XP Workstation	63
	bor between local Gateway User and Remote User	
Using the Rem	ote Desktop Connection for the Importer	68
Appendix C		71
	ent-Descendent Procedure Orders with the Importer	
	'	
Questions		71
How Outside F	Fee-basis Imaging Contractors Work	72
	Importer for DICOM Images for Parent/Descendent Procedures	
Appendix D	)	75
	in Importing of DICOM Objects from Media	
	In importing of Diegon conjects from Media.	
Appendix E		79
	CH_SEND_IMAGE.BAT utility script	

### **Table of Figures**

Figure 1 – Caché Control Panel (before edit going into programmer mode)	5
Figure 2 – Caché Control Panel   Edit User	
Figure 3 – Caché Control Panel with DICOM Gateway login program	6
Figure 4 – VA Import Reconciliation Workflow	
Figure 5 – High Level Importer Workflow	15
Figure 6 – Study Selection and Association	
Figure 7 – Importer Study Selection Screen	19
Figure 8 – Dental Image	22
Figure 9 – Radiology Order File Lookup Screen	28
Figure 10 – CPRS Radiology Study Details	35
Figure 11 – Importer Session Spreadsheet	44
Figure 12 – Enabling Terminal Server on Add/Remove Windows Components	62
Figure 13 – Adding Import Users to the Administrator Group	63
Figure 14 – Starting Remote Desktop Connection	64
Figure 15 – Specifying the Importer DICOM Gateway Terminal Server	64
Figure 16 – Entering User name for Remote Desktop Connection	65
Figure 17 – Setting the Screen Resolution and Color Quality	66
Figure 18 – Setting Network Performance Parameters	67
Figure 19 – Starting Remote Desktop Connection	68
Figure 20 – Specifying the Importer DICOM Gateway Terminal Server	69
Figure 21 – Logging into Terminal Server on Importer DICOM Gateway	69
Figure 22 – Running the Importer Remotely	70

### Chapter 1 Introduction

The Importer is an application on the DICOM Image Gateway that can be used to transfer DICOM objects from studies performed outside of the VA into VistA Imaging. These DICOM objects may be transmitted directly from the outside facility to the VA, as in the case for the studies for the polytrauma patients, or they may be loaded from portable media (CDs or DVDs). Portable media must conform to the DICOM Standard and the Integrating the Healthcare Enterprise (IHE) Portable Data for Imaging (PDI) integration profile. 

1 (Unfortunately, some PACS generate "DICOM" CDs and DVDs in proprietary formats that can't be processed by VistA Imaging – see Appendix D for examples.) Report files that are on the media will not processed at this time but will be dealt with in a future patch.

The Importer can also be used to correct and import locally acquired DICOM objects that were rejected by VistA due to patient and/or study identification information mismatch. This functionality replaces the current DICOM Correct utility on workstations where the Importer software is installed.

### 1.1 Background

The Importer can handle both *ordered* and *unordered* studies.

*Ordered* studies are those that already have corresponding orders on the local VistA system. An example of an ordered study is a contracted study which is ordered by the VA and is performed at an outside facility for a fee. Another example of an ordered study is a locally acquired DICOM study that has erroneous patient and/or study identification information.

*Unordered* studies are those that are unknown or unexpected by to the local VistA system. The classic example is the unordered prior study that was performed while the patient was being treated at an outside facility. The patient brought the CD containing the prior study when s/he went to the local VA for treatment.

The IHE Import Reconciliation Workflow (IRWF) integration profile specifies the industry-standard technique for importing DICOM objects. With the IRWF, patient and study identification information is obtained from a study on the local system and is then used to replace the values of the corresponding data elements in the DICOM objects that are to be imported. To provide an audit trail, the original DICOM values along with the import application attributes are saved elsewhere in the DICOM header. The outside study's DICOM objects, now with the local patient and study identification, are then imported and associated with the study on the local system.

<sup>&</sup>lt;sup>1</sup> IHE Portable Data for Imaging Integration Profile, IHE Technical Framework, vol I, Rev. 8.0, 2007-08-30.

<sup>&</sup>lt;sup>2</sup> IHE Import Reconciliation Workflow, IHE Technical Framework, vol I, Rev. 8.0, 2007-08-30.

When IRWF is used for importing <u>ordered</u> studies, both the patient and the study are already registered on the local system. The patient and study identification information from each local study is used to update the original outside study's DICOM objects. Importation then takes place and each outside study's DICOM objects are associated with the corresponding ordered study on the local system.

The situation is different, however, when IRWF is used for importing <u>unordered</u> studies. First, it is necessary to make sure that the patient is registered in the local system. Then a corresponding equivalent study must be created in the local system for every unordered outside study that is going to be imported. The patient and study identification information for each newly created local study is used to update the original outside study's DICOM objects. Importation then takes place and each outside study's DICOM objects are associated with its corresponding newly created study on the local system.

### 1.2 Importer Workflow Overview

The Importer workflow was designed to take advantage of automation and reduce manual effort. All of the manual steps to import a set of studies are performed first. User input activity and wait time are kept to a minimum. Once all the manual steps have been completed, all the automated import reconciliation steps for the entire set of studies are performed. Similar workflow handles the importing of both unordered and ordered outside studies.

### 1.2.1 Importer Input

The user can select the source of the DICOM objects to be imported. This may either be direct transmission or portable media. A list of studies to be imported from the designated source is then created and presented to the user.

### 1.2.2 Importer Workflow for Unordered Studies (Radiology Only)

It is assumed that the patient is already registered on the local VistA system.

The user can select an *unordered* outside study to be imported. The user can view the study and series information and the images. The user can then identify the patient from the local VistA HIS. Lastly, the user can specify the local VistA radiology procedure and modifier(s) that most closely match what was preformed. The user can repeat this process for each outside study to be imported.

After the user assigns the VistA patient identifiers and specifies the equivalent procedure and modifier(s) to each unordered study to be imported, the rest of the import reconciliation workflow is performed automatically. The corresponding equivalent radiology studies are automatically created on the local VistA system. Each new study is created with the same study date and time as the originally performed study so that the chronological study sequence information is properly maintained. Then the newly created study's patient and study identification information is used to update the values in the DICOM objects for the

corresponding outside study. To provide an audit trail, the original DICOM values along with the import application attributes are saved elsewhere in the DICOM header. The DICOM objects with the new patient and study identification information then are collectively imported and associated with the newly created studies.

(This automatic ordering capability is not supported for Consults.)

### 1.2.3 Importer Workflow for Ordered Studies

The Importer handles both *ordered* Radiology and *ordered* Consult studies.

The user can select an ordered outside study to be imported and view its study and series information and its images. The user can then issue a Modality Worklist query to find the originally ordered study in the local VistA system. The patient and study identification information from that study is then used to update the imported DICOM objects.

The user can alternatively retrieve this information (for a radiology study) from the RADIOLOGY ORDER file (#75.1).

After the user identifies the original studies for each ordered study, the rest of the import reconciliation workflow is performed automatically. The outside study's DICOM objects, now with the VistA patient and study identification, are then imported and associated with the originally ordered study.

Chapter 1 – Introduction

### Chapter 2 Menu Overview and Parameter Setup

### 2.1 Caché Terminal Setup

The Importer application is designed to work using a 132 x 48 character Caché Terminal window. The following steps may be used to set this up.

- 1. Click the blue Caché icon in the system tray and select Terminal.
- 2. If the Caché Terminal session starts in Programmer Mode (with the DICOM> prompt), enter the following MUMPS command: DICOM>D ^MAGDLOGN

To have the Caché Terminal session automatically start the login routine, perform the following steps:

- 2.1 Click the blue Caché icon and select Control Panel.
- 2.2 In the Caché Control Panel window, select Security | User Accounts.




Figure 1 – Caché Control Panel (before edit going into programmer mode)

2.3 Select TRM:, right click and select Edit. Change the Routine from ^%PMODE to ^MAGDLOGN, and click on OK.




Figure 2 – Caché Control Panel | Edit User

2.4 You should now see the Routine for TRM: specified as ^MAGDLOGN.




Figure 3 – Caché Control Panel with DICOM Gateway login program

- 3. In the Caché Terminal session, use Edit | Font to set Font to Terminal, Regular, 9 point. Click on OK.
- 4. Use Edit | Window Size to set Columns to 132 and Rows to 48. Click Save.
- 5. Use Edit | Color to change Foreground and Background colors.
- 6. Set Foreground color to White (255, 255, 255)
- 7. Set Background color to Dark Blue (0, 0, Blue=127). You may have to use Adjust Colors.
- 8. Click Save, and Apply

### 2.2 Enabling the Importer

The Importer runs on a DICOM Image Gateway. Before anyone can use the Importer, run menu option **4-2-2 Update Gateway Configuration Parameters** to enable it:

Use the IMPORTER instead of "DICOM Correct"? NO// Y

### 2.3 Importer Menus

6

The Importer is menu option is **2-12 Import DICOM Objects**.

VistA DICOM Image Gateway

- 1 (Receive PACS Exam Complete Messages)
- 2 (Send PACS Request Image Transfer Messages)
- 3 Process DICOM Images
- 4 Increment DICOM Image Input Pointer
- 5 Display Real-Time Storage Server Statistics
- 6 Display Cumulative Storage Server Statistics
- 7 Display Daily Image Processing Statistics

- 8 Send DICOM Images to Another Storage Server
- 9 Display a DICOM Image Header
- 10 Re-Transmit Images from PACS
- 11 Purge Incomplete Image Information
- 12 Import DICOM Objects
- 13 Validate Failed Image Table

OPTION: 12

Import DICOM Objects

- 1 Import DoD studies that were transmitted directly to the VA
- 2 Import Outside Contracted studies
- 3 CD/DVD Import Unordered studies from DoD facilities
- 4 (CD/DVD Import Unordered studies from other VA facilities)
- 5 CD/DVD Import Unordered studies from non-VA/DoD facilities
- 6 Correct local studies with bad PID and/or study information
- 7 Build Radiology Procedure File
- 8 Build Radiology Modifiers File
- 9 Modify Exam/Report Status Update Parameters

#### OPTION:

The first six options are for importing files, while the last three are for setting up necessary dictionary information.

Use of menu option **4 CD/DVD Import Unordered studies from other VA facilities** is unavailable and may be supported in a future patch.

#### 2.4 Importer Parameter Setup

Menu option **9 Modify Exam/Report Status Update Parameters** is used to establish the parameters that are necessary for advancing the exam status of a study from "Waiting to be Examined" to "Complete" and the report status from "No Report" to "Electronically Filed."

Import DICOM Objects

- 1 Import DoD studies that were transmitted directly to the VA
- 2 Import Outside Contracted studies
- 3 CD/DVD Import Unordered studies from DoD facilities
- 4 (CD/DVD Import Unordered studies from other VA facilities)
- 5 CD/DVD Import Unordered studies from non-VA/DoD facilities
- 6 Correct local studies with bad PID and/or study information
- 7 Build Radiology Procedure File
- 8 Build Radiology Modifiers File
- 9 Modify Exam/Report Status Update Parameters

OPTION: 9

Modify Exam/Report Status Update Parameters

1 Display Exam/Report Status Update Parameters

- 2 Modify Radiology Technologist
- 3 Modify Film Entry
- 4 Modify Diagnostic Code
- 5 Modify Camera/Equipment Room
- 6 Modify CPT Modifier
- 7 Modify All the above Parameters

OPTION:

Menu option 2 identifies the radiology technologist who normally is involved in the examination process. When an unordered study is imported from the outside, however, this person does not participate in the examination process.

Here is the interaction for setting the Radiologist Technologist parameter:

Modify Exam/Report Status Update Parameters

- 1 Display Exam/Report Status Update Parameters
- 2 Modify Radiology Technologist
- 3 Modify Film Entry
- 4 Modify Diagnostic Code
- 5 Modify Camera/Equipment Room
- 6 Modify CPT Modifier
- 7 Modify All the above Parameters

OPTION: 2

Enter the Radiology Technologist: O

Select the Radiology Technologist from this list:

- 1 -- ONESEVEN, PROVIDER
- 2 -- ONETHREENINESEVEN, PROVIDER
- 3 -- ONETHREEONE, PROVIDER
- 4 -- ONETHREETHREE, PROVIDER
- 5 -- ONETHREEZERO, PROVIDER
- 6 -- ONETWO, PROVIDER

Enter the number [1-6]: 6

Radiology Technologist changed to "ONETWO, PROVIDER"

Press <Enter> to continue...

Here is the interaction for setting the Film Entry parameter (specify "NO FILMS"):

Modify Exam/Report Status Update Parameters

- 1 Display Exam/Report Status Update Parameters
- 2 Modify Radiology Technologist
- 3 Modify Film Entry

```
4 Modify Diagnostic Code
5 Modify Camera/Equipment Room
6 Modify CPT Modifier
7 Modify All the above Parameters

OPTION: 3

Enter the Film Size: ?

Select the Film Size from this list:
1 -- 16 x 20
2 -- 4 x 4
3 -- 8 1/2 x 11
4 -- NO FILMS

Enter the number [1-4]: 4

Film Size changed to "NO FILMS"

Press <Enter> to continue...
```

The Radiology Diagnostic Code has minimal meaning for unordered imported studies. Use a neutral value for this. Here is the interaction for setting the Diagnostic Code parameter (specifying "NO ALERT REQUIRED"):

```
Modify Exam/Report Status Update Parameters
      Display Exam/Report Status Update Parameters
      Modify Radiology Technologist
   2
     Modify Film Entry
   3
      Modify Diagnostic Code
      Modify Camera/Equipment Room
      Modify CPT Modifier
   6
      Modify All the above Parameters
  OPTION: 4
Enter the Radiology diagnostic code: ?
Select the Radiology diagnostic code from this list:
  1 -- ABNORMALITY, ATTN. NEEDED
  2 -- CRITICAL ABNORMALITY
  3 -- MAJOR ABNORMALITY, NO ATTN. NEEDED
  4 -- MAJOR ABNORMALITY, PHYSICIAN AWARE
  5 -- MINOR ABNORMALITY
  6 -- NO ALERT REQUIRED
  7 -- NORMAL
  8 -- POSSIBLE MALIGNANCY
  9 -- POSSIBLE MALIGNANCY, FOLLOW-UP NEEDED
```

```
10 -- SIGNIFICANT ABNORMALITY, ATTN NEEDED
11 -- TELERADIOLOGY, NOT YET DICTATED
12 -- UNDICTATED FILMS NOT RETURNED, 3 DAYS
13 -- UNSATISFACTORY/INCOMPLETE EXAM

Enter the number [1-13]: 6

Radiology diagnostic code changed to "NO ALERT REQUIRED"

Press <Enter> to continue...
```

<u>Always</u> use "OUTSIDE STUDY" for the Camera/Equipment/Room (see Appedix). Here is the interaction for setting this parameter:

```
Modify Exam/Report Status Update Parameters
      Display Exam/Report Status Update Parameters
   2 Modify Radiology Technologist
   3 Modify Film Entry
   4 Modify Diagnostic Code
   5 Modify Camera/Equipment Room
      Modify CPT Modifier
      Modify All the above Parameters
  OPTION: 5
Enter the Camera/Equipment/Room: ?
Select the Camera/Equipment/Room from this list:
  1 -- OUTSIDE STUDY (Study performed outside of this facility)
  2 -- PORTABLE (MOBILE PICTURES)
  3 -- TD RAD-EXAM1 (This is the primary x-ray room.)
  4 -- TD RAD-EXAM2 (Sonograms are done in this room.)
  5 -- TD RAD-EXAM3 (Endoscopies are done in this room.)
Enter the number [1-5]: 1
Camera/Equipment/Room changed to "OUTSIDE STUDY"
```

The CPT modifier has minimal meaning for unordered imported studies. Use a neutral value for this. Here is the interaction for setting the CPT Modifier parameter (specifying "TECHNICAL COMPONENT"):

```
Modify Exam/Report Status Update Parameters
```

- 1 Display Exam/Report Status Update Parameters
- 2 Modify Radiology Technologist

Press <Enter> to continue...

```
Modify Film Entry
      Modify Diagnostic Code
   5
      Modify Camera/Equipment Room
      Modify CPT Modifier
      Modify All the above Parameters
  OPTION: 6
Enter the CPT Modifier: T
Select the CPT Modifier from this list:
  1 -- TECHNICAL COMPONENT
  2 -- TEL (LEUKEMIA)
  3 -- TFE3 OR ASPSCR1 (ALVEOLAR SO
  4 -- THIRD/MORE CONCURRENTLY ADMIN INFUSION THERAPY
  5 -- THREE PATIENTS SERVED
  6 -- TP53, COMMONLY CALLED P53
  7 -- TPMT (THIOPURINE METHYLTRANS
  8 -- TRB, T CELL RECEPTOR BETA (L
  9 -- TRG, T CELL RECEPTOR GAMMA (
 10 -- TWIST (SAETHRE-CHOTZEN SYNDR
 11 -- TWO CLASS B FINDINGS
 12 -- TWO PATIENTS SERVED
13 -- TWO SURGEONS
Enter the number [1-13]: 1
CPT Modifier changed to "TECHNICAL COMPONENT"
Press <Enter> to continue...
```

### All of the current parameter values may be displayed using menu option **1 Display Exam/Report Status Update Parameters**, as shown below:

Modify Exam/Report Status Update Parameters

```
2
    Modify Radiology Technologist
    Modify Film Entry
 3
    Modify Diagnostic Code
 4
 5
    Modify Camera/Equipment Room
 6
    Modify CPT Modifier
    Modify All the above Parameters
OPTION: 1
   Radiology Technologist(*): ONETWO,PROVIDER
                Film Size(*): NO FILMS
Radiology diagnostic code(*): NO ALERT REQUIRED
    Camera/Equipment/Room(*): OUTSIDE STUDY
                CPT Modifier: TECHNICAL COMPONENT
```

Display Exam/Report Status Update Parameters

1

#### Chapter 2 – Menu Overview and Parameter Setup

Fields with an asterisk (\*) require values.

Press <Enter> to continue...

- 2.5 Setting up Outside Imaging Locations for the DICOM Importer See Appendix A
- 2.6 Running the Importer from Remote Desktop See Appendix B
- 2.7 Handling Parent-Descendent Procedure Orders with the Importer See Appendix C
- 2.8 Problems Seen in Importing of DICOM Objects from Media See Appendix D
- 2.9 Using the BATCH\_SEND\_IMAGE.BAT utility script See Appendix E

### Chapter 3 Importer Session Operation

The Importer can handle ordered and unordered (prior) studies that are received from direct transmissions or loaded from portable media (CD/DVD). These studies can come from the DoD, a contracted facility, a different VA, or another institution.

The first six Importer menu options specify the operation of the Importer session.

Import DICOM Objects

- 1 Import DoD studies that were transmitted directly to the VA
- 2 Import Outside Contracted studies
- 3 CD/DVD Import Unordered studies from DoD facilities
- 4 (CD/DVD Import Unordered studies from other VA facilities)
- 5 CD/DVD Import Unordered studies from non-VA/DoD facilities
- 6 Correct local studies with bad PID and/or study information
- 7 Build Radiology Procedure File
- 8 Build Radiology Modifiers File
- 9 Modify Exam/Report Status Update Parameters

If a directly transmitted DICOM object fails association because of Patient Name, Patient ID, and/or Accession Number mismatch, menu options 1 and 6 may be used to import these studies. (Menu option 2 may also be used if the outside studies were transmitted from the contracted facility to the VA).

**Note:** The VistA hospital information system DISPLAY STUDIES TO BE IMPORTED option on the IMPORTER menu may be used to view each gateway's list of the transmitted studies that are waiting to be imported.

If the studies are obtained from portable media, menu options 2, 3, 4, and 5 may be used to import their DICOM objects.

The Importer session is divided into two parts, a manual interactive part followed by an automated non-interactive part. In the interactive part, the user selects the studies that are to be imported and defines the association between them and patient and study entities on VistA. In the non-interactive part, the necessary radiology studies are ordered and/or have their status updated, reconciliation takes place, and images are imported, all without any user involvement.

The following two diagrams give high level views of the Importer workflow. Figure 4 – VA Import Reconciliation Workflow is a very high level view of the Importer application. Figure 5 – High Level Importer Workflow illustrates the two major parts of the Importer session. The manual interactive part of the Importer session is "Study Selection and Association" while the automatic part consists of the "Automatic Study Ordering" and "Import Reconciliation" steps.

### **VA Import Reconciliation Workflow**




Figure 4 - VA Import Reconciliation Workflow




Figure 5 – High Level Importer Workflow

### 3.1 Importer Session – Ordered Study Workflow

Import DICOM Objects

- 1 Import DoD studies that were transmitted directly to the VA
- 2 Import Outside Contracted studies
- 3 CD/DVD Import Unordered studies from DoD facilities
- 4 (CD/DVD Import Unordered studies from other VA facilities)
- 5 CD/DVD Import Unordered studies from non-VA/DoD facilities
- 6 Correct local studies with bad PID and/or study information
- 7 Build Radiology Procedure File
- 8 Build Radiology Modifiers File
- 9 Modify Exam/Report Status Update Parameters

There are two Importer options for the processing of ordered studies. The first option (#2) is used to import studies from an outside contracted facility and the other option (#6) is used for correcting local studies with erroneous patient and/or study identification information. The workflow for these two Importer options is identical. (The only difference is that the Imaging file (#2005) field ORIGIN INDEX is "F" for the fee basis contracted studies and "V" for the corrected ones.)

Studies that are to be imported are selected from a list and matched to ordered studies on VistA. After this is done, import reconciliation is performed and they are imported.

### 3.1.1 Importer Study Selection

The Importer Study Selection is used for all imported studies, regardless if they were previously ordered or not, or whether they were directly acquired or obtained from portable media. This workflow allows the user to select importable DICOM studies and match them to existing studies (ordered case) or VistA patients and equivalent procedures (unordered case).

The list of importable DICOM studies is created from either data in the DICOM directory file (DICOMDIR) or from the DICOM objects that have been directly acquire via the network. The list is hierarchically organized. For each study in the list there is an entry for each DICOM series, and within each DICOM series there is an entry for each DICOM object.

The interactive study selection and order association process for both ordered and unordered studies is illustrated below in Figure 6 – Study Selection and Association.

### Study Selection and Association




Figure 6 – Study Selection and Association

### 3.1.1.1 Initial Checking of DICOM Studies

Prior to constructing the list of importable DICOM studies, those DICOM objects that have already been imported are identified and non-supported DICOM objects are flagged. (For more information, see section 3.4 Details of Study Checking at the end of this chapter.)

### 3.1.1.2 Importer Study Selection Screen

The Study Selection screen is presented to the user after checking has been performed and any associated warning messages have been displayed.

The following fields are displayed for each study:

- 1) Selection Number 1:n, identifies study so that it can be selected
- 2) Patient ID
- 3) Patient Name
- 4) Patient DOB
- 5) Patient Sex
- 6) Accession Number
- 7) Study Date
- 8) Study Description
- 9) Number of images, by modality

Figure 7 – Importer Study Selection Screen is an example, using the anonymized patients from the public domain American Dental Association (ADA) 2006 DICOM CD.

Import Studies	that were	Ordered by	y the VA and	d Contracted	to be	Performed	on the Outside

#	Patient ID	Patient Name	Birth	Sex	Accession Number	Date	Description	Images	
1)	000-00-0001		02/01/2002	M		02/20/05		DX=1	
2)	000-00-0001	OUTSIDE-1,PATIENT-1	02/01/2002			01/04/05		IO=1	
3)	000-00-0001	OUTSIDE-1,PATIENT-1	02/01/2002			07/03/05		IO=1	
4)	000-00-0001	OUTSIDE-1,PATIENT-1	02/01/2002			03/17/05		PX=1	
5)	000-00-0001	OUTSIDE-1,PATIENT-1	02/01/2002			06/22/06		XC=1	
6)	000-00-0002	OUTSIDE-2, PATIENT-2	01/01/1900		000000037	07/05/05	Bite Wing X-Ray	IO=2 PX=1	
7)	000-00-0003	OUTSIDE-3, PATIENT-3	12/24/1948	M		03/06/05	3 •	IO=3 XC=2	
8)	000-00-0003	OUTSIDE-3, PATIENT-3	12/24/1948	M		06/20/06		IO=4 XC=3	
9)	000-00-0003	OUTSIDE-3,PATIENT-3	12/24/1948	M		04/17/06		IO=3 PX=1 XC	C=3
10)	000-00-0004	OUTSIDE-4,PATIENT-4	08/04/2006	0		08/04/06		DX=1 IO=2 PX	X=1
11)	000-00-0005	OUTSIDE-5,PATIENT-5	05/13/2005	M	1	05/13/05	OCD Digital X-ray	DX=1	
12)	000-00-0006	OUTSIDE-6,PATIENT-6	05/13/2005	M	1	05/13/05	Sigma Digital X-ray	IO=1	
13)	000-00-0006	OUTSIDE-6,PATIENT-6	05/13/2005	M	1	05/13/05	Sigma Digital X-ray	IO=1	
14)	000-00-0007	OUTSIDE-7,PATIENT-7	12/05/2000	M	1	05/13/05	OPD Digital X-ray	PX=1	
15)	000-00-0007	OUTSIDE-7,PATIENT-7	12/05/2000	M	1	05/13/05	OPD Digital X-ray	PX=1	
16)	8000-00-000	OUTSIDE-8, PATIENT-8	01/01/2006	F	007	06/07/06	Sample for ADA 2006		
17)	8000-00-0008	OUTSIDE-8, PATIENT-8	01/01/2006	F	007	06/07/06	Sample for ADA 2006		
18)	8000-00-000	OUTSIDE-8, PATIENT-8	01/01/2006	F	007	06/07/06	Sample for ADA 2006		
19)	8000-00-000	OUTSIDE-8, PATIENT-8	01/01/2006	F	007	07/20/06	Sample for ADA 2006		
20)	000-00-0008	OUTSIDE-8, PATIENT-8	01/01/2006	F	007	07/20/06	Sample for ADA 2006		
21)	000-00-0009	OUTSIDE-9,PATIENT-9	11/02/1955	F	20	01/17/05	Cosmetic series vid		
22)	000-00-0009	OUTSIDE-9, PATIENT-9	11/02/1955	F	35	03/28/05	2 PA	IO=2	
23)	000-00-0009	OUTSIDE-9, PATIENT-9	11/02/1955	F	36	03/28/05	PA .	IO=1	
24)	000-00-0009	OUTSIDE-9, PATIENT-9	11/02/1955	F	18	01/17/05	Panoramic	PX=1	
25)	000-00-0010	OUTSIDE-10, PATIENT-10	02/14/1972		1432	04/15/05	ProMax Ceph	DX=1	
26)	000-00-0010	OUTSIDE-10, PATIENT-10	02/14/1972		1	04/15/05	Planmeca images	IO=2 PX=1	
27)	000-00-0011	OUTSIDE-11, PATIENT-11	08/07/1970			08/08/06		IO=1	
28)	000-00-0011	OUTSIDE-11, PATIENT-11	08/07/1970			08/08/06		IO=1	
29)	000-00-0011	OUTSIDE-11, PATIENT-11	08/07/1970			08/08/06		IO=1	
30)	000-00-0011	OUTSIDE-11, PATIENT-11	08/07/1970			08/08/06		IO=1	
31)	000-00-0012	OUTSIDE-12, PATIENT-12	01/01/2000	M		03/17/06	FULL MOUTH SERIES (	IO=1	
32)	000-00-0012	OUTSIDE-12, PATIENT-12	01/01/2000	M		03/03/06	FULL MOUTH SERIES (	IO=1	
33)	000-00-0012	OUTSIDE-12, PATIENT-12	01/01/2000	M		03/29/06	FULL MOUTH SERIES (	IO=1	
34)	000-00-0012	OUTSIDE-12, PATIENT-12	01/01/2000	M		06/14/06	PANO 1x2	PX=1	37 0
35)	000-00-0013	OUTSIDE-13, PATIENT-13	03/11/2005	0		08/04/06	ADA DICOM CD	DX=1 IO=1 PX	X=2
36) 37)	000-00-0014 000-00-0015	OUTSIDE-14, PATIENT-14	11/23/1963	0		12/16/05 10/20/05	Study 10/20/2005	IO=2 DX=1	
37)	000-00-0015	OUTSIDE-15, PATIENT-15 OUTSIDE-15, PATIENT-15	11/23/1963			07/06/06	Study 10/20/2005 Study 7/6/2006	ES=1	
39)	000-00-0015	OUTSIDE-15, PATIENT-15 OUTSIDE-15, PATIENT-15	11/23/1963			05/26/06	Study 7/6/2006 Study 5/26/2006	IO=1	
40)	000-00-0015	OUTSIDE-15, PATIENT-15 OUTSIDE-15, PATIENT-15	11/23/1963			10/05/05	Study 3/20/2006 Study 10/5/2005	PX=1	
40)	000-00-0013	OOIDIDE-ID, PAITENI-ID	11/43/1903	U		10/05/05	Scuay 10/5/2005	LV-T	

Please enter the next study to import or enter "C" to continue:

Figure 7 – Importer Study Selection Screen

### 3.1.1.3 Importer Study Selection Interaction

The user is prompted to enter a study number.

```
Please enter study to import or enter "C" to continue:
```

The user may either enter the number of a study (in this case 1 to 40) or may enter a "C" to conclude the manual study selection/order association step and continue with the automatic import reconciliation step.

If the user enters a "C" to continue with the next step, the user is asked if the DICOM objects are now to be imported.

```
Ready to import the DICOM objects? n // Y
```

If the user enters YES, the automatic importation process begins (see 3.1.4 for details). Control is passed back to the Importer menu when it is completed.

```
Ready to import the DICOM objects? n // n
```

If the user enters NO, the importation process is skipped and the user is asked if they wish to exit the Importer session.

```
Exit the application and not import the DICOM objects? n //
```

If the user enters YES, the Importer session is closed and control is passed back to the Importer menu. If the user answers NO, the Importer session continues and control is passed back to the Study Selection screen.

#### 3.1.1.4 Series and Image Selection and Display

Let's assume that the user selects Study 9 and wants to view the images.

When the study is selected, detailed study information is presented to the user along with three options, select the study, view images of the study, or skip the study and select another one.

The user is prompted to enter an action.

```
Is this the correct study/view images? [enter Y, N, R] v//
```

The default is "v" for view. If the user selects "n", control is passed back to the previous study selection screen without any further action. (The "selected study" is still available for selection.)

Let's assume that the user enters "v" for view. The following information is displayed:

If there are more than one series in the study, the user is prompted to enter the series number. (There is no prompt if there is only one series.)

```
There are 3 Series in this Study. Enter a series number:
```

This study contains three series, Series 1 with three External Camera (modality = XC) images, Series 2 with three Inter Oral (modality = IO) images, and Series 3 with one Panoramic x-ray image (modality = PX). The user may select any series and see its images. If the user enters a null series number, control is passed back to the previous "Is this the correct study/view images?" prompt.

Let's assume that the user entered "2" to see the images for Series 2. The following information would be displayed:

```
There are 3 Series in this Study. Enter a series number: 2
There are 3 Images in Series 2. Enter an image number:
```

If there is more than one image in the series, the user is prompted to enter the image number. (There is no prompt if there is only one image.)

```
There are 3 Images in Series 2. Enter an image number:
```

The user may select any image to view. If the user enters a null image number, control is passed back to the previous "Enter a series number:" prompt.

A sequence of images may be entered in <first image #> - <last image #> format, for example "1-3".

Let's assume that the user entered just "1" to see the first image.

```
There are 3 Images in Series 2. Enter an image number: 1
```

Control is immediately passed to the VistA DICOM Viewer (see Figure 8 – Dental Image).

The user can manipulate the image using the viewer or can switch between images if a sequence of images is selected. Control is passed back to the Importer when the user closes the viewer window. If there is more than one image for the series, the user is again presented with the "Enter an image number:" prompt and allowed to view additional images from the series.




Figure 8 – Dental Image

### 3.1.2 Associating the Selected Study with an Ordered Study

After the outside study has been selected is must be associated with an ordered study on VistA. This example illustrates obtaining this information from VistA using a DICOM Modality Worklist. (See the next section 3.1.3 for another way to associate an outside contracted radiology study with the order on VistA.)

When the user selects a study and enters "Y" at the "Is this the correct study/view images?" prompt, control will be passed to the Modality Worklist subroutine. (It is assumed that this is not a contracted radiology study and that menu option **2-12-2 Import Outside Contracted studies** is not being used.)

```
Please enter the next study to import or enter "C" to continue: 9

Patient Name: OUTSIDE-3,PATIENT-3 Patient ID: 000-00-0003 Birth Date: Dec 24, 1948

Sex: M

Accession #: Study Date: 04/17/06 Study Time: 01:00:00 am

Study Instance UID: 2.16.840.1.114059.1.19481224.20060417

Is this the correct study/view images? [enter Y, N, V] v// Y
```

The user must select the site's "Local Modality Worklist" as the DICOM Modality Worklist Query Service Class Provider. (The list itself will depend upon the entries in the site's SCU\_LIST.DIC.)

The user will then need to select the proper Application Entity Title for the study (as defined in WORKLIST.DIC.)

```
Select the Application Entity Title: RAD
```

The user should know the VistA patient id and accession number for the ordered study. Enter these as the Modality Worklist Attribute Key values. (The following example shows an accession number query.)

Enter the Accession Number or the Case Number: 040909-341

### Then enter "Q" to begin the query:

### SELECT MODALITY WORKLIST ATTRIBUTE KEYS PATIENT NAME (1):

PATIENT ID (2):

ACCESSION NUMBER (3) : 040909-341

REQUESTED PROCEDURE ID (4):

MODALITY (5):

START DATE (6):

START TIME (7):

Enter 1-7 to change an item above, "R" to refresh, "Q" to query: q Performing Query... Sending the PDU to the SOP Class Provider via TCP completed!

#### One match should be returned. Here is an example:

Patient Name: PATIENT, FOURONEZERO Race: <unknown> Patient Sex: M Status: OUTPATIENT Pregnancy Status: Not

Pregnant

Patient Identifier: 000-00-0410 ICN/MPI:

9810937512V654749

Admission ID: <unknown> Issuer of Patient ID: USVHA Patient Height: <unknown> Patient Weight: <unknown> Date of Birth: 1 January 1924
3: 605 MORSE STREET CT

Address: 605 MORSE STREET, SALT LAKE CITY, UT 32650

Confidentiality: <unknown> Institution: SALT LAKE CITY

Accession Number: 040909-341 Requested Proc ID: 341

VA Procedure Code: 58 Name: CHEST 2 VIEWS PA&LAT

CPT Code: 71020 Name: CHEST X-RAY

Scheduled: 9 April 2009 at 08:23:04 in RAD Status: SCHEDULED Requesting Physician: IMAGPROVIDERONETWOSIX, ONETWOSIX Priority: ROUTINE

Call Back Number: (301)734-0100
Requesting Service: CARDIOLOGY Attention: <unknown>

Referring Physician: IMAGPROVIDERONETWOSIX, ONETWOSIX

Study UID: 1.2.840.113754.1.40.660.6909590.9177.1.660.40909.341

Reason for Study: TESTING AGAIN

Reason for Study: TESTING AGAIN

MORE OF THE SAME

Study Comments:

If the user enters "No" at the "Is this the correct Patient and Study?" prompt, control will be returned to the Study Selection screen.

If the user enters "Yes" at the prompt, a verification screen will be presented.

Is this the correct Patient and Study? n// y

Is this the correct Patient and Study? n//

Is this the correct patient/study/procedure from VistA? N //

If the user enters "No" at the "Is this the correct patient/study/procedure from VistA?" prompt, control will be returned to the Study Selection screen.

If the user enters "Yes" at the prompt, s/he will be asked if a VA Radiologist will be reading the study (assuming that this is a radiology study and not a consult).

Is this the correct patient/study/procedure from VistA? N // y

```
Will a VA Radiologist read this study? N // N
```

If a VA Radiologist is going to read the study, the Exam Status will be set to "Examined" and the Report Status will be set to "No Report". (The study will also be placed on a VistARad reading worklist.)

If a VA Radiologist is not going to read the study, the Exam Status will be set to "Complete" and Report Status will be set to "Electronically Filed".

Control will then return to the Study Selection screen.

Each study that is to be imported is then identified with a lower-case green "i" on the Study Selection screen, like below:

### 3.1.3 Associating an Outside Contracted Study with the Order on VistA

Import DICOM Objects

- 1 Import DoD studies that were transmitted directly to the VA
- 2 Import Outside Contracted studies
- 3 CD/DVD Import Unordered studies from DoD facilities
- 4 (CD/DVD Import Unordered studies from other VA facilities)
- 5 CD/DVD Import Unordered studies from non-VA/DoD facilities
- 6 Correct local studies with bad PID and/or study information
- 7 Build Radiology Procedure File
- 8 Build Radiology Modifiers File
- 9 Modify Exam/Report Status Update Parameters

Contracted outside studies are entered using menu option **2-12-2 Import Outside Contracted studies**.

The order for a contracted radiology study can be obtained directly from the Radiology Order file (#75.1) on VistA. (See the previous section 3.1.2 for obtaining VistA patient and study information via DICOM Modality Worklist.) The radiology order on VistA will be automatically be registered, if need be, as part of the import reconciliation process. The autoregistration process will use the actual date and time of the study so that the VistA radiology order is created in the database in the chronologically correct sequence.

For a contracted radiology study, when the user selects a study and enters "Y" at the "Is this the correct study/view images?" prompt, s/he will be asked how to lookup the study on VistA.

```
Is this the correct study/view images? [enter Y, N, V] \, v// y Use Modality Worklist or look up in Radiology Order file r // ? Please enter "Modality Worklist", or "Radiology Order file". Use Modality Worklist or look up in Radiology Order file r //
```

If the user enters "m", control will be passed to the Modality Worklist subroutine, as described in the previous section 3.1.2. Let's assume instead that the user enters "R" (or took the default) so that the study will be selected from the Radiology Order file (#75.1).

In order to search the Radiology Order file, the user must first identify the VA patient. The patient is selected exactly as it is in CPRS. The following example shows the short id being used to pick the patient:

```
Enter VA Patient: P1023
```

Is this the correct patient? No // y

Once the patient is selected, the list of active (registered) and pending (not registered) studies is displayed (see Figure 9 – Radiology Order File Lookup Screen below). The user selects the one that corresponds to the contracted outside study.

The user is asked to verify the selection and then control is passed back to the study selection screen. The study that is to be imported will again be identified with a lower-case green "i" on the Study Selection screen.

			Radiology Orders PATIENT,ONEZEROTWOTHREE 000-00-1023		
#	Status	Ordered	Procedure Reason for Study	Accession #	
1)	ACTIVE	08/08/98	CHEST 2 VIEWS PA&LAT		
2)	ACTIVE	08/21/98	ABDOMEN 1 VIEW	660-070505-401	
3)	ACTIVE	08/21/98	HAND 1 OR 2 VIEWS	660-070505-409	
4)	PENDING	08/21/98	CHEST 2 VIEWS PA&LAT		
5)	ACTIVE	08/04/99	ABDOMEN 1 VIEW	660-070505-405	
6)	ACTIVE	01/23/01	ABDOMEN 1 VIEW	660-070505-406	
7)	PENDING	01/23/01	CHEST 2 VIEWS PA&LAT		
8)	ACTIVE	02/21/02	CHEST 2 VIEWS PA&LAT	022102-104	
9)	ACTIVE	01/13/04	CHEST SINGLE VIEW	011304-261	
10)	ACTIVE	04/29/05	CHEST 2 VIEWS PA&LAT	042905-262	
11)	ACTIVE	05/02/05	CHEST 2 VIEWS PA&LAT	050205-260	
12)	ACTIVE	12/27/04	CHEST 2 VIEWS PA&LAT	122704-263	
13)	ACTIVE	10/23/08	HAND 1 OR 2 VIEWS broken	102308-310	
14)	ACTIVE	10/23/09	CHEST 2 VIEWS PA&LAT TESTING NEW RADIOLOGY FUNCTION	ON 660-102309-390	
Please	e enter the	e correspon	nding study: 7		
Proced	#7 Status: PENDING Ordered: 01/23/01 04:28:35 pm Accession Number: Study Date: Procedure Name: CHEST 2 VIEWS PA&LAT Procedure Number: 58 Modifier(s):				
1.100111	CI (D).				

Is this the correct study? y // y

Reason for Study:

	Imported File	VistA
Patient Name:	DENTRIX, VIPERSOFT	PATIENT, ONEZEROTWOTHREE
Patient ID:	801-763-9300	000-00-1023
Date of Birth:	01/01/1900	02/04/1944
Patient Sex:	F	M
Accession Number:	00000037	(automatically generated)
Study Date:	07/05/05	(will be set to same date)
Procedure:	Bite Wing X-Ray	CHEST 2 VIEWS PA&LAT

Is this the correct patient/study/procedure from VistA? N // y

Figure 9 – Radiology Order File Lookup Screen

## 3.1.4 Importing the Ordered Study into VistA

When the user has selected all of the studies that need to be imported, s/he should enter a "C" to the "Please enter the next study to import or enter "C" to continue: "prompt. This will conclude study selection and proceed to the import reconciliation step.

The user will be asked if the DICOM objects are now to be imported.

```
Ready to import the DICOM objects? n // Y
```

If the user enters YES, the automatic importation process begins. Control is passed back to the Importer menu when it is completed.

(If the user enters NO, the importation process is skipped and s/he is asked if they wish to exit the Importer session or return to the Study Selection screen.)

Import reconciliation is entirely automatic. The studies are listed as they are being imported. Here is an example of the output:

The DICOM objects are updated with the correct patient and study information from VistA and are placed onto the Image Input queue of the Gateway. Menu option **2-3 Image Processing** must be run to load them into VistA.

Imported DICOM objects do not need to have corresponding entries in MODALITY.DIC. By default, they are loaded into VistA in DICOM format.

# 3.1.5 Deleting (not Importing) Transmitted DICOM Studies

Let's assume that a set of dummy images were acquired for a test study and were transmitted to the DICOM gateway. This test study appears on the list of importable DICOM studies and needs to be deleted. When the user selects a transmitted study, there are four options that can be used. In addition to the three options listed in previous section 3.1.1.4, there is also an option to delete the study. (This option is not necessary for DICOM studies being imported from portable media.)

```
Import DoD DICOM Studies that were Transmitted to the VA
    Patient ID
                        Patient Name
                                               Birth Sex Accession Number Date
    000-00-0016 OUTSIDE-16, PATIENT-16
                                             03/04/1960 F 030498-12345
Please enter the next study to import or enter "C" to continue: 1
Patient Name: OUTSIDE-16, PATIENT-16 Patient ID: 000-00-0016 Birth Date: Mar 4, 1960 Sex: F
Accession #: 030498-12345 Study Date: 03/19/97 Study Time: 11:47:19 am Referring
Physician: Physician, Referring
Study Instance UID: 2.16.840.1.113662.2.1.486876813333.429.33.100100
Sent by Calling AE Title: DICOM_TEST
Is this the correct study/view images (or delete the study)? [enter Y, N, V (or DEL)] v//?
Please enter one of the following options:
     Yes -- when you are ready to assign this study to a patient
     No -- if you want to skip this study and select a different one
    View -- when you want to display the DICOM objects for this study
  DELete -- should you want to delete this study (you must use "DEL")
Is this the correct study/view images (or delete the study)? [enter Y, N, V (or DEL)] v// del
```

Each study that is to be deleted is then identified with a lower-case green "d" on the Study Selection screen, like below:

Are you sure that you want to delete this study and not import it into VistA?  $n \ // \ y$ 

The final prompt indicates that DICOM objects will be deleted.

```
Ready to delete the DICOM objects? n // Y Deleting #1 .
```

Press <Enter> to continue...

The DICOM objects are removed from the C:\DICOM\IMAGE\_IN folder and from the list of importable DICOM studies.

## 3.2 Correcting Images with Erroneous Patient/Study Identification

Import DICOM Objects

- 1 Import DoD studies that were transmitted directly to the VA
- 2 Import Outside Contracted studies
- 3 CD/DVD Import Unordered studies from DoD facilities
- 4 (CD/DVD Import Unordered studies from other VA facilities)
- 5 CD/DVD Import Unordered studies from non-VA/DoD facilities
- 6 Correct local studies with bad PID and/or study information
- 7 Build Radiology Procedure File
- 8 Build Radiology Modifiers File
- 9 Modify Exam/Report Status Update Parameters

DICOM studies from local modalities that have failed to be associated because of erroneous patient and/or study identification can be corrected and imported by the above method. Use menu option 6 Correct local studies with bad PID and/or study information.

## 3.3 Importer Session – Unordered Study Workflow

An unordered study is one that is unknown to the local VistA system. It may have been performed while the patient was being treated at an outside facility and then was then sent to the local VA when the patient was going there for treatment. For example, a service person has some studies performed in the DoD and then comes to the VA for additional treatment. The DICOM images from the DoD need to be imported into VistA.

Import DICOM Objects

- 1 Import DoD studies that were transmitted directly to the VA
- 2 Import Outside Contracted studies
- 3 CD/DVD Import Unordered studies from DoD facilities
- 4 (CD/DVD Import Unordered studies from other VA facilities)
- 5 CD/DVD Import Unordered studies from non-VA/DoD facilities
- 6 Correct local studies with bad PID and/or study information
- 7 Build Radiology Procedure File
- 8 Build Radiology Modifiers File
- 9 Modify Exam/Report Status Update Parameters

There are four Importer options for the processing of unordered studies. The first option (#1) is used to import studies that were transmitted from the DoD to the VA, like for studies sent from National Naval Medical Center to a VA Polytrauma Center. The other three (#3, #4, and #5) are used to import DICOM objects from portabile media. The workflow for these four Importer options is identical. (The only differences are the source of the DICOM objects and that the Imaging file (#2005) field ORIGIN INDEX is "D" for DoD studies, "V" for VA ones and "N" for the other facilities.)

For the unordered study to be imported into VistA, a new VistA Radiology order needs to be created for it. The corresponding VistA Radiology procedure for the new order should be one that closely matches the procedure that was originally performed. This is necessary so that the study is properly indexed for later lookup. The Importer user needs to choose the appropriate VistA Radiology procedure. To be able to perform this function, the Importer user should be

familiar with radiology procedures in general and have a good working knowledge of VistA Radiology procedures.

When an unordered outside radiology study is to be imported into VistA, an order is automatically placed for it during the import reconciliation step. The VistA order created for the outside radiology study should not accrue workload statistics for VA technical or professional components because the exam was neither performed in the VA nor interpreted by the VA. For this reason, the order created on VistA should use a Radiology imaging location with a CREDIT METHOD = No Credit. This should have been setup by a Radiology/Nuclear Medicine ADPAC (see the Appendix A for details).

#### 3.3.1 Ordering Provider and Ordering Location

At the start of each unordered study Importer session, the user is required to designate an Ordering Provider and Ordering Location that is used for every VistA Radiology order that is created. The site needs to establish a policy for specifying these. One strategy might be (1) for a Polytrauma patient transferred from DoD to the VA, the Ordering Provider is the Chief of the Polytrauma Center and the Ordering Location is the Polytrauma Ward, and (2) for any other patient, the Ordering Provider is the Primary Care Provider and the Ordering Location is the clinic or ward where the patient is being treated.

Here is an example of the interaction:

## 3.3.2 Importer Study Selection

The study selection process for unordered studies is exactly like that for ordered studies (see previous section 3.1.1 for details).

## 3.3.3 Identifying the VistA Patient

Once the study has been chosen, the user must identify the VA patient for the study. The patient is selected exactly as it is in CPRS. The following example shows the short id being used to pick the patient:

Is this the correct patient? No // Y

#### 3.3.4 Selecting the Corresponding VistA Radiology Procedure

After the VA patient has been identified, the user is asked to select the equivalent VistA Radiology procedure that most closely corresponds to the one that was originally performed for the study. This may require the user to view the studies images in order to properly determine the appropriate VistA Radiology procedure. The user may also designate procedure modifier(s).

If the user has selected a VistA Radiology procedure, s/he will then be presented with a verification screen.

Is this the correct patient/study/procedure from VistA? N // Y

Control will then return to the Study Selection screen.

Each study that is to be imported is identified with a lower-case green "i" on the Study Selection screen, like below:

```
8)i 000-00-0003 OUTSIDE-3,PATIENT-3
```

#### 3.3.5 Importing the Unordered Radiology Study into VistA

The importation process for an unordered study is similar to that for an ordered study with one major difference: for the unordered study a corresponding study is automatically created on

VistA. The date of the new study on VistA is set to the date that the original study was performed so that it chronologically collates with the patient's other studies. The accession number will have the date of the original study.

Here is an example of the output:

# 3.3.6 Display of an "Unordered Study" in CPRS

CPRS displays the following information about an "Unordered Study" that has been imported:

- a) Procedure Date and Time of the VistA study is that of the original study
- b) Report Status is "Electronically File" and the Exam Status is "Complete"
- c) Clinical History contains the original data for the imported study, including the Patient Name, Patient ID, Birth Date, Sex, Accession #, Study Date, Study Time, Study Description, and Referring Physician
- d) A list of the DICOM Series, along with the modality and number of images in each

An example of the CPRS display of a VistA order that wa automatically created for an "Unordered Study" is shown in Figure 10 – CPRS Radiology Study Details.




Figure 10 – CPRS Radiology Study Details

#### 3.4 Deleting (not Importing) Transmitted DICOM Studies

Let's assume that a set of dummy images were acquired for a test study and were transmitted to the DICOM gateway. This test study appears on the list of importable DICOM studies and needs to be deleted.

When the user selects a transmitted study, there are four options that can be used. In addition to the three options listed in previous section 3.1.1.4, there is also an option to delete the study. (This option is not necessary for DICOM studies being imported from portable media.)

```
Import DoD DICOM Studies that were Transmitted to the VA
                         Patient Name
                                               Birth Sex Accession Number
    1) 000-00-0016 OUTSIDE-16, PATIENT-16 03/04/1960 F 030498-12345 03/19/97
Please enter the next study to import or enter "C" to continue: 1
Patient Name: OUTSIDE-16, PATIENT-16 Patient ID: 000-00-0016 Birth Date: Mar 4, 1960 Sex: F
Accession #: 030498-12345 Study Date: 03/19/97 Study Time: 11:47:19 am Referring
Physician: Physician, Referring
Study Instance UID: 2.16.840.1.113662.2.1.486876813333.429.33.100100
Sent by Calling AE Title: DICOM_TEST
Is this the correct study/view images (or delete the study)? [enter Y, N, V (or DEL)] v//?
Please enter one of the following options:
     Yes -- when you are ready to assign this study to a patient
     No -- if you want to skip this study and select a different one
    View -- when you want to display the DICOM objects for this study
  DELete -- should you want to delete this study (you must use "DEL")
Is this the correct study/view images (or delete the study)? [enter Y, N, V (or DEL)] v// del
Are you sure that you want to delete this study and not import it into VistA? n // y
```

Each study that is to be deleted is then identified with a lower-case green "d" on the Study Selection screen, like below:

```
1)d 000-00-0016 OUTSIDE-16,PATIENT-16
```

The final prompt indicates that DICOM objects will be deleted.

```
Ready to delete the DICOM objects? n // Y Deleting #1 .

Press <Enter> to continue...
```

The DICOM objects are removed from the C:\DICOM\IMAGE\_IN folder and from the list of importable DICOM studies.

# 3.5 Details of Study Checking

# 3.5.1 Flagging Non-supported DICOM Objects

Each DICOM object is checked against the list of VistA supported SOP Classes and Transfer Syntaxes (stored in SCP\_LIST.DIC). If a DICOM object has a non-VistA supported SOP Class and Transfer Syntax, the user is notified that the DICOM object cannot be imported. The following message will be displayed and the study will be flagged with a question mark ("?") on the Study Selection list.

## 3.5.2 Identifying Studies that have Already Been Imported

Each study's Study Instance UID is checked to determine if the DICOM study is on file on the VistA system. If it is, patient and study identification information for the matching study is retrieved from VistA. In addition, all of the DICOM objects for the study are checked to see if they exist on VistA as well. (This is done by checking if their SOP Instance UIDs are on VistA.)

The Importer application will notify the user when a Study Instance UID of a study to be imported is already on file on VistA. One of the following four situations may have occurred.

1) The study was previously imported and doesn't need to be imported again. (Perhaps the user tried to import the same DICOM objects from the CD a second time.) The study is flagged with a red lowercase "c" on the study selection list.

2) Part of the study has already been imported, while the rest of the study remains to be imported. (Perhaps some of the DICOM objects were transmitted earlier and were imported in a previous session.) The study is flagged with a yellow lowercase "p" on the study selection list.

```
*** There are nnn studies that are partially stored on VistA. ***

*** Each is flagged with a yellow lowercase "p" below. ***
```

- 3) The Study Instance UID is already on file for one or more different studies. The study is highlighted in yellow and flagged with an asterisk "\*" on the study selection list.
- 4) The Study Instance UID is in use for part of the study that has been imported and is also in use for one or more other studies. The study is highlighted in yellow and flagged with an asterisk "\*" on the study selection list.

**Note:** Try to avoid storing DICOM objects in VistA with a duplicate Study Instance UID, if at all possible. In scenario #3, the Importer will automatically replace the original Study Instance UID with a newly created one that is unique. In scenario #4, probably the best strategy would be to continue using the original UID for the DICOM objects that need to be import. Fortunately, the incidence of duplicate Study Instance UID is low, maybe on the order of one in a few thousand.

# Chapter 4 Import Reconciliation Process

Import reconciliation is a standardized process defined by the IHE Radiology Technical Framework (Import Reconciliation Workflow) by which outside DICOM objects have their patient and study information replaced with local system information in order to enable successful importation.

The following data elements in the header of each DICOM object are modified and then the object is placed on the image input queue for processing:

a) Replace the following data elements in the header of each imported DICOM object with values from VistA:

```
(0008,0050) Accession Number
(0010,0010) Patient Name
(0010,0020) Patient ID
(0010,0021) Issuer of Patient ID (always set to "USVHA")
(0010,0030) Date of Birth
(0010,0040) Sex
(0010,1000) Other Patient IDs (set to the VA Integration Control Number)
```

- b) Save the original values for these data elements in the Original Attributes Sequence (0040,0555).
- c) Add a Contributing Equipment Sequence (0018,A001) which includes the Importer User name in the Operator's Name (0008,1070) DICOM element.

Chapter 4 – Import Reconciliation Process

# Chapter 5 Importer Session Log Files

The activity of every Importer Session is record in a log file which can be easily imported into Microsoft Excel for viewing. The log file is located in the directory C:\DICOM\Importer\_Log. (The log files are chronologically named.)

To import this file into Microsoft Excel for viewing, use the following steps:

- Click on the log file (\*.CSV)
  - o MS Excel will automatically launch
- If you have MS Excel 2003
  - o Select all (ctrl-A)
  - o Format
    - Column
      - AutoFit Selection
- If you have MS Excel 2007
  - o Select all (ctrl-A)
  - o Then on the Home tab, in the Cells group, click on Format
  - o Under Cells Size, click on AutoFit Column Width

The Figure 11 – Importer Session Spreadsheet illustrates an Importer log in Excel 2007.




Figure 11 – Importer Session Spreadsheet

# Appendix A

# **Setting up Outside Imaging Locations for the DICOM Importer**

When an unordered prior outside radiology study is imported into VistA, an order is automatically placed for it. The VistA order created for the outside radiology study should not accrue workload statistics for VA technical or professional components because the exam was neither performed in the VA nor interpreted by the VA. For this reason, the order created on VistA should use a Radiology imaging location (entry in file #79.1) with a CREDIT METHOD = No Credit.

The VistA Imaging OUTSIDE IMAGING LOCATION file (#2006.5759) is used by the DICOM Importer to designate the entry in the Radiology IMAGING LOCATIONS file (#79.1) with CREDIT METHOD = No Credit that is to be used for the order created on VistA. There has to be one entry in the VistA Imaging OUTSIDE IMAGING LOCATION file for each enabled Radiology Imaging Type in each division.

The standard Radiology Package tools are used for creating the entries in the Radiology IMAGING LOCATIONS file. These include Parameter Set-up [RA SYSLOC] and Division Parameter Set-up [RA SYSDIV] options of the System Definition Menu [RA SYSDEF] . VistA Imaging provides the IMPORTER menu for constructing the OUTSIDE IMAGING LOCATION file.

A Radiology/Nuclear Medicine ADPAC should perform the modifications to Radiology Division and Imaging Locations files, as only they have the necessary radiology supervision menu privileges.

**Note**: If a site already has a No Credit outside location established for a Radiology Imaging Type for a division, it may be used instead of creating a new one.

# Step 1 – Identify Imaging Types and Divisions that need "No Credit" Imaging Locations

This step is performed by VistA Imaging site personnel.

Log into VistA and select the IMPORTER menu and the Check Outside Imaging Location file option. In this example there are two enabled Image Types (General Radiology and Nuclear Medicine) for the Salt Lake City Division.

```
DVA>D ^XUP

Setting up programmer environment
This is a TEST account.

Terminal Type set to: C-VT100
```

```
Select OPTION NAME: IMPORTER MENU MAG IMPORTER MENU Importer Menu

Build Outside Imaging Location file

Check Outside Imaging Location file

Display Studies to be Imported
```

```
Select Importer Menu Option: Check Outside Imaging Location file
```

```
Checking the Radiology files...
```

Division: SALT LAKE CITY

GENERAL RADIOLOGY - Define "No Credit" Imaging Location!

NUCLEAR MEDICINE - Define "No Credit" Imaging Location!

Please define missing "No Credit" imaging locations for the aforementioned divisions and imaging types (using the Location Parameter Set-up [RA SYSLOC] and Division Parameter Set-up [RA SYSDIV] options of the System Definition Menu ... [RA SYSDEF] options) and then run this option again.

Note that the SALT LAKE CITY Division has two Imaging Types (GENERAL RADIOLOGY and NUCLEAR MEDICINE) that need to have a "NO CREDIT" Imaging Location defined for them. The Radiology ADPAC will do this in Step 2 – Step 4.

If the site already has a No Credit imaging location already defined for the imaging type, the ADPAC does not have to create a new No Credit imaging location. The ADPAC should create the "OUTSIDE STUDY" Camera/Equip/Rm, however, and assign it to the No Credit imaging location.

# Step 2 – Create the OUTSIDE STUDY entry in the CAMERA/EQUIP/RM file

This step is performed by a Radiology ADPAC.

Select the RA SYSDEF menu. Edit the Camera/Equip/Rm parameters. Create a new OUTSIDE STUDY and enter the description "Study performed outside of this facility".

DVA>D ^XUP

Setting up programmer environment This is a TEST account.

Terminal Type set to: C-VT100

Select OPTION NAME: RA SYSDEF

System Definition Menu

← RA SYSDEF

Camera/Equip/Rm Entry/Edit Division Parameter Set-up List of Cameras/Equip/Rms Location Parameter List Location Parameter Set-up Print Division Parameter List

Select System Definition Menu Option: Camera/Equip/Rm Entry/Edit Select Camera/Equip/Room: ?

← CAMERA option

Answer with CAMERA/EQUIP/RM, or DESCRIPTION

Choose from:

PORTABLE MOBILE PICTURES

TD RAD-EXAM1 This is the primary x-ray room. TD RAD-EXAM2 Sonograms are done in this room. TD RAD-EXAM3 Endoscopies are done in this room.

You may enter a new CAMERA/EQUIP/RM, if you wish

Enter a name for this camera/equip/rm, between 1 and 30 characters in length.

Select Camera/Equip/Room: OUTSIDE STUDY

Are you adding 'OUTSIDE STUDY' as a new CAMERA/EQUIP/RM (the 5TH)? No// Y

CAMERA/EQUIP/RM DESCRIPTION: Study performed outside of this facility

CAMERA/EQUIP/RM: OUTSIDE STUDY//

DESCRIPTION: Study performed outside of this facility Replace

Select Camera/Equip/Room: ?

Answer with CAMERA/EQUIP/RM, or DESCRIPTION

Choose from:

OUTSIDE STUDY Study performed outside of this facility

MOBILE PICTURES PORTABLE

TD RAD-EXAM1 This is the primary x-ray room. TD RAD-EXAM2 Sonograms are done in this room. TD RAD-EXAM3 Endoscopies are done in this room.

You may enter a new CAMERA/EQUIP/RM, if you wish

## Appendix A – Setting up Outside Imaging Locations for the DICOM Importer

Enter a name for this camera/equip/rm, between 1 and 30 characters in length.

Select Camera/Equip/Room:

Camera/Equip/Rm Entry/Edit Division Parameter Set-up List of Cameras/Equip/Rms Location Parameter List Location Parameter Set-up Print Division Parameter List

Select System Definition Menu Option:

Do you really want to halt? YES//

# Step 3 – Add new Outside Imaging Locations for each Imaging Type to Divisions

This step is performed by a Radiology ADPAC.

For each division, enter the name of each outside imaging location, one for each enable Imaging Type for the Division. The suggested naming convention is "OUTSIDE <Division Mnemonic> <Imaging Type Abbreviation>".

# Naming Convention Example

In this fictitious example, let us say that the facility has the following Imaging Types enabled:

Imaging Type (from IMAGING TYPE file #79.2)	Abbreviation
GENERAL RADIOLOGY	RAD
NUCLEAR MEDICINE	NUC MED
ULTRASOUND	US
MAGNETIC RESONANCE IMAGING	MRI
CT SCAN	CT
ANGIO/NEURO/INTERVENTIONAL	ANGIO
CARDIOLOGY STUDIES (NUC MED)	CARD
VASCULAR LAB	VAS
MAMMOGRAPHY	MAM

Let us also say that the facility has two divisions "SALT LAKE CITY VAMC" and "CLEAR WATER LAKE VAMC". We chose "SLC" as the division mnemonic for the first, and "CWL" for the mnemonic for the second.

Here is the list of Outside Imaging Locations that we need to add to the system:

OUTSIDE SLC RAD OUTSIDE CWL RAD	
OUTSIDE SLC NUC MED OUTSIDE CWL NUC	MED
OUTSIDE SLC US OUTSIDE CWL US	
OUTSIDE SLC MRI OUTSIDE CWL MRI	
OUTSIDE SLC CT OUTSIDE CWL CT	
OUTSIDE SLC ANGIO OUTSIDE CWL ANGI	O
OUTSIDE SLC CARD OUTSIDE CWL CARD	)
OUTSIDE SLC VAS OUTSIDE CWL VAS	
OUTSIDE SLC MAM OUTSIDE CWL MAM	

End of Naming Convention Example.

In the rest of the examples, for sake of simplicity, there will only be the SALT LAKE CITY Division and two IMAGE TYPES: GENERAL RADIOLOGY and NUCLEAR MEDICINE.

Select the RA SYSDEF menu. Select a Division. Take the defaults until you get to the "Imaging Locations Associated with this Division:" section. Enter the name of each outside imaging location, one for each enable Imaging Type for the Division.

In this example we are creating the Imaging Locations "OUTSIDE SLC RAD" and "OUTSIDE SLC NUC MED" (the others are created the same way). For each location, select HOSPITAL LOCATION TYPE: CLINIC, HOSPITAL LOCATION TYPE EXTENSION: CLINIC//, and the appropriate Imaging Type.

```
DVA>D ^XUP

Setting up programmer environment
This is a TEST account.

Terminal Type set to: C-VT100
```

Select OPTION NAME: RA SYSDEF

System Definition Menu

← RA SYSDEF

← DIVISION option

← Station Number

Camera/Equip/Rm Entry/Edit Division Parameter Set-up List of Cameras/Equip/Rms Location Parameter List Location Parameter Set-up Print Division Parameter List

Select Division: 660 SALT LAKE CITY UT

ALLOW BATCHING OF REPORTS: yes// ALLOW COPYING OF REPORTS: yes//

IMPRESSION REQUIRED ON REPORTS: yes//
ALLOW VERIFYING BY RESIDENTS: yes//
ALLOW RPTS ON CANCELLED CASES?: no//
WARNING ON RPTS NOT YET VERIF?: yes//

Select System Definition Menu Option: DIVision Parameter Set-up

VistA Imaging System V.3.0, Patch 53

```
AUTO E-MAIL TO REQ. PHYS?:
ALLOW E-SIG ON COTS HL7 RPTS: no//
INTERPRETING STAFF REQ'D?: YES//
Miscellaneous Division Parameters:
______
PRINT FLASH CARD FOR EACH EXAM: no//
PRINT JACKET LBLS W/EACH VISIT: no//
CONTRAST REACTION MESSAGE:
RPHARM DOSE WARNING MESSAGE:
 No existing text
 Edit? NO//
HL7 Applications Associated with this Division:
______
Select HL7 RECEIVING APPLICATION:
Imaging Locations Associated with this Division:
_____
Select IMAGING LOCATION: TD-MAINRAD// OUTSIDE SLC RAD
                                                               ← 1<sup>st</sup> New Outside
 Are you adding 'OUTSIDE SLC RAD' as
   a new HOSPITAL LOCATION? No// Y (Yes)
                                                                  Location
  HOSPITAL LOCATION TYPE: ?
    Choose from:
      C
              CLINIC
      Μ
              MODULE
      W
              WARD
      Ζ
               OTHER LOCATION
               NON-CLINIC STOP
      F
              FILE AREA
      Т
              IMAGING
      OR
              OPERATING ROOM
                                                               ← Hospital Location
  HOSPITAL LOCATION TYPE: C CLINIC
  HOSPITAL LOCATION TYPE EXTENSION: CLINIC//
                                                               Type and Extension
  Are you adding 'OUTSIDE SLC RAD' as
                                                                ← New IMAGING
   a new IMAGING LOCATIONS (the 3RD)? No// Y (Yes)
                                                                LOCATIONS file
   IMAGING LOCATIONS TYPE OF IMAGING: ?
       Enter an imaging type for this location.
   Answer with IMAGING TYPE TYPE OF IMAGING, or ABBREVIATION
   Choose from:
   ANGIO/NEURO/INTERVENTIONAL
  CARDIOLOGY STUDIES (NUC MED)
   CT SCAN
  GENERAL RADIOLOGY
  MAGNETIC RESONANCE IMAGING
  MAMMOGRAPHY
  NUCLEAR MEDICINE
  ULTRASOUND
  VASCULAR LAB
  IMAGING LOCATIONS TYPE OF IMAGING: GENERAL RADIOLOGY
                                                          ← Imaging Type
  Are you adding 'OUTSIDE SLC RAD' as
   a new IMAGING LOCATIONS (the 3RD for this RAD/NUC MED DIVISION)? No// Y
  (Yes)
                                                                    \leftarrow 2^{ND} New
Select IMAGING LOCATION: OUTSIDE SLC RAD// OUTSIDE SLC NUC MED
  Are you adding 'OUTSIDE SLC NUC MED' as
                                                                   Outside Location
```

#### Appendix A – Setting up Outside Imaging Locations for the DICOM Importer

a new HOSPITAL LOCATION? No// Y (Yes)
HOSPITAL LOCATION TYPE: C CLINIC
HOSPITAL LOCATION TYPE EXTENSION: CLINIC//
Are you adding 'OUTSIDE SLC NUC MED' as
a new IMAGING LOCATIONS (the 4TH)? No// Y (Yes)
IMAGING LOCATIONS TYPE OF IMAGING: NUCLEAR MEDICINE

\*\* Caution: You are activating a new Imaging Type. \*\*
This means you will have to assign procedures to
this imaging type. Workload reports will be printed
separately for this Imaging Type.

Are you sure? YES

Are you adding 'OUTSIDE SLC NUC MED' as a new IMAGING LOCATIONS (the 4TH for this RAD/NUC MED DIVISION)? No//  ${\bf Y}$  (Yes)

Select IMAGING LOCATION:
Division Parameters have been set!

Imaging Location TD-RAD is OK.

Imaging Location TD-MAINRAD is OK.

Imaging Location file #79.1 entry OUTSIDE SLC RAD has a missing or invalid DSS ID. The Radiology/Nuclear Medicine ADPAC should use the Location Parameter Set-up [RA SYSLOC] option to enter a valid imaging DSS Code for this imaging location.

Imaging Location file #79.1 entry OUTSIDE SLC NUC MED has a missing or invalid DSS ID. The Radiology/Nuclear Medicine ADPAC should use the Location Parameter Set-up [RA SYSLOC] option to enter a valid imaging DSS Code for this imaging location.

Select Division:

#### **Step 4 – Define Outside Imaging Locations Parameters**

This step is performed by a Radiology ADPAC.

Use the Location Parameter Set-up menu option to specify the CAMERA/EQUIP/RM, and set the CREDIT METHOD to "No Credit" and the DSS ID to "X-RAY".

DVA>D ^XUP

Setting up programmer environment This is a TEST account.

Terminal Type set to: C-VT100

Select OPTION NAME: RA SYSDEF

System Definition Menu

← RA SYSDEF

Camera/Equip/Rm Entry/Edit Division Parameter Set-up List of Cameras/Equip/Rms Location Parameter List Location Parameter Set-up Print Division Parameter List

Select System Definition Menu Option: LOC

- 1 Location Parameter List
- 2 Location Parameter Set-up

CHOOSE 1-2: 2 Location Parameter Set-up

← Select Location Set-up option

Select Location: **OUTSIDE SLC RAD**...OK? Yes// Y (Yes)

(GENERAL RADIOLOGY-660)

← Select 1<sup>st</sup> location to update

Imaging Location: OUTSIDE SLC RAD

Flash Card Parameters:

-----

HOW MANY FLASH CARDS PER VISIT: DEFAULT FLASH CARD FORMAT:

No default flash card printer has been assigned. Contact IRM.

Jacket Label Parameters:

-----

As specified in the division parameters, no jacket labels will be printed automatically upon a patient visit.

DEFAULT JACKET LABEL FORMAT:

No default jacket label printer has been assigned. Contact IRM.

Exam Label Parameters:

-----

HOW MANY EXAM LABELS PER EXAM: 1//

#### Appendix A – Setting up Outside Imaging Locations for the DICOM Importer

#### DEFAULT EXAM LABEL FORMAT:

Exam label printer is always the same as the flash card printer.

# Order Entry Parameters:

No default request printer has been assigned. Contact IRM.

#### Report Parameters:

\_\_\_\_\_\_

DEFAULT REPORT HEADER FORMAT:
DEFAULT REPORT FOOTER FORMAT:
REPORT LEFT MARGIN: 10//
REPORT RIGHT MARGIN: 70//
PRINT DX CODES IN REPORT?:
VOICE DICTATION AUTO-PRINT:

No default report printer has been assigned. Contact IRM.

#### Cameras/Equip/Rooms Used by this Location:

\_\_\_\_\_

Select CAMERA/EQUIP/RM: ?
 '^' TO STOP:

You may enter a new CAMERAS/EQUIP/RMS, if you wish Enter all the cameras/equip/rms for this location.

Answer with CAMERA/EQUIP/RM, or DESCRIPTION

Choose from:

OR1 OR RM 1 OR2 OR RM 2

PORTABLE MOBILE PICTURES

RM0 2150 RM2 RM 2151 RM3 CHEST UNIT

OUTSIDE STUDY Study performed outside of this facility

TD RAD-EXAM1 This is the primary x-ray room.

TD RAD-EXAM2 Sonograms are done in this room.

TD RAD-EXAM3 Endoscopies are done in this room.

You may enter a new CAMERA/EQUIP/RM, if you wish Enter a name for this camera/equip/rm, between 1 and 30 characters in length.

Select CAMERA/EQUIP/RM: **OUTSIDE STUDY** Study performed outside of this facility

Are you adding 'OUTSIDE STUDY' as a new CAMERAS/EQUIP/RMS (the 1ST for this IMAGING LOCATIONS)? No//  $\bf Y$  (Yes) Select CAMERA/EQUIP/RM:

Default CPT Modifiers used by this Location:

-----

Your entry cannot be compared with a CPT CODE, so be very sure | that this is the Default CPT Modifier that you want to stuff

```
into every registered exam from this imaging location.
  +-----
Select DEFAULT CPT MODIFIERS(LOC): ?
    You may enter a new DEFAULT CPT MODIFIERS(LOC), if you wish
    VA local modifiers are automatically screened out (rejected.)
  +----+
    Your entry cannot be compared with a CPT CODE, so be very sure
   that this is the Default CPT Modifier that you want to stuff
   into every registered exam from this imaging location.
  +-----+
    Choose a CPT Modifier that should be automatically stuffed
    into the exam record, when the following 2 conditions
    are both met:
      1-There is no default CPT Modifier for this exam's procedure.
      2-This location is the current sign-on (or switched-to) location
       at the time of registration.
    If your entry is invalid, then during exam registration, this
    Default CPT Modifier will NOT be stuffed, instead, an error message
    with the name of the rejected CPT Modifier would be displayed.
Answer with CPT MODIFIER, or NAME, or CODE, or BEGIN CPT RANGE, or
    END CPT RANGE
Do you want the entire CPT MODIFIER List? N
Select DEFAULT CPT MODIFIERS(LOC):
Recipients of the 'Stat' Alert for this Location:
_____
Select STAT REQUEST ALERT RECIPIENTS:
ALLOW 'RELEASED/NOT VERIFIED': no// no
INACTIVE:
TYPE OF IMAGING: GENERAL RADIOLOGY//
CREDIT METHOD: 0// ?
    Enter the type of credit this location will receive for this
examination.
    Choose from:
      0
            Regular Credit
                                                          ← CREDIT METHOD =
             Interpretation Only
         No Credit
                                                              2 "No Credit"
      2
            Technical Component Only
CREDIT METHOD: 0// 2 No Credit
DSS ID: ?
    Enter a valid stop code for this imaging location.
    Only Stop Codes entered in the Imaging Stop Codes file may be selected.
Answer with CLINIC STOP NAME, or AMIS REPORTING STOP CODE
Do you want the entire CLINIC STOP List?
                105
DSS ID: X-RAY
                                   ← DSS ID = "X-RAY"
Imaging Location OUTSIDE SLC RAD is OK.
      _____
```

July 2010

Update the other locations in exactly the same fashion.

## Appendix A – Setting up Outside Imaging Locations for the DICOM Importer

#### Select Location:

Camera/Equip/Rm Entry/Edit Division Parameter Set-up List of Cameras/Equip/Rms Location Parameter List Location Parameter Set-up Print Division Parameter List

Select System Definition Menu Option:

Do you really want to halt? YES//

# Step 5 - "OUTSIDE STUDY" Camera/Equip/Room for all Radiology Imaging Location

This step is performed by a Radiology ADPAC.

Every Radiology Imaging Location that is used for a contract outside imaging facility needs to have "OUTSIDE STUDY" camera/equipment/room assigned to it. When the study is imported, the examination/report status of the case is advanced to either "Examined/No Report" if it is to be read locally or "Complete/Electronically Filed" if it is finished.

If DICOM objects are imported for a study with a Radiology Imaging Location that does not have the "OUTSIDE STUDY" camera/equipment/room assigned to it, a message like the one below will be displayed in the Importer session. The examination/report status will then need to be manually advanced.

# Step 6 – Populate VistA Imaging Outside Imaging Location file

This step is performed by VistA Imaging site personnel.

Log into VistA and select the IMPORTER menu and the Build Outside Imaging Location file option. In this example there are two enabled Image Types (General Radiology and Nuclear Medicine) for the Salt Lake City Division. Both of these now have CREDIT METHOD = "No Credit" entries in the Radiology DIVISION and IMAGING LOCATIONS file. The Build Outside Imaging Location file option will interactively help you add/change entries in the file.

When you have finished, run the Build Outside Imaging Location file option to verify that everything is setup correctly.

```
DVA>D ^XUP
Setting up programmer environment
This is a TEST account.
Terminal Type set to: C-VT100
Select OPTION NAME: IMPORTER MENU MAG IMPORTER MENU Importer Menu
          Build Outside Imaging Location file
          Check Outside Imaging Location file
          Display Studies to be Imported
Select Importer Menu Option: Build Outside Imaging Location file
Checking the Radiology files...
Division: SALT LAKE CITY
Checking the OUTSIDE IMAGING LOCATION file (#2006.5759)...
Division: SALT LAKE CITY
 GENERAL RADIOLOGY - Create record in file #2006.5759!
NUCLEAR MEDICINE - Create record in file #2006.5759!
Please fix the aforementioned problems in the OUTSIDE IMAGING LOCATION file
(#2006.5759) and then run this option again.
OUTSIDE IMAGING LOCATIONS for SALT LAKE CITY (660)
______
GENERAL RADIOLOGY -- (not defined yet)
                                          ← OUTSIDE SLC RAD
OUTSIDE SLC RAD
Use this value?? n// YES
NUCLEAR MEDICINE -- (not defined yet)
                                          ← OUTSIDE SLC NUC MED
OUTSIDE SLC NUC MED
Use this value?? n// YES
          Build Outside Imaging Location file
```

Check Outside Imaging Location file Display Studies to be Imported

58

Select Importer Menu Option: Check Outside Imaging Location file

Checking the Radiology files...

Division: SALT LAKE CITY

Checking the OUTSIDE IMAGING LOCATION file (#2006.5759)...

Division: SALT LAKE CITY
GENERAL RADIOLOGY - OUTSIDE SLC RAD NUCLEAR MEDICINE - OUTSIDE SLC NUC MED

> Build Outside Imaging Location file Check Outside Imaging Location file Display Studies to be Imported

Select Importer Menu Option:

Do you really want to halt? YES//

# Appendix B

# Running the Importer from a Remote Desktop<sup>3</sup>

This Appendix describes the initiation and configuration of Terminal Services on a DICOM Gateway running Microsoft Windows Server 2003 and Microsoft Remote Desktop (RDP) client on Microsoft Windows XP.

The Importer application was written to run on a DICOM Image Gateway. The Microsoft Remote Desktop (RDP) Terminal Service client was used during internal testing for the sake of convenience so that the tester could exercise the Importer from his office when it was running on a Gateway located in the computer room. The operation of the Importer application is the same whether it runs locally or remotely. The only difference is that the DICOM Gateway needs to map the CD/DVD drive on the remote system and that the Importer User must run the Remote Desktop application supplied with the operating system.

**Note**: Imaging support personnel at several sites have found it easier to install a dedicated DICOM Gateway in Radiology for the purpose of running the Importer application.

#### Microsoft Terminal Server Installation

#### Licensing the Terminal Server on Windows Server 2003<sup>4</sup>

Administrator Remote Desktop permits members of the administrator group to remotely access a Windows Server 2003. Each Windows Server 2003 has a license for one administrator RDP connection. This is the license that is typically used for running the Importer Application from a remote workstation.

Terminal Server is the same product with a multi-user license that allows simultaneous access by multiple remote clients to Windows-based programs that run on the server. When you use Terminal Server, the server accepts more than two simultaneous connections by non-administrators. When you use Terminal Server, you can install the Terminal Services Licensing service on any member server. However, you must configure a preferred license server on all terminal servers that have to communicate with non-domain controller license servers that are configured as domain license servers. Enterprise domain license servers that are deployed on non-domain controllers are automatically discovered.

http://www.microsoft.com/windowsserver2003/techinfo/overview/quickstart.mspx#EPC

<sup>&</sup>lt;sup>3</sup> This procedure was provided by Morgan Hammac of the Fayetteville, NC VAMC

<sup>&</sup>lt;sup>4</sup> Guidelines for Deploying Terminal Server,

## **Installing Terminal Server**

- 1. Install Terminal Server on an Importer DICOM Gateway running Windows Server 2003.
  - a. Click Start, click Control Panel, and then double-click Add or Remove Programs.
  - b. Click Add/Remove Windows Components.
  - c. In the dialog box, select **Terminal Server**.
  - d. Click Next. Terminal Server will install
  - e. Follow the instructions in the Windows Components Wizard.

Three questions will be asked about: "Full Security or Relaxed Security", "Location of License Server", and "Client Access License". Use the configuration options that are standard for your site.




Figure 12 – Enabling Terminal Server on Add/Remove Windows Components

- 2. Add the Importer Users to the administrator group on the Importer DICOM Gateway.
  - a. Click Start, click Control Panel, click Administrative Tools, and click Computer Management.
  - b. Add the account of each Importer User to the administrator group.




Figure 13 – Adding Import Users to the Administrator Group

# Setup the Remote Desktop Client on a Windows XP Workstation

Perform the following procedure on each client workstation that will be used to remotely run the Importer application.

- 1. Open the Remote Desktop Connection properties window.
  - a. Click **Start**, click **All Programs**, click **Accessories**, and click **Remote Desktop Connection**. (You may have to install this on the system(s) by local OI&T rules.)




Figure 14 – Starting Remote Desktop Connection



Figure 15 – Specifying the Importer DICOM Gateway Terminal Server

2. Click on **Options** >> and the **General** tab. Enter the IP address of an Importer DICOM Gateway computer, enter the User name and click on **Save**.



Figure 16 – Entering User name for Remote Desktop Connection

3. Click on the **Display** tab. Set the Remote desktop size to 1280 by 1024 pixels and Colors to Highest Quality (32 bit). Check the box for "Display the connection bar when in full screen mode". Note: The Windows XP workstation MUST be able to support 1280 by 1024 pixels and 32 bit color.




Figure 17 – Setting the Screen Resolution and Color Quality

4. (Skip **Local Resources** and **Programs**.) Click on the **Experience** tab. Check all the boxes. (Skip **Advanced**.)




Figure 18 – Setting Network Performance Parameters

5. Close the Remote Desktop Connection window by clicking on in the upper-right corner of the window.

### Division of Labor between local Gateway User and Remote User

The local Gateway User has the "FULL" (MAG DICOM GATEWAY FULL) key while the Remote User (Importer User) has the "VIEW" (MAG DICOM GATEWAY VIEW) key. It is the responsibility of the local Gateway User to start the normal continuously running processing jobs on the gateway, like the MAG\_CSTORE.EXE Storage SCP listeners and the menu option 2-3 Image Processing. The Remote User does not see these processes and with the "VIEW" key, does not have the menu options that are used to control them. S/he only sees the ones that s/he has initiated, in particular the Caché Terminal Importer Session, the DICOM Viewer, and the Laurel Bridge DCF executables.

### **Using the Remote Desktop Connection for the Importer**

1. Click Start, click All Programs, click Accessories, and click Remote Desktop Connection.




Figure 19 – Starting Remote Desktop Connection

2. Enter the name of the Importer DICOM Gateway and click on Connect.



Figure 20 – Specifying the Importer DICOM Gateway Terminal Server

3. The Importer User remote logons to the Importer DICOM Gateway using his/her normal credentials.



Figure 21 – Logging into Terminal Server on Importer DICOM Gateway

- 4. The Importer User needs to map a network drive from the Importer DICOM Gateway to his/her workstation's CD/DVD drive. If multiple people are Importer Users, each should map his/her workstation's CD/DVD drive using a unique drive letter. Enter this drive letter at the **CD/DVD Import Drive:** prompt.
- 5. Follow the Importer instructions in the body of this document.




Figure 22 – Running the Importer Remotely

6. When the Importer User finishes the Importer Session s/he should "Log Off" the Remote Desktop Connection. This will discontinue all the processes. Otherwise the processes will remain active.

# Appendix C

## Handing Parent-Descendent Procedure Orders with the Importer<sup>5</sup>

VistA Radiology supports the ability to order *parent* procedures. Parent procedures are associated with two or more *descendent* procedures. When the patient is registered for the parent procedure, a separate distinct radiology study with its own unique accession number is created for each descendent procedure.

An example of a parent procedure may be "ANKLES, BILATERAL, 2 VIEWS", which is associated with two descendant procedures: "LEFT ANKLE 2 VIEWS" and "RIGHT ANKLE 2 VIEWS". When the patient is registered for ANKLES, BILATERAL, 2 VIEWS, two studies are created, one for LEFT ANKLE 2 VIEWS and one for RIGHT ANKLE 2 VIEWS.

#### **How it Works**

The provider orders the parent procedure through CPRS. This goes into the ORDER file (#100) and when signed the same order is filed in the RAD/NUC MED ORDERS file (#75.1). When this order is registered (for example, by the Importer or the "REG EDIT" menu option) individual studies are created for each descendent procedure and are filed into the EXAMINATIONS (#70.03) multiple which is a sub-file of the RAD/NUC MED PATIENT file (#70). Unique accession numbers are created for the study for each procedure. The study for each procedure is then treated as a separate entity for examination purposes.

Each study for each procedure may have its own unique diagnostic report (that is, be defined an Exam Set – the default) or there may be a single diagnostic report referenced by the descendant procedures (that is, be defined a Print Set). A parent procedure can be designated as a print set by exercising the *Procedure Enter/Edit* option and setting the SINGLE REPORT field (#18) of the parent procedure in the RAD/NUC MED PROCEDURES file (#71) to 'Yes'.

#### Questions

If the VA orders a parent procedure as a fee-basis contracted examination, how does the contractor perform the procedure, how are the resultant DICOM objects created, and how should those DICOM object be imported into VistA?

<sup>&</sup>lt;sup>5</sup> Greg Cebelinski of the Radiology Development Team and Jeannie Jernigan of the Fayetteville, NC VAMC contributed to this document.

### **How Outside Fee-basis Imaging Contractors Work**

The outside imaging contractor can execute an order of "ANKLES, BILATERAL, 2 VIEWS" two different ways:

- The contractor can perform two separate studies, one for the left ankle and one for the right ankle.
- The contractor can also perform a single study which includes both ankles.

Depending upon how the contractor executes the order, there are three different ways to represent the DICOM images for the examination(s).

- 1) If the contractor performs the bilateral ankle order as two separate studies, one for the left ankle and the other for the right ankle, then there will be two different contractor assigned accession numbers, Study Instance UIDs, procedure names, etc. for each study. Each DICOM image must have either the series Laterality (0020,0060) or the Image Laterality (0020,0062) attribute.
- 2) If the contractor performs the bilateral ankle order as one study, there will be one contractor assigned accession number, Study Instance UID, procedure name, etc. for all the DICOM images.
  - a. The contractor may choose to use two separate series, one for the left ankle and the other for the right ankle, and specify the series Laterality (0020,0060) for all of the DICOM images in at series. In this case, each series will have a unique Series Instance UID.
  - b. The contractor may also chose to use one series for all of the ankle images, and specify the Image Laterality (0020,0062) for each DICOM image in the series. In this case, there will be only one Series Instance UID for the entire study.

### How to use the Importer for DICOM Images for Parent/Descendent Procedures

The choice of the method used by the contractor to perform the examination(s) and how the resultant images are represented in DICOM have a great bearing on how the images should be imported into another system.

The Importer does not handle parent/descendent procedures automatically. Some manual effort is required to do this properly. The level of effort depends on how the study was performed and how the DICOM images were created. The following steps may be used as a guideline:

1) Use Importer menu option 2 – Import Outside Contracted studies.

- 2) Using the Importer, determine whether contractor chose to perform the bilateral ankle as one study or as two separate studies.
- 3) If the contractor chose to perform the bilateral ankle as two separate studies, then the images from the left outside study will be imported to the left VistA study and the images from the right side outside study will be imported to the right side VistA study. Follow these steps:
  - a. Select only the LEFT study in the Importer.
  - b. Use the "lookup in Radiology Order file" option.
  - c. Select the patient and the order for the *parent* procedure.
  - d. Import only the LEFT study.
  - e. Exit the Importer application.
  - f. Using the VistA Rad application, the Radiology package to view by case, or CPRS Imaging (local only) Reports tab to confirm that second study for the right side study was created as a result of the bilateral study order (parent/descendant). This study will be in a "waiting for exam" status. Obtain the accession number for it.
  - g. Use Importer menu option 2 Import Outside Contracted studies, and re-enter the Importer application.
  - h. Select the RIGHT side study in the Importer.
  - i. Use the "Modality Worklist" look up option.
  - j. Select the VistA RIGHT side procedure study for the patient by entering its accession number for the query key.
  - k. Import the RIGHT side study.
  - 1. Confirm that the RIGHT side outside study's images are attached to the VistA RIGHT side study and that the study has been completed (Exam Status = "Exam Complete" and Report Status = "Electronically Filed").
- 4) If the contractor performed the bilateral ankle order as one study, there will be only one accession number, Study Instance UID, procedure name, etc. for all the DICOM images. All of the outside images will be imported to the LEFT VistA study. Follow these steps:
  - a. Select the study in the Importer.
  - b. Use the "lookup in Radiology Order file" option.
  - c. Select the patient and the order for the *parent* procedure.
  - d. Import the study.
    - i. The Importer will register the parent procedure and automatically create a study for the LEFT ANKLE 2 VIEWS procedure and a study for the RIGHT ANKLE 2 VIEWS procedure.

- ii. The accession number for the study for the LEFT ANKLE 2 VIEWS procedure will be returned and all the images will be imported to it.
- e. In the Radiology package, insert your local combined standard report stub to indicate images and report have been uploaded to the LEFT study. Then update the status manually in the Radiology package per local policy/procedures.

It is beyond the scope of the Importer to be able identify which images are for the left ankle and import them to the left procedure and identify which images are for the right ankle and import them to the right side procedure. There are two facets of this scenario that are difficult. First, the Importer works at the study level and not the series or image level. Second, DICOM images from one study cannot be filed with two different studies in VistA because this would create duplicate Study Instance UIDs and possibly duplicate Series Instance UIDs (if all the images were in the same series.) This would be a violation of the DICOM Standard and not a limitation of the Importer or VistA Imaging.

**Note**: It is advantageous to require that the outside contracted vendor to perform the bilateral ankle as two separate studies. Then the images can be properly associated with the corresponding two VistA studies.

# Appendix D

### Problems Seen in Importing of DICOM Objects from Media

### **Background**

The DICOM Standard<sup>6</sup> and the IHE Portable Data for Imaging integration profile<sup>7</sup> specify the format and content of media (CDs and DVDs) that is used for interchange of DICOM objects. These specifications cover the DICOM directory file DICOMDIR, the DICOM objects, and additional material on the media (viewers, JPEG files, etc.). It is sad commentary on the state of the art, however, that some vendors still create proprietary media with proprietary viewers on board, despite the fact that these offer no advantage over standard formatted media. This is a major problem in the industry today.

The physician user community is starting to take action about the difficulty viewing and navigating through imaging examinations from some media. The following is an excerpt from a published write up by DICOM expert Dr. David Clunie:<sup>8</sup>

"The American Medical Association, as a consequence of complaints initiated by the American Association of Neurological Surgeons with respect to viewing MRIs, has produced a report from their board of trustees that resulted in the passage of Resolution 539 (A-06) "Development of Standards for MRI Equipment and Interpretation to Improve Patient Safety". Note that the discontent being expressed by the AMA is not confined to neurosurgeons, but involves everyone who receives these media. Further, the emphasis is on safety, specifically, if the media is unreadable or unusable or takes too long to use, then the safety of the patient may be at risk."

"Referring physicians made it abundantly clear that swift, dramatic and effective action by industry and by radiology facilities is expected without delay, and that delay will result in engagement of the regulators and the legislators."

<sup>&</sup>lt;sup>6</sup> PS 3.3-2009 Information Object Definitions, PS 3.10-2009 Media Storage and File Format for Media Interchange and PS 3.11-2009 Media Storage Application Profiles

<sup>&</sup>lt;sup>7</sup> Integrating the Healthcare Enterprise, IHE Radiology Technical Framework, Volume III Transactions (Continued) Revision 9.0 – Final Text, June 27, 2008, Section 4.47 Distribute Imaging Information on Media. http://www.ihe.net/Technical Framework/upload/ihe tf rev9-0ft vol3 2008-06-26.pdf

<sup>8</sup> http://www.dclunie.com/blog/blog/2008/08/is-winter-of-discontent-with-cds.html

"Emphasis would be placed on establishing that the standard of care will be compliance with the IHE PDI specification, and in the absence of any explicit enforcement mechanism, promulgating this as an AMA principle may suffice. Woe betide anyone who then expects to get paid for producing non-compliant media (since payers might not pay for less than the standard of care when they become aware of the issue), or who expects to prevail in the civil courts in the event of a negligence action caused by an unfortunate outcome from inability to read the media."

Elsewhere, Dr. Clunie vehemently singled out Philips/Stentor and Amicas as being the two major PACS vendors who still produce proprietary media with proprietary viewers.<sup>9</sup>

What does this mean for the VA?

First, we ourselves should strive to provide the standard of care specified by the AMA. This is not always easy, as we are a major user of a vendor's product (Philips/Stentor iSite) that fails to produce compliant DICOM media. We need to export our DICOM objects using commercially available IHE PDI compliant products.

Second, we need to require our fee-basis imaging contractors to provide compliant DICOM media. They need to export their DICOM objects using commercially available IHE PDI compliant products.

Third, we need to deal with problematic DICOM media that are received from other locations.

### **Problems**

The Importer is designed to work with media that is formatted according to the DICOM Standard and the IHE Portable Data for Imaging integration profile. DICOM objects on compliant media usually are able to be imported. Those on non-compliant media may not be able to be imported.

<sup>&</sup>lt;sup>9</sup> http://www.dclunie.com/blog/blog/2008/04/requirements-for-office-imaging-system.html

Here is a list of some of the kinds of problems that one should expect to see.

- 1) The media is just blank.
- 2) The media is written improperly. Here is an example where the vendor did not enter the files into the directory correctly:

```
E:\>dir

Volume in drive E has no label.

Volume Serial Number is 8B34-8C3C

Directory of E:\

The parameter is incorrect. 12:00 AM 3,920 DICOMDIR
The parameter is incorrect. 12:00 AM 759 DiscView.htm
The parameter is incorrect. 12:00 AM 240 QuickViewer.txt
The parameter is incorrect. 12:00 AM 105,320 Readmelst.htm

01/15/2010 08:49 AM ODIR> Readmelst_files
The parameter is incorrect. 01/15/2010 08:49 AM 72 autorun.bat
The parameter is incorrect. 01/15/2010 08:49 AM 29 autorun.inf

01/15/2010 08:49 AM ODIR> data

6 File(s) 110,340 bytes4
2 Dir(s) 0 bytes free
```

3) There is no DICOMDIR on the media and the images are kept in separate sub-folders. The Importer cannot be used directly.

If there are DICOM files on the media, they might be able to be sent to a C-Store process on the Importer gateway. The BATCH\_SEND\_IMAGE.BAT batch file (see Appendix E) may be able to be used to extract the DICOM files from the media and send them. Otherwise, it may be necessary to use a commercial product (like Sorna<sup>10</sup>) to read the media and send the DICOM files to the C-Store process.

- 4) There are errors in DICOMDIR.
  - a. The file name of each DICOM object uses forward slashes, rather than back slashes, as required by the DICOM Standard.
  - b. The Referenced SOP Class UID (0004,1510) is wrong. (One vendor used the Media Storage Directory Storage SOP Class. Another vendor used the character string "SOP UID". It must be the same as the SOP Class UID (0008,0016) of the DICOM object.)
  - c. The Referenced Transfer Syntax UID (0004,1512) is absent even though it is a Type 1 (that is, a required data element with a non-null value).

<sup>10</sup> http://www.sorna.com/

- d. The Modality (0008,0060) is a DICOM Series attribute. It may not be defined in DICOMDIR for the Series directory record type even though it is a Type 1.
- 5) DICOMDIR is fine but the DICOM objects are not able to be supported by VistA Imaging.

The subset of DICOM SOP Classes and Transfer Syntaxes that VistA Imaging supports is specified in the SCP\_LIST.DIC dictionary. There are many compression transfer syntaxes that are not yet supported (these will be supported in Patch 99). Here is a sample error message:

- 6) Outside source places multiple patient exams on their same accession number (Skull, hand, foot all done as one exam). This is difficult to handle when this study is imported into VistA.
- 7) The media contains only JPEG2000 images and no DICOM ones.
- 8) One, or more, images in a study are incomplete (for example, only the top half of the chest x-ray). The image(s) fail(s) image processing.
- 9) The exam was not done according to radiology "best practices" standard of care. For example, a fee-basis contracted order for a "knee 3 views" only has two images.
- 10) The outside modality creates a study with a UID that is a duplicate of one which is already on VistA for another patient.
- 11) Some of the expected DICOM header information may be incorrect or blank for example technician name, body location, radiology location.
- 12) The media is a copy of an earlier one. (This one is easy the Importer automatically prevents an imported DICOM object from being imported a second time.)

Thanks to Morgan Hammac of the Fayetteville, NC VAMC for providing information about problems that were seen in the Field Test.

# Appendix E

### Using the BATCH\_SEND\_IMAGE.BAT utility script

Sometimes non-compliant portable media (like CD or DVD) contains DICOM objects but with either a malformed or an absent DICOMDIR file so that they cannot be handled directly by the Importer.

The Importer includes the BATCH\_SEND\_IMAGE.BAT utility script that may help in these situations.

```
C:\>batch_send_image
Batch DICOM File Sender

This batch file can be used to send a set of DICOM to a local C-Store process.

batch_send_image {C-Store host} {port number} {input path} [ {extension} ]

The {input path} can be either a directory on disk or one on portable media.

Only DICOM files should be in the {input path} directory tree.

This will try to send every file in the {input path} to the C-Store process.

An optional extension may be specified -- include the dot, as in ".ext".

If an extension is not specified, none will be used -- DICOM Part-10 compliant.

First Argument is the host ip address or name for the C-Store process.

Second Argument is the path to the folder containing the images, like "D:\DICOM".
```

Here is an example of its usage. The D:\DICOM folder contains 29 DICOM files. None of them have the ".DCM" extension and there is no DICOMDIR file in the root directory.

```
Volume in drive D is 112707_0006
 Volume Serial Number is DA73-57A4
Directory of D:\DICOM
11/27/2007 05:37 AM <DIR>
11/27/2007 05:37 AM <DIR>
11/20/2007 12:20 PM 897,796 NAS00000
                              152,522 NAS00001
11/20/2007 12:20 PM
11/20/2007 12:20 PM
                              310,672 NAS00002
11/20/2007 12:20 PM
                              658,242 NAS00003
                         721,364 NAS00004
287,558 NAS00005
1,034,628 NAS00006
957,628 NAS00007
11/20/2007 12:20 PM
11/20/2007 12:20 PM
11/20/2007 12:20 PM
11/20/2007 12:20 PM
                         1,014,486 NAS00008
1,041,582 NAS00009
11/20/2007 12:20 PM
11/20/2007 12:20 PM
                            967,622 NAS00010
11/20/2007 12:20 PM
11/20/2007 12:20 PM
                               91,742 NAS00011
```

D:\DICOM>dir

#### Appendix E – Using the BATCH\_SEND\_IMAGE.BAT utility script

First, start a MAG\_CSTORE.EXE process listening on port 60100 on the system. Then use the batch script to send these DICOM objects to it.

```
C:\>batch_send_image localhost 60100 d:\dicom
Batch DICOM File Sender
d:\dicom\NAS00000
Store Response
Message ID Resp:1
Data Set Type: 0101
Status: 0000 Status Information:-
      Successful operation
Class UID: 1.2.840.10008.5.1.4.1.1.3.1
Instance UID: 1.2.840.113680.1.103.52974.988052208.123798.1.10
d:\dicom\NAS00028
Store Response
Message ID Resp:1
Data Set Type: 0101
Status: 0000 Status Information:-
      Successful operation
Class UID: 1.2.840.10008.5.1.4.1.1.12.1
Instance UID: 2.16.840.1.113669.632.3.111111.42020013285100.19.5.11.100800
```