27641 QA

4Ch Small Extremity Coil Check 2020-04-29 15:19:31

```
Detected return code: 121


Start Coil Check
-----------------------
Performing Coil Check
-----------------------
Measure Protocol: 005_SmallExtremity_Tra_sn.edb.
Measured Series Loid #1:
Local 2.0.145266076
Measured Series Loid #2:
Local 2.0.145260866
Ordered Series Loid #1:
Local 2.0.145266076
Series Loid for Evaluation 1:
Local 2.0.145266076
Number of images in series: 1
Evaluate image #1
Protocol 005_SmallExtremity_Tra_sn
No Evaluation for SE1_SE2_SE3_SE4 specified in ServiceCoilData.ini.
No Evaluation for SE1_SE2_SE3_SE4@1 specified in ServiceCoilData.ini.
Ordered Series Loid #2:
Local 2.0.145260866
Series Loid for Evaluation 2:
Local 2.0.145260866
Number of images in series: 4
Evaluate image #1
Protocol 005_SmallExtremity_Tra_sn
Evaluation for SE1 specified in ServiceCoilData.ini.
Evaluation for coil element: SE1.
SN evaluation
New Evaluation mode: 0x330 specified
S/N Results
- Brightness              =  878.6 [Specification from  300.0 to 1400.0]
- S/N Ratio               =   82.8 [Specification >   50.0]
Applet Info:
Title:  Series: 2, Image 1/4 of protocol 005_SmallExtremity_Tra_sn.edb
Data type:      short
Width:  256
Height: 256


Results of coil-element SE1 are in Specification!
No Evaluation for SE1@1 specified in ServiceCoilData.ini.
Evaluate image #2
Protocol 005_SmallExtremity_Tra_sn
Evaluation for SE2 specified in ServiceCoilData.ini.
Evaluation for coil element: SE2.
SN evaluation
New Evaluation mode: 0x330 specified
S/N Results
- Brightness              =  635.1 [Specification from    0.0 to    0.0]
```

```
- S/N Ratio                    =    83.8 [Specification >   50.0]
Applet Info:
Title:  Series: 2, Image 2/4 of protocol 005_SmallExtremity_Tra_sn.edb
Data type:       short
Width:  256
Height: 256

Results of coil-element SE2 are in Specification!
No Evaluation for SE2@1 specified in ServiceCoilData.ini.
Evaluate image #3
Protocol 005_SmallExtremity_Tra_sn
Evaluation for SE3 specified in ServiceCoilData.ini.
Evaluation for coil element: SE3.
SN evaluation
New Evaluation mode: 0x330 specified
S/N Results
- Brightness                  =   477.3 [Specification from    0.0 to    0.0]
- S/N Ratio                    =    73.9 [Specification >   50.0]
Applet Info:
Title:  Series: 2, Image 3/4 of protocol 005_SmallExtremity_Tra_sn.edb
Data type:       short
Width:  256
Height: 256

Results of coil-element SE3 are in Specification!
No Evaluation for SE3@1 specified in ServiceCoilData.ini.
Evaluate image #4
Protocol 005_SmallExtremity_Tra_sn
Evaluation for SE4 specified in ServiceCoilData.ini.
Evaluation for coil element: SE4.
SN evaluation
New Evaluation mode: 0x330 specified
S/N Results
- Brightness                  =   820.0 [Specification from    0.0 to    0.0]
- S/N Ratio                    =    88.4 [Specification >   50.0]
Applet Info:
Title:  Series: 2, Image 4/4 of protocol 005_SmallExtremity_Tra_sn.edb
Data type:       short
Width:  256
Height: 256

Results of coil-element SE4 are in Specification!
No Evaluation for SE4@1 specified in ServiceCoilData.ini.
Results are in specification
Program finished succesfully
```